

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041922\
 Data File : BP009953.D
 Acq On : 19 Apr 2022 19:44
 Operator : CG/JU
 Sample : PB144172BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS172

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 04/20/2022
 Supervised By :Yogesh Patel 04/25/2022

Quant Time: Apr 20 01:33:03 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 20 01:32:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.946	152	150101	20.000	ng/u1	0.00
20) Naphthalene-d8	10.752	136	669239	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.575	164	472105	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.328	188	1042692	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.410	240	1038282	20.000	ng/u1	# 0.00
88) Perylene-d12	23.868	264	995084	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.370	96	20118m	5.982	ng/uL	0.00
4) Pyridine-d5	3.781	84	264326	28.505	ng/u1	0.00
7) Phenol-d5	7.105	99	397072	30.132	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.269	67	253160	32.269	ng/u1	0.00
11) 2-Chlorophenol-d4	7.475	132	320750	32.268	ng/u1	0.00
15) 4-Methylphenol-d8	8.652	113	344711	31.335	ng/u1	0.00
21) Nitrobenzene-d5	9.105	128	167028	34.609	ng/u1	0.00
24) 2-Nitrophenol-d4	9.828	143	188587	35.642	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.369	165	349277	31.546	ng/u1	0.00
31) 4-Chloroaniline-d4	10.881	131	427245	27.093	ng/u1	0.00
46) Dimethylphthalate-d6	13.987	166	1188620	32.052	ng/u1	0.00
49) Acenaphthylene-d8	14.269	160	1342652	30.560	ng/u1	0.00
54) 4-Nitrophenol-d4	14.769	143	207826	30.901	ng/u1	0.00
60) Fluorene-d10	15.569	176	1018448	32.646	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.687	200	212733	34.440	ng/u1	0.00
73) Anthracene-d10	17.428	188	1600321	32.056	ng/u1	0.00
81) Pyrene-d10	19.657	212	1940351	34.218	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.710	264	1708669	32.973	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.405	88	39428	11.500	ng/uL#	18
5) Pyridine	3.799	79	279771	28.917	ng/u1#	39
6) Benzaldehyde	7.075	77	252511	31.244	ng/u1	95
8) Phenol	7.134	94	424355	32.171	ng/u1#	95
10) Bis(2-Chloroethyl)ether	7.363	93	330350	32.151	ng/u1#	77
12) 2-Chlorophenol	7.505	128	327367	32.666	ng/u1	94
13) 2-Methylphenol	8.387	108	325481	31.643	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.481	45	484479	31.899	ng/u1#	86
16) Acetophenone	8.769	105	533517	30.782	ng/u1#	81
17) N-Nitroso-di-n-propyla...	8.758	70	284470	31.658	ng/u1#	72
18) 4-Methylphenol	8.716	108	353886	31.770	ng/u1	92
19) Hexachloroethane	9.034	117	134713	31.838	ng/u1	85
22) Nitrobenzene	9.146	77	412238	33.279	ng/u1#	84
23) Isophorone	9.675	82	805270	31.547	ng/u1	96
25) 2-Nitrophenol	9.857	139	198226	35.126	ng/u1#	86
26) 2,4-Dimethylphenol	9.922	107	421469	31.472	ng/u1#	84
27) Bis(2-Chloroethoxy)met...	10.157	93	474608	31.669	ng/u1#	97
29) 2,4-Dichlorophenol	10.399	162	348655	32.481	ng/u1#	82
30) Naphthalene	10.799	128	1101827	32.012	ng/u1	96
32) 4-Chloroaniline	10.904	127	430687	27.599	ng/u1	98
33) Hexachlorobutadiene	11.099	225	245492	31.755	ng/u1	94
34) Caprolactam	11.675	113	77673m	21.881	ng/u1	
35) 4-Chloro-3-methylphenol	12.028	107	414878	31.898	ng/u1#	71
36) 2-Methylnaphthalene	12.410	142	794120	31.581	ng/u1	92

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041922\
 Data File : BP009953.D
 Acq On : 19 Apr 2022 19:44
 Operator : CG/JU
 Sample : PB144172BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS172


Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 04/20/2022
 Supervised By :Yogesh Patel 04/25/2022

Quant Time: Apr 20 01:33:03 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 20 01:32:14 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.628	142	806147	31.751	ng/ul#	95
39) 1,2,4,5-Tetrachloroben...	12.775	216	466113	32.270	ng/ul#	96
40) Hexachlorocyclopentadiene	12.763	237	294327	29.364	ng/ul	96
41) 2,4,6-Trichlorophenol	13.010	196	274948	29.807	ng/ul#	81
42) 2,4,5-Trichlorophenol	13.081	196	286615	28.752	ng/ul#	87
43) 1,1'-Biphenyl	13.410	154	1122911	32.599	ng/ul	93
44) 2-Chloronaphthalene	13.457	162	864873	32.663	ng/ul	93
45) 2-Nitroaniline	13.651	65	290335	37.416	ng/ul#	82
47) Dimethylphthalate	14.034	163	1158365	32.454	ng/ul#	93
48) 2,6-Dinitrotoluene	14.145	165	244279	37.208	ng/ul	96
50) Acenaphthylene	14.298	152	1420872	32.771	ng/ul	95
51) 3-Nitroaniline	14.475	138	216361	31.700	ng/ul#	82
52) Acenaphthene	14.640	153	914240	32.205	ng/ul	96
53) 2,4-Dinitrophenol	14.675	184	106427	26.818	ng/ul#	80
55) 4-Nitrophenol	14.781	109	185794	30.010	ng/ul#	53
56) Dibenzofuran	14.975	168	1340706	32.089	ng/ul	99
57) 2,4-Dinitrotoluene	14.934	165	360587	37.404	ng/ul#	84
58) 2,3,4,6-Tetrachlorophenol	15.204	232	297967	32.783	ng/ul#	88
59) Diethylphthalate	15.398	149	1203314	32.576	ng/ul	93
61) Fluorene	15.628	166	1110925	32.281	ng/ul	91
62) 4-Chlorophenyl-phenyle...	15.622	204	590969	32.296	ng/ul	98
63) 4-Nitroaniline	15.639	138	240745	35.216	ng/ul#	80
66) 4,6-Dinitro-2-methylph...	15.698	198	208354	33.935	ng/ul#	90
67) N-Nitrosodiphenylamine	15.834	169	986140	33.039	ng/ul	94
68) 4-Bromophenyl-phenylether	16.516	248	365752	32.405	ng/ul	95
69) Hexachlorobenzene	16.639	284	419140	32.295	ng/ul#	91
70) Atrazine	16.786	200	400597	32.516	ng/ul	91
71) Pentachlorophenol	16.981	266	245809	27.981	ng/ul#	84
72) Phenanthrene	17.369	178	1825239	32.794	ng/ul	99
74) Anthracene	17.463	178	1851360	32.809	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.381	216	496205	32.916	ng/uL#	86
76) Pentachlorobenzene	14.898	250	477374	32.488	ng/uL	91
77) Carbazole	17.728	167	1648381m	32.324	ng/ul	
78) Di-n-butylphthalate	18.281	149	2107764	33.612	ng/ul#	93
80) Fluoranthene	19.333	202	2255047	34.549	ng/ul#	95
82) Pyrene	19.686	202	2306236	34.850	ng/ul#	93
83) Butylbenzylphthalate	20.557	149	968458	35.674	ng/ul#	79
84) 3,3'-Dichlorobenzidine	21.322	252	684402	30.933	ng/ul#	97
85) Benzo(a)anthracene	21.398	228	2179904	33.118	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.322	149	1424422	34.883	ng/ul#	83
87) Chrysene	21.451	228	2151093	33.872	ng/ul	99
89) Di-n-octyl phthalate	22.263	149	2417368	34.673	ng/ul	100
90) Benzo(b)fluoranthene	23.121	252	2244228	33.945	ng/ul	99
91) Benzo(k)fluoranthene	23.169	252	2049389	33.970	ng/ul#	98
93) Benzo(a)pyrene	23.763	252	2068794	33.758	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.439	276	2081450	32.480	ng/ul#	94
95) Dibenzo(a,h)anthracene	26.457	278	1765602	31.773	ng/ul#	96
96) Benzo(g,h,i)perylene	27.221	276	1733898	32.814	ng/ul#	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 11.500 ng/uL
RT: 3.405 min Scan# 54
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Instrument :
BNA_P
ClientSampleId :
SLCS172




Tgt Ion: 88 Resp: 39428
Ion Ratio Lower Upper
88 100
43 37.7 6.8 10.2
58 83.7 28.4 42.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022




#5
Pyridine
Concen: 28.917 ng/uL
RT: 3.799 min Scan# 121
Delta R.T. -0.006 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44



Tgt Ion: 79 Resp: 279771
Ion Ratio Lower Upper
79 100
52 92.8 35.5 53.3#
51 43.5 20.2 30.4#









#6
Benzaldehyde
Concen: 31.244 ng/ul
RT: 7.075 min Scan# 61
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44

Instrument :
BNA_P
ClientSampleId :
SLCS172

Tgt Ion: 77 Resp: 25251
Ion Ratio Lower Upper
77 100
105 92.5 77.0 115.4
106 89.5 77.0 115.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022



#8
Phenol
Concen: 32.171 ng/ul
RT: 7.134 min Scan# 688
Delta R.T. 0.006 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44


Tgt Ion: 94 Resp: 424355
Ion Ratio Lower Upper
94 100
65 30.7 19.0 28.6
66 43.5 34.6 51.8





#10
Bis(2-Chloroethyl)ether
Concen: 32.151 ng/ul
RT: 7.363 min Scan# 71
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Instrument :
BNA_P
ClientSampleId :
SLCS172




Tgt Ion: 93 Resp: 330350
Ion Ratio Lower Upper
93 100
63 82.7 45.0 67.4
95 32.3 26.3 39.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022




#12
2-Chlorophenol
Concen: 32.666 ng/ul
RT: 7.505 min Scan# 751
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Tgt Ion:128 Resp: 327367
Ion Ratio Lower Upper
128 100
64 52.4 45.3 67.9
130 32.4 29.0 43.6





#13
2-Methylphenol
Concen: 31.643 ng/ul
RT: 8.387 min Scan# 901
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Instrument :
BNA_P
ClientSampleId :
SLCS172



Tgt Ion: 108 Resp: 32548
Ion Ratio Lower Upper
108 100
107 91.5 70.8 106.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022



#14
2,2'-oxybis(1-Chloropropane)
Concen: 31.899 ng/ul
RT: 8.481 min Scan# 917
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44

Tgt Ion: 45 Resp: 484479
Ion Ratio Lower Upper
45 100
77 14.0 17.1 25.7#
79 11.1 13.2 19.8#









#16
Acetophenone
Concen: 30.782 ng/ul
RT: 8.769 min Scan# 965
Delta R.T. 0.006 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44

Instrument :
BNA_P
ClientSampleId :
SLCS172

Tgt Ion: 105 Resp: 53351
Ion Ratio Lower Upper
105 100
77 79.5 53.4 80.2
51 31.3 13.8 20.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022



#17
N-Nitroso-di-n-propylamine
Concen: 31.658 ng/ul
RT: 8.758 min Scan# 964
Delta R.T. 0.006 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44


Tgt Ion: 70 Resp: 284470
Ion Ratio Lower Upper
70 100
42 63.8 31.8 47.8#
101 10.3 15.6 23.4#
130 21.2 13.7 20.5#





#18
4-Methylphenol
Concen: 31.770 ng/ul
RT: 8.716 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Instrument :
BNA_P
ClientSampleId :
SLCS172




Tgt Ion:108 Resp: 35388
Ion Ratio Lower Upper
108 100
107 112.8 83.6 125.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022




#19
Hexachloroethane
Concen: 31.838 ng/ul
RT: 9.034 min Scan# 1011
Delta R.T. 0.006 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Tgt Ion:117 Resp: 134713
Ion Ratio Lower Upper
117 100
201 109.0 106.5 159.7
199 69.0 59.0 88.6





#22
Nitrobenzene
Concen: 33.279 ng/ul
RT: 9.146 min Scan# 1030
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Instrument :
BNA_P
ClientSampleId :
SLCS172




Tgt Ion: 77 Resp: 412238
Ion Ratio Lower Upper
77 100
123 41.8 42.3 63.5
65 14.3 5.8 8.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022




#23
Isophorone
Concen: 31.547 ng/ul
RT: 9.675 min Scan# 1120
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44



Tgt Ion: 82 Resp: 805270
Ion Ratio Lower Upper
82 100
95 8.1 7.8 11.8
138 17.4 12.9 19.3





#25

2-Nitrophenol

Concen: 35.126 ng/ul

RT: 9.857 min Scan# 1151

Delta R.T. -0.000 min

Lab File: BP009953.D

Acq: 19 Apr 2022 19:44

Instrument :

BNA_P

ClientSampleId :

SLCS172





Tgt Ion:139 Resp: 198220

Ion	Ratio	Lower	Upper
139	100		
65	59.7	49.4	74.2
109	36.4	44.7	67.1

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022



#26

2,4-Dimethylphenol

Concen: 31.472 ng/ul

RT: 9.922 min Scan# 1162




Delta R.T. -0.000 min


Lab File: BP009953.D

Acq: 19 Apr 2022 19:44

Tgt Ion:107 Resp: 421469


Ion	Ratio	Lower	Upper
107	100		
121	46.1	34.1	51.1
122	76.2	46.3	69.5





#27
Bis(2-Chloroethoxy)methane
Concen: 31.669 ng/ul
RT: 10.157 min Scan# 1202
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Instrument :
BNA_P
ClientSampleId :
SLCS172




Tgt Ion: 93 Resp: 474608
Ion Ratio Lower Upper
93 100
95 32.1 26.2 39.4
123 10.7 6.2 9.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022




#29
2,4-Dichlorophenol
Concen: 32.481 ng/ul
RT: 10.399 min Scan# 1243
Delta R.T. 0.006 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Tgt Ion:162 Resp: 348655
Ion Ratio Lower Upper
162 100
164 66.9 37.0 55.6#
98 37.2 31.0 46.6





#30
Naphthalene
Concen: 32.012 ng/ul
RT: 10.799 min Scan# 1311
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Instrument :
BNA_P
ClientSampleId :
SLCS172



Tgt Ion:128 Resp: 110182
Ion Ratio Lower Upper
128 100
129 11.2 8.6 13.0
127 12.9 12.2 18.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022



#32
4-Chloroaniline
Concen: 27.599 ng/ul
RT: 10.904 min Scan# 1329
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44


Tgt Ion:127 Resp: 430687
Ion Ratio Lower Upper
127 100
129 31.8 24.6 37.0





#33
Hexachlorobutadiene
Concen: 31.755 ng/ul
RT: 11.099 min Scan# 1362
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Instrument :
BNA_P
ClientSampleId :
SLCS172



Tgt Ion:225 Resp: 245492
Ion Ratio Lower Upper
225 100
223 61.5 53.4 80.0
227 63.6 53.8 80.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022



#34
Caprolactam
Concen: 21.881 ng/ul m
RT: 11.675 min Scan# 1460
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44


Tgt Ion:113 Resp: 77673
Ion Ratio Lower Upper
113 100
55 179.8 58.2 87.2#
56 136.5 68.6 103.0#





#35
4-Chloro-3-methylphenol
Concen: 31.898 ng/ul
RT: 12.028 min Scan# 1520
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Instrument :
BNA_P
ClientSampleId :
SLCS172



Tgt Ion:107 Resp: 414878
Ion Ratio Lower Upper
107 100
144 24.0 15.8 23.8
142 73.4 38.9 58.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022



#36
2-Methylnaphthalene
Concen: 31.581 ng/ul
RT: 12.410 min Scan# 1585
Delta R.T. 0.006 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44


Tgt Ion:142 Resp: 794120
Ion Ratio Lower Upper
142 100
141 87.2 64.3 96.5





#37
1-Methylnaphthalene
Concen: 31.751 ng/ul
RT: 12.628 min Scan# 1622
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Instrument :
BNA_P
ClientSampleId :
SLCS172




Tgt Ion:142 Resp: 80614
Ion Ratio Lower Upper
142 100
141 91.9 69.8 104.8
116 4.2 5.6 8.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 32.270 ng/ul
RT: 12.775 min Scan# 1647
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Tgt Ion:216 Resp: 466113
Ion Ratio Lower Upper
216 100
214 75.3 58.2 87.2
179 21.4 14.2 21.2#
108 17.0 15.3 22.9





#40
Hexachlorocyclopentadiene
Concen: 29.364 ng/ul
RT: 12.763 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Instrument :
BNA_P
ClientSampleId :
SLCS172




Tgt Ion:237 Resp: 294327
Ion Ratio Lower Upper
237 100
235 63.3 53.2 79.8
272 12.2 10.8 16.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022




#41
2,4,6-Trichlorophenol
Concen: 29.807 ng/ul
RT: 13.010 min Scan# 1687
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44



Tgt Ion:196 Resp: 274948
Ion Ratio Lower Upper
196 100
198 98.7 64.1 96.1#
200 30.9 19.9 29.9#









#42
2,4,5-Trichlorophenol
Concen: 28.752 ng/ul
RT: 13.081 min Scan# 1699
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44

Instrument :
BNA_P
ClientSampleId :
SLCS172

Tgt Ion:196 Resp: 286611
Ion Ratio Lower Upper
196 100
198 95.0 68.6 103.0
200 29.0 32.8 49.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022



#43
1,1'-Biphenyl
Concen: 32.599 ng/ul
RT: 13.410 min Scan# 1755
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44


Tgt Ion:154 Resp: 1122911
Ion Ratio Lower Upper
154 100
153 40.4 28.6 43.0
76 12.1 8.5 12.7





#44
2-Chloronaphthalene
Concen: 32.663 ng/ul
RT: 13.457 min Scan# 1763
Delta R.T. 0.006 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Instrument :
BNA_P
ClientSampleId :
SLCS172



Tgt Ion:162 Resp: 86487
Ion Ratio Lower Upper
162 100
164 33.0 27.5 41.3
127 38.9 37.1 55.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022



#45
2-Nitroaniline
Concen: 37.416 ng/ul
RT: 13.651 min Scan# 1796
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44


Tgt Ion: 65 Resp: 290335
Ion Ratio Lower Upper
65 100
92 67.2 77.9 116.9#
138 96.6 81.5 122.3





#47
Dimethylphthalate
Concen: 32.454 ng/ul
RT: 14.034 min Scan# 1861
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Instrument :
BNA_P
ClientSampleId :
SLCS172




Tgt Ion:163 Resp: 115836
Ion Ratio Lower Upper
163 100
194 4.6 4.1 6.1
164 10.2 11.3 16.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022




#48
2,6-Dinitrotoluene
Concen: 37.208 ng/ul
RT: 14.145 min Scan# 1880
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Tgt Ion:165 Resp: 244279
Ion Ratio Lower Upper
165 100
89 60.4 51.7 77.5
121 22.0 18.3 27.5





#50
Acenaphthylene
Concen: 32.771 ng/ul
RT: 14.298 min Scan# 1906
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Instrument :
BNA_P
ClientSampleId :
SLCS172




Tgt Ion:152 Resp: 1420872
Ion Ratio Lower Upper
152 100
151 19.6 18.3 27.5
153 13.1 11.0 16.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022




#51
3-Nitroaniline
Concen: 31.700 ng/ul
RT: 14.475 min Scan# 1936
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Tgt Ion:138 Resp: 216361
Ion Ratio Lower Upper
138 100
108 12.0 15.8 23.8#
92 123.7 117.1 175.7





#52
Acenaphthene
Concen: 32.205 ng/ul
RT: 14.640 min Scan# 1964
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Instrument :
BNA_P
ClientSampleId :
SLCS172




Tgt Ion:153 Resp: 914240
Ion Ratio Lower Upper
153 100
152 46.9 34.3 51.5
154 89.5 70.0 105.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022




#53
2,4-Dinitrophenol
Concen: 26.818 ng/ul
RT: 14.675 min Scan# 1970
Delta R.T. -0.006 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Tgt Ion:184 Resp: 106427
Ion Ratio Lower Upper
184 100
63 75.3 34.6 52.0#
154 70.1 57.7 86.5





#55
4-Nitrophenol
Concen: 30.010 ng/ul
RT: 14.781 min Scan# 1988
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Instrument :
BNA_P
ClientSampleId :
SLCS172



Tgt Ion:109 Resp: 185794
Ion Ratio Lower Upper
109 100
139 111.3 61.4 92.2
65 130.2 65.3 97.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022



#56
Dibenzofuran
Concen: 32.089 ng/ul
RT: 14.975 min Scan# 2021
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44

Tgt Ion:168 Resp: 1340706
Ion Ratio Lower Upper
168 100
139 38.3 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 37.404 ng/ul
RT: 14.934 min Scan# 2014
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44

Instrument :
BNA_P
ClientSampleId :
SLCS172

Tgt Ion:165 Resp: 36058
Ion Ratio Lower Upper
165 100
63 47.3 29.5 44.3
182 3.2 2.8 4.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022



#58
2,3,4,6-Tetrachlorophenol
Concen: 32.783 ng/ul
RT: 15.204 min Scan# 2060
Delta R.T. 0.006 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44


Tgt Ion:232 Resp: 297967
Ion Ratio Lower Upper
232 100
131 40.6 25.4 38.2#
130 2.3 4.2 6.2#
166 29.9 28.1 42.1





#59
Diethylphthalate
Concen: 32.576 ng/ul
RT: 15.398 min Scan# 2093
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Instrument :
BNA_P
ClientSampleId :
SLCS172



Tgt Ion:149 Resp: 1203314
Ion Ratio Lower Upper
149 100
177 22.4 22.0 33.0
150 12.7 10.2 15.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022



#61
Fluorene
Concen: 32.281 ng/ul
RT: 15.628 min Scan# 2132
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44


Tgt Ion:166 Resp: 1110925
Ion Ratio Lower Upper
166 100
165 96.8 70.3 105.5
167 13.2 9.5 14.3





#62
4-Chlorophenyl-phenylether
Concen: 32.296 ng/ul
RT: 15.622 min Scan# 2131
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Instrument :
BNA_P
ClientSampleId :
SLCS172




Tgt Ion: 204 Resp: 590969
Ion Ratio Lower Upper
204 100
206 32.5 27.6 41.4
141 58.6 46.6 69.8

Manual Integrations
APPROVED



Reviewed By : Jagrut Upadhyay 04/20/2022
Supervised By : Yogesh Patel 04/25/2022




#63
4-Nitroaniline
Concen: 35.216 ng/ul
RT: 15.639 min Scan# 2134
Delta R.T. 0.006 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Tgt Ion: 138 Resp: 240745
Ion Ratio Lower Upper
138 100
92 53.8 41.6 62.4
108 91.1 97.6 146.4#





#66
4,6-Dinitro-2-methylphenol
Concen: 33.935 ng/ul
RT: 15.698 min Scan# 2144
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Instrument :
BNA_P
ClientSampleId :
SLCS172



Tgt Ion:198 Resp: 208354
Ion Ratio Lower Upper
198 100
182 4.0 6.0 9.0
77 26.9 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022



#67
N-Nitrosodiphenylamine
Concen: 33.039 ng/ul
RT: 15.834 min Scan# 2167
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44


Tgt Ion:169 Resp: 986140
Ion Ratio Lower Upper
169 100
168 65.7 48.3 72.5
167 34.5 29.6 44.4





#68
4-Bromophenyl-phenylether
Concen: 32.405 ng/ul
RT: 16.516 min Scan# 2283
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Instrument :
BNA_P
ClientSampleId :
SLCS172




Tgt Ion:248 Resp: 365752
Ion Ratio Lower Upper
248 100
250 98.4 79.4 119.0
141 75.7 67.2 100.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022




#69
Hexachlorobenzene
Concen: 32.295 ng/ul
RT: 16.639 min Scan# 2304
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44



Tgt Ion:284 Resp: 419140
Ion Ratio Lower Upper
284 100
142 31.5 19.2 28.8#
249 31.4 27.4 41.2









#70
Atrazine
Concen: 32.516 ng/ul
RT: 16.786 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44

Instrument :
BNA_P
ClientSampleId :
SLCS172

Tgt Ion:200 Resp: 40059
Ion Ratio Lower Upper
200 100
173 24.8 19.2 28.8
215 48.3 32.2 48.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022



#71
Pentachlorophenol
Concen: 27.981 ng/ul
RT: 16.981 min Scan# 2362
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44


Tgt Ion:266 Resp: 245809
Ion Ratio Lower Upper
266 100
264 63.0 57.3 85.9
268 63.6 65.7 98.5#





#72
Phenanthrene
Concen: 32.794 ng/u1
RT: 17.369 min Scan# 2429
Delta R.T. -0.006 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44



Instrument :
BNA_P
ClientSampleId :
SLCS172



Tgt Ion:178 Resp: 1825239
Ion Ratio Lower Upper
178 100
179 14.9 11.7 17.5
176 18.6 14.5 21.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022



#74
Anthracene
Concen: 32.809 ng/u1
RT: 17.463 min Scan# 2444
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44

Tgt Ion:178 Resp: 1851360
Ion Ratio Lower Upper
178 100
179 15.4 12.3 18.5
176 18.1 13.4 20.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 32.916 ng/uL
RT: 13.381 min Scan# 1750
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44

Instrument :
BNA_P
ClientSampleId :
SLCS172

Tgt Ion:216 Resp: 496209
Ion Ratio Lower Upper
216 100
214 78.1 50.4 75.6
218 47.4 35.2 52.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022



#76
Pentachlorobenzene
Concen: 32.488 ng/uL
RT: 14.898 min Scan# 2008
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44


Tgt Ion:250 Resp: 477374
Ion Ratio Lower Upper
250 100
248 62.7 43.5 65.3
252 63.6 46.5 69.7





#77
 Carbazole
 Concen: 32.324 ng/ul m
 RT: 17.728 min Scan# 2489
 Delta R.T. -0.000 min
 Lab File: BP009953.D
 Acq: 19 Apr 2022 19:44




Instrument :
 BNA_P
 ClientSampleId :
 SLCS172



Tgt Ion:167 Resp: 164838
 Ion Ratio Lower Upper
 167 100
 166 21.4 15.8 23.8
 139 13.0 8.3 12.5




Manual Integrations
 APPROVED


Reviewed By :Jagrut Upadhyay 04/20/2022
 Supervised By :Yogesh Patel 04/25/2022



#78
 Di-n-butylphthalate
 Concen: 33.612 ng/ul
 RT: 18.281 min Scan# 2583
 Delta R.T. -0.000 min
 Lab File: BP009953.D
 Acq: 19 Apr 2022 19:44


Tgt Ion:149 Resp: 2107764
 Ion Ratio Lower Upper
 149 100
 150 9.2 9.6 14.4#
 104 5.4 6.2 9.4#





#80
Fluoranthene
Concen: 34.549 ng/u1
RT: 19.333 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Instrument :
BNA_P
ClientSampleId :
SLCS172



Tgt Ion:202 Resp: 225504
Ion Ratio Lower Upper
202 100
101 9.2 9.6 14.4
100 7.2 6.3 9.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022



#82
Pyrene
Concen: 34.850 ng/u1
RT: 19.686 min Scan# 2822
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44


Tgt Ion:202 Resp: 2306236
Ion Ratio Lower Upper
202 100
101 10.9 10.6 15.8
100 9.2 9.8 14.6





#83
Butylbenzylphthalate
Concen: 35.674 ng/ul
RT: 20.557 min Scan# 2970
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Instrument :
BNA_P
ClientSampleId :
SLCS172



Tgt Ion:149 Resp: 968458
Ion Ratio Lower Upper
149 100
91 68.7 71.1 106.7
206 23.6 28.8 43.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022



#84
3,3'-Dichlorobenzidine
Concen: 30.933 ng/ul
RT: 21.322 min Scan# 3100
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44


Tgt Ion:252 Resp: 684402
Ion Ratio Lower Upper
252 100
254 64.3 49.8 74.6
126 9.4 4.3 6.5





#85
Benzo(a)anthracene
Concen: 33.118 ng/ul
RT: 21.398 min Scan# 3113
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Instrument :
BNA_P
ClientSampleId :
SLCS172




Tgt Ion:228 Resp: 2179904
Ion Ratio Lower Upper
228 100
229 19.7 18.2 27.2
226 26.2 22.8 34.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022




#86
Bis(2-ethylhexyl)phthalate
Concen: 34.883 ng/ul
RT: 21.322 min Scan# 3100
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Tgt Ion:149 Resp: 1424422
Ion Ratio Lower Upper
149 100
167 29.0 32.8 49.2#
279 5.5 5.6 8.4#





#87
Chrysene
Concen: 33.872 ng/ul
RT: 21.451 min Scan# 3122
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Instrument :
BNA_P
ClientSampleId :
SLCS172



Tgt Ion:228 Resp: 2151093
Ion Ratio Lower Upper
228 100
226 28.6 22.7 34.1
229 19.7 15.2 22.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022



#89
Di-n-octyl phthalate
Concen: 34.673 ng/ul
RT: 22.263 min Scan# 3260
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44


Tgt Ion:149 Resp: 2417368





#90
Benzo(b)fluoranthene
Concen: 33.945 ng/ul
RT: 23.121 min Scan# 3406
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Instrument :
BNA_P
ClientSampleId :
SLCS172



Tgt Ion:252 Resp: 2244228
Ion Ratio Lower Upper
252 100
253 21.9 17.4 26.0
125 8.0 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022



#91
Benzo(k)fluoranthene
Concen: 33.970 ng/ul
RT: 23.169 min Scan# 3414
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44


Tgt Ion:252 Resp: 2049389
Ion Ratio Lower Upper
252 100
253 21.9 17.9 26.9
125 8.2 5.1 7.7#





#93
Benzo(a)pyrene
Concen: 33.758 ng/ul
RT: 23.763 min Scan# 3515
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44




Instrument :
BNA_P
ClientSampleId :
SLCS172



Tgt Ion:252 Resp: 2068794
Ion Ratio Lower Upper
252 100
253 21.3 17.5 26.3
125 8.8 6.4 9.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022



#94
Indeno(1,2,3-cd)pyrene
Concen: 32.480 ng/ul
RT: 26.439 min Scan# 3970
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44

Tgt Ion:276 Resp: 2081450
Ion Ratio Lower Upper
276 100
138 19.3 12.3 18.5#
277 25.9 19.3 28.9





#95
Dibenzo(a,h)anthracene
Concen: 31.773 ng/ul
RT: 26.457 min Scan# 3972
Delta R.T. 0.006 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44

Instrument :

BNA_P

ClientSampleId :





SLCS172

Tgt Ion:278 Resp: 1765602
Ion Ratio Lower Upper
278 100
139 12.3 6.6 10.0
279 23.7 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 04/20/2022
Supervised By :Yogesh Patel 04/25/2022



#96
Benzo(g,h,i)perylene
Concen: 32.814 ng/ul
RT: 27.221 min Scan# 4103
Delta R.T. -0.000 min
Lab File: BP009953.D
Acq: 19 Apr 2022 19:44

Tgt Ion:276 Resp: 1733898
Ion Ratio Lower Upper
276 100
138 17.4 11.6 17.4
277 23.5 15.9 23.9

