

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041922\
 Data File : BP010004.D
 Acq On : 21 Apr 2022 07:18
 Operator : CG/JU
 Sample : PB144233BS
 Misc :
 ALS Vial : 58 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS233

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 04/21/2022
 Supervised By :Yogesh Patel 04/25/2022

Quant Time: Apr 21 08:05:40 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 20 01:32:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.940	152	106331	20.000	ng/u1	0.00
20) Naphthalene-d8	10.746	136	480984	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.575	164	342463	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.322	188	772745	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.404	240	851441	20.000	ng/u1	# 0.00
88) Perylene-d12	23.857	264	835972	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.369	96	17412m	7.308	ng/uL	0.00
4) Pyridine-d5	3.781	84	225192	34.281	ng/u1	0.00
7) Phenol-d5	7.105	99	338397	36.250	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.269	67	213175	38.358	ng/u1	0.00
11) 2-Chlorophenol-d4	7.475	132	268614	38.146	ng/u1	0.00
15) 4-Methylphenol-d8	8.652	113	295089	37.866	ng/u1	0.00
21) Nitrobenzene-d5	9.099	128	142546	41.097	ng/u1	0.00
24) 2-Nitrophenol-d4	9.828	143	158829	41.767	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.369	165	297061	37.331	ng/u1	0.00
31) 4-Chloroaniline-d4	10.881	131	386274	34.082	ng/u1	0.00
46) Dimethylphthalate-d6	13.981	166	1018930	37.877	ng/u1	0.00
49) Acenaphthylene-d8	14.263	160	1149356	36.064	ng/u1	0.00
54) 4-Nitrophenol-d4	14.763	143	190058	38.957	ng/u1	0.00
60) Fluorene-d10	15.563	176	867017	38.313	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.681	200	176841	38.630	ng/u1	0.00
73) Anthracene-d10	17.422	188	1428360	38.606	ng/u1	0.00
81) Pyrene-d10	19.651	212	1791938	38.535	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.698	264	1715838	39.413	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.405	88	31150	12.825	ng/uL#	15
5) Pyridine	3.805	79	233131	34.015	ng/u1#	44
6) Benzaldehyde	7.075	77	213233	37.245	ng/u1	93
8) Phenol	7.128	94	353559	37.837	ng/u1#	91
10) Bis(2-Chloroethyl)ether	7.363	93	274310	37.687	ng/u1#	79
12) 2-Chlorophenol	7.505	128	269876	38.015	ng/u1	93
13) 2-Methylphenol	8.381	108	270525	37.127	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.469	45	403832	37.534	ng/u1#	84
16) Acetophenone	8.763	105	450666	36.705	ng/u1#	81
17) N-Nitroso-di-n-propyla...	8.752	70	242910	38.160	ng/u1#	74
18) 4-Methylphenol	8.710	108	300854	38.127	ng/u1	91
19) Hexachloroethane	9.028	117	110423	36.840	ng/u1#	80
22) Nitrobenzene	9.140	77	346234	38.890	ng/u1#	84
23) Isophorone	9.669	82	688111	37.508	ng/u1#	97
25) 2-Nitrophenol	9.857	139	163670	40.355	ng/u1#	84
26) 2,4-Dimethylphenol	9.916	107	346495	36.000	ng/u1#	80
27) Bis(2-Chloroethoxy)met...	10.151	93	395371	36.708	ng/u1#	97
29) 2,4-Dichlorophenol	10.393	162	291787	37.823	ng/u1#	84
30) Naphthalene	10.798	128	914482	36.968	ng/u1	96
32) 4-Chloroaniline	10.904	127	381572	34.022	ng/u1	97
33) Hexachlorobutadiene	11.093	225	202901	36.518	ng/u1	92
34) Caprolactam	11.669	113	110012m	43.121	ng/u1	
35) 4-Chloro-3-methylphenol	12.028	107	361429	38.665	ng/u1#	68
36) 2-Methylnaphthalene	12.404	142	665908	36.847	ng/u1	91

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041922\
 Data File : BP010004.D
 Acq On : 21 Apr 2022 07:18
 Operator : CG/JU
 Sample : PB144233BS
 Misc :
 ALS Vial : 58 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS233


Quant Time: Apr 21 08:05:40 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 20 01:32:14 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 04/21/2022
 Supervised By :Yogesh Patel 04/25/2022


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.622	142	677391	37.123	ng/ul#	94
39) 1,2,4,5-Tetrachloroben...	12.775	216	389000	37.127	ng/ul#	91
40) Hexachlorocyclopentadiene	12.757	237	184347	25.354	ng/ul	95
41) 2,4,6-Trichlorophenol	13.010	196	261764	39.120	ng/ul#	83
42) 2,4,5-Trichlorophenol	13.081	196	291397	40.298	ng/ul#	88
43) 1,1'-Biphenyl	13.410	154	943121	37.744	ng/ul	94
44) 2-Chloronaphthalene	13.451	162	722424	37.612	ng/ul	94
45) 2-Nitroaniline	13.645	65	253333	45.006	ng/ul#	79
47) Dimethylphthalate	14.028	163	985424	38.060	ng/ul#	92
48) 2,6-Dinitrotoluene	14.145	165	209717	44.036	ng/ul	92
50) Acenaphthylene	14.292	152	1207097	38.380	ng/ul	95
51) 3-Nitroaniline	14.469	138	201971	40.794	ng/ul#	85
52) Acenaphthene	14.639	153	777753	37.769	ng/ul	95
53) 2,4-Dinitrophenol	14.675	184	101801	35.364	ng/ul#	79
55) 4-Nitrophenol	14.781	109	167585	37.316	ng/ul#	45
56) Dibenzofuran	14.969	168	1148413	37.892	ng/ul	97
57) 2,4-Dinitrotoluene	14.928	165	311494	44.543	ng/ul#	80
58) 2,3,4,6-Tetrachlorophenol	15.198	232	259462	39.353	ng/ul#	88
59) Diethylphthalate	15.392	149	1024418	38.231	ng/ul	94
61) Fluorene	15.622	166	957150	38.342	ng/ul	93
62) 4-Chlorophenyl-phenyle...	15.616	204	495834	37.355	ng/ul	98
63) 4-Nitroaniline	15.633	138	232671	46.919	ng/ul#	80
66) 4,6-Dinitro-2-methylph...	15.698	198	175172	38.497	ng/ul#	93
67) N-Nitrosodiphenylamine	15.828	169	842632	38.093	ng/ul	94
68) 4-Bromophenyl-phenylether	16.510	248	311385	37.226	ng/ul	95
69) Hexachlorobenzene	16.633	284	360795	37.511	ng/ul#	90
70) Atrazine	16.780	200	346906	37.994	ng/ul	92
71) Pentachlorophenol	16.975	266	233058	35.798	ng/ul#	84
72) Phenanthrene	17.369	178	1613555	39.119	ng/ul	99
74) Anthracene	17.457	178	1630197	38.982	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.375	216	418922	37.497	ng/uL#	88
76) Pentachlorobenzene	14.898	250	404310	37.127	ng/uL	91
77) Carbazole	17.722	167	1463838	38.733	ng/ul#	96
78) Di-n-butylphthalate	18.275	149	1849095	39.788	ng/ul#	93
80) Fluoranthene	19.327	202	2044797	38.203	ng/ul	96
82) Pyrene	19.680	202	2118799	39.043	ng/ul#	93
83) Butylbenzylphthalate	20.545	149	900275	40.439	ng/ul#	81
84) 3,3'-Dichlorobenzidine	21.316	252	666631	36.741	ng/ul#	96
85) Benzo(a)anthracene	21.386	228	2120340	39.282	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.310	149	1346661	40.215	ng/ul#	82
87) Chrysene	21.445	228	2084762	40.031	ng/ul	99
89) Di-n-octyl phthalate	22.251	149	2335028	39.867	ng/ul	100
90) Benzo(b)fluoranthene	23.110	252	2224489	40.051	ng/ul	99
91) Benzo(k)fluoranthene	23.157	252	2019467	39.846	ng/ul#	98
93) Benzo(a)pyrene	23.751	252	2053279	39.882	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.415	276	2111872	39.227	ng/ul#	95
95) Dibenzo(a,h)anthracene	26.439	278	1816794	38.917	ng/ul#	96
96) Benzo(g,h,i)perylene	27.203	276	1767464	39.816	ng/ul#	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 12.825 ng/uL
RT: 3.405 min Scan# 54
Delta R.T. -0.000 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion: 88 Resp: 31150
Ion Ratio Lower Upper
88 100
43 38.4 6.8 10.2
58 85.6 28.4 42.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022




#5
Pyridine
Concen: 34.015 ng/uL
RT: 3.805 min Scan# 122
Delta R.T. -0.000 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18



Tgt Ion: 79 Resp: 233131
Ion Ratio Lower Upper
79 100
52 88.8 35.5 53.3#
51 43.1 20.2 30.4#









#6
Benzaldehyde
Concen: 37.245 ng/ul
RT: 7.075 min Scan# 61
Delta R.T. -0.000 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18

Instrument :
BNA_P
ClientSampleId :
SLCS233

Tgt Ion: 77 Resp: 21323
Ion Ratio Lower Upper
77 100
105 93.9 77.0 115.4
106 85.4 77.0 115.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022



#8
Phenol
Concen: 37.837 ng/ul
RT: 7.128 min Scan# 687
Delta R.T. -0.000 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18


Tgt Ion: 94 Resp: 353559
Ion Ratio Lower Upper
94 100
65 34.1 19.0 28.6#
66 44.7 34.6 51.8





#10
Bis(2-Chloroethyl)ether
Concen: 37.687 ng/ul
RT: 7.363 min Scan# 727
Delta R.T. -0.000 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion: 93 Resp: 274310
Ion Ratio Lower Upper
93 100
63 79.6 45.0 67.4
95 31.2 26.3 39.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022




#12
2-Chlorophenol
Concen: 38.015 ng/ul
RT: 7.505 min Scan# 751
Delta R.T. -0.000 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Tgt Ion:128 Resp: 269876
Ion Ratio Lower Upper
128 100
64 51.9 45.3 67.9
130 31.0 29.0 43.6





#13
2-Methylphenol
Concen: 37.127 ng/ul
RT: 8.381 min Scan# 901
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Instrument :
BNA_P
ClientSampleId :
SLCS233



Tgt Ion:108 Resp: 270529
Ion Ratio Lower Upper
108 100
107 89.5 70.8 106.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022



#14
2,2'-oxybis(1-Chloropropane)
Concen: 37.534 ng/ul
RT: 8.469 min Scan# 915
Delta R.T. -0.012 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18


Tgt Ion: 45 Resp: 403832
Ion Ratio Lower Upper
45 100
77 13.6 17.1 25.7#
79 9.6 13.2 19.8#





#16
Acetophenone
Concen: 36.705 ng/ul
RT: 8.763 min Scan# 965
Delta R.T. -0.000 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion: 105 Resp: 450660
Ion Ratio Lower Upper
105 100
77 79.0 53.4 80.2
51 31.0 13.8 20.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022




#17
N-Nitroso-di-n-propylamine
Concen: 38.160 ng/ul
RT: 8.752 min Scan# 963
Delta R.T. -0.000 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Tgt Ion: 70 Resp: 242910
Ion Ratio Lower Upper
70 100
42 61.4 31.8 47.8#
101 9.5 15.6 23.4#
130 20.4 13.7 20.5





#18
4-Methylphenol
Concen: 38.127 ng/ul
RT: 8.710 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion:108 Resp: 300854
Ion Ratio Lower Upper
108 100
107 113.3 83.6 125.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022




#19
Hexachloroethane
Concen: 36.840 ng/ul
RT: 9.028 min Scan# 1010
Delta R.T. -0.000 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Tgt Ion:117 Resp: 110423
Ion Ratio Lower Upper
117 100
201 103.6 106.5 159.7#
199 65.5 59.0 88.6





#22
Nitrobenzene
Concen: 38.890 ng/ul
RT: 9.140 min Scan# 1030
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Instrument :
BNA_P
ClientSampleId :
SLCS233



Tgt Ion: 77 Resp: 346234
Ion Ratio Lower Upper
77 100
123 42.2 42.3 63.5
65 15.2 5.8 8.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022



#23
Isophorone
Concen: 37.508 ng/ul
RT: 9.669 min Scan# 1119
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18

Tgt Ion: 82 Resp: 688111
Ion Ratio Lower Upper
82 100
95 7.5 7.8 11.8
138 16.0 12.9 19.3





#25

2-Nitrophenol

Concen: 40.355 ng/ul

RT: 9.857 min Scan# 1151

Delta R.T. -0.000 min

Lab File: BP010004.D

Acq: 21 Apr 2022 07:18

Instrument :

BNA_P

ClientSampleId :





SLCS233

Tgt Ion:139	Resp: 163670
Ion Ratio	Lower Upper
139 100	
65 59.0	49.4 74.2
109 34.5	44.7 67.1

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022



#26

2,4-Dimethylphenol

Concen: 36.000 ng/ul




RT: 9.916 min Scan# 1161


Delta R.T. -0.006 min

Lab File: BP010004.D

Acq: 21 Apr 2022 07:18

Tgt Ion:107	Resp: 346495
Ion Ratio	Lower Upper
107 100	
121 47.6	34.1 51.1
122 79.9	46.3 69.5









#27
Bis(2-Chloroethoxy)methane
Concen: 36.708 ng/ul
RT: 10.151 min Scan# 1201
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18

Instrument :
BNA_P
ClientSampleId :
SLCS233

Tgt Ion: 93 Resp: 39537
Ion Ratio Lower Upper
93 100
95 33.3 26.2 39.4
123 11.4 6.2 9.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022



#29
2,4-Dichlorophenol
Concen: 37.823 ng/ul
RT: 10.393 min Scan# 1242
Delta R.T. -0.000 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18

Tgt Ion:162 Resp: 291787
Ion Ratio Lower Upper
162 100
164 64.7 37.0 55.6#
98 37.8 31.0 46.6









#30
Naphthalene
Concen: 36.968 ng/ul
RT: 10.798 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18

Instrument :
BNA_P
ClientSampleId :
SLCS233

Tgt Ion:128 Resp: 91448
Ion Ratio Lower Upper
128 100
129 11.3 8.6 13.0
127 13.0 12.2 18.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022



#32
4-Chloroaniline
Concen: 34.022 ng/ul
RT: 10.904 min Scan# 1329
Delta R.T. -0.000 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18


Tgt Ion:127 Resp: 381572
Ion Ratio Lower Upper
127 100
129 32.4 24.6 37.0





#33
Hexachlorobutadiene
Concen: 36.518 ng/ul
RT: 11.093 min Scan# 1361
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion:225 Resp: 20290
Ion Ratio Lower Upper
225 100
223 60.7 53.4 80.0
227 61.2 53.8 80.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022




#34
Caprolactam
Concen: 43.121 ng/ul m
RT: 11.669 min Scan# 1459
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Tgt Ion:113 Resp: 110012
Ion Ratio Lower Upper
113 100
55 180.9 58.2 87.2#
56 145.7 68.6 103.0#





#35
4-Chloro-3-methylphenol
Concen: 38.665 ng/ul
RT: 12.028 min Scan# 1520
Delta R.T. -0.000 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion:107 Resp: 361429
Ion Ratio Lower Upper
107 100
144 23.7 15.8 23.8
142 76.7 38.9 58.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022




#36
2-Methylnaphthalene
Concen: 36.847 ng/ul
RT: 12.404 min Scan# 1584
Delta R.T. -0.000 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Tgt Ion:142 Resp: 665908
Ion Ratio Lower Upper
142 100
141 88.7 64.3 96.5





#37
1-Methylnaphthalene
Concen: 37.123 ng/ul
RT: 12.622 min Scan# 1621
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion:142 Resp: 67739
Ion Ratio Lower Upper
142 100
141 92.4 69.8 104.8
116 4.1 5.6 8.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 37.127 ng/ul
RT: 12.775 min Scan# 1647
Delta R.T. -0.000 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Tgt Ion:216 Resp: 389000
Ion Ratio Lower Upper
216 100
214 81.4 58.2 87.2
179 22.1 14.2 21.2#
108 16.7 15.3 22.9





#40
Hexachlorocyclopentadiene
Concen: 25.354 ng/ul
RT: 12.757 min Scan# 1644
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion:237 Resp: 18434
Ion Ratio Lower Upper
237 100
235 61.9 53.2 79.8
272 12.8 10.8 16.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022




#41
2,4,6-Trichlorophenol
Concen: 39.120 ng/ul
RT: 13.010 min Scan# 1687
Delta R.T. -0.000 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Tgt Ion:196 Resp: 261764
Ion Ratio Lower Upper
196 100
198 97.0 64.1 96.1#
200 30.0 19.9 29.9#





#42
2,4,5-Trichlorophenol
Concen: 40.298 ng/ul
RT: 13.081 min Scan# 1699
Delta R.T. -0.000 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Instrument :
BNA_P
ClientSampleId :
SLCS233



Tgt Ion:196 Resp: 291397
Ion Ratio Lower Upper
196 100
198 95.5 68.6 103.0
200 31.8 32.8 49.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022



#43
1,1'-Biphenyl
Concen: 37.744 ng/ul
RT: 13.410 min Scan# 1755
Delta R.T. -0.000 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18


Tgt Ion:154 Resp: 943121
Ion Ratio Lower Upper
154 100
153 40.2 28.6 43.0
76 11.4 8.5 12.7





#44
2-Chloronaphthalene
Concen: 37.612 ng/ul
RT: 13.451 min Scan# 1762
Delta R.T. -0.000 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion:162 Resp: 722424
Ion Ratio Lower Upper
162 100
164 33.1 27.5 41.3
127 40.4 37.1 55.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022




#45
2-Nitroaniline
Concen: 45.006 ng/ul
RT: 13.645 min Scan# 1795
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Tgt Ion: 65 Resp: 253333
Ion Ratio Lower Upper
65 100
92 64.8 77.9 116.9#
138 93.2 81.5 122.3





#47
Dimethylphthalate
Concen: 38.060 ng/ul
RT: 14.028 min Scan# 1860
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion:163 Resp: 985424
Ion Ratio Lower Upper
163 100
194 4.2 4.1 6.1
164 10.1 11.3 16.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022




#48
2,6-Dinitrotoluene
Concen: 44.036 ng/ul
RT: 14.145 min Scan# 1880
Delta R.T. -0.000 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Tgt Ion:165 Resp: 209717
Ion Ratio Lower Upper
165 100
89 57.0 51.7 77.5
121 20.9 18.3 27.5





#50
Acenaphthylene
Concen: 38.380 ng/ul
RT: 14.292 min Scan# 1906
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Instrument :
BNA_P
ClientSampleId :
SLCS233



Tgt Ion:152 Resp: 120709
Ion Ratio Lower Upper
152 100
151 19.6 18.3 27.5
153 13.2 11.0 16.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022



#51
3-Nitroaniline
Concen: 40.794 ng/ul
RT: 14.469 min Scan# 1935
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18


Tgt Ion:138 Resp: 201971
Ion Ratio Lower Upper
138 100
108 12.4 15.8 23.8#
92 127.2 117.1 175.7





#52
Acenaphthene
Concen: 37.769 ng/ul
RT: 14.639 min Scan# 1964
Delta R.T. -0.000 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion:153 Resp: 77775
Ion Ratio Lower Upper
153 100
152 46.1 34.3 51.5
154 91.5 70.0 105.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022




#53
2,4-Dinitrophenol
Concen: 35.364 ng/ul
RT: 14.675 min Scan# 1970
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Tgt Ion:184 Resp: 101801
Ion Ratio Lower Upper
184 100
63 77.8 34.6 52.0#
154 72.9 57.7 86.5





#55
4-Nitrophenol
Concen: 37.316 ng/ul
RT: 14.781 min Scan# 1988
Delta R.T. -0.000 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Instrument :
BNA_P
ClientSampleId :
SLCS233



Tgt Ion:109 Resp: 167589
Ion Ratio Lower Upper
109 100
139 119.0 61.4 92.2
65 136.0 65.3 97.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022



#56
Dibenzofuran
Concen: 37.892 ng/ul
RT: 14.969 min Scan# 2020
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18

Tgt Ion:168 Resp: 1148413
Ion Ratio Lower Upper
168 100
139 39.1 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 44.543 ng/ul
RT: 14.928 min Scan# 2013
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18

Instrument :
BNA_P
ClientSampleId :
SLCS233

Tgt Ion:165 Resp: 311494
Ion Ratio Lower Upper
165 100
63 50.0 29.5 44.3#
182 3.2 2.8 4.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022



#58
2,3,4,6-Tetrachlorophenol
Concen: 39.353 ng/ul
RT: 15.198 min Scan# 2059
Delta R.T. -0.000 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18

Tgt Ion:232 Resp: 259462
Ion Ratio Lower Upper
232 100
131 42.1 25.4 38.2#
130 2.2 4.2 6.2#
166 31.0 28.1 42.1





#59
Diethylphthalate
Concen: 38.231 ng/ul
RT: 15.392 min Scan# 2092
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18

Instrument :

BNA_P

ClientSampleId :





SLCS233

Tgt Ion:149 Resp: 1024418
Ion Ratio Lower Upper
149 100
177 22.7 22.0 33.0
150 12.7 10.2 15.2

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022



#61
Fluorene
Concen: 38.342 ng/ul
RT: 15.622 min Scan# 2131
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18


Tgt Ion:166 Resp: 957150
Ion Ratio Lower Upper
166 100
165 95.1 70.3 105.5
167 13.4 9.5 14.3





#62
4-Chlorophenyl-phenylether
Concen: 37.355 ng/ul
RT: 15.616 min Scan# 2130
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Instrument : BNA_P
ClientSampleId : SLCS233




Tgt Ion: 204 Resp: 495834
Ion Ratio Lower Upper
204 100
206 32.6 27.6 41.4
141 59.6 46.6 69.8

Manual Integrations
APPROVED



Reviewed By : Jagrut Upadhyay 04/21/2022
Supervised By : Yogesh Patel 04/25/2022




#63
4-Nitroaniline
Concen: 46.919 ng/ul
RT: 15.633 min Scan# 2133
Delta R.T. -0.000 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Tgt Ion: 138 Resp: 232671
Ion Ratio Lower Upper
138 100
92 53.4 41.6 62.4
108 91.4 97.6 146.4





#66
4,6-Dinitro-2-methylphenol
Concen: 38.497 ng/ul
RT: 15.698 min Scan# 2144
Delta R.T. -0.000 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion:198 Resp: 175172
Ion Ratio Lower Upper
198 100
182 3.6 6.0 9.0
77 24.9 17.7 26.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022




#67
N-Nitrosodiphenylamine
Concen: 38.093 ng/ul
RT: 15.828 min Scan# 2166
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Tgt Ion:169 Resp: 842632
Ion Ratio Lower Upper
169 100
168 66.0 48.3 72.5
167 35.2 29.6 44.4





#68
4-Bromophenyl-phenylether
Concen: 37.226 ng/ul
RT: 16.510 min Scan# 211
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Instrument :
BNA_P
ClientSampleId :
SLCS233



Tgt Ion:248 Resp: 311385
Ion Ratio Lower Upper
248 100
250 100.8 79.4 119.0
141 76.1 67.2 100.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022



#69
Hexachlorobenzene
Concen: 37.511 ng/ul
RT: 16.633 min Scan# 2303
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18

Tgt Ion:284 Resp: 360795
Ion Ratio Lower Upper
284 100
142 34.2 19.2 28.8#
249 33.1 27.4 41.2









#70
Atrazine
Concen: 37.994 ng/ul
RT: 16.780 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18

Instrument :
BNA_P
ClientSampleId :
SLCS233

Tgt Ion:200 Resp: 346900
Ion Ratio Lower Upper
200 100
173 25.1 19.2 28.8
215 47.2 32.2 48.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022



#71
Pentachlorophenol
Concen: 35.798 ng/ul
RT: 16.975 min Scan# 2361
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18


Tgt Ion:266 Resp: 233058
Ion Ratio Lower Upper
266 100
264 61.8 57.3 85.9
268 63.7 65.7 98.5





#72
Phenanthrene
Concen: 39.119 ng/u1
RT: 17.369 min Scan# 2428
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Instrument :
BNA_P
ClientSampleId :
SLCS233



Tgt Ion:178 Resp: 161355
Ion Ratio Lower Upper
178 100
179 15.1 11.7 17.5
176 18.4 14.5 21.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022



#74
Anthracene
Concen: 38.982 ng/u1
RT: 17.457 min Scan# 2443
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18

Tgt Ion:178 Resp: 1630197
Ion Ratio Lower Upper
178 100
179 15.2 12.3 18.5
176 17.6 13.4 20.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 37.497 ng/uL
RT: 13.375 min Scan# 1750
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18

Instrument :
BNA_P
ClientSampleId :
SLCS233

Tgt Ion:216 Resp: 418922
Ion Ratio Lower Upper
216 100
214 76.9 50.4 75.6
218 47.0 35.2 52.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022



#76
Pentachlorobenzene
Concen: 37.127 ng/uL
RT: 14.898 min Scan# 2008
Delta R.T. -0.000 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18


Tgt Ion:250 Resp: 404310
Ion Ratio Lower Upper
250 100
248 63.8 43.5 65.3
252 62.4 46.5 69.7





#77
 Carbazole
 Concen: 38.733 ng/ul
 RT: 17.722 min Scan# 2489
 Delta R.T. -0.006 min
 Lab File: BP010004.D
 Acq: 21 Apr 2022 07:18




Instrument :
 BNA_P
 ClientSampleId :
 SLCS233



Tgt Ion:167 Resp: 1463838
 Ion Ratio Lower Upper
 167 100
 166 21.3 15.8 23.8
 139 12.8 8.3 12.5




Manual Integrations
 APPROVED


Reviewed By :Jagrut Upadhyay 04/21/2022
 Supervised By :Yogesh Patel 04/25/2022



#78
 Di-n-butylphthalate
 Concen: 39.788 ng/ul
 RT: 18.275 min Scan# 2582
 Delta R.T. -0.006 min
 Lab File: BP010004.D
 Acq: 21 Apr 2022 07:18


Tgt Ion:149 Resp: 1849095
 Ion Ratio Lower Upper
 149 100
 150 9.0 9.6 14.4#
 104 5.3 6.2 9.4#





#80
Fluoranthene
Concen: 38.203 ng/u1
RT: 19.327 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion:202 Resp: 204479
Ion Ratio Lower Upper
202 100
101 9.9 9.6 14.4
100 7.6 6.3 9.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022




#82
Pyrene
Concen: 39.043 ng/u1
RT: 19.680 min Scan# 2821
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18



Tgt Ion:202 Resp: 2118799
Ion Ratio Lower Upper
202 100
101 10.8 10.6 15.8
100 8.8 9.8 14.6#









#83
Butylbenzylphthalate
Concen: 40.439 ng/ul
RT: 20.545 min Scan# 2969
Delta R.T. -0.012 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18

Instrument :
BNA_P
ClientSampleId :
SLCS233

Tgt Ion:149 Resp: 900275
Ion Ratio Lower Upper
149 100
91 73.3 71.1 106.7
206 21.5 28.8 43.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022



#84
3,3'-Dichlorobenzidine
Concen: 36.741 ng/ul
RT: 21.316 min Scan# 3099
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18


Tgt Ion:252 Resp: 666631
Ion Ratio Lower Upper
252 100
254 65.0 49.8 74.6
126 9.7 4.3 6.5





#85
Benzo(a)anthracene
Concen: 39.282 ng/ul
RT: 21.386 min Scan# 3112
Delta R.T. -0.012 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Instrument :
BNA_P
ClientSampleId :
SLCS233



Tgt Ion:228 Resp: 2120340
Ion Ratio Lower Upper
228 100
229 19.7 18.2 27.2
226 26.3 22.8 34.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 40.215 ng/ul
RT: 21.310 min Scan# 3098
Delta R.T. -0.012 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18


Tgt Ion:149 Resp: 1346661
Ion Ratio Lower Upper
149 100
167 28.2 32.8 49.2#
279 5.0 5.6 8.4#





#87
Chrysene
Concen: 40.031 ng/ul
RT: 21.445 min Scan# 3121
Delta R.T. -0.006 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Instrument :
BNA_P
ClientSampleId :
SLCS233



Tgt Ion:228 Resp: 2084762
Ion Ratio Lower Upper
228 100
226 28.8 22.7 34.1
229 19.9 15.2 22.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022



#89
Di-n-octyl phthalate
Concen: 39.867 ng/ul
RT: 22.251 min Scan# 3258
Delta R.T. -0.012 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18


Tgt Ion:149 Resp: 2335028





#90
Benzo(b)fluoranthene
Concen: 40.051 ng/ul
RT: 23.110 min Scan# 3404
Delta R.T. -0.012 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion:252 Resp: 2224489
Ion Ratio Lower Upper
252 100
253 21.9 17.4 26.0
125 7.9 7.3 10.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022




#91
Benzo(k)fluoranthene
Concen: 39.846 ng/ul
RT: 23.157 min Scan# 3412
Delta R.T. -0.012 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Tgt Ion:252 Resp: 2019467
Ion Ratio Lower Upper
252 100
253 21.8 17.9 26.9
125 8.2 5.1 7.7#





#93
Benzo(a)pyrene
Concen: 39.882 ng/ul
RT: 23.751 min Scan# 3513
Delta R.T. -0.012 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion:252 Resp: 2053279
Ion Ratio Lower Upper
252 100
253 21.6 17.5 26.3
125 8.5 6.4 9.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 39.227 ng/ul
RT: 26.415 min Scan# 3966
Delta R.T. -0.024 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Tgt Ion:276 Resp: 2111872
Ion Ratio Lower Upper
276 100
138 19.8 12.3 18.5#
277 25.2 19.3 28.9





#95
Dibenzo(a,h)anthracene
Concen: 38.917 ng/ul
RT: 26.439 min Scan# 3969
Delta R.T. -0.012 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18




Instrument :
BNA_P
ClientSampleId :
SLCS233



Tgt Ion:278 Resp: 1816794
Ion Ratio Lower Upper
278 100
139 12.6 6.6 10.0#
279 23.6 18.4 27.6#

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 04/21/2022
Supervised By :Yogesh Patel 04/25/2022



#96
Benzo(g,h,i)perylene
Concen: 39.816 ng/ul
RT: 27.203 min Scan# 4100
Delta R.T. -0.018 min
Lab File: BP010004.D
Acq: 21 Apr 2022 07:18

Tgt Ion:276 Resp: 1767464
Ion Ratio Lower Upper
276 100
138 17.5 11.6 17.4#
277 24.1 15.9 23.9#

