

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041923\
 Data File : BP014723.D
 Acq On : 19 Apr 2023 11:07
 Operator : CG/JU
 Sample : PB152181BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK181

Quant Time: Apr 20 01:45:23 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP041823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 19 01:37:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.152	152	181000	20.000	ng/ul	0.00
20) Naphthalene-d8	10.981	136	741994	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.787	164	470204	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.534	188	1061158	20.000	ng/ul	0.00
79) Chrysene-d12	21.610	240	996745	20.000	ng/ul	0.00
88) Perylene-d12	24.192	264	1137795	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.487	96	27152	5.962	ng/uL	0.00
4) Pyridine-d5	3.917	84	302	0.024	ng/ul	0.00
7) Phenol-d5	7.287	99	403429	25.779	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.469	67	284028	29.419	ng/ul	0.00
11) 2-Chlorophenol-d4	7.675	132	338908	29.160	ng/ul	0.00
15) 4-Methylphenol-d8	8.852	113	302146	24.766	ng/ul	0.00
21) Nitrobenzene-d5	9.322	128	143664	30.899	ng/ul	0.00
24) 2-Nitrophenol-d4	10.052	143	125759	29.947	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.587	165	307926	27.006	ng/ul	0.00
31) 4-Chloroaniline-d4	11.116	131	1266	0.073	ng/ul	0.00
46) Dimethylphthalate-d6	14.187	166	1187160	31.886	ng/ul	0.00
49) Acenaphthylene-d8	14.481	160	1234248	28.606	ng/ul	0.00
54) 4-Nitrophenol-d4	14.945	143	133061	22.469	ng/ul	0.00
60) Fluorene-d10	15.775	176	995342	31.156	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.875	200	80283	18.695	ng/ul	0.00
73) Anthracene-d10	17.634	188	1587813	30.973	ng/ul	0.00
81) Pyrene-d10	19.851	212	1883675	32.990	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.021	264	1757195	29.798	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041923\
Data File : BP014723.D
Acq On : 19 Apr 2023 11:07
Operator : CG/JU
Sample : PB152181BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK181

Quant Time: Apr 20 01:45:23 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP041823.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Apr 19 01:37:13 2023
Response via : Initial Calibration

