

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041924\
 Data File : BP020032.D
 Acq On : 19 Apr 2024 17:33
 Operator : MA/JU
 Sample : SST02090
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST020611

Quant Time: Apr 20 00:36:08 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Apr 20 00:34:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.640	152	92905	20.000	ng/u1	0.00
20) Naphthalene-d8	10.410	136	384799	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.316	164	258684	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.157	188	543121	20.000	ng/u1	0.00
79) Chrysene-d12	21.657	240	549654	20.000	ng/u1	0.00
88) Perylene-d12	25.051	264	595828	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.234	96	18526	8.692	ng/uL	0.00
4) Pyridine-d5	3.634	84	135257	20.956	ng/u1	0.00
7) Phenol-d5	6.822	99	163752	20.741	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.993	67	117244	22.398	ng/u1	0.00
11) 2-Chlorophenol-d4	7.175	132	115989	20.622	ng/u1	0.00
15) 4-Methylphenol-d8	8.346	113	126713	20.735	ng/u1	0.00
21) Nitrobenzene-d5	8.799	128	59573	20.272	ng/u1	0.00
24) 2-Nitrophenol-d4	9.510	143	61864	19.164	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.040	165	127194	21.075	ng/u1	0.00
31) 4-Chloroaniline-d4	10.569	131	165749	21.021	ng/u1	0.00
46) Dimethylphthalate-d6	13.728	166	378278	20.797	ng/u1	0.00
49) Acenaphthylene-d8	13.998	160	446739	21.011	ng/u1	0.00
54) 4-Nitrophenol-d4	14.557	143	56830	20.041	ng/u1	0.00
60) Fluorene-d10	15.340	176	331559	21.338	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.487	200	63030	19.421	ng/u1	0.00
73) Anthracene-d10	17.263	188	525951	21.656	ng/u1	0.00
81) Pyrene-d10	19.663	212	621855	19.431	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.810	264	612526	21.238	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.264	88	21740	8.772	ng/uL	100
5) Pyridine	3.652	79	143427	21.594	ng/u1	100
6) Benzaldehyde	6.811	77	118732	29.205	ng/u1	100
8) Phenol	6.846	94	174432	21.043	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.081	93	145015	21.502	ng/u1	100
12) 2-Chlorophenol	7.205	128	122584	20.669	ng/u1	100
13) 2-Methylphenol	8.081	108	128370	21.405	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.158	45	213382	22.701	ng/u1	100
16) Acetophenone	8.463	105	220711	22.078	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.446	70	130457	23.002	ng/u1	100
18) 4-Methylphenol	8.405	108	137330	21.297	ng/u1	100
19) Hexachloroethane	8.693	117	59985	21.548	ng/u1	100
22) Nitrobenzene	8.840	77	191524	22.347	ng/u1	100
23) Isophorone	9.357	82	353680	22.186	ng/u1	100
25) 2-Nitrophenol	9.540	139	67640	19.944	ng/u1	100
26) 2,4-Dimethylphenol	9.599	107	161983	21.354	ng/u1	100
27) Bis(2-Chloroethoxy)met...	9.840	93	199404	21.580	ng/u1	100
29) 2,4-Dichlorophenol	10.069	162	124555	21.019	ng/u1	100
30) Naphthalene	10.457	128	419351	21.132	ng/u1	100
32) 4-Chloroaniline	10.593	127	164583	21.175	ng/u1	100
33) Hexachlorobutadiene	10.728	225	100712	21.451	ng/u1	100
34) Caprolactam	11.399	113	28388	15.328	ng/u1	100
35) 4-Chloro-3-methylphenol	11.722	107	145571	21.862	ng/u1	100
36) 2-Methylnaphthalene	12.087	142	288556	21.609	ng/u1	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041924\
 Data File : BP020032.D
 Acq On : 19 Apr 2024 17:33
 Operator : MA/JU
 Sample : SST02090
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST020611

Quant Time: Apr 20 00:36:08 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Apr 20 00:34:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.310	142	283734	21.400	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.457	216	173337	20.429	ng/ul	100
40) Hexachlorocyclopentadiene	12.422	237	89229	18.517	ng/ul	100
41) 2,4,6-Trichlorophenol	12.722	196	106530	20.387	ng/ul	100
42) 2,4,5-Trichlorophenol	12.798	196	109334	20.382	ng/ul	100
43) 1,1'-Biphenyl	13.128	154	382885	20.560	ng/ul	100
44) 2-Chloronaphthalene	13.169	162	312114	20.662	ng/ul	100
45) 2-Nitroaniline	13.398	65	103435	21.640	ng/ul	100
47) Dimethylphthalate	13.781	163	387976	21.119	ng/ul	100
48) 2,6-Dinitrotoluene	13.910	165	76144	20.504	ng/ul	100
50) Acenaphthylene	14.034	152	494695	20.892	ng/ul	100
51) 3-Nitroaniline	14.251	138	73365	21.519	ng/ul	100
52) Acenaphthene	14.381	153	322907	20.578	ng/ul	100
53) 2,4-Dinitrophenol	14.475	184	39718	18.696	ng/ul	100
55) 4-Nitrophenol	14.575	109	63947	23.436	ng/ul	100
56) Dibenzofuran	14.728	168	452616	21.107	ng/ul	100
57) 2,4-Dinitrotoluene	14.722	165	107029	21.077	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	14.963	232	97666	20.956	ng/ul	100
59) Diethylphthalate	15.181	149	402938	21.925	ng/ul	100
61) Fluorene	15.398	166	381633	21.707	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.398	204	206183	21.774	ng/ul	100
63) 4-Nitroaniline	15.451	138	67208	22.858	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.504	198	66869	19.443	ng/ul	100
67) N-Nitrosodiphenylamine	15.622	169	315329	20.925	ng/ul	100
68) 4-Bromophenyl-phenylether	16.322	248	124081	20.891	ng/ul	100
69) Hexachlorobenzene	16.428	284	133535	20.383	ng/ul	100
70) Atrazine	16.628	200	118281	20.589	ng/ul	100
71) Pentachlorophenol	16.798	266	79349	20.411	ng/ul	100
72) Phenanthrene	17.198	178	596632	21.144	ng/ul	100
74) Anthracene	17.298	178	620941	21.526	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.081	216	172012	19.547	ng/ul	100
76) Pentachlorobenzene	14.640	250	162843	19.909	ng/ul	100
77) Carbazole	17.598	167	538204	21.842	ng/ul	100
78) Di-n-butylphthalate	18.192	149	671697	22.045	ng/ul	100
80) Fluoranthene	19.316	202	730801	19.383	ng/ul	100
82) Pyrene	19.692	202	749766	19.238	ng/ul	100
83) Butylbenzylphthalate	20.675	149	299310	19.205	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.569	252	261337	22.382	ng/ul	100
85) Benzo(a)anthracene	21.633	228	794015	20.807	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.592	149	464790	20.202	ng/ul	100
87) Chrysene	21.710	228	754717	21.144	ng/ul	100
89) Di-n-octyl phthalate	22.898	149	776992	20.726	ng/ul	100
90) Benzo(b)fluoranthene	23.974	252	763933	21.622	ng/ul	100
91) Benzo(k)fluoranthene	24.045	252	767995	21.379	ng/ul	100
93) Benzo(a)pyrene	24.886	252	739607	21.605	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	28.904	276	728747	17.228	ng/ul	100
95) Dibenzo(a,h)anthracene	28.998	278	674319	19.685	ng/ul	100
96) Benzo(g,h,i)perylene	30.098	276	677001	19.847	ng/ul	100





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 8.772 ng/uL
 RT: 3.264 min Scan# 30
 Delta R.T. 0.000 min
 Lab File: BP020032.D
 Acq: 19 Apr 2024 17:33




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020611


Tgt Ion:	88	Resp:	21740
Ion	Ratio	Lower	Upper
88	100		
43	33.3	26.6	40.0
58	76.8	61.4	92.2



#5
 Pyridine
 Concen: 21.594 ng/uL
 RT: 3.652 min Scan# 96
 Delta R.T. 0.000 min
 Lab File: BP020032.D
 Acq: 19 Apr 2024 17:33

Tgt Ion:	79	Resp:	143427
Ion	Ratio	Lower	Upper
79	100		
52	76.2	61.0	91.4
51	39.4	31.5	47.3









#6
Benzaldehyde
Concen: 29.205 ng/ul
RT: 6.811 min Scan# 611
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion: 77 Resp: 118732
Ion Ratio Lower Upper
77 100
105 85.9 68.7 103.1
106 79.6 63.7 95.5



#8
Phenol
Concen: 21.043 ng/ul
RT: 6.846 min Scan# 639
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33

Tgt Ion: 94 Resp: 174432
Ion Ratio Lower Upper
94 100
65 34.7 27.8 41.6
66 50.0 40.0 60.0









#10
Bis(2-Chloroethyl)ether
Concen: 21.502 ng/ul
RT: 7.081 min Scan# 679
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion: 93 Resp: 145015
Ion Ratio Lower Upper
93 100
63 85.7 68.6 102.8
95 33.2 26.6 39.8



#12
2-Chlorophenol
Concen: 20.669 ng/ul
RT: 7.205 min Scan# 700
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33

Tgt Ion: 128 Resp: 122584
Ion Ratio Lower Upper
128 100
64 64.2 51.4 77.0
130 33.1 26.5 39.7





#13

2-Methylphenol

Concen: 21.405 ng/ul

RT: 8.081 min Scan# 849

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33

Instrument :

BNA_P

ClientSampleId :





SSTD020611

Tgt Ion:108 Resp: 128370

Ion Ratio Lower Upper

108 100

107 84.1 67.3 100.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 22.701 ng/ul

RT: 8.158 min Scan# 862

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33




Tgt Ion: 45 Resp: 213382


Ion Ratio Lower Upper

45 100

77 15.7 12.6 18.8

79 11.0 8.8 13.2









#16
Acetophenone
Concen: 22.078 ng/ul
RT: 8.463 min Scan# 914
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:105 Resp: 220711
Ion Ratio Lower Upper
105 100
77 87.9 70.3 105.5
51 31.4 25.1 37.7



#17
N-Nitroso-di-n-propylamine
Concen: 23.002 ng/ul
RT: 8.446 min Scan# 911
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33

Tgt Ion: 70 Resp: 130457
Ion Ratio Lower Upper
70 100
42 59.7 47.8 71.6
101 8.1 6.5 9.7
130 15.6 12.5 18.7









#18
4-Methylphenol
Concen: 21.297 ng/ul
RT: 8.405 min Scan# 904
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:108 Resp: 137330
Ion Ratio Lower Upper
108 100
107 114.4 91.5 137.3



#19
Hexachloroethane
Concen: 21.548 ng/ul
RT: 8.693 min Scan# 953
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33

Tgt Ion:117 Resp: 59985
Ion Ratio Lower Upper
117 100
201 109.7 87.8 131.6
199 68.8 55.0 82.6









#22
Nitrobenzene
Concen: 22.347 ng/ul
RT: 8.840 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion: 77 Resp: 191524
Ion Ratio Lower Upper
77 100
123 35.6 28.5 42.7
65 14.6 11.7 17.5



#23
Isophorone
Concen: 22.186 ng/ul
RT: 9.357 min Scan# 1066
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33

Tgt Ion: 82 Resp: 353680
Ion Ratio Lower Upper
82 100
95 7.5 6.0 9.0
138 13.6 10.9 16.3









#25
2-Nitrophenol
Concen: 19.944 ng/ul
RT: 9.540 min Scan# 1097
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:	139	Resp:	67640
Ion	Ratio	Lower	Upper
139	100		
65	73.2	58.6	87.8
109	43.5	34.8	52.2



#26
2,4-Dimethylphenol
Concen: 21.354 ng/ul
RT: 9.599 min Scan# 1107
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33

Tgt Ion:	107	Resp:	161983
Ion	Ratio	Lower	Upper
107	100		
121	42.5	34.0	51.0
122	73.6	58.9	88.3









#27
 Bis(2-Chloroethoxy)methane
 Concen: 21.580 ng/ul
 RT: 9.840 min Scan# 1148
 Delta R.T. 0.000 min
 Lab File: BP020032.D
 Acq: 19 Apr 2024 17:33




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020611


Tgt Ion: 93 Resp: 199404
 Ion Ratio Lower Upper
 93 100
 95 31.7 25.4 38.0
 123 9.4 7.5 11.3



#29
 2,4-Dichlorophenol
 Concen: 21.019 ng/ul
 RT: 10.069 min Scan# 1187
 Delta R.T. 0.000 min
 Lab File: BP020032.D
 Acq: 19 Apr 2024 17:33

Tgt Ion: 162 Resp: 124555
 Ion Ratio Lower Upper
 162 100
 164 64.6 51.7 77.5
 98 41.0 32.8 49.2





#30

Naphthalene

Concen: 21.132 ng/ul

RT: 10.457 min Scan# 1253

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion:128 Resp: 419351

Ion Ratio Lower Upper

128 100

129 10.9 8.7 13.1

127 13.1 10.5 15.7



#32

4-Chloroaniline

Concen: 21.175 ng/ul

RT: 10.593 min Scan# 1276

Delta R.T. 0.000 min

Lab File: BP020032.D




Acq: 19 Apr 2024 17:33


Tgt Ion:127 Resp: 164583

Ion Ratio Lower Upper

127 100

129 32.4 25.9 38.9





#33

Hexachlorobutadiene

Concen: 21.451 ng/ul

RT: 10.728 min Scan# 1299

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion:225 Resp: 100712

Ion Ratio Lower Upper

225 100

223 63.1 50.5 75.7

227 65.3 52.2 78.4



#34

Caprolactam

Concen: 15.328 ng/ul

RT: 11.399 min Scan# 1413

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33




Tgt Ion:113 Resp: 28388


Ion Ratio Lower Upper

113 100

55 175.6 140.5 210.7

56 142.2 113.8 170.6









#35
4-Chloro-3-methylphenol
Concen: 21.862 ng/ul
RT: 11.722 min Scan# 1468
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:107 Resp: 145571
Ion Ratio Lower Upper
107 100
144 22.5 18.0 27.0
142 70.4 56.3 84.5



#36
2-Methylnaphthalene
Concen: 21.609 ng/ul
RT: 12.087 min Scan# 1530
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33

Tgt Ion:142 Resp: 288556
Ion Ratio Lower Upper
142 100
141 90.1 72.1 108.1









#37
1-Methylnaphthalene
Concen: 21.400 ng/ul
RT: 12.310 min Scan# 1568
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:142 Resp: 283734
Ion Ratio Lower Upper
142 100
141 91.0 72.8 109.2
116 5.0 4.0 6.0



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.429 ng/ul
RT: 12.457 min Scan# 1593
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33

Tgt Ion:216 Resp: 173337
Ion Ratio Lower Upper
216 100
214 80.2 64.2 96.2
179 21.6 17.3 25.9
108 21.0 16.8 25.2





#40

Hexachlorocyclopentadiene

Concen: 18.517 ng/ul

RT: 12.422 min Scan# 1587

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion:237 Resp: 89229

Ion Ratio Lower Upper

237 100

235 60.3 48.2 72.4

272 11.7 9.4 14.0



#41

2,4,6-Trichlorophenol

Concen: 20.387 ng/ul

RT: 12.722 min Scan# 1638

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33




Tgt Ion:196 Resp: 106530


Ion Ratio Lower Upper

196 100

198 98.5 78.8 118.2

200 31.1 24.9 37.3





#42

2,4,5-Trichlorophenol

Concen: 20.382 ng/ul

RT: 12.798 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion:196 Resp: 109334

Ion Ratio Lower Upper

196 100

198 96.9 77.5 116.3

200 30.6 24.5 36.7



#43

1,1'-Biphenyl

Concen: 20.560 ng/ul

RT: 13.128 min Scan# 1707

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33




Tgt Ion:154 Resp: 382885


Ion Ratio Lower Upper

154 100

153 40.3 32.2 48.4

76 14.1 11.3 16.9





#44

2-Chloronaphthalene

Concen: 20.662 ng/ul

RT: 13.169 min Scan# 1714

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion:162 Resp: 312114

Ion Ratio Lower Upper

162 100

164 33.8 27.0 40.6

127 42.1 33.7 50.5



#45

2-Nitroaniline

Concen: 21.640 ng/ul

RT: 13.398 min Scan# 1753

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33




Tgt Ion: 65 Resp: 103435


Ion Ratio Lower Upper

65 100

92 61.7 49.4 74.0

138 77.8 62.2 93.4









#47
Dimethylphthalate
Concen: 21.119 ng/ul
RT: 13.781 min Scan# 1818
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:163 Resp: 387976
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 10.1 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 20.504 ng/ul
RT: 13.910 min Scan# 1840
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33

Tgt Ion:165 Resp: 76144
Ion Ratio Lower Upper
165 100
89 76.2 61.0 91.4
121 23.4 18.7 28.1





#50

Acenaphthylene

Concen: 20.892 ng/ul

RT: 14.034 min Scan# 1861

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion:152 Resp: 494695

Ion Ratio Lower Upper

152 100

151 19.3 15.4 23.2

153 13.3 10.6 16.0



#51

3-Nitroaniline

Concen: 21.519 ng/ul

RT: 14.251 min Scan# 1898

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33




Tgt Ion:138 Resp: 73365


Ion Ratio Lower Upper

138 100

108 13.0 10.4 15.6

92 135.5 108.4 162.6









#52
Acenaphthene
Concen: 20.578 ng/ul
RT: 14.381 min Scan# 1920
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:153 Resp: 322907
Ion Ratio Lower Upper
153 100
152 47.4 37.9 56.9
154 89.3 71.4 107.2



#53
2,4-Dinitrophenol
Concen: 18.696 ng/ul
RT: 14.475 min Scan# 1936
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33

Tgt Ion:184 Resp: 39718
Ion Ratio Lower Upper
184 100
63 81.3 65.0 97.6
154 69.3 55.4 83.2









#55
4-Nitrophenol
Concen: 23.436 ng/ul
RT: 14.575 min Scan# 1953
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:109 Resp: 63947
Ion Ratio Lower Upper
109 100
139 97.4 77.9 116.9
65 126.1 100.9 151.3



#56
Dibenzofuran
Concen: 21.107 ng/ul
RT: 14.728 min Scan# 1979
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33

Tgt Ion:168 Resp: 452616
Ion Ratio Lower Upper
168 100
139 40.8 32.6 49.0









#57
2,4-Dinitrotoluene
Concen: 21.077 ng/ul
RT: 14.722 min Scan# 1978
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:165 Resp: 107029
Ion Ratio Lower Upper
165 100
63 87.8 70.2 105.4
182 2.5 2.0 3.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.956 ng/ul
RT: 14.963 min Scan# 2019
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33

Tgt Ion:232 Resp: 97666
Ion Ratio Lower Upper
232 100
131 43.1 34.5 51.7
130 3.0 2.4 3.6
166 31.8 25.4 38.2





#59

Diethylphthalate

Concen: 21.925 ng/ul

RT: 15.181 min Scan# 2056

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion:149 Resp: 402938

Ion Ratio Lower Upper

149 100

177 21.1 16.9 25.3

150 12.4 9.9 14.9



#61

Fluorene

Concen: 21.707 ng/ul

RT: 15.398 min Scan# 2093

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33




Tgt Ion:166 Resp: 381633


Ion Ratio Lower Upper

166 100

165 97.8 78.2 117.4

167 13.6 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 21.774 ng/ul
RT: 15.398 min Scan# 2093
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:204 Resp: 206183
Ion Ratio Lower Upper
204 100
206 31.4 25.1 37.7
141 61.6 49.3 73.9



#63
4-Nitroaniline
Concen: 22.858 ng/ul
RT: 15.451 min Scan# 2102
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33

Tgt Ion:138 Resp: 67208
Ion Ratio Lower Upper
138 100
92 69.0 55.2 82.8
108 119.2 95.4 143.0









#66
4,6-Dinitro-2-methylphenol
Concen: 19.443 ng/ul
RT: 15.504 min Scan# 2111
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:198 Resp: 66869
Ion Ratio Lower Upper
198 100
182 3.8 3.0 4.6
77 32.0 25.6 38.4



#67
N-Nitrosodiphenylamine
Concen: 20.925 ng/ul
RT: 15.622 min Scan# 2131
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33

Tgt Ion:169 Resp: 315329
Ion Ratio Lower Upper
169 100
168 68.8 55.0 82.6
167 36.9 29.5 44.3









#68
4-Bromophenyl-phenylether
Concen: 20.891 ng/ul
RT: 16.322 min Scan# 211
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:248 Resp: 124081
Ion Ratio Lower Upper
248 100
250 99.0 79.2 118.8
141 81.4 65.1 97.7



#69
Hexachlorobenzene
Concen: 20.383 ng/ul
RT: 16.428 min Scan# 2268
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33

Tgt Ion:284 Resp: 133535
Ion Ratio Lower Upper
284 100
142 38.2 30.6 45.8
249 33.6 26.9 40.3





#70

Atrazine

Concen: 20.589 ng/ul

RT: 16.628 min Scan# 2302

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion:200 Resp: 118281

Ion Ratio Lower Upper

200 100

173 24.1 19.3 28.9

215 45.7 36.6 54.8



#71

Pentachlorophenol

Concen: 20.411 ng/ul

RT: 16.798 min Scan# 2331

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33




Tgt Ion:266 Resp: 79349


Ion Ratio Lower Upper

266 100

264 63.2 50.6 75.8

268 62.9 50.3 75.5





#72

Phenanthrene

Concen: 21.144 ng/ul

RT: 17.198 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion:178 Resp: 596632

Ion Ratio Lower Upper

178 100

179 15.3 12.2 18.4

176 19.4 15.5 23.3



#74

Anthracene

Concen: 21.526 ng/ul

RT: 17.298 min Scan# 2416

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33




Tgt Ion:178 Resp: 620941


Ion Ratio Lower Upper

178 100

179 15.3 12.2 18.4

176 18.2 14.6 21.8









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.547 ng/uL
RT: 13.081 min Scan# 1699
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33




Instrument :
BNA_P
ClientSampleId :
SSTD020611


Tgt Ion:216 Resp: 172012
Ion Ratio Lower Upper
216 100
214 74.1 59.3 88.9
218 47.3 37.8 56.8



#76
Pentachlorobenzene
Concen: 19.909 ng/uL
RT: 14.640 min Scan# 1964
Delta R.T. 0.000 min
Lab File: BP020032.D
Acq: 19 Apr 2024 17:33

Tgt Ion:250 Resp: 162843
Ion Ratio Lower Upper
250 100
248 64.0 51.2 76.8
252 63.6 50.9 76.3









#77
 Carbazole
 Concen: 21.842 ng/ul
 RT: 17.598 min Scan# 2467
 Delta R.T. 0.000 min
 Lab File: BP020032.D
 Acq: 19 Apr 2024 17:33




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020611


Tgt Ion:	167	Resp:	538204
Ion	Ratio	Lower	Upper
167	100		
166	21.8	17.4	26.2
139	13.6	10.9	16.3



#78
 Di-n-butylphthalate
 Concen: 22.045 ng/ul
 RT: 18.192 min Scan# 2568
 Delta R.T. 0.000 min
 Lab File: BP020032.D
 Acq: 19 Apr 2024 17:33

Tgt Ion:	149	Resp:	671697
Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.4	11.0
104	6.6	5.3	7.9





#80

Fluoranthene

Concen: 19.383 ng/ul

RT: 19.316 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion:202 Resp: 730801

Ion Ratio Lower Upper

202 100

101 9.6 7.7 11.5

100 7.1 5.7 8.5



#82

Pyrene

Concen: 19.238 ng/ul

RT: 19.692 min Scan# 2823

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33




Tgt Ion:202 Resp: 749766


Ion Ratio Lower Upper

202 100

101 11.6 9.3 13.9

100 9.5 7.6 11.4





#83

Butylbenzylphthalate

Concen: 19.205 ng/ul

RT: 20.675 min Scan# 2990

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion:149 Resp: 299310

Ion Ratio Lower Upper

149 100

91 87.0 69.6 104.4

206 21.7 17.4 26.0



#84

3,3'-Dichlorobenzidine

Concen: 22.382 ng/ul

RT: 21.569 min Scan# 3142

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33




Tgt Ion:252 Resp: 261337


Ion Ratio Lower Upper

252 100

254 64.4 51.5 77.3

126 9.9 7.9 11.9





#85

Benzo(a)anthracene

Concen: 20.807 ng/ul

RT: 21.633 min Scan# 3153

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion:228 Resp: 794015

Ion Ratio Lower Upper

228 100

229 19.6 15.7 23.5

226 26.7 21.4 32.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 20.202 ng/ul

RT: 21.592 min Scan# 3146

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33




Tgt Ion:149 Resp: 464790


Ion Ratio Lower Upper

149 100

167 26.6 21.3 31.9

279 4.0 3.2 4.8





#87

Chrysene

Concen: 21.144 ng/ul

RT: 21.710 min Scan# 3166

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion:228 Resp: 754717

Ion Ratio Lower Upper

228 100

226 30.3 24.2 36.4

229 19.3 15.4 23.2



#89

Di-n-octyl phthalate

Concen: 20.726 ng/ul




RT: 22.898 min Scan# 3368


Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33

Tgt Ion:149 Resp: 776992





#90

Benzo(b)fluoranthene

Concen: 21.622 ng/ul

RT: 23.974 min Scan# 3551

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion:252 Resp: 763933

Ion Ratio Lower Upper

252 100

253 22.3 17.8 26.8

125 7.2 5.8 8.6



#91

Benzo(k)fluoranthene

Concen: 21.379 ng/ul

RT: 24.045 min Scan# 3563

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33




Tgt Ion:252 Resp: 767995


Ion Ratio Lower Upper

252 100

253 21.9 17.5 26.3

125 7.4 5.9 8.9





#93

Benzo(a)pyrene

Concen: 21.605 ng/ul

RT: 24.886 min Scan# 3706

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion:252 Resp: 739607

Ion Ratio Lower Upper

252 100

253 21.7 17.4 26.0

125 8.8 7.0 10.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 17.228 ng/ul

RT: 28.904 min Scan# 4389

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33




Tgt Ion:276 Resp: 728747


Ion Ratio Lower Upper

276 100

138 14.6 11.7 17.5

277 23.3 18.6 28.0





#95

Dibenzo(a,h)anthracene

Concen: 19.685 ng/ul

RT: 28.998 min Scan# 4405

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33

Instrument :

BNA_P

ClientSampleId :

SSTD020611





Tgt Ion:278 Resp: 674319

Ion Ratio Lower Upper

278 100

139 11.8 9.4 14.2

279 23.8 19.0 28.6



#96

Benzo(g,h,i)perylene

Concen: 19.847 ng/ul

RT: 30.098 min Scan# 4592

Delta R.T. 0.000 min

Lab File: BP020032.D

Acq: 19 Apr 2024 17:33

Tgt Ion:276 Resp: 677001

Ion Ratio Lower Upper

276 100

138 16.9 13.5 20.3

277 24.3 19.4 29.2

