

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041924\
 Data File : BP020036.D
 Acq On : 19 Apr 2024 21:12
 Operator : MA/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SICV615

Quant Time: Apr 20 01:38:57 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Apr 20 01:32:50 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.640	152	91132	20.000	ng/u1	0.00
20) Naphthalene-d8	10.410	136	387053	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.316	164	267218	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.169	188	593460	20.000	ng/u1	0.01
79) Chrysene-d12	21.675	240	584797	20.000	ng/u1	0.02
88) Perylene-d12	25.057	264	637946	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.234	96	18438	8.290	ng/uL	0.00
4) Pyridine-d5	3.634	84	136530	20.855	ng/u1	0.00
7) Phenol-d5	6.822	99	165117	20.789	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.993	67	116346	20.957	ng/u1	0.00
11) 2-Chlorophenol-d4	7.175	132	115033	21.212	ng/u1	0.00
15) 4-Methylphenol-d8	8.346	113	131920	21.263	ng/u1	0.00
21) Nitrobenzene-d5	8.793	128	60260	21.644	ng/u1	0.00
24) 2-Nitrophenol-d4	9.511	143	66002	21.931	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.046	165	128968	21.602	ng/u1	0.00
31) 4-Chloroaniline-d4	10.569	131	175095	21.969	ng/u1	0.00
46) Dimethylphthalate-d6	13.728	166	405650	22.213	ng/u1	0.00
49) Acenaphthylene-d8	14.004	160	459624	21.116	ng/u1	0.00
54) 4-Nitrophenol-d4	14.557	143	66228	23.823	ng/u1	0.00
60) Fluorene-d10	15.345	176	344408	21.186	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.493	200	71436	20.604	ng/u1	0.00
73) Anthracene-d10	17.269	188	554921	20.872	ng/u1	0.00
81) Pyrene-d10	19.675	212	674274	21.455	ng/u1	0.01
92) Benzo(a)pyrene-d12	24.816	264	648229	20.801	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.264	88	21511	8.604	ng/uL	97
5) Pyridine	3.652	79	141289	20.844	ng/u1	97
6) Benzaldehyde	6.811	77	116462	24.528	ng/u1	99
8) Phenol	6.852	94	175963	21.024	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.081	93	140451	20.551	ng/u1	99
12) 2-Chlorophenol	7.205	128	123044	21.533	ng/u1	95
13) 2-Methylphenol	8.081	108	127213	20.740	ng/u1	92
14) 2,2'-oxybis(1-Chloropr...	8.158	45	213067	21.331	ng/u1	99
16) Acetophenone	8.464	105	225768	21.243	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.446	70	131413	21.168	ng/u1	99
18) 4-Methylphenol	8.405	108	137232	21.026	ng/u1	95
19) Hexachloroethane	8.693	117	59389	21.115	ng/u1	96
22) Nitrobenzene	8.840	77	193689	21.347	ng/u1	99
23) Isophorone	9.358	82	364004	21.615	ng/u1	99
25) 2-Nitrophenol	9.540	139	72535	22.628	ng/u1	95
26) 2,4-Dimethylphenol	9.599	107	165296	21.555	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.840	93	200338	21.217	ng/u1	99
29) 2,4-Dichlorophenol	10.075	162	125129	21.524	ng/u1	99
30) Naphthalene	10.458	128	420713	20.961	ng/u1	100
32) 4-Chloroaniline	10.593	127	172340	21.795	ng/u1	99
33) Hexachlorobutadiene	10.728	225	98896	20.688	ng/u1	98
34) Caprolactam	11.405	113	41200	22.613	ng/u1	93
35) 4-Chloro-3-methylphenol	11.722	107	155780	22.351	ng/u1	99
36) 2-Methylnaphthalene	12.087	142	290899	20.993	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041924\
 Data File : BP020036.D
 Acq On : 19 Apr 2024 21:12
 Operator : MA/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SICV615


Quant Time: Apr 20 01:38:57 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Apr 20 01:32:50 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.310	142	290925	21.142	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.457	216	177089	20.335	ng/ul	97
40) Hexachlorocyclopentadiene	12.422	237	92040	18.346	ng/ul	99
41) 2,4,6-Trichlorophenol	12.716	196	110691	21.213	ng/ul	95
42) 2,4,5-Trichlorophenol	12.799	196	114962	21.195	ng/ul	96
43) 1,1'-Biphenyl	13.128	154	398779	20.816	ng/ul	99
44) 2-Chloronaphthalene	13.169	162	318909	20.838	ng/ul	98
45) 2-Nitroaniline	13.399	65	115543	23.238	ng/ul	99
47) Dimethylphthalate	13.781	163	414114	22.083	ng/ul	99
48) 2,6-Dinitrotoluene	13.910	165	80363	22.710	ng/ul	99
50) Acenaphthylene	14.034	152	515407	21.006	ng/ul	98
51) 3-Nitroaniline	14.257	138	78981	23.581	ng/ul	99
52) Acenaphthene	14.381	153	340792	21.071	ng/ul	98
53) 2,4-Dinitrophenol	14.481	184	43422	20.406	ng/ul	95
55) 4-Nitrophenol	14.575	109	71872	22.692	ng/ul	97
56) Dibenzofuran	14.734	168	470928	21.090	ng/ul	100
57) 2,4-Dinitrotoluene	14.728	165	118768	23.709	ng/ul	93
58) 2,3,4,6-Tetrachlorophenol	14.969	232	105288	22.415	ng/ul	97
59) Diethylphthalate	15.187	149	421615	22.389	ng/ul	99
61) Fluorene	15.404	166	393612	21.334	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.404	204	208249	20.774	ng/ul	98
63) 4-Nitroaniline	15.457	138	74142	24.932	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.504	198	74901	20.785	ng/ul	94
67) N-Nitrosodiphenylamine	15.628	169	331734	20.208	ng/ul	96
68) 4-Bromophenyl-phenylether	16.328	248	131968	20.276	ng/ul	97
69) Hexachlorobenzene	16.434	284	144310	20.482	ng/ul	97
70) Atrazine	16.634	200	130356	21.624	ng/ul	95
71) Pentachlorophenol	16.810	266	87682	20.532	ng/ul	98
72) Phenanthrene	17.210	178	640150	20.774	ng/ul	99
74) Anthracene	17.298	178	658728	20.960	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.081	216	179628	19.092	ng/uL	97
76) Pentachlorobenzene	14.640	250	170657	19.275	ng/uL	100
77) Carbazole	17.604	167	577331	21.438	ng/ul	99
78) Di-n-butylphthalate	18.204	149	721026	22.192	ng/ul	99
80) Fluoranthene	19.322	202	785076	21.248	ng/ul	100
82) Pyrene	19.704	202	827071	21.515	ng/ul	99
83) Butylbenzylphthalate	20.686	149	337795	23.046	ng/ul	93
84) 3,3'-Dichlorobenzidine	21.580	252	279793	21.637	ng/ul	98
85) Benzo(a)anthracene	21.651	228	860155	21.264	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.598	149	488413	21.616	ng/ul	99
87) Chrysene	21.716	228	791853	20.786	ng/ul	99
89) Di-n-octyl phthalate	22.904	149	939670	23.832	ng/ul	100
90) Benzo(b)fluoranthene	23.980	252	972761	25.301	ng/ul	99
91) Benzo(k)fluoranthene	24.051	252	1010148	25.628	ng/ul	99
93) Benzo(a)pyrene	24.886	252	762283	20.575	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	28.892	276	409432	9.530	ng/ul	99
95) Dibenzo(a,h)anthracene	28.992	278	732654	20.930	ng/ul	98
96) Benzo(g,h,i)perylene	30.068	276	356558	10.376	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 20 01:38:57 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP041924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Apr 20 01:32:50 2024
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 8.604 ng/uL
 RT: 3.264 min Scan# 30
 Delta R.T. -0.000 min
 Lab File: BP020036.D
 Acq: 19 Apr 2024 21:12




Instrument :
 BNA_P
 ClientSampleId :
 SICV615


Tgt Ion:	88	Resp:	21511
Ion	Ratio	Lower	Upper
88	100		
43	33.9	26.6	40.0
58	79.9	61.4	92.2



#5
 Pyridine
 Concen: 20.844 ng/uL
 RT: 3.652 min Scan# 96
 Delta R.T. 0.000 min
 Lab File: BP020036.D
 Acq: 19 Apr 2024 21:12

Tgt Ion:	79	Resp:	141289
Ion	Ratio	Lower	Upper
79	100		
52	80.3	61.0	91.4
51	39.2	31.5	47.3









#6
Benzaldehyde
Concen: 24.528 ng/ul
RT: 6.811 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12




Instrument :
BNA_P
ClientSampleId :
SICV615


Tgt Ion: 77 Resp: 116462
Ion Ratio Lower Upper
77 100
105 84.8 68.7 103.1
106 78.4 63.7 95.5



#8
Phenol
Concen: 21.024 ng/ul
RT: 6.852 min Scan# 640
Delta R.T. 0.006 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12

Tgt Ion: 94 Resp: 175963
Ion Ratio Lower Upper
94 100
65 34.4 27.8 41.6
66 47.0 40.0 60.0









#10
Bis(2-Chloroethyl)ether
Concen: 20.551 ng/ul
RT: 7.081 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12




Instrument :
BNA_P
ClientSampleId :
SICV615


Tgt Ion: 93 Resp: 140451
Ion Ratio Lower Upper
93 100
63 84.7 68.6 102.8
95 32.1 26.6 39.8



#12
2-Chlorophenol
Concen: 21.533 ng/ul
RT: 7.205 min Scan# 700
Delta R.T. 0.000 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12

Tgt Ion: 128 Resp: 123044
Ion Ratio Lower Upper
128 100
64 59.7 51.4 77.0
130 31.4 26.5 39.7





#13

2-Methylphenol

Concen: 20.740 ng/ul

RT: 8.081 min Scan# 849

Delta R.T. 0.000 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12

Instrument :

BNA_P

ClientSampleId :





SICV615

Tgt Ion:108 Resp: 127213

Ion Ratio Lower Upper

108 100

107 91.4 67.3 100.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 21.331 ng/ul

RT: 8.158 min Scan# 862

Delta R.T. 0.000 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12




Tgt Ion: 45 Resp: 213067


Ion Ratio Lower Upper

45 100

77 16.3 12.6 18.8

79 10.8 8.8 13.2









#16
Acetophenone
Concen: 21.243 ng/ul
RT: 8.464 min Scan# 914
Delta R.T. 0.000 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12




Instrument :
BNA_P
ClientSampleId :
SICV615


Tgt Ion:105 Resp: 225768
Ion Ratio Lower Upper
105 100
77 88.5 70.3 105.5
51 31.5 25.1 37.7



#17
N-Nitroso-di-n-propylamine
Concen: 21.168 ng/ul
RT: 8.446 min Scan# 911
Delta R.T. 0.000 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12

Tgt Ion: 70 Resp: 131413
Ion Ratio Lower Upper
70 100
42 59.8 47.8 71.6
101 9.5 6.5 9.7
130 15.9 12.5 18.7









#18
4-Methylphenol
Concen: 21.026 ng/ul
RT: 8.405 min Scan# 904
Delta R.T. 0.000 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12




Instrument :
BNA_P
ClientSampleId :
SICV615


Tgt Ion:108 Resp: 137232
Ion Ratio Lower Upper
108 100
107 108.9 91.5 137.3



#19
Hexachloroethane
Concen: 21.115 ng/ul
RT: 8.693 min Scan# 953
Delta R.T. 0.000 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12

Tgt Ion:117 Resp: 59389
Ion Ratio Lower Upper
117 100
201 105.4 87.8 131.6
199 64.8 55.0 82.6









#22
Nitrobenzene
Concen: 21.347 ng/ul
RT: 8.840 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12




Instrument :
BNA_P
ClientSampleId :
SICV615


Tgt Ion: 77 Resp: 193689
Ion Ratio Lower Upper
77 100
123 36.8 28.5 42.7
65 14.7 11.7 17.5



#23
Isophorone
Concen: 21.615 ng/ul
RT: 9.358 min Scan# 1066
Delta R.T. 0.000 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12

Tgt Ion: 82 Resp: 364004
Ion Ratio Lower Upper
82 100
95 8.1 6.0 9.0
138 14.0 10.9 16.3









#25
2-Nitrophenol
Concen: 22.628 ng/ul
RT: 9.540 min Scan# 1097
Delta R.T. 0.000 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12




Instrument :
BNA_P
ClientSampleId :
SICV615


Tgt Ion:	139	Resp:	72535
Ion Ratio	Lower	Upper	
139	100		
65	66.8	58.6	87.8
109	43.8	34.8	52.2



#26
2,4-Dimethylphenol
Concen: 21.555 ng/ul
RT: 9.599 min Scan# 1107
Delta R.T. 0.000 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12

Tgt Ion:	107	Resp:	165296
Ion Ratio	Lower	Upper	
107	100		
121	43.6	34.0	51.0
122	74.2	58.9	88.3





#27

Bis(2-Chloroethoxy)methane

Concen: 21.217 ng/ul

RT: 9.840 min Scan# 1148

Delta R.T. 0.000 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12

Instrument :

BNA_P

ClientSampleId :

SICV615





Tgt Ion: 93 Resp: 200338

Ion Ratio Lower Upper

93 100

95 31.4 25.4 38.0

123 9.6 7.5 11.3



#29

2,4-Dichlorophenol

Concen: 21.524 ng/ul

RT: 10.075 min Scan# 1188

Delta R.T. 0.006 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12




Tgt Ion:162 Resp: 125129


Ion Ratio Lower Upper

162 100

164 65.6 51.7 77.5

98 40.1 32.8 49.2









#30
Naphthalene
Concen: 20.961 ng/ul
RT: 10.458 min Scan# 1253
Delta R.T. 0.000 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12




Instrument :
BNA_P
ClientSampleId :
SICV615


Tgt Ion:128 Resp: 420713
Ion Ratio Lower Upper
128 100
129 10.7 8.7 13.1
127 13.2 10.5 15.7



#32
4-Chloroaniline
Concen: 21.795 ng/ul
RT: 10.593 min Scan# 1276
Delta R.T. 0.000 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12

Tgt Ion:127 Resp: 172340
Ion Ratio Lower Upper
127 100
129 31.9 25.9 38.9









#33
Hexachlorobutadiene
Concen: 20.688 ng/ul
RT: 10.728 min Scan# 1299
Delta R.T. 0.000 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12




Instrument :
BNA_P
ClientSampleId :
SICV615


Tgt Ion:225 Resp: 98896
Ion Ratio Lower Upper
225 100
223 63.1 50.5 75.7
227 62.3 52.2 78.4



#34
Caprolactam
Concen: 22.613 ng/ul
RT: 11.405 min Scan# 1414
Delta R.T. 0.006 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12

Tgt Ion:113 Resp: 41200
Ion Ratio Lower Upper
113 100
55 186.6 140.5 210.7
56 149.6 113.8 170.6









#35
4-Chloro-3-methylphenol
Concen: 22.351 ng/ul
RT: 11.722 min Scan# 1468
Delta R.T. 0.000 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12




Instrument :
BNA_P
ClientSampleId :
SICV615


Tgt Ion:107 Resp: 155780
Ion Ratio Lower Upper
107 100
144 22.8 18.0 27.0
142 70.9 56.3 84.5



#36
2-Methylnaphthalene
Concen: 20.993 ng/ul
RT: 12.087 min Scan# 1530
Delta R.T. 0.000 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12

Tgt Ion:142 Resp: 290899
Ion Ratio Lower Upper
142 100
141 92.5 72.1 108.1





#37

1-Methylnaphthalene

Concen: 21.142 ng/ul

RT: 12.310 min Scan# 1568

Delta R.T. 0.000 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12

Instrument :

BNA_P

ClientSampleId :

SICV615





Tgt Ion:142 Resp: 290925

Ion Ratio Lower Upper

142 100

141 93.2 72.8 109.2

116 4.4 4.0 6.0



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.335 ng/ul

RT: 12.457 min Scan# 1593

Delta R.T. 0.000 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12

Tgt Ion:216 Resp: 177089




Ion Ratio Lower Upper


216 100

214 78.0 64.2 96.2

179 20.3 17.3 25.9

108 18.8 16.8 25.2









#40
Hexachlorocyclopentadiene
Concen: 18.346 ng/ul
RT: 12.422 min Scan# 1587
Delta R.T. 0.000 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12




Instrument :
BNA_P
ClientSampleId :
SICV615


Tgt Ion:237 Resp: 92040
Ion Ratio Lower Upper
237 100
235 60.8 48.2 72.4
272 11.5 9.4 14.0



#41
2,4,6-Trichlorophenol
Concen: 21.213 ng/ul
RT: 12.716 min Scan# 1637
Delta R.T. -0.006 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12

Tgt Ion:196 Resp: 110691
Ion Ratio Lower Upper
196 100
198 93.4 78.8 118.2
200 29.5 24.9 37.3





#42

2,4,5-Trichlorophenol

Concen: 21.195 ng/ul

RT: 12.799 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12

Instrument :

BNA_P

ClientSampleId :

SICV615





Tgt Ion:196 Resp: 114962

Ion Ratio Lower Upper

196 100

198 101.9 77.5 116.3

200 31.3 24.5 36.7



#43

1,1'-Biphenyl

Concen: 20.816 ng/ul

RT: 13.128 min Scan# 1707

Delta R.T. 0.000 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12




Tgt Ion:154 Resp: 398779


Ion Ratio Lower Upper

154 100

153 40.9 32.2 48.4

76 13.6 11.3 16.9









#44
2-Chloronaphthalene
Concen: 20.838 ng/ul
RT: 13.169 min Scan# 1714
Delta R.T. 0.000 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12




Instrument :
BNA_P
ClientSampleId :
SICV615


Tgt Ion:162 Resp: 318909
Ion Ratio Lower Upper
162 100
164 34.1 27.0 40.6
127 39.5 33.7 50.5



#45
2-Nitroaniline
Concen: 23.238 ng/ul
RT: 13.399 min Scan# 1753
Delta R.T. 0.000 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12

Tgt Ion: 65 Resp: 115543
Ion Ratio Lower Upper
65 100
92 62.1 49.4 74.0
138 79.4 62.2 93.4





#47

Dimethylphthalate

Concen: 22.083 ng/ul

RT: 13.781 min Scan# 1818

Delta R.T. 0.000 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12

Instrument :

BNA_P

ClientSampleId :

SICV615





Tgt Ion:163 Resp: 414114

Ion Ratio Lower Upper

163 100

194 3.9 3.3 4.9

164 10.4 8.1 12.1



#48

2,6-Dinitrotoluene

Concen: 22.710 ng/ul

RT: 13.910 min Scan# 1840

Delta R.T. 0.000 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12




Tgt Ion:165 Resp: 80363


Ion Ratio Lower Upper

165 100

89 76.9 61.0 91.4

121 24.4 18.7 28.1









#50
Acenaphthylene
Concen: 21.006 ng/ul
RT: 14.034 min Scan# 1861
Delta R.T. 0.000 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12




Instrument :
BNA_P
ClientSampleId :
SICV615


Tgt Ion:152 Resp: 515407
Ion Ratio Lower Upper
152 100
151 20.3 15.4 23.2
153 13.0 10.6 16.0



#51
3-Nitroaniline
Concen: 23.581 ng/ul
RT: 14.257 min Scan# 1899
Delta R.T. 0.006 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12

Tgt Ion:138 Resp: 78981
Ion Ratio Lower Upper
138 100
108 12.8 10.4 15.6
92 136.6 108.4 162.6





#52

Acenaphthene

Concen: 21.071 ng/ul

RT: 14.381 min Scan# 1920

Delta R.T. 0.000 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12

Instrument :

BNA_P

ClientSampleId :

SICV615





Tgt Ion:153 Resp: 340792

Ion Ratio Lower Upper

153 100

152 47.4 37.9 56.9

154 87.0 71.4 107.2



#53

2,4-Dinitrophenol

Concen: 20.406 ng/ul

RT: 14.481 min Scan# 1937

Delta R.T. 0.006 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12




Tgt Ion:184 Resp: 43422


Ion Ratio Lower Upper

184 100

63 86.3 65.0 97.6

154 72.8 55.4 83.2









#55
4-Nitrophenol
Concen: 22.692 ng/ul
RT: 14.575 min Scan# 1953
Delta R.T. 0.000 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12




Instrument :
BNA_P
ClientSampleId :
SICV615


Tgt Ion:109 Resp: 71872
Ion Ratio Lower Upper
109 100
139 93.1 77.9 116.9
65 128.8 100.9 151.3



#56
Dibenzofuran
Concen: 21.090 ng/ul
RT: 14.734 min Scan# 1980
Delta R.T. 0.006 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12

Tgt Ion:168 Resp: 470928
Ion Ratio Lower Upper
168 100
139 40.5 32.6 49.0





#57

2,4-Dinitrotoluene

Concen: 23.709 ng/ul

RT: 14.728 min Scan# 1978

Delta R.T. 0.006 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12

Instrument :

BNA_P

ClientSampleId :

SICV615





Tgt Ion:165 Resp: 118768

Ion Ratio Lower Upper

165 100

63 81.3 70.2 105.4

182 2.6 2.0 3.0



#58

2,3,4,6-Tetrachlorophenol

Concen: 22.415 ng/ul

RT: 14.969 min Scan# 2020

Delta R.T. 0.006 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12

Tgt Ion:232 Resp: 105288




Ion Ratio Lower Upper


232 100

131 42.0 34.5 51.7

130 2.6 2.4 3.6

166 28.8 25.4 38.2









#59
Diethylphthalate
Concen: 22.389 ng/ul
RT: 15.187 min Scan# 2056
Delta R.T. 0.006 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12




Instrument :
BNA_P
ClientSampleId :
SICV615


Tgt Ion:149 Resp: 421615
Ion Ratio Lower Upper
149 100
177 20.5 16.9 25.3
150 12.2 9.9 14.9



#61
Fluorene
Concen: 21.334 ng/ul
RT: 15.404 min Scan# 2094
Delta R.T. 0.006 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12

Tgt Ion:166 Resp: 393612
Ion Ratio Lower Upper
166 100
165 97.0 78.2 117.4
167 13.2 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 20.774 ng/ul
RT: 15.404 min Scan# 2094
Delta R.T. 0.006 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12




Instrument :
BNA_P
ClientSampleId :
SICV615


Tgt Ion:204 Resp: 208249
Ion Ratio Lower Upper
204 100
206 33.2 25.1 37.7
141 62.4 49.3 73.9



#63
4-Nitroaniline
Concen: 24.932 ng/ul
RT: 15.457 min Scan# 2103
Delta R.T. 0.006 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12

Tgt Ion:138 Resp: 74142
Ion Ratio Lower Upper
138 100
92 62.1 55.2 82.8
108 112.3 95.4 143.0









#66
4,6-Dinitro-2-methylphenol
Concen: 20.785 ng/ul
RT: 15.504 min Scan# 2111
Delta R.T. 0.000 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12




Instrument :
BNA_P
ClientSampleId :
SICV615


Tgt Ion:198 Resp: 74901
Ion Ratio Lower Upper
198 100
182 4.0 3.0 4.6
77 35.6 25.6 38.4



#67
N-Nitrosodiphenylamine
Concen: 20.208 ng/ul
RT: 15.628 min Scan# 2132
Delta R.T. 0.006 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12

Tgt Ion:169 Resp: 331734
Ion Ratio Lower Upper
169 100
168 64.8 55.0 82.6
167 35.4 29.5 44.3









#68
4-Bromophenyl-phenylether
Concen: 20.276 ng/ul
RT: 16.328 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12




Instrument :
BNA_P
ClientSampleId :
SICV615


Tgt Ion:248 Resp: 131968
Ion Ratio Lower Upper
248 100
250 94.2 79.2 118.8
141 81.8 65.1 97.7



#69
Hexachlorobenzene
Concen: 20.482 ng/ul
RT: 16.434 min Scan# 2269
Delta R.T. 0.006 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12

Tgt Ion:284 Resp: 144310
Ion Ratio Lower Upper
284 100
142 36.5 30.6 45.8
249 35.7 26.9 40.3





#70

Atrazine

Concen: 21.624 ng/ul

RT: 16.634 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12

Instrument :

BNA_P

ClientSampleId :

SICV615





Tgt Ion:200 Resp: 130356

Ion Ratio Lower Upper

200 100

173 25.1 19.3 28.9

215 50.1 36.6 54.8



#71

Pentachlorophenol

Concen: 20.532 ng/ul

RT: 16.810 min Scan# 2333

Delta R.T. 0.012 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12




Tgt Ion:266 Resp: 87682


Ion Ratio Lower Upper

266 100

264 63.9 50.6 75.8

268 65.3 50.3 75.5





#72

Phenanthrene

Concen: 20.774 ng/ul

RT: 17.210 min Scan# 2401

Delta R.T. 0.012 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12

Instrument :

BNA_P

ClientSampleId :

SICV615





Tgt Ion:178 Resp: 640150

Ion Ratio Lower Upper

178 100

179 15.2 12.2 18.4

176 19.1 15.5 23.3



#74

Anthracene

Concen: 20.960 ng/ul

RT: 17.298 min Scan# 2416

Delta R.T. 0.000 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12




Tgt Ion:178 Resp: 658728


Ion Ratio Lower Upper

178 100

179 15.2 12.2 18.4

176 19.0 14.6 21.8





#75

1,2,3,4-Tetrachlorobenzene

Concen: 19.092 ng/uL

RT: 13.081 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12

Instrument :

BNA_P

ClientSampleId :

SICV615





Tgt Ion:216 Resp: 179628

Ion Ratio Lower Upper

216 100

214 76.8 59.3 88.9

218 48.5 37.8 56.8



#76

Pentachlorobenzene

Concen: 19.275 ng/uL

RT: 14.640 min Scan# 1964

Delta R.T. 0.000 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12




Tgt Ion:250 Resp: 170657


Ion Ratio Lower Upper

250 100

248 64.2 51.2 76.8

252 63.7 50.9 76.3









#77
Carbazole
Concen: 21.438 ng/ul
RT: 17.604 min Scan# 2468
Delta R.T. 0.006 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12




Instrument :
BNA_P
ClientSampleId :
SICV615


Tgt Ion:167 Resp: 577331
Ion Ratio Lower Upper
167 100
166 21.5 17.4 26.2
139 12.7 10.9 16.3



#78
Di-n-butylphthalate
Concen: 22.192 ng/ul
RT: 18.204 min Scan# 2570
Delta R.T. 0.012 min
Lab File: BP020036.D
Acq: 19 Apr 2024 21:12

Tgt Ion:149 Resp: 721026
Ion Ratio Lower Upper
149 100
150 8.8 7.4 11.0
104 6.5 5.3 7.9





#80

Fluoranthene

Concen: 21.248 ng/ul

RT: 19.322 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12

Instrument :

BNA_P

ClientSampleId :

SICV615





Tgt Ion:202 Resp: 785076

Ion Ratio Lower Upper

202 100

101 9.5 7.7 11.5

100 7.1 5.7 8.5



#82

Pyrene

Concen: 21.515 ng/ul

RT: 19.704 min Scan# 2825

Delta R.T. 0.012 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12




Tgt Ion:202 Resp: 827071


Ion Ratio Lower Upper

202 100

101 11.0 9.3 13.9

100 9.3 7.6 11.4





#83

Butylbenzylphthalate

Concen: 23.046 ng/ul

RT: 20.686 min Scan# 2990

Delta R.T. 0.012 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12

Instrument :

BNA_P

ClientSampleId :

SICV615





Tgt Ion:149 Resp: 337795

Ion Ratio Lower Upper

149 100

91 78.8 69.6 104.4

206 21.6 17.4 26.0



#84

3,3'-Dichlorobenzidine

Concen: 21.637 ng/ul

RT: 21.580 min Scan# 3144

Delta R.T. 0.012 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12




Tgt Ion:252 Resp: 279793


Ion Ratio Lower Upper

252 100

254 62.6 51.5 77.3

126 9.5 7.9 11.9





#85

Benzo(a)anthracene

Concen: 21.264 ng/ul

RT: 21.651 min Scan# 3156

Delta R.T. 0.018 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12

Instrument :

BNA_P

ClientSampleId :

SICV615





Tgt Ion:228 Resp: 860155

Ion Ratio Lower Upper

228 100

229 19.2 15.7 23.5

226 26.9 21.4 32.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.616 ng/ul

RT: 21.598 min Scan# 3147

Delta R.T. 0.006 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12




Tgt Ion:149 Resp: 488413


Ion Ratio Lower Upper

149 100

167 27.0 21.3 31.9

279 4.2 3.2 4.8





#87

Chrysene

Concen: 20.786 ng/ul

RT: 21.716 min Scan# 3166

Delta R.T. 0.006 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12

Instrument :

BNA_P

ClientSampleId :

SICV615





Tgt Ion:228 Resp: 791853

Ion Ratio Lower Upper

228 100

226 29.8 24.2 36.4

229 19.5 15.4 23.2



#89

Di-n-octyl phthalate

Concen: 23.832 ng/ul




RT: 22.904 min Scan# 3369


Delta R.T. 0.006 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12

Tgt Ion:149 Resp: 939670





#90

Benzo(b)fluoranthene

Concen: 25.301 ng/ul

RT: 23.980 min Scan# 3552

Delta R.T. 0.006 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12

Instrument :

BNA_P

ClientSampleId :

SICV615





Tgt Ion:252 Resp: 972761

Ion Ratio Lower Upper

252 100

253 22.0 17.8 26.8

125 7.1 5.8 8.6



#91

Benzo(k)fluoranthene

Concen: 25.628 ng/ul

RT: 24.051 min Scan# 3564

Delta R.T. 0.006 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12




Tgt Ion:252 Resp: 1010148


Ion Ratio Lower Upper

252 100

253 21.6 17.5 26.3

125 6.9 5.9 8.9





#93

Benzo(a)pyrene

Concen: 20.575 ng/ul

RT: 24.886 min Scan# 31

Delta R.T. 0.000 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12

Instrument :

BNA_P

ClientSampleId :

SICV615





Tgt Ion:252 Resp: 762283

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.0

125 9.5 7.0 10.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 9.530 ng/ul

RT: 28.892 min Scan# 4387

Delta R.T. -0.012 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12




Tgt Ion:276 Resp: 409432


Ion Ratio Lower Upper

276 100

138 14.4 11.7 17.5

277 23.6 18.6 28.0





#95

Dibenzo(a,h)anthracene

Concen: 20.930 ng/ul

RT: 28.992 min Scan# 4405

Delta R.T. -0.006 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12

Instrument :

BNA_P

ClientSampleId :

SICV615





Tgt Ion:278 Resp: 732654

Ion Ratio Lower Upper

278 100

139 12.3 9.4 14.2

279 25.2 19.0 28.6



#96

Benzo(g,h,i)perylene

Concen: 10.376 ng/ul

RT: 30.068 min Scan# 4587

Delta R.T. -0.029 min

Lab File: BP020036.D

Acq: 19 Apr 2024 21:12

Tgt Ion:276 Resp: 356558

Ion Ratio Lower Upper

276 100

138 15.8 13.5 20.3

277 24.2 19.4 29.2

