

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042122\
 Data File : BP010007.D
 Acq On : 21 Apr 2022 10:31
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020766

Quant Time: Apr 21 14:00:02 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 13:57:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.940	152	163486	20.000	ng/ul	0.00
20) Naphthalene-d8	10.746	136	981498	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.575	164	792150	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.328	188	1800031	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.410	240	1865034	20.000	ng/ul	0.00
88) Perylene-d12	23.863	264	1695497	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.464	96	48	0.013	ng/uL	0.09
4) Pyridine-d5	3.787	84	229436	22.717	ng/ul	0.00
7) Phenol-d5	7.105	99	370440	25.809	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.269	67	218899	25.618	ng/ul	0.00
11) 2-Chlorophenol-d4	7.475	132	264807	24.459	ng/ul	0.00
15) 4-Methylphenol-d8	8.652	113	341149	28.472	ng/ul	0.00
21) Nitrobenzene-d5	9.099	128	159380	22.518	ng/ul	0.00
24) 2-Nitrophenol-d4	9.828	143	188855	24.338	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.369	165	360780	22.218	ng/ul	0.00
31) 4-Chloroaniline-d4	10.881	131	513460	22.201	ng/ul	0.00
46) Dimethylphthalate-d6	13.981	166	1236140	19.866	ng/ul	0.00
49) Acenaphthylene-d8	14.269	160	1489228	20.202	ng/ul	0.00
54) 4-Nitrophenol-d4	14.769	143	230881	20.460	ng/ul	0.00
60) Fluorene-d10	15.569	176	1069865	20.439	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.687	200	220837	20.710	ng/ul	0.00
73) Anthracene-d10	17.428	188	1747574	20.277	ng/ul	0.00
81) Pyrene-d10	19.657	212	2124859	20.861	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.704	264	1797370	20.356	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.405	88	28123	7.531	ng/uL#	21
5) Pyridine	3.805	79	238827	22.664	ng/ul#	37
6) Benzaldehyde	7.075	77	190218	21.610	ng/ul	97
8) Phenol	7.128	94	370853	25.813	ng/ul#	89
10) Bis(2-Chloroethyl)ether	7.358	93	285385	25.501	ng/ul#	77
12) 2-Chlorophenol	7.505	128	265574	24.331	ng/ul	93
13) 2-Methylphenol	8.381	108	304221	27.155	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.469	45	453133	27.392	ng/ul#	83
16) Acetophenone	8.763	105	531357	28.147	ng/ul#	80
17) N-Nitroso-di-n-propyla...	8.752	70	294271	30.067	ng/ul#	72
18) 4-Methylphenol	8.711	108	346067	28.525	ng/ul	91
19) Hexachloroethane	9.028	117	101470	22.018	ng/ul#	81
22) Nitrobenzene	9.146	77	390781	21.510	ng/ul#	85
23) Isophorone	9.675	82	841124	22.468	ng/ul	98
25) 2-Nitrophenol	9.858	139	192580	23.269	ng/ul#	85
26) 2,4-Dimethylphenol	9.922	107	424837	21.631	ng/ul#	80
27) Bis(2-Chloroethoxy)met...	10.152	93	477902	21.744	ng/ul#	97
29) 2,4-Dichlorophenol	10.393	162	342960	21.786	ng/ul#	85
30) Naphthalene	10.799	128	1034616	20.496	ng/ul	97
32) 4-Chloroaniline	10.905	127	509232	22.250	ng/ul	96
33) Hexachlorobutadiene	11.093	225	217713	19.202	ng/ul	97
34) Caprolactam	11.675	113	69567	13.363	ng/ul#	1
35) 4-Chloro-3-methylphenol	12.028	107	443510	23.251	ng/ul#	71
36) 2-Methylnaphthalene	12.404	142	807604	21.899	ng/ul	93

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042122\
 Data File : BP010007.D
 Acq On : 21 Apr 2022 10:31
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020766

Quant Time: Apr 21 14:00:02 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 13:57:49 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.622	142	815682	21.906	ng/ul#	92
39) 1,2,4,5-Tetrachloroben...	12.775	216	468264	19.321	ng/ul	93
40) Hexachlorocyclopentadiene	12.757	237	234452	13.940	ng/ul	95
41) 2,4,6-Trichlorophenol	13.010	196	320709	20.721	ng/ul#	80
42) 2,4,5-Trichlorophenol	13.081	196	352503	21.075	ng/ul#	86
43) 1,1'-Biphenyl	13.410	154	1151620	19.925	ng/ul	93
44) 2-Chloronaphthalene	13.451	162	888215	19.992	ng/ul	93
45) 2-Nitroaniline	13.651	65	316018	24.272	ng/ul#	83
47) Dimethylphthalate	14.028	163	1198897	20.019	ng/ul#	92
48) 2,6-Dinitrotoluene	14.146	165	258857	23.498	ng/ul	95
50) Acenaphthylene	14.298	152	1468830	20.190	ng/ul	95
51) 3-Nitroaniline	14.475	138	212309	18.539	ng/ul#	78
52) Acenaphthene	14.640	153	963459	20.227	ng/ul	98
53) 2,4-Dinitrophenol	14.681	184	144419	21.689	ng/ul#	82
55) 4-Nitrophenol	14.781	109	207920	20.015	ng/ul#	49
56) Dibenzofuran	14.975	168	1420001	20.256	ng/ul	98
57) 2,4-Dinitrotoluene	14.934	165	383142	23.686	ng/ul#	85
58) 2,3,4,6-Tetrachlorophenol	15.198	232	316098	20.727	ng/ul#	86
59) Diethylphthalate	15.393	149	1248873	20.149	ng/ul	93
61) Fluorene	15.622	166	1170674	20.274	ng/ul	91
62) 4-Chlorophenyl-phenyle...	15.616	204	611520	19.917	ng/ul	96
63) 4-Nitroaniline	15.640	138	210573	18.358	ng/ul#	80
66) 4,6-Dinitro-2-methylph...	15.698	198	215078	20.291	ng/ul#	91
67) N-Nitrosodiphenylamine	15.828	169	1041829	20.219	ng/ul	95
68) 4-Bromophenyl-phenylether	16.510	248	384209	19.719	ng/ul	97
69) Hexachlorobenzene	16.634	284	442058	19.730	ng/ul#	89
70) Atrazine	16.787	200	365961	17.207	ng/ul#	91
71) Pentachlorophenol	16.975	266	292226	19.269	ng/ul#	86
72) Phenanthrene	17.369	178	1955006	20.347	ng/ul	99
74) Anthracene	17.463	178	1968719	20.210	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.375	216	512140	19.679	ng/ul#	88
76) Pentachlorobenzene	14.898	250	500088	19.714	ng/ul	91
77) Carbazole	17.728	167	1754424	19.929	ng/ul#	95
78) Di-n-butylphthalate	18.281	149	2231094	20.610	ng/ul#	93
80) Fluoranthene	19.334	202	2407683	20.536	ng/ul#	95
82) Pyrene	19.686	202	2479935	20.862	ng/ul#	93
83) Butylbenzylphthalate	20.551	149	1078213	22.111	ng/ul#	81
84) 3,3'-Dichlorobenzidine	21.322	252	609882	15.346	ng/ul#	95
85) Benzo(a)anthracene	21.392	228	2425465	20.514	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.316	149	1578557	21.521	ng/ul#	82
87) Chrysene	21.445	228	2332643	20.448	ng/ul	99
89) Di-n-octyl phthalate	22.257	149	2713656	22.844	ng/ul	100
90) Benzo(b)fluoranthene	23.116	252	2331075	20.693	ng/ul	99
91) Benzo(k)fluoranthene	23.163	252	2128752	20.709	ng/ul#	98
93) Benzo(a)pyrene	23.757	252	2119371	20.297	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.433	276	2127362	19.483	ng/ul#	94
95) Dibenzo(a,h)anthracene	26.445	278	1819983	19.222	ng/ul#	94
96) Benzo(g,h,i)perylene	27.215	276	1779213	19.762	ng/ul#	91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042122\
 Data File : BP010007.D
 Acq On : 21 Apr 2022 10:31
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020766

Quant Time: Apr 21 14:00:02 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 13:57:49 2022
 Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 7.531 ng/uL
 RT: 3.405 min Scan# 54
 Delta R.T. 0.000 min
 Lab File: BP010007.D
 Acq: 21 Apr 2022 10:31




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020766


Tgt Ion: 88 Resp: 28123
 Ion Ratio Lower Upper
 88 100
 43 35.7 6.8 10.2#
 58 82.3 28.4 42.6#



#5
 Pyridine
 Concen: 22.664 ng/uL
 RT: 3.805 min Scan# 122
 Delta R.T. 0.000 min
 Lab File: BP010007.D
 Acq: 21 Apr 2022 10:31

Tgt Ion: 79 Resp: 238827
 Ion Ratio Lower Upper
 79 100
 52 94.1 35.5 53.3#
 51 44.7 20.2 30.4#









#6
Benzaldehyde
Concen: 21.610 ng/ul
RT: 7.075 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion: 77 Resp: 190218
Ion Ratio Lower Upper
77 100
105 96.4 77.0 115.4
106 90.0 77.0 115.4



#8
Phenol
Concen: 25.813 ng/ul
RT: 7.128 min Scan# 687
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31

Tgt Ion: 94 Resp: 370853
Ion Ratio Lower Upper
94 100
65 33.4 19.0 28.6#
66 47.1 34.6 51.8









#10
Bis(2-Chloroethyl)ether
Concen: 25.501 ng/ul
RT: 7.358 min Scan# 71
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion: 93 Resp: 285385
Ion Ratio Lower Upper
93 100
63 81.9 45.0 67.4#
95 31.6 26.3 39.5



#12
2-Chlorophenol
Concen: 24.331 ng/ul
RT: 7.505 min Scan# 751
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31

Tgt Ion:128 Resp: 265574
Ion Ratio Lower Upper
128 100
64 52.2 45.3 67.9
130 31.7 29.0 43.6





#13

2-Methylphenol

Concen: 27.155 ng/ul

RT: 8.381 min Scan# 900

Delta R.T. 0.000 min

Lab File: BP010007.D

Acq: 21 Apr 2022 10:31

Instrument :

BNA_P

ClientSampleId :





SSTD020766

Tgt Ion:108 Resp: 304221

Ion Ratio Lower Upper

108 100

107 89.9 70.8 106.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 27.392 ng/ul

RT: 8.469 min Scan# 915

Delta R.T. 0.000 min

Lab File: BP010007.D

Acq: 21 Apr 2022 10:31




Tgt Ion: 45 Resp: 453133


Ion Ratio Lower Upper

45 100

77 13.0 17.1 25.7#

79 9.6 13.2 19.8#









#16
Acetophenone
Concen: 28.147 ng/ul
RT: 8.763 min Scan# 965
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:105 Resp: 531357
Ion Ratio Lower Upper
105 100
77 80.3 53.4 80.2#
51 31.5 13.8 20.8#



#17
N-Nitroso-di-n-propylamine
Concen: 30.067 ng/ul
RT: 8.752 min Scan# 963
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31

Tgt Ion: 70 Resp: 294271
Ion Ratio Lower Upper
70 100
42 64.6 31.8 47.8#
101 9.8 15.6 23.4#
130 19.3 13.7 20.5









#18
4-Methylphenol
Concen: 28.525 ng/ul
RT: 8.711 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:108 Resp: 346067
Ion Ratio Lower Upper
108 100
107 113.2 83.6 125.4



#19
Hexachloroethane
Concen: 22.018 ng/ul
RT: 9.028 min Scan# 1010
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31

Tgt Ion:117 Resp: 101470
Ion Ratio Lower Upper
117 100
201 104.8 106.5 159.7#
199 66.3 59.0 88.6









#22
Nitrobenzene
Concen: 21.510 ng/ul
RT: 9.146 min Scan# 110
Delta R.T. 0.006 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion: 77 Resp: 390781
Ion Ratio Lower Upper
77 100
123 42.8 42.3 63.5
65 13.7 5.8 8.8#



#23
Isophorone
Concen: 22.468 ng/ul
RT: 9.675 min Scan# 1120
Delta R.T. 0.006 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31

Tgt Ion: 82 Resp: 841124
Ion Ratio Lower Upper
82 100
95 8.3 7.8 11.8
138 16.8 12.9 19.3









#25
2-Nitrophenol
Concen: 23.269 ng/ul
RT: 9.858 min Scan# 1151
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:139 Resp: 192580
Ion Ratio Lower Upper
139 100
65 59.4 49.4 74.2
109 35.3 44.7 67.1#



#26
2,4-Dimethylphenol
Concen: 21.631 ng/ul
RT: 9.922 min Scan# 1162
Delta R.T. 0.006 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31

Tgt Ion:107 Resp: 424837
Ion Ratio Lower Upper
107 100
121 48.3 34.1 51.1
122 78.6 46.3 69.5#









#27
 Bis(2-Chloroethoxy)methane
 Concen: 21.744 ng/ul
 RT: 10.152 min Scan# 101
 Delta R.T. 0.000 min
 Lab File: BP010007.D
 Acq: 21 Apr 2022 10:31




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020766


Tgt Ion: 93 Resp: 477902
 Ion Ratio Lower Upper
 93 100
 95 32.2 26.2 39.4
 123 10.8 6.2 9.4#



#29
 2,4-Dichlorophenol
 Concen: 21.786 ng/ul
 RT: 10.393 min Scan# 1242
 Delta R.T. 0.000 min
 Lab File: BP010007.D
 Acq: 21 Apr 2022 10:31

Tgt Ion:162 Resp: 342960
 Ion Ratio Lower Upper
 162 100
 164 63.4 37.0 55.6#
 98 37.9 31.0 46.6









#30
Naphthalene
Concen: 20.496 ng/ul
RT: 10.799 min Scan# 1311
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:128 Resp: 1034616
Ion Ratio Lower Upper
128 100
129 10.7 8.6 13.0
127 13.3 12.2 18.4



#32
4-Chloroaniline
Concen: 22.250 ng/ul
RT: 10.905 min Scan# 1329
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31

Tgt Ion:127 Resp: 509232
Ion Ratio Lower Upper
127 100
129 32.9 24.6 37.0









#33
Hexachlorobutadiene
Concen: 19.202 ng/ul
RT: 11.093 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:225 Resp: 217713
Ion Ratio Lower Upper
225 100
223 63.5 53.4 80.0
227 65.6 53.8 80.8



#34
Caprolactam
Concen: 13.363 ng/ul
RT: 11.675 min Scan# 1460
Delta R.T. 0.012 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31

Tgt Ion:113 Resp: 69567
Ion Ratio Lower Upper
113 100
55 202.0 58.2 87.2#
56 147.8 68.6 103.0#









#35
4-Chloro-3-methylphenol
Concen: 23.251 ng/ul
RT: 12.028 min Scan# 1519
Delta R.T. 0.006 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:107 Resp: 443510
Ion Ratio Lower Upper
107 100
144 24.6 15.8 23.8#
142 73.9 38.9 58.3#



#36
2-Methylnaphthalene
Concen: 21.899 ng/ul
RT: 12.404 min Scan# 1584
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31

Tgt Ion:142 Resp: 807604
Ion Ratio Lower Upper
142 100
141 86.9 64.3 96.5









#37
1-Methylnaphthalene
Concen: 21.906 ng/ul
RT: 12.622 min Scan# 1621
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:142 Resp: 815682
Ion Ratio Lower Upper
142 100
141 94.2 69.8 104.8
116 4.1 5.6 8.4#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 19.321 ng/ul
RT: 12.775 min Scan# 1647
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31

Tgt Ion:216 Resp: 468264
Ion Ratio Lower Upper
216 100
214 78.3 58.2 87.2
179 21.1 14.2 21.2
108 16.5 15.3 22.9









#40
Hexachlorocyclopentadiene
Concen: 13.940 ng/ul
RT: 12.757 min Scan# 1644
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:237 Resp: 234452
Ion Ratio Lower Upper
237 100
235 62.0 53.2 79.8
272 13.2 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 20.721 ng/ul
RT: 13.010 min Scan# 1687
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31

Tgt Ion:196 Resp: 320709
Ion Ratio Lower Upper
196 100
198 99.1 64.1 96.1#
200 31.7 19.9 29.9#









#42
2,4,5-Trichlorophenol
Concen: 21.075 ng/ul
RT: 13.081 min Scan# 1699
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:196 Resp: 352503
Ion Ratio Lower Upper
196 100
198 96.7 68.6 103.0
200 30.4 32.8 49.2#



#43
1,1'-Biphenyl
Concen: 19.925 ng/ul
RT: 13.410 min Scan# 1755
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31

Tgt Ion:154 Resp: 1151620
Ion Ratio Lower Upper
154 100
153 40.3 28.6 43.0
76 12.2 8.5 12.7









#44
2-Chloronaphthalene
Concen: 19.992 ng/ul
RT: 13.451 min Scan# 1762
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:162 Resp: 888215
Ion Ratio Lower Upper
162 100
164 32.3 27.5 41.3
127 40.4 37.1 55.7



#45
2-Nitroaniline
Concen: 24.272 ng/ul
RT: 13.651 min Scan# 1796
Delta R.T. 0.006 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31

Tgt Ion: 65 Resp: 316018
Ion Ratio Lower Upper
65 100
92 65.8 77.9 116.9#
138 98.5 81.5 122.3









#47
Dimethylphthalate
Concen: 20.019 ng/ul
RT: 14.028 min Scan# 1860
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:163 Resp: 1198897
Ion Ratio Lower Upper
163 100
194 4.2 4.1 6.1
164 10.0 11.3 16.9#



#48
2,6-Dinitrotoluene
Concen: 23.498 ng/ul
RT: 14.146 min Scan# 1880
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31

Tgt Ion:165 Resp: 258857
Ion Ratio Lower Upper
165 100
89 60.7 51.7 77.5
121 20.1 18.3 27.5









#50
Acenaphthylene
Concen: 20.190 ng/ul
RT: 14.298 min Scan# 1905
Delta R.T. 0.006 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:152 Resp: 1468830
Ion Ratio Lower Upper
152 100
151 19.5 18.3 27.5
153 13.0 11.0 16.6



#51
3-Nitroaniline
Concen: 18.539 ng/ul
RT: 14.475 min Scan# 1936
Delta R.T. 0.006 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31

Tgt Ion:138 Resp: 212309
Ion Ratio Lower Upper
138 100
108 11.2 15.8 23.8#
92 118.6 117.1 175.7









#52
Acenaphthene
Concen: 20.227 ng/ul
RT: 14.640 min Scan# 1964
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:153 Resp: 963459
Ion Ratio Lower Upper
153 100
152 46.6 34.3 51.5
154 88.3 70.0 105.0



#53
2,4-Dinitrophenol
Concen: 21.689 ng/ul
RT: 14.681 min Scan# 1971
Delta R.T. 0.006 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31

Tgt Ion:184 Resp: 144419
Ion Ratio Lower Upper
184 100
63 68.9 34.6 52.0#
154 67.6 57.7 86.5









#55
4-Nitrophenol
Concen: 20.015 ng/ul
RT: 14.781 min Scan# 1988
Delta R.T. 0.006 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:109 Resp: 207920
Ion Ratio Lower Upper
109 100
139 113.6 61.4 92.2#
65 134.3 65.3 97.9#



#56
Dibenzofuran
Concen: 20.256 ng/ul
RT: 14.975 min Scan# 2021
Delta R.T. 0.006 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31

Tgt Ion:168 Resp: 1420001
Ion Ratio Lower Upper
168 100
139 38.9 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 23.686 ng/ul
RT: 14.934 min Scan# 2013
Delta R.T. 0.006 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:165 Resp: 383142
Ion Ratio Lower Upper
165 100
63 46.4 29.5 44.3#
182 3.3 2.8 4.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.727 ng/ul
RT: 15.198 min Scan# 2059
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31

Tgt Ion:232 Resp: 316098
Ion Ratio Lower Upper
232 100
131 45.4 25.4 38.2#
130 2.3 4.2 6.2#
166 31.4 28.1 42.1









#59
Diethylphthalate
Concen: 20.149 ng/ul
RT: 15.393 min Scan# 2092
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:149 Resp: 1248873
Ion Ratio Lower Upper
149 100
177 22.5 22.0 33.0
150 12.7 10.2 15.2



#61
Fluorene
Concen: 20.274 ng/ul
RT: 15.622 min Scan# 2131
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31

Tgt Ion:166 Resp: 1170674
Ion Ratio Lower Upper
166 100
165 96.7 70.3 105.5
167 13.3 9.5 14.3









#62
4-Chlorophenyl-phenylether
Concen: 19.917 ng/ul
RT: 15.616 min Scan# 2130
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:204 Resp: 611520
Ion Ratio Lower Upper
204 100
206 32.8 27.6 41.4
141 61.8 46.6 69.8



#63
4-Nitroaniline
Concen: 18.358 ng/ul
RT: 15.640 min Scan# 2134
Delta R.T. 0.006 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31

Tgt Ion:138 Resp: 210573
Ion Ratio Lower Upper
138 100
92 52.6 41.6 62.4
108 90.4 97.6 146.4#









#66
4,6-Dinitro-2-methylphenol
Concen: 20.291 ng/ul
RT: 15.698 min Scan# 2144
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:198 Resp: 215078
Ion Ratio Lower Upper
198 100
182 3.9 6.0 9.0#
77 26.3 17.7 26.5



#67
N-Nitrosodiphenylamine
Concen: 20.219 ng/ul
RT: 15.828 min Scan# 2166
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31

Tgt Ion:169 Resp: 1041829
Ion Ratio Lower Upper
169 100
168 64.9 48.3 72.5
167 35.3 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 19.719 ng/ul
RT: 16.510 min Scan# 2282
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:248 Resp: 384209
Ion Ratio Lower Upper
248 100
250 98.9 79.4 119.0
141 78.9 67.2 100.8



#69
Hexachlorobenzene
Concen: 19.730 ng/ul
RT: 16.634 min Scan# 2303
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31

Tgt Ion:284 Resp: 442058
Ion Ratio Lower Upper
284 100
142 34.6 19.2 28.8#
249 32.7 27.4 41.2









#70
Atrazine
Concen: 17.207 ng/ul
RT: 16.787 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:200 Resp: 365961
Ion Ratio Lower Upper
200 100
173 25.2 19.2 28.8
215 48.6 32.2 48.4#



#71
Pentachlorophenol
Concen: 19.269 ng/ul
RT: 16.975 min Scan# 2361
Delta R.T. 0.000 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31

Tgt Ion:266 Resp: 292226
Ion Ratio Lower Upper
266 100
264 64.9 57.3 85.9
268 65.4 65.7 98.5#





#72

Phenanthrene

Concen: 20.347 ng/ul

RT: 17.369 min Scan# 2428

Delta R.T. 0.000 min

Lab File: BP010007.D

Acq: 21 Apr 2022 10:31

Instrument :

BNA_P

ClientSampleId :

SSTD020766





Tgt Ion:178 Resp: 1955006

Ion Ratio Lower Upper

178 100

179 15.3 11.7 17.5

176 18.3 14.5 21.7



#74

Anthracene

Concen: 20.210 ng/ul

RT: 17.463 min Scan# 2444

Delta R.T. 0.006 min

Lab File: BP010007.D

Acq: 21 Apr 2022 10:31




Tgt Ion:178 Resp: 1968719


Ion Ratio Lower Upper

178 100

179 15.4 12.3 18.5

176 18.1 13.4 20.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 19.679 ng/uL

RT: 13.375 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP010007.D

Acq: 21 Apr 2022 10:31

Instrument :

BNA_P

ClientSampleId :

SSTD020766





Tgt Ion:216 Resp: 512140

Ion Ratio Lower Upper

216 100

214 75.6 50.4 75.6#

218 47.5 35.2 52.8



#76

Pentachlorobenzene

Concen: 19.714 ng/uL

RT: 14.898 min Scan# 2008

Delta R.T. 0.000 min

Lab File: BP010007.D

Acq: 21 Apr 2022 10:31




Tgt Ion:250 Resp: 500088


Ion Ratio Lower Upper

250 100

248 61.5 43.5 65.3

252 64.5 46.5 69.7





#77

Carbazole

Concen: 19.929 ng/ul

RT: 17.728 min Scan# 2488

Delta R.T. 0.006 min

Lab File: BP010007.D

Acq: 21 Apr 2022 10:31

Instrument :





BNA_P

ClientSampleId :

SSTD020766

Tgt Ion:167 Resp: 1754424

Ion	Ratio	Lower	Upper
167	100		
166	21.4	15.8	23.8
139	12.7	8.3	12.5#



#78

Di-n-butylphthalate

Concen: 20.610 ng/ul

RT: 18.281 min Scan# 2583




Delta R.T. 0.006 min


Lab File: BP010007.D

Acq: 21 Apr 2022 10:31

Tgt Ion:149 Resp: 2231094

Ion	Ratio	Lower	Upper
149	100		
150	9.0	9.6	14.4#
104	5.3	6.2	9.4#









#80
Fluoranthene
Concen: 20.536 ng/u1
RT: 19.334 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:202 Resp: 2407683
Ion Ratio Lower Upper
202 100
101 9.3 9.6 14.4#
100 7.3 6.3 9.5



#82
Pyrene
Concen: 20.862 ng/u1
RT: 19.686 min Scan# 2822
Delta R.T. 0.006 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31

Tgt Ion:202 Resp: 2479935
Ion Ratio Lower Upper
202 100
101 10.3 10.6 15.8#
100 9.1 9.8 14.6#





#83

Butylbenzylphthalate

Concen: 22.111 ng/ul

RT: 20.551 min Scan# 2969

Delta R.T. 0.000 min

Lab File: BP010007.D

Acq: 21 Apr 2022 10:31

Instrument :

BNA_P

ClientSampleId :

SSTD020766





Tgt Ion:149 Resp: 1078213

Ion Ratio Lower Upper

149 100

91 71.8 71.1 106.7

206 23.0 28.8 43.2#



#84

3,3'-Dichlorobenzidine

Concen: 15.346 ng/ul

RT: 21.322 min Scan# 3100

Delta R.T. 0.000 min

Lab File: BP010007.D

Acq: 21 Apr 2022 10:31




Tgt Ion:252 Resp: 609882


Ion Ratio Lower Upper

252 100

254 65.2 49.8 74.6

126 9.6 4.3 6.5#





#85

Benzo(a)anthracene

Concen: 20.514 ng/ul

RT: 21.392 min Scan# 3112

Delta R.T. 0.000 min

Lab File: BP010007.D

Acq: 21 Apr 2022 10:31

Instrument :

BNA_P

ClientSampleId :

SSTD020766





Tgt Ion:228 Resp: 2425465

Ion Ratio Lower Upper

228 100

229 19.5 18.2 27.2

226 26.1 22.8 34.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.521 ng/ul

RT: 21.316 min Scan# 3099

Delta R.T. 0.000 min

Lab File: BP010007.D

Acq: 21 Apr 2022 10:31




Tgt Ion:149 Resp: 1578557


Ion Ratio Lower Upper

149 100

167 28.6 32.8 49.2#

279 5.2 5.6 8.4#





#87

Chrysene

Concen: 20.448 ng/ul

RT: 21.445 min Scan# 3121

Delta R.T. 0.000 min

Lab File: BP010007.D

Acq: 21 Apr 2022 10:31

Instrument :

BNA_P

ClientSampleId :

SSTD020766





Tgt Ion:228 Resp: 2332643

Ion Ratio Lower Upper

228 100

226 29.2 22.7 34.1

229 19.5 15.2 22.8



#89

Di-n-octyl phthalate

Concen: 22.844 ng/ul




RT: 22.257 min Scan# 3259


Delta R.T. 0.000 min

Lab File: BP010007.D

Acq: 21 Apr 2022 10:31

Tgt Ion:149 Resp: 2713656









#90
Benzo(b)fluoranthene
Concen: 20.693 ng/ul
RT: 23.116 min Scan# 3404
Delta R.T. 0.006 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:252 Resp: 2331075
Ion Ratio Lower Upper
252 100
253 21.9 17.4 26.0
125 8.2 7.3 10.9



#91
Benzo(k)fluoranthene
Concen: 20.709 ng/ul
RT: 23.163 min Scan# 3413
Delta R.T. 0.006 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31

Tgt Ion:252 Resp: 2128752
Ion Ratio Lower Upper
252 100
253 21.8 17.9 26.9
125 8.4 5.1 7.7#





#93

Benzo(a)pyrene

Concen: 20.297 ng/ul

RT: 23.757 min Scan# 3

Delta R.T. 0.006 min

Lab File: BP010007.D

Acq: 21 Apr 2022 10:31

Instrument :

BNA_P

ClientSampleId :

SSTD020766





Tgt Ion:252 Resp: 2119371

Ion Ratio Lower Upper

252 100

253 21.8 17.5 26.3

125 9.1 6.4 9.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.483 ng/ul

RT: 26.433 min Scan# 3969

Delta R.T. 0.006 min

Lab File: BP010007.D

Acq: 21 Apr 2022 10:31




Tgt Ion:276 Resp: 2127362


Ion Ratio Lower Upper

276 100

138 20.4 12.3 18.5#

277 25.1 19.3 28.9









#95
Dibenzo(a,h)anthracene
Concen: 19.222 ng/ul
RT: 26.445 min Scan# 3970
Delta R.T. 0.006 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31

Instrument :
BNA_P
ClientSampleId :
SSTD020766

Tgt Ion:278 Resp: 1819983
Ion Ratio Lower Upper
278 100
139 13.0 6.6 10.0#
279 24.4 18.4 27.6



#96
Benzo(g,h,i)perylene
Concen: 19.762 ng/ul
RT: 27.215 min Scan# 4102
Delta R.T. 0.006 min
Lab File: BP010007.D
Acq: 21 Apr 2022 10:31

Tgt Ion:276 Resp: 1779213
Ion Ratio Lower Upper
276 100
138 17.9 11.6 17.4#
277 23.9 15.9 23.9

