

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042122\
 Data File : BP010012.D
 Acq On : 21 Apr 2022 13:19
 Operator : CG/JU
 Sample : SST08027
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST080627

Quant Time: Apr 21 14:25:12 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 13:57:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.940	152	154486	20.000	ng/ul	0.00
20) Naphthalene-d8	10.752	136	685572	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.575	164	475900	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.328	188	1041784	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.410	240	1054480	20.000	ng/ul	# 0.00
88) Perylene-d12	23.868	264	1004342	20.000	ng/ul	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.370	96	110180	31.830	ng/uL	0.00
4) Pyridine-d5	3.781	84	815666	85.465	ng/ul	0.00
7) Phenol-d5	7.111	99	1172799	86.471	ng/ul	0.01
9) Bis-(2-Chloroethyl)eth...	7.269	67	675105	83.610	ng/ul	0.00
11) 2-Chlorophenol-d4	7.475	132	885563	86.559	ng/ul	0.00
15) 4-Methylphenol-d8	8.658	113	955194	84.365	ng/ul	0.00
21) Nitrobenzene-d5	9.105	128	464332	93.921	ng/ul	0.00
24) 2-Nitrophenol-d4	9.828	143	535149	98.732	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.369	165	971376	85.643	ng/ul	0.00
31) 4-Chloroaniline-d4	10.887	131	1333637	82.556	ng/ul	0.00
46) Dimethylphthalate-d6	13.987	166	2948900	78.884	ng/ul	0.00
49) Acenaphthylene-d8	14.269	160	3638956	82.166	ng/ul	0.00
54) 4-Nitrophenol-d4	14.781	143	572702	84.475	ng/ul	0.02
60) Fluorene-d10	15.575	176	2533727	80.570	ng/ul	0.01
65) 4,6-Dinitro-2-methylph...	15.692	200	578050	93.664	ng/ul	0.01
73) Anthracene-d10	17.434	188	4090637	82.010	ng/ul	0.01
81) Pyrene-d10	19.663	212	4879640	84.730	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.716	264	4364490	83.447	ng/ul	0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.405	88	113406	32.138	ng/uL#	22
5) Pyridine	3.805	79	843081	84.666	ng/ul#	46
6) Benzaldehyde	7.075	77	420177	50.515	ng/ul	98
8) Phenol	7.134	94	1177662	86.746	ng/ul#	93
10) Bis(2-Chloroethyl)ether	7.363	93	878095	83.034	ng/ul#	77
12) 2-Chlorophenol	7.511	128	892660	86.546	ng/ul	92
13) 2-Methylphenol	8.387	108	890215	84.090	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.475	45	1271827	81.362	ng/ul#	85
16) Acetophenone	8.769	105	1452656	81.434	ng/ul#	81
17) N-Nitroso-di-n-propyla...	8.769	70	763656	82.572	ng/ul#	75
18) 4-Methylphenol	8.728	108	971245	84.719	ng/ul	94
19) Hexachloroethane	9.028	117	358066	82.224	ng/ul#	81
22) Nitrobenzene	9.152	77	1103124	86.930	ng/ul#	86
23) Isophorone	9.687	82	2174506	83.158	ng/ul#	97
25) 2-Nitrophenol	9.863	139	544442	94.179	ng/ul#	84
26) 2,4-Dimethylphenol	9.922	107	1126946	82.147	ng/ul#	78
27) Bis(2-Chloroethoxy)met...	10.157	93	1230808	80.171	ng/ul#	98
29) 2,4-Dichlorophenol	10.399	162	923752	84.009	ng/ul#	83
30) Naphthalene	10.799	128	2879507	81.668	ng/ul	97
32) 4-Chloroaniline	10.910	127	1314410	82.221	ng/ul	97
33) Hexachlorobutadiene	11.093	225	631589	79.752	ng/ul	95
34) Caprolactam	11.704	113	172287	47.378	ng/ul#	10
35) 4-Chloro-3-methylphenol	12.040	107	1098279	82.430	ng/ul#	69
36) 2-Methylnaphthalene	12.404	142	2061863	80.044	ng/ul	92

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042122\
 Data File : BP010012.D
 Acq On : 21 Apr 2022 13:19
 Operator : CG/JU
 Sample : SST08027
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST080627

Quant Time: Apr 21 14:25:12 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 13:57:49 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.628	142	2086880	80.237	ng/ul#	95
39) 1,2,4,5-Tetrachloroben...	12.775	216	1205410	82.788	ng/ul	94
40) Hexachlorocyclopentadiene	12.757	237	718162	71.077	ng/ul	95
41) 2,4,6-Trichlorophenol	13.010	196	805163	86.591	ng/ul#	83
42) 2,4,5-Trichlorophenol	13.087	196	881621	87.735	ng/ul#	88
43) 1,1'-Biphenyl	13.416	154	2805695	80.801	ng/ul	93
44) 2-Chloronaphthalene	13.457	162	2187187	81.944	ng/ul	93
45) 2-Nitroaniline	13.657	65	772246	98.726	ng/ul#	82
47) Dimethylphthalate	14.034	163	2835498	78.809	ng/ul#	93
48) 2,6-Dinitrotoluene	14.151	165	644666	97.410	ng/ul	93
50) Acenaphthylene	14.298	152	3556388	81.370	ng/ul	96
51) 3-Nitroaniline	14.481	138	585152	85.050	ng/ul#	80
52) Acenaphthene	14.640	153	2292837	80.123	ng/ul	96
53) 2,4-Dinitrophenol	14.687	184	380108	95.019	ng/ul#	80
55) 4-Nitrophenol	14.798	109	487442	78.106	ng/ul#	47
56) Dibenzofuran	14.975	168	3337363	79.241	ng/ul	97
57) 2,4-Dinitrotoluene	14.940	165	912976	93.948	ng/ul#	84
58) 2,3,4,6-Tetrachlorophenol	15.204	232	772864	84.355	ng/ul#	87
59) Diethylphthalate	15.404	149	2935565	78.837	ng/ul	94
61) Fluorene	15.628	166	2730127	78.700	ng/ul	90
62) 4-Chlorophenyl-phenyle...	15.622	204	1430335	77.544	ng/ul	99
63) 4-Nitroaniline	15.651	138	560642	81.357	ng/ul#	79
66) 4,6-Dinitro-2-methylph...	15.710	198	546741	89.125	ng/ul#	93
67) N-Nitrosodiphenylamine	15.834	169	2432688	81.574	ng/ul	95
68) 4-Bromophenyl-phenylether	16.516	248	921643	81.728	ng/ul	92
69) Hexachlorobenzene	16.639	284	1056026	81.439	ng/ul#	92
70) Atrazine	16.792	200	984836	80.007	ng/ul#	88
71) Pentachlorophenol	16.981	266	706600	80.505	ng/ul#	84
72) Phenanthrene	17.375	178	4508520	81.076	ng/ul	99
74) Anthracene	17.469	178	4550284	80.708	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.381	216	1262634	83.830	ng/ul#	85
76) Pentachlorobenzene	14.898	250	1208522	82.318	ng/ul	92
77) Carbazole	17.733	167	4061695	79.718	ng/ul#	95
78) Di-n-butylphthalate	18.281	149	5194813	82.913	ng/ul#	93
80) Fluoranthene	19.333	202	5581221	84.196	ng/ul	97
82) Pyrene	19.692	202	5591827	83.201	ng/ul#	93
83) Butylbenzylphthalate	20.557	149	2481860	90.017	ng/ul#	79
84) 3,3'-Dichlorobenzidine	21.322	252	1516289	67.479	ng/ul#	96
85) Benzo(a)anthracene	21.398	228	5480162	81.979	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.316	149	3608457	87.010	ng/ul#	82
87) Chrysene	21.457	228	5294592	82.090	ng/ul	98
89) Di-n-octyl phthalate	22.263	149	6219148	88.381	ng/ul	100
90) Benzo(b)fluoranthene	23.121	252	5494440	82.341	ng/ul	99
91) Benzo(k)fluoranthene	23.174	252	5184336	85.143	ng/ul#	98
93) Benzo(a)pyrene	23.774	252	5092610	82.333	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.457	276	5275850	81.568	ng/ul#	92
95) Dibenzo(a,h)anthracene	26.468	278	4569166	81.468	ng/ul#	95
96) Benzo(g,h,i)perylene	27.239	276	4367941	81.901	ng/ul#	91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042122\
 Data File : BP010012.D
 Acq On : 21 Apr 2022 13:19
 Operator : CG/JU
 Sample : SSTD080627
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080627

Quant Time: Apr 21 14:25:12 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 13:57:49 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 32.138 ng/uL
RT: 3.405 min Scan# 54
Delta R.T. 0.000 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19




Instrument :
BNA_P
ClientSampleId :
SSTD080627


Tgt Ion: 88 Resp: 113406
Ion Ratio Lower Upper
88 100
43 31.6 6.8 10.2#
58 83.0 28.4 42.6#



#5
Pyridine
Concen: 84.666 ng/uL
RT: 3.805 min Scan# 122
Delta R.T. 0.000 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19

Tgt Ion: 79 Resp: 843081
Ion Ratio Lower Upper
79 100
52 88.3 35.5 53.3#
51 40.7 20.2 30.4#









#6
Benzaldehyde
Concen: 50.515 ng/ul
RT: 7.075 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19




Instrument :
BNA_P
ClientSampleId :
SSTD080627


Tgt Ion: 77 Resp: 420177
Ion Ratio Lower Upper
77 100
105 96.5 77.0 115.4
106 91.9 77.0 115.4



#8
Phenol
Concen: 86.746 ng/ul
RT: 7.134 min Scan# 688
Delta R.T. 0.006 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19

Tgt Ion: 94 Resp: 1177662
Ion Ratio Lower Upper
94 100
65 31.4 19.0 28.6#
66 44.3 34.6 51.8









#10
Bis(2-Chloroethyl)ether
Concen: 83.034 ng/ul
RT: 7.363 min Scan# 71
Delta R.T. 0.006 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19




Instrument :
BNA_P
ClientSampleId :
SSTD080627


Tgt Ion: 93 Resp: 878095
Ion Ratio Lower Upper
93 100
63 82.7 45.0 67.4#
95 33.1 26.3 39.5



#12
2-Chlorophenol
Concen: 86.546 ng/ul
RT: 7.511 min Scan# 752
Delta R.T. 0.006 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19

Tgt Ion:128 Resp: 892660
Ion Ratio Lower Upper
128 100
64 50.2 45.3 67.9
130 32.3 29.0 43.6





#13

2-Methylphenol

Concen: 84.090 ng/ul

RT: 8.387 min Scan# 900

Delta R.T. 0.006 min

Lab File: BP010012.D

Acq: 21 Apr 2022 13:19

Instrument :

BNA_P

ClientSampleId :





SSTD080627

Tgt Ion:108 Resp: 890215

Ion Ratio Lower Upper

108 100

107 89.9 70.8 106.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 81.362 ng/ul

RT: 8.475 min Scan# 916

Delta R.T. 0.006 min

Lab File: BP010012.D

Acq: 21 Apr 2022 13:19




Tgt Ion: 45 Resp: 1271827


Ion Ratio Lower Upper

45 100

77 13.9 17.1 25.7#

79 10.7 13.2 19.8#









#16
Acetophenone
Concen: 81.434 ng/ul
RT: 8.769 min Scan# 965
Delta R.T. 0.006 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19




Instrument :
BNA_P
ClientSampleId :
SSTD080627


Tgt Ion:105 Resp: 1452656
Ion Ratio Lower Upper
105 100
77 80.2 53.4 80.2#
51 30.0 13.8 20.8#



#17
N-Nitroso-di-n-propylamine
Concen: 82.572 ng/ul
RT: 8.769 min Scan# 966
Delta R.T. 0.018 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19

Tgt Ion: 70 Resp: 763656
Ion Ratio Lower Upper
70 100
42 60.8 31.8 47.8#
101 10.3 15.6 23.4#
130 21.2 13.7 20.5#









#18
4-Methylphenol
Concen: 84.719 ng/ul
RT: 8.728 min Scan# 91
Delta R.T. 0.018 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19




Instrument :
BNA_P
ClientSampleId :
SSTD080627


Tgt Ion:108 Resp: 971245
Ion Ratio Lower Upper
108 100
107 111.0 83.6 125.4



#19
Hexachloroethane
Concen: 82.224 ng/ul
RT: 9.028 min Scan# 1010
Delta R.T. 0.000 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19

Tgt Ion:117 Resp: 358066
Ion Ratio Lower Upper
117 100
201 105.2 106.5 159.7#
199 65.8 59.0 88.6









#22
Nitrobenzene
Concen: 86.930 ng/ul
RT: 9.152 min Scan# 1103
Delta R.T. 0.012 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19




Instrument :
BNA_P
ClientSampleId :
SSTD080627


Tgt Ion: 77 Resp: 1103124
Ion Ratio Lower Upper
77 100
123 43.5 42.3 63.5
65 14.6 5.8 8.8#



#23
Isophorone
Concen: 83.158 ng/ul
RT: 9.687 min Scan# 1122
Delta R.T. 0.018 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19

Tgt Ion: 82 Resp: 2174506
Ion Ratio Lower Upper
82 100
95 7.8 7.8 11.8#
138 16.9 12.9 19.3









#25
2-Nitrophenol
Concen: 94.179 ng/ul
RT: 9.863 min Scan# 1151
Delta R.T. 0.006 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19




Instrument :
BNA_P
ClientSampleId :
SSTD080627


Tgt Ion:139 Resp: 544442
Ion Ratio Lower Upper
139 100
65 57.4 49.4 74.2
109 35.9 44.7 67.1#



#26
2,4-Dimethylphenol
Concen: 82.147 ng/ul
RT: 9.922 min Scan# 1162
Delta R.T. 0.006 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19

Tgt Ion:107 Resp: 1126946
Ion Ratio Lower Upper
107 100
121 48.5 34.1 51.1
122 80.7 46.3 69.5#









#27
 Bis(2-Chloroethoxy)methane
 Concen: 80.171 ng/ul
 RT: 10.157 min Scan# 11
 Delta R.T. 0.006 min
 Lab File: BP010012.D
 Acq: 21 Apr 2022 13:19




Instrument :
 BNA_P
 ClientSampleId :
 SSTD080627


Tgt Ion: 93 Resp: 1230808
 Ion Ratio Lower Upper
 93 100
 95 32.5 26.2 39.4
 123 11.0 6.2 9.4#



#29
 2,4-Dichlorophenol
 Concen: 84.009 ng/ul
 RT: 10.399 min Scan# 1243
 Delta R.T. 0.006 min
 Lab File: BP010012.D
 Acq: 21 Apr 2022 13:19

Tgt Ion:162 Resp: 923752
 Ion Ratio Lower Upper
 162 100
 164 65.5 37.0 55.6#
 98 37.8 31.0 46.6









#30
Naphthalene
Concen: 81.668 ng/ul
RT: 10.799 min Scan# 1311
Delta R.T. 0.000 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19




Instrument :
BNA_P
ClientSampleId :
SSTD080627


Tgt Ion:128 Resp: 2879507
Ion Ratio Lower Upper
128 100
129 11.0 8.6 13.0
127 13.4 12.2 18.4



#32
4-Chloroaniline
Concen: 82.221 ng/ul
RT: 10.910 min Scan# 1330
Delta R.T. 0.006 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19

Tgt Ion:127 Resp: 1314410
Ion Ratio Lower Upper
127 100
129 32.6 24.6 37.0









#33
Hexachlorobutadiene
Concen: 79.752 ng/ul
RT: 11.093 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19




Instrument :
BNA_P
ClientSampleId :
SSTD080627


Tgt Ion:225 Resp: 631589
Ion Ratio Lower Upper
225 100
223 62.3 53.4 80.0
227 63.3 53.8 80.8



#34
Caprolactam
Concen: 47.378 ng/ul
RT: 11.704 min Scan# 1465
Delta R.T. 0.041 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19

Tgt Ion:113 Resp: 172287
Ion Ratio Lower Upper
113 100
55 177.3 58.2 87.2#
56 141.4 68.6 103.0#









#35
4-Chloro-3-methylphenol
Concen: 82.430 ng/ul
RT: 12.040 min Scan# 1519
Delta R.T. 0.018 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19




Instrument :
BNA_P
ClientSampleId :
SSTD080627


Tgt Ion:107 Resp: 1098279
Ion Ratio Lower Upper
107 100
144 24.5 15.8 23.8#
142 75.8 38.9 58.3#



#36
2-Methylnaphthalene
Concen: 80.044 ng/ul
RT: 12.404 min Scan# 1584
Delta R.T. 0.000 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19

Tgt Ion:142 Resp: 2061863
Ion Ratio Lower Upper
142 100
141 87.5 64.3 96.5









#37
1-Methylnaphthalene
Concen: 80.237 ng/ul
RT: 12.628 min Scan# 1621
Delta R.T. 0.006 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19




Instrument :
BNA_P
ClientSampleId :
SSTD080627


Tgt Ion:142 Resp: 2086880
Ion Ratio Lower Upper
142 100
141 92.1 69.8 104.8
116 4.0 5.6 8.4#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 82.788 ng/ul
RT: 12.775 min Scan# 1647
Delta R.T. 0.000 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19

Tgt Ion:216 Resp: 1205410
Ion Ratio Lower Upper
216 100
214 78.2 58.2 87.2
179 21.2 14.2 21.2
108 16.9 15.3 22.9









#40
Hexachlorocyclopentadiene
Concen: 71.077 ng/ul
RT: 12.757 min Scan# 1644
Delta R.T. 0.000 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19




Instrument :
BNA_P
ClientSampleId :
SSTD080627


Tgt Ion:237 Resp: 718162
Ion Ratio Lower Upper
237 100
235 62.1 53.2 79.8
272 12.4 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 86.591 ng/ul
RT: 13.010 min Scan# 1687
Delta R.T. 0.000 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19

Tgt Ion:196 Resp: 805163
Ion Ratio Lower Upper
196 100
198 96.1 64.1 96.1
200 30.7 19.9 29.9





#42

2,4,5-Trichlorophenol

Concen: 87.735 ng/ul

RT: 13.087 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP010012.D

Acq: 21 Apr 2022 13:19

Instrument :

BNA_P

ClientSampleId :

SSTD080627





Tgt Ion:196 Resp: 881621

Ion Ratio Lower Upper

196 100

198 94.8 68.6 103.0

200 31.4 32.8 49.2#



#43

1,1'-Biphenyl

Concen: 80.801 ng/ul

RT: 13.416 min Scan# 1756

Delta R.T. 0.006 min

Lab File: BP010012.D

Acq: 21 Apr 2022 13:19




Tgt Ion:154 Resp: 2805695


Ion Ratio Lower Upper

154 100

153 40.4 28.6 43.0

76 12.0 8.5 12.7









#44
2-Chloronaphthalene
Concen: 81.944 ng/ul
RT: 13.457 min Scan# 1762
Delta R.T. 0.006 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19




Instrument :
BNA_P
ClientSampleId :
SSTD080627


Tgt Ion:162 Resp: 2187187
Ion Ratio Lower Upper
162 100
164 32.9 27.5 41.3
127 40.0 37.1 55.7



#45
2-Nitroaniline
Concen: 98.726 ng/ul
RT: 13.657 min Scan# 1797
Delta R.T. 0.012 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19

Tgt Ion: 65 Resp: 772246
Ion Ratio Lower Upper
65 100
92 65.7 77.9 116.9#
138 96.9 81.5 122.3









#47
Dimethylphthalate
Concen: 78.809 ng/ul
RT: 14.034 min Scan# 1861
Delta R.T. 0.006 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19




Instrument :
BNA_P
ClientSampleId :
SSTD080627


Tgt Ion:163 Resp: 2835498
Ion Ratio Lower Upper
163 100
194 4.5 4.1 6.1
164 10.4 11.3 16.9#



#48
2,6-Dinitrotoluene
Concen: 97.410 ng/ul
RT: 14.151 min Scan# 1881
Delta R.T. 0.006 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19

Tgt Ion:165 Resp: 644666
Ion Ratio Lower Upper
165 100
89 58.1 51.7 77.5
121 20.5 18.3 27.5









#50
Acenaphthylene
Concen: 81.370 ng/ul
RT: 14.298 min Scan# 1905
Delta R.T. 0.006 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19




Instrument :
BNA_P
ClientSampleId :
SSTD080627


Tgt Ion:152 Resp: 3556388
Ion Ratio Lower Upper
152 100
151 19.7 18.3 27.5
153 13.4 11.0 16.6



#51
3-Nitroaniline
Concen: 85.050 ng/ul
RT: 14.481 min Scan# 1937
Delta R.T. 0.012 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19

Tgt Ion:138 Resp: 585152
Ion Ratio Lower Upper
138 100
108 11.9 15.8 23.8#
92 120.4 117.1 175.7









#52
Acenaphthene
Concen: 80.123 ng/ul
RT: 14.640 min Scan# 1964
Delta R.T. 0.000 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19




Instrument :
BNA_P
ClientSampleId :
SSTD080627


Tgt Ion:153 Resp: 2292837
Ion Ratio Lower Upper
153 100
152 47.0 34.3 51.5
154 90.0 70.0 105.0



#53
2,4-Dinitrophenol
Concen: 95.019 ng/ul
RT: 14.687 min Scan# 1972
Delta R.T. 0.012 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19

Tgt Ion:184 Resp: 380108
Ion Ratio Lower Upper
184 100
63 71.0 34.6 52.0#
154 66.7 57.7 86.5









#55
4-Nitrophenol
Concen: 78.106 ng/ul
RT: 14.798 min Scan# 1991
Delta R.T. 0.024 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19




Instrument :
BNA_P
ClientSampleId :
SSTD080627


Tgt Ion:109 Resp: 487442
Ion Ratio Lower Upper
109 100
139 118.9 61.4 92.2#
65 131.9 65.3 97.9#



#56
Dibenzofuran
Concen: 79.241 ng/ul
RT: 14.975 min Scan# 2021
Delta R.T. 0.006 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19

Tgt Ion:168 Resp: 3337363
Ion Ratio Lower Upper
168 100
139 39.3 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 93.948 ng/ul
RT: 14.940 min Scan# 2060
Delta R.T. 0.012 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19




Instrument :
BNA_P
ClientSampleId :
SSTD080627


Tgt Ion:165 Resp: 912976
Ion Ratio Lower Upper
165 100
63 47.0 29.5 44.3#
182 3.2 2.8 4.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 84.355 ng/ul
RT: 15.204 min Scan# 2060
Delta R.T. 0.006 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19

Tgt Ion:232 Resp: 772864
Ion Ratio Lower Upper
232 100
131 43.2 25.4 38.2#
130 2.2 4.2 6.2#
166 30.9 28.1 42.1









#59
Diethylphthalate
Concen: 78.837 ng/ul
RT: 15.404 min Scan# 2092
Delta R.T. 0.012 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19




Instrument :
BNA_P
ClientSampleId :
SSTD080627


Tgt Ion:149 Resp: 2935565
Ion Ratio Lower Upper
149 100
177 22.8 22.0 33.0
150 12.6 10.2 15.2



#61
Fluorene
Concen: 78.700 ng/ul
RT: 15.628 min Scan# 2132
Delta R.T. 0.006 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19

Tgt Ion:166 Resp: 2730127
Ion Ratio Lower Upper
166 100
165 97.6 70.3 105.5
167 13.6 9.5 14.3









#62
4-Chlorophenyl-phenylether
Concen: 77.544 ng/ul
RT: 15.622 min Scan# 2131
Delta R.T. 0.006 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19




Instrument :
BNA_P
ClientSampleId :
SSTD080627


Tgt Ion:204 Resp: 1430335
Ion Ratio Lower Upper
204 100
206 33.3 27.6 41.4
141 58.0 46.6 69.8



#63
4-Nitroaniline
Concen: 81.357 ng/ul
RT: 15.651 min Scan# 2136
Delta R.T. 0.018 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19

Tgt Ion:138 Resp: 560642
Ion Ratio Lower Upper
138 100
92 56.0 41.6 62.4
108 90.8 97.6 146.4#









#66
4,6-Dinitro-2-methylphenol
Concen: 89.125 ng/ul
RT: 15.710 min Scan# 2146
Delta R.T. 0.012 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19




Instrument :
BNA_P
ClientSampleId :
SSTD080627


Tgt Ion:198 Resp: 546741
Ion Ratio Lower Upper
198 100
182 3.4 6.0 9.0#
77 24.8 17.7 26.5



#67
N-Nitrosodiphenylamine
Concen: 81.574 ng/ul
RT: 15.834 min Scan# 2167
Delta R.T. 0.006 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19

Tgt Ion:169 Resp: 2432688
Ion Ratio Lower Upper
169 100
168 65.2 48.3 72.5
167 35.2 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 81.728 ng/ul
RT: 16.516 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19




Instrument :
BNA_P
ClientSampleId :
SSTD080627


Tgt Ion:248 Resp: 921643
Ion Ratio Lower Upper
248 100
250 96.9 79.4 119.0
141 70.7 67.2 100.8



#69
Hexachlorobenzene
Concen: 81.439 ng/ul
RT: 16.639 min Scan# 2304
Delta R.T. 0.006 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19

Tgt Ion:284 Resp: 1056026
Ion Ratio Lower Upper
284 100
142 31.7 19.2 28.8#
249 32.4 27.4 41.2









#70
Atrazine
Concen: 80.007 ng/ul
RT: 16.792 min Scan# 21
Delta R.T. 0.012 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19




Instrument :
BNA_P
ClientSampleId :
SSTD080627


Tgt Ion:200 Resp: 984836
Ion Ratio Lower Upper
200 100
173 26.4 19.2 28.8
215 50.0 32.2 48.4#



#71
Pentachlorophenol
Concen: 80.505 ng/ul
RT: 16.981 min Scan# 2362
Delta R.T. 0.006 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19

Tgt Ion:266 Resp: 706600
Ion Ratio Lower Upper
266 100
264 62.9 57.3 85.9
268 64.3 65.7 98.5#





#72

Phenanthrene

Concen: 81.076 ng/ul

RT: 17.375 min Scan# 2428

Delta R.T. 0.006 min

Lab File: BP010012.D

Acq: 21 Apr 2022 13:19

Instrument :

BNA_P

ClientSampleId :

SSTD080627





Tgt Ion:178 Resp: 4508520

Ion Ratio Lower Upper

178 100

179 15.1 11.7 17.5

176 18.9 14.5 21.7



#74

Anthracene

Concen: 80.708 ng/ul

RT: 17.469 min Scan# 2445

Delta R.T. 0.012 min

Lab File: BP010012.D

Acq: 21 Apr 2022 13:19




Tgt Ion:178 Resp: 4550284


Ion Ratio Lower Upper

178 100

179 15.7 12.3 18.5

176 18.5 13.4 20.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 83.830 ng/uL
RT: 13.381 min Scan# 1749
Delta R.T. 0.006 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19




Instrument :
BNA_P
ClientSampleId :
SSTD080627


Tgt Ion:216 Resp: 1262634
Ion Ratio Lower Upper
216 100
214 78.0 50.4 75.6#
218 48.9 35.2 52.8



#76
Pentachlorobenzene
Concen: 82.318 ng/uL
RT: 14.898 min Scan# 2008
Delta R.T. 0.000 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19

Tgt Ion:250 Resp: 1208522
Ion Ratio Lower Upper
250 100
248 61.9 43.5 65.3
252 62.1 46.5 69.7









#77
 Carbazole
 Concen: 79.718 ng/ul
 RT: 17.733 min Scan# 2488
 Delta R.T. 0.012 min
 Lab File: BP010012.D
 Acq: 21 Apr 2022 13:19




Instrument :
 BNA_P
 ClientSampleId :
 SSTD080627


Tgt Ion:167 Resp: 4061695
 Ion Ratio Lower Upper
 167 100
 166 21.9 15.8 23.8
 139 12.8 8.3 12.5#



#78
 Di-n-butylphthalate
 Concen: 82.913 ng/ul
 RT: 18.281 min Scan# 2583
 Delta R.T. 0.006 min
 Lab File: BP010012.D
 Acq: 21 Apr 2022 13:19

Tgt Ion:149 Resp: 5194813
 Ion Ratio Lower Upper
 149 100
 150 9.2 9.6 14.4#
 104 5.4 6.2 9.4#









#80
Fluoranthene
Concen: 84.196 ng/u1
RT: 19.333 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19




Instrument :
BNA_P
ClientSampleId :
SSTD080627


Tgt Ion:202 Resp: 5581221
Ion Ratio Lower Upper
202 100
101 10.1 9.6 14.4
100 7.9 6.3 9.5



#82
Pyrene
Concen: 83.201 ng/u1
RT: 19.692 min Scan# 2823
Delta R.T. 0.012 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19

Tgt Ion:202 Resp: 5591827
Ion Ratio Lower Upper
202 100
101 10.8 10.6 15.8
100 9.3 9.8 14.6#





#83

Butylbenzylphthalate

Concen: 90.017 ng/ul

RT: 20.557 min Scan# 2969

Delta R.T. 0.006 min

Lab File: BP010012.D

Acq: 21 Apr 2022 13:19

Instrument :

BNA_P

ClientSampleId :

SSTD080627





Tgt Ion:149 Resp: 2481860

Ion Ratio Lower Upper

149 100

91 69.2 71.1 106.7#

206 23.9 28.8 43.2#



#84

3,3'-Dichlorobenzidine

Concen: 67.479 ng/ul

RT: 21.322 min Scan# 3100

Delta R.T. 0.000 min

Lab File: BP010012.D

Acq: 21 Apr 2022 13:19




Tgt Ion:252 Resp: 1516289


Ion Ratio Lower Upper

252 100

254 64.7 49.8 74.6

126 9.6 4.3 6.5#





#85

Benzo(a)anthracene

Concen: 81.979 ng/ul

RT: 21.398 min Scan# 3112

Delta R.T. 0.006 min

Lab File: BP010012.D

Acq: 21 Apr 2022 13:19

Instrument :

BNA_P

ClientSampleId :

SSTD080627





Tgt Ion:228 Resp: 5480162

Ion Ratio Lower Upper

228 100

229 20.0 18.2 27.2

226 26.6 22.8 34.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 87.010 ng/ul

RT: 21.316 min Scan# 3099

Delta R.T. 0.000 min

Lab File: BP010012.D

Acq: 21 Apr 2022 13:19




Tgt Ion:149 Resp: 3608457


Ion Ratio Lower Upper

149 100

167 28.3 32.8 49.2#

279 5.0 5.6 8.4#





#87

Chrysene

Concen: 82.090 ng/ul

RT: 21.457 min Scan# 31

Delta R.T. 0.012 min

Lab File: BP010012.D

Acq: 21 Apr 2022 13:19

Instrument :

BNA_P

ClientSampleId :

SSTD080627





Tgt Ion:228 Resp: 5294592

Ion Ratio Lower Upper

228 100

226 29.4 22.7 34.1

229 19.8 15.2 22.8



#89

Di-n-octyl phthalate

Concen: 88.381 ng/ul




RT: 22.263 min Scan# 3260


Delta R.T. 0.006 min

Lab File: BP010012.D

Acq: 21 Apr 2022 13:19

Tgt Ion:149 Resp: 6219148









#90
Benzo(b)fluoranthene
Concen: 82.341 ng/ul
RT: 23.121 min Scan# 3406
Delta R.T. 0.012 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19




Instrument :
BNA_P
ClientSampleId :
SSTD080627


Tgt Ion:252 Resp: 5494440
Ion Ratio Lower Upper
252 100
253 21.8 17.4 26.0
125 8.5 7.3 10.9



#91
Benzo(k)fluoranthene
Concen: 85.143 ng/ul
RT: 23.174 min Scan# 3415
Delta R.T. 0.018 min
Lab File: BP010012.D
Acq: 21 Apr 2022 13:19

Tgt Ion:252 Resp: 5184336
Ion Ratio Lower Upper
252 100
253 21.7 17.9 26.9
125 8.2 5.1 7.7#





#93

Benzo(a)pyrene

Concen: 82.333 ng/ul

RT: 23.774 min Scan# 3

Delta R.T. 0.024 min

Lab File: BP010012.D

Acq: 21 Apr 2022 13:19

Instrument :

BNA_P

ClientSampleId :

SSTD080627





Tgt Ion:252 Resp: 5092610

Ion Ratio Lower Upper

252 100

253 21.8 17.5 26.3

125 9.0 6.4 9.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 81.568 ng/ul

RT: 26.457 min Scan# 3973

Delta R.T. 0.030 min

Lab File: BP010012.D

Acq: 21 Apr 2022 13:19




Tgt Ion:276 Resp: 5275850


Ion Ratio Lower Upper

276 100

138 21.5 12.3 18.5#

277 25.7 19.3 28.9









#95
 Dibenzo(a,h)anthracene
 Concen: 81.468 ng/ul
 RT: 26.468 min Scan# 3975
 Delta R.T. 0.029 min
 Lab File: BP010012.D
 Acq: 21 Apr 2022 13:19

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080627

Tgt Ion:278 Resp: 4569166
 Ion Ratio Lower Upper
 278 100
 139 13.1 6.6 10.0#
 279 23.9 18.4 27.6



#96
 Benzo(g,h,i)perylene
 Concen: 81.901 ng/ul
 RT: 27.239 min Scan# 4106
 Delta R.T. 0.030 min
 Lab File: BP010012.D
 Acq: 21 Apr 2022 13:19

Tgt Ion:276 Resp: 4367941
 Ion Ratio Lower Upper
 276 100
 138 18.1 11.6 17.4#
 277 24.0 15.9 23.9#

