

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042122\
 Data File : BP010068.D
 Acq On : 22 Apr 2022 22:48
 Operator : CG/JU
 Sample : N2490-11
 Misc :
 ALS Vial : 64 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0J34

Manual IntegrationsAPPROVED

Reviewed By :Jagrut Upadhyay 04/25/2022
 Supervised By :mohammad ahmed 04/26/2022

Quant Time: Apr 22 23:32:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 14:49:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.940	152	133466	20.000	ng/ul	0.00
20) Naphthalene-d8	10.746	136	647616	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.569	164	474266	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.322	188	1016616	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.404	240	767835	20.000	ng/ul	0.00
88) Perylene-d12	23.857	264	660001	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.369	96	12316	4.024	ng/uL	0.00
4) Pyridine-d5	3.799	84	72229	8.389	ng/ul	0.01
7) Phenol-d5	7.099	99	69254	5.705	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.263	67	237496	32.603	ng/ul	0.00
11) 2-Chlorophenol-d4	7.469	132	218523	24.305	ng/ul	0.00
15) 4-Methylphenol-d8	8.646	113	145628	14.478	ng/ul	0.00
21) Nitrobenzene-d5	9.099	128	162685	31.585	ng/ul	0.00
24) 2-Nitrophenol-d4	9.822	143	150005	26.666	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.363	165	287220	26.227	ng/ul	0.00
31) 4-Chloroaniline-d4	10.881	131	322389	20.531	ng/ul	0.00
46) Dimethylphthalate-d6	13.975	166	1222905	33.148	ng/ul	0.00
49) Acenaphthylene-d8	14.263	160	1374660	30.787	ng/ul	0.00
54) 4-Nitrophenol-d4	14.775	143	18708m	2.754	ng/ul	0.01
60) Fluorene-d10	15.563	176	1101514	34.710	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.681	200	80898	12.585	ng/ul	0.00
73) Anthracene-d10	17.422	188	1755346	35.690	ng/ul	0.00
81) Pyrene-d10	19.651	212	2014136	46.125	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.698	264	1271734	36.754	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042122\
Data File : BP010068.D
Acq On : 22 Apr 2022 22:48
Operator : CG/JU
Sample : N2490-11
Misc :
ALS Vial : 64 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0J34

Manual IntegrationsAPPROVED

Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022

Quant Time: Apr 22 23:32:50 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Apr 21 14:49:38 2022
Response via : Initial Calibration

