

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042122\
 Data File : BP010044.D
 Acq On : 22 Apr 2022 08:27
 Operator : CG/JU
 Sample : PB144266BS
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS266

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 04/25/2022
 Supervised By :mohammad ahmed 04/26/2022

Quant Time: Apr 22 23:24:39 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 14:49:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.940	152	142708	20.000	ng/ul	0.00
20) Naphthalene-d8	10.746	136	669449	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.569	164	480430	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.322	188	1077069	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.404	240	1096781	20.000	ng/ul	# 0.00
88) Perylene-d12	23.857	264	1021399	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.369	96	20109	6.145	ng/uL	0.00
4) Pyridine-d5	3.781	84	265332	28.820	ng/ul	0.00
7) Phenol-d5	7.099	99	393573	30.320	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.263	67	248484	31.902	ng/ul	0.00
11) 2-Chlorophenol-d4	7.469	132	313901	32.653	ng/ul	0.00
15) 4-Methylphenol-d8	8.646	113	351036	32.640	ng/ul	0.00
21) Nitrobenzene-d5	9.099	128	164264	30.851	ng/ul	0.00
24) 2-Nitrophenol-d4	9.822	143	186554	32.082	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.363	165	350733	30.982	ng/ul	0.00
31) 4-Chloroaniline-d4	10.875	131	453245	27.924	ng/ul	0.00
46) Dimethylphthalate-d6	13.981	166	1189344	31.824	ng/ul	0.00
49) Acenaphthylene-d8	14.263	160	1344206	29.719	ng/ul	0.00
54) 4-Nitrophenol-d4	14.763	143	220324	32.023	ng/ul	0.00
60) Fluorene-d10	15.563	176	1026288	31.924	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.681	200	217469	31.932	ng/ul	0.00
73) Anthracene-d10	17.422	188	1672540	32.098	ng/ul	0.00
81) Pyrene-d10	19.651	212	2047197	32.821	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.698	264	1766162	32.983	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.405	88	38994	11.466	ng/uL#	27
5) Pyridine	3.805	79	289738	30.276	ng/ul#	41
6) Benzaldehyde	7.075	77	253298	36.515	ng/ul	94
8) Phenol	7.128	94	429642	32.941	ng/ul#	93
10) Bis(2-Chloroethyl)ether	7.357	93	327504	32.592	ng/ul#	78
12) 2-Chlorophenol	7.505	128	324576	33.473	ng/ul	94
13) 2-Methylphenol	8.381	108	331411	33.132	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.475	45	495026	33.200	ng/ul#	85
16) Acetophenone	8.763	105	548652	32.698	ng/ul#	79
17) N-Nitroso-di-n-propyla...	8.751	70	293109	33.341	ng/ul#	73
18) 4-Methylphenol	8.710	108	367610	33.378	ng/ul	92
19) Hexachloroethane	9.028	117	133677	32.808	ng/ul	84
22) Nitrobenzene	9.140	77	424831	32.566	ng/ul#	84
23) Isophorone	9.669	82	835883	32.441	ng/ul	96
25) 2-Nitrophenol	9.857	139	204545	33.212	ng/ul#	85
26) 2,4-Dimethylphenol	9.916	107	433345	32.500	ng/ul#	79
27) Bis(2-Chloroethoxy)met...	10.151	93	482631	31.958	ng/ul#	96
29) 2,4-Dichlorophenol	10.393	162	359758	33.060	ng/ul#	84
30) Naphthalene	10.793	128	1124718	32.100	ng/ul	97
32) 4-Chloroaniline	10.898	127	469405	29.347	ng/ul	99
33) Hexachlorobutadiene	11.093	225	242003	31.379	ng/ul	97
34) Caprolactam	11.675	113	132146m	33.759	ng/ul	
35) 4-Chloro-3-methylphenol	12.022	107	444045	33.779	ng/ul#	70
36) 2-Methylnaphthalene	12.398	142	816393	32.127	ng/ul	93

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042122\
 Data File : BP010044.D
 Acq On : 22 Apr 2022 08:27
 Operator : CG/JU
 Sample : PB144266BS
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS266


Quant Time: Apr 22 23:24:39 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 14:49:38 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 04/25/2022
 Supervised By :mohammad ahmed 04/26/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.622	142	833533	32.305	ng/ul#	94
39) 1,2,4,5-Tetrachloroben...	12.769	216	479298	32.291	ng/ul	95
40) Hexachlorocyclopentadiene	12.757	237	246655	29.487	ng/ul	98
41) 2,4,6-Trichlorophenol	13.004	196	322322	33.562	ng/ul#	85
42) 2,4,5-Trichlorophenol	13.075	196	355258	33.999	ng/ul#	87
43) 1,1'-Biphenyl	13.404	154	1155114	32.599	ng/ul	94
44) 2-Chloronaphthalene	13.445	162	887123	32.351	ng/ul	94
45) 2-Nitroaniline	13.645	65	305731	34.384	ng/ul#	81
47) Dimethylphthalate	14.028	163	1197021	32.957	ng/ul#	92
48) 2,6-Dinitrotoluene	14.145	165	255214	34.240	ng/ul	93
50) Acenaphthylene	14.292	152	1471965	32.854	ng/ul	95
51) 3-Nitroaniline	14.469	138	247652	36.677	ng/ul#	83
52) Acenaphthene	14.634	153	949780	32.602	ng/ul	98
53) 2,4-Dinitrophenol	14.675	184	137761	33.075	ng/ul#	82
55) 4-Nitrophenol	14.781	109	201381	33.690	ng/ul#	47
56) Dibenzofuran	14.969	168	1406131	32.596	ng/ul	99
57) 2,4-Dinitrotoluene	14.928	165	377992	34.713	ng/ul#	79
58) 2,3,4,6-Tetrachlorophenol	15.198	232	317583	34.192	ng/ul#	87
59) Diethylphthalate	15.392	149	1260787	34.127	ng/ul#	92
61) Fluorene	15.622	166	1157810	32.690	ng/ul	91
62) 4-Chlorophenyl-phenyle...	15.616	204	607436	32.588	ng/ul	99
63) 4-Nitroaniline	15.633	138	268138	40.811	ng/ul#	79
66) 4,6-Dinitro-2-methylph...	15.698	198	216385	32.931	ng/ul#	94
67) N-Nitrosodiphenylamine	15.828	169	1029802	33.323	ng/ul	95
68) 4-Bromophenyl-phenylether	16.510	248	384088	33.432	ng/ul	95
69) Hexachlorobenzene	16.633	284	439331	32.946	ng/ul#	92
70) Atrazine	16.780	200	414133	32.962	ng/ul#	89
71) Pentachlorophenol	16.975	266	286003	33.678	ng/ul#	84
72) Phenanthrene	17.369	178	1969039	33.635	ng/ul	99
74) Anthracene	17.457	178	1995006	33.834	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.375	216	511788	32.548	ng/uL#	87
76) Pentachlorobenzene	14.892	250	497789	32.283	ng/uL	93
77) Carbazole	17.722	167	1776818	34.053	ng/ul#	96
78) Di-n-butylphthalate	18.275	149	2244188	34.959	ng/ul#	93
80) Fluoranthene	19.327	202	2445466	34.759	ng/ul#	95
82) Pyrene	19.680	202	2490397	34.358	ng/ul#	94
83) Butylbenzylphthalate	20.545	149	1064877	35.541	ng/ul#	82
84) 3,3'-Dichlorobenzidine	21.316	252	758855	38.453	ng/ul#	96
85) Benzo(a)anthracene	21.386	228	2385694	34.275	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.310	149	1555541	35.030	ng/ul#	82
87) Chrysene	21.445	228	2343234	34.190	ng/ul	99
89) Di-n-octyl phthalate	22.251	149	2671395	33.643	ng/ul	100
90) Benzo(b)fluoranthene	23.110	252	2380120	34.662	ng/ul	99
91) Benzo(k)fluoranthene	23.157	252	2192847	33.713	ng/ul#	97
93) Benzo(a)pyrene	23.751	252	2181893	34.361	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.421	276	2216604	35.091	ng/ul#	93
95) Dibenzo(a,h)anthracene	26.439	278	1896383	35.085	ng/ul#	95
96) Benzo(g,h,i)perylene	27.209	276	1866856	35.026	ng/ul	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#2
1,4-Dioxane
Concen: 11.466 ng/uL
RT: 3.405 min Scan# 54
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27

Instrument :
BNA_P
ClientSampleId :
SLCS266

Tgt Ion: 88 Resp: 38994
Ion Ratio Lower Upper
88 100
43 35.3 6.8 10.2
58 78.0 28.4 42.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#5
Pyridine
Concen: 30.276 ng/uL
RT: 3.805 min Scan# 122
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27


Tgt Ion: 79 Resp: 289738
Ion Ratio Lower Upper
79 100
52 90.5 35.5 53.3#
51 44.1 20.2 30.4#





#6
Benzaldehyde
Concen: 36.515 ng/ul
RT: 7.075 min Scan# 61
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Instrument :
BNA_P
ClientSampleId :
SLCS266



Tgt Ion: 77 Resp: 253298
Ion Ratio Lower Upper
77 100
105 92.1 77.0 115.4
106 89.5 77.0 115.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#8
Phenol
Concen: 32.941 ng/ul
RT: 7.128 min Scan# 687
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27


Tgt Ion: 94 Resp: 429642
Ion Ratio Lower Upper
94 100
65 31.8 19.0 28.6#
66 44.5 34.6 51.8





#10
Bis(2-Chloroethyl)ether
Concen: 32.592 ng/ul
RT: 7.357 min Scan# 71
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Instrument :
BNA_P
ClientSampleId :
SLCS266



Tgt Ion: 93 Resp: 327504
Ion Ratio Lower Upper
93 100
63 81.2 45.0 67.4
95 32.1 26.3 39.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#12
2-Chlorophenol
Concen: 33.473 ng/ul
RT: 7.505 min Scan# 751
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27

Tgt Ion:128 Resp: 324576
Ion Ratio Lower Upper
128 100
64 52.3 45.3 67.9
130 32.9 29.0 43.6





#13

2-Methylphenol

Concen: 33.132 ng/ul

RT: 8.381 min Scan# 900

Delta R.T. -0.000 min

Lab File: BP010044.D

Acq: 22 Apr 2022 08:27

Instrument :

BNA_P

ClientSampleId :

SLCS266

Tgt Ion:108 Resp: 33141

Ion Ratio Lower Upper





108 100

107 89.5 70.8 106.2

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#14

2,2'-oxybis(1-Chloropropane)

Concen: 33.200 ng/ul

RT: 8.475 min Scan# 916

Delta R.T. 0.006 min

Lab File: BP010044.D

Acq: 22 Apr 2022 08:27




Tgt Ion: 45 Resp: 495026


Ion Ratio Lower Upper

45 100

77 13.9 17.1 25.7#


79 10.3 13.2 19.8#





#16
Acetophenone
Concen: 32.698 ng/ul
RT: 8.763 min Scan# 90
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Instrument :
BNA_P
ClientSampleId :
SLCS266




Tgt Ion: 105 Resp: 548652
Ion Ratio Lower Upper
105 100
77 81.7 53.4 80.2
51 31.0 13.8 20.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#17
N-Nitroso-di-n-propylamine
Concen: 33.341 ng/ul
RT: 8.751 min Scan# 963
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Tgt Ion: 70 Resp: 293109
Ion Ratio Lower Upper
70 100
42 62.9 31.8 47.8#
101 10.4 15.6 23.4#
130 21.9 13.7 20.5#





#18
4-Methylphenol
Concen: 33.378 ng/ul
RT: 8.710 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Instrument :
BNA_P
ClientSampleId :
SLCS266




Tgt Ion:108 Resp: 367610
Ion Ratio Lower Upper
108 100
107 112.8 83.6 125.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#19
Hexachloroethane
Concen: 32.808 ng/ul
RT: 9.028 min Scan# 1010
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Tgt Ion:117 Resp: 133677
Ion Ratio Lower Upper
117 100
201 108.5 106.5 159.7
199 66.6 59.0 88.6





#22
Nitrobenzene
Concen: 32.566 ng/ul
RT: 9.140 min Scan# 1029
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Instrument : BNA_P
ClientSampleId : SLCS266




Tgt Ion: 77 Resp: 424833
Ion Ratio Lower Upper
77 100
123 42.3 42.3 63.5
65 14.6 5.8 8.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#23
Isophorone
Concen: 32.441 ng/ul
RT: 9.669 min Scan# 1119
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27



Tgt Ion: 82 Resp: 835883
Ion Ratio Lower Upper
82 100
95 8.2 7.8 11.8
138 17.4 12.9 19.3









#25
2-Nitrophenol
Concen: 33.212 ng/ul
RT: 9.857 min Scan# 1151
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27

Instrument :
BNA_P
ClientSampleId :
SLCS266

Tgt Ion:139 Resp: 204549
Ion Ratio Lower Upper
139 100
65 59.2 49.4 74.2
109 35.6 44.7 67.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#26
2,4-Dimethylphenol
Concen: 32.500 ng/ul
RT: 9.916 min Scan# 1161
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27


Tgt Ion:107 Resp: 433345
Ion Ratio Lower Upper
107 100
121 47.9 34.1 51.1
122 80.3 46.3 69.5





#27
Bis(2-Chloroethoxy)methane
Concen: 31.958 ng/ul
RT: 10.151 min Scan# 1201
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Instrument :
BNA_P
ClientSampleId :
SLCS266



Tgt Ion: 93 Resp: 48263
Ion Ratio Lower Upper
93 100
95 34.0 26.2 39.4
123 11.2 6.2 9.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#29
2,4-Dichlorophenol
Concen: 33.060 ng/ul
RT: 10.393 min Scan# 1242
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27


Tgt Ion:162 Resp: 359758
Ion Ratio Lower Upper
162 100
164 65.4 37.0 55.6#
98 38.3 31.0 46.6





#30
Naphthalene
Concen: 32.100 ng/ul
RT: 10.793 min Scan# 1310
Delta R.T. -0.006 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Instrument :
BNA_P
ClientSampleId :
SLCS266



Tgt Ion:128 Resp: 1124718
Ion Ratio Lower Upper
128 100
129 10.9 8.6 13.0
127 13.3 12.2 18.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#32
4-Chloroaniline
Concen: 29.347 ng/ul
RT: 10.898 min Scan# 1328
Delta R.T. -0.006 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27


Tgt Ion:127 Resp: 469405
Ion Ratio Lower Upper
127 100
129 31.4 24.6 37.0





#33
Hexachlorobutadiene
Concen: 31.379 ng/ul
RT: 11.093 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Instrument :
BNA_P
ClientSampleId :
SLCS266




Tgt Ion:225 Resp: 24200
Ion Ratio Lower Upper
225 100
223 65.0 53.4 80.0
227 64.2 53.8 80.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#34
Caprolactam
Concen: 33.759 ng/ul m
RT: 11.675 min Scan# 1460
Delta R.T. 0.012 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Tgt Ion:113 Resp: 132146
Ion Ratio Lower Upper
113 100
55 175.9 58.2 87.2#
56 142.6 68.6 103.0#





#35
4-Chloro-3-methylphenol
Concen: 33.779 ng/ul
RT: 12.022 min Scan# 1519
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Instrument :
BNA_P
ClientSampleId :
SLCS266




Tgt Ion:107 Resp: 444045
Ion Ratio Lower Upper
107 100
144 24.7 15.8 23.8
142 74.0 38.9 58.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#36
2-Methylnaphthalene
Concen: 32.127 ng/ul
RT: 12.398 min Scan# 1583
Delta R.T. -0.006 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Tgt Ion:142 Resp: 816393
Ion Ratio Lower Upper
142 100
141 86.7 64.3 96.5





#37
1-Methylnaphthalene
Concen: 32.305 ng/ul
RT: 12.622 min Scan# 1621
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Instrument :
BNA_P
ClientSampleId :
SLCS266




Tgt Ion:142 Resp: 833533
Ion Ratio Lower Upper
142 100
141 93.0 69.8 104.8
116 3.9 5.6 8.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 32.291 ng/ul
RT: 12.769 min Scan# 1646
Delta R.T. -0.006 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Tgt Ion:216 Resp: 479298
Ion Ratio Lower Upper
216 100
214 76.7 58.2 87.2
179 20.8 14.2 21.2
108 17.0 15.3 22.9





#40
Hexachlorocyclopentadiene
Concen: 29.487 ng/ul
RT: 12.757 min Scan# 1644
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Instrument :
BNA_P
ClientSampleId :
SLCS266




Tgt Ion:237 Resp: 246658
Ion Ratio Lower Upper
237 100
235 65.3 53.2 79.8
272 12.9 10.8 16.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#41
2,4,6-Trichlorophenol
Concen: 33.562 ng/ul
RT: 13.004 min Scan# 1686
Delta R.T. -0.006 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Tgt Ion:196 Resp: 322322
Ion Ratio Lower Upper
196 100
198 94.5 64.1 96.1
200 30.2 19.9 29.9#





#42
2,4,5-Trichlorophenol
Concen: 33.999 ng/ul
RT: 13.075 min Scan# 1698
Delta R.T. -0.006 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Instrument :
BNA_P
ClientSampleId :
SLCS266




Tgt Ion:196 Resp: 355258
Ion Ratio Lower Upper
196 100
198 96.1 68.6 103.0
200 31.1 32.8 49.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#43
1,1'-Biphenyl
Concen: 32.599 ng/ul
RT: 13.404 min Scan# 1754
Delta R.T. -0.006 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27



Tgt Ion:154 Resp: 1155114
Ion Ratio Lower Upper
154 100
153 39.7 28.6 43.0
76 12.3 8.5 12.7









#44
2-Chloronaphthalene
Concen: 32.351 ng/ul
RT: 13.445 min Scan# 1761
Delta R.T. -0.006 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27

Instrument :
BNA_P
ClientSampleId :
SLCS266

Tgt Ion:162 Resp: 88712
Ion Ratio Lower Upper
162 100
164 32.0 27.5 41.3
127 41.3 37.1 55.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#45
2-Nitroaniline
Concen: 34.384 ng/ul
RT: 13.645 min Scan# 1795
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27


Tgt Ion: 65 Resp: 305731
Ion Ratio Lower Upper
65 100
92 65.7 77.9 116.9#
138 95.6 81.5 122.3





#47
Dimethylphthalate
Concen: 32.957 ng/ul
RT: 14.028 min Scan# 1860
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Instrument : BNA_P
ClientSampleId : SLCS266




Tgt Ion:163 Resp: 119702
Ion Ratio Lower Upper
163 100
194 4.5 4.1 6.1
164 9.9 11.3 16.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#48
2,6-Dinitrotoluene
Concen: 34.240 ng/ul
RT: 14.145 min Scan# 1880
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27



Tgt Ion:165 Resp: 255214
Ion Ratio Lower Upper
165 100
89 58.2 51.7 77.5
121 20.6 18.3 27.5









#50
Acenaphthylene
Concen: 32.854 ng/ul
RT: 14.292 min Scan# 1905
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27

Instrument :
BNA_P
ClientSampleId :
SLCS266

Tgt Ion:152 Resp: 147196
Ion Ratio Lower Upper
152 100
151 19.9 18.3 27.5
153 13.0 11.0 16.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#51
3-Nitroaniline
Concen: 36.677 ng/ul
RT: 14.469 min Scan# 1935
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27


Tgt Ion:138 Resp: 247652
Ion Ratio Lower Upper
138 100
108 12.1 15.8 23.8#
92 125.7 117.1 175.7





#52
Acenaphthene
Concen: 32.602 ng/ul
RT: 14.634 min Scan# 1963
Delta R.T. -0.006 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Instrument :
BNA_P
ClientSampleId :
SLCS266




Tgt Ion:153 Resp: 949780
Ion Ratio Lower Upper
153 100
152 45.9 34.3 51.5
154 88.6 70.0 105.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#53
2,4-Dinitrophenol
Concen: 33.075 ng/ul
RT: 14.675 min Scan# 1970
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27



Tgt Ion:184 Resp: 137761
Ion Ratio Lower Upper
184 100
63 70.8 34.6 52.0#
154 69.0 57.7 86.5









#55
4-Nitrophenol
Concen: 33.690 ng/ul
RT: 14.781 min Scan# 1987
Delta R.T. 0.006 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27

Instrument :
BNA_P
ClientSampleId :
SLCS266

Tgt Ion:109 Resp: 20138
Ion Ratio Lower Upper
109 100
139 119.9 61.4 92.2
65 132.1 65.3 97.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#56
Dibenzofuran
Concen: 32.596 ng/ul
RT: 14.969 min Scan# 2020
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27


Tgt Ion:168 Resp: 1406131
Ion Ratio Lower Upper
168 100
139 38.4 30.1 45.1





#57
2,4-Dinitrotoluene
Concen: 34.713 ng/ul
RT: 14.928 min Scan# 2013
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Instrument :
BNA_P
ClientSampleId :
SLCS266




Tgt Ion:165 Resp: 377992
Ion Ratio Lower Upper
165 100
63 50.3 29.5 44.3#
182 3.0 2.8 4.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 34.192 ng/ul
RT: 15.198 min Scan# 2059
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Tgt Ion:232 Resp: 317583
Ion Ratio Lower Upper
232 100
131 43.0 25.4 38.2#
130 2.4 4.2 6.2#
166 30.4 28.1 42.1





#59
Diethylphthalate
Concen: 34.127 ng/ul
RT: 15.392 min Scan# 2092
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Instrument :
BNA_P
ClientSampleId :
SLCS266



Tgt Ion:149 Resp: 126078
Ion Ratio Lower Upper
149 100
177 21.9 22.0 33.0
150 12.3 10.2 15.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#61
Fluorene
Concen: 32.690 ng/ul
RT: 15.622 min Scan# 2131
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27


Tgt Ion:166 Resp: 1157810
Ion Ratio Lower Upper
166 100
165 96.6 70.3 105.5
167 13.8 9.5 14.3





#62
4-Chlorophenyl-phenylether
Concen: 32.588 ng/ul
RT: 15.616 min Scan# 2129
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Instrument :
BNA_P
ClientSampleId :
SLCS266




Tgt Ion:204 Resp: 607430
Ion Ratio Lower Upper
204 100
206 32.4 27.6 41.4
141 58.2 46.6 69.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#63
4-Nitroaniline
Concen: 40.811 ng/ul
RT: 15.633 min Scan# 2133
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Tgt Ion:138 Resp: 268138
Ion Ratio Lower Upper
138 100
92 55.3 41.6 62.4
108 90.8 97.6 146.4#





#66
4,6-Dinitro-2-methylphenol
Concen: 32.931 ng/ul
RT: 15.698 min Scan# 2143
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Instrument :
BNA_P
ClientSampleId :
SLCS266



Tgt Ion:198 Resp: 216385
Ion Ratio Lower Upper
198 100
182 3.7 6.0 9.0
77 24.3 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#67
N-Nitrosodiphenylamine
Concen: 33.323 ng/ul
RT: 15.828 min Scan# 2166
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27


Tgt Ion:169 Resp: 1029802
Ion Ratio Lower Upper
169 100
168 65.5 48.3 72.5
167 35.6 29.6 44.4





#68
4-Bromophenyl-phenylether
Concen: 33.432 ng/ul
RT: 16.510 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Instrument :
BNA_P
ClientSampleId :
SLCS266




Tgt Ion:248 Resp: 384088
Ion Ratio Lower Upper
248 100
250 99.3 79.4 119.0
141 73.7 67.2 100.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#69
Hexachlorobenzene
Concen: 32.946 ng/ul
RT: 16.633 min Scan# 2303
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Tgt Ion:284 Resp: 439331
Ion Ratio Lower Upper
284 100
142 31.7 19.2 28.8#
249 33.1 27.4 41.2





#70
Atrazine
Concen: 32.962 ng/ul
RT: 16.780 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Instrument :
BNA_P
ClientSampleId :
SLCS266




Tgt Ion:200 Resp: 41413
Ion Ratio Lower Upper
200 100
173 26.6 19.2 28.8
215 49.5 32.2 48.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#71
Pentachlorophenol
Concen: 33.678 ng/ul
RT: 16.975 min Scan# 2361
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Tgt Ion:266 Resp: 286003
Ion Ratio Lower Upper
266 100
264 62.4 57.3 85.9
268 64.2 65.7 98.5





#72
Phenanthrene
Concen: 33.635 ng/u1
RT: 17.369 min Scan# 2427
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Instrument :
BNA_P
ClientSampleId :
SLCS266




Tgt Ion:178 Resp: 1969039
Ion Ratio Lower Upper
178 100
179 14.9 11.7 17.5
176 18.5 14.5 21.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#74
Anthracene
Concen: 33.834 ng/u1
RT: 17.457 min Scan# 2443
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Tgt Ion:178 Resp: 1995006
Ion Ratio Lower Upper
178 100
179 15.3 12.3 18.5
176 18.1 13.4 20.2





#75
1,2,3,4-Tetrachlorobenzene
Concen: 32.548 ng/uL
RT: 13.375 min Scan# 1749
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Instrument :
BNA_P
ClientSampleId :
SLCS266




Tgt Ion:216 Resp: 511788
Ion Ratio Lower Upper
216 100
214 76.2 50.4 75.6
218 48.4 35.2 52.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#76
Pentachlorobenzene
Concen: 32.283 ng/uL
RT: 14.892 min Scan# 2007
Delta R.T. -0.006 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Tgt Ion:250 Resp: 497789
Ion Ratio Lower Upper
250 100
248 60.8 43.5 65.3
252 61.8 46.5 69.7





#77
 Carbazole
 Concen: 34.053 ng/ul
 RT: 17.722 min Scan# 2488
 Delta R.T. -0.000 min
 Lab File: BP010044.D
 Acq: 22 Apr 2022 08:27




Instrument :
 BNA_P
 ClientSampleId :
 SLCS266




Tgt Ion:167 Resp: 1776818
 Ion Ratio Lower Upper
 167 100
 166 21.0 15.8 23.8
 139 12.8 8.3 12.5

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
 Supervised By :mohammad ahmed 04/26/2022




#78
 Di-n-butylphthalate
 Concen: 34.959 ng/ul
 RT: 18.275 min Scan# 2582
 Delta R.T. -0.000 min
 Lab File: BP010044.D
 Acq: 22 Apr 2022 08:27




Tgt Ion:149 Resp: 2244188
 Ion Ratio Lower Upper
 149 100
 150 9.2 9.6 14.4#
 104 5.4 6.2 9.4#





#80
Fluoranthene
Concen: 34.759 ng/u1
RT: 19.327 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Instrument :
BNA_P
ClientSampleId :
SLCS266



Tgt Ion:202 Resp: 2445460
Ion Ratio Lower Upper
202 100
101 9.3 9.6 14.4
100 7.4 6.3 9.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#82
Pyrene
Concen: 34.358 ng/u1
RT: 19.680 min Scan# 2821
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27


Tgt Ion:202 Resp: 2490397
Ion Ratio Lower Upper
202 100
101 11.2 10.6 15.8
100 9.3 9.8 14.6#





#83
Butylbenzylphthalate
Concen: 35.541 ng/ul
RT: 20.545 min Scan# 2968
Delta R.T. -0.006 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Instrument :
BNA_P
ClientSampleId :
SLCS266




Tgt Ion:149 Resp: 106487
Ion Ratio Lower Upper
149 100
91 74.6 71.1 106.7
206 21.7 28.8 43.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#84
3,3'-Dichlorobenzidine
Concen: 38.453 ng/ul
RT: 21.316 min Scan# 3099
Delta R.T. -0.006 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Tgt Ion:252 Resp: 758855
Ion Ratio Lower Upper
252 100
254 64.9 49.8 74.6
126 9.6 4.3 6.5





#85
Benzo(a)anthracene
Concen: 34.275 ng/ul
RT: 21.386 min Scan# 3111
Delta R.T. -0.006 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Instrument :
BNA_P
ClientSampleId :
SLCS266




Tgt Ion:228 Resp: 2385694
Ion Ratio Lower Upper
228 100
229 19.9 18.2 27.2
226 25.9 22.8 34.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#86
Bis(2-ethylhexyl)phthalate
Concen: 35.030 ng/ul
RT: 21.310 min Scan# 3098
Delta R.T. -0.006 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Tgt Ion:149 Resp: 1555541
Ion Ratio Lower Upper
149 100
167 28.6 32.8 49.2#
279 4.8 5.6 8.4#





#87
Chrysene
Concen: 34.190 ng/ul
RT: 21.445 min Scan# 3121
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Instrument :
BNA_P
ClientSampleId :
SLCS266



Tgt Ion:228 Resp: 2343234
Ion Ratio Lower Upper
228 100
226 29.1 22.7 34.1
229 19.4 15.2 22.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#89
Di-n-octyl phthalate
Concen: 33.643 ng/ul
RT: 22.251 min Scan# 3258
Delta R.T. -0.006 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27


Tgt Ion:149 Resp: 2671395





#90
Benzo(b)fluoranthene
Concen: 34.662 ng/ul
RT: 23.110 min Scan# 3403
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27




Instrument :
BNA_P
ClientSampleId :
SLCS266



Tgt Ion:252 Resp: 2380120
Ion Ratio Lower Upper
252 100
253 21.7 17.4 26.0
125 8.1 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#91
Benzo(k)fluoranthene
Concen: 33.713 ng/ul
RT: 23.157 min Scan# 3412
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27

Tgt Ion:252 Resp: 2192847
Ion Ratio Lower Upper
252 100
253 21.6 17.9 26.9
125 8.4 5.1 7.7#









#93
Benzo(a)pyrene
Concen: 34.361 ng/ul
RT: 23.751 min Scan# 3513
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27

Instrument :
BNA_P
ClientSampleId :
SLCS266

Tgt Ion:252 Resp: 218189
Ion Ratio Lower Upper
252 100
253 22.1 17.5 26.3
125 9.1 6.4 9.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#94
Indeno(1,2,3-cd)pyrene
Concen: 35.091 ng/ul
RT: 26.421 min Scan# 3967
Delta R.T. -0.006 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27

Tgt Ion:276 Resp: 2216604
Ion Ratio Lower Upper
276 100
138 20.6 12.3 18.5#
277 25.7 19.3 28.9





#95
Dibenzo(a,h)anthracene
Concen: 35.085 ng/ul
RT: 26.439 min Scan# 3969
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27

Instrument :

BNA_P

ClientSampleId :





SLCS266

Tgt Ion:278 Resp: 189638
Ion Ratio Lower Upper
278 100
139 13.0 6.6 10.0
279 24.2 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#96
Benzo(g,h,i)perylene
Concen: 35.026 ng/ul
RT: 27.209 min Scan# 4101
Delta R.T. -0.000 min
Lab File: BP010044.D
Acq: 22 Apr 2022 08:27

Tgt Ion:276 Resp: 1866856
Ion Ratio Lower Upper
276 100
138 17.3 11.6 17.4
277 23.7 15.9 23.9

