

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042122\
 Data File : BP010070.D
 Acq On : 22 Apr 2022 23:55
 Operator : CG/JU
 Sample : N2475-05
 Misc :
 ALS Vial : 66 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 A0BN6

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 04/25/2022
 Supervised By :mohammad ahmed 04/26/2022

Quant Time: Apr 23 01:58:20 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 14:49:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.940	152	141881	20.000	ng/u1	0.00
20) Naphthalene-d8	10.746	136	676701	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.569	164	494690	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.322	188	1084514	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.404	240	833137	20.000	ng/u1	# 0.00
88) Perylene-d12	23.857	264	709272	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.370	96	14392m	4.423	ng/uL	0.00
4) Pyridine-d5	3.805	84	50937m	5.565	ng/u1	0.02
7) Phenol-d5	7.099	99	74995	5.811	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.264	67	234924	30.337	ng/u1	0.00
11) 2-Chlorophenol-d4	7.469	132	224585	23.498	ng/u1	0.00
15) 4-Methylphenol-d8	8.646	113	162682	15.215	ng/u1	0.00
21) Nitrobenzene-d5	9.099	128	161363	29.981	ng/u1	0.00
24) 2-Nitrophenol-d4	9.822	143	154733	26.324	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.363	165	295411	25.816	ng/u1	0.00
31) 4-Chloroaniline-d4	10.875	131	343116	20.912	ng/u1	0.00
46) Dimethylphthalate-d6	13.975	166	1214389	31.558	ng/u1	0.00
49) Acenaphthylene-d8	14.263	160	1358354	29.166	ng/u1	0.00
54) 4-Nitrophenol-d4	14.769	143	19007m	2.683	ng/u1	0.00
60) Fluorene-d10	15.563	176	1071306	32.364	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.675	200	96429	14.062	ng/u1	0.00
73) Anthracene-d10	17.422	188	1709259	32.577	ng/u1	0.00
81) Pyrene-d10	19.651	212	1965730	41.488	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.698	264	1281531	34.465	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042122\
Data File : BP010070.D
Acq On : 22 Apr 2022 23:55
Operator : CG/JU
Sample : N2475-05
Misc :
ALS Vial : 66 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
A0BN6

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022

Quant Time: Apr 23 01:58:20 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Apr 21 14:49:38 2022
Response via : Initial Calibration

