

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042122\
 Data File : BP010076.D
 Acq On : 23 Apr 2022 03:17
 Operator : CG/JU
 Sample : PB144290BS
 Misc :
 ALS Vial : 72 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS290

Quant Time: Apr 23 04:04:44 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 14:49:38 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 04/25/2022
 Supervised By :mohammad ahmed 04/26/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.940	152	150132	20.000	ng/u1	0.00
20) Naphthalene-d8	10.746	136	716550	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.569	164	512702	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.322	188	1118291	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.404	240	1017735	20.000	ng/u1	# 0.00
88) Perylene-d12	23.857	264	869454	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.370	96	18116	5.262	ng/uL	0.00
4) Pyridine-d5	3.781	84	279309	28.838	ng/u1	0.00
7) Phenol-d5	7.099	99	397334	29.096	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.263	67	248230	30.294	ng/u1	0.00
11) 2-Chlorophenol-d4	7.469	132	313487	30.997	ng/u1	0.00
15) 4-Methylphenol-d8	8.646	113	344308	30.431	ng/u1	0.00
21) Nitrobenzene-d5	9.099	128	167952	29.470	ng/u1	0.00
24) 2-Nitrophenol-d4	9.822	143	182598	29.337	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.363	165	349112	28.812	ng/u1	0.00
31) 4-Chloroaniline-d4	10.875	131	440440	25.351	ng/u1	0.00
46) Dimethylphthalate-d6	13.975	166	1138682	28.551	ng/u1	0.00
49) Acenaphthylene-d8	14.263	160	1315217	27.248	ng/u1	0.00
54) 4-Nitrophenol-d4	14.763	143	177750	24.209	ng/u1	0.00
60) Fluorene-d10	15.563	176	991581	28.903	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.681	200	167174	23.642	ng/u1	0.00
73) Anthracene-d10	17.422	188	1570248	29.024	ng/u1	0.00
81) Pyrene-d10	19.651	212	1889522	32.646	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.698	264	1402145	30.761	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.399	88	38467	10.752	ng/uL#	17
5) Pyridine	3.799	79	306613	30.455	ng/u1#	40
6) Benzaldehyde	7.075	77	256594	35.161	ng/u1	96
8) Phenol	7.128	94	438176	31.934	ng/u1#	93
10) Bis(2-Chloroethyl)ether	7.358	93	332635	31.466	ng/u1#	80
12) 2-Chlorophenol	7.499	128	325730	31.931	ng/u1	93
13) 2-Methylphenol	8.381	108	327952	31.164	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.469	45	481744	30.711	ng/u1#	85
16) Acetophenone	8.757	105	545447	30.899	ng/u1#	81
17) N-Nitroso-di-n-propyla...	8.752	70	287849	31.123	ng/u1#	75
18) 4-Methylphenol	8.710	108	367645	31.730	ng/u1	92
19) Hexachloroethane	9.022	117	128741	30.034	ng/u1	85
22) Nitrobenzene	9.140	77	422551	30.262	ng/u1#	85
23) Isophorone	9.669	82	815673	29.576	ng/u1#	97
25) 2-Nitrophenol	9.852	139	199211	30.220	ng/u1#	87
26) 2,4-Dimethylphenol	9.916	107	413276	28.957	ng/u1#	80
27) Bis(2-Chloroethoxy)met...	10.152	93	476375	29.470	ng/u1#	97
29) 2,4-Dichlorophenol	10.393	162	357374	30.682	ng/u1#	83
30) Naphthalene	10.793	128	1128606	30.093	ng/u1	97
32) 4-Chloroaniline	10.899	127	451179	26.353	ng/u1	100
33) Hexachlorobutadiene	11.087	225	237383	28.756	ng/u1	95
34) Caprolactam	11.675	113	111385m	26.585	ng/u1	
35) 4-Chloro-3-methylphenol	12.028	107	426589	30.318	ng/u1#	69
36) 2-Methylnaphthalene	12.398	142	809102	29.747	ng/u1	90

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042122\
 Data File : BP010076.D
 Acq On : 23 Apr 2022 03:17
 Operator : CG/JU
 Sample : PB144290BS
 Misc :
 ALS Vial : 72 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS290

Quant Time: Apr 23 04:04:44 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 14:49:38 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 04/25/2022
 Supervised By :mohammad ahmed 04/26/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.616	142	819048	29.657	ng/ul#	94
39) 1,2,4,5-Tetrachloroben...	12.769	216	471773	29.784	ng/ul	94
40) Hexachlorocyclopentadiene	12.751	237	180145	20.180	ng/ul	94
41) 2,4,6-Trichlorophenol	13.004	196	304814	29.742	ng/ul#	82
42) 2,4,5-Trichlorophenol	13.081	196	327334	29.355	ng/ul#	87
43) 1,1'-Biphenyl	13.404	154	1140896	30.171	ng/ul	93
44) 2-Chloronaphthalene	13.445	162	882015	30.140	ng/ul	93
45) 2-Nitroaniline	13.645	65	297525	31.355	ng/ul#	82
47) Dimethylphthalate	14.028	163	1155980	29.824	ng/ul#	92
48) 2,6-Dinitrotoluene	14.140	165	248552	31.247	ng/ul	94
50) Acenaphthylene	14.292	152	1439080	30.099	ng/ul	95
51) 3-Nitroaniline	14.469	138	234906	32.600	ng/ul#	82
52) Acenaphthene	14.634	153	924593	29.740	ng/ul	97
53) 2,4-Dinitrophenol	14.675	184	79193	17.816	ng/ul#	80
55) 4-Nitrophenol	14.781	109	159046	24.933	ng/ul#	46
56) Dibenzofuran	14.969	168	1384966	30.084	ng/ul	100
57) 2,4-Dinitrotoluene	14.928	165	364957	31.406	ng/ul#	86
58) 2,3,4,6-Tetrachlorophenol	15.198	232	303680	30.637	ng/ul#	87
59) Diethylphthalate	15.392	149	1174521	29.791	ng/ul	94
61) Fluorene	15.622	166	1121587	29.674	ng/ul	91
62) 4-Chlorophenyl-phenyle...	15.610	204	583471	29.332	ng/ul	97
63) 4-Nitroaniline	15.634	138	241002	34.372	ng/ul#	78
66) 4,6-Dinitro-2-methylph...	15.692	198	169366	24.825	ng/ul#	87
67) N-Nitrosodiphenylamine	15.822	169	992280	30.925	ng/ul	95
68) 4-Bromophenyl-phenylether	16.510	248	373337	31.298	ng/ul	93
69) Hexachlorobenzene	16.634	284	421066	30.412	ng/ul#	92
70) Atrazine	16.781	200	383746	29.417	ng/ul#	91
71) Pentachlorophenol	16.975	266	241493	27.388	ng/ul#	83
72) Phenanthrene	17.363	178	1857813	30.565	ng/ul	99
74) Anthracene	17.457	178	1885246	30.794	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.375	216	501339	30.708	ng/uL#	85
76) Pentachlorobenzene	14.892	250	477254	29.811	ng/uL	89
77) Carbazole	17.722	167	1641231	30.295	ng/ul#	95
78) Di-n-butylphthalate	18.275	149	2087867	31.325	ng/ul#	93
80) Fluoranthene	19.327	202	2281329	34.945	ng/ul	96
82) Pyrene	19.680	202	2307579	34.308	ng/ul#	94
83) Butylbenzylphthalate	20.545	149	965239	34.718	ng/ul#	80
84) 3,3'-Dichlorobenzidine	21.316	252	488953	26.701	ng/ul#	96
85) Benzo(a)anthracene	21.386	228	2064860	31.970	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.310	149	1418247	34.418	ng/ul#	82
87) Chrysene	21.445	228	1979543	31.126	ng/ul	99
89) Di-n-octyl phthalate	22.251	149	2382142	35.243	ng/ul	100
90) Benzo(b)fluoranthene	23.110	252	1916284	32.784	ng/ul	99
91) Benzo(k)fluoranthene	23.157	252	1725276	31.160	ng/ul#	98
93) Benzo(a)pyrene	23.751	252	1693959	31.339	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.421	276	1642785	30.552	ng/ul#	94
95) Dibenzo(a,h)anthracene	26.439	278	1373758	29.858	ng/ul#	96
96) Benzo(g,h,i)perylene	27.204	276	1439296	31.723	ng/ul#	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042122\
 Data File : BP010076.D
 Acq On : 23 Apr 2022 03:17
 Operator : CG/JU
 Sample : PB144290BS
 Misc :
 ALS Vial : 72 Sample Multiplier: 1

Instrument :

BNA_P

ClientSampleId :


SLCS290


Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 04/25/2022
 Supervised By :mohammad ahmed 04/26/2022


Quant Time: Apr 23 04:04:44 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 14:49:38 2022
 Response via : Initial Calibration





#2
1,4-Dioxane
Concen: 10.752 ng/uL
RT: 3.399 min Scan# 51
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Instrument :
BNA_P
ClientSampleId :
SLCS290



Tgt Ion: 88 Resp: 3846
Ion Ratio Lower Upper
88 100
43 37.4 6.8 10.2
58 84.6 28.4 42.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#5
Pyridine
Concen: 30.455 ng/uL
RT: 3.799 min Scan# 121
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17


Tgt Ion: 79 Resp: 306613
Ion Ratio Lower Upper
79 100
52 93.3 35.5 53.3#
51 41.5 20.2 30.4#





#6
Benzaldehyde
Concen: 35.161 ng/ul
RT: 7.075 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Instrument :
BNA_P
ClientSampleId :
SLCS290




Tgt Ion: 77 Resp: 256594
Ion Ratio Lower Upper
77 100
105 95.3 77.0 115.4
106 88.6 77.0 115.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#8
Phenol
Concen: 31.934 ng/ul
RT: 7.128 min Scan# 687
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Tgt Ion: 94 Resp: 438176
Ion Ratio Lower Upper
94 100
65 32.0 19.0 28.6#
66 43.8 34.6 51.8





#10
Bis(2-Chloroethyl)ether
Concen: 31.466 ng/ul
RT: 7.358 min Scan# 710
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Instrument :
BNA_P
ClientSampleId :
SLCS290




Tgt Ion: 93 Resp: 332635
Ion Ratio Lower Upper
93 100
63 79.5 45.0 67.4
95 33.3 26.3 39.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#12
2-Chlorophenol
Concen: 31.931 ng/ul
RT: 7.499 min Scan# 750
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17



Tgt Ion:128 Resp: 325730
Ion Ratio Lower Upper
128 100
64 52.3 45.3 67.9
130 31.5 29.0 43.6





#13

2-Methylphenol

Concen: 31.164 ng/ul

RT: 8.381 min Scan# 900

Delta R.T. 0.000 min

Lab File: BP010076.D

Acq: 23 Apr 2022 03:17

Instrument :

BNA_P

ClientSampleId :

SLCS290

Tgt Ion:108 Resp: 32795

Ion Ratio Lower Upper





108 100

107 92.2 70.8 106.2

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#14

2,2'-oxybis(1-Chloropropane)

Concen: 30.711 ng/ul

RT: 8.469 min Scan# 915

Delta R.T. 0.000 min

Lab File: BP010076.D

Acq: 23 Apr 2022 03:17




Tgt Ion: 45 Resp: 481744


Ion Ratio Lower Upper

45 100

77 14.1 17.1 25.7#

79 10.2 13.2 19.8#









#16
Acetophenone
Concen: 30.899 ng/ul
RT: 8.757 min Scan# 964
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17

Instrument :
BNA_P
ClientSampleId :
SLCS290

Tgt Ion:105 Resp: 54544
Ion Ratio Lower Upper
105 100
77 79.5 53.4 80.2
51 30.4 13.8 20.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#17
N-Nitroso-di-n-propylamine
Concen: 31.123 ng/ul
RT: 8.752 min Scan# 963
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17


Tgt Ion: 70 Resp: 287849
Ion Ratio Lower Upper
70 100
42 61.7 31.8 47.8#
101 10.0 15.6 23.4#
130 19.9 13.7 20.5





#18
4-Methylphenol
Concen: 31.730 ng/ul
RT: 8.710 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Instrument :
BNA_P
ClientSampleId :
SLCS290




Tgt Ion:108 Resp: 36764
Ion Ratio Lower Upper
108 100
107 112.5 83.6 125.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#19
Hexachloroethane
Concen: 30.034 ng/ul
RT: 9.022 min Scan# 1009
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Tgt Ion:117 Resp: 128741
Ion Ratio Lower Upper
117 100
201 109.1 106.5 159.7
199 69.1 59.0 88.6





#22
Nitrobenzene
Concen: 30.262 ng/u1
RT: 9.140 min Scan# 1029
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Instrument :
BNA_P
ClientSampleId :
SLCS290




Tgt Ion: 77 Resp: 42255
Ion Ratio Lower Upper
77 100
123 43.0 42.3 63.5
65 15.1 5.8 8.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#23
Isophorone
Concen: 29.576 ng/u1
RT: 9.669 min Scan# 1119
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17



Tgt Ion: 82 Resp: 815673
Ion Ratio Lower Upper
82 100
95 7.8 7.8 11.8#
138 17.0 12.9 19.3









#25
2-Nitrophenol
Concen: 30.220 ng/ul
RT: 9.852 min Scan# 1151
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17

Instrument :
BNA_P
ClientSampleId :
SLCS290

Tgt Ion:139 Resp: 19921
Ion Ratio Lower Upper
139 100
65 62.4 49.4 74.2
109 36.4 44.7 67.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#26
2,4-Dimethylphenol
Concen: 28.957 ng/ul
RT: 9.916 min Scan# 1161
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17


Tgt Ion:107 Resp: 413276
Ion Ratio Lower Upper
107 100
121 46.4 34.1 51.1
122 80.2 46.3 69.5





#27
Bis(2-Chloroethoxy)methane
Concen: 29.470 ng/ul
RT: 10.152 min Scan# 1201
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Instrument :
BNA_P
ClientSampleId :
SLCS290




Tgt Ion: 93 Resp: 476374
Ion Ratio Lower Upper
93 100
95 32.2 26.2 39.4
123 11.2 6.2 9.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#29
2,4-Dichlorophenol
Concen: 30.682 ng/ul
RT: 10.393 min Scan# 1242
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Tgt Ion:162 Resp: 357374
Ion Ratio Lower Upper
162 100
164 66.3 37.0 55.6
98 38.4 31.0 46.6





#30
Naphthalene
Concen: 30.093 ng/u1
RT: 10.793 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Instrument :
BNA_P
ClientSampleId :
SLCS290



Tgt Ion:128 Resp: 1128600
Ion Ratio Lower Upper
128 100
129 11.0 8.6 13.0
127 13.0 12.2 18.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#32
4-Chloroaniline
Concen: 26.353 ng/u1
RT: 10.899 min Scan# 1328
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17

Tgt Ion:127 Resp: 451179
Ion Ratio Lower Upper
127 100
129 30.7 24.6 37.0









#33
Hexachlorobutadiene
Concen: 28.756 ng/ul
RT: 11.087 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17

Instrument :
BNA_P
ClientSampleId :
SLCS290

Tgt Ion:225 Resp: 23738
Ion Ratio Lower Upper
225 100
223 62.8 53.4 80.0
227 63.6 53.8 80.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#34
Caprolactam
Concen: 26.585 ng/ul m
RT: 11.675 min Scan# 1460
Delta R.T. 0.012 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17


Tgt Ion:113 Resp: 111385
Ion Ratio Lower Upper
113 100
55 178.1 58.2 87.2#
56 141.7 68.6 103.0#





#35
4-Chloro-3-methylphenol
Concen: 30.318 ng/ul
RT: 12.028 min Scan# 1520
Delta R.T. 0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Instrument :
BNA_P
ClientSampleId :
SLCS290



Tgt Ion:107 Resp: 426589
Ion Ratio Lower Upper
107 100
144 24.3 15.8 23.8
142 75.1 38.9 58.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#36
2-Methylnaphthalene
Concen: 29.747 ng/ul
RT: 12.398 min Scan# 1583
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17


Tgt Ion:142 Resp: 809102
Ion Ratio Lower Upper
142 100
141 89.1 64.3 96.5





#37
1-Methylnaphthalene
Concen: 29.657 ng/ul
RT: 12.616 min Scan# 1620
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Instrument :
BNA_P
ClientSampleId :
SLCS290




Tgt Ion:142 Resp: 81904
Ion Ratio Lower Upper
142 100
141 92.5 69.8 104.8
116 4.2 5.6 8.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 29.784 ng/ul
RT: 12.769 min Scan# 1646
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Tgt Ion:216 Resp: 471773
Ion Ratio Lower Upper
216 100
214 77.2 58.2 87.2
179 20.6 14.2 21.2
108 16.1 15.3 22.9





#40
Hexachlorocyclopentadiene
Concen: 20.180 ng/ul
RT: 12.751 min Scan# 1643
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Instrument :
BNA_P
ClientSampleId :
SLCS290




Tgt Ion:237 Resp: 18014
Ion Ratio Lower Upper
237 100
235 61.0 53.2 79.8
272 12.4 10.8 16.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#41
2,4,6-Trichlorophenol
Concen: 29.742 ng/ul
RT: 13.004 min Scan# 1686
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Tgt Ion:196 Resp: 304814
Ion Ratio Lower Upper
196 100
198 97.6 64.1 96.1#
200 31.0 19.9 29.9#





#42
2,4,5-Trichlorophenol
Concen: 29.355 ng/ul
RT: 13.081 min Scan# 1699
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Instrument :
BNA_P
ClientSampleId :
SLCS290




Tgt Ion:196 Resp: 327334
Ion Ratio Lower Upper
196 100
198 97.8 68.6 103.0
200 32.2 32.8 49.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#43
1,1'-Biphenyl
Concen: 30.171 ng/ul
RT: 13.404 min Scan# 1754
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Tgt Ion:154 Resp: 1140896
Ion Ratio Lower Upper
154 100
153 40.0 28.6 43.0
76 12.7 8.5 12.7





#44
2-Chloronaphthalene
Concen: 30.140 ng/ul
RT: 13.445 min Scan# 1761
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Instrument :
BNA_P
ClientSampleId :
SLCS290




Tgt Ion:162 Resp: 882015
Ion Ratio Lower Upper
162 100
164 31.8 27.5 41.3
127 40.2 37.1 55.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#45
2-Nitroaniline
Concen: 31.355 ng/ul
RT: 13.645 min Scan# 1795
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Tgt Ion: 65 Resp: 297525
Ion Ratio Lower Upper
65 100
92 66.1 77.9 116.9#
138 96.1 81.5 122.3





#47
Dimethylphthalate
Concen: 29.824 ng/ul
RT: 14.028 min Scan# 1860
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Instrument :
BNA_P
ClientSampleId :
SLCS290




Tgt Ion:163 Resp: 1155980
Ion Ratio Lower Upper
163 100
194 4.3 4.1 6.1
164 9.8 11.3 16.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#48
2,6-Dinitrotoluene
Concen: 31.247 ng/ul
RT: 14.140 min Scan# 1879
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Tgt Ion:165 Resp: 248552
Ion Ratio Lower Upper
165 100
89 59.0 51.7 77.5
121 21.0 18.3 27.5





#50
Acenaphthylene
Concen: 30.099 ng/ul
RT: 14.292 min Scan# 1905
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Instrument :
BNA_P
ClientSampleId :
SLCS290




Tgt Ion:152 Resp: 1439080
Ion Ratio Lower Upper
152 100
151 19.6 18.3 27.5
153 13.0 11.0 16.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#51
3-Nitroaniline
Concen: 32.600 ng/ul
RT: 14.469 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17



Tgt Ion:138 Resp: 234906
Ion Ratio Lower Upper
138 100
108 12.0 15.8 23.8#
92 124.3 117.1 175.7









#52
Acenaphthene
Concen: 29.740 ng/ul
RT: 14.634 min Scan# 1963
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17

Instrument :
BNA_P
ClientSampleId :
SLCS290

Tgt Ion:153 Resp: 92459
Ion Ratio Lower Upper
153 100
152 46.2 34.3 51.5
154 89.1 70.0 105.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#53
2,4-Dinitrophenol
Concen: 17.816 ng/ul
RT: 14.675 min Scan# 1970
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17


Tgt Ion:184 Resp: 79193
Ion Ratio Lower Upper
184 100
63 74.4 34.6 52.0#
154 70.0 57.7 86.5





#55
4-Nitrophenol
Concen: 24.933 ng/ul
RT: 14.781 min Scan# 1988
Delta R.T. 0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Instrument :
BNA_P
ClientSampleId :
SLCS290



Tgt Ion:109 Resp: 159040
Ion Ratio Lower Upper
109 100
139 121.3 61.4 92.2
65 132.0 65.3 97.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#56
Dibenzofuran
Concen: 30.084 ng/ul
RT: 14.969 min Scan# 2020
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17


Tgt Ion:168 Resp: 1384966
Ion Ratio Lower Upper
168 100
139 37.5 30.1 45.1





#57
2,4-Dinitrotoluene
Concen: 31.406 ng/ul
RT: 14.928 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Instrument :
BNA_P
ClientSampleId :
SLCS290




Tgt Ion:165 Resp: 36495
Ion Ratio Lower Upper
165 100
63 46.2 29.5 44.3
182 2.9 2.8 4.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 30.637 ng/ul
RT: 15.198 min Scan# 2059
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Tgt Ion:232 Resp: 303680
Ion Ratio Lower Upper
232 100
131 41.3 25.4 38.2#
130 2.2 4.2 6.2#
166 29.8 28.1 42.1





#59
Diethylphthalate
Concen: 29.791 ng/ul
RT: 15.392 min Scan# 2092
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Instrument :
BNA_P
ClientSampleId :
SLCS290




Tgt Ion:149 Resp: 117452
Ion Ratio Lower Upper
149 100
177 22.7 22.0 33.0
150 12.4 10.2 15.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#61
Fluorene
Concen: 29.674 ng/ul
RT: 15.622 min Scan# 2131
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Tgt Ion:166 Resp: 1121587
Ion Ratio Lower Upper
166 100
165 96.2 70.3 105.5
167 13.8 9.5 14.3





#62
4-Chlorophenyl-phenylether
Concen: 29.332 ng/ul
RT: 15.610 min Scan# 2130
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Instrument :
BNA_P
ClientSampleId :
SLCS290




Tgt Ion:204 Resp: 58347
Ion Ratio Lower Upper
204 100
206 33.3 27.6 41.4
141 60.9 46.6 69.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#63
4-Nitroaniline
Concen: 34.372 ng/ul
RT: 15.634 min Scan# 2133
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Tgt Ion:138 Resp: 241002
Ion Ratio Lower Upper
138 100
92 54.2 41.6 62.4
108 88.0 97.6 146.4





#66
4,6-Dinitro-2-methylphenol
Concen: 24.825 ng/ul
RT: 15.692 min Scan# 2144
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Instrument :
BNA_P
ClientSampleId :
SLCS290




Tgt Ion:198 Resp: 169360
Ion Ratio Lower Upper
198 100
182 4.3 6.0 9.0
77 29.2 17.7 26.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#67
N-Nitrosodiphenylamine
Concen: 30.925 ng/ul
RT: 15.822 min Scan# 2165
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Tgt Ion:169 Resp: 992280
Ion Ratio Lower Upper
169 100
168 65.3 48.3 72.5
167 34.8 29.6 44.4





#68
4-Bromophenyl-phenylether
Concen: 31.298 ng/ul
RT: 16.510 min Scan# 211
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Instrument :
BNA_P
ClientSampleId :
SLCS290




Tgt Ion:248 Resp: 37333
Ion Ratio Lower Upper
248 100
250 96.3 79.4 119.0
141 73.2 67.2 100.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#69
Hexachlorobenzene
Concen: 30.412 ng/ul
RT: 16.634 min Scan# 2303
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Tgt Ion:284 Resp: 421066
Ion Ratio Lower Upper
284 100
142 31.2 19.2 28.8#
249 31.9 27.4 41.2





#70
Atrazine
Concen: 29.417 ng/ul
RT: 16.781 min Scan# 2328
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Instrument :
BNA_P
ClientSampleId :
SLCS290




Tgt Ion:200 Resp: 383740
Ion Ratio Lower Upper
200 100
173 24.8 19.2 28.8
215 48.9 32.2 48.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#71
Pentachlorophenol
Concen: 27.388 ng/ul
RT: 16.975 min Scan# 2361
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17



Tgt Ion:266 Resp: 241493
Ion Ratio Lower Upper
266 100
264 61.5 57.3 85.9
268 62.7 65.7 98.5









#72
Phenanthrene
Concen: 30.565 ng/u1
RT: 17.363 min Scan# 2428
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17

Instrument :
BNA_P
ClientSampleId :
SLCS290

Tgt Ion:178 Resp: 1857811
Ion Ratio Lower Upper
178 100
179 14.9 11.7 17.5
176 18.5 14.5 21.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#74
Anthracene
Concen: 30.794 ng/u1
RT: 17.457 min Scan# 2443
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17


Tgt Ion:178 Resp: 1885246
Ion Ratio Lower Upper
178 100
179 15.4 12.3 18.5
176 17.8 13.4 20.2





#75
1,2,3,4-Tetrachlorobenzene
Concen: 30.708 ng/uL
RT: 13.375 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Instrument :
BNA_P
ClientSampleId :
SLCS290




Tgt Ion:216 Resp: 501339
Ion Ratio Lower Upper
216 100
214 78.6 50.4 75.6
218 49.4 35.2 52.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#76
Pentachlorobenzene
Concen: 29.811 ng/uL
RT: 14.892 min Scan# 2007
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Tgt Ion:250 Resp: 477254
Ion Ratio Lower Upper
250 100
248 63.4 43.5 65.3
252 64.8 46.5 69.7





#77
 Carbazole
 Concen: 30.295 ng/ul
 RT: 17.722 min Scan# 2489
 Delta R.T. -0.000 min
 Lab File: BP010076.D
 Acq: 23 Apr 2022 03:17




Instrument :
 BNA_P
 ClientSampleId :
 SLCS290




Tgt Ion:167 Resp: 164123
 Ion Ratio Lower Upper
 167 100
 166 21.6 15.8 23.8
 139 12.7 8.3 12.5

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
 Supervised By :mohammad ahmed 04/26/2022




#78
 Di-n-butylphthalate
 Concen: 31.325 ng/ul
 RT: 18.275 min Scan# 2582
 Delta R.T. -0.000 min
 Lab File: BP010076.D
 Acq: 23 Apr 2022 03:17



Tgt Ion:149 Resp: 2087867
 Ion Ratio Lower Upper
 149 100
 150 9.1 9.6 14.4#
 104 5.3 6.2 9.4#









#80
Fluoranthene
Concen: 34.945 ng/u1
RT: 19.327 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17

Instrument :
BNA_P
ClientSampleId :
SLCS290

Tgt Ion:202 Resp: 2281329
Ion Ratio Lower Upper
202 100
101 9.6 9.6 14.4
100 7.4 6.3 9.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#82
Pyrene
Concen: 34.308 ng/u1
RT: 19.680 min Scan# 2821
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17

Tgt Ion:202 Resp: 2307579
Ion Ratio Lower Upper
202 100
101 11.2 10.6 15.8
100 9.5 9.8 14.6#









#83
Butylbenzylphthalate
Concen: 34.718 ng/ul
RT: 20.545 min Scan# 2969
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17

Instrument :
BNA_P
ClientSampleId :
SLCS290

Tgt Ion:149 Resp: 965239
Ion Ratio Lower Upper
149 100
91 72.2 71.1 106.7
206 21.5 28.8 43.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#84
3,3'-Dichlorobenzidine
Concen: 26.701 ng/ul
RT: 21.316 min Scan# 3099
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17


Tgt Ion:252 Resp: 488953
Ion Ratio Lower Upper
252 100
254 64.6 49.8 74.6
126 9.7 4.3 6.5





#85
Benzo(a)anthracene
Concen: 31.970 ng/ul
RT: 21.386 min Scan# 3112
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Instrument :
BNA_P
ClientSampleId :
SLCS290



Tgt Ion:228 Resp: 2064860
Ion Ratio Lower Upper
228 100
229 19.5 18.2 27.2
226 26.2 22.8 34.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 34.418 ng/ul
RT: 21.310 min Scan# 3098
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17


Tgt Ion:149 Resp: 1418247
Ion Ratio Lower Upper
149 100
167 28.7 32.8 49.2#
279 5.0 5.6 8.4#





#87
Chrysene
Concen: 31.126 ng/ul
RT: 21.445 min Scan# 3121
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Instrument :
BNA_P
ClientSampleId :
SLCS290



Tgt Ion:228 Resp: 197954
Ion Ratio Lower Upper
228 100
226 29.4 22.7 34.1
229 19.4 15.2 22.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#89
Di-n-octyl phthalate
Concen: 35.243 ng/ul
RT: 22.251 min Scan# 3258
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17


Tgt Ion:149 Resp: 2382142





#90
Benzo(b)fluoranthene
Concen: 32.784 ng/ul
RT: 23.110 min Scan# 3404
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Instrument :
BNA_P
ClientSampleId :
SLCS290




Tgt Ion:252 Resp: 1916284
Ion Ratio Lower Upper
252 100
253 21.4 17.4 26.0
125 8.2 7.3 10.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022




#91
Benzo(k)fluoranthene
Concen: 31.160 ng/ul
RT: 23.157 min Scan# 3412
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Tgt Ion:252 Resp: 1725276
Ion Ratio Lower Upper
252 100
253 21.8 17.9 26.9
125 8.3 5.1 7.7#





#93
Benzo(a)pyrene
Concen: 31.339 ng/ul
RT: 23.751 min Scan# 3513
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Instrument :
BNA_P
ClientSampleId :
SLCS290



Tgt Ion:252 Resp: 1693959
Ion Ratio Lower Upper
252 100
253 21.3 17.5 26.3
125 8.6 6.4 9.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#94
Indeno(1,2,3-cd)pyrene
Concen: 30.552 ng/ul
RT: 26.421 min Scan# 3967
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17


Tgt Ion:276 Resp: 1642785
Ion Ratio Lower Upper
276 100
138 20.3 12.3 18.5#
277 25.2 19.3 28.9





#95
Dibenzo(a,h)anthracene
Concen: 29.858 ng/ul
RT: 26.439 min Scan# 3971
Delta R.T. 0.000 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17




Instrument :
BNA_P
ClientSampleId :
SLCS290




Tgt Ion:278 Resp: 1373758
Ion Ratio Lower Upper
278 100
139 13.0 6.6 10.0#
279 23.1 18.4 27.6

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 04/25/2022
Supervised By :mohammad ahmed 04/26/2022



#96
Benzo(g,h,i)perylene
Concen: 31.723 ng/ul
RT: 27.204 min Scan# 4100
Delta R.T. -0.006 min
Lab File: BP010076.D
Acq: 23 Apr 2022 03:17



Tgt Ion:276 Resp: 1439296
Ion Ratio Lower Upper
276 100
138 18.0 11.6 17.4#
277 24.2 15.9 23.9#

