

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042722\
 Data File : BP010122.D
 Acq On : 27 Apr 2022 15:19
 Operator : CG/JU
 Sample : PB144346BS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS346

Quant Time: Apr 28 03:25:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 14:49:38 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 04/28/2022
 Supervised By :Yogesh Patel 05/05/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.940	152	189659	20.000	ng/u1	0.00
20) Naphthalene-d8	10.746	136	807117	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.575	164	517220	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.328	188	1048143	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.416	240	813456	20.000	ng/u1	# 0.00
88) Perylene-d12	23.874	264	600303	20.000	ng/u1	0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.370	96	28260	6.498	ng/uL	0.00
4) Pyridine-d5	3.781	84	375171	30.663	ng/u1	0.00
7) Phenol-d5	7.105	99	556171	32.239	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.263	67	358005	34.585	ng/u1	0.00
11) 2-Chlorophenol-d4	7.469	132	455886	35.683	ng/u1	0.00
15) 4-Methylphenol-d8	8.652	113	469908	32.877	ng/u1	0.00
21) Nitrobenzene-d5	9.099	128	230819	35.957	ng/u1	0.00
24) 2-Nitrophenol-d4	9.828	143	256542	36.592	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.369	165	458494	33.593	ng/u1	0.00
31) 4-Chloroaniline-d4	10.881	131	540003	27.594	ng/u1	0.00
46) Dimethylphthalate-d6	13.981	166	1396302	34.705	ng/u1	0.00
49) Acenaphthylene-d8	14.269	160	1628055	33.434	ng/u1	0.00
54) 4-Nitrophenol-d4	14.769	143	231795	31.294	ng/u1	0.00
60) Fluorene-d10	15.569	176	1178208	34.043	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.692	200	223830	33.773	ng/u1	0.01
73) Anthracene-d10	17.428	188	1764948	34.806	ng/u1	0.00
81) Pyrene-d10	19.663	212	1987574	42.964	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.716	264	1199436	38.112	ng/u1	0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.405	88	53743	11.891	ng/uL#	14
5) Pyridine	3.799	79	394945	31.053	ng/u1#	42
6) Benzaldehyde	7.075	77	382914	41.535	ng/u1	96
8) Phenol	7.128	94	569216	32.839	ng/u1#	94
10) Bis(2-Chloroethyl)ether	7.358	93	450417	33.727	ng/u1#	79
12) 2-Chlorophenol	7.505	128	443823	34.440	ng/u1	93
13) 2-Methylphenol	8.381	108	431922	32.490	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.475	45	661845	33.399	ng/u1#	86
16) Acetophenone	8.763	105	703674	31.555	ng/u1#	81
17) N-Nitroso-di-n-propyla...	8.757	70	374047	32.015	ng/u1#	75
18) 4-Methylphenol	8.716	108	470420	32.139	ng/u1	93
19) Hexachloroethane	9.022	117	186405	34.423	ng/u1#	80
22) Nitrobenzene	9.140	77	537143	34.152	ng/u1#	86
23) Isophorone	9.675	82	1033905	33.282	ng/u1#	96
25) 2-Nitrophenol	9.857	139	264100	35.568	ng/u1#	82
26) 2,4-Dimethylphenol	9.916	107	525878	32.712	ng/u1#	78
27) Bis(2-Chloroethoxy)met...	10.152	93	617432	33.910	ng/u1#	96
29) 2,4-Dichlorophenol	10.393	162	446437	34.028	ng/u1#	83
30) Naphthalene	10.799	128	1438679	34.057	ng/u1	97
32) 4-Chloroaniline	10.904	127	519506	26.939	ng/u1	98
33) Hexachlorobutadiene	11.093	225	314173	33.788	ng/u1	96
34) Caprolactam	11.681	113	148637m	31.495	ng/u1	
35) 4-Chloro-3-methylphenol	12.028	107	505950	31.923	ng/u1#	70
36) 2-Methylnaphthalene	12.404	142	991335	32.357	ng/u1	91

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042722\
 Data File : BP010122.D
 Acq On : 27 Apr 2022 15:19
 Operator : CG/JU
 Sample : PB144346BS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS346

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 04/28/2022
 Supervised By :Yogesh Patel 05/05/2022

Quant Time: Apr 28 03:25:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 14:49:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.622	142	1011956	32.530	ng/ul#	97
39) 1,2,4,5-Tetrachloroben...	12.775	216	573612	35.896	ng/ul	93
40) Hexachlorocyclopentadiene	12.757	237	313752	34.840	ng/ul	97
41) 2,4,6-Trichlorophenol	13.010	196	368041	35.597	ng/ul#	83
42) 2,4,5-Trichlorophenol	13.087	196	409914m	36.439	ng/ul	
43) 1,1'-Biphenyl	13.410	154	1341954	35.178	ng/ul	93
44) 2-Chloronaphthalene	13.451	162	1036840	35.122	ng/ul	94
45) 2-Nitroaniline	13.651	65	331473	34.627	ng/ul#	84
47) Dimethylphthalate	14.028	163	1320851	33.779	ng/ul#	92
48) 2,6-Dinitrotoluene	14.145	165	283823	35.369	ng/ul	92
50) Acenaphthylene	14.298	152	1666004	34.540	ng/ul	95
51) 3-Nitroaniline	14.475	138	247611	34.063	ng/ul#	81
52) Acenaphthene	14.640	153	1066784	34.014	ng/ul	97
53) 2,4-Dinitrophenol	14.687	184	130831	29.176	ng/ul#	80
55) 4-Nitrophenol	14.787	109	189831	29.499	ng/ul#	44
56) Dibenzofuran	14.975	168	1558190	33.551	ng/ul	99
57) 2,4-Dinitrotoluene	14.940	165	399565	34.084	ng/ul	89
58) 2,3,4,6-Tetrachlorophenol	15.204	232	343720	34.374	ng/ul#	87
59) Diethylphthalate	15.398	149	1341737	33.735	ng/ul#	93
61) Fluorene	15.628	166	1261936	33.095	ng/ul	93
62) 4-Chlorophenyl-phenyle...	15.616	204	666118	33.194	ng/ul	97
63) 4-Nitroaniline	15.639	138	278964m	39.439	ng/ul	
66) 4,6-Dinitro-2-methylph...	15.704	198	213128	33.331	ng/ul#	94
67) N-Nitrosodiphenylamine	15.834	169	1085168	36.083	ng/ul	94
68) 4-Bromophenyl-phenylether	16.516	248	410910	36.754	ng/ul	94
69) Hexachlorobenzene	16.639	284	465007	35.833	ng/ul#	90
70) Atrazine	16.792	200	439028	35.907	ng/ul#	89
71) Pentachlorophenol	16.981	266	286302	34.643	ng/ul#	84
72) Phenanthrene	17.375	178	1994281	35.006	ng/ul	100
74) Anthracene	17.469	178	1987393	34.635	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.381	216	585810	38.284	ng/uL#	87
76) Pentachlorobenzene	14.898	250	549682	36.633	ng/uL	90
77) Carbazole	17.733	167	1722129	33.916	ng/ul#	96
78) Di-n-butylphthalate	18.281	149	2287120	36.611	ng/ul#	93
80) Fluoranthene	19.339	202	2257322	43.260	ng/ul	96
82) Pyrene	19.692	202	2271487	42.253	ng/ul#	94
83) Butylbenzylphthalate	20.557	149	966166	43.478	ng/ul#	82
84) 3,3'-Dichlorobenzidine	21.327	252	565122	38.610	ng/ul#	96
85) Benzo(a)anthracene	21.398	228	1863098	36.090	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.322	149	1384506	42.037	ng/ul#	83
87) Chrysene	21.457	228	1784789	35.112	ng/ul	99
89) Di-n-octyl phthalate	22.269	149	2138443	45.823	ng/ul	100
90) Benzo(b)fluoranthene	23.121	252	1566178	38.808	ng/ul	99
91) Benzo(k)fluoranthene	23.174	252	1410757	36.904	ng/ul#	98
93) Benzo(a)pyrene	23.768	252	1393105	37.329	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.445	276	1504710	40.531	ng/ul#	95
95) Dibenzo(a,h)anthracene	26.462	278	1275413	40.149	ng/ul#	95
96) Benzo(g,h,i)perylene	27.233	276	1295804	41.366	ng/ul#	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042722\
 Data File : BP010122.D
 Acq On : 27 Apr 2022 15:19
 Operator : CG/JU
 Sample : PB144346BS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :

BNA_P

ClientSampleId :


SLCS346


Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 04/28/2022
 Supervised By :Yogesh Patel 05/05/2022


Quant Time: Apr 28 03:25:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 14:49:38 2022
 Response via : Initial Calibration





#2
1,4-Dioxane
Concen: 11.891 ng/uL
RT: 3.405 min Scan# 54
Delta R.T. -0.000 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19




Instrument :
BNA_P
ClientSampleId :
SLCS346




Tgt Ion: 88 Resp: 5374
Ion Ratio Lower Upper
88 100
43 38.7 6.8 10.2
58 85.8 28.4 42.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#5
Pyridine
Concen: 31.053 ng/uL
RT: 3.799 min Scan# 121
Delta R.T. -0.006 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19



Tgt Ion: 79 Resp: 394945
Ion Ratio Lower Upper
79 100
52 91.4 35.5 53.3#
51 41.7 20.2 30.4#









#6
Benzaldehyde
Concen: 41.535 ng/ul
RT: 7.075 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19

Instrument :
BNA_P
ClientSampleId :
SLCS346

Tgt Ion: 77 Resp: 382914
Ion Ratio Lower Upper
77 100
105 94.9 77.0 115.4
106 89.3 77.0 115.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#8
Phenol
Concen: 32.839 ng/ul
RT: 7.128 min Scan# 687
Delta R.T. 0.000 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19


Tgt Ion: 94 Resp: 569216
Ion Ratio Lower Upper
94 100
65 31.6 19.0 28.6#
66 42.4 34.6 51.8





#10
Bis(2-Chloroethyl)ether
Concen: 33.727 ng/ul
RT: 7.358 min Scan# 701
Delta R.T. 0.000 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19




Instrument :
BNA_P
ClientSampleId :
SLCS346




Tgt Ion: 93 Resp: 45041
Ion Ratio Lower Upper
93 100
63 80.1 45.0 67.4
95 31.9 26.3 39.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#12
2-Chlorophenol
Concen: 34.440 ng/ul
RT: 7.505 min Scan# 751
Delta R.T. 0.000 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19



Tgt Ion:128 Resp: 443823
Ion Ratio Lower Upper
128 100
64 51.0 45.3 67.9
130 32.2 29.0 43.6









#13
2-Methylphenol
Concen: 32.490 ng/ul
RT: 8.381 min Scan# 900
Delta R.T. 0.000 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19

Instrument :
BNA_P
ClientSampleId :
SLCS346

Tgt Ion:108 Resp: 43192
Ion Ratio Lower Upper
108 100
107 90.2 70.8 106.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#14
2,2'-oxybis(1-Chloropropane)
Concen: 33.399 ng/ul
RT: 8.475 min Scan# 916
Delta R.T. 0.006 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19

Tgt Ion: 45 Resp: 661845
Ion Ratio Lower Upper
45 100
77 14.3 17.1 25.7#
79 10.6 13.2 19.8#









#16
Acetophenone
Concen: 31.555 ng/ul
RT: 8.763 min Scan# 90
Delta R.T. 0.000 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19

Instrument :
BNA_P
ClientSampleId :
SLCS346

Tgt Ion:105 Resp: 703674
Ion Ratio Lower Upper
105 100
77 79.7 53.4 80.2
51 29.9 13.8 20.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#17
N-Nitroso-di-n-propylamine
Concen: 32.015 ng/ul
RT: 8.757 min Scan# 964
Delta R.T. 0.006 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19


Tgt Ion: 70 Resp: 374047
Ion Ratio Lower Upper
70 100
42 61.3 31.8 47.8#
101 10.9 15.6 23.4#
130 21.3 13.7 20.5#





#18
4-Methylphenol
Concen: 32.139 ng/ul
RT: 8.716 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19




Instrument :
BNA_P
ClientSampleId :
SLCS346




Tgt Ion:108 Resp: 470420
Ion Ratio Lower Upper
108 100
107 111.2 83.6 125.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#19
Hexachloroethane
Concen: 34.423 ng/ul
RT: 9.022 min Scan# 1009
Delta R.T. -0.006 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19



Tgt Ion:117 Resp: 186405
Ion Ratio Lower Upper
117 100
201 103.3 106.5 159.7#
199 65.2 59.0 88.6









#22
Nitrobenzene
Concen: 34.152 ng/ul
RT: 9.140 min Scan# 1028
Delta R.T. 0.000 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19

Instrument :
BNA_P
ClientSampleId :
SLCS346

Tgt Ion: 77 Resp: 53714
Ion Ratio Lower Upper
77 100
123 43.2 42.3 63.5
65 13.9 5.8 8.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#23
Isophorone
Concen: 33.282 ng/ul
RT: 9.675 min Scan# 1120
Delta R.T. 0.006 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19

Tgt Ion: 82 Resp: 1033905
Ion Ratio Lower Upper
82 100
95 7.6 7.8 11.8#
138 17.1 12.9 19.3









#25
2-Nitrophenol
Concen: 35.568 ng/ul
RT: 9.857 min Scan# 1149
Delta R.T. 0.000 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19

Instrument :
BNA_P
ClientSampleId :
SLCS346

Tgt Ion:139 Resp: 264100
Ion Ratio Lower Upper
139 100
65 57.2 49.4 74.2
109 33.5 44.7 67.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#26
2,4-Dimethylphenol
Concen: 32.712 ng/ul
RT: 9.916 min Scan# 1161
Delta R.T. 0.000 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19

Tgt Ion:107 Resp: 525878
Ion Ratio Lower Upper
107 100
121 48.0 34.1 51.1
122 81.6 46.3 69.5









#27
Bis(2-Chloroethoxy)methane
Concen: 33.910 ng/ul
RT: 10.152 min Scan# 1201
Delta R.T. 0.000 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19

Instrument :
BNA_P
ClientSampleId :
SLCS346

Tgt Ion: 93 Resp: 61743
Ion Ratio Lower Upper
93 100
95 31.3 26.2 39.4
123 10.7 6.2 9.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#29
2,4-Dichlorophenol
Concen: 34.028 ng/ul
RT: 10.393 min Scan# 1242
Delta R.T. 0.000 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19


Tgt Ion:162 Resp: 446437
Ion Ratio Lower Upper
162 100
164 66.7 37.0 55.6#
98 38.3 31.0 46.6





#30
Naphthalene
Concen: 34.057 ng/ul
RT: 10.799 min Scan# 1311
Delta R.T. 0.000 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19




Instrument :
BNA_P
ClientSampleId :
SLCS346



Tgt Ion:128 Resp: 1438679
Ion Ratio Lower Upper
128 100
129 11.1 8.6 13.0
127 13.4 12.2 18.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#32
4-Chloroaniline
Concen: 26.939 ng/ul
RT: 10.904 min Scan# 1329
Delta R.T. 0.000 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19

Tgt Ion:127 Resp: 519506
Ion Ratio Lower Upper
127 100
129 31.9 24.6 37.0









#33
Hexachlorobutadiene
Concen: 33.788 ng/ul
RT: 11.093 min Scan# 1361
Delta R.T. 0.000 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19

Instrument :
BNA_P
ClientSampleId :
SLCS346

Tgt Ion:225 Resp: 31417
Ion Ratio Lower Upper
225 100
223 63.0 53.4 80.0
227 63.8 53.8 80.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#34
Caprolactam
Concen: 31.495 ng/ul m
RT: 11.681 min Scan# 1461
Delta R.T. 0.018 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19

Tgt Ion:113 Resp: 148637
Ion Ratio Lower Upper
113 100
55 180.9 58.2 87.2#
56 137.4 68.6 103.0#









#35
4-Chloro-3-methylphenol
Concen: 31.923 ng/ul
RT: 12.028 min Scan# 1518
Delta R.T. 0.006 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19

Instrument :
BNA_P
ClientSampleId :
SLCS346

Tgt Ion:107 Resp: 505950
Ion Ratio Lower Upper
107 100
144 24.7 15.8 23.8
142 74.7 38.9 58.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#36
2-Methylnaphthalene
Concen: 32.357 ng/ul
RT: 12.404 min Scan# 1584
Delta R.T. 0.000 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19


Tgt Ion:142 Resp: 991335
Ion Ratio Lower Upper
142 100
141 88.2 64.3 96.5





#37
1-Methylnaphthalene
Concen: 32.530 ng/ul
RT: 12.622 min Scan# 1621
Delta R.T. 0.000 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19




Instrument :
BNA_P
ClientSampleId :
SLCS346




Tgt Ion:142 Resp: 1011950
Ion Ratio Lower Upper
142 100
141 89.6 69.8 104.8
116 4.0 5.6 8.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 35.896 ng/ul
RT: 12.775 min Scan# 1647
Delta R.T. 0.000 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19




Tgt Ion:216 Resp: 573612
Ion Ratio Lower Upper
216 100
214 78.8 58.2 87.2
179 20.3 14.2 21.2
108 16.0 15.3 22.9





#40
Hexachlorocyclopentadiene
Concen: 34.840 ng/ul
RT: 12.757 min Scan# 1644
Delta R.T. 0.000 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19




Instrument :
BNA_P
ClientSampleId :
SLCS346




Tgt Ion:237 Resp: 313752
Ion Ratio Lower Upper
237 100
235 63.8 53.2 79.8
272 12.8 10.8 16.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#41
2,4,6-Trichlorophenol
Concen: 35.597 ng/ul
RT: 13.010 min Scan# 1687
Delta R.T. 0.000 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19



Tgt Ion:196 Resp: 368041
Ion Ratio Lower Upper
196 100
198 96.3 64.1 96.1#
200 30.3 19.9 29.9#









#42
2,4,5-Trichlorophenol
Concen: 36.439 ng/ul m
RT: 13.087 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19

Instrument :
BNA_P
ClientSampleId :
SLCS346

Tgt Ion:196 Resp: 409914
Ion Ratio Lower Upper
196 100
198 98.4 68.6 103.0
200 31.8 32.8 49.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#43
1,1'-Biphenyl
Concen: 35.178 ng/ul
RT: 13.410 min Scan# 1755
Delta R.T. 0.000 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19


Tgt Ion:154 Resp: 1341954
Ion Ratio Lower Upper
154 100
153 40.0 28.6 43.0
76 12.2 8.5 12.7





#44
2-Chloronaphthalene
Concen: 35.122 ng/ul
RT: 13.451 min Scan# 1760
Delta R.T. 0.000 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19




Instrument :
BNA_P
ClientSampleId :
SLCS346



Tgt Ion:162 Resp: 1036840
Ion Ratio Lower Upper
162 100
164 33.3 27.5 41.3
127 40.2 37.1 55.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#45
2-Nitroaniline
Concen: 34.627 ng/ul
RT: 13.651 min Scan# 1796
Delta R.T. 0.006 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19

Tgt Ion: 65 Resp: 331473
Ion Ratio Lower Upper
65 100
92 66.4 77.9 116.9#
138 99.8 81.5 122.3









#47
Dimethylphthalate
Concen: 33.779 ng/ul
RT: 14.028 min Scan# 1859
Delta R.T. 0.000 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19

Instrument :
BNA_P
ClientSampleId :
SLCS346

Tgt Ion:163 Resp: 132085
Ion Ratio Lower Upper
163 100
194 4.2 4.1 6.1
164 10.3 11.3 16.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#48
2,6-Dinitrotoluene
Concen: 35.369 ng/ul
RT: 14.145 min Scan# 1880
Delta R.T. 0.000 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19


Tgt Ion:165 Resp: 283823
Ion Ratio Lower Upper
165 100
89 57.8 51.7 77.5
121 20.5 18.3 27.5





#50
Acenaphthylene
Concen: 34.540 ng/ul
RT: 14.298 min Scan# 1906
Delta R.T. 0.006 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19




Instrument :
BNA_P
ClientSampleId :
SLCS346




Tgt Ion:152 Resp: 1666004
Ion Ratio Lower Upper
152 100
151 19.5 18.3 27.5
153 13.1 11.0 16.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#51
3-Nitroaniline
Concen: 34.063 ng/ul
RT: 14.475 min Scan# 1936
Delta R.T. 0.006 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19




Tgt Ion:138 Resp: 247611
Ion Ratio Lower Upper
138 100
108 11.7 15.8 23.8#
92 122.0 117.1 175.7





#52
Acenaphthene
Concen: 34.014 ng/ul
RT: 14.640 min Scan# 1964
Delta R.T. 0.000 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19




Instrument :
BNA_P
ClientSampleId :
SLCS346




Tgt Ion:153 Resp: 1066784
Ion Ratio Lower Upper
153 100
152 46.2 34.3 51.5
154 89.8 70.0 105.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#53
2,4-Dinitrophenol
Concen: 29.176 ng/ul
RT: 14.687 min Scan# 1972
Delta R.T. 0.012 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19




Tgt Ion:184 Resp: 130831
Ion Ratio Lower Upper
184 100
63 65.2 34.6 52.0#
154 62.3 57.7 86.5





#55
4-Nitrophenol
Concen: 29.499 ng/u1
RT: 14.787 min Scan# 1986
Delta R.T. 0.012 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19




Instrument :
BNA_P
ClientSampleId :
SLCS346



Tgt Ion:109 Resp: 18983
Ion Ratio Lower Upper
109 100
139 121.8 61.4 92.2
65 135.2 65.3 97.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#56
Dibenzofuran
Concen: 33.551 ng/u1
RT: 14.975 min Scan# 2021
Delta R.T. 0.006 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19

Tgt Ion:168 Resp: 1558190
Ion Ratio Lower Upper
168 100
139 37.3 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 34.084 ng/ul
RT: 14.940 min Scan# 2015
Delta R.T. 0.012 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19

Instrument :
BNA_P
ClientSampleId :
SLCS346

Tgt Ion:165 Resp: 399568
Ion Ratio Lower Upper
165 100
63 44.2 29.5 44.3
182 3.0 2.8 4.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#58
2,3,4,6-Tetrachlorophenol
Concen: 34.374 ng/ul
RT: 15.204 min Scan# 2060
Delta R.T. 0.006 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19

Tgt Ion:232 Resp: 343720
Ion Ratio Lower Upper
232 100
131 42.1 25.4 38.2#
130 1.9 4.2 6.2#
166 29.7 28.1 42.1









#59
Diethylphthalate
Concen: 33.735 ng/ul
RT: 15.398 min Scan# 2093
Delta R.T. 0.006 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19

Instrument :
BNA_P
ClientSampleId :
SLCS346

Tgt Ion:149 Resp: 134173
Ion Ratio Lower Upper
149 100
177 21.9 22.0 33.0
150 12.5 10.2 15.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#61
Fluorene
Concen: 33.095 ng/ul
RT: 15.628 min Scan# 2132
Delta R.T. 0.006 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19

Tgt Ion:166 Resp: 1261936
Ion Ratio Lower Upper
166 100
165 94.6 70.3 105.5
167 13.2 9.5 14.3









#62
4-Chlorophenyl-phenylether
Concen: 33.194 ng/ul
RT: 15.616 min Scan# 2129
Delta R.T. 0.000 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19

Instrument :
BNA_P
ClientSampleId :
SLCS346

Tgt Ion:204 Resp: 666118
Ion Ratio Lower Upper
204 100
206 33.3 27.6 41.4
141 61.2 46.6 69.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#63
4-Nitroaniline
Concen: 39.439 ng/ul m
RT: 15.639 min Scan# 2134
Delta R.T. 0.006 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19

Tgt Ion:138 Resp: 278964
Ion Ratio Lower Upper
138 100
92 55.8 41.6 62.4
108 83.8 97.6 146.4#









#66
4,6-Dinitro-2-methylphenol
Concen: 33.331 ng/ul
RT: 15.704 min Scan# 2143
Delta R.T. 0.006 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19

Instrument :
BNA_P
ClientSampleId :
SLCS346

Tgt Ion:198 Resp: 213128
Ion Ratio Lower Upper
198 100
182 3.7 6.0 9.0
77 24.5 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#67
N-Nitrosodiphenylamine
Concen: 36.083 ng/ul
RT: 15.834 min Scan# 2167
Delta R.T. 0.006 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19


Tgt Ion:169 Resp: 1085168
Ion Ratio Lower Upper
169 100
168 65.7 48.3 72.5
167 34.7 29.6 44.4





#68
4-Bromophenyl-phenylether
Concen: 36.754 ng/ul
RT: 16.516 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19




Instrument :
BNA_P
ClientSampleId :
SLCS346




Tgt Ion:248 Resp: 410910
Ion Ratio Lower Upper
248 100
250 98.6 79.4 119.0
141 72.8 67.2 100.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#69
Hexachlorobenzene
Concen: 35.833 ng/ul
RT: 16.639 min Scan# 2304
Delta R.T. 0.006 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19



Tgt Ion:284 Resp: 465007
Ion Ratio Lower Upper
284 100
142 31.9 19.2 28.8#
249 31.2 27.4 41.2









#70
Atrazine
Concen: 35.907 ng/ul
RT: 16.792 min Scan# 21
Delta R.T. 0.012 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19

Instrument :
BNA_P
ClientSampleId :
SLCS346

Tgt Ion:200 Resp: 439028
Ion Ratio Lower Upper
200 100
173 25.5 19.2 28.8
215 49.7 32.2 48.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#71
Pentachlorophenol
Concen: 34.643 ng/ul
RT: 16.981 min Scan# 2362
Delta R.T. 0.006 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19


Tgt Ion:266 Resp: 286302
Ion Ratio Lower Upper
266 100
264 63.1 57.3 85.9
268 62.7 65.7 98.5





#72
Phenanthrene
Concen: 35.006 ng/u1
RT: 17.375 min Scan# 2428
Delta R.T. 0.006 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19




Instrument :
BNA_P
ClientSampleId :
SLCS346



Tgt Ion:178 Resp: 199428
Ion Ratio Lower Upper
178 100
179 14.8 11.7 17.5
176 18.2 14.5 21.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#74
Anthracene
Concen: 34.635 ng/u1
RT: 17.469 min Scan# 2445
Delta R.T. 0.012 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19

Tgt Ion:178 Resp: 1987393
Ion Ratio Lower Upper
178 100
179 15.1 12.3 18.5
176 17.7 13.4 20.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 38.284 ng/uL
RT: 13.381 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19

Instrument :
BNA_P
ClientSampleId :
SLCS346

Tgt Ion:216 Resp: 585810
Ion Ratio Lower Upper
216 100
214 77.0 50.4 75.6
218 47.6 35.2 52.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#76
Pentachlorobenzene
Concen: 36.633 ng/uL
RT: 14.898 min Scan# 2008
Delta R.T. 0.000 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19


Tgt Ion:250 Resp: 549682
Ion Ratio Lower Upper
250 100
248 63.3 43.5 65.3
252 63.8 46.5 69.7





#77
 Carbazole
 Concen: 33.916 ng/ul
 RT: 17.733 min Scan# 2488
 Delta R.T. 0.011 min
 Lab File: BP010122.D
 Acq: 27 Apr 2022 15:19




Instrument :
 BNA_P
 ClientSampleId :
 SLCS346




Tgt Ion:167 Resp: 1722129
 Ion Ratio Lower Upper
 167 100
 166 21.2 15.8 23.8
 139 12.5 8.3 12.5

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
 Supervised By :Yogesh Patel 05/05/2022




#78
 Di-n-butylphthalate
 Concen: 36.611 ng/ul
 RT: 18.281 min Scan# 2583
 Delta R.T. 0.006 min
 Lab File: BP010122.D
 Acq: 27 Apr 2022 15:19




Tgt Ion:149 Resp: 2287120
 Ion Ratio Lower Upper
 149 100
 150 9.1 9.6 14.4#
 104 5.3 6.2 9.4#





#80
Fluoranthene
Concen: 43.260 ng/u1
RT: 19.339 min Scan# 21
Delta R.T. 0.012 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19




Instrument :
BNA_P
ClientSampleId :
SLCS346




Tgt Ion:202 Resp: 225732
Ion Ratio Lower Upper
202 100
101 9.8 9.6 14.4
100 7.4 6.3 9.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#82
Pyrene
Concen: 42.253 ng/u1
RT: 19.692 min Scan# 2823
Delta R.T. 0.012 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19



Tgt Ion:202 Resp: 2271487
Ion Ratio Lower Upper
202 100
101 11.3 10.6 15.8
100 9.4 9.8 14.6#









#83
Butylbenzylphthalate
Concen: 43.478 ng/ul
RT: 20.557 min Scan# 2968
Delta R.T. 0.006 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19

Instrument :
BNA_P
ClientSampleId :
SLCS346

Tgt Ion:149 Resp: 966160
Ion Ratio Lower Upper
149 100
91 74.1 71.1 106.7
206 22.8 28.8 43.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#84
3,3'-Dichlorobenzidine
Concen: 38.610 ng/ul
RT: 21.327 min Scan# 3101
Delta R.T. 0.006 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19


Tgt Ion:252 Resp: 565122
Ion Ratio Lower Upper
252 100
254 64.7 49.8 74.6
126 9.5 4.3 6.5





#85
Benzo(a)anthracene
Concen: 36.090 ng/ul
RT: 21.398 min Scan# 3111
Delta R.T. 0.006 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19




Instrument :
BNA_P
ClientSampleId :
SLCS346



Tgt Ion:228 Resp: 1863098
Ion Ratio Lower Upper
228 100
229 19.8 18.2 27.2
226 26.5 22.8 34.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 42.037 ng/ul
RT: 21.322 min Scan# 3100
Delta R.T. 0.006 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19


Tgt Ion:149 Resp: 1384506
Ion Ratio Lower Upper
149 100
167 28.9 32.8 49.2#
279 5.4 5.6 8.4#





#87
Chrysene
Concen: 35.112 ng/ul
RT: 21.457 min Scan# 3120
Delta R.T. 0.012 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19




Instrument :
BNA_P
ClientSampleId :
SLCS346



Tgt Ion:228 Resp: 1784789
Ion Ratio Lower Upper
228 100
226 28.8 22.7 34.1
229 20.1 15.2 22.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#89
Di-n-octyl phthalate
Concen: 45.823 ng/ul
RT: 22.269 min Scan# 3261
Delta R.T. 0.012 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19


Tgt Ion:149 Resp: 2138443





#90
Benzo(b)fluoranthene
Concen: 38.808 ng/ul
RT: 23.121 min Scan# 3406
Delta R.T. 0.012 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19




Instrument :
BNA_P
ClientSampleId :
SLCS346




Tgt Ion:252 Resp: 1566178
Ion Ratio Lower Upper
252 100
253 21.7 17.4 26.0
125 8.4 7.3 10.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#91
Benzo(k)fluoranthene
Concen: 36.904 ng/ul
RT: 23.174 min Scan# 3415
Delta R.T. 0.018 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19



Tgt Ion:252 Resp: 1410757
Ion Ratio Lower Upper
252 100
253 21.7 17.9 26.9
125 8.3 5.1 7.7#









#93
Benzo(a)pyrene
Concen: 37.329 ng/ul
RT: 23.768 min Scan# 3516
Delta R.T. 0.018 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19

Instrument :
BNA_P
ClientSampleId :
SLCS346

Tgt Ion:252 Resp: 139310
Ion Ratio Lower Upper
252 100
253 21.7 17.5 26.3
125 9.2 6.4 9.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#94
Indeno(1,2,3-cd)pyrene
Concen: 40.531 ng/ul
RT: 26.445 min Scan# 3971
Delta R.T. 0.018 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19


Tgt Ion:276 Resp: 1504710
Ion Ratio Lower Upper
276 100
138 19.4 12.3 18.5#
277 24.7 19.3 28.9





#95
Dibenzo(a,h)anthracene
Concen: 40.149 ng/ul
RT: 26.462 min Scan# 3969
Delta R.T. 0.024 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19




Instrument :
BNA_P
ClientSampleId :
SLCS346




Tgt Ion:278 Resp: 127541
Ion Ratio Lower Upper
278 100
139 13.4 6.6 10.0
279 23.7 18.4 27.6

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#96
Benzo(g,h,i)perylene
Concen: 41.366 ng/ul
RT: 27.233 min Scan# 4105
Delta R.T. 0.024 min
Lab File: BP010122.D
Acq: 27 Apr 2022 15:19



Tgt Ion:276 Resp: 1295804
Ion Ratio Lower Upper
276 100
138 17.7 11.6 17.4
277 23.4 15.9 23.9

