

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042722\
 Data File : BP010128.D
 Acq On : 27 Apr 2022 20:15
 Operator : CG/JU
 Sample : PB144421BS
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS421

Quant Time: Apr 28 03:26:43 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 14:49:38 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 04/28/2022
 Supervised By :Yogesh Patel 05/05/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.934	152	179527	20.000	ng/u1	0.00
20) Naphthalene-d8	10.734	136	763059	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.563	164	506218	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.322	188	1034785	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.404	240	809215	20.000	ng/u1	# 0.00
88) Perylene-d12	23.857	264	602536	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.363	96	21128	5.132	ng/uL	0.00
4) Pyridine-d5	3.781	84	285931	24.688	ng/u1	0.00
7) Phenol-d5	7.093	99	441377	27.029	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.257	67	287858	29.378	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.463	132	355233	29.374	ng/u1	0.00
15) 4-Methylphenol-d8	8.640	113	379834	28.074	ng/u1	-0.01
21) Nitrobenzene-d5	9.093	128	183477	30.232	ng/u1	0.00
24) 2-Nitrophenol-d4	9.816	143	208242	31.418	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.357	165	371571	28.797	ng/u1	0.00
31) 4-Chloroaniline-d4	10.869	131	481546	26.028	ng/u1	-0.01
46) Dimethylphthalate-d6	13.975	166	1222897	31.055	ng/u1	0.00
49) Acenaphthylene-d8	14.257	160	1381421	28.986	ng/u1	0.00
54) 4-Nitrophenol-d4	14.763	143	203352m	28.051	ng/u1	0.00
60) Fluorene-d10	15.557	176	1025002	30.260	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.680	200	209017	31.945	ng/u1	0.00
73) Anthracene-d10	17.422	188	1613415	32.228	ng/u1	0.00
81) Pyrene-d10	19.651	212	1801919	39.155	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.698	264	1103885	34.946	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.399	88	43304	10.122	ng/uL#	28
5) Pyridine	3.799	79	300389	24.951	ng/u1#	45
6) Benzaldehyde	7.069	77	296400	33.965	ng/u1	94
8) Phenol	7.122	94	443221	27.013	ng/u1#	93
10) Bis(2-Chloroethyl)ether	7.352	93	358456	28.356	ng/u1#	79
12) 2-Chlorophenol	7.493	128	347679	28.502	ng/u1	94
13) 2-Methylphenol	8.375	108	341096	27.106	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.463	45	530702	28.293	ng/u1#	84
16) Acetophenone	8.757	105	562250	26.636	ng/u1#	82
17) N-Nitroso-di-n-propyla...	8.746	70	302161	27.321	ng/u1#	76
18) 4-Methylphenol	8.704	108	373352	26.947	ng/u1	92
19) Hexachloroethane	9.016	117	149966	29.257	ng/u1	81
22) Nitrobenzene	9.134	77	434822	29.243	ng/u1#	85
23) Isophorone	9.663	82	833828	28.391	ng/u1	97
25) 2-Nitrophenol	9.845	139	206114	29.361	ng/u1#	83
26) 2,4-Dimethylphenol	9.910	107	436578	28.725	ng/u1#	81
27) Bis(2-Chloroethoxy)met...	10.145	93	493423	28.664	ng/u1#	96
29) 2,4-Dichlorophenol	10.387	162	360223	29.042	ng/u1#	85
30) Naphthalene	10.787	128	1159476	29.032	ng/u1	97
32) 4-Chloroaniline	10.893	127	456882	25.060	ng/u1	99
33) Hexachlorobutadiene	11.081	225	256683	29.199	ng/u1	95
34) Caprolactam	11.669	113	127993m	28.687	ng/u1	
35) 4-Chloro-3-methylphenol	12.016	107	416874	27.821	ng/u1#	70
36) 2-Methylnaphthalene	12.392	142	814822	28.131	ng/u1	93

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042722\
 Data File : BP010128.D
 Acq On : 27 Apr 2022 20:15
 Operator : CG/JU
 Sample : PB144421BS
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS421


Quant Time: Apr 28 03:26:43 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 14:49:38 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 04/28/2022
 Supervised By :Yogesh Patel 05/05/2022


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.610	142	835815	28.419	ng/ul#	95
39) 1,2,4,5-Tetrachloroben...	12.763	216	479299	30.646	ng/ul#	93
40) Hexachlorocyclopentadiene	12.751	237	266148	30.197	ng/ul	96
41) 2,4,6-Trichlorophenol	12.998	196	307786	30.416	ng/ul#	84
42) 2,4,5-Trichlorophenol	13.075	196	331454	30.105	ng/ul#	86
43) 1,1'-Biphenyl	13.398	154	1122961	30.077	ng/ul	93
44) 2-Chloronaphthalene	13.439	162	868144	30.046	ng/ul	95
45) 2-Nitroaniline	13.639	65	282297	30.131	ng/ul#	79
47) Dimethylphthalate	14.022	163	1145713	29.937	ng/ul#	92
48) 2,6-Dinitrotoluene	14.139	165	243518	31.006	ng/ul	93
50) Acenaphthylene	14.286	152	1408105	29.828	ng/ul	95
51) 3-Nitroaniline	14.469	138	220348	30.971	ng/ul#	79
52) Acenaphthene	14.628	153	905809	29.509	ng/ul	97
53) 2,4-Dinitrophenol	14.675	184	113881	25.948	ng/ul#	79
55) 4-Nitrophenol	14.775	109	166350	26.412	ng/ul#	44
56) Dibenzofuran	14.963	168	1326932	29.193	ng/ul	99
57) 2,4-Dinitrotoluene	14.928	165	356777	31.096	ng/ul#	87
58) 2,3,4,6-Tetrachlorophenol	15.192	232	294983	30.141	ng/ul#	87
59) Diethylphthalate	15.386	149	1188802	30.540	ng/ul	93
61) Fluorene	15.616	166	1084131	29.050	ng/ul	91
62) 4-Chlorophenyl-phenyle...	15.610	204	575482	29.301	ng/ul	98
63) 4-Nitroaniline	15.633	138	238783	34.492	ng/ul#	78
66) 4,6-Dinitro-2-methylph...	15.692	198	194514	30.813	ng/ul#	91
67) N-Nitrosodiphenylamine	15.822	169	950235	32.004	ng/ul	94
68) 4-Bromophenyl-phenylether	16.504	248	352155	31.905	ng/ul	96
69) Hexachlorobenzene	16.627	284	406850	31.756	ng/ul#	90
70) Atrazine	16.780	200	395663	32.778	ng/ul#	90
71) Pentachlorophenol	16.969	266	245800	30.126	ng/ul#	85
72) Phenanthrene	17.363	178	1740740	30.950	ng/ul	99
74) Anthracene	17.457	178	1775822	31.347	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.369	216	488602	32.343	ng/uL#	87
76) Pentachlorobenzene	14.892	250	468037	31.594	ng/uL	90
77) Carbazole	17.722	167	1534756	30.616	ng/ul#	95
78) Di-n-butylphthalate	18.274	149	2064129	33.468	ng/ul#	93
80) Fluoranthene	19.327	202	2041231	39.324	ng/ul	97
82) Pyrene	19.680	202	2054935	38.425	ng/ul#	94
83) Butylbenzylphthalate	20.545	149	866858	39.214	ng/ul#	80
84) 3,3'-Dichlorobenzidine	21.315	252	501324	34.431	ng/ul#	96
85) Benzo(a)anthracene	21.386	228	1689394	32.896	ng/ul	94
86) Bis(2-ethylhexyl)phtha...	21.310	149	1247995	38.091	ng/ul#	82
87) Chrysene	21.445	228	1642058	32.473	ng/ul	100
89) Di-n-octyl phthalate	22.251	149	1932850	41.264	ng/ul	100
90) Benzo(b)fluoranthene	23.110	252	1388162	34.269	ng/ul	100
91) Benzo(k)fluoranthene	23.157	252	1323424	34.491	ng/ul#	97
93) Benzo(a)pyrene	23.751	252	1291140	34.469	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.427	276	1364022	36.605	ng/ul#	93
95) Dibenzo(a,h)anthracene	26.439	278	1148405	36.017	ng/ul#	96
96) Benzo(g,h,i)perylene	27.209	276	1176728	37.425	ng/ul#	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 10.122 ng/uL
RT: 3.399 min Scan# 51
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Instrument :
BNA_P
ClientSampleId :
SLCS421




Tgt Ion: 88 Resp: 43304
Ion Ratio Lower Upper
88 100
43 31.3 6.8 10.2
58 78.7 28.4 42.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#5
Pyridine
Concen: 24.951 ng/uL
RT: 3.799 min Scan# 121
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Tgt Ion: 79 Resp: 300389
Ion Ratio Lower Upper
79 100
52 88.2 35.5 53.3#
51 41.7 20.2 30.4#





#6
Benzaldehyde
Concen: 33.965 ng/ul
RT: 7.069 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Instrument :
BNA_P
ClientSampleId :
SLCS421




Tgt Ion: 77 Resp: 296400
Ion Ratio Lower Upper
77 100
105 93.0 77.0 115.4
106 87.7 77.0 115.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#8
Phenol
Concen: 27.013 ng/ul
RT: 7.122 min Scan# 686
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Tgt Ion: 94 Resp: 443221
Ion Ratio Lower Upper
94 100
65 32.0 19.0 28.6#
66 43.8 34.6 51.8





#10
Bis(2-Chloroethyl)ether
Concen: 28.356 ng/ul
RT: 7.352 min Scan# 725
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Instrument :
BNA_P
ClientSampleId :
SLCS421




Tgt Ion: 93 Resp: 358450
Ion Ratio Lower Upper
93 100
63 79.1 45.0 67.4
95 31.7 26.3 39.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#12
2-Chlorophenol
Concen: 28.502 ng/ul
RT: 7.493 min Scan# 749
Delta R.T. -0.012 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Tgt Ion:128 Resp: 347679
Ion Ratio Lower Upper
128 100
64 53.7 45.3 67.9
130 31.6 29.0 43.6





#13
2-Methylphenol
Concen: 27.106 ng/ul
RT: 8.375 min Scan# 899
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Instrument :
BNA_P
ClientSampleId :
SLCS421




Tgt Ion:108 Resp: 341090
Ion Ratio Lower Upper
108 100
107 88.8 70.8 106.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#14
2,2'-oxybis(1-Chloropropane)
Concen: 28.293 ng/ul
RT: 8.463 min Scan# 914
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Tgt Ion: 45 Resp: 530702
Ion Ratio Lower Upper
45 100
77 13.4 17.1 25.7#
79 10.6 13.2 19.8#





#16
Acetophenone
Concen: 26.636 ng/ul
RT: 8.757 min Scan# 90
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Instrument :
BNA_P
ClientSampleId :
SLCS421




Tgt Ion: 105 Resp: 562250
Ion Ratio Lower Upper
105 100
77 79.2 53.4 80.2
51 29.9 13.8 20.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#17
N-Nitroso-di-n-propylamine
Concen: 27.321 ng/ul
RT: 8.746 min Scan# 962
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Tgt Ion: 70 Resp: 302161
Ion Ratio Lower Upper
70 100
42 59.5 31.8 47.8#
101 9.9 15.6 23.4#
130 20.3 13.7 20.5





#18
4-Methylphenol
Concen: 26.947 ng/ul
RT: 8.704 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Instrument :
BNA_P
ClientSampleId :
SLCS421




Tgt Ion:108 Resp: 373352
Ion Ratio Lower Upper
108 100
107 112.6 83.6 125.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#19
Hexachloroethane
Concen: 29.257 ng/ul
RT: 9.016 min Scan# 1008
Delta R.T. -0.012 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Tgt Ion:117 Resp: 149966
Ion Ratio Lower Upper
117 100
201 106.6 106.5 159.7
199 63.5 59.0 88.6





#22
Nitrobenzene
Concen: 29.243 ng/ul
RT: 9.134 min Scan# 1028
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Instrument :
BNA_P
ClientSampleId :
SLCS421




Tgt Ion: 77 Resp: 434822
Ion Ratio Lower Upper
77 100
123 42.8 42.3 63.5
65 15.3 5.8 8.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#23
Isophorone
Concen: 28.391 ng/ul
RT: 9.663 min Scan# 1118
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15



Tgt Ion: 82 Resp: 833828
Ion Ratio Lower Upper
82 100
95 8.0 7.8 11.8
138 17.1 12.9 19.3









#25
2-Nitrophenol
Concen: 29.361 ng/ul
RT: 9.845 min Scan# 1149
Delta R.T. -0.012 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15

Instrument :
BNA_P
ClientSampleId :
SLCS421

Tgt Ion:139 Resp: 206114
Ion Ratio Lower Upper
139 100
65 58.0 49.4 74.2
109 34.0 44.7 67.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#26
2,4-Dimethylphenol
Concen: 28.725 ng/ul
RT: 9.910 min Scan# 1160
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15


Tgt Ion:107 Resp: 436578
Ion Ratio Lower Upper
107 100
121 46.7 34.1 51.1
122 79.2 46.3 69.5





#27
Bis(2-Chloroethoxy)methane
Concen: 28.664 ng/ul
RT: 10.145 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Instrument :
BNA_P
ClientSampleId :
SLCS421




Tgt Ion: 93 Resp: 49342
Ion Ratio Lower Upper
93 100
95 33.8 26.2 39.4
123 12.1 6.2 9.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#29
2,4-Dichlorophenol
Concen: 29.042 ng/ul
RT: 10.387 min Scan# 1241
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Tgt Ion:162 Resp: 360223
Ion Ratio Lower Upper
162 100
164 62.3 37.0 55.6#
98 36.4 31.0 46.6





#30
Naphthalene
Concen: 29.032 ng/ul
RT: 10.787 min Scan# 1309
Delta R.T. -0.012 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Instrument :
BNA_P
ClientSampleId :
SLCS421




Tgt Ion:128 Resp: 1159470
Ion Ratio Lower Upper
128 100
129 11.0 8.6 13.0
127 13.4 12.2 18.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#32
4-Chloroaniline
Concen: 25.060 ng/ul
RT: 10.893 min Scan# 1327
Delta R.T. -0.012 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Tgt Ion:127 Resp: 456882
Ion Ratio Lower Upper
127 100
129 31.4 24.6 37.0





#33
Hexachlorobutadiene
Concen: 29.199 ng/ul
RT: 11.081 min Scan# 1359
Delta R.T. -0.012 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Instrument :
BNA_P
ClientSampleId :
SLCS421




Tgt Ion:225 Resp: 25668
Ion Ratio Lower Upper
225 100
223 61.9 53.4 80.0
227 63.6 53.8 80.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#34
Caprolactam
Concen: 28.687 ng/ul m
RT: 11.669 min Scan# 1459
Delta R.T. 0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Tgt Ion:113 Resp: 127993
Ion Ratio Lower Upper
113 100
55 186.4 58.2 87.2#
56 142.1 68.6 103.0#





#35
4-Chloro-3-methylphenol
Concen: 27.821 ng/ul
RT: 12.016 min Scan# 1518
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Instrument :
BNA_P
ClientSampleId :
SLCS421



Tgt Ion:107 Resp: 416874
Ion Ratio Lower Upper
107 100
144 24.0 15.8 23.8
142 74.9 38.9 58.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#36
2-Methylnaphthalene
Concen: 28.131 ng/ul
RT: 12.392 min Scan# 1582
Delta R.T. -0.012 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15

Tgt Ion:142 Resp: 814822
Ion Ratio Lower Upper
142 100
141 87.0 64.3 96.5









#37
1-Methylnaphthalene
Concen: 28.419 ng/ul
RT: 12.610 min Scan# 1619
Delta R.T. -0.012 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15

Instrument :
BNA_P
ClientSampleId :
SLCS421

Tgt Ion:142 Resp: 835815
Ion Ratio Lower Upper
142 100
141 92.1 69.8 104.8
116 4.1 5.6 8.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#39
1,2,4,5-Tetrachlorobenzene
Concen: 30.646 ng/ul
RT: 12.763 min Scan# 1645
Delta R.T. -0.012 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15


Tgt Ion:216 Resp: 479299
Ion Ratio Lower Upper
216 100
214 78.7 58.2 87.2
179 21.5 14.2 21.2#
108 16.7 15.3 22.9





#40
Hexachlorocyclopentadiene
Concen: 30.197 ng/ul
RT: 12.751 min Scan# 1643
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Instrument :
BNA_P
ClientSampleId :
SLCS421




Tgt Ion:237 Resp: 266144
Ion Ratio Lower Upper
237 100
235 62.8 53.2 79.8
272 12.5 10.8 16.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#41
2,4,6-Trichlorophenol
Concen: 30.416 ng/ul
RT: 12.998 min Scan# 1685
Delta R.T. -0.012 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Tgt Ion:196 Resp: 307786
Ion Ratio Lower Upper
196 100
198 95.3 64.1 96.1
200 31.1 19.9 29.9#





#42
2,4,5-Trichlorophenol
Concen: 30.105 ng/ul
RT: 13.075 min Scan# 1698
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Instrument :
BNA_P
ClientSampleId :
SLCS421




Tgt Ion:196 Resp: 331454
Ion Ratio Lower Upper
196 100
198 98.2 68.6 103.0
200 31.5 32.8 49.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#43
1,1'-Biphenyl
Concen: 30.077 ng/ul
RT: 13.398 min Scan# 1753
Delta R.T. -0.012 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Tgt Ion:154 Resp: 1122961
Ion Ratio Lower Upper
154 100
153 40.3 28.6 43.0
76 12.3 8.5 12.7





#44
2-Chloronaphthalene
Concen: 30.046 ng/ul
RT: 13.439 min Scan# 1760
Delta R.T. -0.012 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Instrument : BNA_P
ClientSampleId : SLCS421



Tgt Ion:162 Resp: 868144
Ion Ratio Lower Upper
162 100
164 33.5 27.5 41.3
127 41.1 37.1 55.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#45
2-Nitroaniline
Concen: 30.131 ng/ul
RT: 13.639 min Scan# 1794
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15


Tgt Ion: 65 Resp: 282297
Ion Ratio Lower Upper
65 100
92 64.2 77.9 116.9#
138 93.9 81.5 122.3





#47
Dimethylphthalate
Concen: 29.937 ng/ul
RT: 14.022 min Scan# 1859
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Instrument :
BNA_P
ClientSampleId :
SLCS421




Tgt Ion:163 Resp: 114571
Ion Ratio Lower Upper
163 100
194 4.4 4.1 6.1
164 10.0 11.3 16.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#48
2,6-Dinitrotoluene
Concen: 31.006 ng/ul
RT: 14.139 min Scan# 1879
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15



Tgt Ion:165 Resp: 243518
Ion Ratio Lower Upper
165 100
89 58.3 51.7 77.5
121 20.5 18.3 27.5









#50
Acenaphthylene
Concen: 29.828 ng/ul
RT: 14.286 min Scan# 1904
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15

Instrument :
BNA_P
ClientSampleId :
SLCS421

Tgt Ion:152 Resp: 140810
Ion Ratio Lower Upper
152 100
151 19.5 18.3 27.5
153 13.0 11.0 16.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#51
3-Nitroaniline
Concen: 30.971 ng/ul
RT: 14.469 min Scan# 1935
Delta R.T. -0.000 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15


Tgt Ion:138 Resp: 220348
Ion Ratio Lower Upper
138 100
108 11.6 15.8 23.8#
92 119.1 117.1 175.7





#52
Acenaphthene
Concen: 29.509 ng/ul
RT: 14.628 min Scan# 1962
Delta R.T. -0.012 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Instrument :
BNA_P
ClientSampleId :
SLCS421



Tgt Ion:153 Resp: 905809
Ion Ratio Lower Upper
153 100
152 46.4 34.3 51.5
154 89.5 70.0 105.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#53
2,4-Dinitrophenol
Concen: 25.948 ng/ul
RT: 14.675 min Scan# 1970
Delta R.T. -0.000 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15


Tgt Ion:184 Resp: 113881
Ion Ratio Lower Upper
184 100
63 71.4 34.6 52.0#
154 66.1 57.7 86.5





#55
4-Nitrophenol
Concen: 26.412 ng/ul
RT: 14.775 min Scan# 1987
Delta R.T. -0.000 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Instrument :
BNA_P
ClientSampleId :
SLCS421



Tgt Ion:109 Resp: 166350
Ion Ratio Lower Upper
109 100
139 119.8 61.4 92.2
65 136.3 65.3 97.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#56
Dibenzofuran
Concen: 29.193 ng/ul
RT: 14.963 min Scan# 2019
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15


Tgt Ion:168 Resp: 1326932
Ion Ratio Lower Upper
168 100
139 38.4 30.1 45.1





#57
2,4-Dinitrotoluene
Concen: 31.096 ng/ul
RT: 14.928 min Scan# 2013
Delta R.T. -0.000 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Instrument :
BNA_P
ClientSampleId :
SLCS421




Tgt Ion:165 Resp: 356777
Ion Ratio Lower Upper
165 100
63 45.1 29.5 44.3#
182 3.0 2.8 4.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 30.141 ng/ul
RT: 15.192 min Scan# 2058
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15



Tgt Ion:232 Resp: 294983
Ion Ratio Lower Upper
232 100
131 43.6 25.4 38.2#
130 2.3 4.2 6.2#
166 31.0 28.1 42.1






#59
Diethylphthalate
Concen: 30.540 ng/ul
RT: 15.386 min Scan# 2091
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15

Instrument :

BNA_P

ClientSampleId :

SLCS421






Tgt Ion:149 Resp: 118880
Ion Ratio Lower Upper
149 100
177 22.5 22.0 33.0
150 12.6 10.2 15.2

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#61
Fluorene
Concen: 29.050 ng/ul
RT: 15.616 min Scan# 2130
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15


Tgt Ion:166 Resp: 1084131
Ion Ratio Lower Upper
166 100
165 97.2 70.3 105.5
167 13.4 9.5 14.3





#62
4-Chlorophenyl-phenylether
Concen: 29.301 ng/ul
RT: 15.610 min Scan# 2129
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Instrument :
BNA_P
ClientSampleId :
SLCS421




Tgt Ion:204 Resp: 57548
Ion Ratio Lower Upper
204 100
206 33.3 27.6 41.4
141 59.1 46.6 69.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#63
4-Nitroaniline
Concen: 34.492 ng/ul
RT: 15.633 min Scan# 2133
Delta R.T. -0.000 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Tgt Ion:138 Resp: 238783
Ion Ratio Lower Upper
138 100
92 54.4 41.6 62.4
108 88.4 97.6 146.4#





#66
4,6-Dinitro-2-methylphenol
Concen: 30.813 ng/ul
RT: 15.692 min Scan# 2143
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Instrument :
BNA_P
ClientSampleId :
SLCS421




Tgt Ion:198 Resp: 194514
Ion Ratio Lower Upper
198 100
182 3.6 6.0 9.0
77 25.8 17.7 26.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#67
N-Nitrosodiphenylamine
Concen: 32.004 ng/ul
RT: 15.822 min Scan# 2165
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15



Tgt Ion:169 Resp: 950235
Ion Ratio Lower Upper
169 100
168 65.3 48.3 72.5
167 33.9 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 31.905 ng/ul
RT: 16.504 min Scan# 2281
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15

Instrument :
BNA_P
ClientSampleId :
SLCS421

Tgt Ion:248 Resp: 35215
Ion Ratio Lower Upper
248 100
250 98.3 79.4 119.0
141 77.4 67.2 100.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#69
Hexachlorobenzene
Concen: 31.756 ng/ul
RT: 16.627 min Scan# 2302
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15


Tgt Ion:284 Resp: 406850
Ion Ratio Lower Upper
284 100
142 34.2 19.2 28.8#
249 32.5 27.4 41.2





#70
Atrazine
Concen: 32.778 ng/ul
RT: 16.780 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Instrument :
BNA_P
ClientSampleId :
SLCS421



Tgt Ion:200 Resp: 39566
Ion Ratio Lower Upper
200 100
173 24.7 19.2 28.8
215 49.3 32.2 48.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#71
Pentachlorophenol
Concen: 30.126 ng/ul
RT: 16.969 min Scan# 2360
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15


Tgt Ion:266 Resp: 245800
Ion Ratio Lower Upper
266 100
264 62.4 57.3 85.9
268 64.8 65.7 98.5





#72
Phenanthrene
Concen: 30.950 ng/u1
RT: 17.363 min Scan# 2427
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Instrument :
BNA_P
ClientSampleId :
SLCS421




Tgt Ion:178 Resp: 1740740
Ion Ratio Lower Upper
178 100
179 15.4 11.7 17.5
176 18.5 14.5 21.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#74
Anthracene
Concen: 31.347 ng/u1
RT: 17.457 min Scan# 2443
Delta R.T. -0.000 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15



Tgt Ion:178 Resp: 1775822
Ion Ratio Lower Upper
178 100
179 15.6 12.3 18.5
176 18.2 13.4 20.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 32.343 ng/uL
RT: 13.369 min Scan# 1748
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15

Instrument :
BNA_P
ClientSampleId :
SLCS421

Tgt Ion:216 Resp: 488602
Ion Ratio Lower Upper
216 100
214 76.7 50.4 75.6
218 48.4 35.2 52.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#76
Pentachlorobenzene
Concen: 31.594 ng/uL
RT: 14.892 min Scan# 2007
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15


Tgt Ion:250 Resp: 468037
Ion Ratio Lower Upper
250 100
248 62.4 43.5 65.3
252 64.4 46.5 69.7





#77
 Carbazole
 Concen: 30.616 ng/ul
 RT: 17.722 min Scan# 2488
 Delta R.T. -0.000 min
 Lab File: BP010128.D
 Acq: 27 Apr 2022 20:15




Instrument :
 BNA_P
 ClientSampleId :
 SLCS421




Tgt Ion:167 Resp: 1534750
 Ion Ratio Lower Upper
 167 100
 166 21.8 15.8 23.8
 139 13.2 8.3 12.5

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
 Supervised By :Yogesh Patel 05/05/2022




#78
 Di-n-butylphthalate
 Concen: 33.468 ng/ul
 RT: 18.274 min Scan# 2582
 Delta R.T. -0.001 min
 Lab File: BP010128.D
 Acq: 27 Apr 2022 20:15




Tgt Ion:149 Resp: 2064129
 Ion Ratio Lower Upper
 149 100
 150 9.1 9.6 14.4#
 104 5.2 6.2 9.4#





#80
Fluoranthene
Concen: 39.324 ng/u1
RT: 19.327 min Scan# 2761
Delta R.T. 0.000 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Instrument :
BNA_P
ClientSampleId :
SLCS421




Tgt Ion:202 Resp: 204123
Ion Ratio Lower Upper
202 100
101 9.9 9.6 14.4
100 7.8 6.3 9.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#82
Pyrene
Concen: 38.425 ng/u1
RT: 19.680 min Scan# 2821
Delta R.T. -0.000 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15



Tgt Ion:202 Resp: 2054935
Ion Ratio Lower Upper
202 100
101 11.1 10.6 15.8
100 9.2 9.8 14.6#









#83
Butylbenzylphthalate
Concen: 39.214 ng/ul
RT: 20.545 min Scan# 2968
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15

Instrument :
BNA_P
ClientSampleId :
SLCS421

Tgt Ion:149 Resp: 866858
Ion Ratio Lower Upper
149 100
91 72.7 71.1 106.7
206 21.2 28.8 43.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#84
3,3'-Dichlorobenzidine
Concen: 34.431 ng/ul
RT: 21.315 min Scan# 3099
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15


Tgt Ion:252 Resp: 501324
Ion Ratio Lower Upper
252 100
254 64.6 49.8 74.6
126 10.3 4.3 6.5





#85
Benzo(a)anthracene
Concen: 32.896 ng/ul
RT: 21.386 min Scan# 3111
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Instrument :
BNA_P
ClientSampleId :
SLCS421



Tgt Ion:228 Resp: 1689394
Ion Ratio Lower Upper
228 100
229 19.4 18.2 27.2
226 25.7 22.8 34.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 38.091 ng/ul
RT: 21.310 min Scan# 3098
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15


Tgt Ion:149 Resp: 1247995
Ion Ratio Lower Upper
149 100
167 28.6 32.8 49.2#
279 4.7 5.6 8.4#





#87
Chrysene
Concen: 32.473 ng/ul
RT: 21.445 min Scan# 3121
Delta R.T. -0.000 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Instrument :
BNA_P
ClientSampleId :
SLCS421



Tgt Ion:228 Resp: 1642058
Ion Ratio Lower Upper
228 100
226 28.7 22.7 34.1
229 19.2 15.2 22.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#89
Di-n-octyl phthalate
Concen: 41.264 ng/ul
RT: 22.251 min Scan# 3258
Delta R.T. -0.006 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15


Tgt Ion:149 Resp: 1932850





#90
Benzo(b)fluoranthene
Concen: 34.269 ng/ul
RT: 23.110 min Scan# 3404
Delta R.T. -0.000 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Instrument :
BNA_P
ClientSampleId :
SLCS421




Tgt Ion:252 Resp: 1388162
Ion Ratio Lower Upper
252 100
253 21.7 17.4 26.0
125 8.6 7.3 10.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#91
Benzo(k)fluoranthene
Concen: 34.491 ng/ul
RT: 23.157 min Scan# 3412
Delta R.T. -0.000 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Tgt Ion:252 Resp: 1323424
Ion Ratio Lower Upper
252 100
253 21.7 17.9 26.9
125 8.7 5.1 7.7#





#93
Benzo(a)pyrene
Concen: 34.469 ng/ul
RT: 23.751 min Scan# 3513
Delta R.T. -0.000 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15




Instrument :
BNA_P
ClientSampleId :
SLCS421




Tgt Ion:252 Resp: 1291140
Ion Ratio Lower Upper
252 100
253 21.8 17.5 26.3
125 9.5 6.4 9.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 36.605 ng/ul
RT: 26.427 min Scan# 3968
Delta R.T. -0.000 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15



Tgt Ion:276 Resp: 1364022
Ion Ratio Lower Upper
276 100
138 21.2 12.3 18.5#
277 25.6 19.3 28.9





#95
Dibenzo(a,h)anthracene
Concen: 36.017 ng/ul
RT: 26.439 min Scan# 3970
Delta R.T. -0.000 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15

Instrument :

BNA_P

ClientSampleId :





SLCS421

Tgt Ion:278 Resp: 114840
Ion Ratio Lower Upper
278 100
139 13.3 6.6 10.0
279 22.7 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 04/28/2022
Supervised By :Yogesh Patel 05/05/2022



#96
Benzo(g,h,i)perylene
Concen: 37.425 ng/ul
RT: 27.209 min Scan# 4101
Delta R.T. -0.000 min
Lab File: BP010128.D
Acq: 27 Apr 2022 20:15

Tgt Ion:276 Resp: 1176728
Ion Ratio Lower Upper
276 100
138 17.6 11.6 17.4
277 23.8 15.9 23.9

