

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042922\
 Data File : BP010157.D
 Acq On : 29 Apr 2022 18:23
 Operator : CG/JU
 Sample : PB144448BS
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS448

Quant Time: Apr 30 00:31:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 14:49:38 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 05/02/2022
 Supervised By :Yogesh Patel 05/05/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	128068	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.728	136	492276	20.000	ng/u1	-0.02
38) Acenaphthene-d10	14.557	164	294004	20.000	ng/u1	-0.02
64) Phenanthrene-d10	17.310	188	602460	20.000	ng/u1	#-0.01
79) Chrysene-d12	21.398	240	590255	20.000	ng/u1	-0.01
88) Perylene-d12	23.845	264	584211	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.358	96	20781	7.076	ng/uL	-0.01
4) Pyridine-d5	3.775	84	270949	32.794	ng/u1	-0.01
7) Phenol-d5	7.093	99	356291	30.586	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.252	67	232696	33.290	ng/u1	-0.02
11) 2-Chlorophenol-d4	7.458	132	298952	34.653	ng/u1	-0.01
15) 4-Methylphenol-d8	8.640	113	288525	29.894	ng/u1	-0.01
21) Nitrobenzene-d5	9.087	128	145184	37.081	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.810	143	159227	37.237	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.352	165	279963	33.632	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.863	131	330899	27.723	ng/u1	-0.02
46) Dimethylphthalate-d6	13.969	166	812936	35.546	ng/u1	-0.01
49) Acenaphthylene-d8	14.251	160	959283	34.657	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.757	143	136910m	32.518	ng/u1	0.00
60) Fluorene-d10	15.551	176	692987	35.225	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.675	200	141782	37.219	ng/u1	0.00
73) Anthracene-d10	17.410	188	1042901	35.781	ng/u1	-0.01
81) Pyrene-d10	19.645	212	1232879	36.728	ng/u1	-0.01
92) Benzo(a)pyrene-d12	23.692	264	1162941	37.970	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.393	88	43542	14.267	ng/uL#	16
5) Pyridine	3.793	79	293557	34.181	ng/u1#	48
6) Benzaldehyde	7.064	77	249578	40.092	ng/u1	98
8) Phenol	7.116	94	379352	32.410	ng/u1#	94
10) Bis(2-Chloroethyl)ether	7.346	93	303270	33.630	ng/u1#	77
12) 2-Chlorophenol	7.493	128	301659	34.666	ng/u1	94
13) 2-Methylphenol	8.369	108	281683	31.379	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.458	45	443295	33.129	ng/u1#	84
16) Acetophenone	8.752	105	443107	29.426	ng/u1#	79
17) N-Nitroso-di-n-propyla...	8.740	70	235674	29.872	ng/u1#	76
18) 4-Methylphenol	8.699	108	302302	30.586	ng/u1	91
19) Hexachloroethane	9.010	117	131490	35.960	ng/u1	83
22) Nitrobenzene	9.128	77	350927	36.583	ng/u1#	85
23) Isophorone	9.657	82	628200	33.156	ng/u1#	96
25) 2-Nitrophenol	9.840	139	169934	37.523	ng/u1#	81
26) 2,4-Dimethylphenol	9.905	107	328678	33.522	ng/u1#	78
27) Bis(2-Chloroethoxy)met...	10.140	93	387313	34.877	ng/u1#	96
29) 2,4-Dichlorophenol	10.381	162	281696	35.203	ng/u1#	81
30) Naphthalene	10.781	128	927793	36.009	ng/u1	97
32) 4-Chloroaniline	10.887	127	331455	28.180	ng/u1	97
33) Hexachlorobutadiene	11.075	225	210925	37.192	ng/u1	95
34) Caprolactam	11.663	113	85583	29.733	ng/u1#	1
35) 4-Chloro-3-methylphenol	12.016	107	304819	31.533	ng/u1#	69
36) 2-Methylnaphthalene	12.387	142	622157	33.295	ng/u1	91

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042922\
 Data File : BP010157.D
 Acq On : 29 Apr 2022 18:23
 Operator : CG/JU
 Sample : PB144448BS
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS448


Quant Time: Apr 30 00:31:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 14:49:38 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 05/02/2022
 Supervised By :Yogesh Patel 05/05/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.604	142	627278	33.061	ng/ul#	94
39) 1,2,4,5-Tetrachloroben...	12.757	216	356871	39.289	ng/ul	92
40) Hexachlorocyclopentadiene	12.740	237	203944	39.841	ng/ul	98
41) 2,4,6-Trichlorophenol	12.998	196	224541	38.206	ng/ul#	83
42) 2,4,5-Trichlorophenol	13.069	196	243799	38.127	ng/ul#	87
43) 1,1'-Biphenyl	13.393	154	834028	38.462	ng/ul	92
44) 2-Chloronaphthalene	13.434	162	640713	38.181	ng/ul	94
45) 2-Nitroaniline	13.634	65	197437	36.284	ng/ul#	80
47) Dimethylphthalate	14.016	163	783488	35.250	ng/ul#	92
48) 2,6-Dinitrotoluene	14.134	165	166441	36.489	ng/ul	92
50) Acenaphthylene	14.281	152	1021563	37.260	ng/ul	95
51) 3-Nitroaniline	14.457	138	149322	36.137	ng/ul#	82
52) Acenaphthene	14.622	153	651105	36.522	ng/ul	96
53) 2,4-Dinitrophenol	14.669	184	86270	33.846	ng/ul#	80
55) 4-Nitrophenol	14.775	109	119070	32.551	ng/ul#	50
56) Dibenzofuran	14.957	168	946218	35.843	ng/ul	100
57) 2,4-Dinitrotoluene	14.922	165	238464	35.786	ng/ul#	87
58) 2,3,4,6-Tetrachlorophenol	15.187	232	202311	35.593	ng/ul#	88
59) Diethylphthalate	15.381	149	794567	35.146	ng/ul	93
61) Fluorene	15.610	166	764124	35.254	ng/ul	91
62) 4-Chlorophenyl-phenyle...	15.604	204	407872	35.756	ng/ul	98
63) 4-Nitroaniline	15.622	138	166995	41.534	ng/ul#	76
66) 4,6-Dinitro-2-methylph...	15.687	198	132774	36.125	ng/ul#	93
67) N-Nitrosodiphenylamine	15.816	169	658645	38.102	ng/ul	94
68) 4-Bromophenyl-phenylether	16.498	248	246316	38.330	ng/ul	95
69) Hexachlorobenzene	16.622	284	278254	37.305	ng/ul#	90
70) Atrazine	16.769	200	260370	37.049	ng/ul	90
71) Pentachlorophenol	16.963	266	155184m	32.669	ng/ul	
72) Phenanthrene	17.357	178	1198783	36.609	ng/ul	99
74) Anthracene	17.445	178	1202158	36.449	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.363	216	361192	41.067	ng/ul#	87
76) Pentachlorobenzene	14.881	250	329988	38.260	ng/ul#	88
77) Carbazole	17.716	167	1045339	35.816	ng/ul#	95
78) Di-n-butylphthalate	18.269	149	1365352	38.024	ng/ul#	92
80) Fluoranthene	19.322	202	1423252	37.590	ng/ul	97
82) Pyrene	19.675	202	1459785	37.422	ng/ul#	93
83) Butylbenzylphthalate	20.539	149	618616	38.365	ng/ul#	82
84) 3,3'-Dichlorobenzidine	21.310	252	457416	43.069	ng/ul#	96
85) Benzo(a)anthracene	21.380	228	1393408	37.198	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.304	149	927159	38.796	ng/ul#	82
87) Chrysene	21.433	228	1371684	37.189	ng/ul	98
89) Di-n-octyl phthalate	22.245	149	1586979	34.943	ng/ul	100
90) Benzo(b)fluoranthene	23.098	252	1417627	36.095	ng/ul	100
91) Benzo(k)fluoranthene	23.151	252	1408799	37.868	ng/ul#	98
93) Benzo(a)pyrene	23.739	252	1402201	38.608	ng/ul#	97
94) Indeno(1,2,3-cd)pyrene	26.415	276	1599561	44.273	ng/ul#	93
95) Dibenzo(a,h)anthracene	26.427	278	1370594	44.334	ng/ul#	95
96) Benzo(g,h,i)perylene	27.192	276	1392644	45.682	ng/ul#	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#2
1,4-Dioxane
Concen: 14.267 ng/uL
RT: 3.393 min Scan# 51
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23

Instrument :
BNA_P
ClientSampleId :
SLCS448

Tgt Ion: 88 Resp: 43542
Ion Ratio Lower Upper
88 100
43 35.0 6.8 10.2
58 86.2 28.4 42.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#5
Pyridine
Concen: 34.181 ng/uL
RT: 3.793 min Scan# 120
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23

Tgt Ion: 79 Resp: 293557
Ion Ratio Lower Upper
79 100
52 85.8 35.5 53.3#
51 40.9 20.2 30.4#









#6
Benzaldehyde
Concen: 40.092 ng/ul
RT: 7.064 min Scan# 61
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23

Instrument :
BNA_P
ClientSampleId :
SLCS448

Tgt Ion: 77 Resp: 249578
Ion Ratio Lower Upper
77 100
105 96.4 77.0 115.4
106 92.2 77.0 115.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#8
Phenol
Concen: 32.410 ng/ul
RT: 7.116 min Scan# 685
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23


Tgt Ion: 94 Resp: 379352
Ion Ratio Lower Upper
94 100
65 31.9 19.0 28.6
66 43.2 34.6 51.8





#10
Bis(2-Chloroethyl)ether
Concen: 33.630 ng/ul
RT: 7.346 min Scan# 724
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Instrument :
BNA_P
ClientSampleId :
SLCS448




Tgt Ion: 93 Resp: 303270
Ion Ratio Lower Upper
93 100
63 82.3 45.0 67.4
95 32.5 26.3 39.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#12
2-Chlorophenol
Concen: 34.666 ng/ul
RT: 7.493 min Scan# 749
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23



Tgt Ion:128 Resp: 301659
Ion Ratio Lower Upper
128 100
64 51.6 45.3 67.9
130 32.9 29.0 43.6





#13

2-Methylphenol

Concen: 31.379 ng/ul

RT: 8.369 min Scan# 898

Delta R.T. -0.012 min

Lab File: BP010157.D

Acq: 29 Apr 2022 18:23

Instrument :

BNA_P

ClientSampleId :

SLCS448

Tgt Ion:108 Resp: 28168

Ion Ratio Lower Upper

108 100





107 91.6 70.8 106.2

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 05/02/2022

Supervised By :Yogesh Patel 05/05/2022



#14

2,2'-oxybis(1-Chloropropane)

Concen: 33.129 ng/ul

RT: 8.458 min Scan# 913

Delta R.T. -0.012 min

Lab File: BP010157.D

Acq: 29 Apr 2022 18:23




Tgt Ion: 45 Resp: 443295


Ion Ratio Lower Upper

45 100

77 13.1 17.1 25.7#


79 10.3 13.2 19.8#





#16
Acetophenone
Concen: 29.426 ng/ul
RT: 8.752 min Scan# 90
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Instrument :
BNA_P
ClientSampleId :
SLCS448




Tgt Ion: 105 Resp: 44310
Ion Ratio Lower Upper
105 100
77 82.0 53.4 80.2
51 30.6 13.8 20.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#17
N-Nitroso-di-n-propylamine
Concen: 29.872 ng/ul
RT: 8.740 min Scan# 961
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Tgt Ion: 70 Resp: 235674
Ion Ratio Lower Upper
70 100
42 60.0 31.8 47.8#
101 10.7 15.6 23.4#
130 21.4 13.7 20.5#





#18
4-Methylphenol
Concen: 30.586 ng/ul
RT: 8.699 min Scan# 91
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Instrument :
BNA_P
ClientSampleId :
SLCS448




Tgt Ion:108 Resp: 302302
Ion Ratio Lower Upper
108 100
107 113.8 83.6 125.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#19
Hexachloroethane
Concen: 35.960 ng/ul
RT: 9.010 min Scan# 1007
Delta R.T. -0.018 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Tgt Ion:117 Resp: 131490
Ion Ratio Lower Upper
117 100
201 107.3 106.5 159.7
199 67.0 59.0 88.6





#22
Nitrobenzene
Concen: 36.583 ng/ul
RT: 9.128 min Scan# 1027
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Instrument :
BNA_P
ClientSampleId :
SLCS448




Tgt Ion: 77 Resp: 350927
Ion Ratio Lower Upper
77 100
123 43.3 42.3 63.5
65 15.9 5.8 8.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#23
Isophorone
Concen: 33.156 ng/ul
RT: 9.657 min Scan# 1117
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23



Tgt Ion: 82 Resp: 628200
Ion Ratio Lower Upper
82 100
95 7.5 7.8 11.8#
138 17.0 12.9 19.3





#25

2-Nitrophenol

Concen: 37.523 ng/ul

RT: 9.840 min Scan# 1149

Delta R.T. -0.018 min

Lab File: BP010157.D

Acq: 29 Apr 2022 18:23

Instrument :

BNA_P

ClientSampleId :





SLCS448

Tgt Ion:	139	Resp:	169934
Ion Ratio	Lower	Upper	
139	100		
65	56.4	49.4	74.2
109	32.8	44.7	67.1

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#26

2,4-Dimethylphenol

Concen: 33.522 ng/ul




RT: 9.905 min Scan# 1159


Delta R.T. -0.012 min

Lab File: BP010157.D

Acq: 29 Apr 2022 18:23


Tgt Ion:	107	Resp:	328678
Ion Ratio	Lower	Upper	
107	100		
121	48.4	34.1	51.1
122	80.9	46.3	69.5





#27
Bis(2-Chloroethoxy)methane
Concen: 34.877 ng/ul
RT: 10.140 min Scan# 1199
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Instrument :
BNA_P
ClientSampleId :
SLCS448



Tgt Ion: 93 Resp: 38731
Ion Ratio Lower Upper
93 100
95 31.6 26.2 39.4
123 11.3 6.2 9.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#29
2,4-Dichlorophenol
Concen: 35.203 ng/ul
RT: 10.381 min Scan# 1240
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23


Tgt Ion:162 Resp: 281696
Ion Ratio Lower Upper
162 100
164 68.8 37.0 55.6#
98 38.1 31.0 46.6





#30
Naphthalene
Concen: 36.009 ng/ul
RT: 10.781 min Scan# 1308
Delta R.T. -0.018 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Instrument :
BNA_P
ClientSampleId :
SLCS448



Tgt Ion:128 Resp: 92779
Ion Ratio Lower Upper
128 100
129 10.8 8.6 13.0
127 13.1 12.2 18.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#32
4-Chloroaniline
Concen: 28.180 ng/ul
RT: 10.887 min Scan# 1326
Delta R.T. -0.018 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23


Tgt Ion:127 Resp: 331455
Ion Ratio Lower Upper
127 100
129 32.7 24.6 37.0





#33
Hexachlorobutadiene
Concen: 37.192 ng/ul
RT: 11.075 min Scan# 11
Delta R.T. -0.018 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Instrument :
BNA_P
ClientSampleId :
SLCS448




Tgt Ion:225 Resp: 21092
Ion Ratio Lower Upper
225 100
223 62.6 53.4 80.0
227 62.7 53.8 80.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#34
Caprolactam
Concen: 29.733 ng/ul
RT: 11.663 min Scan# 1458
Delta R.T. 0.000 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23



Tgt Ion:113 Resp: 85583
Ion Ratio Lower Upper
113 100
55 195.0 58.2 87.2#
56 148.9 68.6 103.0#









#35
4-Chloro-3-methylphenol
Concen: 31.533 ng/ul
RT: 12.016 min Scan# 1517
Delta R.T. -0.006 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23

Instrument :
BNA_P
ClientSampleId :
SLCS448

Tgt Ion:107 Resp: 304819
Ion Ratio Lower Upper
107 100
144 23.9 15.8 23.8
142 76.1 38.9 58.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#36
2-Methylnaphthalene
Concen: 33.295 ng/ul
RT: 12.387 min Scan# 1581
Delta R.T. -0.018 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23


Tgt Ion:142 Resp: 622157
Ion Ratio Lower Upper
142 100
141 88.7 64.3 96.5





#37
1-Methylnaphthalene
Concen: 33.061 ng/ul
RT: 12.604 min Scan# 1618
Delta R.T. -0.018 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Instrument :
BNA_P
ClientSampleId :
SLCS448




Tgt Ion:142 Resp: 627278
Ion Ratio Lower Upper
142 100
141 92.8 69.8 104.8
116 4.0 5.6 8.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 39.289 ng/ul
RT: 12.757 min Scan# 1644
Delta R.T. -0.018 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Tgt Ion:216 Resp: 356871
Ion Ratio Lower Upper
216 100
214 80.2 58.2 87.2
179 21.1 14.2 21.2
108 16.6 15.3 22.9





#40
Hexachlorocyclopentadiene
Concen: 39.841 ng/ul
RT: 12.740 min Scan# 1642
Delta R.T. -0.018 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Instrument :
BNA_P
ClientSampleId :
SLCS448




Tgt Ion:237 Resp: 203944
Ion Ratio Lower Upper
237 100
235 64.9 53.2 79.8
272 12.6 10.8 16.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#41
2,4,6-Trichlorophenol
Concen: 38.206 ng/ul
RT: 12.998 min Scan# 1685
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Tgt Ion:196 Resp: 224541
Ion Ratio Lower Upper
196 100
198 95.4 64.1 96.1
200 32.0 19.9 29.9#





#42
2,4,5-Trichlorophenol
Concen: 38.127 ng/ul
RT: 13.069 min Scan# 1697
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Instrument :
BNA_P
ClientSampleId :
SLCS448




Tgt Ion:196 Resp: 243799
Ion Ratio Lower Upper
196 100
198 96.0 68.6 103.0
200 29.9 32.8 49.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#43
1,1'-Biphenyl
Concen: 38.462 ng/ul
RT: 13.393 min Scan# 1752
Delta R.T. -0.018 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23



Tgt Ion:154 Resp: 834028
Ion Ratio Lower Upper
154 100
153 41.1 28.6 43.0
76 12.5 8.5 12.7









#44
2-Chloronaphthalene
Concen: 38.181 ng/ul
RT: 13.434 min Scan# 1759
Delta R.T. -0.018 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23

Instrument :
BNA_P
ClientSampleId :
SLCS448

Tgt Ion:162 Resp: 640711
Ion Ratio Lower Upper
162 100
164 33.0 27.5 41.3
127 41.0 37.1 55.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#45
2-Nitroaniline
Concen: 36.284 ng/ul
RT: 13.634 min Scan# 1793
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23


Tgt Ion: 65 Resp: 197437
Ion Ratio Lower Upper
65 100
92 65.1 77.9 116.9#
138 94.7 81.5 122.3





#47
Dimethylphthalate
Concen: 35.250 ng/ul
RT: 14.016 min Scan# 1858
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Instrument :
BNA_P
ClientSampleId :
SLCS448




Tgt Ion:163 Resp: 78348
Ion Ratio Lower Upper
163 100
194 4.2 4.1 6.1
164 10.1 11.3 16.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#48
2,6-Dinitrotoluene
Concen: 36.489 ng/ul
RT: 14.134 min Scan# 1878
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Tgt Ion:165 Resp: 166441
Ion Ratio Lower Upper
165 100
89 57.1 51.7 77.5
121 20.3 18.3 27.5





#50
Acenaphthylene
Concen: 37.260 ng/ul
RT: 14.281 min Scan# 1903
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Instrument : BNA_P
ClientSampleId : SLCS448




Tgt Ion:152 Resp: 102156
Ion Ratio Lower Upper
152 100
151 19.9 18.3 27.5
153 13.0 11.0 16.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#51
3-Nitroaniline
Concen: 36.137 ng/ul
RT: 14.457 min Scan# 1933
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Tgt Ion:138 Resp: 149322
Ion Ratio Lower Upper
138 100
108 12.0 15.8 23.8#
92 124.2 117.1 175.7





#52
Acenaphthene
Concen: 36.522 ng/ul
RT: 14.622 min Scan# 1961
Delta R.T. -0.018 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Instrument :
BNA_P
ClientSampleId :
SLCS448




Tgt Ion:153 Resp: 65110
Ion Ratio Lower Upper
153 100
152 46.1 34.3 51.5
154 90.3 70.0 105.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#53
2,4-Dinitrophenol
Concen: 33.846 ng/ul
RT: 14.669 min Scan# 1969
Delta R.T. -0.006 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Tgt Ion:184 Resp: 86270
Ion Ratio Lower Upper
184 100
63 71.8 34.6 52.0#
154 67.7 57.7 86.5





#55
4-Nitrophenol
Concen: 32.551 ng/ul
RT: 14.775 min Scan# 1985
Delta R.T. 0.000 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Instrument :
BNA_P
ClientSampleId :
SLCS448



Tgt Ion:109 Resp: 119070
Ion Ratio Lower Upper
109 100
139 119.9 61.4 92.2
65 127.1 65.3 97.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#56
Dibenzofuran
Concen: 35.843 ng/ul
RT: 14.957 min Scan# 2018
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23

Tgt Ion:168 Resp: 946218
Ion Ratio Lower Upper
168 100
139 37.9 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 35.786 ng/ul
RT: 14.922 min Scan# 2012
Delta R.T. -0.006 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23

Instrument :
BNA_P
ClientSampleId :
SLCS448

Tgt Ion:165 Resp: 238464
Ion Ratio Lower Upper
165 100
63 45.6 29.5 44.3#
182 3.2 2.8 4.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#58
2,3,4,6-Tetrachlorophenol
Concen: 35.593 ng/ul
RT: 15.187 min Scan# 2057
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23

Tgt Ion:232 Resp: 202311
Ion Ratio Lower Upper
232 100
131 42.9 25.4 38.2#
130 2.1 4.2 6.2#
166 31.9 28.1 42.1









#59
Diethylphthalate
Concen: 35.146 ng/ul
RT: 15.381 min Scan# 2090
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23

Instrument :
BNA_P
ClientSampleId :
SLCS448

Tgt Ion:149 Resp: 79456
Ion Ratio Lower Upper
149 100
177 22.3 22.0 33.0
150 12.2 10.2 15.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#61
Fluorene
Concen: 35.254 ng/ul
RT: 15.610 min Scan# 2129
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23


Tgt Ion:166 Resp: 764124
Ion Ratio Lower Upper
166 100
165 96.6 70.3 105.5
167 13.6 9.5 14.3





#62
4-Chlorophenyl-phenylether
Concen: 35.756 ng/ul
RT: 15.604 min Scan# 2128
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Instrument :
BNA_P
ClientSampleId :
SLCS448




Tgt Ion:204 Resp: 407872
Ion Ratio Lower Upper
204 100
206 32.7 27.6 41.4
141 56.7 46.6 69.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#63
4-Nitroaniline
Concen: 41.534 ng/ul
RT: 15.622 min Scan# 2131
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Tgt Ion:138 Resp: 166995
Ion Ratio Lower Upper
138 100
92 55.7 41.6 62.4
108 85.4 97.6 146.4#





#66
4,6-Dinitro-2-methylphenol
Concen: 36.125 ng/ul
RT: 15.687 min Scan# 2142
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Instrument :
BNA_P
ClientSampleId :
SLCS448




Tgt Ion:198 Resp: 132774
Ion Ratio Lower Upper
198 100
182 4.5 6.0 9.0
77 25.2 17.7 26.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#67
N-Nitrosodiphenylamine
Concen: 38.102 ng/ul
RT: 15.816 min Scan# 2164
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Tgt Ion:169 Resp: 658645
Ion Ratio Lower Upper
169 100
168 66.0 48.3 72.5
167 35.2 29.6 44.4





#68
4-Bromophenyl-phenylether
Concen: 38.330 ng/ul
RT: 16.498 min Scan# 2280
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Instrument :
BNA_P
ClientSampleId :
SLCS448




Tgt Ion:248 Resp: 246310
Ion Ratio Lower Upper
248 100
250 99.0 79.4 119.0
141 75.3 67.2 100.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#69
Hexachlorobenzene
Concen: 37.305 ng/ul
RT: 16.622 min Scan# 2301
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23



Tgt Ion:284 Resp: 278254
Ion Ratio Lower Upper
284 100
142 32.4 19.2 28.8#
249 31.1 27.4 41.2









#70
Atrazine
Concen: 37.049 ng/ul
RT: 16.769 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23

Instrument :
BNA_P
ClientSampleId :
SLCS448

Tgt Ion:200 Resp: 260370
Ion Ratio Lower Upper
200 100
173 26.8 19.2 28.8
215 47.6 32.2 48.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#71
Pentachlorophenol
Concen: 32.669 ng/ul m
RT: 16.963 min Scan# 2359
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23

Tgt Ion:266 Resp: 155184
Ion Ratio Lower Upper
266 100
264 64.2 57.3 85.9
268 65.2 65.7 98.5#









#72
Phenanthrene
Concen: 36.609 ng/u1
RT: 17.357 min Scan# 2426
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23

Instrument :
BNA_P
ClientSampleId :
SLCS448

Tgt Ion:178 Resp: 119878
Ion Ratio Lower Upper
178 100
179 15.4 11.7 17.5
176 18.4 14.5 21.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#74
Anthracene
Concen: 36.449 ng/u1
RT: 17.445 min Scan# 2441
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23

Tgt Ion:178 Resp: 1202158
Ion Ratio Lower Upper
178 100
179 15.8 12.3 18.5
176 18.1 13.4 20.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 41.067 ng/uL
RT: 13.363 min Scan# 11
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23

Instrument :
BNA_P
ClientSampleId :
SLCS448

Tgt Ion:216 Resp: 361192
Ion Ratio Lower Upper
216 100
214 76.0 50.4 75.6
218 49.0 35.2 52.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#76
Pentachlorobenzene
Concen: 38.260 ng/uL
RT: 14.881 min Scan# 2005
Delta R.T. -0.017 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23


Tgt Ion:250 Resp: 329988
Ion Ratio Lower Upper
250 100
248 65.5 43.5 65.3
252 64.4 46.5 69.7





#77
 Carbazole
 Concen: 35.816 ng/ul
 RT: 17.716 min Scan# 2487
 Delta R.T. -0.006 min
 Lab File: BP010157.D
 Acq: 29 Apr 2022 18:23




Instrument :
 BNA_P
 ClientSampleId :
 SLCS448




Tgt Ion:167 Resp: 1045339
 Ion Ratio Lower Upper
 167 100
 166 21.6 15.8 23.8
 139 12.7 8.3 12.5

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
 Supervised By :Yogesh Patel 05/05/2022




#78
 Di-n-butylphthalate
 Concen: 38.024 ng/ul
 RT: 18.269 min Scan# 2581
 Delta R.T. -0.006 min
 Lab File: BP010157.D
 Acq: 29 Apr 2022 18:23




Tgt Ion:149 Resp: 1365352
 Ion Ratio Lower Upper
 149 100
 150 9.0 9.6 14.4#
 104 5.2 6.2 9.4#





#80
Fluoranthene
Concen: 37.590 ng/u1
RT: 19.322 min Scan# 21
Delta R.T. -0.005 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Instrument :
BNA_P
ClientSampleId :
SLCS448




Tgt Ion:202 Resp: 142325
Ion Ratio Lower Upper
202 100
101 9.8 9.6 14.4
100 7.8 6.3 9.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#82
Pyrene
Concen: 37.422 ng/u1
RT: 19.675 min Scan# 2820
Delta R.T. -0.006 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23



Tgt Ion:202 Resp: 1459785
Ion Ratio Lower Upper
202 100
101 10.7 10.6 15.8
100 9.0 9.8 14.6#









#83
Butylbenzylphthalate
Concen: 38.365 ng/ul
RT: 20.539 min Scan# 2968
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23

Instrument :
BNA_P
ClientSampleId :
SLCS448

Tgt Ion:149 Resp: 618610
Ion Ratio Lower Upper
149 100
91 74.2 71.1 106.7
206 21.8 28.8 43.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#84
3,3'-Dichlorobenzidine
Concen: 43.069 ng/ul
RT: 21.310 min Scan# 3098
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23


Tgt Ion:252 Resp: 457416
Ion Ratio Lower Upper
252 100
254 64.6 49.8 74.6
126 9.7 4.3 6.5





#85
Benzo(a)anthracene
Concen: 37.198 ng/ul
RT: 21.380 min Scan# 3110
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Instrument :
BNA_P
ClientSampleId :
SLCS448




Tgt Ion:228 Resp: 1393408
Ion Ratio Lower Upper
228 100
229 19.5 18.2 27.2
226 26.6 22.8 34.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#86
Bis(2-ethylhexyl)phthalate
Concen: 38.796 ng/ul
RT: 21.304 min Scan# 3097
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Tgt Ion:149 Resp: 927159
Ion Ratio Lower Upper
149 100
167 28.1 32.8 49.2#
279 5.2 5.6 8.4#





#87
Chrysene
Concen: 37.189 ng/ul
RT: 21.433 min Scan# 3119
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Instrument :
BNA_P
ClientSampleId :
SLCS448



Tgt Ion:228 Resp: 1371684
Ion Ratio Lower Upper
228 100
226 29.2 22.7 34.1
229 20.3 15.2 22.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#89
Di-n-octyl phthalate
Concen: 34.943 ng/ul
RT: 22.245 min Scan# 3257
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23


Tgt Ion:149 Resp: 1586979





#90
Benzo(b)fluoranthene
Concen: 36.095 ng/ul
RT: 23.098 min Scan# 3402
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Instrument :
BNA_P
ClientSampleId :
SLCS448




Tgt Ion:252 Resp: 141762
Ion Ratio Lower Upper
252 100
253 21.7 17.4 26.0
125 8.8 7.3 10.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#91
Benzo(k)fluoranthene
Concen: 37.868 ng/ul
RT: 23.151 min Scan# 3411
Delta R.T. -0.006 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23



Tgt Ion:252 Resp: 1408799
Ion Ratio Lower Upper
252 100
253 21.6 17.9 26.9
125 8.1 5.1 7.7#





#93
Benzo(a)pyrene
Concen: 38.608 ng/ul
RT: 23.739 min Scan# 3511
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23

Instrument :

BNA_P

ClientSampleId :





SLCS448

Tgt Ion:252 Resp: 140220
Ion Ratio Lower Upper
252 100
253 22.8 17.5 26.3
125 9.7 6.4 9.6

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#94
Indeno(1,2,3-cd)pyrene
Concen: 44.273 ng/ul
RT: 26.415 min Scan# 3966
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23


Tgt Ion:276 Resp: 1599561
Ion Ratio Lower Upper
276 100
138 21.9 12.3 18.5#
277 25.1 19.3 28.9





#95
Dibenzo(a,h)anthracene
Concen: 44.334 ng/ul
RT: 26.427 min Scan# 3967
Delta R.T. -0.012 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23




Instrument :
BNA_P
ClientSampleId :
SLCS448



Tgt Ion:278 Resp: 1370594
Ion Ratio Lower Upper
278 100
139 13.6 6.6 10.0
279 23.7 18.4 27.6

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#96
Benzo(g,h,i)perylene
Concen: 45.682 ng/ul
RT: 27.192 min Scan# 4098
Delta R.T. -0.018 min
Lab File: BP010157.D
Acq: 29 Apr 2022 18:23

Tgt Ion:276 Resp: 1392644
Ion Ratio Lower Upper
276 100
138 17.8 11.6 17.4
277 23.6 15.9 23.9

