




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042925\
Data File : BP024452.D
Acq On : 29 Apr 2025 09:49
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 29 11:27:18 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration



TIC: BP024452.D\data.ms

(70) Pentachlorophenol (C)




16.792min (-0.006) 256843.27 ng

response 479697

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.80	64.63
264.00	62.20	63.79
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042925\
Data File : BP024452.D
Acq On : 29 Apr 2025 09:49
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 29 11:27:18 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration



TIC: BP024452.D\data.ms

(77) Benzidine**19.469min (-0.012) 126073.39 ng****response 2077513**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.50	14.04
183.00	11.50	12.15
0.00	0.00	0.00