

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050824\
 Data File : BP020246.D
 Acq On : 08 May 2024 20:40
 Operator : MA/JU
 Sample : P2377-10
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SVOC-GPC2-BLANK

Quant Time: May 08 22:59:50 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP050124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed May 08 22:57:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.616	152	115669	20.000	ng/u1	0.00
20) Naphthalene-d8	10.387	136	505969	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.304	164	382975	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.139	188	811979	20.000	ng/u1	0.00
79) Chrysene-d12	21.627	240	600138	20.000	ng/u1	-0.01
88) Perylene-d12	24.992	264	662672	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.217	96	47	0.017	ng/uL	0.00
4) Pyridine-d5	3.617	84	424	0.053	ng/u1	0.00
7) Phenol-d5	6.799	99	3	0.000	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.975	67	50	0.008	ng/u1	0.00
11) 2-Chlorophenol-d4	7.164	132	18	0.003	ng/u1	0.00
15) 4-Methylphenol-d8	8.322	113	9	0.001	ng/u1	0.00
21) Nitrobenzene-d5	8.781	128	8	0.002	ng/u1	0.00
24) 2-Nitrophenol-d4	9.493	143	15	0.003	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.910	165	15	0.002	ng/u1	-0.11
31) 4-Chloroaniline-d4	10.546	131	7	0.001	ng/u1	0.00
46) Dimethylphthalate-d6	13.722	166	17	0.001	ng/u1	0.01
49) Acenaphthylene-d8	13.993	160	21	0.001	ng/u1	0.01
54) 4-Nitrophenol-d4	14.551	143	10	0.002	ng/u1	0.00
60) Fluorene-d10	15.310	176	10	0.000	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.475	200	3	0.001	ng/u1	0.00
73) Anthracene-d10	17.239	188	1629	0.046	ng/u1	0.00
81) Pyrene-d10	19.628	212	4	0.000	ng/u1	-0.02
92) Benzo(a)pyrene-d12	24.769	264	193	0.006	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050824\
Data File : BP020246.D
Acq On : 08 May 2024 20:40
Operator : MA/JU
Sample : P2377-10
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SVOC-GPC2-BLANK

Quant Time: May 08 22:59:50 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP050124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed May 08 22:57:55 2024
Response via : Initial Calibration

