

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051622\
 Data File : BP010387.D
 Acq On : 17 May 2022 12:32
 Operator : CG/JU
 Sample : N2713-14MSD
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DBSL3MSD

Quant Time: May 18 00:57:55 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 13 15:04:31 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 05/18/2022
 Supervised By :mohammad ahmed 05/18/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.898	152	146801	20.000	ng/u1	0.00
20) Naphthalene-d8	10.698	136	649427	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.527	164	446122	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.280	188	992567	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.368	240	1002063	20.000	ng/u1	# 0.00
88) Perylene-d12	23.798	264	957985	20.000	ng/u1	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.346	96	15763	4.577	ng/uL	0.00
4) Pyridine-d5	3.763	84	138787	14.192	ng/u1	0.00
7) Phenol-d5	7.063	99	330608	26.832	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.222	67	212394	25.657	ng/u1	0.00
11) 2-Chlorophenol-d4	7.428	132	252650	26.797	ng/u1	0.00
15) 4-Methylphenol-d8	8.604	113	277012	26.455	ng/u1	0.00
21) Nitrobenzene-d5	9.057	128	133522	26.565	ng/u1	0.00
24) 2-Nitrophenol-d4	9.781	143	144259	27.083	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.316	165	278050	27.607	ng/u1	0.00
31) 4-Chloroaniline-d4	10.834	131	346691	23.151	ng/u1	0.00
46) Dimethylphthalate-d6	13.939	166	896436	27.273	ng/u1	0.00
49) Acenaphthylene-d8	14.222	160	1037332	27.361	ng/u1	0.00
54) 4-Nitrophenol-d4	14.727	143	143884m	24.594	ng/u1	0.00
60) Fluorene-d10	15.522	176	767303	26.833	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.645	200	124779	21.495	ng/u1	0.00
73) Anthracene-d10	17.380	188	1215902	27.053	ng/u1	0.00
81) Pyrene-d10	19.615	212	1471473	27.651	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.645	264	1353818	28.144	ng/u1	0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.375	88	34609	9.701	ng/uL#	13
5) Pyridine	3.775	79	244796	24.087	ng/u1#	36
6) Benzaldehyde	7.034	77	203812	31.291	ng/u1	94
8) Phenol	7.087	94	347704	26.158	ng/u1#	93
10) Bis(2-Chloroethyl)ether	7.316	93	266837	25.532	ng/u1#	73
12) 2-Chlorophenol	7.457	128	257963	26.209	ng/u1	97
13) 2-Methylphenol	8.340	108	257816	25.981	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.428	45	432614	25.398	ng/u1#	82
16) Acetophenone	8.722	105	419904	25.523	ng/u1#	78
17) N-Nitroso-di-n-propyla...	8.704	70	226470	25.721	ng/u1#	71
18) 4-Methylphenol	8.669	108	284043	25.899	ng/u1	89
19) Hexachloroethane	8.975	117	106680	24.921	ng/u1	82
22) Nitrobenzene	9.098	77	335515	25.402	ng/u1#	82
23) Isophorone	9.628	82	626464	25.661	ng/u1	97
25) 2-Nitrophenol	9.810	139	152517	26.077	ng/u1#	83
26) 2,4-Dimethylphenol	9.875	107	328738m	25.904	ng/u1	
27) Bis(2-Chloroethoxy)met...	10.104	93	370188	25.467	ng/u1#	98
29) 2,4-Dichlorophenol	10.345	162	270046	26.225	ng/u1#	83
30) Naphthalene	10.745	128	858592	25.312	ng/u1	97
32) 4-Chloroaniline	10.857	127	337742	22.725	ng/u1	96
33) Hexachlorobutadiene	11.039	225	181478	24.871	ng/u1	96
34) Caprolactam	11.628	113	93959	26.146	ng/u1#	1
35) 4-Chloro-3-methylphenol	11.986	107	316387	26.265	ng/u1#	68
36) 2-Methylnaphthalene	12.357	142	604093	25.156	ng/u1	92

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051622\
 Data File : BP010387.D
 Acq On : 17 May 2022 12:32
 Operator : CG/JU
 Sample : N2713-14MSD
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :

BNA_P

ClientSampleId :

DBSL3MSD

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 05/18/2022
 Supervised By :mohammad ahmed 05/18/2022

Quant Time: May 18 00:57:55 2022

Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051322.M


Quant Title : SVOA CALIBRATION

QLast Update : Fri May 13 15:04:31 2022

Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.575	142	624166	25.743	ng/ul#	96
39) 1,2,4,5-Tetrachloroben...	12.728	216	342131	25.131	ng/ul	92
40) Hexachlorocyclopentadiene	12.710	237	170056	22.976	ng/ul	95
41) 2,4,6-Trichlorophenol	12.963	196	227979	26.864	ng/ul#	84
42) 2,4,5-Trichlorophenol	13.039	196	244353	25.788	ng/ul#	86
43) 1,1'-Biphenyl	13.363	154	832032	25.409	ng/ul#	92
44) 2-Chloronaphthalene	13.404	162	648862	25.655	ng/ul	94
45) 2-Nitroaniline	13.610	65	226109	27.164	ng/ul#	78
47) Dimethylphthalate	13.986	163	824752	25.296	ng/ul#	92
48) 2,6-Dinitrotoluene	14.104	165	177403	26.692	ng/ul	96
50) Acenaphthylene	14.251	152	1058416	25.806	ng/ul	95
51) 3-Nitroaniline	14.433	138	163697	26.242	ng/ul#	86
52) Acenaphthene	14.592	153	677060	25.305	ng/ul	96
53) 2,4-Dinitrophenol	14.645	184	73283	19.749	ng/ul#	77
55) 4-Nitrophenol	14.739	109	134078	25.774	ng/ul#	45
56) Dibenzofuran	14.927	168	998997	25.477	ng/ul	99
57) 2,4-Dinitrotoluene	14.892	165	255826	26.738	ng/ul#	80
58) 2,3,4,6-Tetrachlorophenol	15.157	232	212376	26.065	ng/ul#	85
59) Diethylphthalate	15.351	149	852377	25.621	ng/ul	94
61) Fluorene	15.580	166	816139	25.452	ng/ul	92
62) 4-Chlorophenyl-phenyle...	15.574	204	420652	25.219	ng/ul	99
63) 4-Nitroaniline	15.598	138	182612	29.649	ng/ul#	78
66) 4,6-Dinitro-2-methylph...	15.657	198	134649	23.259	ng/ul#	87
67) N-Nitrosodiphenylamine	15.786	169	708629	25.457	ng/ul	95
68) 4-Bromophenyl-phenylether	16.468	248	258971	25.334	ng/ul	95
69) Hexachlorobenzene	16.592	284	291856	24.822	ng/ul#	91
70) Atrazine	16.745	200	276589	24.804	ng/ul#	89
71) Pentachlorophenol	16.933	266	172141	23.333	ng/ul	88
72) Phenanthrene	17.327	178	1329562	25.100	ng/ul	99
74) Anthracene	17.416	178	1340512	25.226	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.328	216	353146	23.411	ng/uL#	85
76) Pentachlorobenzene	14.851	250	337440	24.655	ng/uL	93
77) Carbazole	17.686	167	1184743	25.368	ng/ul#	95
78) Di-n-butylphthalate	18.239	149	1489121	26.180	ng/ul#	93
80) Fluoranthene	19.292	202	1612180	26.069	ng/ul	96
82) Pyrene	19.645	202	1670620	25.886	ng/ul#	95
83) Butylbenzylphthalate	20.515	149	700950	27.162	ng/ul#	81
84) 3,3'-Dichlorobenzidine	21.280	252	473080	23.489	ng/ul#	97
85) Benzo(a)anthracene	21.351	228	1608334	25.667	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.274	149	1048514	27.688	ng/ul#	80
87) Chrysene	21.404	228	1578863	25.616	ng/ul	99
89) Di-n-octyl phthalate	22.209	149	1764079	24.547	ng/ul	100
90) Benzo(b)fluoranthene	23.056	252	1596921	25.009	ng/ul	100
91) Benzo(k)fluoranthene	23.103	252	1543944	25.441	ng/ul#	97
93) Benzo(a)pyrene	23.692	252	1521217	25.642	ng/ul#	98
94) Indeno(1,2,3-cd)pyrene	26.339	276	1472351	26.381	ng/ul#	93
95) Dibenzo(a,h)anthracene	26.350	278	1268851	26.201	ng/ul#	92
96) Benzo(g,h,i)perylene	27.115	276	1221964	26.221	ng/ul#	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 9.701 ng/uL
RT: 3.375 min Scan# 49
Delta R.T. -0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Instrument :
BNA_P
ClientSampleId :
DBSL3MSD



Tgt Ion: 88 Resp: 34609
Ion Ratio Lower Upper
88 100
43 40.8 6.8 10.2
58 85.6 28.4 42.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022



#5
Pyridine
Concen: 24.087 ng/uL
RT: 3.775 min Scan# 117
Delta R.T. -0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32

Tgt Ion: 79 Resp: 244796
Ion Ratio Lower Upper
79 100
52 96.3 35.5 53.3#
51 43.0 20.2 30.4#









#6
Benzaldehyde
Concen: 31.291 ng/ul
RT: 7.034 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32

Instrument :
BNA_P
ClientSampleId :
DBSL3MSD

Tgt Ion: 77 Resp: 20381
Ion Ratio Lower Upper
77 100
105 93.0 77.0 115.4
106 88.5 77.0 115.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022



#8
Phenol
Concen: 26.158 ng/ul
RT: 7.087 min Scan# 680
Delta R.T. -0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32


Tgt Ion: 94 Resp: 347704
Ion Ratio Lower Upper
94 100
65 31.8 19.0 28.6
66 44.6 34.6 51.8





#10
Bis(2-Chloroethyl)ether
Concen: 25.532 ng/ul
RT: 7.316 min Scan# 719
Delta R.T. -0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Instrument :
BNA_P
ClientSampleId :
DBSL3MSD




Tgt Ion: 93 Resp: 26683
Ion Ratio Lower Upper
93 100
63 86.6 45.0 67.4
95 31.7 26.3 39.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022




#12
2-Chlorophenol
Concen: 26.209 ng/ul
RT: 7.457 min Scan# 743
Delta R.T. -0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32



Tgt Ion:128 Resp: 257963
Ion Ratio Lower Upper
128 100
64 56.7 45.3 67.9
130 31.8 29.0 43.6





#13

2-Methylphenol

Concen: 25.981 ng/ul

RT: 8.340 min Scan# 893

Delta R.T. -0.006 min

Lab File: BP010387.D

Acq: 17 May 2022 12:32

Instrument :

BNA_P

ClientSampleId :

DBSL3MSD

Tgt Ion:108 Resp: 257810

Ion Ratio Lower Upper





108 100

107 91.6 70.8 106.2

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022



#14

2,2'-oxybis(1-Chloropropane)

Concen: 25.398 ng/ul

RT: 8.428 min Scan# 908

Delta R.T. -0.000 min

Lab File: BP010387.D

Acq: 17 May 2022 12:32




Tgt Ion: 45 Resp: 432614


Ion Ratio Lower Upper

45 100

77 12.8 17.1 25.7#


79 9.4 13.2 19.8#





#16
Acetophenone
Concen: 25.523 ng/ul
RT: 8.722 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Instrument :
BNA_P
ClientSampleId :
DBSL3MSD



Tgt Ion:105 Resp: 419904
Ion Ratio Lower Upper
105 100
77 81.0 53.4 80.2
51 33.0 13.8 20.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022



#17
N-Nitroso-di-n-propylamine
Concen: 25.721 ng/ul
RT: 8.704 min Scan# 955
Delta R.T. -0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32


Tgt Ion: 70 Resp: 226470
Ion Ratio Lower Upper
70 100
42 64.7 31.8 47.8#
101 8.6 15.6 23.4#
130 21.2 13.7 20.5#





#18
4-Methylphenol
Concen: 25.899 ng/ul
RT: 8.669 min Scan# 949
Delta R.T. -0.000 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Instrument :
BNA_P
ClientSampleId :
DBSL3MSD




Tgt Ion:108 Resp: 28404
Ion Ratio Lower Upper
108 100
107 115.5 83.6 125.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022




#19
Hexachloroethane
Concen: 24.921 ng/ul
RT: 8.975 min Scan# 1001
Delta R.T. -0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Tgt Ion:117 Resp: 106680
Ion Ratio Lower Upper
117 100
201 106.8 106.5 159.7
199 65.0 59.0 88.6





#22
Nitrobenzene
Concen: 25.402 ng/ul
RT: 9.098 min Scan# 1022
Delta R.T. -0.000 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Instrument :
BNA_P
ClientSampleId :
DBSL3MSD




Tgt Ion: 77 Resp: 33551
Ion Ratio Lower Upper
77 100
123 40.8 42.3 63.5
65 14.4 5.8 8.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022




#23
Isophorone
Concen: 25.661 ng/ul
RT: 9.628 min Scan# 1112
Delta R.T. -0.000 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32



Tgt Ion: 82 Resp: 626464
Ion Ratio Lower Upper
82 100
95 7.8 7.8 11.8
138 16.6 12.9 19.3





#25
2-Nitrophenol
Concen: 26.077 ng/ul
RT: 9.810 min Scan# 1143
Delta R.T. -0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32

Instrument :

BNA_P

ClientSampleId :





DBSL3MSD

Tgt Ion:139 Resp: 152517
Ion Ratio Lower Upper
139 100
65 59.6 49.4 74.2
109 32.2 44.7 67.1

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022



#26
2,4-Dimethylphenol
Concen: 25.904 ng/ul m
RT: 9.875 min Scan# 1154
Delta R.T. -0.000 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32


Tgt Ion:107 Resp: 328738
Ion Ratio Lower Upper
107 100
121 48.2 34.1 51.1
122 80.6 46.3 69.5





#27
Bis(2-Chloroethoxy)methane
Concen: 25.467 ng/ul
RT: 10.104 min Scan# 1193
Delta R.T. -0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Instrument :
BNA_P
ClientSampleId :
DBSL3MSD



Tgt Ion: 93 Resp: 370188
Ion Ratio Lower Upper
93 100
95 32.5 26.2 39.4
123 10.4 6.2 9.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022



#29
2,4-Dichlorophenol
Concen: 26.225 ng/ul
RT: 10.345 min Scan# 1234
Delta R.T. -0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32


Tgt Ion:162 Resp: 270046
Ion Ratio Lower Upper
162 100
164 66.6 37.0 55.6#
98 39.7 31.0 46.6





#30
Naphthalene
Concen: 25.312 ng/ul
RT: 10.745 min Scan# 1302
Delta R.T. -0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Instrument :
BNA_P
ClientSampleId :
DBSL3MSD



Tgt Ion:128 Resp: 858592
Ion Ratio Lower Upper
128 100
129 10.7 8.6 13.0
127 13.4 12.2 18.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022



#32
4-Chloroaniline
Concen: 22.725 ng/ul
RT: 10.857 min Scan# 1321
Delta R.T. -0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32


Tgt Ion:127 Resp: 337742
Ion Ratio Lower Upper
127 100
129 32.8 24.6 37.0





#33
Hexachlorobutadiene
Concen: 24.871 ng/ul
RT: 11.039 min Scan# 1352
Delta R.T. -0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Instrument :
BNA_P
ClientSampleId :
DBSL3MSD




Tgt Ion:225 Resp: 181478
Ion Ratio Lower Upper
225 100
223 65.1 53.4 80.0
227 62.3 53.8 80.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022




#34
Caprolactam
Concen: 26.146 ng/ul
RT: 11.628 min Scan# 1452
Delta R.T. -0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32



Tgt Ion:113 Resp: 93959
Ion Ratio Lower Upper
113 100
55 203.2 58.2 87.2#
56 148.3 68.6 103.0#









#35
4-Chloro-3-methylphenol
Concen: 26.265 ng/ul
RT: 11.986 min Scan# 1513
Delta R.T. -0.000 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32

Instrument :
BNA_P
ClientSampleId :
DBSL3MSD

Tgt Ion:107 Resp: 31638
Ion Ratio Lower Upper
107 100
144 23.5 15.8 23.8
142 76.8 38.9 58.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022



#36
2-Methylnaphthalene
Concen: 25.156 ng/ul
RT: 12.357 min Scan# 1576
Delta R.T. -0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32

Tgt Ion:142 Resp: 604093
Ion Ratio Lower Upper
142 100
141 87.6 64.3 96.5





#37
1-Methylnaphthalene
Concen: 25.743 ng/ul
RT: 12.575 min Scan# 1613
Delta R.T. -0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32

Instrument :

BNA_P

ClientSampleId :





DBSL3MSD

Tgt Ion:142 Resp: 624160
Ion Ratio Lower Upper
142 100
141 90.8 69.8 104.8
116 3.9 5.6 8.4

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022



#39
1,2,4,5-Tetrachlorobenzene
Concen: 25.131 ng/ul
RT: 12.728 min Scan# 1639
Delta R.T. -0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32


Tgt Ion:216 Resp: 342131
Ion Ratio Lower Upper
216 100
214 79.8 58.2 87.2
179 20.8 14.2 21.2
108 16.4 15.3 22.9





#40
Hexachlorocyclopentadiene
Concen: 22.976 ng/ul
RT: 12.710 min Scan# 1636
Delta R.T. -0.000 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Instrument :
BNA_P
ClientSampleId :
DBSL3MSD




Tgt Ion:237 Resp: 170050
Ion Ratio Lower Upper
237 100
235 61.6 53.2 79.8
272 13.4 10.8 16.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022




#41
2,4,6-Trichlorophenol
Concen: 26.864 ng/ul
RT: 12.963 min Scan# 1679
Delta R.T. -0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Tgt Ion:196 Resp: 227979
Ion Ratio Lower Upper
196 100
198 95.5 64.1 96.1
200 31.3 19.9 29.9





#42
2,4,5-Trichlorophenol
Concen: 25.788 ng/ul
RT: 13.039 min Scan# 1692
Delta R.T. -0.000 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Instrument :
BNA_P
ClientSampleId :
DBSL3MSD




Tgt Ion:196 Resp: 24435
Ion Ratio Lower Upper
196 100
198 97.8 68.6 103.0
200 30.4 32.8 49.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022




#43
1,1'-Biphenyl
Concen: 25.409 ng/ul
RT: 13.363 min Scan# 1747
Delta R.T. -0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Tgt Ion:154 Resp: 832032
Ion Ratio Lower Upper
154 100
153 40.9 28.6 43.0
76 13.2 8.5 12.7





#44
2-Chloronaphthalene
Concen: 25.655 ng/ul
RT: 13.404 min Scan# 1754
Delta R.T. -0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Instrument :
BNA_P
ClientSampleId :
DBSL3MSD



Tgt Ion:162 Resp: 648862
Ion Ratio Lower Upper
162 100
164 33.2 27.5 41.3
127 40.2 37.1 55.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022



#45
2-Nitroaniline
Concen: 27.164 ng/ul
RT: 13.610 min Scan# 1789
Delta R.T. -0.000 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32

Tgt Ion: 65 Resp: 226109
Ion Ratio Lower Upper
65 100
92 63.9 77.9 116.9#
138 91.0 81.5 122.3





#47
Dimethylphthalate
Concen: 25.296 ng/ul
RT: 13.986 min Scan# 1853
Delta R.T. -0.000 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32

Instrument :

BNA_P

ClientSampleId :





DBSL3MSD

Tgt Ion:163 Resp: 824752
Ion Ratio Lower Upper
163 100
194 4.2 4.1 6.1
164 9.9 11.3 16.9

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022



#48
2,6-Dinitrotoluene
Concen: 26.692 ng/ul
RT: 14.104 min Scan# 1873
Delta R.T. -0.000 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32


Tgt Ion:165 Resp: 177403
Ion Ratio Lower Upper
165 100
89 61.3 51.7 77.5
121 20.9 18.3 27.5





#50
Acenaphthylene
Concen: 25.806 ng/ul
RT: 14.251 min Scan# 1898
Delta R.T. -0.000 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Instrument :
BNA_P
ClientSampleId :
DBSL3MSD



Tgt Ion:152 Resp: 1058410
Ion Ratio Lower Upper
152 100
151 19.5 18.3 27.5
153 13.1 11.0 16.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022



#51
3-Nitroaniline
Concen: 26.242 ng/ul
RT: 14.433 min Scan# 1929
Delta R.T. -0.000 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32


Tgt Ion:138 Resp: 163697
Ion Ratio Lower Upper
138 100
108 11.9 15.8 23.8#
92 128.7 117.1 175.7





#52
Acenaphthene
Concen: 25.305 ng/ul
RT: 14.592 min Scan# 1956
Delta R.T. -0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Instrument :
BNA_P
ClientSampleId :
DBSL3MSD




Tgt Ion:153 Resp: 677060
Ion Ratio Lower Upper
153 100
152 47.8 34.3 51.5
154 89.4 70.0 105.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022




#53
2,4-Dinitrophenol
Concen: 19.749 ng/ul
RT: 14.645 min Scan# 1965
Delta R.T. -0.000 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Tgt Ion:184 Resp: 73283
Ion Ratio Lower Upper
184 100
63 68.2 34.6 52.0#
154 60.4 57.7 86.5





#55
4-Nitrophenol
Concen: 25.774 ng/ul
RT: 14.739 min Scan# 1981
Delta R.T. -0.000 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Instrument :
BNA_P
ClientSampleId :
DBSL3MSD



Tgt Ion:109 Resp: 134078
Ion Ratio Lower Upper
109 100
139 114.1 61.4 92.2
65 140.1 65.3 97.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022



#56
Dibenzofuran
Concen: 25.477 ng/ul
RT: 14.927 min Scan# 2013
Delta R.T. -0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32


Tgt Ion:168 Resp: 998997
Ion Ratio Lower Upper
168 100
139 37.1 30.1 45.1





#57
2,4-Dinitrotoluene
Concen: 26.738 ng/ul
RT: 14.892 min Scan# 2007
Delta R.T. -0.000 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Instrument :
BNA_P
ClientSampleId :
DBSL3MSD




Tgt Ion:165 Resp: 255820
Ion Ratio Lower Upper
165 100
63 49.8 29.5 44.3#
182 2.7 2.8 4.2#

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 26.065 ng/ul
RT: 15.157 min Scan# 2052
Delta R.T. -0.000 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32



Tgt Ion:232 Resp: 212376
Ion Ratio Lower Upper
232 100
131 44.4 25.4 38.2#
130 2.5 4.2 6.2#
166 29.8 28.1 42.1#





#59
Diethylphthalate
Concen: 25.621 ng/ul
RT: 15.351 min Scan# 2085
Delta R.T. -0.000 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32

Instrument :

BNA_P

ClientSampleId :





DBSL3MSD

Tgt Ion:149 Resp: 85237
Ion Ratio Lower Upper
149 100
177 23.3 22.0 33.0
150 12.4 10.2 15.2

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022



#61
Fluorene
Concen: 25.452 ng/ul
RT: 15.580 min Scan# 2124
Delta R.T. -0.000 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32


Tgt Ion:166 Resp: 816139
Ion Ratio Lower Upper
166 100
165 96.3 70.3 105.5
167 13.4 9.5 14.3





#62
4-Chlorophenyl-phenylether
Concen: 25.219 ng/ul
RT: 15.574 min Scan# 2122
Delta R.T. -0.000 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Instrument :
BNA_P
ClientSampleId :
DBSL3MSD




Tgt Ion: 204 Resp: 420652
Ion Ratio Lower Upper
204 100
206 33.8 27.6 41.4
141 57.8 46.6 69.8

Manual Integrations
APPROVED



Reviewed By : Jagrut Upadhyay 05/18/2022
Supervised By : mohammad ahmed 05/18/2022




#63
4-Nitroaniline
Concen: 29.649 ng/ul
RT: 15.598 min Scan# 2127
Delta R.T. -0.000 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Tgt Ion: 138 Resp: 182612
Ion Ratio Lower Upper
138 100
92 52.7 41.6 62.4
108 86.9 97.6 146.4





#66
4,6-Dinitro-2-methylphenol
Concen: 23.259 ng/ul
RT: 15.657 min Scan# 2137
Delta R.T. -0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Instrument :
BNA_P
ClientSampleId :
DBSL3MSD




Tgt Ion:198 Resp: 134649
Ion Ratio Lower Upper
198 100
182 3.9 6.0 9.0
77 28.7 17.7 26.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022




#67
N-Nitrosodiphenylamine
Concen: 25.457 ng/ul
RT: 15.786 min Scan# 2159
Delta R.T. -0.000 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Tgt Ion:169 Resp: 708629
Ion Ratio Lower Upper
169 100
168 65.2 48.3 72.5
167 34.7 29.6 44.4





#68
4-Bromophenyl-phenylether
Concen: 25.334 ng/ul
RT: 16.468 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Instrument :
BNA_P
ClientSampleId :
DBSL3MSD



Tgt Ion:248 Resp: 25897
Ion Ratio Lower Upper
248 100
250 96.5 79.4 119.0
141 77.4 67.2 100.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022



#69
Hexachlorobenzene
Concen: 24.822 ng/ul
RT: 16.592 min Scan# 2296
Delta R.T. -0.000 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32


Tgt Ion:284 Resp: 291856
Ion Ratio Lower Upper
284 100
142 33.3 19.2 28.8#
249 33.0 27.4 41.2





#70
Atrazine
Concen: 24.804 ng/ul
RT: 16.745 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Instrument :
BNA_P
ClientSampleId :
DBSL3MSD



Tgt Ion:200 Resp: 276589
Ion Ratio Lower Upper
200 100
173 27.3 19.2 28.8
215 48.9 32.2 48.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022



#71
Pentachlorophenol
Concen: 23.333 ng/ul
RT: 16.933 min Scan# 2354
Delta R.T. -0.000 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32


Tgt Ion:266 Resp: 172141
Ion Ratio Lower Upper
266 100
264 66.0 57.3 85.9
268 67.0 65.7 98.5





#72
Phenanthrene
Concen: 25.100 ng/u1
RT: 17.327 min Scan# 2420
Delta R.T. -0.000 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Instrument :
BNA_P
ClientSampleId :
DBSL3MSD



Tgt Ion:178 Resp: 1329562
Ion Ratio Lower Upper
178 100
179 15.0 11.7 17.5
176 18.7 14.5 21.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022



#74
Anthracene
Concen: 25.226 ng/u1
RT: 17.416 min Scan# 2436
Delta R.T. -0.000 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32

Tgt Ion:178 Resp: 1340512
Ion Ratio Lower Upper
178 100
179 15.6 12.3 18.5
176 18.1 13.4 20.2





#75
1,2,3,4-Tetrachlorobenzene
Concen: 23.411 ng/uL
RT: 13.328 min Scan# 1741
Delta R.T. -0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32

Instrument :

BNA_P

ClientSampleId :





DBSL3MSD

Tgt Ion:216 Resp: 353140
Ion Ratio Lower Upper
216 100
214 78.2 50.4 75.6
218 48.8 35.2 52.8

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022



#76
Pentachlorobenzene
Concen: 24.655 ng/uL
RT: 14.851 min Scan# 2000
Delta R.T. -0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32

Tgt Ion:250 Resp: 337440
Ion Ratio Lower Upper
250 100
248 61.2 43.5 65.3
252 61.5 46.5 69.7






#77
 Carbazole
 Concen: 25.368 ng/ul
 RT: 17.686 min Scan# 2482
 Delta R.T. -0.000 min
 Lab File: BP010387.D
 Acq: 17 May 2022 12:32

Instrument :

BNA_P

ClientSampleId :

DBSL3MSD






Tgt Ion:167 Resp: 118474
 Ion Ratio Lower Upper
 167 100
 166 21.4 15.8 23.8
 139 13.0 8.3 12.5

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 05/18/2022
 Supervised By :mohammad ahmed 05/18/2022



#78
 Di-n-butylphthalate
 Concen: 26.180 ng/ul
 RT: 18.239 min Scan# 2576
 Delta R.T. -0.000 min
 Lab File: BP010387.D
 Acq: 17 May 2022 12:32


Tgt Ion:149 Resp: 1489121
 Ion Ratio Lower Upper
 149 100
 150 9.0 9.6 14.4#
 104 5.3 6.2 9.4#





#80
Fluoranthene
Concen: 26.069 ng/u1
RT: 19.292 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Instrument :
BNA_P
ClientSampleId :
DBSL3MSD




Tgt Ion:202 Resp: 1612180
Ion Ratio Lower Upper
202 100
101 9.7 9.6 14.4
100 7.8 6.3 9.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022




#82
Pyrene
Concen: 25.886 ng/u1
RT: 19.645 min Scan# 2815
Delta R.T. 0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Tgt Ion:202 Resp: 1670620
Ion Ratio Lower Upper
202 100
101 11.6 10.6 15.8
100 9.6 9.8 14.6#





#83
Butylbenzylphthalate
Concen: 27.162 ng/ul
RT: 20.515 min Scan# 2962
Delta R.T. 0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Instrument :
BNA_P
ClientSampleId :
DBSL3MSD




Tgt Ion:149 Resp: 700950
Ion Ratio Lower Upper
149 100
91 72.1 71.1 106.7
206 22.8 28.8 43.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022




#84
3,3'-Dichlorobenzidine
Concen: 23.489 ng/ul
RT: 21.280 min Scan# 3093
Delta R.T. -0.000 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Tgt Ion:252 Resp: 473080
Ion Ratio Lower Upper
252 100
254 63.6 49.8 74.6
126 10.9 4.3 6.5





#85
Benzo(a)anthracene
Concen: 25.667 ng/ul
RT: 21.351 min Scan# 3105
Delta R.T. 0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Instrument :
BNA_P
ClientSampleId :
DBSL3MSD



Tgt Ion:228 Resp: 1608334
Ion Ratio Lower Upper
228 100
229 19.4 18.2 27.2
226 26.4 22.8 34.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 27.688 ng/ul
RT: 21.274 min Scan# 3092
Delta R.T. 0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32


Tgt Ion:149 Resp: 1048514
Ion Ratio Lower Upper
149 100
167 27.2 32.8 49.2#
279 4.5 5.6 8.4#





#87
Chrysene
Concen: 25.616 ng/ul
RT: 21.404 min Scan# 3114
Delta R.T. 0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Instrument :
BNA_P
ClientSampleId :
DBSL3MSD



Tgt Ion:228 Resp: 157886
Ion Ratio Lower Upper
228 100
226 28.6 22.7 34.1
229 19.8 15.2 22.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022



#89
Di-n-octyl phthalate
Concen: 24.547 ng/ul
RT: 22.209 min Scan# 3251
Delta R.T. 0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32

Tgt Ion:149 Resp: 1764079









#90
Benzo(b)fluoranthene
Concen: 25.009 ng/ul
RT: 23.056 min Scan# 31
Delta R.T. 0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32

Instrument :
BNA_P
ClientSampleId :
DBSL3MSD

Tgt Ion:252 Resp: 159692
Ion Ratio Lower Upper
252 100
253 21.7 17.4 26.0
125 9.1 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022



#91
Benzo(k)fluoranthene
Concen: 25.441 ng/ul
RT: 23.103 min Scan# 3403
Delta R.T. 0.006 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32


Tgt Ion:252 Resp: 1543944
Ion Ratio Lower Upper
252 100
253 21.8 17.9 26.9
125 8.8 5.1 7.7#





#93
Benzo(a)pyrene
Concen: 25.642 ng/ul
RT: 23.692 min Scan# 3503
Delta R.T. 0.012 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Instrument :
BNA_P
ClientSampleId :
DBSL3MSD




Tgt Ion:252 Resp: 1521217
Ion Ratio Lower Upper
252 100
253 22.1 17.5 26.3
125 10.4 6.4 9.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 26.381 ng/ul
RT: 26.339 min Scan# 3953
Delta R.T. 0.017 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Tgt Ion:276 Resp: 1472351
Ion Ratio Lower Upper
276 100
138 21.8 12.3 18.5#
277 24.8 19.3 28.9





#95
Dibenzo(a,h)anthracene
Concen: 26.201 ng/ul
RT: 26.350 min Scan# 3955
Delta R.T. 0.017 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32




Instrument :
BNA_P
ClientSampleId :
DBSL3MSD




Tgt Ion:278 Resp: 126885
Ion Ratio Lower Upper
278 100
139 15.0 6.6 10.0
279 24.7 18.4 27.6

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 05/18/2022
Supervised By :mohammad ahmed 05/18/2022



#96
Benzo(g,h,i)perylene
Concen: 26.221 ng/ul
RT: 27.115 min Scan# 4085
Delta R.T. 0.023 min
Lab File: BP010387.D
Acq: 17 May 2022 12:32



Tgt Ion:276 Resp: 1221964
Ion Ratio Lower Upper
276 100
138 18.1 11.6 17.4
277 23.6 15.9 23.9

