

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052424\
 Data File : BP020447.D
 Acq On : 24 May 2024 10:01
 Operator : MA/JU
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC2.5

Quant Time: May 25 01:11:15 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 25 01:10:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.763	152	331060	20.000	ng	0.00
21) Naphthalene-d8	10.540	136	1384482	20.000	ng	0.00
39) Acenaphthene-d10	14.410	164	963642	20.000	ng	0.00
64) Phenanthrene-d10	17.216	188	1942978	20.000	ng	0.00
76) Chrysene-d12	21.692	240	1498845	20.000	ng	0.00
86) Perylene-d12	25.062	264	1643040	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	0.000	112	0d	0.000	ng	
7) Phenol-d6	0.000	99	0d	0.000	ng	
23) Nitrobenzene-d5	0.000	82	0d	0.000	ng	
42) 2,4,6-Tribromophenol	0.000	330	0d	0.000	ng	
45) 2-Fluorobiphenyl	0.000	172	0d	0.000	ng	
79) Terphenyl-d14	0.000	244	0d	0.000	ng	
Target Compounds						
19) n-Nitroso-di-n-propyla...	8.557	70	44260	2.486	ng	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052424\
Data File : BP020447.D
Acq On : 24 May 2024 10:01
Operator : MA/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Quant Time: May 25 01:11:15 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat May 25 01:10:28 2024
Response via : Initial Calibration

