

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052522\
 Data File : BP010541.D
 Acq On : 25 May 2022 18:08
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020670

Quant Time: May 25 23:27:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 24 05:50:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.881	152	120976	20.000	ng/ul	0.00
20) Naphthalene-d8	10.681	136	538341	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.516	164	384534	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.269	188	886143	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.351	240	947572	20.000	ng/ul	# 0.00
88) Perylene-d12	23.768	264	849640	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.422	96	431	0.152	ng/uL	0.08
4) Pyridine-d5	3.752	84	143034	17.749	ng/ul	0.00
7) Phenol-d5	7.046	99	179217	17.650	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.210	67	127768	18.729	ng/ul	0.00
11) 2-Chlorophenol-d4	7.416	132	145542	18.732	ng/ul	0.00
15) 4-Methylphenol-d8	8.593	113	153819	17.826	ng/ul	0.00
21) Nitrobenzene-d5	9.040	128	78401	18.817	ng/ul	0.00
24) 2-Nitrophenol-d4	9.763	143	83306	18.867	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.304	165	153406	18.374	ng/ul	0.00
31) 4-Chloroaniline-d4	10.881	131	606	0.049	ng/ul	0.05
46) Dimethylphthalate-d6	13.922	166	537319	18.966	ng/ul	0.00
49) Acenaphthylene-d8	14.204	160	606539	18.560	ng/ul	0.00
54) 4-Nitrophenol-d4	14.716	143	86158	17.086	ng/ul	0.00
60) Fluorene-d10	15.510	176	459895	18.659	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.634	200	89908	17.348	ng/ul	0.00
73) Anthracene-d10	17.363	188	751408	18.726	ng/ul	0.00
81) Pyrene-d10	19.598	212	931080	18.502	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.610	264	801460	18.786	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.369	88	23505	7.995	ng/uL#	23
5) Pyridine	3.775	79	152540	18.214	ng/ul#	31
6) Benzaldehyde	7.022	77	113339	21.115	ng/ul	94
8) Phenol	7.075	94	195298	17.829	ng/ul#	88
10) Bis(2-Chloroethyl)ether	7.305	93	160017	18.579	ng/ul#	74
12) 2-Chlorophenol	7.446	128	154203	19.011	ng/ul	97
13) 2-Methylphenol	8.328	108	149042	18.226	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.416	45	282159	20.101	ng/ul#	82
16) Acetophenone	8.704	105	252610	18.632	ng/ul#	73
17) N-Nitroso-di-n-propyla...	8.693	70	143209	19.737	ng/ul#	69
18) 4-Methylphenol	8.652	108	161017	17.816	ng/ul	86
19) Hexachloroethane	8.963	117	69286	19.641	ng/ul	84
22) Nitrobenzene	9.081	77	209117	19.099	ng/ul#	79
23) Isophorone	9.610	82	387332	19.139	ng/ul#	98
25) 2-Nitrophenol	9.799	139	89316	18.423	ng/ul#	86
26) 2,4-Dimethylphenol	9.857	107	200022	19.014	ng/ul#	83
27) Bis(2-Chloroethoxy)met...	10.093	93	225189	18.688	ng/ul#	97
29) 2,4-Dichlorophenol	10.334	162	156030	18.279	ng/ul#	84
30) Naphthalene	10.734	128	519169	18.464	ng/ul	97
32) 4-Chloroaniline	10.846	127	218880	17.767	ng/ul	97
33) Hexachlorobutadiene	11.022	225	113302	18.732	ng/ul	97
34) Caprolactam	11.622	113	45889	15.404	ng/ul#	1
35) 4-Chloro-3-methylphenol	11.969	107	187635	18.791	ng/ul#	71
36) 2-Methylnaphthalene	12.340	142	371350	18.655	ng/ul	91

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052522\
 Data File : BP010541.D
 Acq On : 25 May 2022 18:08
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020670

Quant Time: May 25 23:27:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 24 05:50:40 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.557	142	375181	18.667	ng/ul#	95
39) 1,2,4,5-Tetrachloroben...	12.710	216	209006	17.811	ng/ul#	93
40) Hexachlorocyclopentadiene	12.692	237	135186	21.190	ng/ul	97
41) 2,4,6-Trichlorophenol	12.951	196	134522	18.391	ng/ul#	82
42) 2,4,5-Trichlorophenol	13.022	196	145813	17.853	ng/ul#	87
43) 1,1'-Biphenyl	13.351	154	514650	18.234	ng/ul	94
44) 2-Chloronaphthalene	13.392	162	399595	18.330	ng/ul	95
45) 2-Nitroaniline	13.592	65	143405	19.988	ng/ul#	73
47) Dimethylphthalate	13.975	163	528828	18.818	ng/ul#	91
48) 2,6-Dinitrotoluene	14.092	165	109612	19.134	ng/ul	99
50) Acenaphthylene	14.234	152	660715	18.689	ng/ul	95
51) 3-Nitroaniline	14.422	138	98206	18.265	ng/ul#	84
52) Acenaphthene	14.575	153	427003	18.515	ng/ul	98
53) 2,4-Dinitrophenol	14.634	184	66376	20.753	ng/ul#	78
55) 4-Nitrophenol	14.728	109	82348	18.365	ng/ul#	50
56) Dibenzofuran	14.916	168	628490	18.596	ng/ul	97
57) 2,4-Dinitrotoluene	14.881	165	164670	19.968	ng/ul#	78
58) 2,3,4,6-Tetrachlorophenol	15.145	232	132067	18.805	ng/ul#	85
59) Diethylphthalate	15.334	149	555719	19.379	ng/ul#	92
61) Fluorene	15.563	166	522598	18.908	ng/ul	91
62) 4-Chlorophenyl-phenyle...	15.557	204	268111	18.648	ng/ul	97
63) 4-Nitroaniline	15.581	138	96878	18.248	ng/ul#	80
66) 4,6-Dinitro-2-methylph...	15.645	198	89645	17.345	ng/ul#	85
67) N-Nitrosodiphenylamine	15.769	169	452110	18.193	ng/ul	95
68) 4-Bromophenyl-phenylether	16.451	248	164751	18.053	ng/ul	97
69) Hexachlorobenzene	16.575	284	188828	17.988	ng/ul#	90
70) Atrazine	16.728	200	181210	18.202	ng/ul#	89
71) Pentachlorophenol	16.922	266	109806	16.671	ng/ul#	85
72) Phenanthrene	17.310	178	872931	18.458	ng/ul	99
74) Anthracene	17.398	178	883859	18.630	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.316	216	223217	16.575	ng/uL	89
76) Pentachlorobenzene	14.839	250	217356	17.788	ng/uL	92
77) Carbazole	17.669	167	782348	18.764	ng/ul#	96
78) Di-n-butylphthalate	18.222	149	994575	19.585	ng/ul#	94
80) Fluoranthene	19.274	202	1088023	18.605	ng/ul	97
82) Pyrene	19.627	202	1140814	18.693	ng/ul#	94
83) Butylbenzylphthalate	20.498	149	479456	19.647	ng/ul#	82
84) 3,3'-Dichlorobenzidine	21.269	252	337057	17.698	ng/ul#	96
85) Benzo(a)anthracene	21.333	228	1122663	18.946	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.257	149	712554	19.898	ng/ul#	81
87) Chrysene	21.386	228	1101125	18.892	ng/ul	98
89) Di-n-octyl phthalate	22.186	149	1219880	19.139	ng/ul	100
90) Benzo(b)fluoranthene	23.033	252	1091225	19.269	ng/ul	99
91) Benzo(k)fluoranthene	23.080	252	1007835	18.725	ng/ul#	99
93) Benzo(a)pyrene	23.662	252	990091	18.817	ng/ul#	98
94) Indeno(1,2,3-cd)pyrene	26.286	276	948324	19.159	ng/ul#	95
95) Dibenzo(a,h)anthracene	26.304	278	798461	18.590	ng/ul#	93
96) Benzo(g,h,i)perylene	27.062	276	791672	19.154	ng/ul#	92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052522\
Data File : BP010541.D
Acq On : 25 May 2022 18:08
Operator : CG/JU
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTD020670

Quant Time: May 25 23:27:23 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051322.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue May 24 05:50:40 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.995 ng/uL
RT: 3.369 min Scan# 41
Delta R.T. -0.000 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SSTD020670


Tgt Ion: 88 Resp: 23505
Ion Ratio Lower Upper
88 100
43 38.7 6.8 10.2#
58 79.1 28.4 42.6#



#5
Pyridine
Concen: 18.214 ng/uL
RT: 3.775 min Scan# 117
Delta R.T. -0.000 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08

Tgt Ion: 79 Resp: 152540
Ion Ratio Lower Upper
79 100
52 99.8 35.5 53.3#
51 46.0 20.2 30.4#









#6
Benzaldehyde
Concen: 21.115 ng/ul
RT: 7.022 min Scan# 609
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SSTD020670


Tgt Ion: 77 Resp: 113339
Ion Ratio Lower Upper
77 100
105 93.2 77.0 115.4
106 87.9 77.0 115.4



#8
Phenol
Concen: 17.829 ng/ul
RT: 7.075 min Scan# 678
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08

Tgt Ion: 94 Resp: 195298
Ion Ratio Lower Upper
94 100
65 35.1 19.0 28.6
66 46.6 34.6 51.8









#10
Bis(2-Chloroethyl)ether
Concen: 18.579 ng/ul
RT: 7.305 min Scan# 717
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SSTD020670


Tgt Ion: 93 Resp: 160017
Ion Ratio Lower Upper
93 100
63 85.5 45.0 67.4#
95 31.9 26.3 39.5



#12
2-Chlorophenol
Concen: 19.011 ng/ul
RT: 7.446 min Scan# 741
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08

Tgt Ion:128 Resp: 154203
Ion Ratio Lower Upper
128 100
64 57.9 45.3 67.9
130 32.8 29.0 43.6









#13
2-Methylphenol
Concen: 18.226 ng/ul
RT: 8.328 min Scan# 891
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SSTD020670


Tgt Ion:108 Resp: 149042
Ion Ratio Lower Upper
108 100
107 86.9 70.8 106.2



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.101 ng/ul
RT: 8.416 min Scan# 906
Delta R.T. -0.000 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08

Tgt Ion: 45 Resp: 282159
Ion Ratio Lower Upper
45 100
77 12.3 17.1 25.7#
79 9.8 13.2 19.8#









#16
Acetophenone
Concen: 18.632 ng/ul
RT: 8.704 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SSTD020670


Tgt Ion:105 Resp: 252610
Ion Ratio Lower Upper
105 100
77 85.1 53.4 80.2#
51 36.3 13.8 20.8#



#17
N-Nitroso-di-n-propylamine
Concen: 19.737 ng/ul
RT: 8.693 min Scan# 953
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08

Tgt Ion: 70 Resp: 143209
Ion Ratio Lower Upper
70 100
42 68.2 31.8 47.8#
101 10.2 15.6 23.4#
130 20.6 13.7 20.5#









#18
4-Methylphenol
Concen: 17.816 ng/ul
RT: 8.652 min Scan# 947
Delta R.T. -0.012 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SSTD020670


Tgt Ion:108 Resp: 161017
Ion Ratio Lower Upper
108 100
107 118.7 83.6 125.4



#19
Hexachloroethane
Concen: 19.641 ng/ul
RT: 8.963 min Scan# 999
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08

Tgt Ion:117 Resp: 69286
Ion Ratio Lower Upper
117 100
201 108.6 106.5 159.7
199 67.5 59.0 88.6









#22
Nitrobenzene
Concen: 19.099 ng/ul
RT: 9.081 min Scan# 1109
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SSTD020670


Tgt Ion: 77 Resp: 209117
Ion Ratio Lower Upper
77 100
123 38.0 42.3 63.5#
65 14.6 5.8 8.8#



#23
Isophorone
Concen: 19.139 ng/ul
RT: 9.610 min Scan# 1109
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08

Tgt Ion: 82 Resp: 387332
Ion Ratio Lower Upper
82 100
95 7.6 7.8 11.8#
138 16.2 12.9 19.3









#25
2-Nitrophenol
Concen: 18.423 ng/ul
RT: 9.799 min Scan# 1141
Delta R.T. -0.000 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SSTD020670


Tgt Ion:139 Resp: 89316
Ion Ratio Lower Upper
139 100
65 63.3 49.4 74.2
109 35.8 44.7 67.1#



#26
2,4-Dimethylphenol
Concen: 19.014 ng/ul
RT: 9.857 min Scan# 1151
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08

Tgt Ion:107 Resp: 200022
Ion Ratio Lower Upper
107 100
121 45.6 34.1 51.1
122 77.2 46.3 69.5#









#27
Bis(2-Chloroethoxy)methane
Concen: 18.688 ng/ul
RT: 10.093 min Scan# 1191
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SSTD020670


Tgt Ion: 93 Resp: 225189
Ion Ratio Lower Upper
93 100
95 31.8 26.2 39.4
123 11.1 6.2 9.4#



#29
2,4-Dichlorophenol
Concen: 18.279 ng/ul
RT: 10.334 min Scan# 1232
Delta R.T. -0.000 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08

Tgt Ion:162 Resp: 156030
Ion Ratio Lower Upper
162 100
164 65.1 37.0 55.6#
98 38.0 31.0 46.6









#30
Naphthalene
Concen: 18.464 ng/ul
RT: 10.734 min Scan# 1300
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SSTD020670


Tgt Ion:128 Resp: 519169
Ion Ratio Lower Upper
128 100
129 10.8 8.6 13.0
127 13.0 12.2 18.4



#32
4-Chloroaniline
Concen: 17.767 ng/ul
RT: 10.846 min Scan# 1319
Delta R.T. -0.000 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08

Tgt Ion:127 Resp: 218880
Ion Ratio Lower Upper
127 100
129 32.2 24.6 37.0









#33
Hexachlorobutadiene
Concen: 18.732 ng/ul
RT: 11.022 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SSTD020670


Tgt Ion:225 Resp: 113302
Ion Ratio Lower Upper
225 100
223 65.8 53.4 80.0
227 64.0 53.8 80.8



#34
Caprolactam
Concen: 15.404 ng/ul
RT: 11.622 min Scan# 1451
Delta R.T. -0.000 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08

Tgt Ion:113 Resp: 45889
Ion Ratio Lower Upper
113 100
55 195.5 58.2 87.2#
56 157.9 68.6 103.0#









#35
4-Chloro-3-methylphenol
Concen: 18.791 ng/ul
RT: 11.969 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SSTD020670


Tgt Ion:107 Resp: 187635
Ion Ratio Lower Upper
107 100
144 25.1 15.8 23.8#
142 72.8 38.9 58.3#



#36
2-Methylnaphthalene
Concen: 18.655 ng/ul
RT: 12.340 min Scan# 1573
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08

Tgt Ion:142 Resp: 371350
Ion Ratio Lower Upper
142 100
141 88.7 64.3 96.5









#37
1-Methylnaphthalene
Concen: 18.667 ng/ul
RT: 12.557 min Scan# 1610
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SSTD020670


Tgt Ion:142 Resp: 375181
Ion Ratio Lower Upper
142 100
141 91.9 69.8 104.8
116 4.2 5.6 8.4#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 17.811 ng/ul
RT: 12.710 min Scan# 1636
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08

Tgt Ion:216 Resp: 209006
Ion Ratio Lower Upper
216 100
214 79.2 58.2 87.2
179 22.1 14.2 21.2#
108 18.1 15.3 22.9









#40
Hexachlorocyclopentadiene
Concen: 21.190 ng/ul
RT: 12.692 min Scan# 1633
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SSTD020670


Tgt Ion:237 Resp: 135186
Ion Ratio Lower Upper
237 100
235 64.6 53.2 79.8
272 12.3 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 18.391 ng/ul
RT: 12.951 min Scan# 1677
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08

Tgt Ion:196 Resp: 134522
Ion Ratio Lower Upper
196 100
198 98.0 64.1 96.1#
200 30.5 19.9 29.9#









#42
2,4,5-Trichlorophenol
Concen: 17.853 ng/ul
RT: 13.022 min Scan# 1689
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SSTD020670


Tgt Ion:196 Resp: 145813
Ion Ratio Lower Upper
196 100
198 96.2 68.6 103.0
200 30.4 32.8 49.2#



#43
1,1'-Biphenyl
Concen: 18.234 ng/ul
RT: 13.351 min Scan# 1745
Delta R.T. -0.000 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08

Tgt Ion:154 Resp: 514650
Ion Ratio Lower Upper
154 100
153 39.8 28.6 43.0
76 12.5 8.5 12.7









#44
2-Chloronaphthalene
Concen: 18.330 ng/ul
RT: 13.392 min Scan# 1752
Delta R.T. -0.000 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SSTD020670


Tgt Ion:162 Resp: 399595
Ion Ratio Lower Upper
162 100
164 33.0 27.5 41.3
127 41.6 37.1 55.7



#45
2-Nitroaniline
Concen: 19.988 ng/ul
RT: 13.592 min Scan# 1786
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08

Tgt Ion: 65 Resp: 143405
Ion Ratio Lower Upper
65 100
92 61.1 77.9 116.9#
138 84.7 81.5 122.3









#47
Dimethylphthalate
Concen: 18.818 ng/ul
RT: 13.975 min Scan# 1851
Delta R.T. -0.000 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SSTD020670


Tgt Ion:163 Resp: 528828
Ion Ratio Lower Upper
163 100
194 4.4 4.1 6.1
164 9.6 11.3 16.9#



#48
2,6-Dinitrotoluene
Concen: 19.134 ng/ul
RT: 14.092 min Scan# 1871
Delta R.T. -0.000 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08

Tgt Ion:165 Resp: 109612
Ion Ratio Lower Upper
165 100
89 63.9 51.7 77.5
121 22.6 18.3 27.5









#50
Acenaphthylene
Concen: 18.689 ng/ul
RT: 14.234 min Scan# 1895
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SSTD020670


Tgt Ion:152 Resp: 660715
Ion Ratio Lower Upper
152 100
151 19.2 18.3 27.5
153 13.1 11.0 16.6



#51
3-Nitroaniline
Concen: 18.265 ng/ul
RT: 14.422 min Scan# 1927
Delta R.T. -0.000 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08

Tgt Ion:138 Resp: 98206
Ion Ratio Lower Upper
138 100
108 11.1 15.8 23.8#
92 127.4 117.1 175.7









#52
Acenaphthene
Concen: 18.515 ng/ul
RT: 14.575 min Scan# 1953
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SSTD020670


Tgt Ion:153 Resp: 427003
Ion Ratio Lower Upper
153 100
152 46.2 34.3 51.5
154 88.6 70.0 105.0



#53
2,4-Dinitrophenol
Concen: 20.753 ng/ul
RT: 14.634 min Scan# 1963
Delta R.T. -0.000 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08

Tgt Ion:184 Resp: 66376
Ion Ratio Lower Upper
184 100
63 75.2 34.6 52.0#
154 67.9 57.7 86.5









#55
4-Nitrophenol
Concen: 18.365 ng/ul
RT: 14.728 min Scan# 1979
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SSTD020670


Tgt Ion:109 Resp: 82348
Ion Ratio Lower Upper
109 100
139 109.3 61.4 92.2#
65 136.0 65.3 97.9#



#56
Dibenzofuran
Concen: 18.596 ng/ul
RT: 14.916 min Scan# 2011
Delta R.T. -0.000 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08

Tgt Ion:168 Resp: 628490
Ion Ratio Lower Upper
168 100
139 39.3 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 19.968 ng/ul
RT: 14.881 min Scan# 2005
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SSTD020670


Tgt Ion:165 Resp: 164670
Ion Ratio Lower Upper
165 100
63 51.2 29.5 44.3#
182 2.9 2.8 4.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 18.805 ng/ul
RT: 15.145 min Scan# 2050
Delta R.T. -0.000 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08

Tgt Ion:232 Resp: 132067
Ion Ratio Lower Upper
232 100
131 42.9 25.4 38.2#
130 2.9 4.2 6.2#
166 28.6 28.1 42.1









#59
Diethylphthalate
Concen: 19.379 ng/ul
RT: 15.334 min Scan# 2082
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SSTD020670


Tgt Ion:149 Resp: 555719
Ion Ratio Lower Upper
149 100
177 21.1 22.0 33.0#
150 12.6 10.2 15.2



#61
Fluorene
Concen: 18.908 ng/ul
RT: 15.563 min Scan# 2121
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08

Tgt Ion:166 Resp: 522598
Ion Ratio Lower Upper
166 100
165 97.3 70.3 105.5
167 13.4 9.5 14.3









#62
4-Chlorophenyl-phenylether
Concen: 18.648 ng/ul
RT: 15.557 min Scan# 2120
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SSTD020670


Tgt Ion:204 Resp: 268111
Ion Ratio Lower Upper
204 100
206 32.5 27.6 41.4
141 60.1 46.6 69.8



#63
4-Nitroaniline
Concen: 18.248 ng/ul
RT: 15.581 min Scan# 2124
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08

Tgt Ion:138 Resp: 96878
Ion Ratio Lower Upper
138 100
92 59.8 41.6 62.4
108 95.1 97.6 146.4#









#66
 4,6-Dinitro-2-methylphenol
 Concen: 17.345 ng/ul
 RT: 15.645 min Scan# 2135
 Delta R.T. -0.006 min
 Lab File: BP010541.D
 Acq: 25 May 2022 18:08




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020670


Tgt Ion:198 Resp: 89645
 Ion Ratio Lower Upper
 198 100
 182 4.8 6.0 9.0#
 77 30.7 17.7 26.5#



#67
 N-Nitrosodiphenylamine
 Concen: 18.193 ng/ul
 RT: 15.769 min Scan# 2156
 Delta R.T. -0.006 min
 Lab File: BP010541.D
 Acq: 25 May 2022 18:08

Tgt Ion:169 Resp: 452110
 Ion Ratio Lower Upper
 169 100
 168 64.3 48.3 72.5
 167 34.4 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 18.053 ng/ul
RT: 16.451 min Scan# 211
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SSTD020670


Tgt Ion:248 Resp: 164751
Ion Ratio Lower Upper
248 100
250 97.4 79.4 119.0
141 79.8 67.2 100.8



#69
Hexachlorobenzene
Concen: 17.988 ng/ul
RT: 16.575 min Scan# 2293
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08

Tgt Ion:284 Resp: 188828
Ion Ratio Lower Upper
284 100
142 34.8 19.2 28.8#
249 33.2 27.4 41.2





#70

Atrazine

Concen: 18.202 ng/ul

RT: 16.728 min Scan# 2319

Delta R.T. -0.006 min

Lab File: BP010541.D

Acq: 25 May 2022 18:08

Instrument :

BNA_P

ClientSampleId :

SSTD020670





Tgt Ion:200 Resp: 181210

Ion Ratio Lower Upper

200 100

173 25.4 19.2 28.8

215 50.1 32.2 48.4#



#71

Pentachlorophenol

Concen: 16.671 ng/ul

RT: 16.922 min Scan# 2352

Delta R.T. -0.000 min

Lab File: BP010541.D

Acq: 25 May 2022 18:08




Tgt Ion:266 Resp: 109806


Ion Ratio Lower Upper

266 100

264 63.0 57.3 85.9

268 64.7 65.7 98.5#





#72

Phenanthrene

Concen: 18.458 ng/ul

RT: 17.310 min Scan# 2418

Delta R.T. -0.006 min

Lab File: BP010541.D

Acq: 25 May 2022 18:08

Instrument :

BNA_P

ClientSampleId :

SSTD020670





Tgt Ion:178 Resp: 872931

Ion Ratio Lower Upper

178 100

179 15.3 11.7 17.5

176 18.6 14.5 21.7



#74

Anthracene

Concen: 18.630 ng/ul

RT: 17.398 min Scan# 2433

Delta R.T. -0.006 min

Lab File: BP010541.D

Acq: 25 May 2022 18:08




Tgt Ion:178 Resp: 883859


Ion Ratio Lower Upper

178 100

179 15.6 12.3 18.5

176 17.8 13.4 20.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 16.575 ng/uL

RT: 13.316 min Scan# 1739

Delta R.T. -0.006 min

Lab File: BP010541.D

Acq: 25 May 2022 18:08

Instrument :

BNA_P

ClientSampleId :

SSTD020670





Tgt Ion:216 Resp: 223217

Ion Ratio Lower Upper

216 100

214 75.1 50.4 75.6

218 47.4 35.2 52.8



#76

Pentachlorobenzene

Concen: 17.788 ng/uL

RT: 14.839 min Scan# 1998

Delta R.T. -0.006 min

Lab File: BP010541.D

Acq: 25 May 2022 18:08




Tgt Ion:250 Resp: 217356


Ion Ratio Lower Upper

250 100

248 62.4 43.5 65.3

252 61.4 46.5 69.7









#77
Carbazole
Concen: 18.764 ng/ul
RT: 17.669 min Scan# 2480
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SSTD020670


Tgt Ion:167 Resp: 782348
Ion Ratio Lower Upper
167 100
166 21.0 15.8 23.8
139 13.2 8.3 12.5#



#78
Di-n-butylphthalate
Concen: 19.585 ng/ul
RT: 18.222 min Scan# 2573
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08

Tgt Ion:149 Resp: 994575
Ion Ratio Lower Upper
149 100
150 9.3 9.6 14.4#
104 6.0 6.2 9.4#









#80
Fluoranthene
Concen: 18.605 ng/u1
RT: 19.274 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SSTD020670


Tgt Ion:202 Resp: 1088023
Ion Ratio Lower Upper
202 100
101 10.2 9.6 14.4
100 7.9 6.3 9.5



#82
Pyrene
Concen: 18.693 ng/u1
RT: 19.627 min Scan# 2812
Delta R.T. -0.000 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08

Tgt Ion:202 Resp: 1140814
Ion Ratio Lower Upper
202 100
101 11.2 10.6 15.8
100 9.5 9.8 14.6#









#83
 Butylbenzylphthalate
 Concen: 19.647 ng/ul
 RT: 20.498 min Scan# 2960
 Delta R.T. -0.000 min
 Lab File: BP010541.D
 Acq: 25 May 2022 18:08




Instrument :
 BNA_P
 ClientSampleId :
 SST020670


Tgt Ion:149 Resp: 479456
 Ion Ratio Lower Upper
 149 100
 91 74.7 71.1 106.7
 206 22.2 28.8 43.2#



#84
 3,3'-Dichlorobenzidine
 Concen: 17.698 ng/ul
 RT: 21.269 min Scan# 3091
 Delta R.T. -0.000 min
 Lab File: BP010541.D
 Acq: 25 May 2022 18:08

Tgt Ion:252 Resp: 337057
 Ion Ratio Lower Upper
 252 100
 254 64.7 49.8 74.6
 126 9.9 4.3 6.5#





#85

Benzo(a)anthracene

Concen: 18.946 ng/ul

RT: 21.333 min Scan# 3102

Delta R.T. -0.006 min

Lab File: BP010541.D

Acq: 25 May 2022 18:08

Instrument :

BNA_P

ClientSampleId :

SSTD020670





Tgt Ion:228 Resp: 1122663

Ion Ratio Lower Upper

228 100

229 19.7 18.2 27.2

226 25.9 22.8 34.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 19.898 ng/ul

RT: 21.257 min Scan# 3089

Delta R.T. -0.006 min

Lab File: BP010541.D

Acq: 25 May 2022 18:08




Tgt Ion:149 Resp: 712554


Ion Ratio Lower Upper

149 100

167 27.9 32.8 49.2#

279 4.4 5.6 8.4#









#87
Chrysene
Concen: 18.892 ng/ul
RT: 21.386 min Scan# 3111
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SSTD020670


Tgt Ion:228 Resp: 1101125
Ion Ratio Lower Upper
228 100
226 29.3 22.7 34.1
229 19.9 15.2 22.8



#89
Di-n-octyl phthalate
Concen: 19.139 ng/ul
RT: 22.186 min Scan# 3247
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08

Tgt Ion:149 Resp: 1219880









#90
Benzo(b)fluoranthene
Concen: 19.269 ng/ul
RT: 23.033 min Scan# 3391
Delta R.T. -0.000 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SSTD020670


Tgt Ion:252 Resp: 1091225
Ion Ratio Lower Upper
252 100
253 21.9 17.4 26.0
125 8.7 7.3 10.9



#91
Benzo(k)fluoranthene
Concen: 18.725 ng/ul
RT: 23.080 min Scan# 3399
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08

Tgt Ion:252 Resp: 1007835
Ion Ratio Lower Upper
252 100
253 22.4 17.9 26.9
125 8.2 5.1 7.7#









#93
Benzo(a)pyrene
Concen: 18.817 ng/ul
RT: 23.662 min Scan# 3498
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08




Instrument :
BNA_P
ClientSampleId :
SSTD020670


Tgt Ion:252 Resp: 990091
Ion Ratio Lower Upper
252 100
253 21.9 17.5 26.3
125 10.1 6.4 9.6#



#94
Indeno(1,2,3-cd)pyrene
Concen: 19.159 ng/ul
RT: 26.286 min Scan# 3944
Delta R.T. -0.006 min
Lab File: BP010541.D
Acq: 25 May 2022 18:08

Tgt Ion:276 Resp: 948324
Ion Ratio Lower Upper
276 100
138 20.7 12.3 18.5#
277 24.4 19.3 28.9





#95

Dibenzo(a,h)anthracene

Concen: 18.590 ng/ul

RT: 26.304 min Scan# 3947

Delta R.T. -0.006 min

Lab File: BP010541.D

Acq: 25 May 2022 18:08

Instrument :

BNA_P

ClientSampleId :

SSTD020670





Tgt Ion:278 Resp: 798461

Ion Ratio Lower Upper

278 100

139 14.5 6.6 10.0#

279 24.4 18.4 27.6



#96

Benzo(g,h,i)perylene

Concen: 19.154 ng/ul

RT: 27.062 min Scan# 4076

Delta R.T. -0.006 min

Lab File: BP010541.D

Acq: 25 May 2022 18:08

Tgt Ion:276 Resp: 791672

Ion Ratio Lower Upper

276 100

138 18.9 11.6 17.4#

277 22.9 15.9 23.9

