

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052522\
 Data File : BP010553.D
 Acq On : 26 May 2022 00:54
 Operator : CG/JU
 Sample : N2883-06MSD
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 EW4K6MSD

Quant Time: May 26 02:15:03 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 24 05:50:40 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut
 Upadhyay

05/26/2022
 Supervised By :mohammad
 ahmed

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.881	152	147062	20.000	ng/ul	# 0.00
20) Naphthalene-d8	10.681	136	630716	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.516	164	409487	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.269	188	850968	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.351	240	815109	20.000	ng/ul	0.00
88) Perylene-d12	23.768	264	725523	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.334	96	12434	3.604	ng/uL	0.00
4) Pyridine-d5	3.758	84	94165	9.612	ng/ul	0.00
7) Phenol-d5	7.046	99	257657	20.874	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.210	67	186477	22.486	ng/ul	0.00
11) 2-Chlorophenol-d4	7.416	132	209633	22.195	ng/ul	0.00
15) 4-Methylphenol-d8	8.593	113	224791	21.429	ng/ul	0.00
21) Nitrobenzene-d5	9.040	128	114654	23.488	ng/ul	0.00
24) 2-Nitrophenol-d4	9.763	143	98231	18.989	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.304	165	219334	22.423	ng/ul	0.00
31) 4-Chloroaniline-d4	10.816	131	217793	14.975	ng/ul	-0.01
46) Dimethylphthalate-d6	13.922	166	756795	25.085	ng/ul	0.00
49) Acenaphthylene-d8	14.204	160	889972	25.574	ng/ul	0.00
54) 4-Nitrophenol-d4	14.716	143	43191	8.043	ng/ul	0.00
60) Fluorene-d10	15.510	176	675486	25.736	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.634	200	35803m	7.194	ng/ul	0.00
73) Anthracene-d10	17.363	188	1061224	27.541	ng/ul	0.00
81) Pyrene-d10	19.598	212	1257569	29.051	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.616	264	1008442	27.681	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.370	88	32142	8.993	ng/uL#	20
5) Pyridine	3.775	79	121228	11.907	ng/ul#	35
6) Benzaldehyde	7.022	77	210292m	32.228	ng/ul	
8) Phenol	7.075	94	371937m	27.931	ng/ul	
10) Bis(2-Chloroethyl)ether	7.305	93	255577	24.411	ng/ul#	74
12) 2-Chlorophenol	7.446	128	248216	25.174	ng/ul	97
13) 2-Methylphenol	8.328	108	239919	24.135	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.416	45	425697	24.948	ng/ul#	82
16) Acetophenone	8.705	105	430583	26.126	ng/ul#	78
17) N-Nitroso-di-n-propyla...	8.693	70	221345	25.094	ng/ul#	72
18) 4-Methylphenol	8.652	108	265509	24.167	ng/ul	95
19) Hexachloroethane	8.957	117	108659	25.338	ng/ul#	76
22) Nitrobenzene	9.081	77	320730	25.003	ng/ul#	83
23) Isophorone	9.610	82	613906	25.892	ng/ul#	97
25) 2-Nitrophenol	9.799	139	147584m	25.983	ng/ul	
26) 2,4-Dimethylphenol	9.857	107	254982	20.688	ng/ul#	79
27) Bis(2-Chloroethoxy)met...	10.093	93	358332	25.382	ng/ul#	97
29) 2,4-Dichlorophenol	10.328	162	251685	25.167	ng/ul#	86
30) Naphthalene	10.728	128	856274	25.992	ng/ul	97
32) 4-Chloroaniline	10.846	127	302140	20.933	ng/ul	95
33) Hexachlorobutadiene	11.022	225	183075	25.834	ng/ul	95
34) Caprolactam	11.622	113	93285m	26.728	ng/ul	
35) 4-Chloro-3-methylphenol	11.969	107	318993m	27.267	ng/ul	
36) 2-Methylnaphthalene	12.340	142	608444	26.089	ng/ul	91

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052522\
 Data File : BP010553.D
 Acq On : 26 May 2022 00:54
 Operator : CG/JU
 Sample : N2883-06MSD
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 EW4K6MSD

Manual Integrations
 APPROVED


Reviewed By :Jagrut
 Upadhyay

05/26/2022
 Supervised By :mohammad
 ahmed

Quant Time: May 26 02:15:03 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 24 05:50:40 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.557	142	617956	26.243	ng/ul#	95
39) 1,2,4,5-Tetrachloroben...	12.710	216	340420	27.242	ng/ul	93
40) Hexachlorocyclopentadiene	12.693	237	133685	19.678	ng/ul	95
41) 2,4,6-Trichlorophenol	12.951	196	218760	28.084	ng/ul#	86
42) 2,4,5-Trichlorophenol	13.022	196	239325	27.517	ng/ul#	88
43) 1,1'-Biphenyl	13.351	154	829746	27.606	ng/ul	94
44) 2-Chloronaphthalene	13.393	162	635564	27.378	ng/ul	94
45) 2-Nitroaniline	13.593	65	211998	27.748	ng/ul#	78
47) Dimethylphthalate	13.969	163	801739	26.790	ng/ul#	92
48) 2,6-Dinitrotoluene	14.092	165	171994	28.193	ng/ul	92
50) Acenaphthylene	14.234	152	1034424	27.477	ng/ul	95
51) 3-Nitroaniline	14.422	138	151140	26.397	ng/ul#	80
52) Acenaphthene	14.575	153	676146	27.532	ng/ul	96
53) 2,4-Dinitrophenol	14.634	184	76490	22.458	ng/ul#	77
55) 4-Nitrophenol	14.728	109	116916	24.486	ng/ul#	46
56) Dibenzofuran	14.916	168	981485	27.270	ng/ul	99
57) 2,4-Dinitrotoluene	14.881	165	246185	28.033	ng/ul#	80
58) 2,3,4,6-Tetrachlorophenol	15.145	232	207298	27.718	ng/ul#	86
59) Diethylphthalate	15.334	149	853605	27.954	ng/ul	93
61) Fluorene	15.563	166	821968	27.927	ng/ul	91
62) 4-Chlorophenyl-phenyle...	15.557	204	417103	27.243	ng/ul	97
63) 4-Nitroaniline	15.581	138	161630	28.590	ng/ul#	76
66) 4,6-Dinitro-2-methylph...	15.645	198	129948	26.182	ng/ul#	86
67) N-Nitrosodiphenylamine	15.769	169	725969	30.420	ng/ul	95
68) 4-Bromophenyl-phenylether	16.451	248	255795	29.188	ng/ul	99
69) Hexachlorobenzene	16.575	284	291737	28.940	ng/ul#	90
70) Atrazine	16.728	200	291798	30.522	ng/ul#	90
71) Pentachlorophenol	16.922	266	148576	23.489	ng/ul#	84
72) Phenanthrene	17.310	178	1321392	29.096	ng/ul	98
74) Anthracene	17.398	178	1316655	28.900	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.316	216	360153	27.849	ng/uL#	85
76) Pentachlorobenzene	14.840	250	348281	29.682	ng/uL	91
77) Carbazole	17.669	167	1170289	29.228	ng/ul#	95
78) Di-n-butylphthalate	18.222	149	1550597	31.797	ng/ul#	93
80) Fluoranthene	19.275	202	1575754	31.324	ng/ul	97
82) Pyrene	19.628	202	1617627	30.814	ng/ul	95
83) Butylbenzylphthalate	20.498	149	729977	34.774	ng/ul#	82
84) 3,3'-Dichlorobenzidine	21.263	252	464508	28.354	ng/ul#	95
85) Benzo(a)anthracene	21.333	228	1535007	30.115	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.257	149	1125111	36.525	ng/ul#	81
87) Chrysene	21.386	228	1489272	29.704	ng/ul	99
89) Di-n-octyl phthalate	22.186	149	1863810	34.244	ng/ul	100
90) Benzo(b)fluoranthene	23.033	252	1409130m	29.139	ng/ul	
91) Benzo(k)fluoranthene	23.080	252	1353682	29.453	ng/ul#	97
93) Benzo(a)pyrene	23.668	252	1300534	28.946	ng/ul#	98
94) Indeno(1,2,3-cd)pyrene	26.292	276	1285524	30.414	ng/ul#	92
95) Dibenzo(a,h)anthracene	26.304	278	1112291	30.327	ng/ul#	95
96) Benzo(g,h,i)perylene	27.057	276	827076	23.434	ng/ul#	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 8.993 ng/uL
RT: 3.370 min Scan# 48
Delta R.T. 0.000 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD






Tgt Ion: 88 Resp: 3214
Ion Ratio Lower Upper
88 100
43 39.0 6.8 10.2
58 81.3 28.4 42.6

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#5
Pyridine
Concen: 11.907 ng/uL
RT: 3.775 min Scan# 117
Delta R.T. 0.000 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54


Tgt Ion: 79 Resp: 121228
Ion Ratio Lower Upper
79 100
52 95.7 35.5 53.3#
51 45.8 20.2 30.4#





#6
Benzaldehyde
Concen: 32.228 ng/ul m
RT: 7.022 min Scan# 669
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD






Tgt Ion: 77 Resp: 21029
Ion Ratio Lower Upper
77 100
105 96.1 77.0 115.4
106 91.3 77.0 115.4

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#8
Phenol
Concen: 27.931 ng/ul m
RT: 7.075 min Scan# 678
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54


Tgt Ion: 94 Resp: 371937
Ion Ratio Lower Upper
94 100
65 32.8 19.0 28.6#
66 43.3 34.6 51.8





#10
Bis(2-Chloroethyl)ether
Concen: 24.411 ng/ul
RT: 7.305 min Scan# 717
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD






Tgt Ion: 93 Resp: 25557
Ion Ratio Lower Upper
93 100
63 85.8 45.0 67.4
95 32.8 26.3 39.5

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#12
2-Chlorophenol
Concen: 25.174 ng/ul
RT: 7.446 min Scan# 741
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Tgt Ion:128 Resp: 248216
Ion Ratio Lower Upper
128 100
64 56.6 45.3 67.9
130 32.0 29.0 43.6





#13
2-Methylphenol
Concen: 24.135 ng/ul
RT: 8.328 min Scan# 891
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :

BNA_P

ClientSampleId :

EW4K6MSD

Tgt Ion:108 Resp: 239919
Ion Ratio Lower Upper
108 100
107 87.4 70.8 106.2





Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay

05/26/2022




Supervised By :mohammad
ahmed


05/27/2022



#14
2,2'-oxybis(1-Chloropropane)
Concen: 24.948 ng/ul
RT: 8.416 min Scan# 906
Delta R.T. 0.000 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54


Tgt Ion: 45 Resp: 425697
Ion Ratio Lower Upper
45 100
77 12.5 17.1 25.7#
79 9.6 13.2 19.8#





#16
Acetophenone
Concen: 26.126 ng/ul
RT: 8.705 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD






Tgt Ion:105 Resp: 43058
Ion Ratio Lower Upper
105 100
77 81.8 53.4 80.2
51 32.1 13.8 20.8

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#17
N-Nitroso-di-n-propylamine
Concen: 25.094 ng/ul
RT: 8.693 min Scan# 953
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54


Tgt Ion: 70 Resp: 221345
Ion Ratio Lower Upper
70 100
42 64.6 31.8 47.8#
101 10.0 15.6 23.4#
130 21.1 13.7 20.5#





#18
4-Methylphenol
Concen: 24.167 ng/ul
RT: 8.652 min Scan# 946
Delta R.T. -0.012 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD






Tgt Ion:108 Resp: 265509
Ion Ratio Lower Upper
108 100
107 110.0 83.6 125.4

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#19
Hexachloroethane
Concen: 25.338 ng/ul
RT: 8.957 min Scan# 998
Delta R.T. -0.012 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54


Tgt Ion:117 Resp: 108659
Ion Ratio Lower Upper
117 100
201 96.3 106.5 159.7#
199 63.7 59.0 88.6





#22
Nitrobenzene
Concen: 25.003 ng/u1
RT: 9.081 min Scan# 1019
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD






Tgt Ion: 77 Resp: 320730
Ion Ratio Lower Upper
77 100
123 40.8 42.3 63.5
65 14.4 5.8 8.8

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#23
Isophorone
Concen: 25.892 ng/u1
RT: 9.610 min Scan# 1109
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54


Tgt Ion: 82 Resp: 613906
Ion Ratio Lower Upper
82 100
95 7.7 7.8 11.8
138 15.8 12.9 19.3





#25
2-Nitrophenol
Concen: 25.983 ng/ul m
RT: 9.799 min Scan# 1141
Delta R.T. 0.000 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD






Tgt Ion:139 Resp: 147584
Ion Ratio Lower Upper
139 100
65 58.1 49.4 74.2
109 32.1 44.7 67.1

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#26
2,4-Dimethylphenol
Concen: 20.688 ng/ul
RT: 9.857 min Scan# 1151
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54


Tgt Ion:107 Resp: 254982
Ion Ratio Lower Upper
107 100
121 48.0 34.1 51.1
122 80.5 46.3 69.5





#27
Bis(2-Chloroethoxy)methane
Concen: 25.382 ng/ul
RT: 10.093 min Scan# 1191
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD






Tgt Ion: 93 Resp: 35833
Ion Ratio Lower Upper
93 100
95 33.1 26.2 39.4
123 12.0 6.2 9.4

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#29
2,4-Dichlorophenol
Concen: 25.167 ng/ul
RT: 10.328 min Scan# 1231
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54


Tgt Ion:162 Resp: 251685
Ion Ratio Lower Upper
162 100
164 63.0 37.0 55.6#
98 38.9 31.0 46.6





#30
Naphthalene
Concen: 25.992 ng/ul
RT: 10.728 min Scan# 1299
Delta R.T. -0.012 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD






Tgt Ion:128 Resp: 856274
Ion Ratio Lower Upper
128 100
129 10.7 8.6 13.0
127 13.1 12.2 18.4

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#32
4-Chloroaniline
Concen: 20.933 ng/ul
RT: 10.846 min Scan# 1319
Delta R.T. 0.000 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54


Tgt Ion:127 Resp: 302140
Ion Ratio Lower Upper
127 100
129 33.4 24.6 37.0





#33
Hexachlorobutadiene
Concen: 25.834 ng/ul
RT: 11.022 min Scan# 1349
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD






Tgt Ion:225 Resp: 183075
Ion Ratio Lower Upper
225 100
223 61.1 53.4 80.0
227 64.4 53.8 80.8

Manual Integrations
APPROVED


Reviewed By :Jagrut
Upadhyay

05/26/2022
Supervised By :mohammad
ahmed



05/27/2022




#34
Caprolactam
Concen: 26.728 ng/ul m
RT: 11.622 min Scan# 1451
Delta R.T. 0.000 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54




Tgt Ion:113 Resp: 93285
Ion Ratio Lower Upper
113 100
55 195.5 58.2 87.2#
56 150.7 68.6 103.0#





#35
4-Chloro-3-methylphenol
Concen: 27.267 ng/ul m
RT: 11.969 min Scan# 1510
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD






Tgt Ion:107 Resp: 31899
Ion Ratio Lower Upper
107 100
144 25.1 15.8 23.8
142 73.5 38.9 58.3

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#36
2-Methylnaphthalene
Concen: 26.089 ng/ul
RT: 12.340 min Scan# 1573
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54


Tgt Ion:142 Resp: 608444
Ion Ratio Lower Upper
142 100
141 88.3 64.3 96.5





#37
1-Methylnaphthalene
Concen: 26.243 ng/ul
RT: 12.557 min Scan# 1610
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD






Tgt Ion:142 Resp: 617950
Ion Ratio Lower Upper
142 100
141 91.5 69.8 104.8
116 4.0 5.6 8.4

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#39
1,2,4,5-Tetrachlorobenzene
Concen: 27.242 ng/ul
RT: 12.710 min Scan# 1636
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54


Tgt Ion:216 Resp: 340420
Ion Ratio Lower Upper
216 100
214 78.8 58.2 87.2
179 20.8 14.2 21.2
108 15.8 15.3 22.9





#40
Hexachlorocyclopentadiene
Concen: 19.678 ng/ul
RT: 12.693 min Scan# 1633
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD






Tgt Ion:237 Resp: 133685
Ion Ratio Lower Upper
237 100
235 61.9 53.2 79.8
272 13.3 10.8 16.2

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#41
2,4,6-Trichlorophenol
Concen: 28.084 ng/ul
RT: 12.951 min Scan# 1677
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54


Tgt Ion:196 Resp: 218760
Ion Ratio Lower Upper
196 100
198 93.1 64.1 96.1
200 30.9 19.9 29.9#





#42
2,4,5-Trichlorophenol
Concen: 27.517 ng/ul
RT: 13.022 min Scan# 1689
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD






Tgt Ion:196 Resp: 239325
Ion Ratio Lower Upper
196 100
198 95.4 68.6 103.0
200 30.6 32.8 49.2

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#43
1,1'-Biphenyl
Concen: 27.606 ng/ul
RT: 13.351 min Scan# 1745
Delta R.T. 0.000 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54


Tgt Ion:154 Resp: 829746
Ion Ratio Lower Upper
154 100
153 39.3 28.6 43.0
76 12.3 8.5 12.7





#44
2-Chloronaphthalene
Concen: 27.378 ng/ul
RT: 13.393 min Scan# 1752
Delta R.T. 0.000 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD






Tgt Ion:162 Resp: 635564
Ion Ratio Lower Upper
162 100
164 33.2 27.5 41.3
127 40.4 37.1 55.7

Manual Integrations
APPROVED


Reviewed By :Jagrut
Upadhyay

05/26/2022
Supervised By :mohammad
ahmed



05/27/2022




#45
2-Nitroaniline
Concen: 27.748 ng/ul
RT: 13.593 min Scan# 1786
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54




Tgt Ion: 65 Resp: 211998
Ion Ratio Lower Upper
65 100
92 64.9 77.9 116.9#
138 90.6 81.5 122.3





#47
Dimethylphthalate
Concen: 26.790 ng/ul
RT: 13.969 min Scan# 1850
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD






Tgt Ion:163 Resp: 801739
Ion Ratio Lower Upper
163 100
194 4.0 4.1 6.1
164 10.3 11.3 16.9

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#48
2,6-Dinitrotoluene
Concen: 28.193 ng/ul
RT: 14.092 min Scan# 1871
Delta R.T. 0.000 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Tgt Ion:165 Resp: 171994
Ion Ratio Lower Upper
165 100
89 57.7 51.7 77.5
121 20.3 18.3 27.5





#50
Acenaphthylene
Concen: 27.477 ng/ul
RT: 14.234 min Scan# 1895
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :

BNA_P

ClientSampleId :

EW4K6MSD





Tgt Ion:152 Resp: 1034424
Ion Ratio Lower Upper
152 100
151 19.7 18.3 27.5
153 13.3 11.0 16.6

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#51
3-Nitroaniline
Concen: 26.397 ng/ul
RT: 14.422 min Scan# 1927
Delta R.T. 0.000 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54


Tgt Ion:138 Resp: 151140
Ion Ratio Lower Upper
138 100
108 12.5 15.8 23.8#
92 121.0 117.1 175.7





#52
Acenaphthene
Concen: 27.532 ng/ul
RT: 14.575 min Scan# 1953
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD






Tgt Ion:153 Resp: 676140
Ion Ratio Lower Upper
153 100
152 47.0 34.3 51.5
154 90.5 70.0 105.0

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#53
2,4-Dinitrophenol
Concen: 22.458 ng/ul
RT: 14.634 min Scan# 1963
Delta R.T. 0.000 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54


Tgt Ion:184 Resp: 76490
Ion Ratio Lower Upper
184 100
63 75.4 34.6 52.0#
154 67.1 57.7 86.5





#55
4-Nitrophenol
Concen: 24.486 ng/ul
RT: 14.728 min Scan# 1979
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD






Tgt Ion:109 Resp: 116910
Ion Ratio Lower Upper
109 100
139 117.7 61.4 92.2
65 135.9 65.3 97.9

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#56
Dibenzofuran
Concen: 27.270 ng/ul
RT: 14.916 min Scan# 2011
Delta R.T. 0.000 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54


Tgt Ion:168 Resp: 981485
Ion Ratio Lower Upper
168 100
139 38.3 30.1 45.1





#57
2,4-Dinitrotoluene
Concen: 28.033 ng/ul
RT: 14.881 min Scan# 2005
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD






Tgt Ion:165 Resp: 246185
Ion Ratio Lower Upper
165 100
63 50.1 29.5 44.3#
182 3.4 2.8 4.2

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#58
2,3,4,6-Tetrachlorophenol
Concen: 27.718 ng/ul
RT: 15.145 min Scan# 2050
Delta R.T. 0.000 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Tgt Ion:232 Resp: 207298
Ion Ratio Lower Upper
232 100
131 43.3 25.4 38.2#
130 2.7 4.2 6.2#
166 29.7 28.1 42.1





#59
Diethylphthalate
Concen: 27.954 ng/ul
RT: 15.334 min Scan# 2082
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :

BNA_P

ClientSampleId :

EW4K6MSD





Tgt Ion:149 Resp: 85360
Ion Ratio Lower Upper
149 100
177 22.3 22.0 33.0
150 12.9 10.2 15.2

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#61
Fluorene
Concen: 27.927 ng/ul
RT: 15.563 min Scan# 2121
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54


Tgt Ion:166 Resp: 821968
Ion Ratio Lower Upper
166 100
165 96.5 70.3 105.5
167 13.9 9.5 14.3





#62
4-Chlorophenyl-phenylether
Concen: 27.243 ng/ul
RT: 15.557 min Scan# 2120
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD






Tgt Ion:204 Resp: 41710
Ion Ratio Lower Upper
204 100
206 33.0 27.6 41.4
141 60.5 46.6 69.8

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#63
4-Nitroaniline
Concen: 28.590 ng/ul
RT: 15.581 min Scan# 2124
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54


Tgt Ion:138 Resp: 161630
Ion Ratio Lower Upper
138 100
92 55.6 41.6 62.4
108 86.1 97.6 146.4#





#66
4,6-Dinitro-2-methylphenol
Concen: 26.182 ng/ul
RT: 15.645 min Scan# 2135
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD






Tgt Ion:198 Resp: 12994
Ion Ratio Lower Upper
198 100
182 4.8 6.0 9.0
77 30.1 17.7 26.5

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#67
N-Nitrosodiphenylamine
Concen: 30.420 ng/ul
RT: 15.769 min Scan# 2156
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54


Tgt Ion:169 Resp: 725969
Ion Ratio Lower Upper
169 100
168 64.9 48.3 72.5
167 35.1 29.6 44.4





#68
4-Bromophenyl-phenylether
Concen: 29.188 ng/ul
RT: 16.451 min Scan# 211
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD






Tgt Ion:248 Resp: 25579
Ion Ratio Lower Upper
248 100
250 98.3 79.4 119.0
141 82.3 67.2 100.8

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#69
Hexachlorobenzene
Concen: 28.940 ng/ul
RT: 16.575 min Scan# 2293
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54


Tgt Ion:284 Resp: 291737
Ion Ratio Lower Upper
284 100
142 35.1 19.2 28.8#
249 33.2 27.4 41.2





#70
Atrazine
Concen: 30.522 ng/ul
RT: 16.728 min Scan# 2319
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD






Tgt Ion:200 Resp: 291798
Ion Ratio Lower Upper
200 100
173 27.0 19.2 28.8
215 48.4 32.2 48.4

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#71
Pentachlorophenol
Concen: 23.489 ng/ul
RT: 16.922 min Scan# 2352
Delta R.T. 0.000 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54


Tgt Ion:266 Resp: 148576
Ion Ratio Lower Upper
266 100
264 63.0 57.3 85.9
268 64.0 65.7 98.5





#72
Phenanthrene
Concen: 29.096 ng/u1
RT: 17.310 min Scan# 2418
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD






Tgt Ion:178 Resp: 1321392
Ion Ratio Lower Upper
178 100
179 15.7 11.7 17.5
176 18.5 14.5 21.7

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#74
Anthracene
Concen: 28.900 ng/u1
RT: 17.398 min Scan# 2433
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54


Tgt Ion:178 Resp: 1316655
Ion Ratio Lower Upper
178 100
179 15.3 12.3 18.5
176 17.9 13.4 20.2





#75
1,2,3,4-Tetrachlorobenzene
Concen: 27.849 ng/uL
RT: 13.316 min Scan# 1739
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD






Tgt Ion:216 Resp: 36015
Ion Ratio Lower Upper
216 100
214 79.0 50.4 75.6
218 49.1 35.2 52.8

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#76
Pentachlorobenzene
Concen: 29.682 ng/uL
RT: 14.840 min Scan# 1998
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Tgt Ion:250 Resp: 348281
Ion Ratio Lower Upper
250 100
248 62.0 43.5 65.3
252 63.4 46.5 69.7






#77
 Carbazole
 Concen: 29.228 ng/ul
 RT: 17.669 min Scan# 2479
 Delta R.T. -0.006 min
 Lab File: BP010553.D
 Acq: 26 May 2022 00:54

Instrument :

BNA_P

ClientSampleId :

EW4K6MSD



Tgt Ion:167 Resp: 1170289
 Ion Ratio Lower Upper
 167 100
 166 21.7 15.8 23.8
 139 13.3 8.3 12.5




Manual Integrations

APPROVED


Reviewed By :Jagrut
 Upadhyay

05/26/2022
 Supervised By :mohammad
 ahmed



05/27/2022




#78
 Di-n-butylphthalate
 Concen: 31.797 ng/ul
 RT: 18.222 min Scan# 2573
 Delta R.T. -0.006 min
 Lab File: BP010553.D
 Acq: 26 May 2022 00:54




Tgt Ion:149 Resp: 1550597
 Ion Ratio Lower Upper
 149 100
 150 9.1 9.6 14.4#
 104 5.6 6.2 9.4#





#80
Fluoranthene
Concen: 31.324 ng/u1
RT: 19.275 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD






Tgt Ion:202 Resp: 1575754
Ion Ratio Lower Upper
202 100
101 10.2 9.6 14.4
100 7.9 6.3 9.5

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#82
Pyrene
Concen: 30.814 ng/u1
RT: 19.628 min Scan# 2812
Delta R.T. 0.000 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54


Tgt Ion:202 Resp: 1617627
Ion Ratio Lower Upper
202 100
101 11.4 10.6 15.8
100 9.9 9.8 14.6





#83
Butylbenzylphthalate
Concen: 34.774 ng/ul
RT: 20.498 min Scan# 2960
Delta R.T. 0.000 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD






Tgt Ion:149 Resp: 72997
Ion Ratio Lower Upper
149 100
91 72.9 71.1 106.7
206 23.1 28.8 43.2

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#84
3,3'-Dichlorobenzidine
Concen: 28.354 ng/ul
RT: 21.263 min Scan# 3090
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Tgt Ion:252 Resp: 464508
Ion Ratio Lower Upper
252 100
254 64.9 49.8 74.6
126 12.6 4.3 6.5






#85
Benzo(a)anthracene
Concen: 30.115 ng/ul
RT: 21.333 min Scan# 3102
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :

BNA_P

ClientSampleId :

EW4K6MSD



Tgt Ion:228 Resp: 153500
Ion Ratio Lower Upper
228 100
229 19.8 18.2 27.2
226 26.6 22.8 34.2




Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay

05/26/2022




Supervised By :mohammad
ahmed


05/27/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 36.525 ng/ul
RT: 21.257 min Scan# 3089
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54


Tgt Ion:149 Resp: 1125111
Ion Ratio Lower Upper
149 100
167 27.9 32.8 49.2#
279 4.6 5.6 8.4#





#87
Chrysene
Concen: 29.704 ng/ul
RT: 21.386 min Scan# 3111
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD





Tgt Ion:228 Resp: 1489272
Ion Ratio Lower Upper
228 100
226 28.7 22.7 34.1
229 20.0 15.2 22.8

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#89
Di-n-octyl phthalate
Concen: 34.244 ng/ul
RT: 22.186 min Scan# 3247
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54


Tgt Ion:149 Resp: 1863810





#90
Benzo(b)fluoranthene
Concen: 29.139 ng/ul m
RT: 23.033 min Scan# 3391
Delta R.T. 0.000 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD






Tgt Ion:252 Resp: 1409130
Ion Ratio Lower Upper
252 100
253 23.0 17.4 26.0
125 11.0 7.3 10.9

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#91
Benzo(k)fluoranthene
Concen: 29.453 ng/ul
RT: 23.080 min Scan# 3399
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54


Tgt Ion:252 Resp: 1353682
Ion Ratio Lower Upper
252 100
253 22.4 17.9 26.9
125 10.4 5.1 7.7





#93
Benzo(a)pyrene
Concen: 28.946 ng/ul
RT: 23.668 min Scan# 3498
Delta R.T. 0.000 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :
BNA_P
ClientSampleId :
EW4K6MSD






Tgt Ion:252 Resp: 1300534
Ion Ratio Lower Upper
252 100
253 21.9 17.5 26.3
125 11.1 6.4 9.6

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/26/2022
Supervised By :mohammad
ahmed


05/27/2022



#94
Indeno(1,2,3-cd)pyrene
Concen: 30.414 ng/ul
RT: 26.292 min Scan# 3945
Delta R.T. 0.000 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Tgt Ion:276 Resp: 1285524
Ion Ratio Lower Upper
276 100
138 21.6 12.3 18.5#
277 25.6 19.3 28.9






#95
Dibenzo(a,h)anthracene
Concen: 30.327 ng/ul
RT: 26.304 min Scan# 3947
Delta R.T. -0.006 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54

Instrument :

BNA_P

ClientSampleId :

EW4K6MSD



Tgt Ion:278 Resp: 111229
Ion Ratio Lower Upper
278 100
139 14.2 6.6 10.0#
279 23.3 18.4 27.6#




Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay


05/26/2022

Supervised By :mohammad
ahmed

05/27/2022



#96
Benzo(g,h,i)perylene
Concen: 23.434 ng/ul
RT: 27.057 min Scan# 4075
Delta R.T. -0.012 min
Lab File: BP010553.D
Acq: 26 May 2022 00:54



Tgt Ion:276 Resp: 827076
Ion Ratio Lower Upper
276 100
138 18.9 11.6 17.4#
277 24.0 15.9 23.9#

