

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052523\
 Data File : BP015298.D
 Acq On : 25 May 2023 19:28
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: May 26 04:51:44 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052423.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 25 01:41:43 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	105	0.00
2	1,4-Dioxane	40.000	40.289	-0.7	100	0.00
3	Pyridine	40.000	43.705	-9.3	109	0.00
4	n-Nitrosodimethylamine	40.000	43.827	-9.6	112	0.00
5 S	2-Fluorophenol	80.000	81.954	-2.4	104	0.00
6	Aniline	40.000	41.943	-4.9	108	0.00
7 S	Phenol-d6	80.000	82.554	-3.2	107	0.00
8	2-Chlorophenol	40.000	41.068	-2.7	107	0.00
9	Benzaldehyde	40.000	40.587	-1.5	105	0.00
10 C	Phenol	40.000	41.062	-2.7	107	0.00
11	bis(2-Chloroethyl)ether	40.000	41.427	-3.6	108	0.00
12	1,3-Dichlorobenzene	40.000	39.830	0.4	104	0.00
13 C	1,4-Dichlorobenzene	40.000	40.271	-0.7	106	0.00
14	1,2-Dichlorobenzene	40.000	39.965	0.1	105	0.00
15	Benzyl Alcohol	40.000	43.170	-7.9	109	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	42.744	-6.9	112	0.00
17	2-Methylphenol	40.000	42.026	-5.1	108	0.00
18	Hexachloroethane	40.000	40.606	-1.5	107	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	43.139	-7.8	110	0.00
20	3+4-Methylphenols	40.000	43.027	-7.6	109	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	110	0.00
22	Acetophenone	40.000	40.509	-1.3	110	0.00
23 S	Nitrobenzene-d5	80.000	82.963	-3.7	111	0.00
24	Nitrobenzene	40.000	41.219	-3.0	111	0.00
25	Isophorone	40.000	41.015	-2.5	107	0.00
26 C	2-Nitrophenol	40.000	41.973	-4.9	110	0.00
27	2,4-Dimethylphenol	40.000	41.150	-2.9	108	0.00
28	bis(2-Chloroethoxy)methane	40.000	42.058	-5.1	111	0.00
29 C	2,4-Dichlorophenol	40.000	41.524	-3.8	109	0.00
30	1,2,4-Trichlorobenzene	40.000	40.218	-0.5	110	0.00
31	Naphthalene	40.000	40.418	-1.0	111	0.00
32	Benzoic acid	40.000	37.726	5.7	101	0.00
33	4-Chloroaniline	40.000	41.611	-4.0	107	0.00
34 C	Hexachlorobutadiene	40.000	40.016	-0.0	110	0.00
35	Caprolactam	40.000	36.669	8.3	90	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.958	-2.4	105	0.00
37	2-Methylnaphthalene	40.000	40.668	-1.7	109	0.00
38	1-Methylnaphthalene	40.000	40.933	-2.3	109	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	104	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.013	-5.0	109	0.00
41 P	Hexachlorocyclopentadiene	40.000	44.812	-12.0	108	0.00
42 S	2,4,6-Tribromophenol	80.000	76.149	4.8	92	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.878	-7.2	106	0.00
44	2,4,5-Trichlorophenol	40.000	42.604	-6.5	105	0.00
45 S	2-Fluorobiphenyl	80.000	84.833	-6.0	109	0.00
46	1,1'-Biphenyl	40.000	42.017	-5.0	108	0.00
47	2-Chloronaphthalene	40.000	42.013	-5.0	108	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052523\
 Data File : BP015298.D
 Acq On : 25 May 2023 19:28
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: May 26 04:51:44 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052423.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 25 01:41:43 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.559	-6.4	100	0.00
49	Acenaphthylene	40.000	41.438	-3.6	103	0.00
50	Dimethylphthalate	40.000	39.391	1.5	98	0.00
51	2,6-Dinitrotoluene	40.000	40.223	-0.6	97	0.00
52 C	Acenaphthene	40.000	41.219	-3.0	104	0.00
53	3-Nitroaniline	40.000	40.882	-2.2	94	0.00
54 P	2,4-Dinitrophenol	40.000	36.618	8.5	92	0.00
55	Dibenzofuran	40.000	40.722	-1.8	103	0.00
56 P	4-Nitrophenol	40.000	36.443	8.9	88	0.00
57	2,4-Dinitrotoluene	40.000	40.081	-0.2	93	0.00
58	Fluorene	40.000	40.070	-0.2	100	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.554	1.1	95	0.00
60	Diethylphthalate	40.000	39.174	2.1	96	0.00
61	4-Chlorophenyl-phenylether	40.000	40.334	-0.8	102	0.00
62	4-Nitroaniline	40.000	35.971	10.1	87	0.00
63	Azobenzene	40.000	41.220	-3.0	102	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	90	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.647	-9.1	89	0.00
66 c	n-Nitrosodiphenylamine	40.000	43.673	-9.2	97	0.00
67	4-Bromophenyl-phenylether	40.000	43.636	-9.1	97	0.00
68	Hexachlorobenzene	40.000	42.129	-5.3	94	0.00
69	Atrazine	40.000	41.284	-3.2	87	0.00
70 C	Pentachlorophenol	40.000	42.483	-6.2	89	0.00
71	Phenanthrene	40.000	41.390	-3.5	92	0.00
72	Anthracene	40.000	41.822	-4.6	92	0.00
73	Carbazole	40.000	40.846	-2.1	87	0.00
74	Di-n-butylphthalate	40.000	41.414	-3.5	88	0.00
75 C	Fluoranthene	40.000	39.131	2.2	84	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	80	0.00
77	Benzydine	40.000	44.352	-10.9	74	0.00
78	Pyrene	40.000	42.588	-6.5	84	0.00
79 S	Terphenyl-d14	80.000	84.453	-5.6	85	0.00
80	Butylbenzylphthalate	40.000	42.647	-6.6	80	0.00
81	Benzo(a)anthracene	40.000	41.506	-3.8	80	0.00
82	3,3'-Dichlorobenzidine	40.000	41.434	-3.6	77	0.00
83	Chrysene	40.000	41.333	-3.3	81	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.950	-7.4	80	0.00
85 c	Di-n-octyl phthalate	40.000	42.881	-7.2	79	0.00
86 I	Perylene-d12	20.000	20.000	0.0	78	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.350	-3.4	79	0.00
88	Benzo(b)fluoranthene	40.000	40.792	-2.0	80	0.00
89	Benzo(k)fluoranthene	40.000	38.683	3.3	77	0.00
90 C	Benzo(a)pyrene	40.000	40.075	-0.2	78	0.00
91	Dibenzo(a,h)anthracene	40.000	41.823	-4.6	79	0.00
92	Benzo(g,h,i)perylene	40.000	41.398	-3.5	78	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052523\
Data File : BP015298.D
Acq On : 25 May 2023 19:28
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: May 26 04:51:44 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052423.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 25 01:41:43 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0