

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060723\
 Data File : BP015389.D
 Acq On : 07 Jun 2023 09:35
 Operator : MA/JU
 Sample : SSTD00583
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD005625

Quant Time: Jun 07 23:10:15 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jun 07 13:01:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.105	152	125982	20.000	ng/u1	0.00
20) Naphthalene-d8	10.928	136	574487	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.745	164	420594	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.498	188	997573	20.000	ng/u1	0.00
79) Chrysene-d12	21.580	240	1037017	20.000	ng/u1	0.00
88) Perylene-d12	24.139	264	1171195	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.417	96	6135	1.827	ng/uL	0.00
4) Pyridine-d5	3.870	84	28396	3.213	ng/u1	0.02
7) Phenol-d5	7.258	99	46366	3.995	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.422	67	31656	4.567	ng/u1	0.00
11) 2-Chlorophenol-d4	7.628	132	36447	4.331	ng/u1	0.00
15) 4-Methylphenol-d8	8.816	113	38187	4.009	ng/u1	0.00
21) Nitrobenzene-d5	9.281	128	19451	4.313	ng/u1	0.00
24) 2-Nitrophenol-d4	10.010	143	21296	4.135	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.557	165	40023	4.260	ng/u1	0.00
31) 4-Chloroaniline-d4	11.075	131	49843	3.696	ng/u1	0.00
46) Dimethylphthalate-d6	14.146	166	154784	4.667	ng/u1	0.00
49) Acenaphthylene-d8	14.440	160	162418	4.479	ng/u1	0.00
54) 4-Nitrophenol-d4	14.957	143	14660	2.493	ng/u1	0.01
60) Fluorene-d10	15.740	176	133537	4.775	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.863	200	21384	3.384	ng/u1	0.00
73) Anthracene-d10	17.598	188	213575	4.706	ng/u1	0.00
81) Pyrene-d10	19.822	212	254569	4.587	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.969	264	269894	4.593	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.452	88	6573	1.859	ng/uL	100
5) Pyridine	3.887	79	33065	3.609	ng/u1#	84
6) Benzaldehyde	7.240	77	12630	2.909	ng/u1	99
8) Phenol	7.281	94	49891	4.167	ng/u1	95
10) Bis(2-Chloroethyl)ether	7.516	93	42217	4.459	ng/u1	98
12) 2-Chlorophenol	7.663	128	38890	4.374	ng/u1	97
13) 2-Methylphenol	8.552	108	37579	4.079	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.628	45	58552	4.667	ng/u1	98
16) Acetophenone	8.934	105	68910	4.533	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.910	70	37935	4.624	ng/u1	96
18) 4-Methylphenol	8.881	108	42744	4.217	ng/u1	99
19) Hexachloroethane	9.193	117	17079	4.546	ng/u1	95
22) Nitrobenzene	9.322	77	51823	4.333	ng/u1	99
23) Isophorone	9.846	82	106928	4.466	ng/u1	99
25) 2-Nitrophenol	10.040	139	22994	4.135	ng/u1	95
26) 2,4-Dimethylphenol	10.093	107	52175	4.424	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.328	93	63596	4.568	ng/u1	98
29) 2,4-Dichlorophenol	10.581	162	38188	4.088	ng/u1	98
30) Naphthalene	10.981	128	144853	4.653	ng/u1	98
32) 4-Chloroaniline	11.099	127	50238	3.600	ng/u1	98
33) Hexachlorobutadiene	11.257	225	29458	4.676	ng/u1	98
34) Caprolactam	11.863	113	13386	3.822	ng/u1	86
35) 4-Chloro-3-methylphenol	12.210	107	48796	4.311	ng/u1	97
36) 2-Methylnaphthalene	12.581	142	103898	4.747	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060723\
 Data File : BP015389.D
 Acq On : 07 Jun 2023 09:35
 Operator : MA/JU
 Sample : SST00583
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SST005625


Quant Time: Jun 07 23:10:15 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jun 07 13:01:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.799	142	104302	4.734	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.946	216	57605	4.398	ng/ul	98
40) Hexachlorocyclopentadiene	12.922	237	6420	1.196	ng/ul	94
41) 2,4,6-Trichlorophenol	13.187	196	34263	3.931	ng/ul	97
42) 2,4,5-Trichlorophenol	13.263	196	39110	4.117	ng/ul	99
43) 1,1'-Biphenyl	13.581	154	149117	4.585	ng/ul	99
44) 2-Chloronaphthalene	13.628	162	115795	4.538	ng/ul	99
45) 2-Nitroaniline	13.834	65	30477	3.732	ng/ul	97
47) Dimethylphthalate	14.193	163	161977	4.730	ng/ul	99
48) 2,6-Dinitrotoluene	14.322	165	28347	4.040	ng/ul	98
50) Acenaphthylene	14.469	152	184009	4.582	ng/ul	99
51) 3-Nitroaniline	14.663	138	18949	3.272	ng/ul#	98
52) Acenaphthene	14.810	153	129636	4.670	ng/ul	98
53) 2,4-Dinitrophenol	14.887	184	8063	1.911	ng/ul	92
55) 4-Nitrophenol	14.969	109	15281	2.667	ng/ul	88
56) Dibenzofuran	15.145	168	186660	4.744	ng/ul	99
57) 2,4-Dinitrotoluene	15.110	165	42054	4.075	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.375	232	36362	4.240	ng/ul	95
59) Diethylphthalate	15.551	149	164986	4.701	ng/ul	99
61) Fluorene	15.792	166	156248	4.871	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.781	204	79949	4.946	ng/ul	99
63) 4-Nitroaniline	15.822	138	16264	2.982	ng/ul	92
66) 4,6-Dinitro-2-methylph...	15.881	198	21733	3.372	ng/ul#	96
67) N-Nitrosodiphenylamine	15.998	169	132728	4.690	ng/ul	99
68) 4-Bromophenyl-phenylether	16.681	248	48639	4.557	ng/ul	96
69) Hexachlorobenzene	16.798	284	58959	4.683	ng/ul	93
70) Atrazine	16.945	200	49303	4.363	ng/ul	97
71) Pentachlorophenol	17.169	266	17300	2.446	ng/ul	89
72) Phenanthrene	17.545	178	252017	4.708	ng/ul	99
74) Anthracene	17.634	178	260022	4.787	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.551	216	60065	4.363	ng/uL	98
76) Pentachlorobenzene	15.063	250	66733	4.629	ng/uL	99
77) Carbazole	17.904	167	229821	4.699	ng/ul	99
78) Di-n-butylphthalate	18.434	149	284216	4.580	ng/ul	99
80) Fluoranthene	19.498	202	310942	4.599	ng/ul	100
82) Pyrene	19.851	202	331226	4.735	ng/ul	100
83) Butylbenzylphthalate	20.704	149	133309	4.312	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.492	252	95254	3.888	ng/ul	99
85) Benzo(a)anthracene	21.563	228	337255	4.652	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.463	149	198547	4.345	ng/ul	99
87) Chrysene	21.622	228	321706	4.695	ng/ul	99
89) Di-n-octyl phthalate	22.433	149	344935	4.481	ng/ul	100
90) Benzo(b)fluoranthene	23.345	252	360745	4.772	ng/ul	98
91) Benzo(k)fluoranthene	23.398	252	343768	4.704	ng/ul	99
93) Benzo(a)pyrene	24.021	252	306744	4.654	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.845	276	378596	4.438	ng/ul	99
95) Dibenzo(a,h)anthracene	26.862	278	306997	4.417	ng/ul	97
96) Benzo(g,h,i)perylene	27.674	276	314341	4.472	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 07 23:10:15 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jun 07 13:01:04 2023
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 1.859 ng/uL
 RT: 3.452 min Scan# 41
 Delta R.T. 0.000 min
 Lab File: BP015389.D
 Acq: 07 Jun 2023 09:35




Instrument :
 BNA_P
 ClientSampleId :
 SSTD005625


Tgt Ion: 88 Resp: 6573
 Ion Ratio Lower Upper
 88 100
 43 30.0 23.9 35.9
 58 64.3 51.4 77.2



#5
 Pyridine
 Concen: 3.609 ng/uL
 RT: 3.887 min Scan# 119
 Delta R.T. 0.012 min
 Lab File: BP015389.D
 Acq: 07 Jun 2023 09:35

Tgt Ion: 79 Resp: 33065
 Ion Ratio Lower Upper
 79 100
 52 89.2 59.5 89.3
 51 45.6 29.8 44.6#









#6
Benzaldehyde
Concen: 2.909 ng/ul
RT: 7.240 min Scan# 688
Delta R.T. 0.006 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35




Instrument :
BNA_P
ClientSampleId :
SSTD005625


Tgt Ion: 77	Resp: 12630
Ion Ratio	Lower Upper
77 100	
105 95.8	77.3 115.9
106 93.1	73.3 109.9



#8
Phenol
Concen: 4.167 ng/ul
RT: 7.281 min Scan# 696
Delta R.T. -0.000 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35

Tgt Ion: 94	Resp: 49891
Ion Ratio	Lower Upper
94 100	
65 31.1	23.5 35.3
66 38.5	34.2 51.4









#10
 Bis(2-Chloroethyl)ether
 Concen: 4.459 ng/ul
 RT: 7.516 min Scan# 710
 Delta R.T. -0.000 min
 Lab File: BP015389.D
 Acq: 07 Jun 2023 09:35

Instrument :
 BNA_P
 ClientSampleId :
 SSTD005625

Tgt Ion: 93 Resp: 42217




Ion	Ratio	Lower	Upper
93	100		
63	76.8	62.6	93.8
95	33.1	25.5	38.3




#12
 2-Chlorophenol
 Concen: 4.374 ng/ul
 RT: 7.663 min Scan# 761
 Delta R.T. 0.006 min
 Lab File: BP015389.D
 Acq: 07 Jun 2023 09:35

Tgt Ion: 128 Resp: 38890

Ion	Ratio	Lower	Upper
128	100		
64	52.4	40.2	60.4
130	32.3	24.5	36.7





#13

2-Methylphenol

Concen: 4.079 ng/ul

RT: 8.552 min Scan# 911

Delta R.T. 0.006 min

Lab File: BP015389.D

Acq: 07 Jun 2023 09:35

Instrument :

BNA_P

ClientSampleId :





SSTD005625

Tgt Ion:108 Resp: 37579

Ion Ratio Lower Upper

108 100

107 88.0 71.2 106.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 4.667 ng/ul

RT: 8.628 min Scan# 925

Delta R.T. 0.000 min

Lab File: BP015389.D

Acq: 07 Jun 2023 09:35




Tgt Ion: 45 Resp: 58552


Ion Ratio Lower Upper

45 100

77 14.1 11.8 17.8

79 12.2 8.7 13.1









#16
Acetophenone
Concen: 4.533 ng/ul
RT: 8.934 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35




Instrument :
BNA_P
ClientSampleId :
SSTD005625


Tgt Ion:	105	Resp:	68910
Ion Ratio	Lower	Upper	
105	100		
77	84.1	64.9	97.3
51	32.1	23.5	35.3



#17
N-Nitroso-di-n-propylamine
Concen: 4.624 ng/ul
RT: 8.910 min Scan# 973
Delta R.T. -0.006 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35

Tgt Ion:	70	Resp:	37935
Ion Ratio	Lower	Upper	
70	100		
42	61.1	46.6	69.8
101	9.0	7.8	11.6
130	18.1	15.9	23.9









#18
4-Methylphenol
Concen: 4.217 ng/ul
RT: 8.881 min Scan# 90
Delta R.T. -0.000 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35




Instrument :
BNA_P
ClientSampleId :
SSTD005625


Tgt Ion:108 Resp: 42744
Ion Ratio Lower Upper
108 100
107 112.3 89.3 133.9



#19
Hexachloroethane
Concen: 4.546 ng/ul
RT: 9.193 min Scan# 1021
Delta R.T. -0.000 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35

Tgt Ion:117 Resp: 17079
Ion Ratio Lower Upper
117 100
201 110.1 82.8 124.2
199 65.4 50.6 76.0









#22
Nitrobenzene
Concen: 4.333 ng/ul
RT: 9.322 min Scan# 1043
Delta R.T. 0.000 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35




Instrument :
BNA_P
ClientSampleId :
SSTD005625


Tgt Ion: 77	Resp: 51823
Ion Ratio	Lower Upper
77 100	
123 42.4	33.5 50.3
65 13.1	11.6 17.4



#23
Isophorone
Concen: 4.466 ng/ul
RT: 9.846 min Scan# 1132
Delta R.T. -0.006 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35

Tgt Ion: 82	Resp: 106928
Ion Ratio	Lower Upper
82 100	
95 7.6	5.8 8.8
138 16.1	13.1 19.7









#25
2-Nitrophenol
Concen: 4.135 ng/ul
RT: 10.040 min Scan# 1165
Delta R.T. -0.000 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35




Instrument :
BNA_P
ClientSampleId :
SSTD005625


Tgt Ion:	139	Resp:	22994
Ion	Ratio	Lower	Upper
139	100		
65	63.2	46.9	70.3
109	39.1	30.2	45.2



#26
2,4-Dimethylphenol
Concen: 4.424 ng/ul
RT: 10.093 min Scan# 1174
Delta R.T. -0.000 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35

Tgt Ion:	107	Resp:	52175
Ion	Ratio	Lower	Upper
107	100		
121	49.0	39.0	58.6
122	77.2	64.8	97.2





#27

Bis(2-Chloroethoxy)methane

Concen: 4.568 ng/ul

RT: 10.328 min Scan# 1214

Delta R.T. -0.006 min

Lab File: BP015389.D

Acq: 07 Jun 2023 09:35

Instrument :

BNA_P

ClientSampleId :

SSTD005625





Tgt Ion: 93 Resp: 63596

Ion Ratio Lower Upper

93 100

95 32.8 25.7 38.5

123 9.4 8.6 13.0



#29

2,4-Dichlorophenol

Concen: 4.088 ng/ul

RT: 10.581 min Scan# 1257

Delta R.T. 0.006 min

Lab File: BP015389.D

Acq: 07 Jun 2023 09:35




Tgt Ion:162 Resp: 38188


Ion Ratio Lower Upper

162 100

164 65.6 50.6 76.0

98 38.5 30.2 45.2





#30

Naphthalene

Concen: 4.653 ng/ul

RT: 10.981 min Scan# 11

Delta R.T. 0.000 min

Lab File: BP015389.D

Acq: 07 Jun 2023 09:35

Instrument :

BNA_P

ClientSampleId :

SSTD005625





Tgt Ion:128 Resp: 144853

Ion Ratio Lower Upper

128 100

129 12.2 8.6 13.0

127 13.1 10.5 15.7



#32

4-Chloroaniline

Concen: 3.600 ng/ul

RT: 11.099 min Scan# 1345

Delta R.T. 0.006 min

Lab File: BP015389.D




Acq: 07 Jun 2023 09:35


Tgt Ion:127 Resp: 50238

Ion Ratio Lower Upper

127 100

129 33.9 26.1 39.1









#33
Hexachlorobutadiene
Concen: 4.676 ng/ul
RT: 11.257 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35




Instrument :
BNA_P
ClientSampleId :
SSTD005625


Tgt Ion:225 Resp: 29458
Ion Ratio Lower Upper
225 100
223 62.4 52.0 78.0
227 64.3 50.6 75.8



#34
Caprolactam
Concen: 3.822 ng/ul
RT: 11.863 min Scan# 1475
Delta R.T. -0.006 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35

Tgt Ion:113 Resp: 13386
Ion Ratio Lower Upper
113 100
55 177.7 127.9 191.9
56 143.6 101.1 151.7









#35
4-Chloro-3-methylphenol
Concen: 4.311 ng/ul
RT: 12.210 min Scan# 1534
Delta R.T. 0.000 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35




Instrument :
BNA_P
ClientSampleId :
SSTD005625


Tgt Ion:107 Resp: 48796
Ion Ratio Lower Upper
107 100
144 25.7 19.4 29.0
142 77.8 60.4 90.6



#36
2-Methylnaphthalene
Concen: 4.747 ng/ul
RT: 12.581 min Scan# 1597
Delta R.T. 0.000 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35

Tgt Ion:142 Resp: 103898
Ion Ratio Lower Upper
142 100
141 89.9 72.0 108.0









#37
1-Methylnaphthalene
Concen: 4.734 ng/ul
RT: 12.799 min Scan# 1634
Delta R.T. 0.000 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35




Instrument :
BNA_P
ClientSampleId :
SSTD005625


Tgt Ion:142 Resp: 104302
Ion Ratio Lower Upper
142 100
141 91.5 71.9 107.9
116 4.5 3.4 5.2



#39
1,2,4,5-Tetrachlorobenzene
Concen: 4.398 ng/ul
RT: 12.946 min Scan# 1659
Delta R.T. 0.000 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35

Tgt Ion:216 Resp: 57605
Ion Ratio Lower Upper
216 100
214 79.5 62.2 93.2
179 20.9 17.4 26.2
108 17.8 14.2 21.4





#40

Hexachlorocyclopentadiene

Concen: 1.196 ng/ul

RT: 12.922 min Scan# 1655

Delta R.T. -0.000 min

Lab File: BP015389.D

Acq: 07 Jun 2023 09:35

Instrument :

BNA_P

ClientSampleId :

SSTD005625





Tgt Ion:237 Resp: 6420

Ion Ratio Lower Upper

237 100

235 64.9 48.2 72.2

272 11.4 10.3 15.5



#41

2,4,6-Trichlorophenol

Concen: 3.931 ng/ul

RT: 13.187 min Scan# 1700

Delta R.T. 0.000 min

Lab File: BP015389.D

Acq: 07 Jun 2023 09:35




Tgt Ion:196 Resp: 34263


Ion Ratio Lower Upper

196 100

198 100.7 77.8 116.6

200 34.1 25.8 38.6









#42
2,4,5-Trichlorophenol
Concen: 4.117 ng/ul
RT: 13.263 min Scan# 1712
Delta R.T. 0.006 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35




Instrument :
BNA_P
ClientSampleId :
SSTD005625


Tgt Ion:196 Resp: 39110
Ion Ratio Lower Upper
196 100
198 95.9 75.5 113.3
200 30.7 24.7 37.1



#43
1,1'-Biphenyl
Concen: 4.585 ng/ul
RT: 13.581 min Scan# 1767
Delta R.T. -0.000 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35

Tgt Ion:154 Resp: 149117
Ion Ratio Lower Upper
154 100
153 39.3 32.2 48.2
76 13.5 10.4 15.6









#44
2-Chloronaphthalene
Concen: 4.538 ng/ul
RT: 13.628 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35




Instrument :
BNA_P
ClientSampleId :
SSTD005625


Tgt Ion:162 Resp: 115795
Ion Ratio Lower Upper
162 100
164 33.7 27.0 40.6
127 39.7 32.2 48.4



#45
2-Nitroaniline
Concen: 3.732 ng/ul
RT: 13.834 min Scan# 1810
Delta R.T. -0.000 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35

Tgt Ion: 65 Resp: 30477
Ion Ratio Lower Upper
65 100
92 65.8 55.4 83.2
138 100.3 78.8 118.2









#47
Dimethylphthalate
Concen: 4.730 ng/ul
RT: 14.193 min Scan# 1871
Delta R.T. -0.006 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35




Instrument :
BNA_P
ClientSampleId :
SSTD005625


Tgt Ion:163 Resp: 161977
Ion Ratio Lower Upper
163 100
194 3.6 3.2 4.8
164 10.0 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 4.040 ng/ul
RT: 14.322 min Scan# 1893
Delta R.T. 0.000 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35

Tgt Ion:165 Resp: 28347
Ion Ratio Lower Upper
165 100
89 63.8 51.5 77.3
121 23.4 17.3 25.9









#50
Acenaphthylene
Concen: 4.582 ng/ul
RT: 14.469 min Scan# 1918
Delta R.T. 0.000 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35




Instrument :
BNA_P
ClientSampleId :
SSTD005625


Tgt Ion:152 Resp: 184009
Ion Ratio Lower Upper
152 100
151 19.5 16.0 24.0
153 13.0 10.6 15.8



#51
3-Nitroaniline
Concen: 3.272 ng/ul
RT: 14.663 min Scan# 1951
Delta R.T. 0.006 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35

Tgt Ion:138 Resp: 18949
Ion Ratio Lower Upper
138 100
108 13.3 8.4 12.6#
92 121.2 98.0 147.0









#52
Acenaphthene
Concen: 4.670 ng/ul
RT: 14.810 min Scan# 1976
Delta R.T. -0.000 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35




Instrument :
BNA_P
ClientSampleId :
SSTD005625


Tgt Ion:153 Resp: 129636
Ion Ratio Lower Upper
153 100
152 48.6 38.3 57.5
154 90.1 69.8 104.6



#53
2,4-Dinitrophenol
Concen: 1.911 ng/ul
RT: 14.887 min Scan# 1989
Delta R.T. 0.018 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35

Tgt Ion:184 Resp: 8063
Ion Ratio Lower Upper
184 100
63 70.6 59.0 88.4
154 62.1 57.6 86.4









#55
4-Nitrophenol
Concen: 2.667 ng/ul
RT: 14.969 min Scan# 2001
Delta R.T. 0.012 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35




Instrument :
BNA_P
ClientSampleId :
SSTD005625


Tgt Ion:109 Resp: 15281
Ion Ratio Lower Upper
109 100
139 121.4 89.1 133.7
65 146.3 103.1 154.7



#56
Dibenzofuran
Concen: 4.744 ng/ul
RT: 15.145 min Scan# 2033
Delta R.T. 0.000 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35

Tgt Ion:168 Resp: 186660
Ion Ratio Lower Upper
168 100
139 38.2 31.2 46.8









#57
2,4-Dinitrotoluene
Concen: 4.075 ng/ul
RT: 15.110 min Scan# 2027
Delta R.T. -0.000 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35




Instrument :
BNA_P
ClientSampleId :
SSTD005625


Tgt Ion:165 Resp: 42054
Ion Ratio Lower Upper
165 100
63 52.2 39.1 58.7
182 2.7 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 4.240 ng/ul
RT: 15.375 min Scan# 2072
Delta R.T. 0.006 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35

Tgt Ion:232 Resp: 36362
Ion Ratio Lower Upper
232 100
131 44.5 37.8 56.8
130 2.6 1.9 2.9
166 29.4 26.2 39.2









#59
Diethylphthalate
Concen: 4.701 ng/ul
RT: 15.551 min Scan# 2102
Delta R.T. -0.006 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35




Instrument :
BNA_P
ClientSampleId :
SSTD005625


Tgt Ion:149 Resp: 164986
Ion Ratio Lower Upper
149 100
177 21.9 17.8 26.8
150 12.4 10.1 15.1



#61
Fluorene
Concen: 4.871 ng/ul
RT: 15.792 min Scan# 2143
Delta R.T. -0.000 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35

Tgt Ion:166 Resp: 156248
Ion Ratio Lower Upper
166 100
165 100.2 79.1 118.7
167 13.3 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 4.946 ng/ul
RT: 15.781 min Scan# 2141
Delta R.T. -0.000 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35




Instrument :
BNA_P
ClientSampleId :
SSTD005625


Tgt Ion:204 Resp: 79949
Ion Ratio Lower Upper
204 100
206 32.8 26.6 40.0
141 61.4 49.8 74.8



#63
4-Nitroaniline
Concen: 2.982 ng/ul
RT: 15.822 min Scan# 2148
Delta R.T. 0.006 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35

Tgt Ion:138 Resp: 16264
Ion Ratio Lower Upper
138 100
92 62.2 44.6 67.0
108 105.8 78.4 117.6









#66
4,6-Dinitro-2-methylphenol
Concen: 3.372 ng/ul
RT: 15.881 min Scan# 2158
Delta R.T. -0.000 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35




Instrument :
BNA_P
ClientSampleId :
SSTD005625


Tgt Ion:198 Resp: 21733
Ion Ratio Lower Upper
198 100
182 6.0 3.0 4.4#
77 27.4 20.5 30.7



#67
N-Nitrosodiphenylamine
Concen: 4.690 ng/ul
RT: 15.998 min Scan# 2178
Delta R.T. 0.000 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35

Tgt Ion:169 Resp: 132728
Ion Ratio Lower Upper
169 100
168 67.6 53.0 79.4
167 36.4 28.9 43.3









#68
4-Bromophenyl-phenylether
Concen: 4.557 ng/ul
RT: 16.681 min Scan# 2294
Delta R.T. 0.000 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35




Instrument :
BNA_P
ClientSampleId :
SSTD005625


Tgt Ion:248 Resp: 48639
Ion Ratio Lower Upper
248 100
250 101.0 78.3 117.5
141 80.5 61.1 91.7



#69
Hexachlorobenzene
Concen: 4.683 ng/ul
RT: 16.798 min Scan# 2314
Delta R.T. -0.006 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35

Tgt Ion:284 Resp: 58959
Ion Ratio Lower Upper
284 100
142 36.0 25.4 38.2
249 34.5 24.8 37.2





#70

Atrazine

Concen: 4.363 ng/ul

RT: 16.945 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP015389.D

Acq: 07 Jun 2023 09:35

Instrument :

BNA_P

ClientSampleId :

SSTD005625





Tgt Ion:200 Resp: 49303

Ion Ratio Lower Upper

200 100

173 27.6 19.8 29.6

215 48.0 37.4 56.0



#71

Pentachlorophenol

Concen: 2.446 ng/ul

RT: 17.169 min Scan# 2377

Delta R.T. 0.018 min

Lab File: BP015389.D

Acq: 07 Jun 2023 09:35




Tgt Ion:266 Resp: 17300


Ion Ratio Lower Upper

266 100

264 54.8 49.6 74.4

268 54.2 51.0 76.6





#72

Phenanthrene

Concen: 4.708 ng/ul

RT: 17.545 min Scan# 2441

Delta R.T. -0.000 min

Lab File: BP015389.D

Acq: 07 Jun 2023 09:35

Instrument :

BNA_P

ClientSampleId :

SSTD005625





Tgt Ion:178 Resp: 252017

Ion Ratio Lower Upper

178 100

179 15.6 12.3 18.5

176 19.0 15.4 23.0



#74

Anthracene

Concen: 4.787 ng/ul

RT: 17.634 min Scan# 2456

Delta R.T. -0.000 min

Lab File: BP015389.D

Acq: 07 Jun 2023 09:35




Tgt Ion:178 Resp: 260022


Ion Ratio Lower Upper

178 100

179 15.2 12.4 18.6

176 18.2 15.0 22.4





#75

1,2,3,4-Tetrachlorobenzene

Concen: 4.363 ng/uL

RT: 13.551 min Scan# 1762

Delta R.T. 0.000 min

Lab File: BP015389.D

Acq: 07 Jun 2023 09:35

Instrument :

BNA_P

ClientSampleId :

SSTD005625





Tgt Ion:216 Resp: 60065

Ion Ratio Lower Upper

216 100

214 79.7 62.1 93.1

218 48.8 39.0 58.4



#76

Pentachlorobenzene

Concen: 4.629 ng/uL

RT: 15.063 min Scan# 2019

Delta R.T. 0.000 min

Lab File: BP015389.D

Acq: 07 Jun 2023 09:35




Tgt Ion:250 Resp: 66733


Ion Ratio Lower Upper

250 100

248 62.3 49.0 73.4

252 63.2 49.9 74.9









#77
 Carbazole
 Concen: 4.699 ng/ul
 RT: 17.904 min Scan# 2502
 Delta R.T. 0.000 min
 Lab File: BP015389.D
 Acq: 07 Jun 2023 09:35




Instrument :
 BNA_P
 ClientSampleId :
 SSTD005625


Tgt Ion:167 Resp: 229821
 Ion Ratio Lower Upper
 167 100
 166 21.0 17.2 25.8
 139 13.1 10.6 16.0



#78
 Di-n-butylphthalate
 Concen: 4.580 ng/ul
 RT: 18.434 min Scan# 2592
 Delta R.T. 0.000 min
 Lab File: BP015389.D
 Acq: 07 Jun 2023 09:35

Tgt Ion:149 Resp: 284216
 Ion Ratio Lower Upper
 149 100
 150 8.5 7.1 10.7
 104 5.5 4.5 6.7





#80

Fluoranthene

Concen: 4.599 ng/ul

RT: 19.498 min Scan# 2773

Delta R.T. 0.000 min

Lab File: BP015389.D

Acq: 07 Jun 2023 09:35

Instrument :

BNA_P

ClientSampleId :

SSTD005625





Tgt Ion:202 Resp: 310942

Ion Ratio Lower Upper

202 100

101 10.8 8.6 13.0

100 7.9 6.6 9.8



#82

Pyrene

Concen: 4.735 ng/ul

RT: 19.851 min Scan# 2833

Delta R.T. 0.000 min

Lab File: BP015389.D

Acq: 07 Jun 2023 09:35




Tgt Ion:202 Resp: 331226


Ion Ratio Lower Upper

202 100

101 12.5 9.9 14.9

100 10.7 8.4 12.6





#83

Butylbenzylphthalate

Concen: 4.312 ng/ul

RT: 20.704 min Scan# 2978

Delta R.T. 0.000 min

Lab File: BP015389.D

Acq: 07 Jun 2023 09:35

Instrument :

BNA_P

ClientSampleId :

SSTD005625





Tgt Ion:149 Resp: 133309

Ion Ratio Lower Upper

149 100

91 71.7 60.4 90.6

206 21.8 16.4 24.6



#84

3,3'-Dichlorobenzidine

Concen: 3.888 ng/ul

RT: 21.492 min Scan# 3112

Delta R.T. -0.000 min

Lab File: BP015389.D

Acq: 07 Jun 2023 09:35




Tgt Ion:252 Resp: 95254


Ion Ratio Lower Upper

252 100

254 64.7 51.6 77.4

126 11.7 8.5 12.7









#85
Benzo(a)anthracene
Concen: 4.652 ng/ul
RT: 21.563 min Scan# 3125
Delta R.T. -0.006 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35




Instrument :
BNA_P
ClientSampleId :
SSTD005625


Tgt Ion:228 Resp: 337255
Ion Ratio Lower Upper
228 100
229 19.6 15.8 23.6
226 27.1 21.8 32.6



#86
Bis(2-ethylhexyl)phthalate
Concen: 4.345 ng/ul
RT: 21.463 min Scan# 3107
Delta R.T. 0.000 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35

Tgt Ion:149 Resp: 198547
Ion Ratio Lower Upper
149 100
167 26.6 21.8 32.8
279 4.6 3.7 5.5









#87
Chrysene
Concen: 4.695 ng/ul
RT: 21.622 min Scan# 3134
Delta R.T. 0.000 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35




Instrument :
BNA_P
ClientSampleId :
SSTD005625


Tgt Ion:228 Resp: 321706
Ion Ratio Lower Upper
228 100
226 29.3 23.9 35.9
229 19.2 15.1 22.7



#89
Di-n-octyl phthalate
Concen: 4.481 ng/ul
RT: 22.433 min Scan# 3272
Delta R.T. 0.000 min
Lab File: BP015389.D
Acq: 07 Jun 2023 09:35

Tgt Ion:149 Resp: 344935





#90

Benzo(b)fluoranthene

Concen: 4.772 ng/ul

RT: 23.345 min Scan# 3428

Delta R.T. -0.006 min

Lab File: BP015389.D

Acq: 07 Jun 2023 09:35

Instrument :

BNA_P

ClientSampleId :

SSTD005625





Tgt Ion:252 Resp: 360745

Ion Ratio Lower Upper

252 100

253 21.1 17.4 26.2

125 9.8 7.3 10.9



#91

Benzo(k)fluoranthene

Concen: 4.704 ng/ul

RT: 23.398 min Scan# 3436

Delta R.T. -0.006 min

Lab File: BP015389.D

Acq: 07 Jun 2023 09:35




Tgt Ion:252 Resp: 343768


Ion Ratio Lower Upper

252 100

253 22.3 17.7 26.5

125 9.8 7.3 10.9





#93

Benzo(a)pyrene

Concen: 4.654 ng/ul

RT: 24.021 min Scan# 3

Delta R.T. -0.006 min

Lab File: BP015389.D

Acq: 07 Jun 2023 09:35

Instrument :

BNA_P

ClientSampleId :

SSTD005625





Tgt Ion:252 Resp: 306744

Ion Ratio Lower Upper

252 100

253 22.5 17.8 26.6

125 10.3 8.1 12.1



#94

Indeno(1,2,3-cd)pyrene

Concen: 4.438 ng/ul

RT: 26.845 min Scan# 4022

Delta R.T. 0.000 min

Lab File: BP015389.D

Acq: 07 Jun 2023 09:35




Tgt Ion:276 Resp: 378596


Ion Ratio Lower Upper

276 100

138 22.5 17.3 25.9

277 25.3 20.2 30.2





#95

Dibenzo(a,h)anthracene

Concen: 4.417 ng/ul

RT: 26.862 min Scan# 4025

Delta R.T. 0.000 min

Lab File: BP015389.D

Acq: 07 Jun 2023 09:35

Instrument :

BNA_P

ClientSampleId :

SSTD005625





Tgt Ion:278 Resp: 306997

Ion Ratio Lower Upper

278 100

139 14.9 11.4 17.0

279 25.5 18.7 28.1



#96

Benzo(g,h,i)perylene

Concen: 4.472 ng/ul

RT: 27.674 min Scan# 4163

Delta R.T. -0.012 min

Lab File: BP015389.D

Acq: 07 Jun 2023 09:35

Tgt Ion:276 Resp: 314341

Ion Ratio Lower Upper

276 100

138 20.0 15.6 23.4

277 25.1 19.0 28.6

