

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060723\
 Data File : BP015393.D
 Acq On : 07 Jun 2023 11:50
 Operator : MA/JU
 Sample : SST08087
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST080629

Quant Time: Jun 07 23:12:38 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jun 07 13:01:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.104	152	152647	20.000	ng/u1	0.00
20) Naphthalene-d8	10.934	136	681838	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.745	164	443000	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.504	188	1011781	20.000	ng/u1	0.00
79) Chrysene-d12	21.592	240	980217	20.000	ng/u1	0.00
88) Perylene-d12	24.145	264	1207582	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.411	96	128870	31.675	ng/uL	0.00
4) Pyridine-d5	3.846	84	884643	82.619	ng/u1	0.00
7) Phenol-d5	7.257	99	1172089	83.347	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.428	67	667030	79.422	ng/u1	0.00
11) 2-Chlorophenol-d4	7.628	132	856690	84.023	ng/u1	0.00
15) 4-Methylphenol-d8	8.828	113	944655	81.848	ng/u1	0.01
21) Nitrobenzene-d5	9.287	128	452818	84.590	ng/u1	0.00
24) 2-Nitrophenol-d4	10.010	143	530409	86.764	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.551	165	936200	83.953	ng/u1	0.00
31) 4-Chloroaniline-d4	11.075	131	1263691	78.948	ng/u1	0.00
46) Dimethylphthalate-d6	14.157	166	2761818	79.060	ng/u1	0.00
49) Acenaphthylene-d8	14.445	160	3082225	80.691	ng/u1	0.00
54) 4-Nitrophenol-d4	14.957	143	531125	85.738	ng/u1	0.01
60) Fluorene-d10	15.745	176	2281991	77.465	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.875	200	546702	85.293	ng/u1	0.01
73) Anthracene-d10	17.610	188	3628344	78.818	ng/u1	0.01
81) Pyrene-d10	19.827	212	4248016	80.972	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.986	264	4941846	81.560	ng/u1	0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.446	88	136832	31.944	ng/uL	93
5) Pyridine	3.869	79	921537	83.016	ng/u1	99
6) Benzaldehyde	7.234	77	501111	95.256	ng/u1	99
8) Phenol	7.287	94	1196919	82.499	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.522	93	919688	80.176	ng/u1	99
12) 2-Chlorophenol	7.663	128	901017	83.640	ng/u1	98
13) 2-Methylphenol	8.552	108	924621	82.831	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.640	45	1204522	79.240	ng/u1	98
16) Acetophenone	8.940	105	1467274	79.660	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.934	70	771656	77.621	ng/u1	98
18) 4-Methylphenol	8.893	108	998975	81.338	ng/u1	98
19) Hexachloroethane	9.193	117	365675	80.329	ng/u1	98
22) Nitrobenzene	9.328	77	1175264	82.787	ng/u1	98
23) Isophorone	9.863	82	2292660	80.676	ng/u1	100
25) 2-Nitrophenol	10.046	139	556983	84.387	ng/u1	97
26) 2,4-Dimethylphenol	10.098	107	1139654	81.420	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.334	93	1305649	79.026	ng/u1	99
29) 2,4-Dichlorophenol	10.581	162	930821	83.952	ng/u1	98
30) Naphthalene	10.987	128	2933610	79.399	ng/u1	100
32) 4-Chloroaniline	11.098	127	1319503	79.669	ng/u1	99
33) Hexachlorobutadiene	11.263	225	598347	80.020	ng/u1	97
34) Caprolactam	11.904	113	195239	46.966	ng/u1	95
35) 4-Chloro-3-methylphenol	12.216	107	1091179	81.226	ng/u1	99
36) 2-Methylnaphthalene	12.581	142	2022601	77.869	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060723\
 Data File : BP015393.D
 Acq On : 07 Jun 2023 11:50
 Operator : MA/JU
 Sample : SST08087
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST080629

Quant Time: Jun 07 23:12:38 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jun 07 13:01:04 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.804	142	2018880	77.198	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.945	216	1141880	82.777	ng/ul	99
40) Hexachlorocyclopentadiene	12.922	237	561280	99.262	ng/ul	95
41) 2,4,6-Trichlorophenol	13.187	196	794992	86.588	ng/ul	99
42) 2,4,5-Trichlorophenol	13.263	196	860807	86.038	ng/ul	98
43) 1,1'-Biphenyl	13.587	154	2731018	79.717	ng/ul	100
44) 2-Chloronaphthalene	13.628	162	2165572	80.568	ng/ul	99
45) 2-Nitroaniline	13.839	65	758174	88.134	ng/ul	98
47) Dimethylphthalate	14.204	163	2830996	78.490	ng/ul	100
48) 2,6-Dinitrotoluene	14.334	165	632243	85.559	ng/ul	98
50) Acenaphthylene	14.475	152	3342337	79.014	ng/ul	100
51) 3-Nitroaniline	14.663	138	544177	89.207	ng/ul	100
52) Acenaphthene	14.816	153	2296011	78.522	ng/ul	99
53) 2,4-Dinitrophenol	14.869	184	417309	93.879	ng/ul	99
55) 4-Nitrophenol	14.975	109	517603	85.779	ng/ul	93
56) Dibenzofuran	15.145	168	3177498	76.671	ng/ul	98
57) 2,4-Dinitrotoluene	15.122	165	902951	83.079	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.375	232	750514	83.081	ng/ul	97
59) Diethylphthalate	15.563	149	2868825	77.607	ng/ul	100
61) Fluorene	15.798	166	2538146	75.129	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.786	204	1289896	75.757	ng/ul	100
63) 4-Nitroaniline	15.833	138	530371	92.319	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.886	198	550887	84.269	ng/ul	94
67) N-Nitrosodiphenylamine	16.004	169	2235464	77.880	ng/ul	99
68) 4-Bromophenyl-phenylether	16.686	248	876092	80.937	ng/ul	99
69) Hexachlorobenzene	16.804	284	1027222	80.451	ng/ul	96
70) Atrazine	16.963	200	926713	80.848	ng/ul	99
71) Pentachlorophenol	17.151	266	635598	88.593	ng/ul	99
72) Phenanthrene	17.551	178	4249384	78.263	ng/ul	100
74) Anthracene	17.645	178	4250536	77.155	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.551	216	1175812	84.203	ng/uL	99
76) Pentachlorobenzene	15.069	250	1171082	80.099	ng/uL	98
77) Carbazole	17.910	167	3931990	79.266	ng/ul	99
78) Di-n-butylphthalate	18.439	149	4940170	78.494	ng/ul	99
80) Fluoranthene	19.504	202	5165160	80.826	ng/ul	98
82) Pyrene	19.857	202	5226598	79.054	ng/ul	98
83) Butylbenzylphthalate	20.710	149	2430382	83.170	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.498	252	1925406	83.143	ng/ul	99
85) Benzo(a)anthracene	21.574	228	5446186	79.478	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.463	149	3536004	81.867	ng/ul	100
87) Chrysene	21.633	228	5095369	78.668	ng/ul	98
89) Di-n-octyl phthalate	22.439	149	6310039	79.506	ng/ul	100
90) Benzo(b)fluoranthene	23.362	252	6126949	78.602	ng/ul	100
91) Benzo(k)fluoranthene	23.415	252	5991775	79.519	ng/ul	100
93) Benzo(a)pyrene	24.045	252	5445020	80.120	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.886	276	7464284	84.867	ng/ul	98
95) Dibenzo(a,h)anthracene	26.898	278	6073276	84.757	ng/ul	99
96) Benzo(g,h,i)perylene	27.721	276	6148613	84.830	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060723\
 Data File : BP015393.D
 Acq On : 07 Jun 2023 11:50
 Operator : MA/JU
 Sample : SSTD08087
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080629

Quant Time: Jun 07 23:12:38 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jun 07 13:01:04 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 31.944 ng/uL
RT: 3.446 min Scan# 44
Delta R.T. -0.006 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD080629


Tgt Ion: 88 Resp: 136832
Ion Ratio Lower Upper
88 100
43 30.3 23.9 35.9
58 71.5 51.4 77.2



#5
Pyridine
Concen: 83.016 ng/uL
RT: 3.869 min Scan# 116
Delta R.T. -0.006 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50

Tgt Ion: 79 Resp: 921537
Ion Ratio Lower Upper
79 100
52 73.9 59.5 89.3
51 35.8 29.8 44.6









#6
Benzaldehyde
Concen: 95.256 ng/ul
RT: 7.234 min Scan# 688
Delta R.T. -0.000 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD080629


Tgt Ion: 77 Resp: 501111
Ion Ratio Lower Upper
77 100
105 95.9 77.3 115.9
106 90.2 73.3 109.9



#8
Phenol
Concen: 82.499 ng/ul
RT: 7.287 min Scan# 697
Delta R.T. 0.006 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50

Tgt Ion: 94 Resp: 1196919
Ion Ratio Lower Upper
94 100
65 30.7 23.5 35.3
66 43.2 34.2 51.4









#10
Bis(2-Chloroethyl)ether
Concen: 80.176 ng/ul
RT: 7.522 min Scan# 71
Delta R.T. 0.006 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD080629


Tgt Ion: 93 Resp: 919688
Ion Ratio Lower Upper
93 100
63 76.7 62.6 93.8
95 32.3 25.5 38.3



#12
2-Chlorophenol
Concen: 83.640 ng/ul
RT: 7.663 min Scan# 761
Delta R.T. 0.006 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50

Tgt Ion: 128 Resp: 901017
Ion Ratio Lower Upper
128 100
64 48.9 40.2 60.4
130 32.1 24.5 36.7









#13
2-Methylphenol
Concen: 82.831 ng/ul
RT: 8.552 min Scan# 911
Delta R.T. 0.006 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD080629


Tgt Ion:108 Resp: 924621
Ion Ratio Lower Upper
108 100
107 88.1 71.2 106.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 79.240 ng/ul
RT: 8.640 min Scan# 927
Delta R.T. 0.012 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50

Tgt Ion: 45 Resp: 1204522
Ion Ratio Lower Upper
45 100
77 15.5 11.8 17.8
79 11.9 8.7 13.1









#16
Acetophenone
Concen: 79.660 ng/ul
RT: 8.940 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD080629


Tgt Ion:105 Resp: 1467274
Ion Ratio Lower Upper
105 100
77 81.0 64.9 97.3
51 28.8 23.5 35.3



#17
N-Nitroso-di-n-propylamine
Concen: 77.621 ng/ul
RT: 8.934 min Scan# 977
Delta R.T. 0.017 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50

Tgt Ion: 70 Resp: 771656
Ion Ratio Lower Upper
70 100
42 59.5 46.6 69.8
101 9.9 7.8 11.6
130 19.1 15.9 23.9









#18
4-Methylphenol
Concen: 81.338 ng/ul
RT: 8.893 min Scan# 91
Delta R.T. 0.012 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD080629


Tgt Ion:108 Resp: 998975
Ion Ratio Lower Upper
108 100
107 113.3 89.3 133.9



#19
Hexachloroethane
Concen: 80.329 ng/ul
RT: 9.193 min Scan# 1021
Delta R.T. -0.000 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50

Tgt Ion:117 Resp: 365675
Ion Ratio Lower Upper
117 100
201 102.4 82.8 124.2
199 66.9 50.6 76.0









#22
Nitrobenzene
Concen: 82.787 ng/ul
RT: 9.328 min Scan# 1043
Delta R.T. 0.006 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD080629


Tgt Ion: 77 Resp: 1175264
Ion Ratio Lower Upper
77 100
123 43.1 33.5 50.3
65 14.2 11.6 17.4



#23
Isophorone
Concen: 80.676 ng/ul
RT: 9.863 min Scan# 1135
Delta R.T. 0.012 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50

Tgt Ion: 82 Resp: 2292660
Ion Ratio Lower Upper
82 100
95 7.6 5.8 8.8
138 16.3 13.1 19.7









#25
2-Nitrophenol
Concen: 84.387 ng/ul
RT: 10.046 min Scan# 1165
Delta R.T. 0.006 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD080629


Tgt Ion:139 Resp: 556983
Ion Ratio Lower Upper
139 100
65 55.5 46.9 70.3
109 36.8 30.2 45.2



#26
2,4-Dimethylphenol
Concen: 81.420 ng/ul
RT: 10.098 min Scan# 1175
Delta R.T. 0.006 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50

Tgt Ion:107 Resp: 1139654
Ion Ratio Lower Upper
107 100
121 47.9 39.0 58.6
122 80.4 64.8 97.2









#27
 Bis(2-Chloroethoxy)methane
 Concen: 79.026 ng/ul
 RT: 10.334 min Scan# 1215
 Delta R.T. -0.000 min
 Lab File: BP015393.D
 Acq: 07 Jun 2023 11:50

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080629

Tgt Ion: 93 Resp: 1305649




Ion	Ratio	Lower	Upper
93	100		
95	32.7	25.7	38.5
123	10.7	8.6	13.0




#29
 2,4-Dichlorophenol
 Concen: 83.952 ng/ul
 RT: 10.581 min Scan# 1257
 Delta R.T. 0.006 min
 Lab File: BP015393.D
 Acq: 07 Jun 2023 11:50

Tgt Ion: 162 Resp: 930821

Ion	Ratio	Lower	Upper
162	100		
164	65.5	50.6	76.0
98	37.5	30.2	45.2









#30
Naphthalene
Concen: 79.399 ng/ul
RT: 10.987 min Scan# 1326
Delta R.T. 0.006 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD080629


Tgt Ion:128 Resp: 2933610
Ion Ratio Lower Upper
128 100
129 10.6 8.6 13.0
127 13.1 10.5 15.7



#32
4-Chloroaniline
Concen: 79.669 ng/ul
RT: 11.098 min Scan# 1345
Delta R.T. 0.006 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50

Tgt Ion:127 Resp: 1319503
Ion Ratio Lower Upper
127 100
129 32.0 26.1 39.1





#33

Hexachlorobutadiene

Concen: 80.020 ng/ul

RT: 11.263 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP015393.D

Acq: 07 Jun 2023 11:50

Instrument :

BNA_P

ClientSampleId :

SSTD080629





Tgt Ion:225 Resp: 598347

Ion Ratio Lower Upper

225 100

223 62.0 52.0 78.0

227 62.3 50.6 75.8



#34

Caprolactam

Concen: 46.966 ng/ul

RT: 11.904 min Scan# 1482

Delta R.T. 0.035 min

Lab File: BP015393.D

Acq: 07 Jun 2023 11:50




Tgt Ion:113 Resp: 195239


Ion Ratio Lower Upper

113 100

55 169.1 127.9 191.9

56 128.8 101.1 151.7









#35
4-Chloro-3-methylphenol
Concen: 81.226 ng/ul
RT: 12.216 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD080629


Tgt Ion:107 Resp: 1091179
Ion Ratio Lower Upper
107 100
144 24.1 19.4 29.0
142 74.1 60.4 90.6



#36
2-Methylnaphthalene
Concen: 77.869 ng/ul
RT: 12.581 min Scan# 1597
Delta R.T. -0.000 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50

Tgt Ion:142 Resp: 2022601
Ion Ratio Lower Upper
142 100
141 88.5 72.0 108.0





#37

1-Methylnaphthalene

Concen: 77.198 ng/ul

RT: 12.804 min Scan# 1635

Delta R.T. 0.006 min

Lab File: BP015393.D

Acq: 07 Jun 2023 11:50

Instrument :

BNA_P

ClientSampleId :

SSTD080629





Tgt Ion:142 Resp: 2018880

Ion Ratio Lower Upper

142 100

141 92.6 71.9 107.9

116 4.4 3.4 5.2



#39

1,2,4,5-Tetrachlorobenzene

Concen: 82.777 ng/ul

RT: 12.945 min Scan# 1659

Delta R.T. -0.000 min

Lab File: BP015393.D

Acq: 07 Jun 2023 11:50

Tgt Ion:216 Resp: 1141880




Ion Ratio Lower Upper


216 100

214 78.5 62.2 93.2

179 21.8 17.4 26.2

108 18.1 14.2 21.4





#40

Hexachlorocyclopentadiene

Concen: 99.262 ng/ul

RT: 12.922 min Scan# 1655

Delta R.T. -0.000 min

Lab File: BP015393.D

Acq: 07 Jun 2023 11:50

Instrument :

BNA_P

ClientSampleId :

SSTD080629





Tgt Ion:237 Resp: 561280

Ion Ratio Lower Upper

237 100

235 64.1 48.2 72.2

272 12.3 10.3 15.5



#41

2,4,6-Trichlorophenol

Concen: 86.588 ng/ul

RT: 13.187 min Scan# 1700

Delta R.T. -0.000 min

Lab File: BP015393.D

Acq: 07 Jun 2023 11:50




Tgt Ion:196 Resp: 794992


Ion Ratio Lower Upper

196 100

198 96.7 77.8 116.6

200 30.8 25.8 38.6





#42

2,4,5-Trichlorophenol

Concen: 86.038 ng/ul

RT: 13.263 min Scan# 1713

Delta R.T. 0.006 min

Lab File: BP015393.D

Acq: 07 Jun 2023 11:50

Instrument :

BNA_P

ClientSampleId :

SSTD080629





Tgt Ion:196 Resp: 860807

Ion Ratio Lower Upper

196 100

198 97.0 75.5 113.3

200 31.2 24.7 37.1



#43

1,1'-Biphenyl

Concen: 79.717 ng/ul

RT: 13.587 min Scan# 1768

Delta R.T. 0.006 min

Lab File: BP015393.D

Acq: 07 Jun 2023 11:50




Tgt Ion:154 Resp: 2731018


Ion Ratio Lower Upper

154 100

153 40.3 32.2 48.2

76 12.8 10.4 15.6









#44
2-Chloronaphthalene
Concen: 80.568 ng/ul
RT: 13.628 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD080629


Tgt Ion:162 Resp: 2165572
Ion Ratio Lower Upper
162 100
164 32.6 27.0 40.6
127 40.8 32.2 48.4



#45
2-Nitroaniline
Concen: 88.134 ng/ul
RT: 13.839 min Scan# 1811
Delta R.T. 0.006 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50

Tgt Ion: 65 Resp: 758174
Ion Ratio Lower Upper
65 100
92 68.4 55.4 83.2
138 100.8 78.8 118.2









#47
Dimethylphthalate
Concen: 78.490 ng/ul
RT: 14.204 min Scan# 1873
Delta R.T. 0.006 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD080629


Tgt Ion:163 Resp: 2830996
Ion Ratio Lower Upper
163 100
194 4.0 3.2 4.8
164 10.3 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 85.559 ng/ul
RT: 14.334 min Scan# 1895
Delta R.T. 0.012 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50

Tgt Ion:165 Resp: 632243
Ion Ratio Lower Upper
165 100
89 62.7 51.5 77.3
121 21.2 17.3 25.9









#50
Acenaphthylene
Concen: 79.014 ng/ul
RT: 14.475 min Scan# 1918
Delta R.T. 0.006 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD080629


Tgt Ion:152 Resp: 3342337
Ion Ratio Lower Upper
152 100
151 20.0 16.0 24.0
153 13.1 10.6 15.8



#51
3-Nitroaniline
Concen: 89.207 ng/ul
RT: 14.663 min Scan# 1951
Delta R.T. 0.006 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50

Tgt Ion:138 Resp: 544177
Ion Ratio Lower Upper
138 100
108 11.0 8.4 12.6
92 122.6 98.0 147.0





#52

Acenaphthene

Concen: 78.522 ng/ul

RT: 14.816 min Scan# 1976

Delta R.T. 0.006 min

Lab File: BP015393.D

Acq: 07 Jun 2023 11:50

Instrument :

BNA_P

ClientSampleId :

SSTD080629





Tgt Ion:153 Resp: 2296011

Ion Ratio Lower Upper

153 100

152 47.7 38.3 57.5

154 88.3 69.8 104.6



#53

2,4-Dinitrophenol

Concen: 93.879 ng/ul

RT: 14.869 min Scan# 1986

Delta R.T. -0.000 min

Lab File: BP015393.D

Acq: 07 Jun 2023 11:50




Tgt Ion:184 Resp: 417309


Ion Ratio Lower Upper

184 100

63 74.8 59.0 88.4

154 71.0 57.6 86.4









#55
4-Nitrophenol
Concen: 85.779 ng/ul
RT: 14.975 min Scan# 2004
Delta R.T. 0.017 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD080629


Tgt Ion:109 Resp: 517603
Ion Ratio Lower Upper
109 100
139 106.9 89.1 133.7
65 118.2 103.1 154.7



#56
Dibenzofuran
Concen: 76.671 ng/ul
RT: 15.145 min Scan# 2033
Delta R.T. -0.000 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50

Tgt Ion:168 Resp: 3177498
Ion Ratio Lower Upper
168 100
139 40.3 31.2 46.8





#57

2,4-Dinitrotoluene

Concen: 83.079 ng/ul

RT: 15.122 min Scan# 2027

Delta R.T. 0.012 min

Lab File: BP015393.D

Acq: 07 Jun 2023 11:50

Instrument :

BNA_P

ClientSampleId :

SSTD080629





Tgt Ion:165 Resp: 902951

Ion Ratio Lower Upper

165 100

63 48.3 39.1 58.7

182 2.7 2.2 3.4



#58

2,3,4,6-Tetrachlorophenol

Concen: 83.081 ng/ul

RT: 15.375 min Scan# 2072

Delta R.T. 0.006 min

Lab File: BP015393.D

Acq: 07 Jun 2023 11:50

Tgt Ion:232 Resp: 750514




Ion Ratio Lower Upper


232 100

131 44.3 37.8 56.8

130 2.4 1.9 2.9

166 31.5 26.2 39.2









#59
Diethylphthalate
Concen: 77.607 ng/ul
RT: 15.563 min Scan# 2104
Delta R.T. 0.006 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD080629


Tgt Ion:149 Resp: 2868825
Ion Ratio Lower Upper
149 100
177 22.2 17.8 26.8
150 12.7 10.1 15.1



#61
Fluorene
Concen: 75.129 ng/ul
RT: 15.798 min Scan# 2144
Delta R.T. 0.006 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50

Tgt Ion:166 Resp: 2538146
Ion Ratio Lower Upper
166 100
165 99.3 79.1 118.7
167 13.6 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 75.757 ng/ul
RT: 15.786 min Scan# 2141
Delta R.T. 0.006 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD080629


Tgt Ion:204 Resp: 1289896
Ion Ratio Lower Upper
204 100
206 32.9 26.6 40.0
141 62.5 49.8 74.8



#63
4-Nitroaniline
Concen: 92.319 ng/ul
RT: 15.833 min Scan# 2150
Delta R.T. 0.017 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50

Tgt Ion:138 Resp: 530371
Ion Ratio Lower Upper
138 100
92 56.1 44.6 67.0
108 97.5 78.4 117.6









#66
4,6-Dinitro-2-methylphenol
Concen: 84.269 ng/ul
RT: 15.886 min Scan# 2158
Delta R.T. 0.006 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD080629


Tgt Ion:198 Resp: 550887
Ion Ratio Lower Upper
198 100
182 3.4 3.0 4.4
77 28.7 20.5 30.7



#67
N-Nitrosodiphenylamine
Concen: 77.880 ng/ul
RT: 16.004 min Scan# 2179
Delta R.T. 0.006 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50

Tgt Ion:169 Resp: 2235464
Ion Ratio Lower Upper
169 100
168 66.7 53.0 79.4
167 36.7 28.9 43.3









#68
4-Bromophenyl-phenylether
Concen: 80.937 ng/ul
RT: 16.686 min Scan# 2295
Delta R.T. 0.006 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD080629


Tgt Ion:248 Resp: 876092
Ion Ratio Lower Upper
248 100
250 97.2 78.3 117.5
141 74.7 61.1 91.7



#69
Hexachlorobenzene
Concen: 80.451 ng/ul
RT: 16.804 min Scan# 2315
Delta R.T. -0.000 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50

Tgt Ion:284 Resp: 1027222
Ion Ratio Lower Upper
284 100
142 35.2 25.4 38.2
249 31.6 24.8 37.2





#70

Atrazine

Concen: 80.848 ng/ul

RT: 16.963 min Scan# 21

Delta R.T. 0.012 min

Lab File: BP015393.D

Acq: 07 Jun 2023 11:50

Instrument :

BNA_P

ClientSampleId :

SSTD080629





Tgt Ion:200 Resp: 926713

Ion Ratio Lower Upper

200 100

173 25.0 19.8 29.6

215 45.6 37.4 56.0



#71

Pentachlorophenol

Concen: 88.593 ng/ul

RT: 17.151 min Scan# 2374

Delta R.T. -0.000 min

Lab File: BP015393.D

Acq: 07 Jun 2023 11:50




Tgt Ion:266 Resp: 635598


Ion Ratio Lower Upper

266 100

264 63.7 49.6 74.4

268 63.5 51.0 76.6





#72

Phenanthrene

Concen: 78.263 ng/ul

RT: 17.551 min Scan# 2441

Delta R.T. 0.006 min

Lab File: BP015393.D

Acq: 07 Jun 2023 11:50

Instrument :

BNA_P

ClientSampleId :

SSTD080629





Tgt Ion:178 Resp: 4249384

Ion Ratio Lower Upper

178 100

179 15.3 12.3 18.5

176 19.4 15.4 23.0



#74

Anthracene

Concen: 77.155 ng/ul

RT: 17.645 min Scan# 2458

Delta R.T. 0.012 min

Lab File: BP015393.D

Acq: 07 Jun 2023 11:50




Tgt Ion:178 Resp: 4250536


Ion Ratio Lower Upper

178 100

179 15.3 12.4 18.6

176 18.8 15.0 22.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 84.203 ng/uL
RT: 13.551 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD080629


Tgt Ion:216 Resp: 1175812
Ion Ratio Lower Upper
216 100
214 78.2 62.1 93.1
218 47.5 39.0 58.4



#76
Pentachlorobenzene
Concen: 80.099 ng/uL
RT: 15.069 min Scan# 2020
Delta R.T. 0.006 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50

Tgt Ion:250 Resp: 1171082
Ion Ratio Lower Upper
250 100
248 63.2 49.0 73.4
252 63.8 49.9 74.9





#77

Carbazole

Concen: 79.266 ng/ul

RT: 17.910 min Scan# 2502

Delta R.T. 0.006 min

Lab File: BP015393.D

Acq: 07 Jun 2023 11:50

Instrument :

BNA_P

ClientSampleId :

SSTD080629





Tgt Ion:167 Resp: 3931990

Ion Ratio Lower Upper

167 100

166 21.8 17.2 25.8

139 13.4 10.6 16.0



#78

Di-n-butylphthalate

Concen: 78.494 ng/ul

RT: 18.439 min Scan# 2593

Delta R.T. 0.006 min

Lab File: BP015393.D

Acq: 07 Jun 2023 11:50




Tgt Ion:149 Resp: 4940170


Ion Ratio Lower Upper

149 100

150 9.2 7.1 10.7

104 5.8 4.5 6.7





#80

Fluoranthene

Concen: 80.826 ng/ul

RT: 19.504 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP015393.D

Acq: 07 Jun 2023 11:50

Instrument :

BNA_P

ClientSampleId :

SSTD080629





Tgt Ion:202 Resp: 5165160

Ion Ratio Lower Upper

202 100

101 9.8 8.6 13.0

100 7.5 6.6 9.8



#82

Pyrene

Concen: 79.054 ng/ul

RT: 19.857 min Scan# 2834

Delta R.T. 0.006 min

Lab File: BP015393.D

Acq: 07 Jun 2023 11:50




Tgt Ion:202 Resp: 5226598


Ion Ratio Lower Upper

202 100

101 11.7 9.9 14.9

100 9.5 8.4 12.6





#83

Butylbenzylphthalate

Concen: 83.170 ng/ul

RT: 20.710 min Scan# 2979

Delta R.T. 0.006 min

Lab File: BP015393.D

Acq: 07 Jun 2023 11:50

Instrument :

BNA_P

ClientSampleId :

SSTD080629





Tgt Ion:149 Resp: 2430382

Ion Ratio Lower Upper

149 100

91 74.9 60.4 90.6

206 21.9 16.4 24.6



#84

3,3'-Dichlorobenzidine

Concen: 83.143 ng/ul

RT: 21.498 min Scan# 3113

Delta R.T. 0.006 min

Lab File: BP015393.D

Acq: 07 Jun 2023 11:50




Tgt Ion:252 Resp: 1925406


Ion Ratio Lower Upper

252 100

254 64.8 51.6 77.4

126 10.0 8.5 12.7









#85
Benzo(a)anthracene
Concen: 79.478 ng/ul
RT: 21.574 min Scan# 3125
Delta R.T. 0.006 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD080629


Tgt Ion:228 Resp: 5446186
Ion Ratio Lower Upper
228 100
229 19.8 15.8 23.6
226 27.5 21.8 32.6



#86
Bis(2-ethylhexyl)phthalate
Concen: 81.867 ng/ul
RT: 21.463 min Scan# 3107
Delta R.T. -0.000 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50

Tgt Ion:149 Resp: 3536004
Ion Ratio Lower Upper
149 100
167 27.2 21.8 32.8
279 4.6 3.7 5.5





#87

Chrysene

Concen: 78.668 ng/ul

RT: 21.633 min Scan# 3134

Delta R.T. 0.012 min

Lab File: BP015393.D

Acq: 07 Jun 2023 11:50

Instrument :

BNA_P

ClientSampleId :

SSTD080629





Tgt Ion:228 Resp: 5095369

Ion Ratio Lower Upper

228 100

226 30.7 23.9 35.9

229 19.7 15.1 22.7



#89

Di-n-octyl phthalate

Concen: 79.506 ng/ul




RT: 22.439 min Scan# 3273


Delta R.T. 0.006 min

Lab File: BP015393.D

Acq: 07 Jun 2023 11:50

Tgt Ion:149 Resp: 6310039









#90
Benzo(b)fluoranthene
Concen: 78.602 ng/ul
RT: 23.362 min Scan# 3428
Delta R.T. 0.012 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD080629


Tgt Ion:252 Resp: 6126949
Ion Ratio Lower Upper
252 100
253 21.8 17.4 26.2
125 8.8 7.3 10.9



#91
Benzo(k)fluoranthene
Concen: 79.519 ng/ul
RT: 23.415 min Scan# 3439
Delta R.T. 0.012 min
Lab File: BP015393.D
Acq: 07 Jun 2023 11:50

Tgt Ion:252 Resp: 5991775
Ion Ratio Lower Upper
252 100
253 22.2 17.7 26.5
125 8.8 7.3 10.9





#93

Benzo(a)pyrene

Concen: 80.120 ng/ul

RT: 24.045 min Scan# 3

Delta R.T. 0.017 min

Lab File: BP015393.D

Acq: 07 Jun 2023 11:50

Instrument :

BNA_P

ClientSampleId :

SSTD080629





Tgt Ion:252 Resp: 5445020

Ion Ratio Lower Upper

252 100

253 21.8 17.8 26.6

125 9.6 8.1 12.1



#94

Indeno(1,2,3-cd)pyrene

Concen: 84.867 ng/ul

RT: 26.886 min Scan# 4029

Delta R.T. 0.041 min

Lab File: BP015393.D

Acq: 07 Jun 2023 11:50




Tgt Ion:276 Resp: 7464284


Ion Ratio Lower Upper

276 100

138 22.8 17.3 25.9

277 26.2 20.2 30.2





#95

Dibenzo(a,h)anthracene

Concen: 84.757 ng/ul

RT: 26.898 min Scan# 4025

Delta R.T. 0.035 min

Lab File: BP015393.D

Acq: 07 Jun 2023 11:50

Instrument :

BNA_P

ClientSampleId :

SSTD080629





Tgt Ion:278 Resp: 6073276

Ion Ratio Lower Upper

278 100

139 13.4 11.4 17.0

279 23.7 18.7 28.1



#96

Benzo(g,h,i)perylene

Concen: 84.830 ng/ul

RT: 27.721 min Scan# 4171

Delta R.T. 0.035 min

Lab File: BP015393.D

Acq: 07 Jun 2023 11:50

Tgt Ion:276 Resp: 6148613

Ion Ratio Lower Upper

276 100

138 18.8 15.6 23.4

277 23.5 19.0 28.6

