

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060723\
 Data File : BP015411.D
 Acq On : 07 Jun 2023 22:03
 Operator : MA/JU
 Sample : 02970-10
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SVOC-GPC2-BLANK

Quant Time: Jun 08 00:14:24 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jun 07 23:46:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.099	152	143924	20.000	ng/u1	0.00
20) Naphthalene-d8	10.928	136	622313	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.739	164	468561	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.498	188	1110114	20.000	ng/u1	0.00
79) Chrysene-d12	21.580	240	1118173	20.000	ng/u1	0.00
88) Perylene-d12	24.133	264	1199824	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.411	96	170	0.044	ng/uL	0.00
4) Pyridine-d5	3.846	84	250	0.025	ng/u1	0.00
7) Phenol-d5	7.240	99	12	0.001	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	7.422	67	158	0.020	ng/u1	0.00
11) 2-Chlorophenol-d4	7.622	132	19	0.002	ng/u1	0.00
15) 4-Methylphenol-d8	8.810	113	8	0.001	ng/u1	0.00
21) Nitrobenzene-d5	9.281	128	37	0.008	ng/u1	0.00
24) 2-Nitrophenol-d4	10.022	143	22	0.004	ng/u1	0.02
28) 2,4-Dichlorophenol-d3	10.557	165	48	0.005	ng/u1	0.00
31) 4-Chloroaniline-d4	11.063	131	5	0.000	ng/u1	0.00
46) Dimethylphthalate-d6	14.145	166	8	0.000	ng/u1	0.00
49) Acenaphthylene-d8	14.439	160	18	0.000	ng/u1	0.00
54) 4-Nitrophenol-d4	14.951	143	13	0.002	ng/u1	0.00
60) Fluorene-d10	15.722	176	9	0.000	ng/u1	-0.02
65) 4,6-Dinitro-2-methylph...	15.851	200	10	0.001	ng/u1	-0.01
73) Anthracene-d10	17.498	188	1110114	21.979	ng/u1	-0.10
81) Pyrene-d10	19.810	212	64	0.001	ng/u1	-0.01
92) Benzo(a)pyrene-d12	24.133	264	991274	16.466	ng/u1	0.16

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060723\
Data File : BP015411.D
Acq On : 07 Jun 2023 22:03
Operator : MA/JU
Sample : 02970-10
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SVOC-GPC2-BLANK

Quant Time: Jun 08 00:14:24 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jun 07 23:46:15 2023
Response via : Initial Calibration

