

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060723\
 Data File : BP015412.D
 Acq On : 07 Jun 2023 22:36
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020737

Quant Time: Jun 08 00:14:32 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jun 07 23:46:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.099	152	148271	20.000	ng/u1	0.00
20) Naphthalene-d8	10.928	136	693112	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.740	164	478495	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.498	188	1128119	20.000	ng/u1	0.00
79) Chrysene-d12	21.580	240	1116222	20.000	ng/u1	0.00
88) Perylene-d12	24.133	264	1254107	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.411	96	33434	8.350	ng/uL	0.00
4) Pyridine-d5	3.852	84	200392	19.267	ng/u1	0.00
7) Phenol-d5	7.252	99	285430	20.896	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.417	67	179454	21.998	ng/u1	0.00
11) 2-Chlorophenol-d4	7.622	132	214418	21.650	ng/u1	0.00
15) 4-Methylphenol-d8	8.811	113	237708	21.204	ng/u1	0.00
21) Nitrobenzene-d5	9.275	128	117521	21.597	ng/u1	0.00
24) 2-Nitrophenol-d4	10.005	143	134354	21.620	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.546	165	245960	21.697	ng/u1	0.00
31) 4-Chloroaniline-d4	11.063	131	356566	21.914	ng/u1	0.00
46) Dimethylphthalate-d6	14.146	166	805915	21.359	ng/u1	0.00
49) Acenaphthylene-d8	14.440	160	887770	21.517	ng/u1	0.00
54) 4-Nitrophenol-d4	14.940	143	136151	20.150	ng/u1	0.00
60) Fluorene-d10	15.734	176	681692	21.424	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.857	200	144021	20.152	ng/u1	0.00
73) Anthracene-d10	17.598	188	1094432	21.322	ng/u1	0.00
81) Pyrene-d10	19.822	212	1306660	21.872	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.969	264	1334657	21.210	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.446	88	36003	8.512	ng/uL	90
5) Pyridine	3.870	79	223913	20.766	ng/u1	97
6) Benzaldehyde	7.228	77	108188	20.721	ng/u1	97
8) Phenol	7.281	94	293808	20.849	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.511	93	241580	21.682	ng/u1	99
12) 2-Chlorophenol	7.658	128	227912	21.781	ng/u1	97
13) 2-Methylphenol	8.546	108	231221	21.325	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.634	45	331504	22.452	ng/u1	99
16) Acetophenone	8.934	105	395284	22.094	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.911	70	216132	22.382	ng/u1	98
18) 4-Methylphenol	8.875	108	257871	21.616	ng/u1	98
19) Hexachloroethane	9.187	117	98270	22.224	ng/u1	98
22) Nitrobenzene	9.316	77	309893	21.474	ng/u1	97
23) Isophorone	9.846	82	616435	21.339	ng/u1	99
25) 2-Nitrophenol	10.034	139	146217	21.793	ng/u1	99
26) 2,4-Dimethylphenol	10.087	107	304821	21.423	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.328	93	359409	21.400	ng/u1	99
29) 2,4-Dichlorophenol	10.569	162	247140	21.927	ng/u1	98
30) Naphthalene	10.975	128	796296	21.201	ng/u1	100
32) 4-Chloroaniline	11.093	127	360982	21.441	ng/u1	99
33) Hexachlorobutadiene	11.257	225	163997	21.575	ng/u1	98
34) Caprolactam	11.863	113	89322	21.053	ng/u1	98
35) 4-Chloro-3-methylphenol	12.204	107	300296	21.990	ng/u1	98
36) 2-Methylnaphthalene	12.575	142	564533	21.381	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060723\
 Data File : BP015412.D
 Acq On : 07 Jun 2023 22:36
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020737


Quant Time: Jun 08 00:14:32 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jun 07 23:46:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.793	142	571174	21.485	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.940	216	320552	21.514	ng/ul	97
40) Hexachlorocyclopentadiene	12.916	237	123540	20.227	ng/ul	97
41) 2,4,6-Trichlorophenol	13.181	196	217093	21.891	ng/ul	97
42) 2,4,5-Trichlorophenol	13.257	196	236985	21.930	ng/ul	99
43) 1,1'-Biphenyl	13.575	154	786768	21.262	ng/ul	99
44) 2-Chloronaphthalene	13.622	162	619841	21.350	ng/ul	99
45) 2-Nitroaniline	13.828	65	210932	22.701	ng/ul	96
47) Dimethylphthalate	14.193	163	839310	21.544	ng/ul	99
48) 2,6-Dinitrotoluene	14.322	165	176287	22.086	ng/ul	98
50) Acenaphthylene	14.463	152	985471	21.569	ng/ul	100
51) 3-Nitroaniline	14.651	138	130141	19.751	ng/ul	96
52) Acenaphthene	14.804	153	676407	21.417	ng/ul	99
53) 2,4-Dinitrophenol	14.863	184	83303	17.350	ng/ul	96
55) 4-Nitrophenol	14.957	109	131727	20.211	ng/ul	95
56) Dibenzofuran	15.140	168	961183	21.472	ng/ul	100
57) 2,4-Dinitrotoluene	15.110	165	258468	22.017	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.363	232	214087	21.941	ng/ul	99
59) Diethylphthalate	15.551	149	859811	21.534	ng/ul	99
61) Fluorene	15.793	166	786860	21.563	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.781	204	398440	21.665	ng/ul	97
63) 4-Nitroaniline	15.816	138	120712	19.314	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.875	198	147589	20.248	ng/ul	96
67) N-Nitrosodiphenylamine	15.993	169	688259	21.505	ng/ul	100
68) 4-Bromophenyl-phenylether	16.675	248	258127	21.388	ng/ul	97
69) Hexachlorobenzene	16.798	284	304458	21.386	ng/ul	97
70) Atrazine	16.945	200	276459	21.632	ng/ul	98
71) Pentachlorophenol	17.145	266	160490	19.945	ng/ul	98
72) Phenanthrene	17.539	178	1288048	21.276	ng/ul	99
74) Anthracene	17.634	178	1305768	21.258	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.546	216	333015	21.389	ng/uL	98
76) Pentachlorobenzene	15.057	250	349062	21.413	ng/uL	98
77) Carbazole	17.898	167	1205093	21.789	ng/ul	99
78) Di-n-butylphthalate	18.428	149	1552967	22.130	ng/ul	99
80) Fluoranthene	19.492	202	1604727	22.052	ng/ul	99
82) Pyrene	19.851	202	1652467	21.949	ng/ul	98
83) Butylbenzylphthalate	20.698	149	762551	22.916	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.486	252	543324	20.603	ng/ul	99
85) Benzo(a)anthracene	21.563	228	1690982	21.670	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.457	149	1145455	23.289	ng/ul	99
87) Chrysene	21.616	228	1574671	21.349	ng/ul	99
89) Di-n-octyl phthalate	22.427	149	2010413	24.391	ng/ul	100
90) Benzo(b)fluoranthene	23.345	252	1734343	21.424	ng/ul	100
91) Benzo(k)fluoranthene	23.398	252	1707845	21.825	ng/ul	99
93) Benzo(a)pyrene	24.022	252	1482734	21.008	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.833	276	1776227	19.446	ng/ul	99
95) Dibenzo(a,h)anthracene	26.851	278	1466226	19.703	ng/ul	99
96) Benzo(g,h,i)perylene	27.668	276	1411356	18.749	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 08 00:14:32 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jun 07 23:46:15 2023
Response via : Initial Calibration





#2

1,4-Dioxane

Concen: 8.512 ng/uL

RT: 3.446 min Scan# 44

Delta R.T. -0.006 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36

Instrument :

BNA_P

ClientSampleId :

SSTD020737





Tgt Ion: 88 Resp: 36003

Ion Ratio Lower Upper

88 100

43 32.6 23.9 35.9

58 73.6 51.4 77.2



#5

Pyridine

Concen: 20.766 ng/uL

RT: 3.870 min Scan# 116

Delta R.T. -0.006 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36




Tgt Ion: 79 Resp: 223913


Ion Ratio Lower Upper

79 100

52 77.0 59.5 89.3

51 38.4 29.8 44.6









#6
Benzaldehyde
Concen: 20.721 ng/ul
RT: 7.228 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36




Instrument :
BNA_P
ClientSampleId :
SSTD020737


Tgt Ion: 77 Resp: 108188
Ion Ratio Lower Upper
77 100
105 94.3 77.3 115.9
106 88.0 73.3 109.9



#8
Phenol
Concen: 20.849 ng/ul
RT: 7.281 min Scan# 696
Delta R.T. 0.000 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36

Tgt Ion: 94 Resp: 293808
Ion Ratio Lower Upper
94 100
65 31.1 23.5 35.3
66 43.0 34.2 51.4









#10
Bis(2-Chloroethyl)ether
Concen: 21.682 ng/ul
RT: 7.511 min Scan# 71
Delta R.T. -0.006 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36




Instrument :
BNA_P
ClientSampleId :
SSTD020737


Tgt Ion: 93 Resp: 241580
Ion Ratio Lower Upper
93 100
63 77.5 62.6 93.8
95 33.0 25.5 38.3



#12
2-Chlorophenol
Concen: 21.781 ng/ul
RT: 7.658 min Scan# 760
Delta R.T. 0.000 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36

Tgt Ion:128 Resp: 227912
Ion Ratio Lower Upper
128 100
64 49.0 40.2 60.4
130 32.9 24.5 36.7









#13
2-Methylphenol
Concen: 21.325 ng/ul
RT: 8.546 min Scan# 911
Delta R.T. 0.000 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36




Instrument :
BNA_P
ClientSampleId :
SSTD020737


Tgt Ion:108 Resp: 231221
Ion Ratio Lower Upper
108 100
107 89.4 71.2 106.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 22.452 ng/ul
RT: 8.634 min Scan# 926
Delta R.T. 0.006 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36

Tgt Ion: 45 Resp: 331504
Ion Ratio Lower Upper
45 100
77 15.0 11.8 17.8
79 11.5 8.7 13.1









#16
Acetophenone
Concen: 22.094 ng/ul
RT: 8.934 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36




Instrument :
BNA_P
ClientSampleId :
SSTD020737


Tgt Ion:105 Resp: 395284
Ion Ratio Lower Upper
105 100
77 80.3 64.9 97.3
51 29.0 23.5 35.3



#17
N-Nitroso-di-n-propylamine
Concen: 22.382 ng/ul
RT: 8.911 min Scan# 973
Delta R.T. -0.006 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36

Tgt Ion: 70 Resp: 216132
Ion Ratio Lower Upper
70 100
42 59.7 46.6 69.8
101 9.1 7.8 11.6
130 20.0 15.9 23.9









#18
4-Methylphenol
Concen: 21.616 ng/ul
RT: 8.875 min Scan# 90
Delta R.T. -0.006 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36




Instrument :
BNA_P
ClientSampleId :
SSTD020737


Tgt Ion:108 Resp: 257871
Ion Ratio Lower Upper
108 100
107 114.0 89.3 133.9



#19
Hexachloroethane
Concen: 22.224 ng/ul
RT: 9.187 min Scan# 1020
Delta R.T. -0.006 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36

Tgt Ion:117 Resp: 98270
Ion Ratio Lower Upper
117 100
201 101.1 82.8 124.2
199 64.4 50.6 76.0









#22
Nitrobenzene
Concen: 21.474 ng/ul
RT: 9.316 min Scan# 1042
Delta R.T. -0.006 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36




Instrument :
BNA_P
ClientSampleId :
SSTD020737


Tgt Ion: 77 Resp: 309893
Ion Ratio Lower Upper
77 100
123 40.0 33.5 50.3
65 13.5 11.6 17.4



#23
Isophorone
Concen: 21.339 ng/ul
RT: 9.846 min Scan# 1132
Delta R.T. -0.006 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36

Tgt Ion: 82 Resp: 616435
Ion Ratio Lower Upper
82 100
95 7.9 5.8 8.8
138 16.4 13.1 19.7









#25
2-Nitrophenol
Concen: 21.793 ng/ul
RT: 10.034 min Scan# 1164
Delta R.T. -0.006 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36




Instrument :
BNA_P
ClientSampleId :
SSTD020737


Tgt Ion:139 Resp: 146217
Ion Ratio Lower Upper
139 100
65 59.0 46.9 70.3
109 38.5 30.2 45.2



#26
2,4-Dimethylphenol
Concen: 21.423 ng/ul
RT: 10.087 min Scan# 1173
Delta R.T. -0.006 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36

Tgt Ion:107 Resp: 304821
Ion Ratio Lower Upper
107 100
121 48.5 39.0 58.6
122 80.6 64.8 97.2





#27

Bis(2-Chloroethoxy)methane

Concen: 21.400 ng/ul

RT: 10.328 min Scan# 1214

Delta R.T. -0.006 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36

Instrument :

BNA_P

ClientSampleId :

SSTD020737





Tgt Ion: 93 Resp: 359409

Ion Ratio Lower Upper

93 100

95 31.3 25.7 38.5

123 10.2 8.6 13.0



#29

2,4-Dichlorophenol

Concen: 21.927 ng/ul

RT: 10.569 min Scan# 1255

Delta R.T. -0.006 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36




Tgt Ion:162 Resp: 247140


Ion Ratio Lower Upper

162 100

164 65.5 50.6 76.0

98 38.7 30.2 45.2









#30
Naphthalene
Concen: 21.201 ng/ul
RT: 10.975 min Scan# 1324
Delta R.T. -0.006 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36




Instrument :
BNA_P
ClientSampleId :
SSTD020737


Tgt Ion:128 Resp: 796296
Ion Ratio Lower Upper
128 100
129 10.7 8.6 13.0
127 13.3 10.5 15.7



#32
4-Chloroaniline
Concen: 21.441 ng/ul
RT: 11.093 min Scan# 1344
Delta R.T. 0.000 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36

Tgt Ion:127 Resp: 360982
Ion Ratio Lower Upper
127 100
129 32.3 26.1 39.1





#33

Hexachlorobutadiene

Concen: 21.575 ng/ul

RT: 11.257 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36

Instrument :

BNA_P

ClientSampleId :

SSTD020737





Tgt Ion:225 Resp: 163997

Ion Ratio Lower Upper

225 100

223 62.6 52.0 78.0

227 64.0 50.6 75.8



#34

Caprolactam

Concen: 21.053 ng/ul

RT: 11.863 min Scan# 1475

Delta R.T. -0.006 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36




Tgt Ion:113 Resp: 89322


Ion Ratio Lower Upper

113 100

55 162.6 127.9 191.9

56 124.8 101.1 151.7









#35
4-Chloro-3-methylphenol
Concen: 21.990 ng/ul
RT: 12.204 min Scan# 1533
Delta R.T. -0.006 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36




Instrument :
BNA_P
ClientSampleId :
SSTD020737


Tgt Ion:107 Resp: 300296
Ion Ratio Lower Upper
107 100
144 24.6 19.4 29.0
142 73.7 60.4 90.6



#36
2-Methylnaphthalene
Concen: 21.381 ng/ul
RT: 12.575 min Scan# 1596
Delta R.T. -0.006 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36

Tgt Ion:142 Resp: 564533
Ion Ratio Lower Upper
142 100
141 89.9 72.0 108.0





#37

1-Methylnaphthalene

Concen: 21.485 ng/ul

RT: 12.793 min Scan# 1633

Delta R.T. -0.006 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36

Instrument :

BNA_P

ClientSampleId :

SSTD020737





Tgt Ion:142 Resp: 571174

Ion Ratio Lower Upper

142 100

141 91.1 71.9 107.9

116 4.3 3.4 5.2



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.514 ng/ul

RT: 12.940 min Scan# 1658

Delta R.T. -0.006 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36

Tgt Ion:216 Resp: 320552




Ion Ratio Lower Upper


216 100

214 81.0 62.2 93.2

179 21.2 17.4 26.2

108 19.2 14.2 21.4









#40
Hexachlorocyclopentadiene
Concen: 20.227 ng/ul
RT: 12.916 min Scan# 1654
Delta R.T. -0.006 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36




Instrument :
BNA_P
ClientSampleId :
SSTD020737


Tgt Ion:237 Resp: 123540
Ion Ratio Lower Upper
237 100
235 62.4 48.2 72.2
272 12.0 10.3 15.5



#41
2,4,6-Trichlorophenol
Concen: 21.891 ng/ul
RT: 13.181 min Scan# 1699
Delta R.T. -0.006 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36

Tgt Ion:196 Resp: 217093
Ion Ratio Lower Upper
196 100
198 100.4 77.8 116.6
200 31.6 25.8 38.6





#42

2,4,5-Trichlorophenol

Concen: 21.930 ng/ul

RT: 13.257 min Scan# 1712

Delta R.T. 0.000 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36

Instrument :

BNA_P

ClientSampleId :

SSTD020737





Tgt Ion:196 Resp: 236985

Ion Ratio Lower Upper

196 100

198 95.5 75.5 113.3

200 30.8 24.7 37.1



#43

1,1'-Biphenyl

Concen: 21.262 ng/ul

RT: 13.575 min Scan# 1766

Delta R.T. -0.006 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36




Tgt Ion:154 Resp: 786768


Ion Ratio Lower Upper

154 100

153 40.2 32.2 48.2

76 13.9 10.4 15.6





#44

2-Chloronaphthalene

Concen: 21.350 ng/ul

RT: 13.622 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36

Instrument :

BNA_P

ClientSampleId :

SSTD020737





Tgt Ion:162 Resp: 619841

Ion Ratio Lower Upper

162 100

164 33.3 27.0 40.6

127 39.6 32.2 48.4



#45

2-Nitroaniline

Concen: 22.701 ng/ul

RT: 13.828 min Scan# 1809

Delta R.T. -0.006 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36




Tgt Ion: 65 Resp: 210932


Ion Ratio Lower Upper

65 100

92 65.5 55.4 83.2

138 95.2 78.8 118.2









#47
Dimethylphthalate
Concen: 21.544 ng/ul
RT: 14.193 min Scan# 1871
Delta R.T. -0.006 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36




Instrument :
BNA_P
ClientSampleId :
SSTD020737


Tgt Ion:163 Resp: 839310
Ion Ratio Lower Upper
163 100
194 3.9 3.2 4.8
164 10.6 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 22.086 ng/ul
RT: 14.322 min Scan# 1893
Delta R.T. 0.000 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36

Tgt Ion:165 Resp: 176287
Ion Ratio Lower Upper
165 100
89 62.0 51.5 77.3
121 21.3 17.3 25.9









#50
Acenaphthylene
Concen: 21.569 ng/ul
RT: 14.463 min Scan# 1917
Delta R.T. -0.006 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36




Instrument :
BNA_P
ClientSampleId :
SSTD020737


Tgt Ion:152 Resp: 985471
Ion Ratio Lower Upper
152 100
151 20.0 16.0 24.0
153 12.8 10.6 15.8



#51
3-Nitroaniline
Concen: 19.751 ng/ul
RT: 14.651 min Scan# 1949
Delta R.T. -0.006 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36

Tgt Ion:138 Resp: 130141
Ion Ratio Lower Upper
138 100
108 11.8 8.4 12.6
92 126.7 98.0 147.0





#52

Acenaphthene

Concen: 21.417 ng/ul

RT: 14.804 min Scan# 1975

Delta R.T. -0.006 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36

Instrument :

BNA_P

ClientSampleId :

SSTD020737





Tgt Ion:153 Resp: 676407

Ion Ratio Lower Upper

153 100

152 46.5 38.3 57.5

154 87.4 69.8 104.6



#53

2,4-Dinitrophenol

Concen: 17.350 ng/ul

RT: 14.863 min Scan# 1985

Delta R.T. -0.006 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36




Tgt Ion:184 Resp: 83303


Ion Ratio Lower Upper

184 100

63 75.5 59.0 88.4

154 67.9 57.6 86.4









#55
4-Nitrophenol
Concen: 20.211 ng/ul
RT: 14.957 min Scan# 2001
Delta R.T. 0.000 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36




Instrument :
BNA_P
ClientSampleId :
SSTD020737


Tgt Ion:109 Resp: 131727
Ion Ratio Lower Upper
109 100
139 108.8 89.1 133.7
65 120.0 103.1 154.7



#56
Dibenzofuran
Concen: 21.472 ng/ul
RT: 15.140 min Scan# 2032
Delta R.T. -0.006 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36

Tgt Ion:168 Resp: 961183
Ion Ratio Lower Upper
168 100
139 39.0 31.2 46.8









#57
2,4-Dinitrotoluene
Concen: 22.017 ng/ul
RT: 15.110 min Scan# 2027
Delta R.T. 0.000 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36




Instrument :
BNA_P
ClientSampleId :
SSTD020737


Tgt Ion:165 Resp: 258468
Ion Ratio Lower Upper
165 100
63 48.4 39.1 58.7
182 2.5 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.941 ng/ul
RT: 15.363 min Scan# 2070
Delta R.T. -0.006 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36

Tgt Ion:232 Resp: 214087
Ion Ratio Lower Upper
232 100
131 47.5 37.8 56.8
130 2.6 1.9 2.9
166 33.4 26.2 39.2









#59
Diethylphthalate
Concen: 21.534 ng/ul
RT: 15.551 min Scan# 2103
Delta R.T. -0.006 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36




Instrument :
BNA_P
ClientSampleId :
SSTD020737


Tgt Ion:149 Resp: 859811
Ion Ratio Lower Upper
149 100
177 22.2 17.8 26.8
150 12.1 10.1 15.1



#61
Fluorene
Concen: 21.563 ng/ul
RT: 15.793 min Scan# 2143
Delta R.T. 0.000 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36

Tgt Ion:166 Resp: 786860
Ion Ratio Lower Upper
166 100
165 97.2 79.1 118.7
167 12.7 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 21.665 ng/ul
RT: 15.781 min Scan# 2141
Delta R.T. 0.000 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36




Instrument :
BNA_P
ClientSampleId :
SSTD020737


Tgt Ion:204 Resp: 398440
Ion Ratio Lower Upper
204 100
206 32.1 26.6 40.0
141 59.4 49.8 74.8



#63
4-Nitroaniline
Concen: 19.314 ng/ul
RT: 15.816 min Scan# 2147
Delta R.T. 0.000 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36

Tgt Ion:138 Resp: 120712
Ion Ratio Lower Upper
138 100
92 57.1 44.6 67.0
108 101.4 78.4 117.6









#66
4,6-Dinitro-2-methylphenol
Concen: 20.248 ng/ul
RT: 15.875 min Scan# 2158
Delta R.T. -0.006 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36




Instrument :
BNA_P
ClientSampleId :
SSTD020737


Tgt Ion:198 Resp: 147589
Ion Ratio Lower Upper
198 100
182 4.2 3.0 4.4
77 27.7 20.5 30.7



#67
N-Nitrosodiphenylamine
Concen: 21.505 ng/ul
RT: 15.993 min Scan# 2177
Delta R.T. -0.006 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36

Tgt Ion:169 Resp: 688259
Ion Ratio Lower Upper
169 100
168 66.6 53.0 79.4
167 36.4 28.9 43.3









#68
4-Bromophenyl-phenylether
Concen: 21.388 ng/ul
RT: 16.675 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36




Instrument :
BNA_P
ClientSampleId :
SSTD020737


Tgt Ion:248 Resp: 258127
Ion Ratio Lower Upper
248 100
250 98.9 78.3 117.5
141 81.4 61.1 91.7



#69
Hexachlorobenzene
Concen: 21.386 ng/ul
RT: 16.798 min Scan# 2314
Delta R.T. -0.006 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36

Tgt Ion:284 Resp: 304458
Ion Ratio Lower Upper
284 100
142 34.6 25.4 38.2
249 31.9 24.8 37.2





#70

Atrazine

Concen: 21.632 ng/ul

RT: 16.945 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36

Instrument :

BNA_P

ClientSampleId :

SSTD020737





Tgt Ion:200 Resp: 276459

Ion Ratio Lower Upper

200 100

173 26.4 19.8 29.6

215 45.8 37.4 56.0



#71

Pentachlorophenol

Concen: 19.945 ng/ul

RT: 17.145 min Scan# 2373

Delta R.T. -0.006 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36




Tgt Ion:266 Resp: 160490


Ion Ratio Lower Upper

266 100

264 63.3 49.6 74.4

268 62.6 51.0 76.6





#72

Phenanthrene

Concen: 21.276 ng/ul

RT: 17.539 min Scan# 2440

Delta R.T. -0.006 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36

Instrument :

BNA_P

ClientSampleId :

SSTD020737





Tgt Ion:178 Resp: 1288048

Ion Ratio Lower Upper

178 100

179 15.1 12.3 18.5

176 19.3 15.4 23.0



#74

Anthracene

Concen: 21.258 ng/ul

RT: 17.634 min Scan# 2456

Delta R.T. 0.000 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36




Tgt Ion:178 Resp: 1305768


Ion Ratio Lower Upper

178 100

179 15.5 12.4 18.6

176 18.8 15.0 22.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 21.389 ng/uL
RT: 13.546 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36




Instrument :
BNA_P
ClientSampleId :
SSTD020737


Tgt Ion:216 Resp: 333015
Ion Ratio Lower Upper
216 100
214 79.5 62.1 93.1
218 48.2 39.0 58.4



#76
Pentachlorobenzene
Concen: 21.413 ng/uL
RT: 15.057 min Scan# 2018
Delta R.T. -0.006 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36

Tgt Ion:250 Resp: 349062
Ion Ratio Lower Upper
250 100
248 63.2 49.0 73.4
252 62.9 49.9 74.9





#77

Carbazole

Concen: 21.789 ng/ul

RT: 17.898 min Scan# 2501

Delta R.T. -0.006 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36

Instrument :

BNA_P

ClientSampleId :

SSTD020737





Tgt Ion:167 Resp: 1205093

Ion Ratio Lower Upper

167 100

166 21.1 17.2 25.8

139 13.0 10.6 16.0



#78

Di-n-butylphthalate

Concen: 22.130 ng/ul

RT: 18.428 min Scan# 2591

Delta R.T. -0.006 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36




Tgt Ion:149 Resp: 1552967


Ion Ratio Lower Upper

149 100

150 9.3 7.1 10.7

104 5.8 4.5 6.7









#80
Fluoranthene
Concen: 22.052 ng/ul
RT: 19.492 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36




Instrument :
BNA_P
ClientSampleId :
SSTD020737


Tgt Ion:202 Resp: 1604727
Ion Ratio Lower Upper
202 100
101 11.0 8.6 13.0
100 8.4 6.6 9.8



#82
Pyrene
Concen: 21.949 ng/ul
RT: 19.851 min Scan# 2833
Delta R.T. 0.000 min
Lab File: BP015412.D
Acq: 07 Jun 2023 22:36

Tgt Ion:202 Resp: 1652467
Ion Ratio Lower Upper
202 100
101 11.6 9.9 14.9
100 9.6 8.4 12.6





#83

Butylbenzylphthalate

Concen: 22.916 ng/ul

RT: 20.698 min Scan# 2978

Delta R.T. -0.006 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36

Instrument :

BNA_P

ClientSampleId :

SSTD020737





Tgt Ion:149 Resp: 762551

Ion Ratio Lower Upper

149 100

91 76.1 60.4 90.6

206 20.3 16.4 24.6



#84

3,3'-Dichlorobenzidine

Concen: 20.603 ng/ul

RT: 21.486 min Scan# 3111

Delta R.T. -0.006 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36




Tgt Ion:252 Resp: 543324


Ion Ratio Lower Upper

252 100

254 65.2 51.6 77.4

126 11.0 8.5 12.7





#85

Benzo(a)anthracene

Concen: 21.670 ng/ul

RT: 21.563 min Scan# 31

Delta R.T. -0.006 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36

Instrument :

BNA_P

ClientSampleId :

SSTD020737





Tgt Ion:228 Resp: 1690982

Ion Ratio Lower Upper

228 100

229 19.5 15.8 23.6

226 26.5 21.8 32.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 23.289 ng/ul

RT: 21.457 min Scan# 3106

Delta R.T. -0.006 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36




Tgt Ion:149 Resp: 1145455


Ion Ratio Lower Upper

149 100

167 26.7 21.8 32.8

279 4.3 3.7 5.5





#87

Chrysene

Concen: 21.349 ng/ul

RT: 21.616 min Scan# 31

Delta R.T. -0.006 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36

Instrument :

BNA_P

ClientSampleId :

SSTD020737





Tgt Ion:228 Resp: 1574671

Ion Ratio Lower Upper

228 100

226 29.9 23.9 35.9

229 19.5 15.1 22.7



#89

Di-n-octyl phthalate

Concen: 24.391 ng/ul




RT: 22.427 min Scan# 3271


Delta R.T. -0.006 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36

Tgt Ion:149 Resp: 2010413





#90

Benzo(b)fluoranthene

Concen: 21.424 ng/ul

RT: 23.345 min Scan# 3427

Delta R.T. -0.006 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36

Instrument :

BNA_P

ClientSampleId :

SSTD020737





Tgt Ion:252 Resp: 1734343

Ion Ratio Lower Upper

252 100

253 21.6 17.4 26.2

125 9.1 7.3 10.9



#91

Benzo(k)fluoranthene

Concen: 21.825 ng/ul

RT: 23.398 min Scan# 3436

Delta R.T. -0.006 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36




Tgt Ion:252 Resp: 1707845


Ion Ratio Lower Upper

252 100

253 21.6 17.7 26.5

125 8.9 7.3 10.9





#93

Benzo(a)pyrene

Concen: 21.008 ng/ul

RT: 24.022 min Scan# 3

Delta R.T. -0.006 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36

Instrument :

BNA_P

ClientSampleId :

SSTD020737





Tgt Ion:252 Resp: 1482734

Ion Ratio Lower Upper

252 100

253 21.8 17.8 26.6

125 10.1 8.1 12.1



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.446 ng/ul

RT: 26.833 min Scan# 4020

Delta R.T. -0.012 min

Lab File: BP015412.D

Acq: 07 Jun 2023 22:36




Tgt Ion:276 Resp: 1776227


Ion Ratio Lower Upper

276 100

138 21.9 17.3 25.9

277 24.6 20.2 30.2









#95
 Dibenzo(a,h)anthracene
 Concen: 19.703 ng/ul
 RT: 26.851 min Scan# 4025
 Delta R.T. -0.012 min
 Lab File: BP015412.D
 Acq: 07 Jun 2023 22:36

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020737

Tgt Ion:278 Resp: 1466226
 Ion Ratio Lower Upper
 278 100
 139 14.7 11.4 17.0
 279 23.6 18.7 28.1



#96
 Benzo(g,h,i)perylene
 Concen: 18.749 ng/ul
 RT: 27.668 min Scan# 4162
 Delta R.T. -0.018 min
 Lab File: BP015412.D
 Acq: 07 Jun 2023 22:36

Tgt Ion:276 Resp: 1411356
 Ion Ratio Lower Upper
 276 100
 138 19.6 15.6 23.4
 277 23.7 19.0 28.6

