

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060823\
 Data File : BP015426.D
 Acq On : 08 Jun 2023 09:07
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020739

Quant Time: Jun 09 01:14:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jun 07 23:46:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.099	152	131024	20.000	ng/u1	0.00
20) Naphthalene-d8	10.928	136	606989	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.740	164	423713	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.498	188	1017299	20.000	ng/u1	0.00
79) Chrysene-d12	21.580	240	1043758	20.000	ng/u1	0.00
88) Perylene-d12	24.133	264	1176300	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.411	96	30216	8.540	ng/uL	0.00
4) Pyridine-d5	3.846	84	192926	20.991	ng/u1	0.00
7) Phenol-d5	7.246	99	247205	20.480	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.416	67	155088	21.513	ng/u1	0.00
11) 2-Chlorophenol-d4	7.622	132	188812	21.575	ng/u1	0.00
15) 4-Methylphenol-d8	8.810	113	211608	21.360	ng/u1	0.00
21) Nitrobenzene-d5	9.275	128	102331	21.473	ng/u1	0.00
24) 2-Nitrophenol-d4	10.004	143	117229	21.541	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.540	165	218205	21.980	ng/u1	-0.01
31) 4-Chloroaniline-d4	11.063	131	310314	21.777	ng/u1	0.00
46) Dimethylphthalate-d6	14.145	166	718917	21.517	ng/u1	0.00
49) Acenaphthylene-d8	14.434	160	790033	21.624	ng/u1	0.00
54) 4-Nitrophenol-d4	14.940	143	118661	19.832	ng/u1	0.00
60) Fluorene-d10	15.734	176	612983	21.756	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.857	200	123260	19.126	ng/u1	0.00
73) Anthracene-d10	17.598	188	995558	21.509	ng/u1	0.00
81) Pyrene-d10	19.822	212	1200700	21.493	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.968	264	1240903	21.024	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.446	88	30619	8.192	ng/uL	87
5) Pyridine	3.869	79	199859	20.975	ng/u1	97
6) Benzaldehyde	7.228	77	89012	19.293	ng/u1	100
8) Phenol	7.275	94	262464	21.076	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.510	93	212054	21.537	ng/u1	99
12) 2-Chlorophenol	7.652	128	200995	21.737	ng/u1	99
13) 2-Methylphenol	8.546	108	203611	21.250	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.634	45	284225	21.784	ng/u1	97
16) Acetophenone	8.934	105	349374	22.098	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.910	70	187800	22.008	ng/u1	97
18) 4-Methylphenol	8.875	108	228117	21.639	ng/u1	98
19) Hexachloroethane	9.187	117	85698	21.932	ng/u1	97
22) Nitrobenzene	9.316	77	275120	21.770	ng/u1	99
23) Isophorone	9.846	82	537633	21.252	ng/u1	100
25) 2-Nitrophenol	10.034	139	128723	21.907	ng/u1	99
26) 2,4-Dimethylphenol	10.087	107	273440	21.944	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.328	93	312537	21.249	ng/u1	98
29) 2,4-Dichlorophenol	10.569	162	217531	22.039	ng/u1	99
30) Naphthalene	10.975	128	702166	21.348	ng/u1	100
32) 4-Chloroaniline	11.093	127	324365	21.999	ng/u1	99
33) Hexachlorobutadiene	11.257	225	146058	21.942	ng/u1	94
34) Caprolactam	11.863	113	77464	20.848	ng/u1	90
35) 4-Chloro-3-methylphenol	12.204	107	266839	22.312	ng/u1	100
36) 2-Methylnaphthalene	12.575	142	500612	21.650	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060823\
 Data File : BP015426.D
 Acq On : 08 Jun 2023 09:07
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020739


Quant Time: Jun 09 01:14:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jun 07 23:46:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.793	142	504983	21.691	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.940	216	288937	21.899	ng/ul	98
40) Hexachlorocyclopentadiene	12.916	237	118100	21.837	ng/ul	98
41) 2,4,6-Trichlorophenol	13.181	196	194685	22.170	ng/ul	97
42) 2,4,5-Trichlorophenol	13.251	196	212989	22.257	ng/ul	97
43) 1,1'-Biphenyl	13.575	154	709742	21.660	ng/ul	100
44) 2-Chloronaphthalene	13.622	162	550311	21.406	ng/ul	99
45) 2-Nitroaniline	13.828	65	183725	22.329	ng/ul	97
47) Dimethylphthalate	14.192	163	744380	21.577	ng/ul	99
48) 2,6-Dinitrotoluene	14.322	165	154766	21.897	ng/ul	99
50) Acenaphthylene	14.463	152	877554	21.690	ng/ul	99
51) 3-Nitroaniline	14.651	138	115700	19.830	ng/ul	99
52) Acenaphthene	14.804	153	602416	21.540	ng/ul	99
53) 2,4-Dinitrophenol	14.863	184	67441	15.862	ng/ul	96
55) 4-Nitrophenol	14.957	109	119167	20.648	ng/ul	97
56) Dibenzofuran	15.140	168	861720	21.739	ng/ul	97
57) 2,4-Dinitrotoluene	15.110	165	229735	22.100	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.363	232	192575	22.288	ng/ul	99
59) Diethylphthalate	15.551	149	778149	22.009	ng/ul	99
61) Fluorene	15.792	166	708943	21.940	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.775	204	361504	22.198	ng/ul	100
63) 4-Nitroaniline	15.816	138	107439	19.413	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.875	198	128523	19.553	ng/ul	94
67) N-Nitrosodiphenylamine	15.992	169	618369	21.426	ng/ul	100
68) 4-Bromophenyl-phenylether	16.675	248	234318	21.530	ng/ul	98
69) Hexachlorobenzene	16.798	284	278786	21.716	ng/ul	98
70) Atrazine	16.945	200	248346	21.549	ng/ul	99
71) Pentachlorophenol	17.145	266	147590	20.340	ng/ul	97
72) Phenanthrene	17.539	178	1169754	21.427	ng/ul	100
74) Anthracene	17.633	178	1185058	21.394	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.545	216	297671	21.201	ng/ul	99
76) Pentachlorobenzene	15.057	250	312918	21.287	ng/ul	97
77) Carbazole	17.898	167	1087617	21.807	ng/ul	100
78) Di-n-butylphthalate	18.428	149	1399270	22.112	ng/ul	99
80) Fluoranthene	19.492	202	1460288	21.460	ng/ul	99
82) Pyrene	19.851	202	1517364	21.553	ng/ul	97
83) Butylbenzylphthalate	20.704	149	693098	22.275	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.486	252	509889	20.678	ng/ul	98
85) Benzo(a)anthracene	21.563	228	1569340	21.508	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.457	149	1045657	22.736	ng/ul	99
87) Chrysene	21.622	228	1484187	21.520	ng/ul	99
89) Di-n-octyl phthalate	22.427	149	1825980	23.619	ng/ul	100
90) Benzo(b)fluoranthene	23.345	252	1682442	22.158	ng/ul	99
91) Benzo(k)fluoranthene	23.398	252	1558694	21.236	ng/ul	99
93) Benzo(a)pyrene	24.021	252	1394116	21.059	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.833	276	1663062	19.411	ng/ul	99
95) Dibenzo(a,h)anthracene	26.851	278	1379811	19.768	ng/ul	99
96) Benzo(g,h,i)perylene	27.668	276	1331626	18.860	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 09 01:14:05 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jun 07 23:46:15 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.192 ng/uL
RT: 3.446 min Scan# 44
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion: 88 Resp: 30619
Ion Ratio Lower Upper
88 100
43 33.0 23.9 35.9
58 76.9 51.4 77.2



#5
Pyridine
Concen: 20.975 ng/uL
RT: 3.869 min Scan# 116
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07

Tgt Ion: 79 Resp: 199859
Ion Ratio Lower Upper
79 100
52 77.4 59.5 89.3
51 37.7 29.8 44.6









#6
Benzaldehyde
Concen: 19.293 ng/ul
RT: 7.228 min Scan# 612
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:	77	Resp:	89012
Ion	Ratio	Lower	Upper
77	100		
105	96.3	77.3	115.9
106	91.9	73.3	109.9



#8
Phenol
Concen: 21.076 ng/ul
RT: 7.275 min Scan# 695
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07

Tgt Ion:	94	Resp:	262464
Ion	Ratio	Lower	Upper
94	100		
65	31.4	23.5	35.3
66	44.4	34.2	51.4









#10
Bis(2-Chloroethyl)ether
Concen: 21.537 ng/ul
RT: 7.510 min Scan# 735
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion: 93 Resp: 212054
Ion Ratio Lower Upper
93 100
63 76.8 62.6 93.8
95 31.5 25.5 38.3



#12
2-Chlorophenol
Concen: 21.737 ng/ul
RT: 7.652 min Scan# 759
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07

Tgt Ion:128 Resp: 200995
Ion Ratio Lower Upper
128 100
64 50.3 40.2 60.4
130 29.5 24.5 36.7









#13
2-Methylphenol
Concen: 21.250 ng/ul
RT: 8.546 min Scan# 911
Delta R.T. -0.000 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:108 Resp: 203611
Ion Ratio Lower Upper
108 100
107 90.8 71.2 106.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 21.784 ng/ul
RT: 8.634 min Scan# 926
Delta R.T. 0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07

Tgt Ion: 45 Resp: 284225
Ion Ratio Lower Upper
45 100
77 16.0 11.8 17.8
79 12.3 8.7 13.1









#16
Acetophenone
Concen: 22.098 ng/ul
RT: 8.934 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:105 Resp: 349374
Ion Ratio Lower Upper
105 100
77 80.4 64.9 97.3
51 29.2 23.5 35.3



#17
N-Nitroso-di-n-propylamine
Concen: 22.008 ng/ul
RT: 8.910 min Scan# 973
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07

Tgt Ion: 70 Resp: 187800
Ion Ratio Lower Upper
70 100
42 61.5 46.6 69.8
101 9.1 7.8 11.6
130 19.5 15.9 23.9









#18
4-Methylphenol
Concen: 21.639 ng/ul
RT: 8.875 min Scan# 90
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:108 Resp: 228117
Ion Ratio Lower Upper
108 100
107 109.5 89.3 133.9



#19
Hexachloroethane
Concen: 21.932 ng/ul
RT: 9.187 min Scan# 1020
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07

Tgt Ion:117 Resp: 85698
Ion Ratio Lower Upper
117 100
201 99.2 82.8 124.2
199 64.9 50.6 76.0









#22
Nitrobenzene
Concen: 21.770 ng/ul
RT: 9.316 min Scan# 1042
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion: 77 Resp: 275120
Ion Ratio Lower Upper
77 100
123 41.3 33.5 50.3
65 13.2 11.6 17.4



#23
Isophorone
Concen: 21.252 ng/ul
RT: 9.846 min Scan# 1132
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07

Tgt Ion: 82 Resp: 537633
Ion Ratio Lower Upper
82 100
95 7.4 5.8 8.8
138 16.6 13.1 19.7









#25
2-Nitrophenol
Concen: 21.907 ng/ul
RT: 10.034 min Scan# 1164
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:139 Resp: 128723
Ion Ratio Lower Upper
139 100
65 58.2 46.9 70.3
109 37.1 30.2 45.2



#26
2,4-Dimethylphenol
Concen: 21.944 ng/ul
RT: 10.087 min Scan# 1173
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07

Tgt Ion:107 Resp: 273440
Ion Ratio Lower Upper
107 100
121 47.9 39.0 58.6
122 77.6 64.8 97.2









#27
 Bis(2-Chloroethoxy)methane
 Concen: 21.249 ng/ul
 RT: 10.328 min Scan# 1214
 Delta R.T. -0.006 min
 Lab File: BP015426.D
 Acq: 08 Jun 2023 09:07




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020739


Tgt Ion: 93 Resp: 312537
 Ion Ratio Lower Upper
 93 100
 95 33.9 25.7 38.5
 123 10.8 8.6 13.0



#29
 2,4-Dichlorophenol
 Concen: 22.039 ng/ul
 RT: 10.569 min Scan# 1255
 Delta R.T. -0.006 min
 Lab File: BP015426.D
 Acq: 08 Jun 2023 09:07

Tgt Ion: 162 Resp: 217531
 Ion Ratio Lower Upper
 162 100
 164 64.3 50.6 76.0
 98 38.7 30.2 45.2









#30
Naphthalene
Concen: 21.348 ng/ul
RT: 10.975 min Scan# 1324
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:128 Resp: 702166
Ion Ratio Lower Upper
128 100
129 10.8 8.6 13.0
127 13.0 10.5 15.7



#32
4-Chloroaniline
Concen: 21.999 ng/ul
RT: 11.093 min Scan# 1344
Delta R.T. -0.000 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07

Tgt Ion:127 Resp: 324365
Ion Ratio Lower Upper
127 100
129 32.2 26.1 39.1









#33
Hexachlorobutadiene
Concen: 21.942 ng/ul
RT: 11.257 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:225 Resp: 146058
Ion Ratio Lower Upper
225 100
223 59.5 52.0 78.0
227 59.0 50.6 75.8



#34
Caprolactam
Concen: 20.848 ng/ul
RT: 11.863 min Scan# 1475
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07

Tgt Ion:113 Resp: 77464
Ion Ratio Lower Upper
113 100
55 174.5 127.9 191.9
56 135.6 101.1 151.7









#35
4-Chloro-3-methylphenol
Concen: 22.312 ng/ul
RT: 12.204 min Scan# 1533
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:107 Resp: 266839
Ion Ratio Lower Upper
107 100
144 24.8 19.4 29.0
142 75.3 60.4 90.6



#36
2-Methylnaphthalene
Concen: 21.650 ng/ul
RT: 12.575 min Scan# 1596
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07

Tgt Ion:142 Resp: 500612
Ion Ratio Lower Upper
142 100
141 88.9 72.0 108.0





#37

1-Methylnaphthalene

Concen: 21.691 ng/ul

RT: 12.793 min Scan# 1633

Delta R.T. -0.006 min

Lab File: BP015426.D

Acq: 08 Jun 2023 09:07

Instrument :

BNA_P

ClientSampleId :

SSTD020739





Tgt Ion:142 Resp: 504983

Ion Ratio Lower Upper

142 100

141 91.1 71.9 107.9

116 4.4 3.4 5.2



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.899 ng/ul

RT: 12.940 min Scan# 1658

Delta R.T. -0.006 min

Lab File: BP015426.D

Acq: 08 Jun 2023 09:07

Tgt Ion:216 Resp: 288937




Ion Ratio Lower Upper


216 100

214 75.6 62.2 93.2

179 20.5 17.4 26.2

108 16.9 14.2 21.4









#40
Hexachlorocyclopentadiene
Concen: 21.837 ng/ul
RT: 12.916 min Scan# 1654
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:237 Resp: 118100
Ion Ratio Lower Upper
237 100
235 62.4 48.2 72.2
272 13.1 10.3 15.5



#41
2,4,6-Trichlorophenol
Concen: 22.170 ng/ul
RT: 13.181 min Scan# 1699
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07

Tgt Ion:196 Resp: 194685
Ion Ratio Lower Upper
196 100
198 94.1 77.8 116.6
200 31.8 25.8 38.6





#42

2,4,5-Trichlorophenol

Concen: 22.257 ng/ul

RT: 13.251 min Scan# 1711

Delta R.T. -0.006 min

Lab File: BP015426.D

Acq: 08 Jun 2023 09:07

Instrument :

BNA_P

ClientSampleId :

SSTD020739





Tgt Ion:196 Resp: 212989

Ion Ratio Lower Upper

196 100

198 98.2 75.5 113.3

200 30.1 24.7 37.1



#43

1,1'-Biphenyl

Concen: 21.660 ng/ul

RT: 13.575 min Scan# 1766

Delta R.T. -0.006 min

Lab File: BP015426.D

Acq: 08 Jun 2023 09:07




Tgt Ion:154 Resp: 709742


Ion Ratio Lower Upper

154 100

153 40.1 32.2 48.2

76 13.3 10.4 15.6









#44
2-Chloronaphthalene
Concen: 21.406 ng/ul
RT: 13.622 min Scan# 1774
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:162 Resp: 550311
Ion Ratio Lower Upper
162 100
164 32.7 27.0 40.6
127 40.8 32.2 48.4



#45
2-Nitroaniline
Concen: 22.329 ng/ul
RT: 13.828 min Scan# 1809
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07

Tgt Ion: 65 Resp: 183725
Ion Ratio Lower Upper
65 100
92 66.0 55.4 83.2
138 96.5 78.8 118.2









#47
Dimethylphthalate
Concen: 21.577 ng/ul
RT: 14.192 min Scan# 1871
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:163 Resp: 744380
Ion Ratio Lower Upper
163 100
194 3.9 3.2 4.8
164 10.0 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 21.897 ng/ul
RT: 14.322 min Scan# 1893
Delta R.T. -0.000 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07

Tgt Ion:165 Resp: 154766
Ion Ratio Lower Upper
165 100
89 65.2 51.5 77.3
121 21.3 17.3 25.9





#50

Acenaphthylene

Concen: 21.690 ng/ul

RT: 14.463 min Scan# 1917

Delta R.T. -0.006 min

Lab File: BP015426.D

Acq: 08 Jun 2023 09:07

Instrument :

BNA_P

ClientSampleId :

SSTD020739





Tgt Ion:152 Resp: 877554

Ion Ratio Lower Upper

152 100

151 19.8 16.0 24.0

153 12.9 10.6 15.8



#51

3-Nitroaniline

Concen: 19.830 ng/ul

RT: 14.651 min Scan# 1949

Delta R.T. -0.006 min

Lab File: BP015426.D

Acq: 08 Jun 2023 09:07




Tgt Ion:138 Resp: 115700


Ion Ratio Lower Upper

138 100

108 11.5 8.4 12.6

92 123.8 98.0 147.0





#52

Acenaphthene

Concen: 21.540 ng/ul

RT: 14.804 min Scan# 1975

Delta R.T. -0.006 min

Lab File: BP015426.D

Acq: 08 Jun 2023 09:07

Instrument :

BNA_P

ClientSampleId :

SSTD020739





Tgt Ion:153 Resp: 602416

Ion Ratio Lower Upper

153 100

152 46.6 38.3 57.5

154 88.2 69.8 104.6



#53

2,4-Dinitrophenol

Concen: 15.862 ng/ul

RT: 14.863 min Scan# 1985

Delta R.T. -0.006 min

Lab File: BP015426.D

Acq: 08 Jun 2023 09:07




Tgt Ion:184 Resp: 67441


Ion Ratio Lower Upper

184 100

63 79.1 59.0 88.4

154 71.4 57.6 86.4









#55
4-Nitrophenol
Concen: 20.648 ng/ul
RT: 14.957 min Scan# 2001
Delta R.T. -0.000 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:109 Resp: 119167
Ion Ratio Lower Upper
109 100
139 110.1 89.1 133.7
65 123.0 103.1 154.7



#56
Dibenzofuran
Concen: 21.739 ng/ul
RT: 15.140 min Scan# 2032
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07

Tgt Ion:168 Resp: 861720
Ion Ratio Lower Upper
168 100
139 40.6 31.2 46.8





#57

2,4-Dinitrotoluene

Concen: 22.100 ng/ul

RT: 15.110 min Scan# 2027

Delta R.T. -0.000 min

Lab File: BP015426.D

Acq: 08 Jun 2023 09:07

Instrument :

BNA_P

ClientSampleId :

SSTD020739





Tgt Ion:165 Resp: 229735

Ion Ratio Lower Upper

165 100

63 48.7 39.1 58.7

182 2.6 2.2 3.4



#58

2,3,4,6-Tetrachlorophenol

Concen: 22.288 ng/ul

RT: 15.363 min Scan# 2070

Delta R.T. -0.006 min

Lab File: BP015426.D

Acq: 08 Jun 2023 09:07

Tgt Ion:232 Resp: 192575




Ion Ratio Lower Upper


232 100

131 48.6 37.8 56.8

130 2.8 1.9 2.9

166 33.0 26.2 39.2









#59
Diethylphthalate
Concen: 22.009 ng/ul
RT: 15.551 min Scan# 2102
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:149 Resp: 778149
Ion Ratio Lower Upper
149 100
177 22.0 17.8 26.8
150 12.7 10.1 15.1



#61
Fluorene
Concen: 21.940 ng/ul
RT: 15.792 min Scan# 2143
Delta R.T. -0.000 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07

Tgt Ion:166 Resp: 708943
Ion Ratio Lower Upper
166 100
165 98.1 79.1 118.7
167 13.5 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 22.198 ng/ul
RT: 15.775 min Scan# 2140
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:204 Resp: 361504
Ion Ratio Lower Upper
204 100
206 33.1 26.6 40.0
141 62.1 49.8 74.8



#63
4-Nitroaniline
Concen: 19.413 ng/ul
RT: 15.816 min Scan# 2147
Delta R.T. -0.000 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07

Tgt Ion:138 Resp: 107439
Ion Ratio Lower Upper
138 100
92 55.2 44.6 67.0
108 101.8 78.4 117.6









#66
 4,6-Dinitro-2-methylphenol
 Concen: 19.553 ng/ul
 RT: 15.875 min Scan# 2157
 Delta R.T. -0.006 min
 Lab File: BP015426.D
 Acq: 08 Jun 2023 09:07




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020739


Tgt Ion:198 Resp: 128523
 Ion Ratio Lower Upper
 198 100
 182 4.0 3.0 4.4
 77 29.3 20.5 30.7



#67
 N-Nitrosodiphenylamine
 Concen: 21.426 ng/ul
 RT: 15.992 min Scan# 2177
 Delta R.T. -0.006 min
 Lab File: BP015426.D
 Acq: 08 Jun 2023 09:07

Tgt Ion:169 Resp: 618369
 Ion Ratio Lower Upper
 169 100
 168 66.6 53.0 79.4
 167 36.0 28.9 43.3









#68
4-Bromophenyl-phenylether
Concen: 21.530 ng/ul
RT: 16.675 min Scan# 2293
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:248 Resp: 234318
Ion Ratio Lower Upper
248 100
250 96.5 78.3 117.5
141 79.6 61.1 91.7



#69
Hexachlorobenzene
Concen: 21.716 ng/ul
RT: 16.798 min Scan# 2314
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07

Tgt Ion:284 Resp: 278786
Ion Ratio Lower Upper
284 100
142 33.5 25.4 38.2
249 30.5 24.8 37.2





#70

Atrazine

Concen: 21.549 ng/ul

RT: 16.945 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP015426.D

Acq: 08 Jun 2023 09:07

Instrument :

BNA_P

ClientSampleId :

SSTD020739





Tgt Ion:200 Resp: 248346

Ion Ratio Lower Upper

200 100

173 25.5 19.8 29.6

215 45.7 37.4 56.0



#71

Pentachlorophenol

Concen: 20.340 ng/ul

RT: 17.145 min Scan# 2373

Delta R.T. -0.006 min

Lab File: BP015426.D

Acq: 08 Jun 2023 09:07




Tgt Ion:266 Resp: 147590


Ion Ratio Lower Upper

266 100

264 66.1 49.6 74.4

268 64.6 51.0 76.6





#72

Phenanthrene

Concen: 21.427 ng/ul

RT: 17.539 min Scan# 2440

Delta R.T. -0.006 min

Lab File: BP015426.D

Acq: 08 Jun 2023 09:07

Instrument :

BNA_P

ClientSampleId :

SSTD020739





Tgt Ion:178 Resp: 1169754

Ion Ratio Lower Upper

178 100

179 15.6 12.3 18.5

176 19.4 15.4 23.0



#74

Anthracene

Concen: 21.394 ng/ul

RT: 17.633 min Scan# 2456

Delta R.T. -0.000 min

Lab File: BP015426.D

Acq: 08 Jun 2023 09:07




Tgt Ion:178 Resp: 1185058


Ion Ratio Lower Upper

178 100

179 15.5 12.4 18.6

176 19.0 15.0 22.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 21.201 ng/uL
RT: 13.545 min Scan# 1761
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:216 Resp: 297671
Ion Ratio Lower Upper
216 100
214 78.6 62.1 93.1
218 47.9 39.0 58.4



#76
Pentachlorobenzene
Concen: 21.287 ng/uL
RT: 15.057 min Scan# 2018
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07

Tgt Ion:250 Resp: 312918
Ion Ratio Lower Upper
250 100
248 64.2 49.0 73.4
252 64.8 49.9 74.9





#77

Carbazole

Concen: 21.807 ng/ul

RT: 17.898 min Scan# 2501

Delta R.T. -0.006 min

Lab File: BP015426.D

Acq: 08 Jun 2023 09:07

Instrument :





BNA_P

ClientSampleId :

SSTD020739

Tgt Ion:167 Resp: 1087617

Ion	Ratio	Lower	Upper
167	100		
166	21.3	17.2	25.8
139	13.4	10.6	16.0



#78

Di-n-butylphthalate

Concen: 22.112 ng/ul

RT: 18.428 min Scan# 2591




Delta R.T. -0.006 min


Lab File: BP015426.D

Acq: 08 Jun 2023 09:07

Tgt Ion:149 Resp: 1399270

Ion	Ratio	Lower	Upper
149	100		
150	9.4	7.1	10.7
104	5.9	4.5	6.7





#80

Fluoranthene

Concen: 21.460 ng/ul

RT: 19.492 min Scan# 2772

Delta R.T. -0.006 min

Lab File: BP015426.D

Acq: 08 Jun 2023 09:07

Instrument :

BNA_P

ClientSampleId :

SSTD020739





Tgt Ion:202 Resp: 1460288

Ion Ratio Lower Upper

202 100

101 10.4 8.6 13.0

100 7.9 6.6 9.8



#82

Pyrene

Concen: 21.553 ng/ul

RT: 19.851 min Scan# 2833

Delta R.T. -0.000 min

Lab File: BP015426.D

Acq: 08 Jun 2023 09:07




Tgt Ion:202 Resp: 1517364


Ion Ratio Lower Upper

202 100

101 11.3 9.9 14.9

100 9.6 8.4 12.6





#83

Butylbenzylphthalate

Concen: 22.275 ng/ul

RT: 20.704 min Scan# 2978

Delta R.T. -0.000 min

Lab File: BP015426.D

Acq: 08 Jun 2023 09:07

Instrument :

BNA_P

ClientSampleId :

SSTD020739





Tgt Ion:149 Resp: 693098

Ion Ratio Lower Upper

149 100

91 73.8 60.4 90.6

206 20.9 16.4 24.6



#84

3,3'-Dichlorobenzidine

Concen: 20.678 ng/ul

RT: 21.486 min Scan# 3111

Delta R.T. -0.006 min

Lab File: BP015426.D

Acq: 08 Jun 2023 09:07




Tgt Ion:252 Resp: 509889


Ion Ratio Lower Upper

252 100

254 65.7 51.6 77.4

126 11.3 8.5 12.7





#85

Benzo(a)anthracene

Concen: 21.508 ng/ul

RT: 21.563 min Scan# 3124

Delta R.T. -0.006 min

Lab File: BP015426.D

Acq: 08 Jun 2023 09:07

Instrument :

BNA_P

ClientSampleId :

SSTD020739





Tgt Ion:228 Resp: 1569340

Ion Ratio Lower Upper

228 100

229 19.5 15.8 23.6

226 27.1 21.8 32.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 22.736 ng/ul

RT: 21.457 min Scan# 3106

Delta R.T. -0.006 min

Lab File: BP015426.D

Acq: 08 Jun 2023 09:07




Tgt Ion:149 Resp: 1045657


Ion Ratio Lower Upper

149 100

167 26.7 21.8 32.8

279 4.0 3.7 5.5





#87

Chrysene

Concen: 21.520 ng/ul

RT: 21.622 min Scan# 3134

Delta R.T. -0.000 min

Lab File: BP015426.D

Acq: 08 Jun 2023 09:07

Instrument :

BNA_P

ClientSampleId :

SSTD020739





Tgt Ion:228 Resp: 1484187

Ion Ratio Lower Upper

228 100

226 29.5 23.9 35.9

229 19.3 15.1 22.7



#89

Di-n-octyl phthalate

Concen: 23.619 ng/ul




RT: 22.427 min Scan# 3271


Delta R.T. -0.006 min

Lab File: BP015426.D

Acq: 08 Jun 2023 09:07

Tgt Ion:149 Resp: 1825980









#90
Benzo(b)fluoranthene
Concen: 22.158 ng/ul
RT: 23.345 min Scan# 3427
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:252 Resp: 1682442
Ion Ratio Lower Upper
252 100
253 21.6 17.4 26.2
125 9.5 7.3 10.9



#91
Benzo(k)fluoranthene
Concen: 21.236 ng/ul
RT: 23.398 min Scan# 3436
Delta R.T. -0.006 min
Lab File: BP015426.D
Acq: 08 Jun 2023 09:07

Tgt Ion:252 Resp: 1558694
Ion Ratio Lower Upper
252 100
253 21.6 17.7 26.5
125 8.9 7.3 10.9





#93

Benzo(a)pyrene

Concen: 21.059 ng/ul

RT: 24.021 min Scan# 3542

Delta R.T. -0.006 min

Lab File: BP015426.D

Acq: 08 Jun 2023 09:07

Instrument :

BNA_P

ClientSampleId :

SSTD020739





Tgt Ion:252 Resp: 1394116

Ion Ratio Lower Upper

252 100

253 21.7 17.8 26.6

125 9.9 8.1 12.1



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.411 ng/ul

RT: 26.833 min Scan# 4020

Delta R.T. -0.012 min

Lab File: BP015426.D

Acq: 08 Jun 2023 09:07




Tgt Ion:276 Resp: 1663062


Ion Ratio Lower Upper

276 100

138 20.9 17.3 25.9

277 24.8 20.2 30.2





#95

Dibenzo(a,h)anthracene

Concen: 19.768 ng/ul

RT: 26.851 min Scan# 4023

Delta R.T. -0.012 min

Lab File: BP015426.D

Acq: 08 Jun 2023 09:07

Instrument :

BNA_P

ClientSampleId :

SSTD020739





Tgt Ion:278 Resp: 1379811

Ion Ratio Lower Upper

278 100

139 14.4 11.4 17.0

279 24.1 18.7 28.1



#96

Benzo(g,h,i)perylene

Concen: 18.860 ng/ul

RT: 27.668 min Scan# 4162

Delta R.T. -0.018 min

Lab File: BP015426.D

Acq: 08 Jun 2023 09:07

Tgt Ion:276 Resp: 1331626

Ion Ratio Lower Upper

276 100

138 19.1 15.6 23.4

277 23.8 19.0 28.6

