

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060823\
 Data File : BP015427.D
 Acq On : 08 Jun 2023 09:53
 Operator : MA/JU
 Sample : PB153179BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK179

Quant Time: Jun 09 01:14:43 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jun 07 23:46:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.099	152	133278	20.000	ng/ul	0.00
20) Naphthalene-d8	10.928	136	609211	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.745	164	437473	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.504	188	1041759	20.000	ng/ul	0.00
79) Chrysene-d12	21.580	240	1058551	20.000	ng/ul	0.00
88) Perylene-d12	24.139	264	1117278	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.411	96	20016	5.561	ng/uL	0.00
4) Pyridine-d5	3.846	84	242252	25.912	ng/ul	0.00
7) Phenol-d5	7.252	99	322546	26.269	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.416	67	214395	29.237	ng/ul	0.00
11) 2-Chlorophenol-d4	7.622	132	257148	28.886	ng/ul	0.00
15) 4-Methylphenol-d8	8.810	113	282786	28.062	ng/ul	0.00
21) Nitrobenzene-d5	9.275	128	138641	28.987	ng/ul	0.00
24) 2-Nitrophenol-d4	10.005	143	159842	29.264	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.546	165	269371	27.035	ng/ul	0.00
31) 4-Chloroaniline-d4	11.069	131	401914	28.103	ng/ul	0.00
46) Dimethylphthalate-d6	14.151	166	1074895	31.159	ng/ul	0.00
49) Acenaphthylene-d8	14.440	160	1121689	29.736	ng/ul	0.00
54) 4-Nitrophenol-d4	14.945	143	148486	24.036	ng/ul	0.00
60) Fluorene-d10	15.740	176	871108	29.945	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.863	200	161487	24.469	ng/ul	0.00
73) Anthracene-d10	17.604	188	1481569	31.258	ng/ul	0.00
81) Pyrene-d10	19.828	212	1803976	31.841	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.974	264	1761310	31.418	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060823\
Data File : BP015427.D
Acq On : 08 Jun 2023 09:53
Operator : MA/JU
Sample : PB153179BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK179

Quant Time: Jun 09 01:14:43 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jun 07 23:46:15 2023
Response via : Initial Calibration

