

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060823\
 Data File : BP015438.D
 Acq On : 08 Jun 2023 17:11
 Operator : MA/JU
 Sample : PB153136BS
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS136

Quant Time: Jun 09 01:19:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jun 07 23:46:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.099	152	107630	20.000	ng/u1	0.00
20) Naphthalene-d8	10.922	136	496290	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.740	164	347381	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.498	188	822379	20.000	ng/u1	0.00
79) Chrysene-d12	21.580	240	821521	20.000	ng/u1	0.00
88) Perylene-d12	24.133	264	872325	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.411	96	26626	9.161	ng/uL	0.00
4) Pyridine-d5	3.846	84	363738	48.179	ng/u1	0.00
7) Phenol-d5	7.252	99	464998	46.896	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.416	67	286916	48.451	ng/u1	0.00
11) 2-Chlorophenol-d4	7.622	132	362064	50.363	ng/u1	0.00
15) 4-Methylphenol-d8	8.816	113	403847	49.626	ng/u1	0.00
21) Nitrobenzene-d5	9.275	128	193181	49.580	ng/u1	0.00
24) 2-Nitrophenol-d4	10.004	143	227452	51.117	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.546	165	405885	50.005	ng/u1	0.00
31) 4-Chloroaniline-d4	11.063	131	524620	45.029	ng/u1	0.00
46) Dimethylphthalate-d6	14.145	166	1393076	50.855	ng/u1	0.00
49) Acenaphthylene-d8	14.440	160	1518144	50.684	ng/u1	0.00
54) 4-Nitrophenol-d4	14.945	143	254722	51.926	ng/u1	0.00
60) Fluorene-d10	15.734	176	1160569	50.242	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.863	200	259561	49.821	ng/u1	0.00
73) Anthracene-d10	17.598	188	1840923	49.200	ng/u1	0.00
81) Pyrene-d10	19.822	212	2295762	52.213	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.968	264	2309088	52.755	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.446	88	53125	17.304	ng/uL	90
5) Pyridine	3.869	79	347260	44.367	ng/u1	96
6) Benzaldehyde	7.228	77	262034	69.138	ng/u1	98
8) Phenol	7.281	94	465195	45.475	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.510	93	356698	44.102	ng/u1	99
12) 2-Chlorophenol	7.658	128	346511	45.620	ng/u1	98
13) 2-Methylphenol	8.546	108	357151	45.377	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.622	45	480763	44.856	ng/u1	99
16) Acetophenone	8.934	105	590363	45.457	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.916	70	319132	45.528	ng/u1	98
18) 4-Methylphenol	8.875	108	397432	45.894	ng/u1	98
19) Hexachloroethane	9.187	117	149632	46.618	ng/u1	98
22) Nitrobenzene	9.316	77	477672	46.228	ng/u1	99
23) Isophorone	9.846	82	933731	45.141	ng/u1	99
25) 2-Nitrophenol	10.034	139	224559	46.742	ng/u1	99
26) 2,4-Dimethylphenol	10.087	107	445245	43.702	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.328	93	541489	45.027	ng/u1	99
29) 2,4-Dichlorophenol	10.569	162	379942	47.079	ng/u1	97
30) Naphthalene	10.975	128	1195995	44.472	ng/u1	99
32) 4-Chloroaniline	11.087	127	450660	37.383	ng/u1	99
33) Hexachlorobutadiene	11.251	225	246486	45.288	ng/u1	97
34) Caprolactam	11.875	113	86953	28.622	ng/u1	99
35) 4-Chloro-3-methylphenol	12.204	107	467791	47.840	ng/u1	100
36) 2-Methylnaphthalene	12.575	142	853185	45.128	ng/u1	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060823\
 Data File : BP015438.D
 Acq On : 08 Jun 2023 17:11
 Operator : MA/JU
 Sample : PB153136BS
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS136

Quant Time: Jun 09 01:19:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jun 07 23:46:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.793	142	858142	45.081	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.940	216	491029	45.393	ng/ul	99
40) Hexachlorocyclopentadiene	12.910	237	220822	49.801	ng/ul	96
41) 2,4,6-Trichlorophenol	13.181	196	335596	46.613	ng/ul	98
42) 2,4,5-Trichlorophenol	13.257	196	368209	46.933	ng/ul	98
43) 1,1'-Biphenyl	13.575	154	1223041	45.527	ng/ul	100
44) 2-Chloronaphthalene	13.622	162	956538	45.383	ng/ul	99
45) 2-Nitroaniline	13.834	65	335197	49.691	ng/ul	98
47) Dimethylphthalate	14.192	163	1304617	46.127	ng/ul	99
48) 2,6-Dinitrotoluene	14.322	165	285269	49.230	ng/ul	100
50) Acenaphthylene	14.469	152	1510037	45.524	ng/ul	99
51) 3-Nitroaniline	14.651	138	276031	57.705	ng/ul	97
52) Acenaphthene	14.804	153	1041311	45.415	ng/ul	99
53) 2,4-Dinitrophenol	14.863	184	161845	46.431	ng/ul	97
55) 4-Nitrophenol	14.957	109	235249	49.718	ng/ul	92
56) Dibenzofuran	15.140	168	1485001	45.695	ng/ul	98
57) 2,4-Dinitrotoluene	15.110	165	413047	48.465	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.363	232	327329	46.209	ng/ul	100
59) Diethylphthalate	15.551	149	1369044	47.230	ng/ul	99
61) Fluorene	15.792	166	1213459	45.805	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.775	204	620026	46.438	ng/ul	97
63) 4-Nitroaniline	15.822	138	287938	63.460	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.875	198	242383	45.616	ng/ul	91
67) N-Nitrosodiphenylamine	15.992	169	1066762	45.723	ng/ul	100
68) 4-Bromophenyl-phenylether	16.675	248	409928	46.593	ng/ul	98
69) Hexachlorobenzene	16.798	284	480174	46.268	ng/ul	95
70) Atrazine	16.945	200	215210	23.099	ng/ul	98
71) Pentachlorophenol	17.145	266	272986	46.538	ng/ul	100
72) Phenanthrene	17.539	178	2045344	46.346	ng/ul	99
74) Anthracene	17.633	178	2024116	45.203	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.545	216	516259	45.485	ng/ul	99
76) Pentachlorobenzene	15.057	250	521706	43.902	ng/ul	97
77) Carbazole	17.898	167	1877035	46.555	ng/ul	100
78) Di-n-butylphthalate	18.428	149	2454459	47.980	ng/ul	99
80) Fluoranthene	19.492	202	2557356	47.749	ng/ul	100
82) Pyrene	19.851	202	2608984	47.085	ng/ul	97
83) Butylbenzylphthalate	20.698	149	1221310	49.868	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.486	252	920673	47.436	ng/ul	100
85) Benzo(a)anthracene	21.563	228	2678329	46.636	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.457	149	1842605	50.901	ng/ul	100
87) Chrysene	21.622	228	2496096	45.982	ng/ul	99
89) Di-n-octyl phthalate	22.427	149	3274660	57.118	ng/ul	100
90) Benzo(b)fluoranthene	23.345	252	2720425	48.313	ng/ul	100
91) Benzo(k)fluoranthene	23.398	252	2661757	48.901	ng/ul	100
93) Benzo(a)pyrene	24.021	252	2311056	47.075	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.839	276	2736654	43.073	ng/ul	99
95) Dibenzo(a,h)anthracene	26.851	278	2274887	43.949	ng/ul	99
96) Benzo(g,h,i)perylene	27.668	276	2205167	42.116	ng/ul	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 17.304 ng/uL
 RT: 3.446 min Scan# 44
 Delta R.T. -0.006 min
 Lab File: BP015438.D
 Acq: 08 Jun 2023 17:11




Instrument :
 BNA_P
 ClientSampleId :
 SLCS136


Tgt Ion: 88 Resp: 53125
 Ion Ratio Lower Upper
 88 100
 43 32.2 23.9 35.9
 58 74.7 51.4 77.2



#5
 Pyridine
 Concen: 44.367 ng/uL
 RT: 3.869 min Scan# 116
 Delta R.T. -0.006 min
 Lab File: BP015438.D
 Acq: 08 Jun 2023 17:11

Tgt Ion: 79 Resp: 347260
 Ion Ratio Lower Upper
 79 100
 52 78.8 59.5 89.3
 51 38.0 29.8 44.6









#6
Benzaldehyde
Concen: 69.138 ng/ul
RT: 7.228 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11




Instrument :
BNA_P
ClientSampleId :
SLCS136


Tgt Ion: 77 Resp: 262034
Ion Ratio Lower Upper
77 100
105 96.7 77.3 115.9
106 88.4 73.3 109.9



#8
Phenol
Concen: 45.475 ng/ul
RT: 7.281 min Scan# 696
Delta R.T. -0.000 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11

Tgt Ion: 94 Resp: 465195
Ion Ratio Lower Upper
94 100
65 30.7 23.5 35.3
66 44.0 34.2 51.4









#10
Bis(2-Chloroethyl)ether
Concen: 44.102 ng/ul
RT: 7.510 min Scan# 710
Delta R.T. -0.006 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11




Instrument :
BNA_P
ClientSampleId :
SLCS136


Tgt Ion:	93	Resp:	356698
Ion	Ratio	Lower	Upper
93	100		
63	78.8	62.6	93.8
95	32.6	25.5	38.3



#12
2-Chlorophenol
Concen: 45.620 ng/ul
RT: 7.658 min Scan# 760
Delta R.T. -0.000 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11

Tgt Ion:	128	Resp:	346511
Ion	Ratio	Lower	Upper
128	100		
64	49.6	40.2	60.4
130	32.6	24.5	36.7





#13

2-Methylphenol

Concen: 45.377 ng/ul

RT: 8.546 min Scan# 911

Delta R.T. -0.000 min

Lab File: BP015438.D

Acq: 08 Jun 2023 17:11

Instrument :

BNA_P

ClientSampleId :





SLCS136

Tgt Ion:108 Resp: 357151

Ion Ratio Lower Upper

108 100

107 89.6 71.2 106.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 44.856 ng/ul

RT: 8.622 min Scan# 924

Delta R.T. -0.006 min

Lab File: BP015438.D

Acq: 08 Jun 2023 17:11




Tgt Ion: 45 Resp: 480763


Ion Ratio Lower Upper

45 100

77 13.9 11.8 17.8

79 10.8 8.7 13.1









#16
Acetophenone
Concen: 45.457 ng/ul
RT: 8.934 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11




Instrument :
BNA_P
ClientSampleId :
SLCS136


Tgt Ion:105 Resp: 590363
Ion Ratio Lower Upper
105 100
77 82.3 64.9 97.3
51 30.6 23.5 35.3



#17
N-Nitroso-di-n-propylamine
Concen: 45.528 ng/ul
RT: 8.916 min Scan# 974
Delta R.T. -0.000 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11

Tgt Ion: 70 Resp: 319132
Ion Ratio Lower Upper
70 100
42 60.1 46.6 69.8
101 9.7 7.8 11.6
130 19.2 15.9 23.9









#18
4-Methylphenol
Concen: 45.894 ng/ul
RT: 8.875 min Scan# 90
Delta R.T. -0.006 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11




Instrument :
BNA_P
ClientSampleId :
SLCS136


Tgt Ion:108 Resp: 397432
Ion Ratio Lower Upper
108 100
107 114.0 89.3 133.9



#19
Hexachloroethane
Concen: 46.618 ng/ul
RT: 9.187 min Scan# 1020
Delta R.T. -0.006 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11

Tgt Ion:117 Resp: 149632
Ion Ratio Lower Upper
117 100
201 101.3 82.8 124.2
199 64.7 50.6 76.0









#22
Nitrobenzene
Concen: 46.228 ng/ul
RT: 9.316 min Scan# 1042
Delta R.T. -0.006 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11




Instrument :
BNA_P
ClientSampleId :
SLCS136


Tgt Ion: 77 Resp: 477672
Ion Ratio Lower Upper
77 100
123 41.2 33.5 50.3
65 14.4 11.6 17.4



#23
Isophorone
Concen: 45.141 ng/ul
RT: 9.846 min Scan# 1132
Delta R.T. -0.006 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11

Tgt Ion: 82 Resp: 933731
Ion Ratio Lower Upper
82 100
95 7.8 5.8 8.8
138 16.0 13.1 19.7









#25
2-Nitrophenol
Concen: 46.742 ng/ul
RT: 10.034 min Scan# 1164
Delta R.T. -0.006 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11




Instrument :
BNA_P
ClientSampleId :
SLCS136


Tgt Ion:139 Resp: 224559
Ion Ratio Lower Upper
139 100
65 58.7 46.9 70.3
109 38.5 30.2 45.2



#26
2,4-Dimethylphenol
Concen: 43.702 ng/ul
RT: 10.087 min Scan# 1173
Delta R.T. -0.006 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11

Tgt Ion:107 Resp: 445245
Ion Ratio Lower Upper
107 100
121 46.7 39.0 58.6
122 78.2 64.8 97.2









#27
 Bis(2-Chloroethoxy)methane
 Concen: 45.027 ng/ul
 RT: 10.328 min Scan# 1214
 Delta R.T. -0.006 min
 Lab File: BP015438.D
 Acq: 08 Jun 2023 17:11




Instrument :
 BNA_P
 ClientSampleId :
 SLCS136


Tgt Ion: 93 Resp: 541489
 Ion Ratio Lower Upper
 93 100
 95 32.6 25.7 38.5
 123 10.7 8.6 13.0



#29
 2,4-Dichlorophenol
 Concen: 47.079 ng/ul
 RT: 10.569 min Scan# 1255
 Delta R.T. -0.006 min
 Lab File: BP015438.D
 Acq: 08 Jun 2023 17:11

Tgt Ion: 162 Resp: 379942
 Ion Ratio Lower Upper
 162 100
 164 66.3 50.6 76.0
 98 37.5 30.2 45.2









#30
Naphthalene
Concen: 44.472 ng/ul
RT: 10.975 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11




Instrument :
BNA_P
ClientSampleId :
SLCS136


Tgt Ion:128 Resp: 1195995
Ion Ratio Lower Upper
128 100
129 10.6 8.6 13.0
127 13.4 10.5 15.7



#32
4-Chloroaniline
Concen: 37.383 ng/ul
RT: 11.087 min Scan# 1343
Delta R.T. -0.006 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11

Tgt Ion:127 Resp: 450660
Ion Ratio Lower Upper
127 100
129 32.0 26.1 39.1





#33

Hexachlorobutadiene

Concen: 45.288 ng/ul

RT: 11.251 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP015438.D

Acq: 08 Jun 2023 17:11

Instrument :

BNA_P

ClientSampleId :

SLCS136





Tgt Ion:225 Resp: 246486

Ion Ratio Lower Upper

225 100

223 63.2 52.0 78.0

227 65.9 50.6 75.8



#34

Caprolactam

Concen: 28.622 ng/ul

RT: 11.875 min Scan# 1477

Delta R.T. 0.006 min

Lab File: BP015438.D

Acq: 08 Jun 2023 17:11




Tgt Ion:113 Resp: 86953


Ion Ratio Lower Upper

113 100

55 162.7 127.9 191.9

56 126.7 101.1 151.7









#35
4-Chloro-3-methylphenol
Concen: 47.840 ng/ul
RT: 12.204 min Scan# 1533
Delta R.T. -0.006 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11




Instrument :
BNA_P
ClientSampleId :
SLCS136


Tgt Ion:107 Resp: 467791
Ion Ratio Lower Upper
107 100
144 24.4 19.4 29.0
142 75.9 60.4 90.6



#36
2-Methylnaphthalene
Concen: 45.128 ng/ul
RT: 12.575 min Scan# 1596
Delta R.T. -0.006 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11

Tgt Ion:142 Resp: 853185
Ion Ratio Lower Upper
142 100
141 89.7 72.0 108.0





#37

1-Methylnaphthalene

Concen: 45.081 ng/ul

RT: 12.793 min Scan# 1633

Delta R.T. -0.006 min

Lab File: BP015438.D

Acq: 08 Jun 2023 17:11

Instrument :

BNA_P

ClientSampleId :

SLCS136





Tgt Ion:142 Resp: 858142

Ion Ratio Lower Upper

142 100

141 93.3 71.9 107.9

116 4.4 3.4 5.2



#39

1,2,4,5-Tetrachlorobenzene

Concen: 45.393 ng/ul

RT: 12.940 min Scan# 1658

Delta R.T. -0.006 min

Lab File: BP015438.D

Acq: 08 Jun 2023 17:11

Tgt Ion:216 Resp: 491029




Ion Ratio Lower Upper


216 100

214 78.1 62.2 93.2

179 22.1 17.4 26.2

108 18.3 14.2 21.4









#40
Hexachlorocyclopentadiene
Concen: 49.801 ng/ul
RT: 12.910 min Scan# 1653
Delta R.T. -0.012 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11




Instrument :
BNA_P
ClientSampleId :
SLCS136


Tgt Ion:237 Resp: 220822
Ion Ratio Lower Upper
237 100
235 64.3 48.2 72.2
272 12.7 10.3 15.5



#41
2,4,6-Trichlorophenol
Concen: 46.613 ng/ul
RT: 13.181 min Scan# 1699
Delta R.T. -0.006 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11

Tgt Ion:196 Resp: 335596
Ion Ratio Lower Upper
196 100
198 95.8 77.8 116.6
200 30.6 25.8 38.6









#42
2,4,5-Trichlorophenol
Concen: 46.933 ng/ul
RT: 13.257 min Scan# 1711
Delta R.T. -0.000 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11




Instrument :
BNA_P
ClientSampleId :
SLCS136


Tgt Ion:196 Resp: 368209
Ion Ratio Lower Upper
196 100
198 96.5 75.5 113.3
200 31.6 24.7 37.1



#43
1,1'-Biphenyl
Concen: 45.527 ng/ul
RT: 13.575 min Scan# 1766
Delta R.T. -0.006 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11

Tgt Ion:154 Resp: 1223041
Ion Ratio Lower Upper
154 100
153 40.2 32.2 48.2
76 13.2 10.4 15.6





#44

2-Chloronaphthalene

Concen: 45.383 ng/ul

RT: 13.622 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP015438.D

Acq: 08 Jun 2023 17:11

Instrument :

BNA_P

ClientSampleId :

SLCS136





Tgt Ion:162 Resp: 956538

Ion Ratio Lower Upper

162 100

164 33.0 27.0 40.6

127 40.7 32.2 48.4



#45

2-Nitroaniline

Concen: 49.691 ng/ul

RT: 13.834 min Scan# 1810

Delta R.T. -0.000 min

Lab File: BP015438.D

Acq: 08 Jun 2023 17:11




Tgt Ion: 65 Resp: 335197


Ion Ratio Lower Upper

65 100

92 66.4 55.4 83.2

138 98.0 78.8 118.2





#47

Dimethylphthalate

Concen: 46.127 ng/ul

RT: 14.192 min Scan# 1871

Delta R.T. -0.006 min

Lab File: BP015438.D

Acq: 08 Jun 2023 17:11

Instrument :

BNA_P

ClientSampleId :

SLCS136





Tgt Ion:163 Resp: 1304617

Ion Ratio Lower Upper

163 100

194 4.0 3.2 4.8

164 10.5 8.2 12.2



#48

2,6-Dinitrotoluene

Concen: 49.230 ng/ul

RT: 14.322 min Scan# 1893

Delta R.T. -0.000 min

Lab File: BP015438.D

Acq: 08 Jun 2023 17:11




Tgt Ion:165 Resp: 285269


Ion Ratio Lower Upper

165 100

89 64.7 51.5 77.3

121 21.8 17.3 25.9









#50
Acenaphthylene
Concen: 45.524 ng/ul
RT: 14.469 min Scan# 1918
Delta R.T. -0.000 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11




Instrument :
BNA_P
ClientSampleId :
SLCS136


Tgt Ion:152 Resp: 1510037
Ion Ratio Lower Upper
152 100
151 19.4 16.0 24.0
153 12.7 10.6 15.8



#51
3-Nitroaniline
Concen: 57.705 ng/ul
RT: 14.651 min Scan# 1949
Delta R.T. -0.006 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11

Tgt Ion:138 Resp: 276031
Ion Ratio Lower Upper
138 100
108 11.6 8.4 12.6
92 125.6 98.0 147.0





#52

Acenaphthene

Concen: 45.415 ng/ul

RT: 14.804 min Scan# 1975

Delta R.T. -0.006 min

Lab File: BP015438.D

Acq: 08 Jun 2023 17:11

Instrument :

BNA_P

ClientSampleId :

SLCS136





Tgt Ion:153 Resp: 1041311

Ion Ratio Lower Upper

153 100

152 47.1 38.3 57.5

154 87.4 69.8 104.6



#53

2,4-Dinitrophenol

Concen: 46.431 ng/ul

RT: 14.863 min Scan# 1985

Delta R.T. -0.006 min

Lab File: BP015438.D

Acq: 08 Jun 2023 17:11




Tgt Ion:184 Resp: 161845


Ion Ratio Lower Upper

184 100

63 76.7 59.0 88.4

154 74.1 57.6 86.4









#55
4-Nitrophenol
Concen: 49.718 ng/ul
RT: 14.957 min Scan# 2001
Delta R.T. -0.000 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11




Instrument :
BNA_P
ClientSampleId :
SLCS136


Tgt Ion:109 Resp: 235249
Ion Ratio Lower Upper
109 100
139 102.8 89.1 133.7
65 118.8 103.1 154.7



#56
Dibenzofuran
Concen: 45.695 ng/ul
RT: 15.140 min Scan# 2032
Delta R.T. -0.006 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11

Tgt Ion:168 Resp: 1485001
Ion Ratio Lower Upper
168 100
139 40.1 31.2 46.8









#57
2,4-Dinitrotoluene
Concen: 48.465 ng/ul
RT: 15.110 min Scan# 2027
Delta R.T. -0.000 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11




Instrument :
BNA_P
ClientSampleId :
SLCS136


Tgt Ion:165 Resp: 413047
Ion Ratio Lower Upper
165 100
63 48.8 39.1 58.7
182 2.7 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 46.209 ng/ul
RT: 15.363 min Scan# 2070
Delta R.T. -0.006 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11

Tgt Ion:232 Resp: 327329
Ion Ratio Lower Upper
232 100
131 47.7 37.8 56.8
130 2.7 1.9 2.9
166 32.6 26.2 39.2





#59

Diethylphthalate

Concen: 47.230 ng/ul

RT: 15.551 min Scan# 2102

Delta R.T. -0.006 min

Lab File: BP015438.D

Acq: 08 Jun 2023 17:11

Instrument :

BNA_P

ClientSampleId :

SLCS136





Tgt Ion:149 Resp: 1369044

Ion Ratio Lower Upper

149 100

177 21.6 17.8 26.8

150 12.4 10.1 15.1



#61

Fluorene

Concen: 45.805 ng/ul

RT: 15.792 min Scan# 2143

Delta R.T. -0.000 min

Lab File: BP015438.D

Acq: 08 Jun 2023 17:11




Tgt Ion:166 Resp: 1213459


Ion Ratio Lower Upper

166 100

165 97.6 79.1 118.7

167 12.8 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 46.438 ng/ul
RT: 15.775 min Scan# 2140
Delta R.T. -0.006 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11




Instrument :
BNA_P
ClientSampleId :
SLCS136


Tgt Ion:204 Resp: 620026
Ion Ratio Lower Upper
204 100
206 33.5 26.6 40.0
141 65.4 49.8 74.8



#63
4-Nitroaniline
Concen: 63.460 ng/ul
RT: 15.822 min Scan# 2148
Delta R.T. 0.006 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11

Tgt Ion:138 Resp: 287938
Ion Ratio Lower Upper
138 100
92 56.1 44.6 67.0
108 103.6 78.4 117.6









#66
4,6-Dinitro-2-methylphenol
Concen: 45.616 ng/ul
RT: 15.875 min Scan# 2157
Delta R.T. -0.006 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11




Instrument :
BNA_P
ClientSampleId :
SLCS136


Tgt Ion:198 Resp: 242383
Ion Ratio Lower Upper
198 100
182 3.5 3.0 4.4
77 30.5 20.5 30.7



#67
N-Nitrosodiphenylamine
Concen: 45.723 ng/ul
RT: 15.992 min Scan# 2177
Delta R.T. -0.006 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11

Tgt Ion:169 Resp: 1066762
Ion Ratio Lower Upper
169 100
168 66.4 53.0 79.4
167 36.2 28.9 43.3









#68
4-Bromophenyl-phenylether
Concen: 46.593 ng/ul
RT: 16.675 min Scan# 2293
Delta R.T. -0.006 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11




Instrument :
BNA_P
ClientSampleId :
SLCS136


Tgt Ion:	248	Resp:	409928
Ion	Ratio	Lower	Upper
248	100		
250	97.2	78.3	117.5
141	79.7	61.1	91.7



#69
Hexachlorobenzene
Concen: 46.268 ng/ul
RT: 16.798 min Scan# 2314
Delta R.T. -0.006 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11

Tgt Ion:	284	Resp:	480174
Ion	Ratio	Lower	Upper
284	100		
142	35.2	25.4	38.2
249	33.3	24.8	37.2









#70
Atrazine
Concen: 23.099 ng/ul
RT: 16.945 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11




Instrument :
BNA_P
ClientSampleId :
SLCS136


Tgt Ion:200 Resp: 215210
Ion Ratio Lower Upper
200 100
173 26.0 19.8 29.6
215 47.7 37.4 56.0



#71
Pentachlorophenol
Concen: 46.538 ng/ul
RT: 17.145 min Scan# 2373
Delta R.T. -0.006 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11

Tgt Ion:266 Resp: 272986
Ion Ratio Lower Upper
266 100
264 62.2 49.6 74.4
268 64.2 51.0 76.6





#72

Phenanthrene

Concen: 46.346 ng/ul

RT: 17.539 min Scan# 2440

Delta R.T. -0.006 min

Lab File: BP015438.D

Acq: 08 Jun 2023 17:11

Instrument :

BNA_P

ClientSampleId :

SLCS136





Tgt Ion:178 Resp: 2045344

Ion Ratio Lower Upper

178 100

179 15.4 12.3 18.5

176 18.7 15.4 23.0



#74

Anthracene

Concen: 45.203 ng/ul

RT: 17.633 min Scan# 2456

Delta R.T. -0.000 min

Lab File: BP015438.D

Acq: 08 Jun 2023 17:11




Tgt Ion:178 Resp: 2024116


Ion Ratio Lower Upper

178 100

179 15.4 12.4 18.6

176 19.1 15.0 22.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 45.485 ng/uL
RT: 13.545 min Scan# 1761
Delta R.T. -0.006 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11




Instrument :
BNA_P
ClientSampleId :
SLCS136


Tgt Ion:216 Resp: 516259
Ion Ratio Lower Upper
216 100
214 77.0 62.1 93.1
218 47.0 39.0 58.4



#76
Pentachlorobenzene
Concen: 43.902 ng/uL
RT: 15.057 min Scan# 2018
Delta R.T. -0.006 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11

Tgt Ion:250 Resp: 521706
Ion Ratio Lower Upper
250 100
248 64.5 49.0 73.4
252 64.5 49.9 74.9





#77

Carbazole

Concen: 46.555 ng/ul

RT: 17.898 min Scan# 2501

Delta R.T. -0.006 min

Lab File: BP015438.D

Acq: 08 Jun 2023 17:11

Instrument :

BNA_P

ClientSampleId :

SLCS136





Tgt Ion:167 Resp: 1877035

Ion Ratio Lower Upper

167 100

166 21.5 17.2 25.8

139 13.8 10.6 16.0



#78

Di-n-butylphthalate

Concen: 47.980 ng/ul

RT: 18.428 min Scan# 2591

Delta R.T. -0.006 min

Lab File: BP015438.D

Acq: 08 Jun 2023 17:11




Tgt Ion:149 Resp: 2454459


Ion Ratio Lower Upper

149 100

150 9.2 7.1 10.7

104 6.0 4.5 6.7









#80
Fluoranthene
Concen: 47.749 ng/ul
RT: 19.492 min Scan# 2772
Delta R.T. -0.006 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11




Instrument :
BNA_P
ClientSampleId :
SLCS136


Tgt Ion:202 Resp: 2557356
Ion Ratio Lower Upper
202 100
101 10.8 8.6 13.0
100 8.0 6.6 9.8



#82
Pyrene
Concen: 47.085 ng/ul
RT: 19.851 min Scan# 2833
Delta R.T. -0.000 min
Lab File: BP015438.D
Acq: 08 Jun 2023 17:11

Tgt Ion:202 Resp: 2608984
Ion Ratio Lower Upper
202 100
101 11.4 9.9 14.9
100 9.5 8.4 12.6





#83

Butylbenzylphthalate

Concen: 49.868 ng/ul

RT: 20.698 min Scan# 2977

Delta R.T. -0.006 min

Lab File: BP015438.D

Acq: 08 Jun 2023 17:11

Instrument :

BNA_P

ClientSampleId :

SLCS136





Tgt Ion:149 Resp: 1221310

Ion Ratio Lower Upper

149 100

91 76.2 60.4 90.6

206 20.0 16.4 24.6



#84

3,3'-Dichlorobenzidine

Concen: 47.436 ng/ul

RT: 21.486 min Scan# 3111

Delta R.T. -0.006 min

Lab File: BP015438.D

Acq: 08 Jun 2023 17:11




Tgt Ion:252 Resp: 920673


Ion Ratio Lower Upper

252 100

254 64.5 51.6 77.4

126 10.4 8.5 12.7





#85

Benzo(a)anthracene

Concen: 46.636 ng/ul

RT: 21.563 min Scan# 3124

Delta R.T. -0.006 min

Lab File: BP015438.D

Acq: 08 Jun 2023 17:11

Instrument :

BNA_P

ClientSampleId :

SLCS136





Tgt Ion:228 Resp: 2678329

Ion Ratio Lower Upper

228 100

229 19.6 15.8 23.6

226 27.2 21.8 32.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 50.901 ng/ul

RT: 21.457 min Scan# 3106

Delta R.T. -0.006 min

Lab File: BP015438.D

Acq: 08 Jun 2023 17:11




Tgt Ion:149 Resp: 1842605


Ion Ratio Lower Upper

149 100

167 27.4 21.8 32.8

279 4.5 3.7 5.5





#87

Chrysene

Concen: 45.982 ng/ul

RT: 21.622 min Scan# 3134

Delta R.T. -0.000 min

Lab File: BP015438.D

Acq: 08 Jun 2023 17:11

Instrument :

BNA_P

ClientSampleId :

SLCS136





Tgt Ion:228 Resp: 2496096

Ion Ratio Lower Upper

228 100

226 30.3 23.9 35.9

229 19.7 15.1 22.7



#89

Di-n-octyl phthalate

Concen: 57.118 ng/ul




RT: 22.427 min Scan# 3271


Delta R.T. -0.006 min

Lab File: BP015438.D

Acq: 08 Jun 2023 17:11

Tgt Ion:149 Resp: 3274660





#90

Benzo(b)fluoranthene

Concen: 48.313 ng/ul

RT: 23.345 min Scan# 3427

Delta R.T. -0.006 min

Lab File: BP015438.D

Acq: 08 Jun 2023 17:11

Instrument :

BNA_P

ClientSampleId :

SLCS136





Tgt Ion:252 Resp: 2720425

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.2

125 9.0 7.3 10.9



#91

Benzo(k)fluoranthene

Concen: 48.901 ng/ul

RT: 23.398 min Scan# 3436

Delta R.T. -0.006 min

Lab File: BP015438.D

Acq: 08 Jun 2023 17:11




Tgt Ion:252 Resp: 2661757


Ion Ratio Lower Upper

252 100

253 21.8 17.7 26.5

125 9.1 7.3 10.9





#93

Benzo(a)pyrene

Concen: 47.075 ng/ul

RT: 24.021 min Scan# 3

Delta R.T. -0.006 min

Lab File: BP015438.D

Acq: 08 Jun 2023 17:11

Instrument :

BNA_P

ClientSampleId :

SLCS136





Tgt Ion:252 Resp: 2311056

Ion Ratio Lower Upper

252 100

253 21.6 17.8 26.6

125 9.5 8.1 12.1



#94

Indeno(1,2,3-cd)pyrene

Concen: 43.073 ng/ul

RT: 26.839 min Scan# 4021

Delta R.T. -0.006 min

Lab File: BP015438.D

Acq: 08 Jun 2023 17:11




Tgt Ion:276 Resp: 2736654


Ion Ratio Lower Upper

276 100

138 22.2 17.3 25.9

277 25.3 20.2 30.2









#95
 Dibenzo(a,h)anthracene
 Concen: 43.949 ng/ul
 RT: 26.851 min Scan# 4023
 Delta R.T. -0.012 min
 Lab File: BP015438.D
 Acq: 08 Jun 2023 17:11

Instrument :
 BNA_P
 ClientSampleId :
 SLCS136

Tgt Ion:278 Resp: 2274887

Ion	Ratio	Lower	Upper
278	100		
139	14.0	11.4	17.0
279	24.0	18.7	28.1



#96
 Benzo(g,h,i)perylene
 Concen: 42.116 ng/ul
 RT: 27.668 min Scan# 4162
 Delta R.T. -0.018 min
 Lab File: BP015438.D
 Acq: 08 Jun 2023 17:11

Tgt Ion:276 Resp: 2205167

Ion	Ratio	Lower	Upper
276	100		
138	18.5	15.6	23.4
277	23.8	19.0	28.6

