

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060923\
 Data File : BP015449.D
 Acq On : 09 Jun 2023 12:47
 Operator : MA/JU
 Sample : PB153136BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK136

Quant Time: Jun 09 22:56:43 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jun 07 23:46:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.099	152	204368	20.000	ng/ul	0.00
20) Naphthalene-d8	10.928	136	920736	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.745	164	639309	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.504	188	1514647	20.000	ng/ul	0.00
79) Chrysene-d12	21.586	240	1505501	20.000	ng/ul	0.00
88) Perylene-d12	24.145	264	1582244	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.405	96	31738	5.751	ng/uL	0.00
4) Pyridine-d5	3.846	84	409869	28.591	ng/ul	0.00
7) Phenol-d5	7.252	99	542234	28.800	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.416	67	352171	31.320	ng/ul	0.00
11) 2-Chlorophenol-d4	7.622	132	432630	31.693	ng/ul	0.00
15) 4-Methylphenol-d8	8.810	113	470592	30.455	ng/ul	0.00
21) Nitrobenzene-d5	9.275	128	232142	32.114	ng/ul	0.00
24) 2-Nitrophenol-d4	10.004	143	270647	32.785	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.546	165	458498	30.447	ng/ul	0.00
31) 4-Chloroaniline-d4	11.069	131	673739	31.170	ng/ul	0.00
46) Dimethylphthalate-d6	14.151	166	1715442	34.027	ng/ul	0.00
49) Acenaphthylene-d8	14.439	160	1816983	32.961	ng/ul	0.00
54) 4-Nitrophenol-d4	14.945	143	270501	29.963	ng/ul	0.00
60) Fluorene-d10	15.739	176	1402666	32.994	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.863	200	290516	30.277	ng/ul	0.00
73) Anthracene-d10	17.604	188	2344036	34.014	ng/ul	0.00
81) Pyrene-d10	19.827	212	2835362	35.188	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.980	264	2796810	35.229	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060923\
Data File : BP015449.D
Acq On : 09 Jun 2023 12:47
Operator : MA/JU
Sample : PB153136BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK136

Quant Time: Jun 09 22:56:43 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jun 07 23:46:15 2023
Response via : Initial Calibration

