

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060923\
 Data File : BP015452.D
 Acq On : 09 Jun 2023 15:11
 Operator : MA/JU
 Sample : PB153326BL
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK326

Quant Time: Jun 09 23:14:19 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jun 07 23:46:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.099	152	175498	20.000	ng/u1	0.00
20) Naphthalene-d8	10.922	136	808264	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.739	164	575700	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.498	188	1354558	20.000	ng/u1	0.00
79) Chrysene-d12	21.580	240	1355566	20.000	ng/u1	0.00
88) Perylene-d12	24.133	264	1395200	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.411	96	25707	5.424	ng/uL	0.00
4) Pyridine-d5	3.846	84	342487	27.821	ng/u1	0.00
7) Phenol-d5	7.246	99	433751	26.828	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.416	67	284507	29.465	ng/u1	0.00
11) 2-Chlorophenol-d4	7.622	132	343297	29.286	ng/u1	0.00
15) 4-Methylphenol-d8	8.810	113	380446	28.671	ng/u1	0.00
21) Nitrobenzene-d5	9.275	128	184507	29.076	ng/u1	0.00
24) 2-Nitrophenol-d4	9.998	143	217507	30.015	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.546	165	356839	26.994	ng/u1	0.00
31) 4-Chloroaniline-d4	11.063	131	551292	29.054	ng/u1	0.00
46) Dimethylphthalate-d6	14.145	166	1387720	30.568	ng/u1	0.00
49) Acenaphthylene-d8	14.434	160	1480605	29.827	ng/u1	0.00
54) 4-Nitrophenol-d4	14.939	143	195460	24.043	ng/u1	0.00
60) Fluorene-d10	15.733	176	1133471	29.608	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.857	200	225411	26.268	ng/u1	0.00
73) Anthracene-d10	17.598	188	1896613	30.774	ng/u1	0.00
81) Pyrene-d10	19.821	212	2334279	32.174	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.968	264	2205893	31.510	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060923\
Data File : BP015452.D
Acq On : 09 Jun 2023 15:11
Operator : MA/JU
Sample : PB153326BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK326

Quant Time: Jun 09 23:14:19 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jun 07 23:46:15 2023
Response via : Initial Calibration

