

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP061024\  
 Data File : BP020580.D  
 Acq On : 10 Jun 2024 16:55  
 Operator : MA/JU  
 Sample : P2784-04  
 Misc :  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 WCS-TPD

Quant Time: Jun 10 17:36:20 2024  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP052924.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu May 30 03:41:03 2024  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc    | Units | Dev(Min) |
|-----------------------------|--------|------|----------|---------|-------|----------|
| -----                       |        |      |          |         |       |          |
| Internal Standards          |        |      |          |         |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 7.758  | 152  | 179730   | 20.000  | ng    | -0.01    |
| 21) Naphthalene-d8          | 10.534 | 136  | 688841   | 20.000  | ng    | 0.00     |
| 39) Acenaphthene-d10        | 14.381 | 164  | 346955   | 20.000  | ng    | -0.01    |
| 64) Phenanthrene-d10        | 17.204 | 188  | 932006   | 20.000  | ng    | 0.00     |
| 76) Chrysene-d12            | 21.663 | 240  | 678040   | 20.000  | ng    | 0.00     |
| 86) Perylene-d12            | 25.045 | 264  | 195555   | 20.000  | ng    | 0.02     |
| System Monitoring Compounds |        |      |          |         |       |          |
| 5) 2-Fluorophenol           | 5.375  | 112  | 1399338  | 139.401 | ng    | 0.00     |
| 7) Phenol-d6                | 6.934  | 99   | 1402703  | 99.086  | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 8.905  | 82   | 1208816  | 96.061  | ng    | -0.02    |
| 42) 2,4,6-Tribromophenol    | 15.904 | 330  | 846517   | 235.097 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 12.993 | 172  | 2834313  | 121.779 | ng    | -0.01    |
| 79) Terphenyl-d14           | 19.939 | 244  | 4841562  | 118.870 | ng    | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP061024\  
Data File : BP020580.D  
Acq On : 10 Jun 2024 16:55  
Operator : MA/JU  
Sample : P2784-04  
Misc :  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
BNA\_P  
ClientSampleId :  
WCS-TPD

Quant Time: Jun 10 17:36:20 2024  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP052924.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu May 30 03:41:03 2024  
Response via : Initial Calibration

