

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061221\
 Data File : BP005894.D
 Acq On : 12 Jun 2021 18:13
 Operator : CG/JU
 Sample : M2691-04MSD
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ETGI-357MSD

Quant Time: Jun 14 02:09:41 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 11 13:53:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.940	152	83013	20.000	ng	0.00
21) Naphthalene-d8	10.746	136	317661	20.000	ng	0.00
39) Acenaphthene-d10	14.569	164	188407	20.000	ng	0.00
64) Phenanthrene-d10	17.310	188	383510	20.000	ng	-0.01
76) Chrysene-d12	21.386	240	433030	20.000	ng	0.00
86) Perylene-d12	23.798	264	500947	20.000	ng	-0.02

System Monitoring Compounds						
5) 2-Fluorophenol	5.511	112	627120	121.221	ng	0.00
7) Phenol-d6	7.122	99	748321	111.600	ng	0.01
23) Nitrobenzene-d5	9.104	82	457838	70.662	ng	0.00
42) 2,4,6-Tribromophenol	16.057	330	372686	115.268	ng	0.00
45) 2-Fluorobiphenyl	13.198	172	915789	63.827	ng	0.00
79) Terphenyl-d14	19.857	244	1515912	65.550	ng	-0.01

Target Compounds					Qvalue	
2) 1,4-Dioxane	3.369	88	75169	32.162	ng	98
3) Pyridine	3.787	79	204578	37.242	ng	98
4) n-Nitrosodimethylamine	3.693	42	90213	35.330	ng	# 89
6) Aniline	7.263	93	214601	28.667	ng	99
8) 2-Chlorophenol	7.505	128	256109	43.177	ng	99
9) Benzaldehyde	7.075	77	158147	55.325	ng	97
10) Phenol	7.146	94	294822	44.013	ng	100
11) bis(2-Chloroethyl)ether	7.363	93	207896	40.957	ng	98
12) 1,3-Dichlorobenzene	7.828	146	265749	40.067	ng	99
13) 1,4-Dichlorobenzene	7.975	146	275068	40.667	ng	99
14) 1,2-Dichlorobenzene	8.293	146	261251	40.407	ng	98
15) Benzyl Alcohol	8.187	79	219861	43.532	ng	97
16) 2,2'-oxybis(1-Chloropr...	8.475	45	168752	35.964	ng	94
17) 2-Methylphenol	8.399	107	201490	44.148	ng	98
18) Hexachloroethane	9.022	117	98501	40.259	ng	95
19) n-Nitroso-di-n-propyla...	8.752	70	160634	38.941	ng	95
20) 3+4-Methylphenols	8.728	107	268654	43.577	ng	99
22) Acetophenone	8.769	105	344532	41.206	ng	98
24) Nitrobenzene	9.146	77	251892	37.438	ng	99
25) Isophorone	9.675	82	426781	35.668	ng	99
26) 2-Nitrophenol	9.857	139	135086	42.080	ng	94
27) 2,4-Dimethylphenol	9.928	122	223477	46.728	ng	98
28) bis(2-Chloroethoxy)met...	10.157	93	262249	41.995	ng	99
29) 2,4-Dichlorophenol	10.404	162	239124	41.605	ng	99
30) 1,2,4-Trichlorobenzene	10.610	180	249710	38.667	ng	99
31) Naphthalene	10.793	128	715042	38.904	ng	100
32) Benzoic acid	10.104	122	182562	50.560	ng	98
33) 4-Chloroaniline	10.910	127	104919	14.130	ng	97
34) Hexachlorobutadiene	11.081	225	166672	38.156	ng	98
35) Caprolactam	11.710	113	71123	42.962	ng	95
36) 4-Chloro-3-methylphenol	12.045	107	242389	41.534	ng	98
37) 2-Methylnaphthalene	12.398	142	510669	39.024	ng	99
38) 1-Methylnaphthalene	12.622	142	468938	38.044	ng	99
40) 1,2,4,5-Tetrachloroben...	12.769	216	274614	41.409	ng	99
41) Hexachlorocyclopentadiene	12.745	237	244822	73.923	ng	99
43) 2,4,6-Trichlorophenol	13.010	196	185317	40.740	ng	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061221\
 Data File : BP005894.D
 Acq On : 12 Jun 2021 18:13
 Operator : CG/JU
 Sample : M2691-04MSD
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ETGI-357MSD

Quant Time: Jun 14 02:09:41 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 11 13:53:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.092	196	200464	40.921	ng	98
46) 1,1'-Biphenyl	13.404	154	613921	40.414	ng	98
47) 2-Chloronaphthalene	13.451	162	494134	39.835	ng	99
48) 2-Nitroaniline	13.657	65	136024	41.706	ng	96
49) Acenaphthylene	14.292	152	777213	40.840	ng	99
50) Dimethylphthalate	14.028	163	611667	39.486	ng	99
51) 2,6-Dinitrotoluene	14.145	165	135777	38.563	ng	96
52) Acenaphthene	14.634	154	456973	39.541	ng	99
53) 3-Nitroaniline	14.481	138	62321	18.263	ng	97
54) 2,4-Dinitrophenol	14.686	184	63703	33.426	ng	96
55) Dibenzofuran	14.963	168	731785	38.503	ng	98
56) 4-Nitrophenol	14.810	139	228757	82.702	ng	98
57) 2,4-Dinitrotoluene	14.934	165	189677	40.278	ng	99
58) Fluorene	15.610	166	581597	39.274	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.198	232	166930m	38.664	ng	
60) Diethylphthalate	15.386	149	622307	40.040	ng	99
61) 4-Chlorophenyl-phenyle...	15.604	204	296099	38.560	ng	97
62) 4-Nitroaniline	15.639	138	127152	36.200	ng	99
63) Azobenzene	15.898	77	522847	37.184	ng	96
65) 4,6-Dinitro-2-methylph...	15.698	198	60571	21.492	ng	98
66) n-Nitrosodiphenylamine	15.822	169	509265	40.044	ng	98
67) 4-Bromophenyl-phenylether	16.498	248	195585	39.734	ng	96
68) Hexachlorobenzene	16.616	284	238595	40.438	ng	98
69) Atrazine	16.775	200	193231	48.836	ng	99
70) Pentachlorophenol	16.969	266	253705	77.333	ng	99
71) Phenanthrene	17.357	178	938316	40.741	ng	100
72) Anthracene	17.445	178	955350	42.147	ng	100
73) Carbazole	17.716	167	857170	39.387	ng	100
74) Di-n-butylphthalate	18.263	149	1083254	42.962	ng	100
75) Fluoranthene	19.316	202	1239011	44.298	ng	98
77) Benzidine	19.492	184	839672	93.188	ng	99
78) Pyrene	19.669	202	1272616	42.097	ng	99
80) Butylbenzylphthalate	20.533	149	536962	43.174	ng	95
81) Benzo(a)anthracene	21.368	228	1254264	40.126	ng	99
82) 3,3'-Dichlorobenzidine	21.304	252	424069	39.240	ng	99
83) Chrysene	21.421	228	1230602	40.646	ng	99
84) Bis(2-ethylhexyl)phtha...	21.286	149	883877	48.457	ng	99
85) Di-n-octyl phthalate	22.210	149	1411381	43.253	ng	99
87) Indeno(1,2,3-cd)pyrene	26.351	276	1700227	39.689	ng	100
88) Benzo(b)fluoranthene	23.063	252	1426584	41.579	ng	99
89) Benzo(k)fluoranthene	23.110	252	1318600	39.633	ng	99
90) Benzo(a)pyrene	23.692	252	1370693	42.493	ng	99
91) Dibenzo(a,h)anthracene	26.362	278	1432943	38.864	ng	99
92) Benzo(g,h,i)perylene	27.127	276	1434206	39.376	ng	98





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.940 min Scan# 825
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:152 Resp: 83013
Ion Ratio Lower Upper
152 100
150 154.2 125.6 188.4
115 60.5 48.5 72.7



#2
1,4-Dioxane
Concen: 32.162 ng
RT: 3.369 min Scan# 48
Delta R.T. -0.012 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion: 88 Resp: 75169
Ion Ratio Lower Upper
88 100
58 60.0 46.7 70.1
43 25.5 20.3 30.5









#3
Pyridine
Concen: 37.242 ng
RT: 3.787 min Scan# 119
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion: 79 Resp: 204578
Ion Ratio Lower Upper
79 100
52 51.9 41.7 62.5
51 27.9 24.1 36.1



#4
n-Nitrosodimethylamine
Concen: 35.330 ng
RT: 3.693 min Scan# 103
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion: 42 Resp: 90213
Ion Ratio Lower Upper
42 100
74 145.0 105.4 158.0
44 6.7 4.2 6.4#









#5
2-Fluorophenol
Concen: 121.221 ng
RT: 5.511 min Scan# 412
Delta R.T. -0.000 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:112 Resp: 627120
Ion Ratio Lower Upper
112 100
64 50.8 41.4 62.2
63 28.4 23.8 35.8



#6
Aniline
Concen: 28.667 ng
RT: 7.263 min Scan# 710
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion: 93 Resp: 214601
Ion Ratio Lower Upper
93 100
66 38.0 29.7 44.5
65 18.7 15.3 22.9









#7
Phenol-d6
Concen: 111.600 ng
RT: 7.122 min Scan# 686
Delta R.T. 0.012 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD

Tgt Ion: 99 Resp: 748321




Ion	Ratio	Lower	Upper
99	100		
42	15.2	13.4	20.0
71	38.3	30.6	46.0




#8
2-Chlorophenol
Concen: 43.177 ng
RT: 7.505 min Scan# 751
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion: 128 Resp: 256109

Ion	Ratio	Lower	Upper
128	100		
130	31.8	12.4	52.4
64	39.2	20.3	60.3









#9
Benzaldehyde
Concen: 55.325 ng
RT: 7.075 min Scan# 678
Delta R.T. -0.012 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion: 77 Resp: 158147
Ion Ratio Lower Upper
77 100
105 101.8 76.7 116.7
106 97.5 76.2 116.2



#10
Phenol
Concen: 44.013 ng
RT: 7.146 min Scan# 690
Delta R.T. 0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion: 94 Resp: 294822
Ion Ratio Lower Upper
94 100
65 28.6 8.8 48.8
66 41.0 20.7 60.7









#11
bis(2-Chloroethyl)ether
Concen: 40.957 ng
RT: 7.363 min Scan# 727
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion: 93 Resp: 207896
Ion Ratio Lower Upper
93 100
63 64.0 46.2 86.2
95 33.0 12.2 52.2



#12
1,3-Dichlorobenzene
Concen: 40.067 ng
RT: 7.828 min Scan# 806
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:146 Resp: 265749
Ion Ratio Lower Upper
146 100
148 64.8 51.5 77.3
75 31.7 25.7 38.5









#13
1,4-Dichlorobenzene
Concen: 40.667 ng
RT: 7.975 min Scan# 831
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:146 Resp: 275068
Ion Ratio Lower Upper
146 100
148 63.1 51.3 76.9
111 40.9 33.0 49.6



#14
1,2-Dichlorobenzene
Concen: 40.407 ng
RT: 8.293 min Scan# 885
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:146 Resp: 261251
Ion Ratio Lower Upper
146 100
148 65.0 50.3 75.5
111 43.3 33.7 50.5









#15
Benzyl Alcohol
Concen: 43.532 ng
RT: 8.187 min Scan# 867
Delta R.T. -0.000 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD

Tgt Ion: 79 Resp: 219861




Ion	Ratio	Lower	Upper
79	100		
108	74.3	62.5	93.7
77	64.5	52.9	79.3




#16
2,2'-oxybis(1-Chloropropane)
Concen: 35.964 ng
RT: 8.475 min Scan# 916
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion: 45 Resp: 168752

Ion	Ratio	Lower	Upper
45	100		
77	25.3	2.0	42.0
79	18.4	0.0	36.7









#17
2-Methylphenol
Concen: 44.148 ng
RT: 8.399 min Scan# 903
Delta R.T. 0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:	107	Resp:	201490
Ion	Ratio	Lower	Upper
107	100		
108	109.4	87.7	131.5
77	46.9	40.2	60.2
79	47.1	39.0	58.4



#18
Hexachloroethane
Concen: 40.259 ng
RT: 9.022 min Scan# 1009
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:	117	Resp:	98501
Ion	Ratio	Lower	Upper
117	100		
119	98.7	76.2	114.2
201	110.9	83.4	125.0









#19
n-Nitroso-di-n-propylamine
Concen: 38.941 ng
RT: 8.752 min Scan# 963
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD

Tgt Ion: 70 Resp: 160634




Ion	Ratio	Lower	Upper
70	100		
42	45.8	40.8	61.2
101	11.2	8.8	13.2
130	24.3	18.9	28.3




#20
3+4-Methylphenols
Concen: 43.577 ng
RT: 8.728 min Scan# 959
Delta R.T. 0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion: 107 Resp: 268654

Ion	Ratio	Lower	Upper
107	100		
108	89.9	71.0	111.0
77	36.6	16.6	56.6
79	31.7	12.7	52.7









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.746 min Scan# 1302
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:	136	Resp:	317661
Ion	Ratio	Lower	Upper
136	100		
137	10.6	8.6	12.8
54	5.1	4.6	6.8
68	4.6	3.6	5.4



#22
Acetophenone
Concen: 41.206 ng
RT: 8.769 min Scan# 966
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:	105	Resp:	344532
Ion	Ratio	Lower	Upper
105	100		
71	1.9	1.9	2.9
51	21.2	18.4	27.6
120	22.9	17.9	26.9









#23
Nitrobenzene-d5
Concen: 70.662 ng
RT: 9.104 min Scan# 1023
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion: 82 Resp: 457838
Ion Ratio Lower Upper
82 100
128 43.9 35.7 53.5
54 40.8 35.0 52.4



#24
Nitrobenzene
Concen: 37.438 ng
RT: 9.146 min Scan# 1030
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion: 77 Resp: 251892
Ion Ratio Lower Upper
77 100
123 47.7 38.2 57.2
65 13.4 11.4 17.2









#25
Isophorone
Concen: 35.668 ng
RT: 9.675 min Scan# 1120
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:	82	Resp:	426781
Ion	Ratio	Lower	Upper
82	100		
95	8.5	6.5	9.7
138	19.1	15.0	22.4



#26
2-Nitrophenol
Concen: 42.080 ng
RT: 9.857 min Scan# 1151
Delta R.T. -0.012 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:	139	Resp:	135086
Ion	Ratio	Lower	Upper
139	100		
109	33.9	28.8	43.2
65	46.4	41.0	61.6









#27
2,4-Dimethylphenol
Concen: 46.728 ng
RT: 9.928 min Scan# 1163
Delta R.T. -0.000 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:122 Resp: 223477
Ion Ratio Lower Upper
122 100
107 120.3 94.5 141.7
121 61.3 47.4 71.2



#28
bis(2-Chloroethoxy)methane
Concen: 41.995 ng
RT: 10.157 min Scan# 1202
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion: 93 Resp: 262249
Ion Ratio Lower Upper
93 100
95 33.1 25.9 38.9
123 12.7 10.2 15.2









#29
2,4-Dichlorophenol
Concen: 41.605 ng
RT: 10.404 min Scan# 1244
Delta R.T. -0.000 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:	162	Resp:	239124
Ion	Ratio	Lower	Upper
162	100		
164	63.2	44.9	84.9
98	34.4	14.3	54.3



#30
1,2,4-Trichlorobenzene
Concen: 38.667 ng
RT: 10.610 min Scan# 1279
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:	180	Resp:	249710
Ion	Ratio	Lower	Upper
180	100		
182	96.3	77.9	116.9
145	31.1	24.7	37.1









#31
Naphthalene
Concen: 38.904 ng
RT: 10.793 min Scan# 1310
Delta R.T. -0.012 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:128 Resp: 715042
Ion Ratio Lower Upper
128 100
129 11.2 9.0 13.6
127 13.5 10.7 16.1



#32
Benzoic acid
Concen: 50.560 ng
RT: 10.104 min Scan# 1193
Delta R.T. 0.023 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:122 Resp: 182562
Ion Ratio Lower Upper
122 100
105 126.4 103.2 143.2
77 95.1 76.2 116.2









#33
4-Chloroaniline
Concen: 14.130 ng
RT: 10.910 min Scan# 1330
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:127	Resp: 104919
Ion Ratio	Lower Upper
127 100	
129 34.4	26.1 39.1
65 29.3	22.5 33.7
92 20.7	15.0 22.6



#34
Hexachlorobutadiene
Concen: 38.156 ng
RT: 11.081 min Scan# 1359
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:225	Resp: 166672
Ion Ratio	Lower Upper
225 100	
223 61.9	51.0 76.6
227 65.4	51.4 77.2









#35
Caprolactam
Concen: 42.962 ng
RT: 11.710 min Scan# 1466
Delta R.T. 0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:	113	Resp:	71123
Ion	Ratio	Lower	Upper
113	100		
55	110.1	98.9	138.9
56	92.6	71.7	111.7



#36
4-Chloro-3-methylphenol
Concen: 41.534 ng
RT: 12.045 min Scan# 1523
Delta R.T. 0.012 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:	107	Resp:	242389
Ion	Ratio	Lower	Upper
107	100		
144	26.5	21.1	31.7
142	80.1	66.0	99.0









#37
2-Methylnaphthalene
Concen: 39.024 ng
RT: 12.398 min Scan# 1583
Delta R.T. -0.012 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:142 Resp: 510669
Ion Ratio Lower Upper
142 100
141 89.0 71.8 107.6
115 38.6 31.6 47.4



#38
1-Methylnaphthalene
Concen: 38.044 ng
RT: 12.622 min Scan# 1621
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:142 Resp: 468938
Ion Ratio Lower Upper
142 100
141 91.1 73.3 109.9
116 4.3 3.5 5.3









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.569 min Scan# 1952
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:164 Resp: 188407
Ion Ratio Lower Upper
164 100
162 98.8 80.2 120.2
160 43.2 34.5 51.7



#40
1,2,4,5-Tetrachlorobenzene
Concen: 41.409 ng
RT: 12.769 min Scan# 1646
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:216 Resp: 274614
Ion Ratio Lower Upper
216 100
214 78.1 61.9 92.9
179 21.6 16.8 25.2
108 19.2 14.9 22.3









#41
Hexachlorocyclopentadiene
Concen: 73.923 ng
RT: 12.745 min Scan# 1642
Delta R.T. -0.012 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:	237	Resp:	244822
Ion	Ratio	Lower	Upper
237	100		
235	63.3	44.5	84.5
272	13.1	0.0	32.5



#42
2,4,6-Tribromophenol
Concen: 115.268 ng
RT: 16.057 min Scan# 2205
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:	330	Resp:	372686
Ion	Ratio	Lower	Upper
330	100		
332	97.3	77.8	116.8
141	29.9	25.4	38.2









#43
2,4,6-Trichlorophenol
Concen: 40.740 ng
RT: 13.010 min Scan# 1687
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Instrument :
BNA_P
ClientSampled :
ETGI-357MSD

Tgt Ion:196 Resp: 185317




Ion	Ratio	Lower	Upper
196	100		
198	97.8	79.5	119.3
200	31.7	25.0	37.6




#44
2,4,5-Trichlorophenol
Concen: 40.921 ng
RT: 13.092 min Scan# 1701
Delta R.T. 0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:196 Resp: 200464

Ion	Ratio	Lower	Upper
196	100		
198	95.3	74.7	112.1
97	48.2	37.4	56.0
132	26.9	20.0	30.0









#45
2-Fluorobiphenyl
Concen: 63.827 ng
RT: 13.198 min Scan# 1719
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD

Tgt Ion:172 Resp: 915789




Ion	Ratio	Lower	Upper
172	100		
171	34.5	27.9	41.9
170	23.3	18.6	27.8




#46
1,1'-Biphenyl
Concen: 40.414 ng
RT: 13.404 min Scan# 1754
Delta R.T. -0.012 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:154 Resp: 613921

Ion	Ratio	Lower	Upper
154	100		
153	41.6	20.4	60.4
76	10.5	0.0	30.8









#47
2-Chloronaphthalene
Concen: 39.835 ng
RT: 13.451 min Scan# 1762
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:162 Resp: 494134
Ion Ratio Lower Upper
162 100
127 39.9 31.4 47.0
164 32.5 26.6 39.8



#48
2-Nitroaniline
Concen: 41.706 ng
RT: 13.657 min Scan# 1797
Delta R.T. -0.000 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion: 65 Resp: 136024
Ion Ratio Lower Upper
65 100
92 73.8 57.7 86.5
138 121.0 92.1 138.1









#49
Acenaphthylene
Concen: 40.840 ng
RT: 14.292 min Scan# 1905
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:152 Resp: 777213
Ion Ratio Lower Upper
152 100
151 20.3 16.2 24.2
153 12.6 10.6 16.0



#50
Dimethylphthalate
Concen: 39.486 ng
RT: 14.028 min Scan# 1860
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:163 Resp: 611667
Ion Ratio Lower Upper
163 100
194 3.9 3.4 5.2
164 10.4 8.2 12.2









#51
2,6-Dinitrotoluene
Concen: 38.563 ng
RT: 14.145 min Scan# 1880
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:165 Resp: 135777
Ion Ratio Lower Upper
165 100
63 55.1 41.3 61.9
89 57.9 44.8 67.2



#52
Acenaphthene
Concen: 39.541 ng
RT: 14.634 min Scan# 1963
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:154 Resp: 456973
Ion Ratio Lower Upper
154 100
153 113.1 89.4 134.2
152 52.8 43.0 64.4









#53
3-Nitroaniline
Concen: 18.263 ng
RT: 14.481 min Scan# 1937
Delta R.T. -0.000 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:138 Resp: 62321
Ion Ratio Lower Upper
138 100
108 11.4 8.4 12.6
92 106.7 87.8 131.8



#54
2,4-Dinitrophenol
Concen: 33.426 ng
RT: 14.686 min Scan# 1972
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:184 Resp: 63703
Ion Ratio Lower Upper
184 100
63 61.1 47.0 70.6
154 69.7 52.3 78.5









#55
Dibenzenofuran
Concen: 38.503 ng
RT: 14.963 min Scan# 2019
Delta R.T. -0.012 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:168 Resp: 731785
Ion Ratio Lower Upper
168 100
139 40.6 31.0 46.6
169 13.3 10.8 16.2



#56
4-Nitrophenol
Concen: 82.702 ng
RT: 14.810 min Scan# 1993
Delta R.T. 0.023 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:139 Resp: 228757
Ion Ratio Lower Upper
139 100
109 90.4 72.3 112.3
65 90.1 72.5 112.5









#57
2,4-Dinitrotoluene
Concen: 40.278 ng
RT: 14.934 min Scan# 2014
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD

Tgt Ion:165 Resp: 189677




Ion	Ratio	Lower	Upper
165	100		
63	39.4	32.3	48.5
89	70.2	55.9	83.9
182	2.7	2.2	3.4




#58
Fluorene
Concen: 39.274 ng
RT: 15.610 min Scan# 2129
Delta R.T. -0.012 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:166 Resp: 581597

Ion	Ratio	Lower	Upper
166	100		
165	98.1	78.0	117.0
167	13.6	10.5	15.7









#59
2,3,4,6-Tetrachlorophenol
Concen: 38.664 ng m
RT: 15.198 min Scan# 2059
Delta R.T. -0.000 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:232 Resp: 166930
Ion Ratio Lower Upper
232 100
131 40.2 32.2 48.4
130 2.7 2.0 3.0
166 25.0 19.4 29.2



#60
Diethylphthalate
Concen: 40.040 ng
RT: 15.386 min Scan# 2091
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:149 Resp: 622307
Ion Ratio Lower Upper
149 100
177 21.6 17.5 26.3
150 12.8 9.9 14.9









#61
4-Chlorophenyl-phenylether
Concen: 38.560 ng
RT: 15.604 min Scan# 2128
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:	204	Resp:	296099
Ion	Ratio	Lower	Upper
204	100		
206	32.5	26.6	40.0
141	63.5	48.6	73.0



#62
4-Nitroaniline
Concen: 36.200 ng
RT: 15.639 min Scan# 2134
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:	138	Resp:	127152
Ion	Ratio	Lower	Upper
138	100		
92	52.0	31.2	71.2
108	100.1	78.6	118.6









#63
Azobenzene
Concen: 37.184 ng
RT: 15.898 min Scan# 2178
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD

Tgt Ion: 77 Resp: 522847




Ion	Ratio	Lower	Upper
77	100		
182	29.3	6.0	46.0
105	20.7	0.0	38.8
51	28.2	8.8	48.8




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.310 min Scan# 2418
Delta R.T. -0.012 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion: 188 Resp: 383510

Ion	Ratio	Lower	Upper
188	100		
94	7.1	5.6	8.4
80	7.8	6.3	9.5









#65
4,6-Dinitro-2-methylphenol
Concen: 21.492 ng
RT: 15.698 min Scan# 2144
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampled :
ETGI-357MSD


Tgt Ion:198	Resp:	60571
Ion Ratio	Lower	Upper
198	100	
51	32.5	14.0 54.0
105	40.4	19.5 59.5



#66
n-Nitrosodiphenylamine
Concen: 40.044 ng
RT: 15.822 min Scan# 2165
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:169	Resp:	509265
Ion Ratio	Lower	Upper
169	100	
168	66.1	54.4 81.6
167	36.3	29.0 43.6









#67
4-Bromophenyl-phenylether
Concen: 39.734 ng
RT: 16.498 min Scan# 2280
Delta R.T. -0.012 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:248	Resp:	195585
Ion Ratio	Lower	Upper
248	100	
250	99.4	79.8 119.8
141	80.0	57.8 86.6



#68
Hexachlorobenzene
Concen: 40.438 ng
RT: 16.616 min Scan# 2300
Delta R.T. -0.012 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:284	Resp:	238595
Ion Ratio	Lower	Upper
284	100	
142	32.5	26.9 40.3
249	29.7	24.5 36.7









#69
Atrazine
Concen: 48.836 ng
RT: 16.775 min Scan# 2327
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:200	Resp: 193231
Ion Ratio	Lower Upper
200 100	
173 25.8	4.1 44.1
215 47.9	27.7 67.7



#70
Pentachlorophenol
Concen: 77.333 ng
RT: 16.969 min Scan# 2360
Delta R.T. -0.000 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:266	Resp: 253705
Ion Ratio	Lower Upper
266 100	
268 64.7	51.6 77.4
264 61.3	50.3 75.5









#71
Phenanthrene
Concen: 40.741 ng
RT: 17.357 min Scan# 2426
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:178 Resp: 938316
Ion Ratio Lower Upper
178 100
176 18.9 15.1 22.7
179 15.1 12.4 18.6



#72
Anthracene
Concen: 42.147 ng
RT: 17.445 min Scan# 2441
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:178 Resp: 955350
Ion Ratio Lower Upper
178 100
176 18.7 14.9 22.3
179 15.3 12.2 18.2









#73
 Carbazole
 Concen: 39.387 ng
 RT: 17.716 min Scan# 2487
 Delta R.T. -0.006 min
 Lab File: BP005894.D
 Acq: 12 Jun 2021 18:13




Instrument :
 BNA_P
 ClientSampleId :
 ETGI-357MSD


Tgt Ion:	167	Resp:	857170
Ion	Ratio	Lower	Upper
167	100		
166	21.5	17.1	25.7
139	13.7	10.7	16.1



#74
 Di-n-butylphthalate
 Concen: 42.962 ng
 RT: 18.263 min Scan# 2580
 Delta R.T. -0.006 min
 Lab File: BP005894.D
 Acq: 12 Jun 2021 18:13

Tgt Ion:	149	Resp:	1083254
Ion	Ratio	Lower	Upper
149	100		
150	9.0	7.3	10.9
104	5.5	4.5	6.7









#75
Fluoranthene
Concen: 44.298 ng
RT: 19.316 min Scan# 2759
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD

Tgt Ion:202 Resp: 1239011




Ion	Ratio	Lower	Upper
202	100		
101	8.4	0.0	27.6
203	17.2	0.0	37.8




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.386 min Scan# 3111
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:240 Resp: 433030

Ion	Ratio	Lower	Upper
240	100		
120	8.6	6.5	9.7
236	25.3	20.6	31.0









#77
Benzidine
Concen: 93.188 ng
RT: 19.492 min Scan# 2789
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:184 Resp: 839672
Ion Ratio Lower Upper
184 100
185 14.1 11.0 16.4
183 11.9 9.5 14.3



#78
Pyrene
Concen: 42.097 ng
RT: 19.669 min Scan# 2819
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:202 Resp: 1272616
Ion Ratio Lower Upper
202 100
200 21.1 16.3 24.5
203 17.6 14.2 21.2









#79
 Terphenyl-d14
 Concen: 65.550 ng
 RT: 19.857 min Scan# 2851
 Delta R.T. -0.012 min
 Lab File: BP005894.D
 Acq: 12 Jun 2021 18:13




Instrument :
 BNA_P
 ClientSampleId :
 ETGI-357MSD


Tgt Ion:244 Resp: 1515912
 Ion Ratio Lower Upper
 244 100
 212 7.6 6.0 9.0
 122 8.8 7.1 10.7



#80
 Butylbenzylphthalate
 Concen: 43.174 ng
 RT: 20.533 min Scan# 2966
 Delta R.T. -0.006 min
 Lab File: BP005894.D
 Acq: 12 Jun 2021 18:13

Tgt Ion:149 Resp: 536962
 Ion Ratio Lower Upper
 149 100
 91 64.2 54.6 81.8
 206 23.1 16.7 25.1









#81
Benzo(a)anthracene
Concen: 40.126 ng
RT: 21.368 min Scan# 3108
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:228 Resp: 1254264
Ion Ratio Lower Upper
228 100
226 26.8 21.6 32.4
229 20.0 15.8 23.6



#82
3,3'-Dichlorobenzidine
Concen: 39.240 ng
RT: 21.304 min Scan# 3097
Delta R.T. -0.000 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:252 Resp: 424069
Ion Ratio Lower Upper
252 100
254 64.3 50.6 76.0
126 9.0 7.0 10.6









#83
Chrysene
Concen: 40.646 ng
RT: 21.421 min Scan# 3117
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:228 Resp: 1230602
Ion Ratio Lower Upper
228 100
226 30.1 23.8 35.6
229 19.8 15.4 23.2



#84
Bis(2-ethylhexyl)phthalate
Concen: 48.457 ng
RT: 21.286 min Scan# 3094
Delta R.T. -0.012 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:149 Resp: 883877
Ion Ratio Lower Upper
149 100
167 27.6 21.9 32.9
279 5.1 4.3 6.5









#85
Di-n-octyl phthalate
Concen: 43.253 ng
RT: 22.210 min Scan# 3251
Delta R.T. -0.012 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:149 Resp: 1411381
Ion Ratio Lower Upper
149 100
167 1.6 1.4 2.0
43 6.4 5.6 8.4



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.798 min Scan# 3521
Delta R.T. -0.018 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:264 Resp: 500947
Ion Ratio Lower Upper
264 100
260 23.8 18.6 28.0
265 21.8 17.6 26.4









#87
Indeno(1,2,3-cd)pyrene
Concen: 39.689 ng
RT: 26.351 min Scan# 3955
Delta R.T. -0.006 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:276 Resp: 1700227
Ion Ratio Lower Upper
276 100
138 19.3 15.3 22.9
277 25.6 20.4 30.6



#88
Benzo(b)fluoranthene
Concen: 41.579 ng
RT: 23.063 min Scan# 3396
Delta R.T. -0.012 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:252 Resp: 1426584
Ion Ratio Lower Upper
252 100
253 21.7 17.5 26.3
125 8.1 6.2 9.4









#89
Benzo(k)fluoranthene
Concen: 39.633 ng
RT: 23.110 min Scan# 3404
Delta R.T. -0.012 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13




Instrument :
BNA_P
ClientSampleId :
ETGI-357MSD


Tgt Ion:252 Resp: 1318600
Ion Ratio Lower Upper
252 100
253 22.2 17.6 26.4
125 8.1 6.0 9.0



#90
Benzo(a)pyrene
Concen: 42.493 ng
RT: 23.692 min Scan# 3503
Delta R.T. -0.012 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:252 Resp: 1370693
Ion Ratio Lower Upper
252 100
253 22.0 17.9 26.9
125 9.1 6.5 9.7









#91
Dibenzo(a,h)anthracene
Concen: 38.864 ng
RT: 26.362 min Scan# 3957
Delta R.T. -0.018 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Instrument :
BNA_P
ClientSampled :
ETGI-357MSD

Tgt Ion:278 Resp: 1432943
Ion Ratio Lower Upper
278 100
139 12.6 9.3 13.9
279 23.9 18.8 28.2



#92
Benzo(g,h,i)perylene
Concen: 39.376 ng
RT: 27.127 min Scan# 4087
Delta R.T. -0.024 min
Lab File: BP005894.D
Acq: 12 Jun 2021 18:13

Tgt Ion:276 Resp: 1434206
Ion Ratio Lower Upper
276 100
277 23.9 18.9 28.3
138 17.1 12.6 18.8

