

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061423\
 Data File : BP015522.D
 Acq On : 14 Jun 2023 17:45
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020750

Quant Time: Jun 14 23:10:27 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 09 23:33:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.087	152	156287	20.000	ng/u1	0.00
20) Naphthalene-d8	10.916	136	729393	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.733	164	495375	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.492	188	1154935	20.000	ng/u1	0.00
79) Chrysene-d12	21.574	240	1132051	20.000	ng/u1	0.00
88) Perylene-d12	24.121	264	1208762	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.405	96	33458	7.927	ng/uL	0.00
4) Pyridine-d5	3.846	84	233292	21.280	ng/u1	0.00
7) Phenol-d5	7.240	99	288589	20.043	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.410	67	178903	20.805	ng/u1	0.00
11) 2-Chlorophenol-d4	7.616	132	211798	20.289	ng/u1	0.00
15) 4-Methylphenol-d8	8.804	113	229817	19.448	ng/u1	0.00
21) Nitrobenzene-d5	9.269	128	115731	20.210	ng/u1	0.00
24) 2-Nitrophenol-d4	9.993	143	134479	20.564	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.534	165	248946	20.869	ng/u1	0.00
31) 4-Chloroaniline-d4	11.057	131	347386	20.288	ng/u1	0.00
46) Dimethylphthalate-d6	14.139	166	796899	20.400	ng/u1	0.00
49) Acenaphthylene-d8	14.428	160	883643	20.687	ng/u1	0.00
54) 4-Nitrophenol-d4	14.939	143	122007	17.441	ng/u1	0.00
60) Fluorene-d10	15.728	176	683451	20.748	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.851	200	135174	18.475	ng/u1	0.00
73) Anthracene-d10	17.586	188	1093704	20.813	ng/u1	0.00
81) Pyrene-d10	19.810	212	1311935	21.653	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.951	264	1243177	20.497	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.440	88	34466	7.731	ng/uL	92
5) Pyridine	3.863	79	237477	20.895	ng/u1	91
6) Benzaldehyde	7.222	77	98167	17.838	ng/u1	94
8) Phenol	7.269	94	303434	20.427	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.504	93	240410	20.470	ng/u1	98
12) 2-Chlorophenol	7.646	128	226079	20.498	ng/u1	98
13) 2-Methylphenol	8.534	108	224556	19.648	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.628	45	318367	20.456	ng/u1	97
16) Acetophenone	8.922	105	385546	20.444	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.904	70	204442	20.086	ng/u1	97
18) 4-Methylphenol	8.869	108	250149	19.893	ng/u1	98
19) Hexachloroethane	9.175	117	94295	20.232	ng/u1	95
22) Nitrobenzene	9.310	77	316645	20.851	ng/u1	99
23) Isophorone	9.834	82	607704	19.990	ng/u1	98
25) 2-Nitrophenol	10.028	139	144897	20.522	ng/u1	100
26) 2,4-Dimethylphenol	10.081	107	302484	20.201	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.316	93	355647	20.122	ng/u1	98
29) 2,4-Dichlorophenol	10.563	162	248245	20.930	ng/u1	95
30) Naphthalene	10.969	128	796741	20.158	ng/u1	100
32) 4-Chloroaniline	11.081	127	361969	20.430	ng/u1	99
33) Hexachlorobutadiene	11.240	225	164449	20.559	ng/u1	98
34) Caprolactam	11.863	113	82161	18.401	ng/u1	95
35) 4-Chloro-3-methylphenol	12.198	107	290741	20.231	ng/u1	99
36) 2-Methylnaphthalene	12.569	142	566362	20.383	ng/u1	97


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061423\
 Data File : BP015522.D
 Acq On : 14 Jun 2023 17:45
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020750

Quant Time: Jun 14 23:10:27 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 09 23:33:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.787	142	577105	20.629	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.934	216	323276	20.957	ng/ul	98
40) Hexachlorocyclopentadiene	12.904	237	134885	21.332	ng/ul	97
41) 2,4,6-Trichlorophenol	13.169	196	215212	20.962	ng/ul	99
42) 2,4,5-Trichlorophenol	13.245	196	227442	20.329	ng/ul	97
43) 1,1'-Biphenyl	13.569	154	789574	20.611	ng/ul	99
44) 2-Chloronaphthalene	13.610	162	618461	20.576	ng/ul	97
45) 2-Nitroaniline	13.822	65	202800	21.082	ng/ul	99
47) Dimethylphthalate	14.186	163	819327	20.314	ng/ul	99
48) 2,6-Dinitrotoluene	14.316	165	169784	20.547	ng/ul	99
50) Acenaphthylene	14.457	152	961446	20.326	ng/ul	100
51) 3-Nitroaniline	14.645	138	119127	17.464	ng/ul	95
52) Acenaphthene	14.798	153	670711	20.513	ng/ul	99
53) 2,4-Dinitrophenol	14.857	184	106561	21.438	ng/ul	95
55) 4-Nitrophenol	14.951	109	119149	17.658	ng/ul	95
56) Dibenzofuran	15.133	168	955877	20.626	ng/ul	99
57) 2,4-Dinitrotoluene	15.104	165	248640	20.458	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.357	232	215457	21.329	ng/ul	98
59) Diethylphthalate	15.545	149	853429	20.646	ng/ul	99
61) Fluorene	15.780	166	781131	20.677	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.769	204	392620	20.621	ng/ul	99
63) 4-Nitroaniline	15.810	138	108108	16.708	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.869	198	139405	18.682	ng/ul	93
67) N-Nitrosodiphenylamine	15.986	169	675879	20.628	ng/ul	99
68) 4-Bromophenyl-phenylether	16.669	248	253328	20.503	ng/ul	96
69) Hexachlorobenzene	16.786	284	297986	20.445	ng/ul	94
70) Atrazine	16.939	200	268259	20.503	ng/ul	99
71) Pentachlorophenol	17.139	266	165540	20.095	ng/ul	99
72) Phenanthrene	17.533	178	1276323	20.593	ng/ul	100
74) Anthracene	17.622	178	1292864	20.559	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.534	216	334708	20.998	ng/uL	100
76) Pentachlorobenzene	15.051	250	344609	20.649	ng/uL	99
77) Carbazole	17.892	167	1160591	20.497	ng/ul	99
78) Di-n-butylphthalate	18.422	149	1516642	21.111	ng/ul	100
80) Fluoranthene	19.486	202	1591673	21.566	ng/ul	100
82) Pyrene	19.839	202	1639411	21.471	ng/ul	100
83) Butylbenzylphthalate	20.692	149	737561	21.855	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.480	252	492066	18.399	ng/ul	99
85) Benzo(a)anthracene	21.557	228	1647283	20.815	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.451	149	1098651	22.025	ng/ul	99
87) Chrysene	21.610	228	1515896	20.265	ng/ul	99
89) Di-n-octyl phthalate	22.421	149	1921855	24.192	ng/ul	100
90) Benzo(b)fluoranthene	23.333	252	1611016	20.647	ng/ul	99
91) Benzo(k)fluoranthene	23.386	252	1599955	21.213	ng/ul	99
93) Benzo(a)pyrene	24.009	252	1371932	20.167	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.815	276	1619459	18.395	ng/ul	99
95) Dibenzo(a,h)anthracene	26.833	278	1332225	18.574	ng/ul	99
96) Benzo(g,h,i)perylene	27.645	276	1264344	17.427	ng/ul	98





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 7.731 ng/uL
RT: 3.440 min Scan# 41
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45




Instrument :
BNA_P
ClientSampleId :
SSTD020750


Tgt Ion:	88	Resp:	34466
Ion	Ratio	Lower	Upper
88	100		
43	33.2	23.9	35.9
58	71.0	51.4	77.2



#5
Pyridine
Concen: 20.895 ng/uL
RT: 3.863 min Scan# 115
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45

Tgt Ion:	79	Resp:	237477
Ion	Ratio	Lower	Upper
79	100		
52	84.1	59.5	89.3
51	39.0	29.8	44.6









#6
Benzaldehyde
Concen: 17.838 ng/ul
RT: 7.222 min Scan# 686
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45




Instrument :
BNA_P
ClientSampleId :
SSTD020750


Tgt Ion: 77 Resp: 98167
Ion Ratio Lower Upper
77 100
105 90.8 77.3 115.9
106 85.3 73.3 109.9



#8
Phenol
Concen: 20.427 ng/ul
RT: 7.269 min Scan# 694
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45

Tgt Ion: 94 Resp: 303434
Ion Ratio Lower Upper
94 100
65 29.9 23.5 35.3
66 42.6 34.2 51.4









#10
Bis(2-Chloroethyl)ether
Concen: 20.470 ng/ul
RT: 7.504 min Scan# 734
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45




Instrument :
BNA_P
ClientSampleId :
SSTD020750


Tgt Ion: 93 Resp: 240410
Ion Ratio Lower Upper
93 100
63 80.2 62.6 93.8
95 33.5 25.5 38.3



#12
2-Chlorophenol
Concen: 20.498 ng/ul
RT: 7.646 min Scan# 758
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45

Tgt Ion:128 Resp: 226079
Ion Ratio Lower Upper
128 100
64 51.8 40.2 60.4
130 31.6 24.5 36.7









#13
2-Methylphenol
Concen: 19.648 ng/ul
RT: 8.534 min Scan# 909
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45




Instrument :
BNA_P
ClientSampleId :
SSTD020750


Tgt Ion:108 Resp: 224556
Ion Ratio Lower Upper
108 100
107 86.5 71.2 106.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.456 ng/ul
RT: 8.628 min Scan# 925
Delta R.T. 0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45

Tgt Ion: 45 Resp: 318367
Ion Ratio Lower Upper
45 100
77 15.7 11.8 17.8
79 12.2 8.7 13.1









#16
Acetophenone
Concen: 20.444 ng/ul
RT: 8.922 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45




Instrument :
BNA_P
ClientSampleId :
SSTD020750


Tgt Ion:105 Resp: 385546
Ion Ratio Lower Upper
105 100
77 81.7 64.9 97.3
51 29.8 23.5 35.3



#17
N-Nitroso-di-n-propylamine
Concen: 20.086 ng/ul
RT: 8.904 min Scan# 972
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45

Tgt Ion: 70 Resp: 204442
Ion Ratio Lower Upper
70 100
42 61.0 46.6 69.8
101 9.8 7.8 11.6
130 19.5 15.9 23.9





#18

4-Methylphenol

Concen: 19.893 ng/ul

RT: 8.869 min Scan# 90

Delta R.T. -0.006 min

Lab File: BP015522.D

Acq: 14 Jun 2023 17:45

Instrument :

BNA_P

ClientSampleId :





SSTD020750

Tgt Ion:108 Resp: 250149

Ion Ratio Lower Upper

108 100

107 113.3 89.3 133.9



#19

Hexachloroethane

Concen: 20.232 ng/ul

RT: 9.175 min Scan# 1018

Delta R.T. -0.012 min

Lab File: BP015522.D

Acq: 14 Jun 2023 17:45




Tgt Ion:117 Resp: 94295


Ion Ratio Lower Upper

117 100

201 106.6 82.8 124.2

199 70.0 50.6 76.0









#22
Nitrobenzene
Concen: 20.851 ng/ul
RT: 9.310 min Scan# 1041
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45




Instrument :
BNA_P
ClientSampleId :
SSTD020750


Tgt Ion: 77 Resp: 316645
Ion Ratio Lower Upper
77 100
123 41.4 33.5 50.3
65 13.8 11.6 17.4



#23
Isophorone
Concen: 19.990 ng/ul
RT: 9.834 min Scan# 1130
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45

Tgt Ion: 82 Resp: 607704
Ion Ratio Lower Upper
82 100
95 7.6 5.8 8.8
138 15.4 13.1 19.7









#25
2-Nitrophenol
Concen: 20.522 ng/ul
RT: 10.028 min Scan# 1163
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45




Instrument :
BNA_P
ClientSampleId :
SSTD020750


Tgt Ion:139 Resp: 144897
Ion Ratio Lower Upper
139 100
65 58.6 46.9 70.3
109 38.0 30.2 45.2



#26
2,4-Dimethylphenol
Concen: 20.201 ng/ul
RT: 10.081 min Scan# 1172
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45

Tgt Ion:107 Resp: 302484
Ion Ratio Lower Upper
107 100
121 46.5 39.0 58.6
122 79.0 64.8 97.2









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.122 ng/ul
 RT: 10.316 min Scan# 1212
 Delta R.T. -0.006 min
 Lab File: BP015522.D
 Acq: 14 Jun 2023 17:45

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020750

Tgt Ion: 93 Resp: 355647




Ion	Ratio	Lower	Upper
93	100		
95	33.2	25.7	38.5
123	10.9	8.6	13.0




#29
 2,4-Dichlorophenol
 Concen: 20.930 ng/ul
 RT: 10.563 min Scan# 1254
 Delta R.T. -0.006 min
 Lab File: BP015522.D
 Acq: 14 Jun 2023 17:45

Tgt Ion:162 Resp: 248245

Ion	Ratio	Lower	Upper
162	100		
164	68.9	50.6	76.0
98	39.3	30.2	45.2





#30

Naphthalene

Concen: 20.158 ng/ul

RT: 10.969 min Scan# 1323

Delta R.T. -0.006 min

Lab File: BP015522.D

Acq: 14 Jun 2023 17:45

Instrument :

BNA_P

ClientSampleId :

SSTD020750





Tgt Ion:128 Resp: 796741

Ion Ratio Lower Upper

128 100

129 11.1 8.6 13.0

127 13.1 10.5 15.7



#32

4-Chloroaniline

Concen: 20.430 ng/ul

RT: 11.081 min Scan# 1342

Delta R.T. -0.006 min

Lab File: BP015522.D




Acq: 14 Jun 2023 17:45


Tgt Ion:127 Resp: 361969

Ion Ratio Lower Upper

127 100

129 32.3 26.1 39.1





#33

Hexachlorobutadiene

Concen: 20.559 ng/ul

RT: 11.240 min Scan# 1

Delta R.T. -0.012 min

Lab File: BP015522.D

Acq: 14 Jun 2023 17:45

Instrument :

BNA_P

ClientSampleId :

SSTD020750





Tgt Ion:225 Resp: 164449

Ion Ratio Lower Upper

225 100

223 61.9 52.0 78.0

227 62.9 50.6 75.8



#34

Caprolactam

Concen: 18.401 ng/ul

RT: 11.863 min Scan# 1475

Delta R.T. -0.006 min

Lab File: BP015522.D

Acq: 14 Jun 2023 17:45




Tgt Ion:113 Resp: 82161


Ion Ratio Lower Upper

113 100

55 167.9 127.9 191.9

56 130.9 101.1 151.7









#35
4-Chloro-3-methylphenol
Concen: 20.231 ng/ul
RT: 12.198 min Scan# 1532
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45




Instrument :
BNA_P
ClientSampleId :
SSTD020750


Tgt Ion:107 Resp: 290741
Ion Ratio Lower Upper
107 100
144 24.0 19.4 29.0
142 74.8 60.4 90.6



#36
2-Methylnaphthalene
Concen: 20.383 ng/ul
RT: 12.569 min Scan# 1595
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45

Tgt Ion:142 Resp: 566362
Ion Ratio Lower Upper
142 100
141 87.1 72.0 108.0









#37
1-Methylnaphthalene
Concen: 20.629 ng/ul
RT: 12.787 min Scan# 1632
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45




Instrument :
BNA_P
ClientSampleId :
SSTD020750


Tgt Ion:142 Resp: 577105
Ion Ratio Lower Upper
142 100
141 92.7 71.9 107.9
116 4.3 3.4 5.2



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.957 ng/ul
RT: 12.934 min Scan# 1657
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45

Tgt Ion:216 Resp: 323276
Ion Ratio Lower Upper
216 100
214 79.1 62.2 93.2
179 21.2 17.4 26.2
108 17.2 14.2 21.4









#40
Hexachlorocyclopentadiene
Concen: 21.332 ng/ul
RT: 12.904 min Scan# 1652
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45




Instrument :
BNA_P
ClientSampleId :
SSTD020750


Tgt Ion:237 Resp: 134885
Ion Ratio Lower Upper
237 100
235 62.7 48.2 72.2
272 12.1 10.3 15.5



#41
2,4,6-Trichlorophenol
Concen: 20.962 ng/ul
RT: 13.169 min Scan# 1697
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45

Tgt Ion:196 Resp: 215212
Ion Ratio Lower Upper
196 100
198 96.2 77.8 116.6
200 30.7 25.8 38.6





#42

2,4,5-Trichlorophenol

Concen: 20.329 ng/ul

RT: 13.245 min Scan# 1710

Delta R.T. -0.006 min

Lab File: BP015522.D

Acq: 14 Jun 2023 17:45

Instrument :

BNA_P

ClientSampleId :

SSTD020750





Tgt Ion:196 Resp: 227442

Ion Ratio Lower Upper

196 100

198 98.3 75.5 113.3

200 31.1 24.7 37.1



#43

1,1'-Biphenyl

Concen: 20.611 ng/ul

RT: 13.569 min Scan# 1765

Delta R.T. -0.006 min

Lab File: BP015522.D

Acq: 14 Jun 2023 17:45




Tgt Ion:154 Resp: 789574


Ion Ratio Lower Upper

154 100

153 39.4 32.2 48.2

76 12.7 10.4 15.6





#44

2-Chloronaphthalene

Concen: 20.576 ng/ul

RT: 13.610 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP015522.D

Acq: 14 Jun 2023 17:45

Instrument :

BNA_P

ClientSampleId :

SSTD020750





Tgt Ion:162 Resp: 618461

Ion Ratio Lower Upper

162 100

164 32.4 27.0 40.6

127 42.2 32.2 48.4



#45

2-Nitroaniline

Concen: 21.082 ng/ul

RT: 13.822 min Scan# 1808

Delta R.T. -0.006 min

Lab File: BP015522.D

Acq: 14 Jun 2023 17:45




Tgt Ion: 65 Resp: 202800


Ion Ratio Lower Upper

65 100

92 68.0 55.4 83.2

138 98.5 78.8 118.2









#47
Dimethylphthalate
Concen: 20.314 ng/ul
RT: 14.186 min Scan# 1870
Delta R.T. -0.000 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45




Instrument :
BNA_P
ClientSampleId :
SSTD020750


Tgt Ion:163 Resp: 819327
Ion Ratio Lower Upper
163 100
194 3.9 3.2 4.8
164 9.8 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 20.547 ng/ul
RT: 14.316 min Scan# 1892
Delta R.T. -0.000 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45

Tgt Ion:165 Resp: 169784
Ion Ratio Lower Upper
165 100
89 63.8 51.5 77.3
121 21.4 17.3 25.9









#50
Acenaphthylene
Concen: 20.326 ng/ul
RT: 14.457 min Scan# 1916
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45




Instrument :
BNA_P
ClientSampleId :
SSTD020750


Tgt Ion:152 Resp: 961446
Ion Ratio Lower Upper
152 100
151 19.8 16.0 24.0
153 13.2 10.6 15.8



#51
3-Nitroaniline
Concen: 17.464 ng/ul
RT: 14.645 min Scan# 1948
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45

Tgt Ion:138 Resp: 119127
Ion Ratio Lower Upper
138 100
108 12.2 8.4 12.6
92 128.1 98.0 147.0









#52
Acenaphthene
Concen: 20.513 ng/ul
RT: 14.798 min Scan# 1974
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45




Instrument :
BNA_P
ClientSampleId :
SSTD020750


Tgt Ion:153 Resp: 670711
Ion Ratio Lower Upper
153 100
152 47.7 38.3 57.5
154 88.1 69.8 104.6



#53
2,4-Dinitrophenol
Concen: 21.438 ng/ul
RT: 14.857 min Scan# 1984
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45

Tgt Ion:184 Resp: 106561
Ion Ratio Lower Upper
184 100
63 80.7 59.0 88.4
154 73.7 57.6 86.4









#55
4-Nitrophenol
Concen: 17.658 ng/ul
RT: 14.951 min Scan# 2000
Delta R.T. -0.000 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45




Instrument :
BNA_P
ClientSampleId :
SSTD020750


Tgt Ion:109 Resp: 119149
Ion Ratio Lower Upper
109 100
139 108.5 89.1 133.7
65 120.7 103.1 154.7



#56
Dibenzofuran
Concen: 20.626 ng/ul
RT: 15.133 min Scan# 2031
Delta R.T. -0.000 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45

Tgt Ion:168 Resp: 955877
Ion Ratio Lower Upper
168 100
139 39.5 31.2 46.8









#57
2,4-Dinitrotoluene
Concen: 20.458 ng/ul
RT: 15.104 min Scan# 2026
Delta R.T. -0.000 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45




Instrument :
BNA_P
ClientSampleId :
SSTD020750


Tgt Ion:165 Resp: 248640
Ion Ratio Lower Upper
165 100
63 48.7 39.1 58.7
182 2.4 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.329 ng/ul
RT: 15.357 min Scan# 2069
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45

Tgt Ion:232 Resp: 215457
Ion Ratio Lower Upper
232 100
131 44.5 37.8 56.8
130 2.6 1.9 2.9
166 32.6 26.2 39.2









#59
Diethylphthalate
Concen: 20.646 ng/ul
RT: 15.545 min Scan# 2101
Delta R.T. -0.000 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45




Instrument :
BNA_P
ClientSampleId :
SSTD020750


Tgt Ion:149 Resp: 853429
Ion Ratio Lower Upper
149 100
177 21.6 17.8 26.8
150 12.3 10.1 15.1



#61
Fluorene
Concen: 20.677 ng/ul
RT: 15.780 min Scan# 2141
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45

Tgt Ion:166 Resp: 781131
Ion Ratio Lower Upper
166 100
165 98.0 79.1 118.7
167 13.9 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 20.621 ng/ul
RT: 15.769 min Scan# 2139
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45




Instrument :
BNA_P
ClientSampleId :
SSTD020750


Tgt Ion:204 Resp: 392620
Ion Ratio Lower Upper
204 100
206 32.0 26.6 40.0
141 61.8 49.8 74.8



#63
4-Nitroaniline
Concen: 16.708 ng/ul
RT: 15.810 min Scan# 2146
Delta R.T. -0.000 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45

Tgt Ion:138 Resp: 108108
Ion Ratio Lower Upper
138 100
92 55.4 44.6 67.0
108 98.4 78.4 117.6









#66
4,6-Dinitro-2-methylphenol
Concen: 18.682 ng/ul
RT: 15.869 min Scan# 2156
Delta R.T. -0.000 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45




Instrument :
BNA_P
ClientSampleId :
SSTD020750


Tgt Ion:198 Resp: 139405
Ion Ratio Lower Upper
198 100
182 3.8 3.0 4.4
77 29.6 20.5 30.7



#67
N-Nitrosodiphenylamine
Concen: 20.628 ng/ul
RT: 15.986 min Scan# 2176
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45

Tgt Ion:169 Resp: 675879
Ion Ratio Lower Upper
169 100
168 65.6 53.0 79.4
167 36.2 28.9 43.3









#68
4-Bromophenyl-phenylether
Concen: 20.503 ng/ul
RT: 16.669 min Scan# 211
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45




Instrument :
BNA_P
ClientSampleId :
SSTD020750


Tgt Ion:248 Resp: 253328
Ion Ratio Lower Upper
248 100
250 100.8 78.3 117.5
141 80.0 61.1 91.7



#69
Hexachlorobenzene
Concen: 20.445 ng/ul
RT: 16.786 min Scan# 2312
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45

Tgt Ion:284 Resp: 297986
Ion Ratio Lower Upper
284 100
142 37.0 25.4 38.2
249 32.5 24.8 37.2





#70

Atrazine

Concen: 20.503 ng/ul

RT: 16.939 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP015522.D

Acq: 14 Jun 2023 17:45

Instrument :

BNA_P

ClientSampleId :

SSTD020750





Tgt Ion:200 Resp: 268259

Ion Ratio Lower Upper

200 100

173 25.8 19.8 29.6

215 46.4 37.4 56.0



#71

Pentachlorophenol

Concen: 20.095 ng/ul

RT: 17.139 min Scan# 2372

Delta R.T. -0.000 min

Lab File: BP015522.D

Acq: 14 Jun 2023 17:45




Tgt Ion:266 Resp: 165540


Ion Ratio Lower Upper

266 100

264 62.6 49.6 74.4

268 63.0 51.0 76.6





#72

Phenanthrene

Concen: 20.593 ng/ul

RT: 17.533 min Scan# 2439

Delta R.T. -0.006 min

Lab File: BP015522.D

Acq: 14 Jun 2023 17:45

Instrument :

BNA_P

ClientSampleId :

SSTD020750





Tgt Ion:178 Resp: 1276323

Ion Ratio Lower Upper

178 100

179 15.2 12.3 18.5

176 19.2 15.4 23.0



#74

Anthracene

Concen: 20.559 ng/ul

RT: 17.622 min Scan# 2454

Delta R.T. -0.006 min

Lab File: BP015522.D

Acq: 14 Jun 2023 17:45




Tgt Ion:178 Resp: 1292864


Ion Ratio Lower Upper

178 100

179 15.7 12.4 18.6

176 18.4 15.0 22.4





#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.998 ng/uL

RT: 13.534 min Scan# 1760

Delta R.T. -0.006 min

Lab File: BP015522.D

Acq: 14 Jun 2023 17:45

Instrument :

BNA_P

ClientSampleId :

SSTD020750





Tgt Ion:216 Resp: 334708

Ion Ratio Lower Upper

216 100

214 78.0 62.1 93.1

218 48.6 39.0 58.4



#76

Pentachlorobenzene

Concen: 20.649 ng/uL

RT: 15.051 min Scan# 2017

Delta R.T. -0.006 min

Lab File: BP015522.D

Acq: 14 Jun 2023 17:45




Tgt Ion:250 Resp: 344609


Ion Ratio Lower Upper

250 100

248 61.7 49.0 73.4

252 63.6 49.9 74.9









#77
 Carbazole
 Concen: 20.497 ng/ul
 RT: 17.892 min Scan# 2500
 Delta R.T. -0.006 min
 Lab File: BP015522.D
 Acq: 14 Jun 2023 17:45




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020750


Tgt Ion:167 Resp: 1160591
 Ion Ratio Lower Upper
 167 100
 166 21.7 17.2 25.8
 139 13.6 10.6 16.0



#78
 Di-n-butylphthalate
 Concen: 21.111 ng/ul
 RT: 18.422 min Scan# 2590
 Delta R.T. -0.006 min
 Lab File: BP015522.D
 Acq: 14 Jun 2023 17:45

Tgt Ion:149 Resp: 1516642
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.1 10.7
 104 5.8 4.5 6.7





#80

Fluoranthene

Concen: 21.566 ng/ul

RT: 19.486 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP015522.D

Acq: 14 Jun 2023 17:45

Instrument :

BNA_P

ClientSampleId :

SSTD020750





Tgt Ion:202 Resp: 1591673

Ion Ratio Lower Upper

202 100

101 10.8 8.6 13.0

100 8.3 6.6 9.8



#82

Pyrene

Concen: 21.471 ng/ul

RT: 19.839 min Scan# 2831

Delta R.T. -0.006 min

Lab File: BP015522.D

Acq: 14 Jun 2023 17:45




Tgt Ion:202 Resp: 1639411


Ion Ratio Lower Upper

202 100

101 12.6 9.9 14.9

100 10.5 8.4 12.6





#83

Butylbenzylphthalate

Concen: 21.855 ng/ul

RT: 20.692 min Scan# 2977

Delta R.T. -0.006 min

Lab File: BP015522.D

Acq: 14 Jun 2023 17:45

Instrument :

BNA_P

ClientSampleId :

SSTD020750





Tgt Ion:149 Resp: 737561

Ion Ratio Lower Upper

149 100

91 77.2 60.4 90.6

206 19.8 16.4 24.6



#84

3,3'-Dichlorobenzidine

Concen: 18.399 ng/ul

RT: 21.480 min Scan# 3110

Delta R.T. -0.006 min

Lab File: BP015522.D

Acq: 14 Jun 2023 17:45




Tgt Ion:252 Resp: 492066


Ion Ratio Lower Upper

252 100

254 63.7 51.6 77.4

126 11.1 8.5 12.7





#85

Benzo(a)anthracene

Concen: 20.815 ng/ul

RT: 21.557 min Scan# 3123

Delta R.T. -0.000 min

Lab File: BP015522.D

Acq: 14 Jun 2023 17:45

Instrument :

BNA_P

ClientSampleId :

SSTD020750





Tgt Ion:228 Resp: 1647283

Ion Ratio Lower Upper

228 100

229 19.8 15.8 23.6

226 26.9 21.8 32.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 22.025 ng/ul

RT: 21.451 min Scan# 3105

Delta R.T. -0.000 min

Lab File: BP015522.D

Acq: 14 Jun 2023 17:45




Tgt Ion:149 Resp: 1098651


Ion Ratio Lower Upper

149 100

167 26.9 21.8 32.8

279 4.2 3.7 5.5





#87

Chrysene

Concen: 20.265 ng/ul

RT: 21.610 min Scan# 3132

Delta R.T. -0.006 min

Lab File: BP015522.D

Acq: 14 Jun 2023 17:45

Instrument :

BNA_P

ClientSampleId :

SSTD020750





Tgt Ion:228 Resp: 1515896

Ion Ratio Lower Upper

228 100

226 30.0 23.9 35.9

229 19.6 15.1 22.7



#89

Di-n-octyl phthalate

Concen: 24.192 ng/ul




RT: 22.421 min Scan# 3270


Delta R.T. -0.000 min

Lab File: BP015522.D

Acq: 14 Jun 2023 17:45

Tgt Ion:149 Resp: 1921855









#90
Benzo(b)fluoranthene
Concen: 20.647 ng/ul
RT: 23.333 min Scan# 3426
Delta R.T. -0.006 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45




Instrument :
BNA_P
ClientSampleId :
SSTD020750


Tgt Ion:252 Resp: 1611016
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.2
125 9.6 7.3 10.9



#91
Benzo(k)fluoranthene
Concen: 21.213 ng/ul
RT: 23.386 min Scan# 3434
Delta R.T. -0.000 min
Lab File: BP015522.D
Acq: 14 Jun 2023 17:45

Tgt Ion:252 Resp: 1599955
Ion Ratio Lower Upper
252 100
253 21.6 17.7 26.5
125 9.3 7.3 10.9





#93

Benzo(a)pyrene

Concen: 20.167 ng/ul

RT: 24.009 min Scan# 3

Delta R.T. -0.000 min

Lab File: BP015522.D

Acq: 14 Jun 2023 17:45

Instrument :

BNA_P

ClientSampleId :

SSTD020750





Tgt Ion:252 Resp: 1371932

Ion Ratio Lower Upper

252 100

253 21.9 17.8 26.6

125 10.3 8.1 12.1



#94

Indeno(1,2,3-cd)pyrene

Concen: 18.395 ng/ul

RT: 26.815 min Scan# 4017

Delta R.T. -0.012 min

Lab File: BP015522.D

Acq: 14 Jun 2023 17:45




Tgt Ion:276 Resp: 1619459


Ion Ratio Lower Upper

276 100

138 22.0 17.3 25.9

277 25.0 20.2 30.2





#95

Dibenzo(a,h)anthracene

Concen: 18.574 ng/ul

RT: 26.833 min Scan# 4020

Delta R.T. -0.006 min

Lab File: BP015522.D

Acq: 14 Jun 2023 17:45

Instrument :

BNA_P

ClientSampleId :

SSTD020750





Tgt Ion:278 Resp: 1332225

Ion Ratio Lower Upper

278 100

139 14.4 11.4 17.0

279 23.0 18.7 28.1



#96

Benzo(g,h,i)perylene

Concen: 17.427 ng/ul

RT: 27.645 min Scan# 4158

Delta R.T. -0.012 min

Lab File: BP015522.D

Acq: 14 Jun 2023 17:45

Tgt Ion:276 Resp: 1264344

Ion Ratio Lower Upper

276 100

138 18.6 15.6 23.4

277 22.8 19.0 28.6

