

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061423\
 Data File : BP015542.D
 Acq On : 15 Jun 2023 05:31
 Operator : MA/JU
 Sample : PB153338BL
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK338

Quant Time: Jun 15 06:02:42 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jun 15 05:36:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.087	152	126235	20.000	ng/u1	0.00
20) Naphthalene-d8	10.910	136	602218	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.728	164	430839	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.486	188	1005256	20.000	ng/u1	0.00
79) Chrysene-d12	21.569	240	941681	20.000	ng/u1	0.00
88) Perylene-d12	24.115	264	940466	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.405	96	20787	6.098	ng/uL	0.00
4) Pyridine-d5	3.846	84	247540	27.955	ng/u1	0.00
7) Phenol-d5	7.240	99	371890	31.978	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.405	67	240414	34.615	ng/u1	0.00
11) 2-Chlorophenol-d4	7.610	132	293182	34.771	ng/u1	0.00
15) 4-Methylphenol-d8	8.799	113	322362	33.774	ng/u1	0.00
21) Nitrobenzene-d5	9.263	128	155537	32.897	ng/u1	0.00
24) 2-Nitrophenol-d4	9.993	143	185652	34.384	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.534	165	313933	31.874	ng/u1	0.00
31) 4-Chloroaniline-d4	11.057	131	380235	26.895	ng/u1	0.00
46) Dimethylphthalate-d6	14.134	166	1193761	35.137	ng/u1	0.00
49) Acenaphthylene-d8	14.422	160	1282392	34.520	ng/u1	0.00
54) 4-Nitrophenol-d4	14.940	143	142786	23.469	ng/u1	0.00
60) Fluorene-d10	15.722	176	984717	34.371	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.851	200	181566	28.511	ng/u1	0.00
73) Anthracene-d10	17.586	188	1625619	35.542	ng/u1	0.00
81) Pyrene-d10	19.810	212	1979996	39.285	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.951	264	1739755	36.868	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061423\
Data File : BP015542.D
Acq On : 15 Jun 2023 05:31
Operator : MA/JU
Sample : PB153338BL
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK338

Quant Time: Jun 15 06:02:42 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jun 15 05:36:46 2023
Response via : Initial Calibration

