




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061423\
 Data File : BP015553.D
 Acq On : 15 Jun 2023 13:37
 Operator : MA/JU
 Sample : 03088-10
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DCFR6

Manual IntegrationsAPPROVED

Quant Time: Jun 15 22:04:56 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jun 15 05:36:46 2023
 Response via : Initial Calibration

Reviewed By :Yogesh Patel 06/16/2023
 Supervised By :mohammad ahmed 06/16/2023



TIC: BP015553.D\data.ms

(54) 4-Nitrophenol-d4 (S)

14.916min (-0.018) 0.02 ng/u1

response 92

Ion	Exp%	Act%
143.00	100.00	100.00
113.00	86.60	48.31#
41.00	42.20	401.69#
42.00	29.40	152.54#




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061423\
 Data File : BP015553.D
 Acq On : 15 Jun 2023 13:37
 Operator : MA/JU
 Sample : 03088-10
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DCFR6

Manual IntegrationsAPPROVED

Quant Time: Jun 15 22:04:56 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jun 15 05:36:46 2023
 Response via : Initial Calibration

Reviewed By :Yogesh Patel 06/16/2023
 Supervised By :mohammad ahmed 06/16/2023



TIC: BP015553.D\data.ms

(54) 4-Nitrophenol-d4 (S)

14.963min (+ 0.029) 6.34 ng/ul m

response 36392

Ion	Exp%	Act%
143.00	100.00	100.00
113.00	86.60	79.88
41.00	42.20	53.03#
42.00	29.40	27.36

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061423\
 Data File : BP015553.D
 Acq On : 15 Jun 2023 13:37
 Operator : MA/JU
 Sample : 03088-10
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DCFR6

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 06/16/2023
 Supervised By :mohammad ahmed 06/16/2023

Quant Time: Jun 15 22:46:18 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jun 15 05:36:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.087	152	119092	20.000	ng/ul	0.00
20) Naphthalene-d8	10.910	136	580286	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.728	164	406730	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.486	188	953026	20.000	ng/ul	0.00
79) Chrysene-d12	21.569	240	927426	20.000	ng/ul	0.00
88) Perylene-d12	24.116	264	965137	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.405	96	5280	1.642	ng/uL	0.00
4) Pyridine-d5	3.875	84	13837	1.656	ng/ul	0.03
7) Phenol-d5	7.246	99	114723	10.456	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.411	67	73264	11.181	ng/ul	0.00
11) 2-Chlorophenol-d4	7.616	132	94265	11.850	ng/ul	0.00
15) 4-Methylphenol-d8	8.805	113	71530	7.944	ng/ul	0.00
21) Nitrobenzene-d5	9.263	128	49227	10.805	ng/ul	0.00
24) 2-Nitrophenol-d4	9.993	143	58477	11.240	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.540	165	101227	10.666	ng/ul	0.00
31) 4-Chloroaniline-d4	11.063	131	89919	6.601	ng/ul	0.00
46) Dimethylphthalate-d6	14.134	166	382778	11.935	ng/ul	0.00
49) Acenaphthylene-d8	14.422	160	414263	11.812	ng/ul	0.00
54) 4-Nitrophenol-d4	14.963	143	36392m	6.336	ng/ul	0.03
60) Fluorene-d10	15.722	176	326580	12.075	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.857	200	44362	7.348	ng/ul	0.00
73) Anthracene-d10	17.587	188	503830	11.619	ng/ul	0.00
81) Pyrene-d10	19.810	212	620624	12.503	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.945	264	470753	9.721	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061423\
Data File : BP015553.D
Acq On : 15 Jun 2023 13:37
Operator : MA/JU
Sample : 03088-10
Misc :
ALS Vial : 47 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DCFR6

Manual IntegrationsAPPROVED

Quant Time: Jun 15 22:46:18 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jun 15 05:36:46 2023
Response via : Initial Calibration

Reviewed By :Yogesh Patel 06/16/2023
Supervised By :mohammad ahmed 06/16/2023

