

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061423\
 Data File : BP015568.D
 Acq On : 15 Jun 2023 22:06
 Operator : MA/JU
 Sample : SP6218
 Misc : SFAM SURROGATES
 ALS Vial : 62 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SP6218

Quant Time: Jun 15 23:29:31 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jun 15 05:36:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.087	152	118787	20.000	ng/ul	0.00
20) Naphthalene-d8	10.910	136	554437	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.728	164	381595	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.486	188	897345	20.000	ng/ul	0.00
79) Chrysene-d12	21.569	240	911501	20.000	ng/ul	0.00
88) Perylene-d12	24.110	264	970384	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.405	96	45612	14.219	ng/uL	0.00
4) Pyridine-d5	3.840	84	693519	83.231	ng/ul	0.00
7) Phenol-d5	7.240	99	796127	72.749	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.410	67	493872	75.566	ng/ul	0.00
11) 2-Chlorophenol-d4	7.610	132	620076	78.152	ng/ul	0.00
15) 4-Methylphenol-d8	8.804	113	702868	78.258	ng/ul	0.00
21) Nitrobenzene-d5	9.263	128	344464	79.135	ng/ul	0.00
24) 2-Nitrophenol-d4	9.993	143	404123	81.297	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.534	165	673483	74.272	ng/ul	0.00
31) 4-Chloroaniline-d4	11.051	131	991561	76.181	ng/ul	0.00
46) Dimethylphthalate-d6	14.139	166	2295849	76.297	ng/ul	0.00
49) Acenaphthylene-d8	14.428	160	2575295	78.269	ng/ul	0.00
54) 4-Nitrophenol-d4	14.934	143	372993	69.218	ng/ul	0.00
60) Fluorene-d10	15.722	176	1931623	76.123	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.851	200	385421	67.799	ng/ul	0.00
73) Anthracene-d10	17.586	188	3111413	76.208	ng/ul	0.00
81) Pyrene-d10	19.810	212	3784015	77.565	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.951	264	3797501	77.994	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061423\
Data File : BP015568.D
Acq On : 15 Jun 2023 22:06
Operator : MA/JU
Sample : SP6218
Misc : SFAM SURROGATES
ALS Vial : 62 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SP6218

Quant Time: Jun 15 23:29:31 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jun 15 05:36:46 2023
Response via : Initial Calibration

