

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061522\
 Data File : BP010730.D
 Acq On : 15 Jun 2022 14:45
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 16 00:02:24 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 16 00:00:39 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	85	0.00
2	1,4-Dioxane	0.466	0.415	10.9	75	0.00
3	Pyridine	1.317	1.218	7.5	76	0.00
4	n-Nitrosodimethylamine	0.735	0.861	-17.1	95	0.00
5 S	2-Fluorophenol	1.159	1.101	5.0	79	0.00
6	Aniline	1.853	1.740	6.1	77	0.00
7 S	Phenol-d6	1.585	1.511	4.7	78	0.00
8	2-Chlorophenol	1.258	1.226	2.5	81	0.00
9	Benzaldehyde	0.958	0.790	17.5	77	0.00
10 C	Phenol	1.635	1.532	6.3	76	0.00
11	bis(2-Chloroethyl)ether	1.256	1.160	7.6	78	0.00
12	1,3-Dichlorobenzene	1.440	1.444	-0.3	85	0.00
13 C	1,4-Dichlorobenzene	1.474	1.498	-1.6	84	0.00
14	1,2-Dichlorobenzene	1.419	1.427	-0.6	85	0.00
15	Benzyl Alcohol	1.257	1.229	2.2	80	0.00
16	2,2'-oxybis(1-Chloropropane	2.112	2.141	-1.4	85	0.00
17	2-Methylphenol	1.104	1.080	2.2	81	0.00
18	Hexachloroethane	0.572	0.580	-1.4	86	0.00
19 P	n-Nitroso-di-n-propylamine	1.118	1.091	2.4	81	0.00
20	3+4-Methylphenols	1.515	1.480	2.3	80	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	86	0.00
22	Acetophenone	0.533	0.524	1.7	83	0.00
23 S	Nitrobenzene-d5	0.425	0.435	-2.4	85	0.00
24	Nitrobenzene	0.429	0.426	0.7	83	0.00
25	Isophorone	0.727	0.706	2.9	81	0.00
26 C	2-Nitrophenol	0.176	0.181	-2.8	83	0.00
27	2,4-Dimethylphenol	0.258	0.256	0.8	82	0.00
28	bis(2-Chloroethoxy)methane	0.391	0.370	5.4	80	0.00
29 C	2,4-Dichlorophenol	0.304	0.322	-5.9	88	0.00
30	1,2,4-Trichlorobenzene	0.352	0.384	-9.1	92	0.00
31	Naphthalene	1.067	1.074	-0.7	84	0.00
32	Benzoic acid	0.190	0.195	-2.6	80	0.00
33	4-Chloroaniline	0.419	0.424	-1.2	83	0.00
34 C	Hexachlorobutadiene	0.239	0.283	-18.4	99	0.00
35	Caprolactam	0.104	0.100	3.8	80	0.00
36 C	4-Chloro-3-methylphenol	0.363	0.372	-2.5	85	0.00
37	2-Methylnaphthalene	0.744	0.784	-5.4	88	0.00
38	1-Methylnaphthalene	0.732	0.762	-4.1	87	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	92	0.00
40	1,2,4,5-Tetrachlorobenzene	0.615	0.665	-8.1	97	0.00
41 P	Hexachlorocyclopentadiene	0.255	0.318	-24.7	107	0.00
42 S	2,4,6-Tribromophenol	0.272	0.337	-23.9	110	0.00
43 C	2,4,6-Trichlorophenol	0.399	0.424	-6.3	93	0.00
44	2,4,5-Trichlorophenol	0.444	0.479	-7.9	94	0.00
45 S	2-Fluorobiphenyl	1.390	1.441	-3.7	94	0.00
46	1,1'-Biphenyl	1.475	1.474	0.1	91	0.00
47	2-Chloronaphthalene	1.154	1.176	-1.9	92	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061522\
 Data File : BP010730.D
 Acq On : 15 Jun 2022 14:45
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 16 00:02:24 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 16 00:00:39 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.428	0.428	0.0	88	0.00
49	Acenaphthylene	1.830	1.834	-0.2	90	0.00
50	Dimethylphthalate	1.520	1.515	0.3	91	0.00
51	2,6-Dinitrotoluene	0.324	0.327	-0.9	90	0.00
52 C	Acenaphthene	1.110	1.116	-0.5	91	0.00
53	3-Nitroaniline	0.332	0.325	2.1	84	0.00
54 P	2,4-Dinitrophenol	0.181	0.192	-6.1	94	0.00
55	Dibenzofuran	1.784	1.832	-2.7	93	0.00
56 P	4-Nitrophenol	0.248	0.242	2.4	85	0.00
57	2,4-Dinitrotoluene	0.451	0.465	-3.1	90	0.00
58	Fluorene	1.486	1.539	-3.6	94	0.00
59	2,3,4,6-Tetrachlorophenol	0.404	0.460	-13.9	101	0.00
60	Diethylphthalate	1.557	1.584	-1.7	93	0.00
61	4-Chlorophenyl-phenylether	0.754	0.811	-7.6	97	0.00
62	4-Nitroaniline	0.340	0.351	-3.2	89	0.00
63	Azobenzene	1.561	1.584	-1.5	91	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
65	4,6-Dinitro-2-methylphenol	0.126	0.132	-4.8	96	0.00
66 c	n-Nitrosodiphenylamine	0.584	0.586	-0.3	94	0.00
67	4-Bromophenyl-phenylether	0.218	0.241	-10.6	103	0.00
68	Hexachlorobenzene	0.252	0.283	-12.3	108	0.00
69	Atrazine	0.208	0.219	-5.3	100	0.00
70 C	Pentachlorophenol	0.137	0.156	-13.9	103	0.00
71	Phenanthrene	1.100	1.113	-1.2	95	0.00
72	Anthracene	1.063	1.077	-1.3	95	0.00
73	Carbazole	0.949	0.949	0.0	92	0.00
74	Di-n-butylphthalate	1.245	1.224	1.7	93	0.00
75 C	Fluoranthene	1.301	1.359	-4.5	97	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
77	Benzydine	0.391	0.395	-1.0	83	0.00
78	Pyrene	1.349	1.298	3.8	98	0.00
79 S	Terphenyl-d14	1.070	1.089	-1.8	104	0.00
80	Butylbenzylphthalate	0.578	0.526	9.0	92	0.00
81	Benzo(a)anthracene	1.330	1.330	0.0	100	0.00
82	3,3'-Dichlorobenzidine	0.388	0.389	-0.3	99	0.00
83	Chrysene	1.270	1.289	-1.5	104	0.00
84	Bis(2-ethylhexyl)phthalate	0.821	0.762	7.2	94	0.00
85 c	Di-n-octyl phthalate	1.380	1.272	7.8	93	0.00
86 I	Perylene-d12	1.000	1.000	0.0	102	0.00
87	Indeno(1,2,3-cd)pyrene	1.455	1.501	-3.2	102	0.00
88	Benzo(b)fluoranthene	1.248	1.270	-1.8	103	0.00
89	Benzo(k)fluoranthene	1.272	1.299	-2.1	101	0.00
90 C	Benzo(a)pyrene	1.045	1.063	-1.7	101	0.00
91	Dibenzo(a,h)anthracene	1.217	1.269	-4.3	104	0.00
92	Benzo(g,h,i)perylene	1.206	1.242	-3.0	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061522\
Data File : BP010730.D
Acq On : 15 Jun 2022 14:45
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jun 16 00:02:24 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 16 00:00:39 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0