

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061623\
 Data File : BP015574.D
 Acq On : 16 Jun 2023 11:17
 Operator : MA/JU
 Sample : 03220-04
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 CONR0

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/19/2023
 Supervised By :mohammad ahmed 06/20/2023

Quant Time: Jun 16 22:20:10 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jun 15 05:36:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.087	152	138632	20.000	ng/u1	0.00
20) Naphthalene-d8	10.910	136	634492	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.728	164	445087	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.487	188	1045930	20.000	ng/u1	0.00
79) Chrysene-d12	21.569	240	960433	20.000	ng/u1	0.00
88) Perylene-d12	24.116	264	975594	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.405	96	14924	3.986	ng/uL	0.00
4) Pyridine-d5	3.852	84	85740	8.817	ng/u1	0.00
7) Phenol-d5	7.246	99	62913	4.926	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.405	67	234561	30.752	ng/u1	0.00
11) 2-Chlorophenol-d4	7.611	132	207089	22.364	ng/u1	0.00
15) 4-Methylphenol-d8	8.799	113	139032	13.264	ng/u1	0.00
21) Nitrobenzene-d5	9.263	128	152252	30.564	ng/u1	0.00
24) 2-Nitrophenol-d4	9.993	143	166439	29.258	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.534	165	265277	25.564	ng/u1	0.00
31) 4-Chloroaniline-d4	11.058	131	367164	24.650	ng/u1	0.00
46) Dimethylphthalate-d6	14.134	166	1212892	34.557	ng/u1	0.00
49) Acenaphthylene-d8	14.422	160	1232890	32.125	ng/u1	0.00
54) 4-Nitrophenol-d4	14.987	143	15839m	2.520	ng/u1	0.05
60) Fluorene-d10	15.722	176	1013073	34.229	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.851	200	183374	27.675	ng/u1	0.00
73) Anthracene-d10	17.587	188	1784743	37.504	ng/u1	0.00
81) Pyrene-d10	19.810	212	2181745	42.443	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.951	264	1879416	38.394	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061623\
Data File : BP015574.D
Acq On : 16 Jun 2023 11:17
Operator : MA/JU
Sample : 03220-04
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
CONR0

Quant Time: Jun 16 22:20:10 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jun 15 05:36:46 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/19/2023
Supervised By :mohammad ahmed 06/20/2023

