

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP061720\
 Data File : BP002432.D
 Acq On : 17 Jun 2020 22:32
 Operator : CG/JU
 Sample : L3024-16
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 RO-6

Quant Time: Jun 18 01:30:10 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP060520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 05 15:08:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.59	152	140769	20.00	ng	0.00
21) Naphthalene-d8	10.36	136	585014	20.00	ng	0.00
39) Acenaphthene-d10	14.23	164	364492	20.00	ng	0.00
64) Phenanthrene-d10	16.99	188	841381	20.00	ng	-0.01
76) Chrysene-d12	21.09	240	861416	20.00	ng	-0.02
86) Perylene-d12	23.29	264	902296	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.20	112	721576	94.85	ng	0.00
7) Phenol-d6	6.78	99	987005	97.45	ng	0.00
23) Nitrobenzene-d5	8.73	82	625403	63.84	ng	-0.01
42) 2,4,6-Tribromophenol	15.73	330	334546	95.87	ng	0.00
45) 2-Fluorobiphenyl	12.85	172	1384374	63.86	ng	0.00
79) Terphenyl-d14	19.58	244	2283582	59.96	ng	0.00
Target Compounds						
50) Dimethylphthalate	13.70	163	130526	5.226	ng	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP061720\
Data File : BP002432.D
Acq On : 17 Jun 2020 22:32
Operator : CG/JU
Sample : L3024-16
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
RO-6

Quant Time: Jun 18 01:30:10 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP060520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 05 15:08:53 2020
Response via : Initial Calibration

