

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061821\
 Data File : BP005973.D
 Acq On : 18 Jun 2021 15:59
 Operator : CG/JU
 Sample : M2687-19
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DUP-02

Quant Time: Jun 18 16:30:29 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 16 11:35:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	70625	20.000	ng	0.00
21) Naphthalene-d8	10.728	136	246921	20.000	ng	0.00
39) Acenaphthene-d10	14.551	164	143402	20.000	ng	0.00
64) Phenanthrene-d10	17.298	188	312906	20.000	ng	0.00
76) Chrysene-d12	21.375	240	395868	20.000	ng	0.00
86) Perylene-d12	23.780	264	474070	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.499	112	576745	131.039	ng	0.00
7) Phenol-d6	7.099	99	594361	104.187	ng	0.00
23) Nitrobenzene-d5	9.093	82	454093	90.162	ng	0.00
42) 2,4,6-Tribromophenol	16.045	330	373608	151.818	ng	0.00
45) 2-Fluorobiphenyl	13.181	172	998515	91.434	ng	0.00
79) Terphenyl-d14	19.845	244	1969465	93.157	ng	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061821\
Data File : BP005973.D
Acq On : 18 Jun 2021 15:59
Operator : CG/JU
Sample : M2687-19
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DUP-02

Quant Time: Jun 18 16:30:29 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 16 11:35:30 2021
Response via : Initial Calibration

