

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061821\
 Data File : BP005970.D
 Acq On : 18 Jun 2021 14:14
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 18 14:50:25 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 16 11:35:30 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	83	0.00
2	1,4-Dioxane	40.000	40.467	-1.2	88	0.00
3	Pyridine	40.000	45.968	-14.9	95	0.00
4	n-Nitrosodimethylamine	40.000	39.477	1.3	86	0.00
5 S	2-Fluorophenol	80.000	86.146	-7.7	93	0.00
6	Aniline	40.000	43.809	-9.5	95	0.00
7 S	Phenol-d6	80.000	87.152	-8.9	94	0.00
8	2-Chlorophenol	40.000	43.714	-9.3	96	0.00
9	Benzaldehyde	40.000	46.560	-16.4	92	0.00
10 C	Phenol	40.000	43.731	-9.3	96	0.00
11	bis(2-Chloroethyl)ether	40.000	44.345	-10.9	97	0.00
12	1,3-Dichlorobenzene	40.000	43.342	-8.4	95	0.00
13 C	1,4-Dichlorobenzene	40.000	43.068	-7.7	95	0.00
14	1,2-Dichlorobenzene	40.000	43.307	-8.3	95	0.00
15	Benzyl Alcohol	40.000	42.526	-6.3	91	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.687	-4.2	91	-0.01
17	2-Methylphenol	40.000	44.565	-11.4	97	0.00
18	Hexachloroethane	40.000	42.898	-7.2	93	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	43.920	-9.8	95	0.00
20	3+4-Methylphenols	40.000	44.397	-11.0	96	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	83	0.00
22	Acetophenone	40.000	42.636	-6.6	93	0.00
23 S	Nitrobenzene-d5	80.000	82.778	-3.5	90	0.00
24	Nitrobenzene	40.000	41.181	-3.0	92	0.00
25	Isophorone	40.000	41.976	-4.9	93	0.00
26 C	2-Nitrophenol	40.000	42.985	-7.5	95	0.00
27	2,4-Dimethylphenol	40.000	42.106	-5.3	95	0.00
28	bis(2-Chloroethoxy)methane	40.000	43.364	-8.4	99	0.00
29 C	2,4-Dichlorophenol	40.000	41.438	-3.6	93	0.00
30	1,2,4-Trichlorobenzene	40.000	41.092	-2.7	94	0.00
31	Naphthalene	40.000	42.177	-5.4	96	0.00
32	Benzoic acid	40.000	37.998	5.0	82	0.00
33	4-Chloroaniline	40.000	42.654	-6.6	96	0.00
34 C	Hexachlorobutadiene	40.000	41.202	-3.0	95	0.00
35	Caprolactam	40.000	43.573	-8.9	92	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.881	-7.2	97	0.00
37	2-Methylnaphthalene	40.000	41.921	-4.8	96	0.00
38	1-Methylnaphthalene	40.000	42.316	-5.8	96	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	82	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.541	-3.9	90	0.00
41 P	Hexachlorocyclopentadiene	40.000	35.610	11.0	80	0.00
42 S	2,4,6-Tribromophenol	80.000	85.846	-7.3	93	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.836	-2.1	89	0.00
44	2,4,5-Trichlorophenol	40.000	40.828	-2.1	89	0.00
45 S	2-Fluorobiphenyl	80.000	83.480	-4.4	92	0.00
46	1,1'-Biphenyl	40.000	42.032	-5.1	91	0.00
47	2-Chloronaphthalene	40.000	41.855	-4.6	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061821\
 Data File : BP005970.D
 Acq On : 18 Jun 2021 14:14
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 18 14:50:25 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 16 11:35:30 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.514	-6.3	93	0.00
49	Acenaphthylene	40.000	42.288	-5.7	94	0.00
50	Dimethylphthalate	40.000	41.855	-4.6	94	0.00
51	2,6-Dinitrotoluene	40.000	41.688	-4.2	92	0.00
52 C	Acenaphthene	40.000	41.752	-4.4	94	0.00
53	3-Nitroaniline	40.000	43.417	-8.5	95	0.00
54 P	2,4-Dinitrophenol	40.000	41.699	-4.2	92	0.00
55	Dibenzofuran	40.000	40.983	-2.5	92	0.00
56 P	4-Nitrophenol	40.000	38.363	4.1	82	0.00
57	2,4-Dinitrotoluene	40.000	42.481	-6.2	92	0.00
58	Fluorene	40.000	41.371	-3.4	93	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.058	-2.6	90	0.00
60	Diethylphthalate	40.000	41.371	-3.4	92	0.00
61	4-Chlorophenyl-phenylether	40.000	41.165	-2.9	93	0.00
62	4-Nitroaniline	40.000	43.412	-8.5	95	0.00
63	Azobenzene	40.000	40.620	-1.5	90	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	79	0.00
65	4,6-Dinitro-2-methylphenol	40.000	37.040	7.4	78	0.00
66 c	n-Nitrosodiphenylamine	40.000	42.594	-6.5	93	0.00
67	4-Bromophenyl-phenylether	40.000	42.515	-6.3	92	0.00
68	Hexachlorobenzene	40.000	42.263	-5.7	92	0.00
69	Atrazine	40.000	36.081	9.8	67	0.00
70 C	Pentachlorophenol	40.000	41.127	-2.8	81	0.00
71	Phenanthrene	40.000	42.066	-5.2	91	-0.01
72	Anthracene	40.000	42.534	-6.3	92	0.00
73	Carbazole	40.000	43.119	-7.8	93	0.00
74	Di-n-butylphthalate	40.000	43.247	-8.1	93	0.00
75 C	Fluoranthene	40.000	43.008	-7.5	94	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	89	0.00
77	Benzydine	40.000	37.040	7.4	72	0.00
78	Pyrene	40.000	38.926	2.7	95	0.00
79 S	Terphenyl-d14	80.000	81.149	-1.4	98	0.00
80	Butylbenzylphthalate	40.000	41.879	-4.7	100	0.00
81	Benzo(a)anthracene	40.000	41.042	-2.6	101	0.00
82	3,3'-Dichlorobenzidine	40.000	42.134	-5.3	105	0.00
83	Chrysene	40.000	41.445	-3.6	102	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.643	-4.1	101	0.00
85 c	Di-n-octyl phthalate	40.000	41.989	-5.0	101	0.00
86 I	Perylene-d12	20.000	20.000	0.0	87	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	43.618	-9.0	104	-0.01
88	Benzo(b)fluoranthene	40.000	43.256	-8.1	102	-0.01
89	Benzo(k)fluoranthene	40.000	44.582	-11.5	106	0.00
90 C	Benzo(a)pyrene	40.000	43.659	-9.1	104	0.00
91	Dibenzo(a,h)anthracene	40.000	43.535	-8.8	104	-0.01
92	Benzo(g,h,i)perylene	40.000	42.727	-6.8	102	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061821\
Data File : BP005970.D
Acq On : 18 Jun 2021 14:14
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jun 18 14:50:25 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 16 11:35:30 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0