

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062023\
 Data File : BP015618.D
 Acq On : 20 Jun 2023 13:05
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020758

Quant Time: Jun 20 22:25:30 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jun 15 05:36:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.081	152	102288	20.000	ng/ul	0.00
20) Naphthalene-d8	10.910	136	458538	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.728	164	308047	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.486	188	746314	20.000	ng/ul	0.00
79) Chrysene-d12	21.574	240	736647	20.000	ng/ul	0.00
88) Perylene-d12	24.121	264	802685	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.405	96	19699	7.131	ng/uL	0.00
4) Pyridine-d5	3.846	84	143368	19.981	ng/ul	0.00
7) Phenol-d5	7.246	99	194296	20.618	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.405	67	113826	20.225	ng/ul	0.00
11) 2-Chlorophenol-d4	7.610	132	144334	21.125	ng/ul	0.00
15) 4-Methylphenol-d8	8.804	113	159924	20.678	ng/ul	0.00
21) Nitrobenzene-d5	9.263	128	78150	21.708	ng/ul	0.00
24) 2-Nitrophenol-d4	9.993	143	90363	21.980	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.540	165	162469	21.664	ng/ul	0.00
31) 4-Chloroaniline-d4	11.057	131	237955	22.105	ng/ul	0.00
46) Dimethylphthalate-d6	14.134	166	506954	20.870	ng/ul	0.00
49) Acenaphthylene-d8	14.428	160	591080	22.253	ng/ul	0.00
54) 4-Nitrophenol-d4	14.951	143	73757	16.955	ng/ul	0.02
60) Fluorene-d10	15.722	176	446013	21.774	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.857	200	89461	18.922	ng/ul	0.00
73) Anthracene-d10	17.586	188	744998	21.940	ng/ul	0.00
81) Pyrene-d10	19.810	212	927779	23.532	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.957	264	853934	21.202	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.440	88	21179	7.259	ng/uL	98
5) Pyridine	3.869	79	146951	19.755	ng/ul	99
6) Benzaldehyde	7.222	77	64284	17.847	ng/ul	98
8) Phenol	7.275	94	202728	20.853	ng/ul	98
10) Bis(2-Chloroethyl)ether	7.499	93	152646	19.859	ng/ul	99
12) 2-Chlorophenol	7.646	128	153731	21.296	ng/ul	99
13) 2-Methylphenol	8.540	108	155935	20.847	ng/ul	96
14) 2,2'-oxybis(1-Chloropr...	8.610	45	124725	12.245	ng/ul	97
16) Acetophenone	8.922	105	257084	20.829	ng/ul	99
17) N-Nitroso-di-n-propyla...	8.899	70	137592	20.654	ng/ul	98
18) 4-Methylphenol	8.869	108	171844	20.880	ng/ul	99
19) Hexachloroethane	9.169	117	63787	20.911	ng/ul	99
22) Nitrobenzene	9.304	77	209326	21.926	ng/ul	98
23) Isophorone	9.834	82	398692	20.862	ng/ul	99
25) 2-Nitrophenol	10.022	139	96083	21.646	ng/ul	99
26) 2,4-Dimethylphenol	10.081	107	204212	21.694	ng/ul	98
27) Bis(2-Chloroethoxy)met...	10.316	93	222315	20.009	ng/ul	98
29) 2,4-Dichlorophenol	10.563	162	162659	21.815	ng/ul	98
30) Naphthalene	10.963	128	533524	21.472	ng/ul	100
32) 4-Chloroaniline	11.081	127	241038	21.641	ng/ul	99
33) Hexachlorobutadiene	11.234	225	108086	21.494	ng/ul	99
34) Caprolactam	11.869	113	60310	21.486	ng/ul	94
35) 4-Chloro-3-methylphenol	12.204	107	199471	22.079	ng/ul	100
36) 2-Methylnaphthalene	12.563	142	368262	21.082	ng/ul	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062023\
 Data File : BP015618.D
 Acq On : 20 Jun 2023 13:05
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020758

Quant Time: Jun 20 22:25:30 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jun 15 05:36:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.781	142	369867	21.030	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.928	216	212179	22.120	ng/ul	99
40) Hexachlorocyclopentadiene	12.898	237	48463	12.325	ng/ul	99
41) 2,4,6-Trichlorophenol	13.175	196	141258	22.126	ng/ul	99
42) 2,4,5-Trichlorophenol	13.251	196	157526	22.642	ng/ul	95
43) 1,1'-Biphenyl	13.563	154	515740	21.649	ng/ul	100
44) 2-Chloronaphthalene	13.610	162	408363	21.849	ng/ul	98
45) 2-Nitroaniline	13.822	65	137046	22.910	ng/ul	95
47) Dimethylphthalate	14.181	163	532321	21.224	ng/ul	100
48) 2,6-Dinitrotoluene	14.310	165	113139	22.018	ng/ul	99
50) Acenaphthylene	14.451	152	644183	21.900	ng/ul	100
51) 3-Nitroaniline	14.645	138	90665	21.374	ng/ul#	97
52) Acenaphthene	14.792	153	450107	22.137	ng/ul	99
53) 2,4-Dinitrophenol	14.869	184	47070	15.228	ng/ul	97
55) 4-Nitrophenol	14.963	109	75714	18.045	ng/ul	98
56) Dibenzofuran	15.128	168	635787	22.062	ng/ul	98
57) 2,4-Dinitrotoluene	15.104	165	168948	22.355	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.357	232	139608	22.225	ng/ul	99
59) Diethylphthalate	15.539	149	550512	21.417	ng/ul	98
61) Fluorene	15.781	166	525940	22.388	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.769	204	259206	21.893	ng/ul	98
63) 4-Nitroaniline	15.810	138	55615	13.822	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.875	198	91681	19.013	ng/ul	94
67) N-Nitrosodiphenylamine	15.986	169	447106	21.117	ng/ul	98
68) 4-Bromophenyl-phenylether	16.669	248	167954	21.035	ng/ul	100
69) Hexachlorobenzene	16.786	284	208243	22.111	ng/ul	97
70) Atrazine	16.939	200	180696	21.372	ng/ul	98
71) Pentachlorophenol	17.139	266	115228	21.646	ng/ul	98
72) Phenanthrene	17.533	178	874463	21.834	ng/ul	100
74) Anthracene	17.622	178	898732	22.116	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.534	216	215675	20.939	ng/uL	99
76) Pentachlorobenzene	15.051	250	228453	21.184	ng/uL	99
77) Carbazole	17.892	167	824964	22.546	ng/ul	99
78) Di-n-butylphthalate	18.416	149	1009162	21.738	ng/ul	99
80) Fluoranthene	19.486	202	1125376	23.433	ng/ul	98
82) Pyrene	19.839	202	1164096	23.429	ng/ul	99
83) Butylbenzylphthalate	20.692	149	521950	23.768	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.480	252	369037	21.205	ng/ul	100
85) Benzo(a)anthracene	21.557	228	1123384	21.814	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.445	149	745068	22.954	ng/ul	99
87) Chrysene	21.610	228	1055390	21.682	ng/ul	99
89) Di-n-octyl phthalate	22.416	149	1321271	25.046	ng/ul	100
90) Benzo(b)fluoranthene	23.339	252	1113922	21.499	ng/ul	99
91) Benzo(k)fluoranthene	23.392	252	1068503	21.333	ng/ul	99
93) Benzo(a)pyrene	24.010	252	950554	21.042	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.821	276	1240616	21.221	ng/ul	99
95) Dibenzo(a,h)anthracene	26.833	278	1021903	21.455	ng/ul	100
96) Benzo(g,h,i)perylene	27.662	276	1007970	20.921	ng/ul	97





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 7.259 ng/uL
RT: 3.440 min Scan# 43
Delta R.T. -0.000 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:	88	Resp:	21179
Ion	Ratio	Lower	Upper
88	100		
43	32.1	23.9	35.9
58	65.5	51.4	77.2



#5
Pyridine
Concen: 19.755 ng/uL
RT: 3.869 min Scan# 116
Delta R.T. 0.006 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05

Tgt Ion:	79	Resp:	146951
Ion	Ratio	Lower	Upper
79	100		
52	74.3	59.5	89.3
51	38.1	29.8	44.6









#6
Benzaldehyde
Concen: 17.847 ng/ul
RT: 7.222 min Scan# 61
Delta R.T. -0.000 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion: 77 Resp: 64284
Ion Ratio Lower Upper
77 100
105 96.8 77.3 115.9
106 87.1 73.3 109.9



#8
Phenol
Concen: 20.853 ng/ul
RT: 7.275 min Scan# 695
Delta R.T. 0.006 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05

Tgt Ion: 94 Resp: 202728
Ion Ratio Lower Upper
94 100
65 31.0 23.5 35.3
66 43.8 34.2 51.4









#10
Bis(2-Chloroethyl)ether
Concen: 19.859 ng/ul
RT: 7.499 min Scan# 733
Delta R.T. -0.006 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion: 93 Resp: 152646
Ion Ratio Lower Upper
93 100
63 78.6 62.6 93.8
95 32.4 25.5 38.3



#12
2-Chlorophenol
Concen: 21.296 ng/ul
RT: 7.646 min Scan# 758
Delta R.T. -0.000 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05

Tgt Ion:128 Resp: 153731
Ion Ratio Lower Upper
128 100
64 50.1 40.2 60.4
130 31.5 24.5 36.7





#13

2-Methylphenol

Concen: 20.847 ng/ul

RT: 8.540 min Scan# 910

Delta R.T. 0.006 min

Lab File: BP015618.D

Acq: 20 Jun 2023 13:05

Instrument :

BNA_P

ClientSampleId :





SSTD020758

Tgt Ion:108 Resp: 155935

Ion Ratio Lower Upper

108 100

107 93.0 71.2 106.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 12.245 ng/ul

RT: 8.610 min Scan# 922

Delta R.T. -0.006 min

Lab File: BP015618.D

Acq: 20 Jun 2023 13:05




Tgt Ion: 45 Resp: 124725


Ion Ratio Lower Upper

45 100

77 15.6 11.8 17.8

79 12.1 8.7 13.1









#16
Acetophenone
Concen: 20.829 ng/ul
RT: 8.922 min Scan# 975
Delta R.T. -0.000 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:105 Resp: 257084
Ion Ratio Lower Upper
105 100
77 82.2 64.9 97.3
51 29.6 23.5 35.3



#17
N-Nitroso-di-n-propylamine
Concen: 20.654 ng/ul
RT: 8.899 min Scan# 971
Delta R.T. -0.006 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05

Tgt Ion: 70 Resp: 137592
Ion Ratio Lower Upper
70 100
42 56.8 46.6 69.8
101 10.4 7.8 11.6
130 19.3 15.9 23.9









#18
4-Methylphenol
Concen: 20.880 ng/ul
RT: 8.869 min Scan# 966
Delta R.T. 0.006 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:108 Resp: 171844
Ion Ratio Lower Upper
108 100
107 110.6 89.3 133.9



#19
Hexachloroethane
Concen: 20.911 ng/ul
RT: 9.169 min Scan# 1017
Delta R.T. -0.006 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05

Tgt Ion:117 Resp: 63787
Ion Ratio Lower Upper
117 100
201 102.3 82.8 124.2
199 63.8 50.6 76.0









#22
Nitrobenzene
Concen: 21.926 ng/ul
RT: 9.304 min Scan# 1040
Delta R.T. -0.000 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion: 77 Resp: 209326
Ion Ratio Lower Upper
77 100
123 40.4 33.5 50.3
65 14.4 11.6 17.4



#23
Isophorone
Concen: 20.862 ng/ul
RT: 9.834 min Scan# 1130
Delta R.T. -0.000 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05

Tgt Ion: 82 Resp: 398692
Ion Ratio Lower Upper
82 100
95 8.0 5.8 8.8
138 16.7 13.1 19.7









#25
2-Nitrophenol
Concen: 21.646 ng/ul
RT: 10.022 min Scan# 1162
Delta R.T. -0.000 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:139 Resp: 96083
Ion Ratio Lower Upper
139 100
65 59.5 46.9 70.3
109 37.4 30.2 45.2



#26
2,4-Dimethylphenol
Concen: 21.694 ng/ul
RT: 10.081 min Scan# 1172
Delta R.T. 0.006 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05

Tgt Ion:107 Resp: 204212
Ion Ratio Lower Upper
107 100
121 47.5 39.0 58.6
122 78.6 64.8 97.2





#27

Bis(2-Chloroethoxy)methane

Concen: 20.009 ng/ul

RT: 10.316 min Scan# 1212

Delta R.T. -0.000 min

Lab File: BP015618.D

Acq: 20 Jun 2023 13:05

Instrument :

BNA_P

ClientSampleId :

SSTD020758





Tgt Ion: 93 Resp: 222315

Ion Ratio Lower Upper

93 100

95 31.0 25.7 38.5

123 10.9 8.6 13.0



#29

2,4-Dichlorophenol

Concen: 21.815 ng/ul

RT: 10.563 min Scan# 1254

Delta R.T. 0.006 min

Lab File: BP015618.D

Acq: 20 Jun 2023 13:05




Tgt Ion:162 Resp: 162659


Ion Ratio Lower Upper

162 100

164 65.3 50.6 76.0

98 38.8 30.2 45.2









#30
Naphthalene
Concen: 21.472 ng/ul
RT: 10.963 min Scan# 1322
Delta R.T. -0.000 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:128 Resp: 533524
Ion Ratio Lower Upper
128 100
129 10.7 8.6 13.0
127 13.1 10.5 15.7



#32
4-Chloroaniline
Concen: 21.641 ng/ul
RT: 11.081 min Scan# 1342
Delta R.T. -0.000 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05

Tgt Ion:127 Resp: 241038
Ion Ratio Lower Upper
127 100
129 32.0 26.1 39.1





#33

Hexachlorobutadiene

Concen: 21.494 ng/ul

RT: 11.234 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP015618.D

Acq: 20 Jun 2023 13:05

Instrument :

BNA_P

ClientSampleId :

SSTD020758





Tgt Ion:225 Resp: 108086

Ion Ratio Lower Upper

225 100

223 65.6 52.0 78.0

227 64.4 50.6 75.8



#34

Caprolactam

Concen: 21.486 ng/ul

RT: 11.869 min Scan# 1476

Delta R.T. 0.006 min

Lab File: BP015618.D

Acq: 20 Jun 2023 13:05




Tgt Ion:113 Resp: 60310


Ion Ratio Lower Upper

113 100

55 170.2 127.9 191.9

56 131.3 101.1 151.7









#35
4-Chloro-3-methylphenol
Concen: 22.079 ng/ul
RT: 12.204 min Scan# 1533
Delta R.T. 0.012 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:107 Resp: 199471
Ion Ratio Lower Upper
107 100
144 23.9 19.4 29.0
142 75.1 60.4 90.6



#36
2-Methylnaphthalene
Concen: 21.082 ng/ul
RT: 12.563 min Scan# 1594
Delta R.T. -0.000 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05

Tgt Ion:142 Resp: 368262
Ion Ratio Lower Upper
142 100
141 89.0 72.0 108.0





#37

1-Methylnaphthalene

Concen: 21.030 ng/ul

RT: 12.781 min Scan# 1631

Delta R.T. -0.000 min

Lab File: BP015618.D

Acq: 20 Jun 2023 13:05

Instrument :

BNA_P

ClientSampleId :

SSTD020758





Tgt Ion:142 Resp: 369867

Ion Ratio Lower Upper

142 100

141 92.7 71.9 107.9

116 4.3 3.4 5.2



#39

1,2,4,5-Tetrachlorobenzene

Concen: 22.120 ng/ul

RT: 12.928 min Scan# 1656

Delta R.T. -0.000 min

Lab File: BP015618.D

Acq: 20 Jun 2023 13:05

Tgt Ion:216 Resp: 212179




Ion Ratio Lower Upper


216 100

214 77.3 62.2 93.2

179 22.5 17.4 26.2

108 17.3 14.2 21.4









#40
Hexachlorocyclopentadiene
Concen: 12.325 ng/ul
RT: 12.898 min Scan# 1651
Delta R.T. -0.000 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:237 Resp: 48463
Ion Ratio Lower Upper
237 100
235 61.2 48.2 72.2
272 12.5 10.3 15.5



#41
2,4,6-Trichlorophenol
Concen: 22.126 ng/ul
RT: 13.175 min Scan# 1698
Delta R.T. 0.006 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05

Tgt Ion:196 Resp: 141258
Ion Ratio Lower Upper
196 100
198 98.2 77.8 116.6
200 31.1 25.8 38.6





#42

2,4,5-Trichlorophenol

Concen: 22.642 ng/ul

RT: 13.251 min Scan# 1711

Delta R.T. 0.006 min

Lab File: BP015618.D

Acq: 20 Jun 2023 13:05

Instrument :

BNA_P

ClientSampleId :

SSTD020758





Tgt Ion:196 Resp: 157526

Ion Ratio Lower Upper

196 100

198 99.9 75.5 113.3

200 32.2 24.7 37.1



#43

1,1'-Biphenyl

Concen: 21.649 ng/ul

RT: 13.563 min Scan# 1764

Delta R.T. -0.000 min

Lab File: BP015618.D

Acq: 20 Jun 2023 13:05




Tgt Ion:154 Resp: 515740


Ion Ratio Lower Upper

154 100

153 40.4 32.2 48.2

76 13.0 10.4 15.6









#44
2-Chloronaphthalene
Concen: 21.849 ng/ul
RT: 13.610 min Scan# 1772
Delta R.T. -0.000 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:162 Resp: 408363
Ion Ratio Lower Upper
162 100
164 31.9 27.0 40.6
127 40.4 32.2 48.4



#45
2-Nitroaniline
Concen: 22.910 ng/ul
RT: 13.822 min Scan# 1808
Delta R.T. -0.000 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05

Tgt Ion: 65 Resp: 137046
Ion Ratio Lower Upper
65 100
92 64.6 55.4 83.2
138 93.4 78.8 118.2









#47
Dimethylphthalate
Concen: 21.224 ng/ul
RT: 14.181 min Scan# 1869
Delta R.T. -0.000 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:163 Resp: 532321
Ion Ratio Lower Upper
163 100
194 4.1 3.2 4.8
164 10.2 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 22.018 ng/ul
RT: 14.310 min Scan# 1891
Delta R.T. -0.000 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05

Tgt Ion:165 Resp: 113139
Ion Ratio Lower Upper
165 100
89 63.4 51.5 77.3
121 21.9 17.3 25.9









#50
Acenaphthylene
Concen: 21.900 ng/ul
RT: 14.451 min Scan# 1915
Delta R.T. -0.000 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:152 Resp: 644183
Ion Ratio Lower Upper
152 100
151 20.1 16.0 24.0
153 13.1 10.6 15.8



#51
3-Nitroaniline
Concen: 21.374 ng/ul
RT: 14.645 min Scan# 1948
Delta R.T. -0.000 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05

Tgt Ion:138 Resp: 90665
Ion Ratio Lower Upper
138 100
108 12.7 8.4 12.6#
92 126.1 98.0 147.0









#52
Acenaphthene
Concen: 22.137 ng/ul
RT: 14.792 min Scan# 1973
Delta R.T. -0.000 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:153 Resp: 450107
Ion Ratio Lower Upper
153 100
152 46.1 38.3 57.5
154 87.4 69.8 104.6



#53
2,4-Dinitrophenol
Concen: 15.228 ng/ul
RT: 14.869 min Scan# 1986
Delta R.T. 0.012 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05

Tgt Ion:184 Resp: 47070
Ion Ratio Lower Upper
184 100
63 76.8 59.0 88.4
154 70.6 57.6 86.4









#55
4-Nitrophenol
Concen: 18.045 ng/ul
RT: 14.963 min Scan# 2002
Delta R.T. 0.018 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:109 Resp: 75714
Ion Ratio Lower Upper
109 100
139 107.7 89.1 133.7
65 130.4 103.1 154.7



#56
Dibenzofuran
Concen: 22.062 ng/ul
RT: 15.128 min Scan# 2030
Delta R.T. -0.000 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05

Tgt Ion:168 Resp: 635787
Ion Ratio Lower Upper
168 100
139 40.1 31.2 46.8









#57
2,4-Dinitrotoluene
Concen: 22.355 ng/ul
RT: 15.104 min Scan# 2026
Delta R.T. 0.006 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:165 Resp: 168948
Ion Ratio Lower Upper
165 100
63 49.8 39.1 58.7
182 2.7 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 22.225 ng/ul
RT: 15.357 min Scan# 2069
Delta R.T. -0.000 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05

Tgt Ion:232 Resp: 139608
Ion Ratio Lower Upper
232 100
131 46.5 37.8 56.8
130 2.6 1.9 2.9
166 32.5 26.2 39.2









#59
Diethylphthalate
Concen: 21.417 ng/ul
RT: 15.539 min Scan# 2100
Delta R.T. -0.000 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:149 Resp: 550512
Ion Ratio Lower Upper
149 100
177 21.3 17.8 26.8
150 12.0 10.1 15.1



#61
Fluorene
Concen: 22.388 ng/ul
RT: 15.781 min Scan# 2141
Delta R.T. -0.000 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05

Tgt Ion:166 Resp: 525940
Ion Ratio Lower Upper
166 100
165 97.5 79.1 118.7
167 13.6 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 21.893 ng/ul
RT: 15.769 min Scan# 2139
Delta R.T. -0.000 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:204 Resp: 259206
Ion Ratio Lower Upper
204 100
206 32.4 26.6 40.0
141 61.1 49.8 74.8



#63
4-Nitroaniline
Concen: 13.822 ng/ul
RT: 15.810 min Scan# 2146
Delta R.T. 0.006 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05

Tgt Ion:138 Resp: 55615
Ion Ratio Lower Upper
138 100
92 56.7 44.6 67.0
108 95.2 78.4 117.6









#66
4,6-Dinitro-2-methylphenol
Concen: 19.013 ng/ul
RT: 15.875 min Scan# 2157
Delta R.T. 0.012 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:198 Resp: 91681
Ion Ratio Lower Upper
198 100
182 4.0 3.0 4.4
77 29.1 20.5 30.7



#67
N-Nitrosodiphenylamine
Concen: 21.117 ng/ul
RT: 15.986 min Scan# 2176
Delta R.T. 0.006 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05

Tgt Ion:169 Resp: 447106
Ion Ratio Lower Upper
169 100
168 68.2 53.0 79.4
167 36.2 28.9 43.3









#68
4-Bromophenyl-phenylether
Concen: 21.035 ng/ul
RT: 16.669 min Scan# 2112
Delta R.T. 0.006 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:248 Resp: 167954
Ion Ratio Lower Upper
248 100
250 98.4 78.3 117.5
141 76.3 61.1 91.7



#69
Hexachlorobenzene
Concen: 22.111 ng/ul
RT: 16.786 min Scan# 2312
Delta R.T. -0.000 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05

Tgt Ion:284 Resp: 208243
Ion Ratio Lower Upper
284 100
142 35.2 25.4 38.2
249 30.6 24.8 37.2









#70
 Atrazine
 Concen: 21.372 ng/ul
 RT: 16.939 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BP015618.D
 Acq: 20 Jun 2023 13:05




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020758


Tgt Ion:	200	Resp:	180696
Ion Ratio	Lower	Upper	
200	100		
173	26.8	19.8	29.6
215	47.0	37.4	56.0



#71
 Pentachlorophenol
 Concen: 21.646 ng/ul
 RT: 17.139 min Scan# 2372
 Delta R.T. 0.006 min
 Lab File: BP015618.D
 Acq: 20 Jun 2023 13:05

Tgt Ion:	266	Resp:	115228
Ion Ratio	Lower	Upper	
266	100		
264	63.3	49.6	74.4
268	62.0	51.0	76.6





#72

Phenanthrene

Concen: 21.834 ng/ul

RT: 17.533 min Scan# 2439

Delta R.T. 0.006 min

Lab File: BP015618.D

Acq: 20 Jun 2023 13:05

Instrument :

BNA_P

ClientSampleId :

SSTD020758





Tgt Ion:178 Resp: 874463

Ion Ratio Lower Upper

178 100

179 15.2 12.3 18.5

176 19.1 15.4 23.0



#74

Anthracene

Concen: 22.116 ng/ul

RT: 17.622 min Scan# 2454

Delta R.T. -0.000 min

Lab File: BP015618.D

Acq: 20 Jun 2023 13:05




Tgt Ion:178 Resp: 898732


Ion Ratio Lower Upper

178 100

179 15.1 12.4 18.6

176 18.3 15.0 22.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.939 ng/uL
RT: 13.534 min Scan# 1759
Delta R.T. -0.000 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:216 Resp: 215675
Ion Ratio Lower Upper
216 100
214 77.7 62.1 93.1
218 49.5 39.0 58.4



#76
Pentachlorobenzene
Concen: 21.184 ng/uL
RT: 15.051 min Scan# 2017
Delta R.T. 0.006 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05

Tgt Ion:250 Resp: 228453
Ion Ratio Lower Upper
250 100
248 62.0 49.0 73.4
252 63.5 49.9 74.9









#77
 Carbazole
 Concen: 22.546 ng/ul
 RT: 17.892 min Scan# 2500
 Delta R.T. -0.000 min
 Lab File: BP015618.D
 Acq: 20 Jun 2023 13:05




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020758


Tgt Ion:167 Resp: 824964
 Ion Ratio Lower Upper
 167 100
 166 21.2 17.2 25.8
 139 13.2 10.6 16.0



#78
 Di-n-butylphthalate
 Concen: 21.738 ng/ul
 RT: 18.416 min Scan# 2589
 Delta R.T. -0.000 min
 Lab File: BP015618.D
 Acq: 20 Jun 2023 13:05

Tgt Ion:149 Resp: 1009162
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.1 10.7
 104 5.8 4.5 6.7









#80
Fluoranthene
Concen: 23.433 ng/ul
RT: 19.486 min Scan# 2771
Delta R.T. -0.000 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05




Instrument :
BNA_P
ClientSampleId :
SSTD020758


Tgt Ion:202 Resp: 1125376
Ion Ratio Lower Upper
202 100
101 10.2 8.6 13.0
100 7.6 6.6 9.8



#82
Pyrene
Concen: 23.429 ng/ul
RT: 19.839 min Scan# 2831
Delta R.T. -0.000 min
Lab File: BP015618.D
Acq: 20 Jun 2023 13:05

Tgt Ion:202 Resp: 1164096
Ion Ratio Lower Upper
202 100
101 12.2 9.9 14.9
100 10.0 8.4 12.6





#83

Butylbenzylphthalate

Concen: 23.768 ng/ul

RT: 20.692 min Scan# 2976

Delta R.T. -0.000 min

Lab File: BP015618.D

Acq: 20 Jun 2023 13:05

Instrument :

BNA_P

ClientSampleId :

SSTD020758





Tgt Ion:149 Resp: 521950

Ion Ratio Lower Upper

149 100

91 75.0 60.4 90.6

206 21.0 16.4 24.6



#84

3,3'-Dichlorobenzidine

Concen: 21.205 ng/ul

RT: 21.480 min Scan# 3110

Delta R.T. -0.000 min

Lab File: BP015618.D

Acq: 20 Jun 2023 13:05




Tgt Ion:252 Resp: 369037


Ion Ratio Lower Upper

252 100

254 64.9 51.6 77.4

126 10.6 8.5 12.7





#85

Benzo(a)anthracene

Concen: 21.814 ng/ul

RT: 21.557 min Scan# 3123

Delta R.T. 0.006 min

Lab File: BP015618.D

Acq: 20 Jun 2023 13:05

Instrument :

BNA_P

ClientSampleId :

SSTD020758





Tgt Ion:228 Resp: 1123384

Ion Ratio Lower Upper

228 100

229 19.7 15.8 23.6

226 27.0 21.8 32.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 22.954 ng/ul

RT: 21.445 min Scan# 3104

Delta R.T. -0.000 min

Lab File: BP015618.D

Acq: 20 Jun 2023 13:05




Tgt Ion:149 Resp: 745068


Ion Ratio Lower Upper

149 100

167 26.9 21.8 32.8

279 4.3 3.7 5.5





#87

Chrysene

Concen: 21.682 ng/ul

RT: 21.610 min Scan# 3132

Delta R.T. -0.000 min

Lab File: BP015618.D

Acq: 20 Jun 2023 13:05

Instrument :

BNA_P

ClientSampleId :

SSTD020758





Tgt Ion:228 Resp: 1055390

Ion Ratio Lower Upper

228 100

226 30.3 23.9 35.9

229 19.8 15.1 22.7



#89

Di-n-octyl phthalate

Concen: 25.046 ng/ul




RT: 22.416 min Scan# 3269


Delta R.T. -0.000 min

Lab File: BP015618.D

Acq: 20 Jun 2023 13:05

Tgt Ion:149 Resp: 1321271





#90

Benzo(b)fluoranthene

Concen: 21.499 ng/ul

RT: 23.339 min Scan# 3426

Delta R.T. 0.012 min

Lab File: BP015618.D

Acq: 20 Jun 2023 13:05

Instrument :

BNA_P

ClientSampleId :

SSTD020758





Tgt Ion:252 Resp: 1113922

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.2

125 8.0 7.3 10.9



#91

Benzo(k)fluoranthene

Concen: 21.333 ng/ul

RT: 23.392 min Scan# 3435

Delta R.T. 0.012 min

Lab File: BP015618.D

Acq: 20 Jun 2023 13:05




Tgt Ion:252 Resp: 1068503


Ion Ratio Lower Upper

252 100

253 21.4 17.7 26.5

125 8.6 7.3 10.9





#93

Benzo(a)pyrene

Concen: 21.042 ng/ul

RT: 24.010 min Scan# 3540

Delta R.T. 0.006 min

Lab File: BP015618.D

Acq: 20 Jun 2023 13:05

Instrument :

BNA_P

ClientSampleId :

SSTD020758





Tgt Ion:252 Resp: 950554

Ion Ratio Lower Upper

252 100

253 21.8 17.8 26.6

125 9.6 8.1 12.1



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.221 ng/ul

RT: 26.821 min Scan# 4018

Delta R.T. 0.012 min

Lab File: BP015618.D

Acq: 20 Jun 2023 13:05




Tgt Ion:276 Resp: 1240616


Ion Ratio Lower Upper

276 100

138 20.2 17.3 25.9

277 25.0 20.2 30.2





#95

Dibenzo(a,h)anthracene

Concen: 21.455 ng/ul

RT: 26.833 min Scan# 4020

Delta R.T. 0.006 min

Lab File: BP015618.D

Acq: 20 Jun 2023 13:05

Instrument :

BNA_P

ClientSampleId :

SSTD020758





Tgt Ion:278 Resp: 1021903

Ion Ratio Lower Upper

278 100

139 13.7 11.4 17.0

279 23.4 18.7 28.1



#96

Benzo(g,h,i)perylene

Concen: 20.921 ng/ul

RT: 27.662 min Scan# 4161

Delta R.T. 0.023 min

Lab File: BP015618.D

Acq: 20 Jun 2023 13:05

Tgt Ion:276 Resp: 1007970

Ion Ratio Lower Upper

276 100

138 17.3 15.6 23.4

277 22.7 19.0 28.6

