

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062223\
 Data File : BP015640.D
 Acq On : 22 Jun 2023 11:10
 Operator : MA/JU
 Sample : PB153402BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK402

Quant Time: Jun 22 22:14:49 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jun 22 22:01:49 2023
 Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|-------------------------------|--------|------|----------|--------|---------|----------|
| Internal Standards | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 8.081 | 152 | 96374 | 20.000 | ng/ul | 0.00 |
| 20) Naphthalene-d8 | 10.910 | 136 | 420274 | 20.000 | ng/ul # | 0.00 |
| 38) Acenaphthene-d10 | 14.728 | 164 | 284881 | 20.000 | ng/ul | 0.00 |
| 64) Phenanthrene-d10 | 17.492 | 188 | 644014 | 20.000 | ng/ul | 0.00 |
| 79) Chrysene-d12 | 21.580 | 240 | 557847 | 20.000 | ng/ul | 0.00 |
| 88) Perylene-d12 | 24.139 | 264 | 527619 | 20.000 | ng/ul | 0.00 |
| System Monitoring Compounds | | | | | | |
| 3) 1,4-Dioxane-d8 | 3.399 | 96 | 18756 | 7.207 | ng/uL | 0.00 |
| 4) Pyridine-d5 | 3.840 | 84 | 234171 | 34.639 | ng/ul | 0.00 |
| 7) Phenol-d5 | 7.240 | 99 | 302183 | 34.035 | ng/ul | 0.00 |
| 9) Bis-(2-Chloroethyl)eth... | 7.405 | 67 | 200797 | 37.869 | ng/ul | 0.00 |
| 11) 2-Chlorophenol-d4 | 7.605 | 132 | 253959 | 39.452 | ng/ul | 0.00 |
| 15) 4-Methylphenol-d8 | 8.805 | 113 | 261602 | 35.901 | ng/ul | 0.00 |
| 21) Nitrobenzene-d5 | 9.263 | 128 | 131350 | 39.808 | ng/ul | 0.00 |
| 24) 2-Nitrophenol-d4 | 9.987 | 143 | 149385 | 39.645 | ng/ul | 0.00 |
| 28) 2,4-Dichlorophenol-d3 | 10.534 | 165 | 242653 | 35.302 | ng/ul | 0.00 |
| 31) 4-Chloroaniline-d4 | 11.057 | 131 | 377854 | 38.298 | ng/ul | 0.00 |
| 46) Dimethylphthalate-d6 | 14.134 | 166 | 953199 | 42.431 | ng/ul | 0.00 |
| 49) Acenaphthylene-d8 | 14.428 | 160 | 1006914 | 40.991 | ng/ul | 0.00 |
| 54) 4-Nitrophenol-d4 | 14.951 | 143 | 129500 | 32.191 | ng/ul | 0.00 |
| 60) Fluorene-d10 | 15.728 | 176 | 755963 | 39.906 | ng/ul | 0.00 |
| 65) 4,6-Dinitro-2-methylph... | 15.863 | 200 | 122905 | 30.125 | ng/ul | 0.00 |
| 73) Anthracene-d10 | 17.592 | 188 | 1204823 | 41.118 | ng/ul | 0.00 |
| 81) Pyrene-d10 | 19.816 | 212 | 1398214 | 46.831 | ng/ul | 0.00 |
| 92) Benzo(a)pyrene-d12 | 23.969 | 264 | 1133404 | 42.812 | ng/ul | 0.00 |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062223\
 Data File : BP015640.D
 Acq On : 22 Jun 2023 11:10
 Operator : MA/JU
 Sample : PB153402BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK402

Quant Time: Jun 22 22:14:49 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jun 22 22:01:49 2023
 Response via : Initial Calibration

