

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062322\
 Data File : BP010775.D
 Acq On : 23 Jun 2022 15:17
 Operator : CG/JU
 Sample : SSTD01011
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD010611

Quant Time: Jun 23 17:55:26 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jun 23 17:53:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.722	152	391883	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.522	136	1614345	20.000	ng/u1	#-0.01
38) Acenaphthene-d10	14.381	164	875962	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.145	188	1617438	20.000	ng/u1	#-0.01
79) Chrysene-d12	21.257	240	1373353	20.000	ng/u1	#-0.01
88) Perylene-d12	23.633	264	1470599	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.211	96	38466	3.914	ng/uL	0.00
4) Pyridine-d5	3.611	84	258139	9.996	ng/u1	0.00
7) Phenol-d5	6.893	99	286811	9.722	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.063	67	195685	10.410	ng/u1	0.00
11) 2-Chlorophenol-d4	7.258	132	239447	9.943	ng/u1	0.00
15) 4-Methylphenol-d8	8.434	113	226955	9.858	ng/u1	-0.01
21) Nitrobenzene-d5	8.887	128	87336	10.230	ng/u1	0.00
24) 2-Nitrophenol-d4	9.604	143	65567	9.710	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.140	165	190610	9.095	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.663	131	323076	9.555	ng/u1	0.00
46) Dimethylphthalate-d6	13.804	166	558212	9.555	ng/u1	0.00
49) Acenaphthylene-d8	14.075	160	721279	9.527	ng/u1	0.00
54) 4-Nitrophenol-d4	14.575	143	61871	7.897	ng/u1	-0.01
60) Fluorene-d10	15.381	176	485339	9.297	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.492	200	29536	6.658	ng/u1	-0.01
73) Anthracene-d10	17.245	188	682464	9.441	ng/u1	0.00
81) Pyrene-d10	19.492	212	701481	10.054	ng/u1	-0.01
92) Benzo(a)pyrene-d12	23.480	264	661009	9.093	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.240	88	42288	4.145	ng/uL#	92
5) Pyridine	3.628	79	261901	9.905	ng/u1#	74
6) Benzaldehyde	6.869	77	182650	12.399	ng/u1	96
8) Phenol	6.916	94	313756	9.983	ng/u1#	86
10) Bis(2-Chloroethyl)ether	7.157	93	255051	10.134	ng/u1	90
12) 2-Chlorophenol	7.287	128	249995	9.955	ng/u1	92
13) 2-Methylphenol	8.169	108	231685	9.869	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.316	45	3726	0.111	ng/u1#	14
16) Acetophenone	8.546	105	375354	10.494	ng/u1#	90
17) N-Nitroso-di-n-propyla...	8.534	70	181692	10.453	ng/u1	91
18) 4-Methylphenol	8.493	108	242507	10.021	ng/u1	96
19) Hexachloroethane	8.804	117	102113	10.126	ng/u1	85
22) Nitrobenzene	8.928	77	228168	10.522	ng/u1	93
23) Isophorone	9.451	82	499513	9.735	ng/u1	97
25) 2-Nitrophenol	9.634	139	80887	9.683	ng/u1#	81
26) 2,4-Dimethylphenol	9.704	107	264671	9.887	ng/u1	95
27) Bis(2-Chloroethoxy)met...	9.946	93	317986	9.771	ng/u1	97
29) 2,4-Dichlorophenol	10.163	162	201179	9.265	ng/u1	99
30) Naphthalene	10.575	128	799716	9.695	ng/u1	99
32) 4-Chloroaniline	10.687	127	322606	9.691	ng/u1	99
33) Hexachlorobutadiene	10.863	225	131708	8.466	ng/u1	96
34) Caprolactam	11.434	113	54601	8.427	ng/u1	86
35) 4-Chloro-3-methylphenol	11.816	107	213466	9.934	ng/u1	93
36) 2-Methylnaphthalene	12.193	142	514004	9.458	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062322\
 Data File : BP010775.D
 Acq On : 23 Jun 2022 15:17
 Operator : CG/JU
 Sample : SST01011
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SST010611


Quant Time: Jun 23 17:55:26 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jun 23 17:53:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.416	142	511431	9.447	ng/ul#	98
39) 1,2,4,5-Tetrachloroben...	12.557	216	240116	8.973	ng/ul	98
40) Hexachlorocyclopentadiene	12.545	237	128620	7.873	ng/ul	99
41) 2,4,6-Trichlorophenol	12.804	196	128504	9.192	ng/ul	96
42) 2,4,5-Trichlorophenol	12.869	196	135992	9.204	ng/ul	99
43) 1,1'-Biphenyl	13.216	154	660496	9.870	ng/ul	99
44) 2-Chloronaphthalene	13.251	162	494774	9.634	ng/ul	100
45) 2-Nitroaniline	13.457	65	83319	11.135	ng/ul#	85
47) Dimethylphthalate	13.845	163	560766	9.611	ng/ul	98
48) 2,6-Dinitrotoluene	13.963	165	67550	10.022	ng/ul	93
50) Acenaphthylene	14.098	152	798212	9.761	ng/ul	98
51) 3-Nitroaniline	14.287	138	75396	8.808	ng/ul#	90
52) Acenaphthene	14.445	153	513068	9.723	ng/ul	99
53) 2,4-Dinitrophenol	14.487	184	18786	6.544	ng/ul	92
55) 4-Nitrophenol	14.592	109	57661	9.336	ng/ul#	80
56) Dibenzofuran	14.781	168	709324	9.568	ng/ul	95
57) 2,4-Dinitrotoluene	14.745	165	86931	9.427	ng/ul	89
58) 2,3,4,6-Tetrachlorophenol	15.010	232	94575	9.009	ng/ul#	94
59) Diethylphthalate	15.222	149	550706	9.801	ng/ul	98
61) Fluorene	15.434	166	561007	9.503	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.439	204	262093	9.027	ng/ul	98
63) 4-Nitroaniline	15.451	138	77393	9.591	ng/ul	88
66) 4,6-Dinitro-2-methylph...	15.510	198	31229	6.469	ng/ul	95
67) N-Nitrosodiphenylamine	15.651	169	466837	9.873	ng/ul	98
68) 4-Bromophenyl-phenylether	16.339	248	151228	9.030	ng/ul	98
69) Hexachlorobenzene	16.439	284	178195	9.075	ng/ul	95
70) Atrazine	16.616	200	143060	8.863	ng/ul	97
71) Pentachlorophenol	16.786	266	70325	7.775	ng/ul	97
72) Phenanthrene	17.186	178	817013	9.639	ng/ul	99
74) Anthracene	17.281	178	827678	9.714	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.169	216	256742	9.443	ng/ul	98
76) Pentachlorobenzene	14.698	250	220138	9.291	ng/ul	98
77) Carbazole	17.551	167	740472	9.857	ng/ul	100
78) Di-n-butylphthalate	18.133	149	851455	10.026	ng/ul	99
80) Fluoranthene	19.169	202	862524	10.202	ng/ul#	88
82) Pyrene	19.522	202	884669	10.308	ng/ul#	84
83) Butylbenzylphthalate	20.427	149	258603	9.461	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.186	252	243464	8.835	ng/ul	99
85) Benzo(a)anthracene	21.245	228	799220	9.475	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.204	149	442242	9.623	ng/ul#	96
87) Chrysene	21.298	228	772549	9.499	ng/ul	99
89) Di-n-octyl phthalate	22.127	149	682966	9.154	ng/ul	100
90) Benzo(b)fluoranthene	22.910	252	827752	9.200	ng/ul#	96
91) Benzo(k)fluoranthene	22.957	252	791759	9.397	ng/ul#	95
93) Benzo(a)pyrene	23.527	252	815994	9.233	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	26.098	276	1009102	9.263	ng/ul#	91
95) Dibenzo(a,h)anthracene	26.127	278	871086	9.352	ng/ul#	93
96) Benzo(g,h,i)perylene	26.851	276	859928	9.299	ng/ul#	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 23 17:55:26 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jun 23 17:53:52 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 4.145 ng/uL
RT: 3.240 min Scan# 27
Delta R.T. -0.006 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion: 88 Resp: 42288
Ion Ratio Lower Upper
88 100
43 32.5 35.8 53.6#
58 83.5 68.1 102.1



#5
Pyridine
Concen: 9.905 ng/uL
RT: 3.628 min Scan# 92
Delta R.T. 0.000 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17

Tgt Ion: 79 Resp: 261901
Ion Ratio Lower Upper
79 100
52 76.1 83.8 125.8#
51 36.5 41.0 61.4#









#6
Benzaldehyde
Concen: 12.399 ng/ul
RT: 6.869 min Scan# 644
Delta R.T. -0.006 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion: 77 Resp: 182650
Ion Ratio Lower Upper
77 100
105 99.3 76.0 114.0
106 95.0 73.2 109.8



#8
Phenol
Concen: 9.983 ng/ul
RT: 6.916 min Scan# 651
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17

Tgt Ion: 94 Resp: 313756
Ion Ratio Lower Upper
94 100
65 28.0 27.4 41.2
66 36.3 37.4 56.2#









#10
Bis(2-Chloroethyl)ether
Concen: 10.134 ng/ul
RT: 7.157 min Scan# 693
Delta R.T. -0.006 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion: 93	Resp: 255051
Ion Ratio	Lower Upper
93 100	
63 78.6	72.4 108.6
95 32.6	27.3 40.9



#12
2-Chlorophenol
Concen: 9.955 ng/ul
RT: 7.287 min Scan# 714
Delta R.T. -0.006 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17

Tgt Ion:128	Resp: 249995
Ion Ratio	Lower Upper
128 100	
64 47.3	44.8 67.2
130 32.3	25.8 38.6





#13

2-Methylphenol

Concen: 9.869 ng/ul

RT: 8.169 min Scan# 8

Delta R.T. -0.006 min

Lab File: BP010775.D

Acq: 23 Jun 2022 15:17

Instrument :

BNA_P

ClientSampleId :





SSTD010611

Tgt Ion:108 Resp: 231685

Ion Ratio Lower Upper

108 100

107 89.7 71.4 107.0



#14

2,2'-oxybis(1-Chloropropane)

Concen: 0.111 ng/ul

RT: 8.316 min Scan# 889

Delta R.T. 0.041 min

Lab File: BP010775.D

Acq: 23 Jun 2022 15:17




Tgt Ion: 45 Resp: 3726


Ion Ratio Lower Upper

45 100

77 54.7 11.2 16.8#

79 30.9 6.9 10.3#









#16
Acetophenone
Concen: 10.494 ng/ul
RT: 8.546 min Scan# 91
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:105 Resp: 375354
Ion Ratio Lower Upper
105 100
77 78.4 67.5 101.3
51 26.8 29.6 44.4#



#17
N-Nitroso-di-n-propylamine
Concen: 10.453 ng/ul
RT: 8.534 min Scan# 926
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17

Tgt Ion: 70 Resp: 181692
Ion Ratio Lower Upper
70 100
42 57.8 53.8 80.8
101 9.6 7.8 11.6
130 20.7 15.4 23.2









#18
4-Methylphenol
Concen: 10.021 ng/ul
RT: 8.493 min Scan# 91
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:108 Resp: 242507
Ion Ratio Lower Upper
108 100
107 114.6 88.3 132.5



#19
Hexachloroethane
Concen: 10.126 ng/ul
RT: 8.804 min Scan# 972
Delta R.T. -0.006 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17

Tgt Ion:117 Resp: 102113
Ion Ratio Lower Upper
117 100
201 91.7 85.7 128.5
199 58.5 57.4 86.0









#22
Nitrobenzene
Concen: 10.522 ng/ul
RT: 8.928 min Scan# 994
Delta R.T. -0.006 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion: 77 Resp: 228168
Ion Ratio Lower Upper
77 100
123 44.6 31.5 47.3
65 13.7 11.4 17.2



#23
Isophorone
Concen: 9.735 ng/ul
RT: 9.451 min Scan# 1082
Delta R.T. -0.006 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17

Tgt Ion: 82 Resp: 499513
Ion Ratio Lower Upper
82 100
95 7.1 6.6 10.0
138 18.2 13.8 20.6









#25
2-Nitrophenol
Concen: 9.683 ng/ul
RT: 9.634 min Scan# 1113
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:139 Resp: 80887
Ion Ratio Lower Upper
139 100
65 51.2 54.2 81.4#
109 28.2 30.6 45.8#



#26
2,4-Dimethylphenol
Concen: 9.887 ng/ul
RT: 9.704 min Scan# 1125
Delta R.T. -0.006 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17

Tgt Ion:107 Resp: 264671
Ion Ratio Lower Upper
107 100
121 52.5 39.1 58.7
122 86.6 66.1 99.1





#27

Bis(2-Chloroethoxy)methane

Concen: 9.771 ng/ul

RT: 9.946 min Scan# 1167

Delta R.T. -0.006 min

Lab File: BP010775.D

Acq: 23 Jun 2022 15:17

Instrument :

BNA_P

ClientSampleId :

SSTD010611





Tgt Ion: 93 Resp: 317986

Ion Ratio Lower Upper

93 100

95 31.5 26.6 39.8

123 11.0 7.8 11.8



#29

2,4-Dichlorophenol

Concen: 9.265 ng/ul

RT: 10.163 min Scan# 1203

Delta R.T. -0.012 min

Lab File: BP010775.D

Acq: 23 Jun 2022 15:17




Tgt Ion:162 Resp: 201179


Ion Ratio Lower Upper

162 100

164 66.3 52.5 78.7

98 38.7 31.4 47.0









#30
Naphthalene
Concen: 9.695 ng/ul
RT: 10.575 min Scan# 1273
Delta R.T. -0.006 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:128 Resp: 799716
Ion Ratio Lower Upper
128 100
129 10.8 8.7 13.1
127 13.6 10.2 15.4



#32
4-Chloroaniline
Concen: 9.691 ng/ul
RT: 10.687 min Scan# 1292
Delta R.T. -0.006 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17

Tgt Ion:127 Resp: 322606
Ion Ratio Lower Upper
127 100
129 31.9 26.2 39.2









#33
Hexachlorobutadiene
Concen: 8.466 ng/ul
RT: 10.863 min Scan# 1419
Delta R.T. -0.006 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:225 Resp: 131708
Ion Ratio Lower Upper
225 100
223 62.5 53.7 80.5
227 64.1 49.4 74.2



#34
Caprolactam
Concen: 8.427 ng/ul
RT: 11.434 min Scan# 1419
Delta R.T. -0.018 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17

Tgt Ion:113 Resp: 54601
Ion Ratio Lower Upper
113 100
55 173.6 153.8 230.8
56 119.5 110.2 165.4









#35
4-Chloro-3-methylphenol
Concen: 9.934 ng/ul
RT: 11.816 min Scan# 1484
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:107 Resp: 213466
Ion Ratio Lower Upper
107 100
144 25.4 18.6 27.8
142 81.5 59.6 89.4



#36
2-Methylnaphthalene
Concen: 9.458 ng/ul
RT: 12.193 min Scan# 1548
Delta R.T. -0.006 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17

Tgt Ion:142 Resp: 514004
Ion Ratio Lower Upper
142 100
141 89.0 70.1 105.1









#37
1-Methylnaphthalene
Concen: 9.447 ng/ul
RT: 12.416 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:142 Resp: 511431
Ion Ratio Lower Upper
142 100
141 93.3 73.3 109.9
116 3.9 4.0 6.0#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 8.973 ng/ul
RT: 12.557 min Scan# 1610
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17

Tgt Ion:216 Resp: 240116
Ion Ratio Lower Upper
216 100
214 78.3 63.0 94.6
179 21.0 18.1 27.1
108 20.4 13.6 20.4









#40
Hexachlorocyclopentadiene
Concen: 7.873 ng/ul
RT: 12.545 min Scan# 1608
Delta R.T. -0.006 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:237 Resp: 128620
Ion Ratio Lower Upper
237 100
235 65.4 52.8 79.2
272 14.5 11.5 17.3



#41
2,4,6-Trichlorophenol
Concen: 9.192 ng/ul
RT: 12.804 min Scan# 1652
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17

Tgt Ion:196 Resp: 128504
Ion Ratio Lower Upper
196 100
198 93.3 78.2 117.4
200 30.5 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 9.204 ng/ul
RT: 12.869 min Scan# 1663
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:196 Resp: 135992
Ion Ratio Lower Upper
196 100
198 97.8 77.0 115.4
200 31.5 24.8 37.2



#43
1,1'-Biphenyl
Concen: 9.870 ng/ul
RT: 13.216 min Scan# 1722
Delta R.T. -0.006 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17

Tgt Ion:154 Resp: 660496
Ion Ratio Lower Upper
154 100
153 39.8 31.9 47.9
76 13.3 9.2 13.8









#44
2-Chloronaphthalene
Concen: 9.634 ng/ul
RT: 13.251 min Scan# 1728
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:162 Resp: 494774
Ion Ratio Lower Upper
162 100
164 32.3 25.5 38.3
127 40.0 31.9 47.9



#45
2-Nitroaniline
Concen: 11.135 ng/ul
RT: 13.457 min Scan# 1763
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17

Tgt Ion: 65 Resp: 83319
Ion Ratio Lower Upper
65 100
92 67.2 49.0 73.6
138 100.8 66.4 99.6#









#47
Dimethylphthalate
Concen: 9.611 ng/ul
RT: 13.845 min Scan# 1829
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:163 Resp: 560766
Ion Ratio Lower Upper
163 100
194 3.6 3.4 5.2
164 10.3 7.6 11.4



#48
2,6-Dinitrotoluene
Concen: 10.022 ng/ul
RT: 13.963 min Scan# 1849
Delta R.T. -0.006 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17

Tgt Ion:165 Resp: 67550
Ion Ratio Lower Upper
165 100
89 60.8 52.9 79.3
121 27.5 19.4 29.2









#50
Acenaphthylene
Concen: 9.761 ng/ul
RT: 14.098 min Scan# 1874
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:152 Resp: 798212
Ion Ratio Lower Upper
152 100
151 19.4 16.5 24.7
153 12.8 10.1 15.1



#51
3-Nitroaniline
Concen: 8.808 ng/ul
RT: 14.287 min Scan# 1904
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17

Tgt Ion:138 Resp: 75396
Ion Ratio Lower Upper
138 100
108 29.9 11.0 16.4#
92 123.2 104.6 156.8





#52

Acenaphthene

Concen: 9.723 ng/ul

RT: 14.445 min Scan# 1931

Delta R.T. -0.006 min

Lab File: BP010775.D

Acq: 23 Jun 2022 15:17

Instrument :

BNA_P

ClientSampleId :

SSTD010611





Tgt Ion:153 Resp: 513068

Ion Ratio Lower Upper

153 100

152 47.8 36.2 54.2

154 88.3 70.8 106.2



#53

2,4-Dinitrophenol

Concen: 6.544 ng/ul

RT: 14.487 min Scan# 1938

Delta R.T. -0.012 min

Lab File: BP010775.D

Acq: 23 Jun 2022 15:17




Tgt Ion:184 Resp: 18786


Ion Ratio Lower Upper

184 100

63 79.1 73.8 110.6

154 70.7 57.2 85.8









#55
4-Nitrophenol
Concen: 9.336 ng/ul
RT: 14.592 min Scan# 1956
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:109 Resp: 57661
Ion Ratio Lower Upper
109 100
139 135.6 75.6 113.4#
65 139.1 107.6 161.4



#56
Dibenzofuran
Concen: 9.568 ng/ul
RT: 14.781 min Scan# 1988
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17

Tgt Ion:168 Resp: 709324
Ion Ratio Lower Upper
168 100
139 38.1 32.8 49.2









#57
2,4-Dinitrotoluene
Concen: 9.427 ng/ul
RT: 14.745 min Scan# 1984
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:165 Resp: 86931
Ion Ratio Lower Upper
165 100
63 50.0 47.0 70.6
182 3.1 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 9.009 ng/ul
RT: 15.010 min Scan# 2027
Delta R.T. -0.006 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17

Tgt Ion:232 Resp: 94575
Ion Ratio Lower Upper
232 100
131 50.6 35.8 53.8
130 2.8 1.8 2.8#
166 31.3 23.7 35.5









#59
Diethylphthalate
Concen: 9.801 ng/ul
RT: 15.222 min Scan# 2063
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:149 Resp: 550706
Ion Ratio Lower Upper
149 100
177 21.1 17.6 26.4
150 12.6 10.3 15.5



#61
Fluorene
Concen: 9.503 ng/ul
RT: 15.434 min Scan# 2099
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17

Tgt Ion:166 Resp: 561007
Ion Ratio Lower Upper
166 100
165 97.2 77.4 116.2
167 13.1 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 9.027 ng/ul
RT: 15.439 min Scan# 2100
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:204 Resp: 262093
Ion Ratio Lower Upper
204 100
206 32.2 26.0 39.0
141 63.8 49.3 73.9



#63
4-Nitroaniline
Concen: 9.591 ng/ul
RT: 15.451 min Scan# 2102
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17

Tgt Ion:138 Resp: 77393
Ion Ratio Lower Upper
138 100
92 57.2 46.2 69.2
108 87.7 85.8 128.8









#66
4,6-Dinitro-2-methylphenol
Concen: 6.469 ng/ul
RT: 15.510 min Scan# 2114
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:198 Resp: 31229
Ion Ratio Lower Upper
198 100
182 6.1 4.3 6.5
77 32.5 23.7 35.5



#67
N-Nitrosodiphenylamine
Concen: 9.873 ng/ul
RT: 15.651 min Scan# 2136
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17

Tgt Ion:169 Resp: 466837
Ion Ratio Lower Upper
169 100
168 67.5 52.6 78.8
167 36.4 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 9.030 ng/ul
RT: 16.339 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:248 Resp: 151228
Ion Ratio Lower Upper
248 100
250 100.2 80.8 121.2
141 82.8 63.9 95.9



#69
Hexachlorobenzene
Concen: 9.075 ng/ul
RT: 16.439 min Scan# 2270
Delta R.T. -0.006 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17

Tgt Ion:284 Resp: 178195
Ion Ratio Lower Upper
284 100
142 37.8 27.8 41.6
249 29.8 25.4 38.2





#70

Atrazine

Concen: 8.863 ng/ul

RT: 16.616 min Scan# 21

Delta R.T. -0.012 min

Lab File: BP010775.D

Acq: 23 Jun 2022 15:17

Instrument :

BNA_P

ClientSampleId :

SSTD010611





Tgt Ion:200 Resp: 143060

Ion Ratio Lower Upper

200 100

173 28.3 20.7 31.1

215 46.6 38.6 57.8



#71

Pentachlorophenol

Concen: 7.775 ng/ul

RT: 16.786 min Scan# 2329

Delta R.T. -0.012 min

Lab File: BP010775.D

Acq: 23 Jun 2022 15:17




Tgt Ion:266 Resp: 70325


Ion Ratio Lower Upper

266 100

264 62.4 51.4 77.2

268 64.5 53.5 80.3









#72
Phenanthrene
Concen: 9.639 ng/ul
RT: 17.186 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:178 Resp: 817013
Ion Ratio Lower Upper
178 100
179 15.5 12.7 19.1
176 19.0 14.6 22.0



#74
Anthracene
Concen: 9.714 ng/ul
RT: 17.281 min Scan# 2413
Delta R.T. -0.006 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17

Tgt Ion:178 Resp: 827678
Ion Ratio Lower Upper
178 100
179 14.9 12.2 18.2
176 18.6 14.5 21.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 9.443 ng/uL
RT: 13.169 min Scan# 1714
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:216 Resp: 256742
Ion Ratio Lower Upper
216 100
214 78.5 65.0 97.4
218 48.0 39.1 58.7



#76
Pentachlorobenzene
Concen: 9.291 ng/uL
RT: 14.698 min Scan# 1974
Delta R.T. -0.006 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17

Tgt Ion:250 Resp: 220138
Ion Ratio Lower Upper
250 100
248 61.2 49.8 74.6
252 62.5 51.2 76.8









#77
 Carbazole
 Concen: 9.857 ng/ul
 RT: 17.551 min Scan# 2461
 Delta R.T. -0.012 min
 Lab File: BP010775.D
 Acq: 23 Jun 2022 15:17




Instrument :
 BNA_P
 ClientSampleId :
 SSTD010611


Tgt Ion:167 Resp: 740472
 Ion Ratio Lower Upper
 167 100
 166 21.0 16.9 25.3
 139 13.3 10.6 16.0



#78
 Di-n-butylphthalate
 Concen: 10.026 ng/ul
 RT: 18.133 min Scan# 2558
 Delta R.T. -0.012 min
 Lab File: BP010775.D
 Acq: 23 Jun 2022 15:17

Tgt Ion:149 Resp: 851455
 Ion Ratio Lower Upper
 149 100
 150 8.8 7.1 10.7
 104 5.1 4.6 6.8









#80
Fluoranthene
Concen: 10.202 ng/ul
RT: 19.169 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:202 Resp: 862524
Ion Ratio Lower Upper
202 100
101 13.9 7.1 10.7#
100 10.6 5.7 8.5#



#82
Pyrene
Concen: 10.308 ng/ul
RT: 19.522 min Scan# 2794
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17

Tgt Ion:202 Resp: 884669
Ion Ratio Lower Upper
202 100
101 16.3 7.9 11.9#
100 13.3 6.7 10.1#









#83
Butylbenzylphthalate
Concen: 9.461 ng/ul
RT: 20.427 min Scan# 2949
Delta R.T. -0.006 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:149 Resp: 258603
Ion Ratio Lower Upper
149 100
91 72.9 60.0 90.0
206 19.9 17.4 26.2



#84
3,3'-Dichlorobenzidine
Concen: 8.835 ng/ul
RT: 21.186 min Scan# 3077
Delta R.T. -0.006 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17

Tgt Ion:252 Resp: 243464
Ion Ratio Lower Upper
252 100
254 63.7 51.0 76.4
126 12.0 8.6 12.8





#85

Benzo(a)anthracene

Concen: 9.475 ng/ul

RT: 21.245 min Scan# 3088

Delta R.T. -0.006 min

Lab File: BP010775.D

Acq: 23 Jun 2022 15:17

Instrument :

BNA_P

ClientSampleId :

SSTD010611





Tgt Ion:228 Resp: 799220

Ion Ratio Lower Upper

228 100

229 19.4 15.0 22.4

226 26.3 20.6 31.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 9.623 ng/ul

RT: 21.204 min Scan# 3080

Delta R.T. -0.006 min

Lab File: BP010775.D

Acq: 23 Jun 2022 15:17




Tgt Ion:149 Resp: 442242


Ion Ratio Lower Upper

149 100

167 26.1 22.6 33.8

279 4.2 4.2 6.4#









#87
Chrysene
Concen: 9.499 ng/ul
RT: 21.298 min Scan# 3097
Delta R.T. -0.006 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:228 Resp: 772549
Ion Ratio Lower Upper
228 100
226 29.7 23.6 35.4
229 18.9 15.6 23.4



#89
Di-n-octyl phthalate
Concen: 9.154 ng/ul
RT: 22.127 min Scan# 3237
Delta R.T. -0.006 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17

Tgt Ion:149 Resp: 682966









#90
Benzo(b)fluoranthene
Concen: 9.200 ng/ul
RT: 22.910 min Scan# 3370
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:252 Resp: 827752
Ion Ratio Lower Upper
252 100
253 22.6 17.8 26.6
125 11.9 6.2 9.2#



#91
Benzo(k)fluoranthene
Concen: 9.397 ng/ul
RT: 22.957 min Scan# 3378
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17

Tgt Ion:252 Resp: 791759
Ion Ratio Lower Upper
252 100
253 21.8 17.0 25.4
125 12.6 6.2 9.4#









#93
Benzo(a)pyrene
Concen: 9.233 ng/ul
RT: 23.527 min Scan# 3477
Delta R.T. -0.012 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:252 Resp: 815994
Ion Ratio Lower Upper
252 100
253 22.1 17.4 26.0
125 13.9 6.6 10.0#



#94
Indeno(1,2,3-cd)pyrene
Concen: 9.263 ng/ul
RT: 26.098 min Scan# 3912
Delta R.T. -0.023 min
Lab File: BP010775.D
Acq: 23 Jun 2022 15:17

Tgt Ion:276 Resp: 1009102
Ion Ratio Lower Upper
276 100
138 27.7 14.6 22.0#
277 24.7 19.6 29.4









#95
 Dibenzo(a,h)anthracene
 Concen: 9.352 ng/ul
 RT: 26.127 min Scan# 3921
 Delta R.T. -0.023 min
 Lab File: BP010775.D
 Acq: 23 Jun 2022 15:17

Instrument :
 BNA_P
 ClientSampleId :
 SSTD010611

Tgt Ion:278 Resp: 871086
 Ion Ratio Lower Upper
 278 100
 139 19.3 9.4 14.2#
 279 23.8 18.9 28.3



#96
 Benzo(g,h,i)perylene
 Concen: 9.299 ng/ul
 RT: 26.851 min Scan# 4040
 Delta R.T. -0.023 min
 Lab File: BP010775.D
 Acq: 23 Jun 2022 15:17

Tgt Ion:276 Resp: 859928
 Ion Ratio Lower Upper
 276 100
 138 26.0 12.7 19.1#
 277 24.0 18.2 27.4

