



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062322\
 Data File : BP010773.D
 Acq On : 23 Jun 2022 14:10
 Operator : CG/JU
 Sample : DFTPP660
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP660

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Title : SVOA CALIBRATION
 Last Update : Thu Jun 23 18:30:35 2022



AutoFind: Scans 2471, 2472, 2473; Background Corrected with Scan 2466

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.6	253709	PASS
68	69	0.00	2	1.7	4322	PASS
69	198	0.00	100	42.2	256863	PASS
70	69	0.00	2	0.5	1300	PASS
127	198	10	80	51.7	314887	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	609303	PASS
199	198	5	9	6.8	41707	PASS
275	198	10	60	27.3	166603	PASS
365	198	1	100	3.7	22247	PASS
441	198	0.01	100	15.3	93346	PASS
442	198	50	100	96.7	589259	PASS
443	442	15	24	19.6	115459	PASS

m/z	Abundance
36.00	829.0
36.80	125.0
37.90	433.0
39.00	891.0
40.00	2273.0
41.10	1802.0
42.00	544.0
43.00	1395.0
44.00	11536.0
45.00	407.0
46.00	122.0
47.00	201.0
48.00	77.0
49.00	751.0
49.90	379.0
51.00	584.0
52.10	242.0
52.90	302.0
54.00	298.0
55.10	1384.0
56.00	604.0
57.00	950.0
58.00	219.0
59.00	270.0
60.00	68.0
60.70	7.0
61.30	22.0
61.80	24.0
63.10	254.0
64.00	216.0
64.90	453.0
65.90	502.0
67.00	777.0
68.00	531.0
69.00	716.0
69.90	271.0
70.10	278.0
70.90	229.0
72.10	58.0
73.00	1120.0
74.00	162.0
74.90	290.0
76.20	109.0
77.00	624.0
77.90	386.0
78.90	474.0
80.00	250.0
81.10	379.0
81.90	420.0
83.00	397.0
83.90	809.0
84.90	333.0
85.90	433.0
87.00	84.0
87.90	149.0
88.80	92.0
89.20	61.0
90.10	11.0
91.00	735.0
91.90	231.0
93.00	168.0
94.00	480.0
95.20	334.0
96.10	932.0
97.10	576.0
98.10	715.0
98.90	96.0
99.20	57.0
99.90	103.0
101.00	35.0
102.00	127.0
102.90	372.0
103.80	224.0
104.90	344.0
105.80	64.0
107.00	284.0
107.80	185.0
108.90	103.0
109.90	102.0
111.10	312.0
112.00	270.0
112.70	18.0
113.90	20.0
115.00	311.0
115.80	14.0
116.20	57.0
117.10	230.0
118.10	88.0
119.00	368.0
120.00	106.0
120.80	303.0
122.10	74.0

Instrument :
BNA_P
ClientSampleId :
DFTPP660

122.90	75.0
123.90	24.0
125.10	207.0
125.90	164.0
127.20	164.0
128.20	49.0
128.90	141.0
129.80	124.0
130.90	184.0
131.90	135.0
132.90	704.0
133.80	114.0
135.10	292.0
136.10	108.0
137.10	63.0
138.10	70.0
139.00	81.0
139.90	44.0
141.00	90.0
142.00	28.0
142.70	25.0
143.30	57.0
145.10	195.0
145.90	42.0
147.10	403.0
148.10	99.0
148.90	401.0
149.60	18.0
150.60	23.0
152.10	67.0
153.00	78.0
154.20	80.0
155.00	111.0
156.00	50.0
157.00	143.0
157.70	19.0
158.80	37.0
160.10	17.0
160.80	126.0
163.00	277.0
163.70	22.0
165.00	182.0
165.80	156.0
167.00	209.0
167.90	46.0
168.90	33.0
170.10	28.0
171.10	21.0
172.10	13.0
173.10	23.0
174.20	12.0
174.80	57.0
175.10	65.0
175.80	81.0
177.00	295.0
177.90	83.0
179.00	124.0
180.00	51.0
181.00	36.0
182.30	7.0
183.00	16.0
184.10	49.0
185.40	13.0
185.70	13.0
186.90	22.0
189.00	102.0
190.10	89.0
191.10	734.0
191.90	207.0
193.00	928.0
194.10	68.0
194.70	20.0
196.20	23.0
197.10	49.0
197.90	41.0
200.20	42.0
201.10	24.0
203.10	45.0
204.10	11.0
204.90	97.0
205.90	17.0
207.10	6237.0
208.00	1359.0
208.90	829.0
210.10	288.0
211.10	100.0
211.80	6.0
212.40	12.0
213.20	1.0
213.80	12.0
215.20	15.0
216.40	0.0
217.00	28.0
217.90	13.0
218.60	9.0
219.60	16.0

Instrument :
BNA_P
ClientSampleId :
DFTPP660

220.20	17.0
221.10	409.0
222.00	54.0
223.00	55.0
224.10	31.0
224.80	7.0
226.80	24.0
227.40	11.0
227.90	14.0
228.90	3.0
229.40	13.0
230.20	3.0
231.00	10.0
231.90	2.0
232.70	7.0
234.30	15.0
235.10	16.0
235.90	8.0
236.90	8.0
237.40	7.0
238.60	10.0
239.00	14.0
240.20	50.0
241.90	13.0
243.80	22.0
245.30	7.0
246.30	11.0
246.70	15.0
247.90	15.0
248.80	25.0
249.70	43.0
251.00	52.0
251.60	5.0
252.80	24.0
253.40	3.0
254.10	18.0
255.00	137.0
256.40	8.0
257.20	12.0
257.90	7.0
259.40	10.0
260.90	13.0
262.60	8.0
264.10	27.0
264.90	200.0
266.00	53.0
267.00	228.0
268.90	40.0
269.30	12.0
270.20	25.0
272.00	8.0
272.70	7.0
273.60	8.0
274.70	24.0
276.70	7.0
278.30	9.0
278.80	8.0
279.30	4.0
281.00	636.0
282.00	219.0
283.00	277.0
283.90	87.0
284.70	10.0
286.20	6.0
287.70	18.0
289.00	12.0
290.90	9.0
292.80	20.0
294.00	8.0
295.00	52.0
295.90	16.0
297.30	7.0
299.20	11.0
300.30	7.0
300.90	19.0
301.90	7.0
302.80	8.0
303.50	8.0
305.00	7.0
306.00	7.0
307.10	21.0
308.00	11.0
309.10	15.0
310.00	17.0
310.50	7.0
311.00	39.0
311.80	19.0
312.90	21.0
314.50	8.0
316.10	10.0
317.70	6.0
319.30	8.0
320.60	15.0
322.30	7.0
323.10	1.0
323.90	21.0

Instrument :
BNA_P
ClientSampleId :
DFTPP660

325.10	12.0
327.10	11.0
327.90	20.0
328.50	8.0
330.20	8.0
330.90	13.0
331.70	9.0
332.10	10.0
333.50	8.0
335.00	12.0
336.60	7.0
337.10	11.0
338.10	8.0
339.10	13.0
339.90	12.0
340.90	119.0
342.00	14.0
342.30	15.0
343.10	58.0
344.80	22.0
345.90	11.0
346.60	4.0
347.50	9.0
348.60	27.0
350.80	7.0
352.30	10.0
353.30	4.0
353.90	28.0
354.80	102.0
355.20	66.0
356.30	19.0
357.10	35.0
357.80	4.0
359.00	16.0
359.60	11.0
360.10	15.0
360.70	11.0
361.40	12.0
361.80	8.0
363.30	9.0
363.90	7.0
364.90	17.0
366.40	8.0
368.00	6.0
368.70	9.0
369.50	9.0
370.30	10.0
370.80	23.0
372.70	9.0
374.30	7.0
376.00	11.0
377.50	9.0
378.20	8.0
378.90	6.0
380.50	8.0
382.20	18.0
383.80	7.0
385.00	19.0
387.10	12.0
387.70	18.0
388.40	9.0
389.80	17.0
391.60	7.0
392.20	6.0
393.20	7.0
395.10	16.0
396.40	20.0
398.10	18.0
399.20	18.0
400.10	6.0
400.90	50.0
403.00	15.0
404.20	9.0
404.60	8.0
405.80	6.0
407.40	16.0
408.90	7.0
410.20	10.0
410.50	8.0
412.10	9.0
413.00	8.0
413.70	10.0
415.00	43.0
417.20	27.0
418.60	23.0
420.10	12.0
420.70	8.0
421.60	19.0
423.20	8.0
424.10	3.0
424.60	8.0
426.20	11.0
428.10	12.0
428.60	5.0
429.20	7.0
430.00	11.0

Instrument :
BNA_P
ClientSampleId :
DFTPP660

430.90	8.0
432.00	13.0
432.50	7.0
434.20	13.0
435.70	10.0
437.10	7.0
438.90	7.0
440.40	7.0
441.20	17.0
442.00	10.0
442.50	4.0
443.60	8.0
444.70	15.0
446.60	8.0
448.30	13.0
449.20	0.0
449.80	9.0
451.40	7.0
452.10	2.0
453.00	6.0
454.70	21.0
456.10	9.0
457.60	8.0
458.80	13.0
459.30	9.0
461.00	7.0
461.50	7.0
462.80	21.0
463.50	19.0
464.10	15.0
465.10	6.0
465.60	8.0
466.90	9.0
468.50	18.0
470.10	20.0
472.00	9.0
473.30	5.0
474.80	12.0
475.70	15.0
476.70	13.0
478.10	7.0
479.60	19.0
480.00	17.0
480.60	5.0
481.40	8.0
482.80	6.0
484.50	9.0
486.10	6.0
487.60	8.0
489.30	17.0
490.20	9.0
491.00	20.0
492.10	18.0
493.30	8.0
494.00	8.0
495.40	6.0
497.00	7.0
498.30	7.0
498.60	7.0
500.20	13.0
500.80	2.0
501.70	18.0
503.50	11.0
504.90	6.0
506.50	7.0
508.00	10.0
509.60	6.0
511.30	8.0
513.10	14.0
514.30	12.0
515.90	8.0
517.50	8.0
519.10	8.0
519.90	2.0
520.90	17.0
522.20	7.0
523.80	6.0
525.40	7.0
526.90	7.0
528.40	8.0
529.20	12.0
530.10	8.0
531.70	7.0
533.10	13.0
534.90	17.0
535.80	6.0
536.50	7.0
537.90	20.0
539.60	7.0
541.00	12.0
542.70	7.0
544.30	9.0
545.80	9.0
547.40	8.0
548.90	37.0

Instrument :
BNA_P
ClientSampleId :
DFTPP660

m/z	Abundance
35.00	163.0
36.00	662.0
37.00	1084.0
38.10	3065.0
39.10	17016.0
40.00	2821.0
41.00	1881.0
42.00	929.0
43.10	1590.0
44.00	9831.0
45.00	901.0
46.00	150.0
47.00	175.0
47.90	157.0
49.10	2418.0
50.00	61016.0
51.00	245376.0
52.00	12379.0
53.00	552.0
54.10	362.0
55.00	2897.0
56.00	7200.0
57.00	17056.0
58.00	980.0
58.90	328.0
59.90	289.0
61.00	3252.0
62.00	3145.0
63.00	9005.0
64.00	1721.0
65.00	4806.0
66.00	477.0
67.00	959.0
68.00	4681.0
69.00	241280.0
70.00	2018.0
70.90	314.0
71.90	222.0
73.00	3354.0
74.00	26096.0
75.00	38464.0
76.00	13335.0
77.00	281856.0
78.00	19464.0
79.00	17872.0
80.00	13625.0
81.00	19120.0
82.00	4813.0
83.00	5045.0
83.90	1352.0
85.00	3584.0
86.00	5316.0
87.00	2960.0
87.90	1110.0
88.90	672.0
89.80	155.0
91.00	4729.0
91.90	4544.0
93.00	29864.0
94.00	2138.0
95.00	687.0
96.00	2613.0
97.00	985.0
98.00	22552.0
99.00	18120.0
100.00	1626.0
101.00	9797.0
101.90	826.0
103.00	3916.0
104.00	6787.0
105.00	5831.0
106.00	2018.0
107.00	76856.0
108.00	12942.0
109.00	2404.0
110.00	140928.0
111.00	22376.0
112.00	3072.0
113.00	1179.0
114.20	173.0
115.00	580.0
116.00	4528.0
117.00	66976.0
117.90	5620.0
119.00	958.0
120.00	1180.0
120.90	394.0
121.90	5179.0
123.00	8098.0
124.00	3517.0
125.00	3513.0
127.00	272448.0

Instrument :
BNA_P
ClientSampleId :
DFTPP660

128.00	21808.0
129.00	103976.0
130.00	9266.0
131.00	2259.0
132.00	936.0
132.90	918.0
134.00	3044.0
135.00	8724.0
136.00	3535.0
137.00	4461.0
137.90	1112.0
139.00	705.0
140.00	1456.0
141.00	13213.0
142.00	4638.0
143.00	3073.0
144.00	899.0
145.00	801.0
146.00	2369.0
147.00	7439.0
148.00	15336.0
149.00	3943.0
150.00	873.0
151.10	1934.0
153.00	4494.0
154.00	3519.0
155.00	7438.0
156.00	11376.0
156.90	2025.0
158.00	2640.0
159.00	2022.0
160.00	3923.0
161.00	6419.0
162.00	1943.0
163.00	812.0
164.00	783.0
164.90	5512.0
166.00	3392.0
167.00	27264.0
168.00	16077.0
169.00	2766.0
169.90	834.0
170.30	594.0
170.90	927.0
172.00	2513.0
173.00	3092.0
174.00	5210.0
175.00	9941.0
176.00	3273.0
177.00	4724.0
177.90	1718.0
179.00	17504.0
180.00	13524.0
181.00	6004.0
182.00	1200.0
182.60	435.0
183.10	475.0
183.90	1498.0
185.00	8841.0
186.00	68800.0
187.00	19880.0
188.00	2288.0
189.00	4566.0
189.90	819.0
191.00	2686.0
191.90	5379.0
193.00	6919.0
194.00	1379.0
194.90	1289.0
196.00	14722.0
198.00	484736.0
199.00	34664.0
200.00	2829.0
201.50	2250.0
203.00	3163.0
204.00	16378.0
205.00	28232.0
206.00	125688.0
207.00	24744.0
208.00	5058.0
209.10	2355.0
210.00	2146.0
211.10	4569.0
213.00	598.0
213.70	196.0
215.00	1332.0
216.00	2438.0
217.00	32712.0
218.00	4644.0
218.90	406.0
221.00	19208.0
221.70	4265.0
223.00	7257.0
224.00	68272.0
225.00	17952.0
226.00	2053.0

Instrument :
BNA_P
ClientSampleId :
DFTPP660

227.00	27560.0
228.00	4491.0
229.00	6189.0
230.00	701.0
230.90	2126.0
232.00	493.0
232.90	462.0
234.00	1623.0
235.00	2214.0
235.90	1422.0
236.90	2844.0
237.90	524.0
238.90	1043.0
240.00	739.0
241.00	1646.0
242.00	3671.0
243.00	3938.0
244.00	52224.0
245.00	7745.0
246.00	9937.0
247.00	2019.0
247.90	571.0
249.00	2270.0
250.10	484.0
250.90	321.0
251.90	546.0
253.00	839.0
255.00	259712.0
256.00	39648.0
257.00	3373.0
258.00	14455.0
259.00	2382.0
260.00	480.0
260.90	572.0
261.80	46.0
263.00	189.0
264.00	415.0
265.00	6127.0
265.90	955.0
267.00	279.0
268.00	112.0
269.00	99.0
270.00	412.0
271.00	490.0
272.00	551.0
273.00	6885.0
274.00	21240.0
275.00	117664.0
276.00	16440.0
277.00	9049.0
277.90	1899.0
278.90	335.0
279.90	75.0
281.00	968.0
281.90	757.0
283.00	1413.0
284.00	951.0
285.00	1622.0
286.00	179.0
287.00	106.0
287.80	206.0
288.90	510.0
290.00	281.0
290.90	231.0
291.90	593.0
292.90	2513.0
293.80	752.0
296.00	32496.0
297.00	4398.0
297.90	321.0
299.10	145.0
299.80	13.0
300.90	300.0
301.90	663.0
303.00	3767.0
304.10	1163.0
305.00	137.0
306.00	59.0
306.60	18.0
307.20	24.0
307.90	586.0
309.10	490.0
309.90	359.0
310.80	189.0
312.00	79.0
312.90	321.0
313.90	1446.0
314.90	3569.0
316.00	2374.0
317.00	387.0
317.80	12.0
318.60	16.0
319.20	22.0
320.20	105.0
321.10	948.0
322.00	531.0

Instrument :
BNA_P
ClientSampleId :
DFTPP660

323.00	10024.0
324.00	2051.0
324.90	223.0
325.90	203.0
326.90	2160.0
328.00	1178.0
329.00	228.0
329.90	129.0
331.90	592.0
333.00	1070.0
334.00	6440.0
335.00	1960.0
335.90	227.0
336.80	21.0
337.90	7.0
338.90	159.0
339.80	113.0
340.90	1047.0
342.00	490.0
343.00	86.0
344.10	54.0
344.90	103.0
345.90	2615.0
346.90	863.0
347.90	50.0
348.90	15.0
349.60	12.0
350.20	47.0
351.90	2883.0
353.00	2246.0
354.00	3150.0
355.00	722.0
355.90	192.0
356.50	21.0
357.00	33.0
358.10	92.0
359.00	345.0
359.90	83.0
360.50	18.0
361.10	45.0
361.70	15.0
362.30	14.0
362.80	62.0
364.90	13391.0
366.00	2125.0
366.70	214.0
367.80	7.0
369.90	311.0
371.00	912.0
372.00	4934.0
372.90	1443.0
374.00	276.0
375.70	15.0
376.90	177.0
378.00	10.0
379.00	20.0
380.40	8.0
381.10	7.0
381.90	71.0
382.90	1573.0
383.80	458.0
384.80	251.0
386.00	53.0
386.60	9.0
387.10	14.0
387.40	11.0
388.00	17.0
388.40	13.0
390.00	732.0
390.90	468.0
392.00	311.0
392.90	22.0
394.80	119.0
395.70	15.0
396.20	20.0
397.10	44.0
397.90	15.0
398.60	12.0
399.30	7.0
399.70	12.0
400.70	191.0
402.00	2179.0
403.00	3458.0
404.00	1120.0
405.10	249.0
406.00	56.0
407.20	12.0
407.60	7.0
408.40	6.0
408.90	14.0
410.00	97.0
410.70	17.0
411.90	7.0
413.00	7.0
413.40	8.0
414.90	147.0

Instrument :
BNA_P
ClientSampleId :
DFTPP660

415.90	35.0
416.60	6.0
417.10	2.0
418.10	10.0
418.60	16.0
421.00	2610.0
422.00	2781.0
423.00	18368.0
424.00	4251.0
425.00	587.0
426.00	12.0
427.60	15.0
428.40	3.0
429.20	25.0
430.10	2.0
430.80	10.0
432.30	7.0
432.90	25.0
433.90	13.0
434.50	16.0
435.40	16.0
437.40	19.0
438.80	112.0
439.20	89.0
441.00	54408.0
442.00	335040.0
443.00	70600.0
444.00	6496.0
445.00	551.0
445.70	3.0
446.50	9.0
447.10	16.0
448.00	7.0
449.50	8.0
451.20	8.0
451.90	0.0
452.70	8.0
453.70	9.0
454.40	9.0
455.70	11.0
456.60	10.0
457.40	7.0
459.20	22.0
459.60	16.0
460.80	11.0
461.70	17.0
462.70	21.0
463.80	18.0
465.30	9.0
467.00	6.0
468.00	1.0
468.60	10.0
470.00	7.0
471.10	9.0
471.70	7.0
473.30	7.0
474.90	92.0
475.90	7.0
476.40	7.0
477.90	8.0
478.80	15.0
479.80	23.0
481.10	8.0
482.70	7.0
483.50	11.0
484.30	7.0
485.80	6.0
486.50	4.0
487.30	7.0
488.10	2.0
489.10	9.0
490.20	12.0
491.20	7.0
492.20	8.0
493.20	5.0
493.80	6.0
495.30	9.0
496.90	8.0
497.40	9.0
498.60	13.0
499.50	8.0
499.90	7.0
501.80	21.0
503.00	18.0
503.80	9.0
505.00	18.0
506.20	7.0
507.80	9.0
509.40	8.0
511.00	5.0
512.60	7.0
514.10	8.0
514.70	8.0
515.70	7.0
517.40	7.0
518.90	7.0

Instrument :
BNA_P
ClientSampleId :
DFTPP660

520.80	21.0
522.00	8.0
523.60	8.0
525.60	9.0
526.80	7.0
528.30	8.0
529.80	12.0
531.40	8.0
533.20	8.0
534.60	8.0
535.30	4.0
536.20	8.0
537.00	7.0
537.70	9.0
539.30	9.0
541.00	8.0
542.60	17.0
544.00	8.0
545.60	7.0
547.00	16.0
548.90	8.0
549.70	12.0

Instrument :
BNA_P
ClientSampleId :
DFTPP660

m/z	Abundance
36.00	706.0
37.00	1477.0
38.00	3855.0
39.00	23256.0
40.00	2018.0
41.10	2284.0
41.90	788.0
43.00	1325.0
44.00	9510.0
45.00	906.0
45.90	182.0
47.00	219.0
48.00	374.0
49.10	2694.0
50.00	79504.0
51.00	307008.0
52.00	15953.0
53.00	1292.0
54.00	329.0
55.00	2997.0
56.00	9775.0
57.00	22344.0
58.00	1075.0
58.90	373.0
59.90	333.0
61.00	4623.0
62.00	4577.0
63.00	12761.0
64.00	1945.0
65.00	6737.0
66.10	683.0
67.10	1154.0
68.00	5638.0
69.00	316096.0
70.00	2031.0
71.10	519.0
72.00	134.0
73.10	3576.0
74.00	34096.0
75.00	50056.0
76.00	17920.0
77.00	364672.0
78.00	25136.0
79.00	24976.0
80.00	17848.0
81.00	24768.0
82.00	6436.0
83.00	5685.0
83.90	1327.0
85.00	4754.0
86.00	7210.0
87.00	3184.0
88.00	1578.0
88.90	787.0
90.10	219.0
91.00	6321.0
92.00	6898.0
93.00	38632.0
94.00	2775.0
95.00	1061.0
96.00	2746.0
97.00	1211.0
98.00	31528.0
99.00	25840.0
100.00	2127.0
101.00	13341.0
101.90	816.0
103.00	4932.0
104.00	9139.0
105.00	8711.0
106.00	2946.0
107.00	105384.0
108.00	16808.0
109.00	3332.0
110.00	200512.0
111.00	30272.0
112.00	3754.0
113.00	1351.0
113.80	240.0
115.00	780.0
116.00	6266.0
117.00	88632.0
118.00	6702.0
119.00	965.0
119.90	1454.0
120.90	554.0
122.00	7291.0
123.00	11539.0
124.00	5469.0
125.00	4637.0
127.00	384640.0
128.00	29856.0

Instrument :
BNA_P
ClientSampleId :
DFTPP660

129.00	148160.0
130.00	12589.0
131.00	2481.0
132.00	1231.0
133.00	1339.0
134.00	4563.0
135.00	12471.0
136.00	4974.0
137.00	6248.0
137.90	1474.0
139.00	716.0
140.00	2033.0
141.00	19712.0
142.00	6713.0
143.00	4745.0
144.00	1333.0
145.00	1492.0
146.00	3834.0
147.00	10679.0
148.00	23880.0
149.00	5000.0
149.90	1458.0
151.00	2354.0
151.80	1544.0
153.00	6617.0
154.00	4429.0
155.00	10290.0
156.00	15793.0
157.10	3030.0
158.00	3238.0
159.00	2992.0
160.00	5627.0
161.00	7953.0
162.00	2463.0
163.00	1057.0
164.00	1067.0
165.00	7317.0
166.00	5020.0
167.00	36984.0
168.00	20648.0
169.00	3519.0
170.10	1070.0
171.00	1591.0
171.90	3372.0
173.00	4596.0
174.00	7672.0
175.00	13923.0
176.00	4274.0
177.00	6165.0
178.00	2428.0
179.00	28152.0
180.00	19456.0
181.00	9420.0
182.00	1339.0
182.90	980.0
183.90	2036.0
185.00	13715.0
186.00	103048.0
187.00	30232.0
188.00	3085.0
189.00	7544.0
190.00	1440.0
191.00	4060.0
192.00	8796.0
193.00	10104.0
194.00	2277.0
195.00	1558.0
196.00	21288.0
198.00	753728.0
199.00	49448.0
200.00	3641.0
201.40	3660.0
203.00	4419.0
204.00	26720.0
205.00	46184.0
206.00	197056.0
207.00	32056.0
208.00	8119.0
209.00	2619.0
210.20	3374.0
211.00	7574.0
212.80	492.0
213.90	312.0
214.90	1713.0
216.00	4192.0
217.00	49200.0
218.00	6966.0
218.90	628.0
221.00	33080.0
223.00	10559.0
224.00	109064.0
225.00	27752.0
226.00	2728.0
227.00	44720.0
228.00	6848.0
229.00	9023.0

Instrument :
BNA_P
ClientSampleId :
DFTPP660

230.00	1423.0
231.00	3892.0
232.00	699.0
233.10	889.0
234.00	2894.0
235.00	3347.0
236.00	2208.0
237.00	3803.0
237.90	781.0
239.00	2058.0
239.90	1285.0
241.00	2308.0
242.00	5878.0
243.00	6400.0
244.00	85024.0
245.00	11740.0
246.00	16207.0
247.00	3107.0
247.90	703.0
249.00	3578.0
250.00	657.0
251.00	697.0
252.00	708.0
252.90	1935.0
255.00	441280.0
256.00	63136.0
257.00	4945.0
258.00	24416.0
259.00	4122.0
259.90	850.0
260.90	806.0
261.80	61.0
263.00	165.0
264.00	701.0
265.00	10584.0
266.00	1500.0
266.80	348.0
267.90	217.0
269.00	214.0
269.90	412.0
271.00	1000.0
272.10	1514.0
273.00	14590.0
274.00	34664.0
275.00	200832.0
276.00	26752.0
277.00	16848.0
277.90	2993.0
279.00	716.0
279.90	99.0
281.00	1399.0
282.00	641.0
283.00	1653.0
284.00	1612.0
285.00	2943.0
285.90	519.0
286.80	89.0
287.10	104.0
287.90	273.0
289.00	640.0
289.90	750.0
290.90	577.0
291.90	821.0
293.00	3799.0
293.90	1309.0
296.00	56400.0
297.00	8302.0
298.00	553.0
298.80	97.0
300.10	65.0
301.00	771.0
302.00	1038.0
303.00	6171.0
304.00	1629.0
305.00	213.0
306.00	56.0
307.00	111.0
307.90	658.0
308.90	495.0
310.00	699.0
311.00	218.0
311.90	219.0
312.90	589.0
314.10	2737.0
314.90	5549.0
316.00	3624.0
316.90	531.0
317.80	6.0
318.40	23.0
319.00	149.0
320.00	343.0
320.90	1754.0
321.90	899.0
323.00	18296.0
324.00	3477.0
325.00	412.0

Instrument :
BNA_P
ClientSampleId :
DFTPP660

326.00	501.0
327.00	4019.0
327.90	1886.0
328.90	274.0
330.00	83.0
331.20	120.0
331.90	1551.0
333.00	1940.0
334.00	11758.0
335.00	3374.0
335.90	507.0
336.80	7.0
337.30	26.0
338.10	12.0
338.80	353.0
339.90	313.0
341.00	2107.0
342.00	580.0
343.00	92.0
343.70	13.0
346.00	3989.0
346.90	895.0
347.90	244.0
348.40	1.0
350.10	71.0
351.00	285.0
352.00	5380.0
353.00	4109.0
354.00	6331.0
355.00	1495.0
356.00	315.0
356.90	15.0
358.00	123.0
359.00	571.0
360.00	85.0
360.80	155.0
362.30	36.0
362.80	85.0
363.20	73.0
364.90	27632.0
366.00	4076.0
366.90	358.0
367.70	11.0
368.60	12.0
368.90	16.0
370.00	610.0
371.00	1241.0
372.00	10367.0
373.00	2421.0
373.90	247.0
374.90	41.0
375.60	0.0
376.90	275.0
377.70	57.0
378.30	8.0
378.90	16.0
379.40	8.0
381.00	22.0
383.00	3226.0
384.00	592.0
384.90	295.0
385.90	53.0
387.10	14.0
387.90	52.0
390.00	1417.0
390.90	911.0
392.00	897.0
393.00	168.0
393.50	18.0
394.80	77.0
395.50	26.0
396.10	47.0
397.00	125.0
398.10	16.0
398.70	13.0
400.00	25.0
400.90	613.0
402.00	4349.0
403.00	5735.0
403.90	2385.0
405.00	439.0
406.00	69.0
407.60	13.0
408.20	4.0
408.90	19.0
410.00	210.0
410.80	31.0
412.50	7.0
414.00	14.0
415.00	281.0
415.90	112.0
416.80	7.0
417.20	16.0
418.40	18.0
419.30	12.0
421.00	5016.0

Instrument :
BNA_P
ClientSampleId :
DFTPP660

422.00	4880.0
423.00	35752.0
424.00	7485.0
425.00	1219.0
426.00	34.0
426.80	10.0
427.60	7.0
428.10	8.0
429.00	17.0
430.00	35.0
431.00	10.0
432.90	8.0
433.80	5.0
434.50	9.0
435.30	10.0
436.00	17.0
437.20	23.0
438.10	33.0
439.20	208.0
441.00	106296.0
442.00	681280.0
443.00	133568.0
444.00	14452.0
445.00	646.0
446.00	78.0
446.90	13.0
448.60	8.0
450.20	7.0
451.10	0.0
451.90	7.0
453.30	9.0
454.80	17.0
456.50	12.0
457.90	15.0
458.70	12.0
459.70	10.0
461.00	24.0
461.80	15.0
462.60	15.0
464.60	14.0
465.90	8.0
467.40	6.0
468.00	9.0
468.80	17.0
470.60	9.0
472.20	8.0
472.60	10.0
473.80	9.0
474.90	87.0
475.50	28.0
476.20	15.0
477.10	23.0
478.50	14.0
480.10	6.0
481.70	10.0
482.30	1.0
483.20	7.0
484.00	3.0
484.80	8.0
486.40	8.0
488.00	9.0
489.00	23.0
489.80	5.0
491.10	7.0
492.80	14.0
494.20	8.0
495.80	8.0
497.40	9.0
498.90	6.0
500.30	24.0
502.20	15.0
503.00	14.0
503.60	13.0
504.70	1.0
505.20	16.0
506.90	9.0
507.50	3.0
508.50	7.0
510.00	9.0
511.60	8.0
513.10	8.0
514.80	7.0
515.90	14.0
516.30	7.0
517.90	8.0
519.60	21.0
519.90	20.0
521.20	15.0
522.70	7.0
523.40	4.0
524.10	7.0
525.50	22.0
527.40	10.0
528.90	8.0
530.50	8.0
532.10	7.0

Instrument :
BNA_P
ClientSampleId :
DFTPP660

533.10	12.0
533.60	7.0
535.10	18.0
536.00	1.0
536.70	9.0
537.70	1.0
538.20	7.0
539.90	6.0
541.10	13.0
541.40	8.0
543.00	9.0
544.60	9.0
546.20	9.0
546.90	3.0
547.80	17.0
549.50	9.0

Instrument :
BNA_P
ClientSampleId :
DFTPP660

m/z	Abundance
36.00	668.0
37.00	1193.0
38.10	3082.0
39.00	15275.0
40.00	2479.0
41.00	1913.0
42.00	618.0
43.00	1354.0
44.00	9475.0
44.90	594.0
46.10	110.0
46.90	203.0
48.00	189.0
49.10	2076.0
50.00	56528.0
51.00	210496.0
52.10	11334.0
53.00	901.0
54.10	377.0
55.00	2542.0
56.00	7173.0
57.00	14362.0
57.90	534.0
59.00	429.0
59.80	210.0
61.00	2941.0
62.00	3561.0
63.00	8911.0
64.00	1428.0
65.00	4716.0
66.00	607.0
67.00	748.0
68.00	4239.0
69.00	215360.0
70.00	1498.0
71.10	598.0
72.10	161.0
73.00	2698.0
74.00	24744.0
75.00	35528.0
76.10	12517.0
77.00	259136.0
78.00	18064.0
79.00	16864.0
80.00	12535.0
81.00	18664.0
82.00	5042.0
83.00	4782.0
84.10	973.0
85.00	3531.0
86.00	5198.0
87.00	2672.0
87.90	838.0
88.90	652.0
90.00	326.0
91.00	5015.0
92.00	4453.0
93.00	28584.0
94.00	2555.0
95.00	659.0
96.00	1914.0
96.90	691.0
98.00	23008.0
99.00	17440.0
100.00	1685.0
101.00	9011.0
102.00	928.0
103.00	3849.0
104.00	5798.0
105.00	6492.0
106.10	1867.0
107.00	77840.0
108.00	11562.0
109.10	2120.0
110.00	144704.0
111.00	22128.0
112.00	2883.0
113.10	823.0
114.10	292.0
115.00	614.0
116.00	4449.0
117.00	67176.0
118.00	5280.0
118.90	1112.0
120.00	1116.0
121.00	470.0
122.00	5672.0
123.00	7975.0
123.90	3548.0
125.00	3595.0
127.00	288064.0
128.00	21112.0

Instrument :
BNA_P
ClientSampleId :
DFTPP660

129.00	108256.0
130.00	8748.0
131.00	1852.0
132.00	1197.0
133.00	1175.0
134.00	3449.0
135.00	9260.0
136.00	3829.0
137.00	5487.0
137.90	1051.0
138.90	606.0
139.90	1464.0
141.00	14737.0
142.00	4986.0
143.00	3334.0
144.00	691.0
145.00	753.0
146.00	2809.0
147.00	7930.0
148.00	18400.0
149.00	3435.0
150.00	1297.0
151.00	1933.0
151.90	1300.0
153.00	4760.0
154.00	3495.0
155.00	7806.0
156.00	12176.0
157.00	2478.0
157.90	2654.0
159.00	1926.0
160.00	4328.0
161.00	6590.0
161.90	2190.0
163.00	760.0
164.10	812.0
165.00	5458.0
166.00	4514.0
167.00	29112.0
168.00	16632.0
169.00	2691.0
169.90	710.0
170.90	1582.0
171.90	2832.0
173.00	3400.0
174.00	5552.0
175.00	11416.0
176.00	3049.0
177.00	5056.0
177.90	1900.0
179.00	22304.0
180.00	14767.0
181.00	6859.0
182.00	1098.0
183.00	755.0
184.00	1875.0
185.00	9816.0
186.00	80400.0
187.00	22752.0
188.00	2420.0
189.00	5106.0
190.00	823.0
190.90	3458.0
192.00	7278.0
193.00	8362.0
194.00	1796.0
195.00	957.0
196.00	17520.0
198.00	589568.0
199.00	41008.0
200.00	2921.0
201.40	3083.0
203.00	4450.0
204.00	20608.0
205.00	36224.0
206.00	158208.0
207.00	26584.0
208.00	6513.0
209.00	2511.0
210.10	3074.0
211.00	5935.0
212.90	372.0
214.00	41.0
214.90	1837.0
216.00	3228.0
217.00	39768.0
218.00	5703.0
219.10	524.0
221.00	26144.0
221.80	5478.0
223.00	9742.0
224.00	86896.0
225.00	22184.0
226.10	2791.0
227.00	36608.0
228.00	4984.0

Instrument :
BNA_P
ClientSampleId :
DFTPP660

229.00	8802.0
230.00	1052.0
231.00	2890.0
231.90	532.0
233.00	769.0
233.90	2266.0
234.90	2986.0
236.00	1789.0
237.00	3100.0
238.00	435.0
238.90	1328.0
240.00	1535.0
241.00	2232.0
242.00	5100.0
243.10	5489.0
244.00	75160.0
245.00	9643.0
246.00	14810.0
247.00	2908.0
248.00	760.0
249.00	2932.0
250.00	418.0
250.90	856.0
251.90	831.0
253.00	1144.0
255.00	382016.0
256.00	56312.0
257.00	4279.0
258.00	21072.0
259.00	3481.0
259.80	575.0
261.00	509.0
262.00	45.0
262.90	235.0
264.00	654.0
265.00	8924.0
265.90	1256.0
267.10	472.0
267.90	173.0
269.00	127.0
270.00	371.0
271.00	895.0
272.00	1286.0
273.00	11986.0
274.00	31320.0
275.00	181312.0
276.00	24536.0
277.00	13963.0
278.00	2445.0
279.00	765.0
279.80	89.0
281.00	1028.0
281.90	331.0
282.20	347.0
283.00	1835.0
283.90	983.0
285.00	2998.0
286.00	694.0
286.80	22.0
287.90	330.0
288.90	713.0
289.90	629.0
290.90	229.0
292.10	588.0
293.00	3099.0
294.00	767.0
296.00	48152.0
297.00	6861.0
297.80	526.0
300.10	89.0
301.00	584.0
302.00	997.0
303.00	6265.0
304.00	1490.0
305.00	197.0
306.00	38.0
306.90	180.0
307.90	971.0
309.00	336.0
309.90	671.0
311.00	232.0
311.90	210.0
312.90	595.0
314.10	2654.0
315.00	5597.0
316.00	3240.0
317.00	417.0
318.00	25.0
318.90	122.0
320.00	303.0
321.00	1995.0
322.00	856.0
323.00	17104.0
323.90	2920.0
324.90	586.0
325.90	248.0

Instrument :
BNA_P
ClientSampleId :
DFTPP660

327.00	3595.0
328.00	1815.0
328.90	291.0
329.70	30.0
330.10	28.0
332.00	1247.0
333.00	1682.0
334.00	11134.0
335.00	2554.0
336.00	283.0
337.20	13.0
337.70	15.0
338.90	289.0
340.00	348.0
341.00	2183.0
341.90	437.0
343.10	73.0
344.00	12.0
345.90	4380.0
346.90	864.0
347.80	139.0
350.20	101.0
350.90	261.0
352.00	5597.0
353.00	3718.0
354.00	5596.0
355.00	1342.0
356.00	197.0
356.90	75.0
358.00	69.0
359.00	371.0
359.80	126.0
360.90	119.0
362.20	26.0
362.80	61.0
365.00	25768.0
366.00	3847.0
366.80	479.0
367.90	10.0
368.80	20.0
369.90	571.0
371.00	1296.0
372.00	9938.0
373.00	2418.0
374.10	357.0
374.90	73.0
377.00	198.0
377.90	65.0
378.50	14.0
380.00	7.0
381.60	20.0
382.90	2753.0
384.00	887.0
384.90	254.0
386.20	16.0
387.20	7.0
387.90	59.0
389.10	94.0
389.90	1308.0
391.10	1118.0
392.00	633.0
393.00	128.0
393.60	18.0
394.20	11.0
395.10	109.0
395.60	40.0
397.00	97.0
397.70	10.0
398.90	9.0
399.90	25.0
401.00	779.0
402.00	4554.0
403.00	6190.0
403.90	2633.0
404.90	423.0
406.20	13.0
406.70	8.0
407.70	9.0
408.20	11.0
408.70	9.0
410.00	232.0
411.20	20.0
412.40	15.0
412.90	8.0
413.90	8.0
414.90	271.0
415.80	184.0
416.70	21.0
418.00	13.0
418.60	10.0
419.00	8.0
421.00	5345.0
422.00	5635.0
423.00	40672.0
424.00	8501.0
425.10	880.0

Instrument :
BNA_P
ClientSampleId :
DFTPP660

426.00	43.0
426.50	1.0
427.20	60.0
428.10	19.0
429.00	43.0
430.00	33.0
430.60	2.0
431.70	12.0
432.40	12.0
433.90	25.0
434.90	10.0
435.90	14.0
436.90	28.0
437.50	5.0
438.10	83.0
441.00	119384.0
442.00	751488.0
443.00	142208.0
444.00	14262.0
445.00	699.0
446.00	71.0
447.60	14.0
449.20	8.0
450.70	9.0
451.20	5.0
452.30	7.0
453.90	10.0
455.60	8.0
457.10	7.0
457.70	14.0
458.60	13.0
459.30	0.0
460.10	20.0
461.20	9.0
461.60	7.0
462.90	15.0
465.00	7.0
466.20	12.0
468.20	14.0
469.10	1.0
469.70	9.0
471.20	7.0
471.90	1.0
472.80	9.0
473.80	8.0
474.40	6.0
475.00	33.0
476.00	25.0
477.00	17.0
477.60	13.0
479.10	9.0
480.70	7.0
481.80	9.0
482.30	6.0
483.80	8.0
485.40	8.0
487.10	14.0
488.90	19.0
490.20	21.0
491.00	20.0
491.60	8.0
492.30	2.0
493.30	6.0
494.80	9.0
496.40	9.0
497.90	6.0
499.50	8.0
501.00	7.0
503.10	23.0
504.30	9.0
505.10	13.0
505.80	8.0
507.40	9.0
509.00	8.0
510.60	9.0
512.10	8.0
513.60	16.0
515.20	25.0
516.20	0.0
516.90	9.0
518.30	10.0
520.10	9.0
521.60	7.0
522.00	9.0
523.10	9.0
524.70	8.0
526.10	8.0
527.80	7.0
529.40	8.0
531.00	7.0
532.70	11.0
534.20	7.0
535.00	11.0
536.00	16.0
537.30	8.0
538.80	6.0

Instrument :
BNA_P
ClientSampleId :
DFTPP660

540.50	6.0
542.10	7.0
544.00	12.0
545.30	16.0
547.00	14.0
548.30	8.0
548.70	18.0

Instrument :
BNA_P
ClientSampleId :
DFTPP660