

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062323\
 Data File : BP015668.D
 Acq On : 23 Jun 2023 13:13
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020763

Quant Time: Jun 23 22:34:08 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jun 22 22:01:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.081	152	141383	20.000	ng/u1	0.00
20) Naphthalene-d8	10.910	136	575457	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.728	164	341095	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.492	188	630841	20.000	ng/u1	0.00
79) Chrysene-d12	21.586	240	519522	20.000	ng/u1	0.00
88) Perylene-d12	24.151	264	657956	20.000	ng/u1	0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.399	96	30296	7.935	ng/uL	0.00
4) Pyridine-d5	3.840	84	203568	20.526	ng/u1	0.00
7) Phenol-d5	7.246	99	253937	19.496	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.405	67	153467	19.729	ng/u1	0.00
11) 2-Chlorophenol-d4	7.610	132	195514	20.703	ng/u1	0.00
15) 4-Methylphenol-d8	8.805	113	202607	18.953	ng/u1	0.00
21) Nitrobenzene-d5	9.263	128	97920	21.674	ng/u1	0.00
24) 2-Nitrophenol-d4	9.993	143	114256	22.145	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.534	165	200757	21.331	ng/u1	0.00
31) 4-Chloroaniline-d4	11.057	131	253608	18.773	ng/u1	0.00
46) Dimethylphthalate-d6	14.134	166	563635	20.955	ng/u1	0.00
49) Acenaphthylene-d8	14.428	160	646288	21.974	ng/u1	0.00
54) 4-Nitrophenol-d4	14.957	143	67563	14.027	ng/u1	0.00
60) Fluorene-d10	15.728	176	454752	20.049	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.869	200	77923	19.498	ng/u1	0.00
73) Anthracene-d10	17.592	188	623019	21.706	ng/u1	0.00
81) Pyrene-d10	19.822	212	626780	22.541	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.980	264	688719	20.862	ng/u1	0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.434	88	33001	8.183	ng/uL	92
5) Pyridine	3.864	79	208229	20.253	ng/u1	97
6) Benzaldehyde	7.216	77	84490	16.971	ng/u1	98
8) Phenol	7.275	94	262179	19.511	ng/u1	94
10) Bis(2-Chloroethyl)ether	7.499	93	207113	19.494	ng/u1	99
12) 2-Chlorophenol	7.640	128	206109	20.657	ng/u1	96
13) 2-Methylphenol	8.534	108	199849	19.330	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.616	45	276490	19.638	ng/u1	95
16) Acetophenone	8.916	105	332896	19.513	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.899	70	179836	19.531	ng/u1	99
18) 4-Methylphenol	8.869	108	217938	19.158	ng/u1	100
19) Hexachloroethane	9.163	117	91114	21.610	ng/u1	98
22) Nitrobenzene	9.305	77	261549	21.830	ng/u1	99
23) Isophorone	9.834	82	499342	20.820	ng/u1	99
25) 2-Nitrophenol	10.022	139	119917	21.527	ng/u1	95
26) 2,4-Dimethylphenol	10.075	107	259669	21.981	ng/u1	95
27) Bis(2-Chloroethoxy)met...	10.310	93	281475	20.186	ng/u1	99
29) 2,4-Dichlorophenol	10.563	162	205259	21.935	ng/u1	99
30) Naphthalene	10.957	128	670621	21.506	ng/u1	100
32) 4-Chloroaniline	11.081	127	257171	18.398	ng/u1	97
33) Hexachlorobutadiene	11.228	225	147796	23.420	ng/u1	98
34) Caprolactam	11.863	113	60577	17.197	ng/u1	92
35) 4-Chloro-3-methylphenol	12.204	107	227455	20.061	ng/u1	97
36) 2-Methylnaphthalene	12.563	142	442887	20.203	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062323\
 Data File : BP015668.D
 Acq On : 23 Jun 2023 13:13
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020763

Quant Time: Jun 23 22:34:08 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jun 22 22:01:49 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.781	142	447281	20.265	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.928	216	252034	23.729	ng/ul	99
40) Hexachlorocyclopentadiene	12.893	237	66878	15.361	ng/ul	98
41) 2,4,6-Trichlorophenol	13.175	196	160211	22.663	ng/ul	99
42) 2,4,5-Trichlorophenol	13.251	196	170116	22.083	ng/ul	98
43) 1,1'-Biphenyl	13.563	154	593616	22.504	ng/ul	99
44) 2-Chloronaphthalene	13.610	162	471516	22.783	ng/ul	99
45) 2-Nitroaniline	13.828	65	144483	21.813	ng/ul	98
47) Dimethylphthalate	14.181	163	588937	21.207	ng/ul	99
48) 2,6-Dinitrotoluene	14.316	165	119413	20.987	ng/ul	99
50) Acenaphthylene	14.457	152	716009	21.984	ng/ul	99
51) 3-Nitroaniline	14.651	138	96407	20.525	ng/ul	91
52) Acenaphthene	14.792	153	488528	21.699	ng/ul	99
53) 2,4-Dinitrophenol	14.881	184	47207	13.793	ng/ul	99
55) 4-Nitrophenol	14.975	109	70036	15.074	ng/ul	91
56) Dibenzofuran	15.134	168	665747	20.863	ng/ul	100
57) 2,4-Dinitrotoluene	15.116	165	157243	18.790	ng/ul	87
58) 2,3,4,6-Tetrachlorophenol	15.363	232	135982	19.550	ng/ul	99
59) Diethylphthalate	15.539	149	583796	20.511	ng/ul	98
61) Fluorene	15.781	166	531437	20.430	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.769	204	276287	21.074	ng/ul	98
63) 4-Nitroaniline	15.822	138	50267	11.283	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.887	198	79810	19.581	ng/ul	92
67) N-Nitrosodiphenylamine	15.986	169	445935	24.917	ng/ul	98
68) 4-Bromophenyl-phenylether	16.669	248	168927	25.030	ng/ul	96
69) Hexachlorobenzene	16.792	284	189816	23.843	ng/ul	95
70) Atrazine	16.945	200	165501	23.158	ng/ul	98
71) Pentachlorophenol	17.151	266	80456	17.880	ng/ul	97
72) Phenanthrene	17.539	178	739574	21.846	ng/ul	98
74) Anthracene	17.628	178	743702	21.651	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.534	216	256091	29.414	ng/uL	100
76) Pentachlorobenzene	15.051	250	247543	27.156	ng/uL	98
77) Carbazole	17.904	167	627164	20.278	ng/ul	100
78) Di-n-butylphthalate	18.422	149	881954	22.475	ng/ul	99
80) Fluoranthene	19.498	202	771900	22.790	ng/ul	98
82) Pyrene	19.851	202	790794	22.568	ng/ul	97
83) Butylbenzylphthalate	20.704	149	341933	22.078	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.492	252	265289	21.614	ng/ul	99
85) Benzo(a)anthracene	21.569	228	782662	21.550	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.457	149	481755	21.045	ng/ul	99
87) Chrysene	21.627	228	741080	21.588	ng/ul	99
89) Di-n-octyl phthalate	22.427	149	840825	19.444	ng/ul	100
90) Benzo(b)fluoranthene	23.357	252	852138	20.064	ng/ul	99
91) Benzo(k)fluoranthene	23.410	252	864162	21.049	ng/ul	99
93) Benzo(a)pyrene	24.033	252	772486	20.862	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.868	276	1075772	22.449	ng/ul	98
95) Dibenzo(a,h)anthracene	26.874	278	893581	22.888	ng/ul	99
96) Benzo(g,h,i)perylene	27.709	276	882229	22.339	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062323\
Data File : BP015668.D
Acq On : 23 Jun 2023 13:13
Operator : MA/JU
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTD020763

Quant Time: Jun 23 22:34:08 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jun 22 22:01:49 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.183 ng/uL
RT: 3.434 min Scan# 42
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:	88	Resp:	33001
Ion	Ratio	Lower	Upper
88	100		
43	32.3	23.9	35.9
58	72.0	51.4	77.2



#5
Pyridine
Concen: 20.253 ng/uL
RT: 3.864 min Scan# 115
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13

Tgt Ion:	79	Resp:	208229
Ion	Ratio	Lower	Upper
79	100		
52	75.7	59.5	89.3
51	40.1	29.8	44.6









#6
Benzaldehyde
Concen: 16.971 ng/ul
RT: 7.216 min Scan# 685
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion: 77 Resp: 84490
Ion Ratio Lower Upper
77 100
105 95.8 77.3 115.9
106 88.5 73.3 109.9



#8
Phenol
Concen: 19.511 ng/ul
RT: 7.275 min Scan# 695
Delta R.T. 0.006 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13

Tgt Ion: 94 Resp: 262179
Ion Ratio Lower Upper
94 100
65 31.2 23.5 35.3
66 47.3 34.2 51.4









#10
Bis(2-Chloroethyl)ether
Concen: 19.494 ng/ul
RT: 7.499 min Scan# 733
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion: 93 Resp: 207113
Ion Ratio Lower Upper
93 100
63 78.4 62.6 93.8
95 33.3 25.5 38.3



#12
2-Chlorophenol
Concen: 20.657 ng/ul
RT: 7.640 min Scan# 757
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13

Tgt Ion:128 Resp: 206109
Ion Ratio Lower Upper
128 100
64 53.2 40.2 60.4
130 33.3 24.5 36.7









#13
2-Methylphenol
Concen: 19.330 ng/ul
RT: 8.534 min Scan# 909
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:108 Resp: 199849
Ion Ratio Lower Upper
108 100
107 89.7 71.2 106.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 19.638 ng/ul
RT: 8.616 min Scan# 923
Delta R.T. 0.006 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13

Tgt Ion: 45 Resp: 276490
Ion Ratio Lower Upper
45 100
77 16.4 11.8 17.8
79 13.0 8.7 13.1









#16
Acetophenone
Concen: 19.513 ng/ul
RT: 8.916 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:105 Resp: 332896
Ion Ratio Lower Upper
105 100
77 83.9 64.9 97.3
51 31.1 23.5 35.3



#17
N-Nitroso-di-n-propylamine
Concen: 19.531 ng/ul
RT: 8.899 min Scan# 971
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13

Tgt Ion: 70 Resp: 179836
Ion Ratio Lower Upper
70 100
42 59.0 46.6 69.8
101 10.3 7.8 11.6
130 19.4 15.9 23.9









#18
4-Methylphenol
Concen: 19.158 ng/ul
RT: 8.869 min Scan# 966
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:108 Resp: 217938
Ion Ratio Lower Upper
108 100
107 111.8 89.3 133.9



#19
Hexachloroethane
Concen: 21.610 ng/ul
RT: 9.163 min Scan# 1016
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13

Tgt Ion:117 Resp: 91114
Ion Ratio Lower Upper
117 100
201 103.7 82.8 124.2
199 68.1 50.6 76.0









#22
Nitrobenzene
Concen: 21.830 ng/ul
RT: 9.305 min Scan# 1040
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion: 77 Resp: 261549
Ion Ratio Lower Upper
77 100
123 41.1 33.5 50.3
65 13.9 11.6 17.4



#23
Isophorone
Concen: 20.820 ng/ul
RT: 9.834 min Scan# 1130
Delta R.T. 0.006 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13

Tgt Ion: 82 Resp: 499342
Ion Ratio Lower Upper
82 100
95 7.8 5.8 8.8
138 16.5 13.1 19.7









#25
2-Nitrophenol
Concen: 21.527 ng/ul
RT: 10.022 min Scan# 1162
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:139 Resp: 119917
Ion Ratio Lower Upper
139 100
65 61.0 46.9 70.3
109 42.2 30.2 45.2



#26
2,4-Dimethylphenol
Concen: 21.981 ng/ul
RT: 10.075 min Scan# 1171
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13

Tgt Ion:107 Resp: 259669
Ion Ratio Lower Upper
107 100
121 45.8 39.0 58.6
122 77.0 64.8 97.2









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.186 ng/ul
 RT: 10.310 min Scan# 1211
 Delta R.T. 0.000 min
 Lab File: BP015668.D
 Acq: 23 Jun 2023 13:13




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020763


Tgt Ion: 93 Resp: 281475
 Ion Ratio Lower Upper
 93 100
 95 31.8 25.7 38.5
 123 10.5 8.6 13.0



#29
 2,4-Dichlorophenol
 Concen: 21.935 ng/ul
 RT: 10.563 min Scan# 1254
 Delta R.T. 0.006 min
 Lab File: BP015668.D
 Acq: 23 Jun 2023 13:13

Tgt Ion: 162 Resp: 205259
 Ion Ratio Lower Upper
 162 100
 164 63.8 50.6 76.0
 98 36.1 30.2 45.2









#30
Naphthalene
Concen: 21.506 ng/ul
RT: 10.957 min Scan# 1321
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:128 Resp: 670621
Ion Ratio Lower Upper
128 100
129 10.8 8.6 13.0
127 13.4 10.5 15.7



#32
4-Chloroaniline
Concen: 18.398 ng/ul
RT: 11.081 min Scan# 1342
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13

Tgt Ion:127 Resp: 257171
Ion Ratio Lower Upper
127 100
129 31.2 26.1 39.1





#33

Hexachlorobutadiene

Concen: 23.420 ng/ul

RT: 11.228 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP015668.D

Acq: 23 Jun 2023 13:13

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:225 Resp: 147796

Ion Ratio Lower Upper

225 100

223 62.3 52.0 78.0

227 63.1 50.6 75.8



#34

Caprolactam

Concen: 17.197 ng/ul

RT: 11.863 min Scan# 1475

Delta R.T. -0.006 min

Lab File: BP015668.D

Acq: 23 Jun 2023 13:13




Tgt Ion:113 Resp: 60577


Ion Ratio Lower Upper

113 100

55 167.5 127.9 191.9

56 137.7 101.1 151.7









#35
4-Chloro-3-methylphenol
Concen: 20.061 ng/ul
RT: 12.204 min Scan# 1533
Delta R.T. 0.006 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:107 Resp: 227455
Ion Ratio Lower Upper
107 100
144 23.1 19.4 29.0
142 73.2 60.4 90.6



#36
2-Methylnaphthalene
Concen: 20.203 ng/ul
RT: 12.563 min Scan# 1594
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13

Tgt Ion:142 Resp: 442887
Ion Ratio Lower Upper
142 100
141 88.0 72.0 108.0









#37
1-Methylnaphthalene
Concen: 20.265 ng/ul
RT: 12.781 min Scan# 1631
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:142 Resp: 447281
Ion Ratio Lower Upper
142 100
141 90.8 71.9 107.9
116 4.5 3.4 5.2



#39
1,2,4,5-Tetrachlorobenzene
Concen: 23.729 ng/ul
RT: 12.928 min Scan# 1656
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13

Tgt Ion:216 Resp: 252034
Ion Ratio Lower Upper
216 100
214 78.6 62.2 93.2
179 21.9 17.4 26.2
108 17.7 14.2 21.4









#40
Hexachlorocyclopentadiene
Concen: 15.361 ng/ul
RT: 12.893 min Scan# 1650
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:	237	Resp:	66878
Ion	Ratio	Lower	Upper
237	100		
235	61.1	48.2	72.2
272	11.5	10.3	15.5



#41
2,4,6-Trichlorophenol
Concen: 22.663 ng/ul
RT: 13.175 min Scan# 1698
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13

Tgt Ion:	196	Resp:	160211
Ion	Ratio	Lower	Upper
196	100		
198	98.7	77.8	116.6
200	31.8	25.8	38.6





#42

2,4,5-Trichlorophenol

Concen: 22.083 ng/ul

RT: 13.251 min Scan# 1711

Delta R.T. 0.000 min

Lab File: BP015668.D

Acq: 23 Jun 2023 13:13

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:196 Resp: 170116

Ion Ratio Lower Upper

196 100

198 96.3 75.5 113.3

200 32.1 24.7 37.1



#43

1,1'-Biphenyl

Concen: 22.504 ng/ul

RT: 13.563 min Scan# 1764

Delta R.T. 0.000 min

Lab File: BP015668.D

Acq: 23 Jun 2023 13:13




Tgt Ion:154 Resp: 593616


Ion Ratio Lower Upper

154 100

153 40.9 32.2 48.2

76 13.1 10.4 15.6









#44
2-Chloronaphthalene
Concen: 22.783 ng/ul
RT: 13.610 min Scan# 1772
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:162 Resp: 471516
Ion Ratio Lower Upper
162 100
164 33.3 27.0 40.6
127 41.5 32.2 48.4



#45
2-Nitroaniline
Concen: 21.813 ng/ul
RT: 13.828 min Scan# 1809
Delta R.T. 0.006 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13

Tgt Ion: 65 Resp: 144483
Ion Ratio Lower Upper
65 100
92 67.7 55.4 83.2
138 96.8 78.8 118.2





#47

Dimethylphthalate

Concen: 21.207 ng/ul

RT: 14.181 min Scan# 1869

Delta R.T. 0.000 min

Lab File: BP015668.D

Acq: 23 Jun 2023 13:13

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:163 Resp: 588937

Ion Ratio Lower Upper

163 100

194 4.0 3.2 4.8

164 9.8 8.2 12.2



#48

2,6-Dinitrotoluene

Concen: 20.987 ng/ul

RT: 14.316 min Scan# 1892

Delta R.T. 0.006 min

Lab File: BP015668.D

Acq: 23 Jun 2023 13:13




Tgt Ion:165 Resp: 119413


Ion Ratio Lower Upper

165 100

89 65.1 51.5 77.3

121 21.7 17.3 25.9





#50

Acenaphthylene

Concen: 21.984 ng/ul

RT: 14.457 min Scan# 1916

Delta R.T. 0.006 min

Lab File: BP015668.D

Acq: 23 Jun 2023 13:13

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:152 Resp: 716009

Ion Ratio Lower Upper

152 100

151 19.5 16.0 24.0

153 13.0 10.6 15.8



#51

3-Nitroaniline

Concen: 20.525 ng/ul

RT: 14.651 min Scan# 1949

Delta R.T. 0.000 min

Lab File: BP015668.D

Acq: 23 Jun 2023 13:13




Tgt Ion:138 Resp: 96407


Ion Ratio Lower Upper

138 100

108 11.8 8.4 12.6

92 133.3 98.0 147.0









#52
Acenaphthene
Concen: 21.699 ng/ul
RT: 14.792 min Scan# 1973
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:153 Resp: 488528
Ion Ratio Lower Upper
153 100
152 46.9 38.3 57.5
154 88.5 69.8 104.6



#53
2,4-Dinitrophenol
Concen: 13.793 ng/ul
RT: 14.881 min Scan# 1988
Delta R.T. 0.006 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13

Tgt Ion:184 Resp: 47207
Ion Ratio Lower Upper
184 100
63 71.9 59.0 88.4
154 71.6 57.6 86.4









#55
4-Nitrophenol
Concen: 15.074 ng/ul
RT: 14.975 min Scan# 2004
Delta R.T. 0.012 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:109 Resp: 70036
Ion Ratio Lower Upper
109 100
139 100.7 89.1 133.7
65 118.5 103.1 154.7



#56
Dibenzofuran
Concen: 20.863 ng/ul
RT: 15.134 min Scan# 2031
Delta R.T. 0.006 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13

Tgt Ion:168 Resp: 665747
Ion Ratio Lower Upper
168 100
139 39.2 31.2 46.8









#57
2,4-Dinitrotoluene
Concen: 18.790 ng/ul
RT: 15.116 min Scan# 2028
Delta R.T. 0.006 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:165 Resp: 157243
Ion Ratio Lower Upper
165 100
63 58.3 39.1 58.7
182 2.6 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.550 ng/ul
RT: 15.363 min Scan# 2070
Delta R.T. 0.006 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13

Tgt Ion:232 Resp: 135982
Ion Ratio Lower Upper
232 100
131 47.3 37.8 56.8
130 2.8 1.9 2.9
166 31.8 26.2 39.2









#59
Diethylphthalate
Concen: 20.511 ng/ul
RT: 15.539 min Scan# 2100
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:149 Resp: 583796
Ion Ratio Lower Upper
149 100
177 21.1 17.8 26.8
150 12.3 10.1 15.1



#61
Fluorene
Concen: 20.430 ng/ul
RT: 15.781 min Scan# 2141
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13

Tgt Ion:166 Resp: 531437
Ion Ratio Lower Upper
166 100
165 99.1 79.1 118.7
167 13.2 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 21.074 ng/ul
RT: 15.769 min Scan# 2139
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:204 Resp: 276287
Ion Ratio Lower Upper
204 100
206 32.3 26.6 40.0
141 64.3 49.8 74.8



#63
4-Nitroaniline
Concen: 11.283 ng/ul
RT: 15.822 min Scan# 2148
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13

Tgt Ion:138 Resp: 50267
Ion Ratio Lower Upper
138 100
92 60.9 44.6 67.0
108 104.9 78.4 117.6









#66
4,6-Dinitro-2-methylphenol
Concen: 19.581 ng/ul
RT: 15.887 min Scan# 2159
Delta R.T. 0.012 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:198 Resp: 79810
Ion Ratio Lower Upper
198 100
182 4.3 3.0 4.4
77 30.2 20.5 30.7



#67
N-Nitrosodiphenylamine
Concen: 24.917 ng/ul
RT: 15.986 min Scan# 2176
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13

Tgt Ion:169 Resp: 445935
Ion Ratio Lower Upper
169 100
168 68.2 53.0 79.4
167 36.8 28.9 43.3









#68
4-Bromophenyl-phenylether
Concen: 25.030 ng/ul
RT: 16.669 min Scan# 2292
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:248 Resp: 168927
Ion Ratio Lower Upper
248 100
250 94.8 78.3 117.5
141 79.9 61.1 91.7



#69
Hexachlorobenzene
Concen: 23.843 ng/ul
RT: 16.792 min Scan# 2313
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13

Tgt Ion:284 Resp: 189816
Ion Ratio Lower Upper
284 100
142 36.3 25.4 38.2
249 32.3 24.8 37.2









#70
Atrazine
Concen: 23.158 ng/ul
RT: 16.945 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:200 Resp: 165501
Ion Ratio Lower Upper
200 100
173 25.8 19.8 29.6
215 47.5 37.4 56.0



#71
Pentachlorophenol
Concen: 17.880 ng/ul
RT: 17.151 min Scan# 2374
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13

Tgt Ion:266 Resp: 80456
Ion Ratio Lower Upper
266 100
264 60.5 49.6 74.4
268 60.1 51.0 76.6





#72

Phenanthrene

Concen: 21.846 ng/ul

RT: 17.539 min Scan# 2440

Delta R.T. 0.006 min

Lab File: BP015668.D

Acq: 23 Jun 2023 13:13

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:178 Resp: 739574

Ion Ratio Lower Upper

178 100

179 14.8 12.3 18.5

176 18.5 15.4 23.0



#74

Anthracene

Concen: 21.651 ng/ul

RT: 17.628 min Scan# 2455

Delta R.T. 0.000 min

Lab File: BP015668.D

Acq: 23 Jun 2023 13:13




Tgt Ion:178 Resp: 743702


Ion Ratio Lower Upper

178 100

179 15.3 12.4 18.6

176 18.4 15.0 22.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 29.414 ng/uL
RT: 13.534 min Scan# 1759
Delta R.T. 0.006 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:216 Resp: 256091
Ion Ratio Lower Upper
216 100
214 78.0 62.1 93.1
218 48.8 39.0 58.4



#76
Pentachlorobenzene
Concen: 27.156 ng/uL
RT: 15.051 min Scan# 2017
Delta R.T. 0.006 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13

Tgt Ion:250 Resp: 247543
Ion Ratio Lower Upper
250 100
248 62.4 49.0 73.4
252 64.4 49.9 74.9





#77

Carbazole

Concen: 20.278 ng/ul

RT: 17.904 min Scan# 2502

Delta R.T. 0.006 min

Lab File: BP015668.D

Acq: 23 Jun 2023 13:13

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:167 Resp: 627164

Ion Ratio Lower Upper

167 100

166 21.4 17.2 25.8

139 13.4 10.6 16.0



#78

Di-n-butylphthalate

Concen: 22.475 ng/ul

RT: 18.422 min Scan# 2590

Delta R.T. 0.000 min

Lab File: BP015668.D

Acq: 23 Jun 2023 13:13




Tgt Ion:149 Resp: 881954


Ion Ratio Lower Upper

149 100

150 9.3 7.1 10.7

104 5.8 4.5 6.7





#80

Fluoranthene

Concen: 22.790 ng/ul

RT: 19.498 min Scan# 2773

Delta R.T. 0.006 min

Lab File: BP015668.D

Acq: 23 Jun 2023 13:13

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:202 Resp: 771900

Ion Ratio Lower Upper

202 100

101 10.1 8.6 13.0

100 7.6 6.6 9.8



#82

Pyrene

Concen: 22.568 ng/ul

RT: 19.851 min Scan# 2833

Delta R.T. 0.006 min

Lab File: BP015668.D

Acq: 23 Jun 2023 13:13




Tgt Ion:202 Resp: 790794


Ion Ratio Lower Upper

202 100

101 11.1 9.9 14.9

100 9.7 8.4 12.6





#83

Butylbenzylphthalate

Concen: 22.078 ng/ul

RT: 20.704 min Scan# 2978

Delta R.T. 0.006 min

Lab File: BP015668.D

Acq: 23 Jun 2023 13:13

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:149 Resp: 341933

Ion Ratio Lower Upper

149 100

91 74.1 60.4 90.6

206 21.9 16.4 24.6



#84

3,3'-Dichlorobenzidine

Concen: 21.614 ng/ul

RT: 21.492 min Scan# 3112

Delta R.T. 0.000 min

Lab File: BP015668.D

Acq: 23 Jun 2023 13:13




Tgt Ion:252 Resp: 265289


Ion Ratio Lower Upper

252 100

254 63.9 51.6 77.4

126 11.0 8.5 12.7









#85
Benzo(a)anthracene
Concen: 21.550 ng/ul
RT: 21.569 min Scan# 3125
Delta R.T. 0.006 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:228 Resp: 782662
Ion Ratio Lower Upper
228 100
229 19.5 15.8 23.6
226 25.7 21.8 32.6



#86
Bis(2-ethylhexyl)phthalate
Concen: 21.045 ng/ul
RT: 21.457 min Scan# 3106
Delta R.T. 0.000 min
Lab File: BP015668.D
Acq: 23 Jun 2023 13:13

Tgt Ion:149 Resp: 481755
Ion Ratio Lower Upper
149 100
167 27.7 21.8 32.8
279 4.7 3.7 5.5





#87

Chrysene

Concen: 21.588 ng/ul

RT: 21.627 min Scan# 3135

Delta R.T. 0.006 min

Lab File: BP015668.D

Acq: 23 Jun 2023 13:13

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:228 Resp: 741080

Ion Ratio Lower Upper

228 100

226 29.6 23.9 35.9

229 19.7 15.1 22.7



#89

Di-n-octyl phthalate

Concen: 19.444 ng/ul




RT: 22.427 min Scan# 3271


Delta R.T. 0.006 min

Lab File: BP015668.D

Acq: 23 Jun 2023 13:13

Tgt Ion:149 Resp: 840825





#90

Benzo(b)fluoranthene

Concen: 20.064 ng/ul

RT: 23.357 min Scan# 3429

Delta R.T. 0.012 min

Lab File: BP015668.D

Acq: 23 Jun 2023 13:13

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:252 Resp: 852138

Ion Ratio Lower Upper

252 100

253 21.9 17.4 26.2

125 8.6 7.3 10.9



#91

Benzo(k)fluoranthene

Concen: 21.049 ng/ul

RT: 23.410 min Scan# 3438

Delta R.T. 0.012 min

Lab File: BP015668.D

Acq: 23 Jun 2023 13:13




Tgt Ion:252 Resp: 864162


Ion Ratio Lower Upper

252 100

253 21.9 17.7 26.5

125 8.3 7.3 10.9





#93

Benzo(a)pyrene

Concen: 20.862 ng/ul

RT: 24.033 min Scan# 3544

Delta R.T. 0.012 min

Lab File: BP015668.D

Acq: 23 Jun 2023 13:13

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:252 Resp: 772486

Ion Ratio Lower Upper

252 100

253 21.7 17.8 26.6

125 9.5 8.1 12.1



#94

Indeno(1,2,3-cd)pyrene

Concen: 22.449 ng/ul

RT: 26.868 min Scan# 4026

Delta R.T. 0.030 min

Lab File: BP015668.D

Acq: 23 Jun 2023 13:13




Tgt Ion:276 Resp: 1075772


Ion Ratio Lower Upper

276 100

138 21.4 17.3 25.9

277 26.4 20.2 30.2





#95

Dibenzo(a,h)anthracene

Concen: 22.888 ng/ul

RT: 26.874 min Scan# 4027

Delta R.T. 0.024 min

Lab File: BP015668.D

Acq: 23 Jun 2023 13:13

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:278 Resp: 893581

Ion Ratio Lower Upper

278 100

139 13.2 11.4 17.0

279 23.8 18.7 28.1



#96

Benzo(g,h,i)perylene

Concen: 22.339 ng/ul

RT: 27.709 min Scan# 4169

Delta R.T. 0.030 min

Lab File: BP015668.D

Acq: 23 Jun 2023 13:13

Tgt Ion:276 Resp: 882229

Ion Ratio Lower Upper

276 100

138 17.4 15.6 23.4

277 23.5 19.0 28.6

