

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062323\
 Data File : BP015697.D
 Acq On : 24 Jun 2023 06:09
 Operator : MA/JU
 Sample : 03085-07
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DCFP3

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/26/2023
 Supervised By :mohammad ahmed 06/27/2023

Quant Time: Jun 24 06:45:42 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jun 22 22:01:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.075	152	138458	20.000	ng/u1	0.00
20) Naphthalene-d8	10.904	136	523126	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.728	164	304043	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.492	188	673563	20.000	ng/u1	0.00
79) Chrysene-d12	21.580	240	669056	20.000	ng/u1	0.00
88) Perylene-d12	24.139	264	792846	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.399	96	10122	2.707	ng/uL	0.00
4) Pyridine-d5	0.000	84	0d	0.000	ng/u1	
7) Phenol-d5	7.246	99	147983	11.601	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.399	67	104304	13.692	ng/u1	0.00
11) 2-Chlorophenol-d4	7.610	132	131238	14.191	ng/u1	0.00
15) 4-Methylphenol-d8	8.810	113	68352	6.529	ng/u1	0.00
21) Nitrobenzene-d5	9.263	128	62286	15.166	ng/u1	0.00
24) 2-Nitrophenol-d4	9.993	143	70674	15.068	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.551	165	105767	12.362	ng/u1	0.02
31) 4-Chloroaniline-d4	11.075	131	85472	6.960	ng/u1	0.02
46) Dimethylphthalate-d6	14.134	166	381372	15.907	ng/u1	0.00
49) Acenaphthylene-d8	14.422	160	436329	16.643	ng/u1	0.00
54) 4-Nitrophenol-d4	14.992	143	37452m	8.723	ng/u1	0.04
60) Fluorene-d10	15.722	176	322842	15.968	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.875	200	38964m	9.131	ng/u1	0.01
73) Anthracene-d10	17.592	188	476940	15.563	ng/u1	0.00
81) Pyrene-d10	19.816	212	632560	17.665	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.968	264	635077	15.964	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062323\
Data File : BP015697.D
Acq On : 24 Jun 2023 06:09
Operator : MA/JU
Sample : 03085-07
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DCFP3

Quant Time: Jun 24 06:45:42 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP060723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jun 22 22:01:49 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/26/2023
Supervised By :mohammad ahmed 06/27/2023

