

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062623\
 Data File : BP015716.D
 Acq On : 26 Jun 2023 16:11
 Operator : MA/JU
 Sample : SST04092
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST040635

Quant Time: Jun 27 00:01:54 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 26 23:54:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.063	152	143230	20.000	ng/u1	0.00
20) Naphthalene-d8	10.893	136	595492	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.710	164	369318	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.475	188	774275	20.000	ng/u1	0.00
79) Chrysene-d12	21.563	240	553329	20.000	ng/u1	0.00
88) Perylene-d12	24.116	264	536826	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.387	96	62990	15.823	ng/uL	0.00
4) Pyridine-d5	3.828	84	420871	41.939	ng/u1	0.00
7) Phenol-d5	7.228	99	500695	40.856	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.387	67	311314	41.060	ng/u1	0.00
11) 2-Chlorophenol-d4	7.593	132	378759	40.943	ng/u1	0.00
15) 4-Methylphenol-d8	8.793	113	399618	41.277	ng/u1	0.00
21) Nitrobenzene-d5	9.246	128	193208	42.454	ng/u1	0.00
24) 2-Nitrophenol-d4	9.975	143	226437	42.631	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.516	165	400765	42.341	ng/u1	0.00
31) 4-Chloroaniline-d4	11.040	131	511640	42.080	ng/u1	0.00
46) Dimethylphthalate-d6	14.122	166	1157386	40.545	ng/u1	0.00
49) Acenaphthylene-d8	14.410	160	1305855	40.695	ng/u1	0.00
54) 4-Nitrophenol-d4	14.940	143	171264	45.465	ng/u1	0.00
60) Fluorene-d10	15.710	176	948107	40.338	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.851	200	199139	42.035	ng/u1	0.00
73) Anthracene-d10	17.575	188	1431186	40.466	ng/u1	0.00
81) Pyrene-d10	19.804	212	1465589	40.563	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.945	264	1113579	41.122	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.423	88	68730	16.043	ng/uL	98
5) Pyridine	3.846	79	431227	41.076	ng/u1	98
6) Benzaldehyde	7.205	77	228377	47.540	ng/u1	96
8) Phenol	7.258	94	513685	40.393	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.481	93	412095	40.728	ng/u1	98
12) 2-Chlorophenol	7.628	128	399571	40.656	ng/u1	99
13) 2-Methylphenol	8.522	108	394164	41.051	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.599	45	565614	40.927	ng/u1	99
16) Acetophenone	8.905	105	655249	41.173	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.887	70	357500	40.446	ng/u1	98
18) 4-Methylphenol	8.858	108	426626	41.354	ng/u1	100
19) Hexachloroethane	9.146	117	183787	40.488	ng/u1	97
22) Nitrobenzene	9.293	77	535331	41.888	ng/u1	97
23) Isophorone	9.816	82	1005301	41.106	ng/u1	100
25) 2-Nitrophenol	10.005	139	238641	42.334	ng/u1	99
26) 2,4-Dimethylphenol	10.063	107	510296	41.180	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.293	93	574191	41.589	ng/u1	99
29) 2,4-Dichlorophenol	10.546	162	398844	42.385	ng/u1	100
30) Naphthalene	10.940	128	1299098	40.130	ng/u1	99
32) 4-Chloroaniline	11.063	127	533811	42.257	ng/u1	99
33) Hexachlorobutadiene	11.210	225	284334	40.011	ng/u1	97
34) Caprolactam	11.863	113	124817	43.996	ng/u1	96
35) 4-Chloro-3-methylphenol	12.187	107	462139	41.956	ng/u1	98
36) 2-Methylnaphthalene	12.546	142	868247	40.742	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062623\
 Data File : BP015716.D
 Acq On : 26 Jun 2023 16:11
 Operator : MA/JU
 Sample : SST04092
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST040635

Quant Time: Jun 27 00:01:54 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 26 23:54:28 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.763	142	872141	40.118	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.910	216	496745	40.816	ng/ul	99
40) Hexachlorocyclopentadiene	12.875	237	150452	41.323	ng/ul	97
41) 2,4,6-Trichlorophenol	13.157	196	324275	42.161	ng/ul	95
42) 2,4,5-Trichlorophenol	13.240	196	350308	43.989	ng/ul	97
43) 1,1'-Biphenyl	13.546	154	1181509	40.390	ng/ul	99
44) 2-Chloronaphthalene	13.593	162	933945	40.528	ng/ul	100
45) 2-Nitroaniline	13.810	65	316020	44.052	ng/ul	95
47) Dimethylphthalate	14.169	163	1209543	40.943	ng/ul	99
48) 2,6-Dinitrotoluene	14.304	165	256463	42.798	ng/ul	100
50) Acenaphthylene	14.440	152	1441012	40.756	ng/ul	99
51) 3-Nitroaniline	14.640	138	209668	44.618	ng/ul	93
52) Acenaphthene	14.781	153	986451	40.233	ng/ul	98
53) 2,4-Dinitrophenol	14.863	184	142394	45.088	ng/ul	94
55) 4-Nitrophenol	14.957	109	177831	42.528	ng/ul	96
56) Dibenzofuran	15.116	168	1373792	40.363	ng/ul	99
57) 2,4-Dinitrotoluene	15.098	165	353711	43.525	ng/ul	89
58) 2,3,4,6-Tetrachlorophenol	15.345	232	299012	42.671	ng/ul	99
59) Diethylphthalate	15.528	149	1229805	41.137	ng/ul	98
61) Fluorene	15.763	166	1108351	40.147	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.751	204	569905	40.344	ng/ul	100
63) 4-Nitroaniline	15.810	138	155721	49.011	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.869	198	204088	42.361	ng/ul	98
67) N-Nitrosodiphenylamine	15.969	169	940901	40.593	ng/ul	98
68) 4-Bromophenyl-phenylether	16.651	248	363173	41.075	ng/ul	95
69) Hexachlorobenzene	16.775	284	418231	40.089	ng/ul	99
70) Atrazine	16.928	200	369066	41.596	ng/ul	99
71) Pentachlorophenol	17.128	266	208122	41.855	ng/ul	97
72) Phenanthrene	17.516	178	1678568	39.906	ng/ul	99
74) Anthracene	17.610	178	1703228	40.298	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.516	216	504759	40.063	ng/uL	98
76) Pentachlorobenzene	15.034	250	502653	39.530	ng/uL	97
77) Carbazole	17.887	167	1462220	41.046	ng/ul	100
78) Di-n-butylphthalate	18.404	149	1955945	41.415	ng/ul	99
80) Fluoranthene	19.475	202	1834997	40.865	ng/ul	100
82) Pyrene	19.833	202	1842630	40.227	ng/ul	100
83) Butylbenzylphthalate	20.680	149	773765	41.406	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.475	252	492562	39.379	ng/ul	99
85) Benzo(a)anthracene	21.545	228	1567938	40.282	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.433	149	1049785	40.876	ng/ul	98
87) Chrysene	21.604	228	1469588	39.794	ng/ul	99
89) Di-n-octyl phthalate	22.398	149	1642812	41.874	ng/ul	100
90) Benzo(b)fluoranthene	23.327	252	1420328	41.177	ng/ul	98
91) Benzo(k)fluoranthene	23.380	252	1428795	40.997	ng/ul	99
93) Benzo(a)pyrene	24.004	252	1228311	40.754	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.815	276	1656303	40.481	ng/ul	98
95) Dibenzo(a,h)anthracene	26.827	278	1361461	40.511	ng/ul	100
96) Benzo(g,h,i)perylene	27.651	276	1369697	40.871	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062623\
 Data File : BP015716.D
 Acq On : 26 Jun 2023 16:11
 Operator : MA/JU
 Sample : SSTD04092
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD040635

Quant Time: Jun 27 00:01:54 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 26 23:54:28 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 16.043 ng/uL
RT: 3.423 min Scan# 40
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11




Instrument :
BNA_P
ClientSampleId :
SSTD040635


Tgt Ion: 88 Resp: 68730
Ion Ratio Lower Upper
88 100
43 34.3 27.4 41.2
58 77.7 59.8 89.8



#5
Pyridine
Concen: 41.076 ng/uL
RT: 3.846 min Scan# 112
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11

Tgt Ion: 79 Resp: 431227
Ion Ratio Lower Upper
79 100
52 79.5 65.5 98.3
51 39.4 31.8 47.8









#6
Benzaldehyde
Concen: 47.540 ng/ul
RT: 7.205 min Scan# 683
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11




Instrument :
BNA_P
ClientSampleId :
SSTD040635


Tgt Ion: 77 Resp: 228377
Ion Ratio Lower Upper
77 100
105 94.5 73.4 110.0
106 88.2 66.6 100.0



#8
Phenol
Concen: 40.393 ng/ul
RT: 7.258 min Scan# 692
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11

Tgt Ion: 94 Resp: 513685
Ion Ratio Lower Upper
94 100
65 32.0 26.4 39.6
66 45.7 37.3 55.9









#10
 Bis(2-Chloroethyl)ether
 Concen: 40.728 ng/ul
 RT: 7.481 min Scan# 730
 Delta R.T. 0.000 min
 Lab File: BP015716.D
 Acq: 26 Jun 2023 16:11

Instrument :
 BNA_P
 ClientSampleId :
 SSTD040635

Tgt Ion: 93 Resp: 412095




Ion	Ratio	Lower	Upper
93	100		
63	80.3	62.8	94.2
95	33.4	25.6	38.4




#12
 2-Chlorophenol
 Concen: 40.656 ng/ul
 RT: 7.628 min Scan# 755
 Delta R.T. 0.000 min
 Lab File: BP015716.D
 Acq: 26 Jun 2023 16:11

Tgt Ion: 128 Resp: 399571

Ion	Ratio	Lower	Upper
128	100		
64	52.0	41.2	61.8
130	32.7	26.0	39.0





#13

2-Methylphenol

Concen: 41.051 ng/ul

RT: 8.522 min Scan# 907

Delta R.T. 0.000 min

Lab File: BP015716.D

Acq: 26 Jun 2023 16:11

Instrument :

BNA_P

ClientSampleId :





SSTD040635

Tgt Ion:108 Resp: 394164

Ion Ratio Lower Upper

108 100

107 91.7 73.8 110.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 40.927 ng/ul

RT: 8.599 min Scan# 920

Delta R.T. 0.006 min

Lab File: BP015716.D

Acq: 26 Jun 2023 16:11




Tgt Ion: 45 Resp: 565614


Ion Ratio Lower Upper

45 100

77 15.2 12.4 18.6

79 11.7 9.3 13.9









#16
Acetophenone
Concen: 41.173 ng/ul
RT: 8.905 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11




Instrument :
BNA_P
ClientSampleId :
SSTD040635


Tgt Ion:105 Resp: 655249
Ion Ratio Lower Upper
105 100
77 83.4 65.0 97.4
51 31.3 25.1 37.7



#17
N-Nitroso-di-n-propylamine
Concen: 40.446 ng/ul
RT: 8.887 min Scan# 969
Delta R.T. 0.006 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11

Tgt Ion: 70 Resp: 357500
Ion Ratio Lower Upper
70 100
42 61.7 48.2 72.2
101 9.5 7.6 11.4
130 18.9 13.9 20.9









#18
4-Methylphenol
Concen: 41.354 ng/ul
RT: 8.858 min Scan# 90
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11




Instrument :
BNA_P
ClientSampleId :
SSTD040635


Tgt Ion:108 Resp: 426626
Ion Ratio Lower Upper
108 100
107 113.6 91.3 136.9



#19
Hexachloroethane
Concen: 40.488 ng/ul
RT: 9.146 min Scan# 1013
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11

Tgt Ion:117 Resp: 183787
Ion Ratio Lower Upper
117 100
201 101.4 84.2 126.2
199 62.9 51.8 77.8









#22
Nitrobenzene
Concen: 41.888 ng/ul
RT: 9.293 min Scan# 1038
Delta R.T. 0.006 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11




Instrument :
BNA_P
ClientSampleId :
SSTD040635


Tgt Ion: 77 Resp: 535331
Ion Ratio Lower Upper
77 100
123 41.2 31.0 46.4
65 14.2 11.3 16.9



#23
Isophorone
Concen: 41.106 ng/ul
RT: 9.816 min Scan# 1127
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11

Tgt Ion: 82 Resp: 1005301
Ion Ratio Lower Upper
82 100
95 7.7 6.3 9.5
138 15.4 12.3 18.5









#25
2-Nitrophenol
Concen: 42.334 ng/ul
RT: 10.005 min Scan# 1159
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11




Instrument :
BNA_P
ClientSampleId :
SSTD040635


Tgt Ion:139 Resp: 238641
Ion Ratio Lower Upper
139 100
65 61.9 50.2 75.2
109 40.4 31.3 46.9



#26
2,4-Dimethylphenol
Concen: 41.180 ng/ul
RT: 10.063 min Scan# 1169
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11

Tgt Ion:107 Resp: 510296
Ion Ratio Lower Upper
107 100
121 45.4 38.2 57.2
122 76.4 62.1 93.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 41.589 ng/ul
 RT: 10.293 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: BP015716.D
 Acq: 26 Jun 2023 16:11




Instrument :
 BNA_P
 ClientSampleId :
 SSTD040635


Tgt Ion: 93 Resp: 574191
 Ion Ratio Lower Upper
 93 100
 95 33.4 26.6 39.8
 123 10.4 8.6 13.0



#29
 2,4-Dichlorophenol
 Concen: 42.385 ng/ul
 RT: 10.546 min Scan# 1251
 Delta R.T. -0.006 min
 Lab File: BP015716.D
 Acq: 26 Jun 2023 16:11

Tgt Ion: 162 Resp: 398844
 Ion Ratio Lower Upper
 162 100
 164 66.4 53.0 79.6
 98 39.2 31.3 46.9









#30
Naphthalene
Concen: 40.130 ng/ul
RT: 10.940 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11




Instrument :
BNA_P
ClientSampleId :
SSTD040635


Tgt Ion:128 Resp: 1299098
Ion Ratio Lower Upper
128 100
129 11.1 8.6 12.8
127 13.6 10.3 15.5



#32
4-Chloroaniline
Concen: 42.257 ng/ul
RT: 11.063 min Scan# 1339
Delta R.T. -0.006 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11

Tgt Ion:127 Resp: 533811
Ion Ratio Lower Upper
127 100
129 31.5 24.7 37.1





#33

Hexachlorobutadiene

Concen: 40.011 ng/ul

RT: 11.210 min Scan# 1364

Delta R.T. 0.000 min

Lab File: BP015716.D

Acq: 26 Jun 2023 16:11

Instrument :

BNA_P

ClientSampleId :

SSTD040635





Tgt Ion:225 Resp: 284334

Ion Ratio Lower Upper

225 100

223 64.6 50.4 75.6

227 67.5 51.0 76.4



#34

Caprolactam

Concen: 43.996 ng/ul

RT: 11.863 min Scan# 1475

Delta R.T. 0.006 min

Lab File: BP015716.D

Acq: 26 Jun 2023 16:11




Tgt Ion:113 Resp: 124817


Ion Ratio Lower Upper

113 100

55 182.8 142.8 214.2

56 138.7 116.9 175.3









#35
4-Chloro-3-methylphenol
Concen: 41.956 ng/ul
RT: 12.187 min Scan# 1530
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11




Instrument :
BNA_P
ClientSampleId :
SSTD040635


Tgt Ion:107 Resp: 462139
Ion Ratio Lower Upper
107 100
144 23.2 18.4 27.6
142 71.2 58.6 87.8



#36
2-Methylnaphthalene
Concen: 40.742 ng/ul
RT: 12.546 min Scan# 1591
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11

Tgt Ion:142 Resp: 868247
Ion Ratio Lower Upper
142 100
141 89.0 73.2 109.8





#37

1-Methylnaphthalene

Concen: 40.118 ng/ul

RT: 12.763 min Scan# 1628

Delta R.T. 0.000 min

Lab File: BP015716.D

Acq: 26 Jun 2023 16:11

Instrument :

BNA_P

ClientSampleId :

SSTD040635





Tgt Ion:142 Resp: 872141

Ion Ratio Lower Upper

142 100

141 94.2 73.1 109.7

116 4.6 4.0 6.0



#39

1,2,4,5-Tetrachlorobenzene

Concen: 40.816 ng/ul

RT: 12.910 min Scan# 1653

Delta R.T. 0.000 min

Lab File: BP015716.D

Acq: 26 Jun 2023 16:11

Tgt Ion:216 Resp: 496745




Ion Ratio Lower Upper


216 100

214 78.5 62.4 93.6

179 22.3 17.6 26.4

108 18.6 15.5 23.3









#40
Hexachlorocyclopentadiene
Concen: 41.323 ng/ul
RT: 12.875 min Scan# 1647
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11




Instrument :
BNA_P
ClientSampleId :
SSTD040635


Tgt Ion:237 Resp: 150452
Ion Ratio Lower Upper
237 100
235 63.0 48.6 73.0
272 12.5 10.5 15.7



#41
2,4,6-Trichlorophenol
Concen: 42.161 ng/ul
RT: 13.157 min Scan# 1695
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11

Tgt Ion:196 Resp: 324275
Ion Ratio Lower Upper
196 100
198 98.6 75.4 113.2
200 32.6 23.6 35.4









#42
2,4,5-Trichlorophenol
Concen: 43.989 ng/ul
RT: 13.240 min Scan# 1709
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11




Instrument :
BNA_P
ClientSampleId :
SSTD040635


Tgt Ion:196 Resp: 350308
Ion Ratio Lower Upper
196 100
198 96.7 74.9 112.3
200 32.6 24.3 36.5



#43
1,1'-Biphenyl
Concen: 40.390 ng/ul
RT: 13.546 min Scan# 1761
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11

Tgt Ion:154 Resp: 1181509
Ion Ratio Lower Upper
154 100
153 40.4 32.8 49.2
76 13.4 10.7 16.1









#44
2-Chloronaphthalene
Concen: 40.528 ng/ul
RT: 13.593 min Scan# 1769
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11




Instrument :
BNA_P
ClientSampleId :
SSTD040635


Tgt Ion:162 Resp: 933945
Ion Ratio Lower Upper
162 100
164 32.5 26.1 39.1
127 42.2 33.4 50.0



#45
2-Nitroaniline
Concen: 44.052 ng/ul
RT: 13.810 min Scan# 1806
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11

Tgt Ion: 65 Resp: 316020
Ion Ratio Lower Upper
65 100
92 67.2 49.5 74.3
138 92.3 71.4 107.2









#47
Dimethylphthalate
Concen: 40.943 ng/ul
RT: 14.169 min Scan# 1867
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11




Instrument :
BNA_P
ClientSampleId :
SSTD040635


Tgt Ion:163 Resp: 1209543
Ion Ratio Lower Upper
163 100
194 3.9 3.1 4.7
164 10.0 7.7 11.5



#48
2,6-Dinitrotoluene
Concen: 42.798 ng/ul
RT: 14.304 min Scan# 1890
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11

Tgt Ion:165 Resp: 256463
Ion Ratio Lower Upper
165 100
89 67.7 54.1 81.1
121 22.0 18.2 27.2









#50
Acenaphthylene
Concen: 40.756 ng/ul
RT: 14.440 min Scan# 1913
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11




Instrument :
BNA_P
ClientSampleId :
SSTD040635


Tgt Ion:152 Resp: 1441012
Ion Ratio Lower Upper
152 100
151 19.7 16.2 24.2
153 12.6 10.5 15.7



#51
3-Nitroaniline
Concen: 44.618 ng/ul
RT: 14.640 min Scan# 1947
Delta R.T. -0.006 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11

Tgt Ion:138 Resp: 209668
Ion Ratio Lower Upper
138 100
108 12.1 9.1 13.7
92 128.9 96.8 145.2





#52

Acenaphthene

Concen: 40.233 ng/ul

RT: 14.781 min Scan# 1971

Delta R.T. 0.006 min

Lab File: BP015716.D

Acq: 26 Jun 2023 16:11

Instrument :

BNA_P

ClientSampleId :

SSTD040635





Tgt Ion:153 Resp: 986451

Ion Ratio Lower Upper

153 100

152 47.8 39.4 59.2

154 87.6 71.4 107.2



#53

2,4-Dinitrophenol

Concen: 45.088 ng/ul

RT: 14.863 min Scan# 1985

Delta R.T. -0.006 min

Lab File: BP015716.D

Acq: 26 Jun 2023 16:11




Tgt Ion:184 Resp: 142394


Ion Ratio Lower Upper

184 100

63 81.4 71.4 107.0

154 73.6 61.1 91.7









#55
4-Nitrophenol
Concen: 42.528 ng/ul
RT: 14.957 min Scan# 2001
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11




Instrument :
BNA_P
ClientSampleId :
SSTD040635


Tgt Ion:109 Resp: 177831
Ion Ratio Lower Upper
109 100
139 103.0 81.3 121.9
65 118.8 99.9 149.9



#56
Dibenzofuran
Concen: 40.363 ng/ul
RT: 15.116 min Scan# 2028
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11

Tgt Ion:168 Resp: 1373792
Ion Ratio Lower Upper
168 100
139 39.7 32.2 48.4









#57
2,4-Dinitrotoluene
Concen: 43.525 ng/ul
RT: 15.098 min Scan# 20
Delta R.T. -0.006 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11




Instrument :
BNA_P
ClientSampleId :
SSTD040635


Tgt Ion:165 Resp: 353711
Ion Ratio Lower Upper
165 100
63 61.5 56.8 85.2
182 2.5 2.0 3.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 42.671 ng/ul
RT: 15.345 min Scan# 2067
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11

Tgt Ion:232 Resp: 299012
Ion Ratio Lower Upper
232 100
131 47.2 38.2 57.2
130 2.7 2.2 3.2
166 32.2 25.6 38.4









#59
Diethylphthalate
Concen: 41.137 ng/ul
RT: 15.528 min Scan# 2097
Delta R.T. 0.006 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11




Instrument :
BNA_P
ClientSampleId :
SSTD040635


Tgt Ion:149 Resp: 1229805
Ion Ratio Lower Upper
149 100
177 21.8 16.8 25.2
150 12.4 10.3 15.5



#61
Fluorene
Concen: 40.147 ng/ul
RT: 15.763 min Scan# 2138
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11

Tgt Ion:166 Resp: 1108351
Ion Ratio Lower Upper
166 100
165 99.0 78.7 118.1
167 13.2 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 40.344 ng/ul
RT: 15.751 min Scan# 2136
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11




Instrument :
BNA_P
ClientSampleId :
SSTD040635


Tgt Ion:204 Resp: 569905
Ion Ratio Lower Upper
204 100
206 32.8 26.0 39.0
141 64.5 51.8 77.8



#63
4-Nitroaniline
Concen: 49.011 ng/ul
RT: 15.810 min Scan# 2146
Delta R.T. -0.006 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11

Tgt Ion:138 Resp: 155721
Ion Ratio Lower Upper
138 100
92 58.6 50.1 75.1
108 111.8 85.5 128.3









#66
4,6-Dinitro-2-methylphenol
Concen: 42.361 ng/ul
RT: 15.869 min Scan# 2157
Delta R.T. -0.006 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11




Instrument :
BNA_P
ClientSampleId :
SSTD040635


Tgt Ion:198 Resp: 204088
Ion Ratio Lower Upper
198 100
182 3.9 3.4 5.2
77 31.4 24.4 36.6



#67
N-Nitrosodiphenylamine
Concen: 40.593 ng/ul
RT: 15.969 min Scan# 2173
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11

Tgt Ion:169 Resp: 940901
Ion Ratio Lower Upper
169 100
168 65.7 54.2 81.4
167 36.6 29.7 44.5









#68
4-Bromophenyl-phenylether
Concen: 41.075 ng/ul
RT: 16.651 min Scan# 2289
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11




Instrument :
BNA_P
ClientSampleId :
SSTD040635


Tgt Ion:248 Resp: 363173
Ion Ratio Lower Upper
248 100
250 99.2 84.6 126.8
141 80.8 67.8 101.6



#69
Hexachlorobenzene
Concen: 40.089 ng/ul
RT: 16.775 min Scan# 2310
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11

Tgt Ion:284 Resp: 418231
Ion Ratio Lower Upper
284 100
142 35.2 27.8 41.8
249 31.5 26.1 39.1





#70

Atrazine

Concen: 41.596 ng/ul

RT: 16.928 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP015716.D

Acq: 26 Jun 2023 16:11

Instrument :

BNA_P

ClientSampleId :

SSTD040635





Tgt Ion:200 Resp: 369066

Ion Ratio Lower Upper

200 100

173 25.6 19.8 29.8

215 45.9 36.8 55.2



#71

Pentachlorophenol

Concen: 41.855 ng/ul

RT: 17.128 min Scan# 2370

Delta R.T. -0.006 min

Lab File: BP015716.D

Acq: 26 Jun 2023 16:11




Tgt Ion:266 Resp: 208122


Ion Ratio Lower Upper

266 100

264 62.8 53.4 80.2

268 64.3 51.4 77.2





#72

Phenanthrene

Concen: 39.906 ng/ul

RT: 17.516 min Scan# 2436

Delta R.T. 0.000 min

Lab File: BP015716.D

Acq: 26 Jun 2023 16:11

Instrument :

BNA_P

ClientSampleId :

SSTD040635





Tgt Ion:178 Resp: 1678568

Ion Ratio Lower Upper

178 100

179 15.3 12.4 18.6

176 19.1 15.6 23.4



#74

Anthracene

Concen: 40.298 ng/ul

RT: 17.610 min Scan# 2452

Delta R.T. 0.000 min

Lab File: BP015716.D

Acq: 26 Jun 2023 16:11




Tgt Ion:178 Resp: 1703228


Ion Ratio Lower Upper

178 100

179 15.2 12.1 18.1

176 18.5 14.6 22.0





#75

1,2,3,4-Tetrachlorobenzene

Concen: 40.063 ng/uL

RT: 13.516 min Scan# 1756

Delta R.T. 0.000 min

Lab File: BP015716.D

Acq: 26 Jun 2023 16:11

Instrument :

BNA_P

ClientSampleId :

SSTD040635





Tgt Ion:216 Resp: 504759

Ion Ratio Lower Upper

216 100

214 80.1 62.0 93.0

218 48.3 38.4 57.6



#76

Pentachlorobenzene

Concen: 39.530 ng/uL

RT: 15.034 min Scan# 2014

Delta R.T. 0.000 min

Lab File: BP015716.D

Acq: 26 Jun 2023 16:11




Tgt Ion:250 Resp: 502653


Ion Ratio Lower Upper

250 100

248 65.8 50.7 76.1

252 66.5 51.2 76.8





#77

Carbazole

Concen: 41.046 ng/ul

RT: 17.887 min Scan# 2499

Delta R.T. 0.000 min

Lab File: BP015716.D

Acq: 26 Jun 2023 16:11

Instrument :

BNA_P

ClientSampleId :

SSTD040635





Tgt Ion:167 Resp: 1462220

Ion Ratio Lower Upper

167 100

166 21.6 17.4 26.0

139 13.8 11.0 16.4



#78

Di-n-butylphthalate

Concen: 41.415 ng/ul

RT: 18.404 min Scan# 2587

Delta R.T. 0.000 min

Lab File: BP015716.D

Acq: 26 Jun 2023 16:11




Tgt Ion:149 Resp: 1955945


Ion Ratio Lower Upper

149 100

150 9.2 7.5 11.3

104 6.2 4.6 7.0





#80

Fluoranthene

Concen: 40.865 ng/ul

RT: 19.475 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP015716.D

Acq: 26 Jun 2023 16:11

Instrument :

BNA_P

ClientSampleId :

SSTD040635





Tgt Ion:202 Resp: 1834997

Ion Ratio Lower Upper

202 100

101 10.4 8.3 12.5

100 8.3 6.6 9.8



#82

Pyrene

Concen: 40.227 ng/ul

RT: 19.833 min Scan# 2830

Delta R.T. 0.000 min

Lab File: BP015716.D

Acq: 26 Jun 2023 16:11




Tgt Ion:202 Resp: 1842630


Ion Ratio Lower Upper

202 100

101 11.6 9.3 13.9

100 9.7 7.6 11.4





#83

Butylbenzylphthalate

Concen: 41.406 ng/ul

RT: 20.680 min Scan# 2974

Delta R.T. 0.000 min

Lab File: BP015716.D

Acq: 26 Jun 2023 16:11

Instrument :

BNA_P

ClientSampleId :

SSTD040635





Tgt Ion:149 Resp: 773765

Ion Ratio Lower Upper

149 100

91 77.3 59.8 89.6

206 20.6 16.4 24.6



#84

3,3'-Dichlorobenzidine

Concen: 39.379 ng/ul

RT: 21.475 min Scan# 3109

Delta R.T. 0.000 min

Lab File: BP015716.D

Acq: 26 Jun 2023 16:11




Tgt Ion:252 Resp: 492562


Ion Ratio Lower Upper

252 100

254 64.4 51.0 76.6

126 10.8 8.6 13.0





#85

Benzo(a)anthracene

Concen: 40.282 ng/ul

RT: 21.545 min Scan# 3121

Delta R.T. 0.000 min

Lab File: BP015716.D

Acq: 26 Jun 2023 16:11

Instrument :

BNA_P

ClientSampleId :

SSTD040635





Tgt Ion:228 Resp: 1567938

Ion Ratio Lower Upper

228 100

229 19.9 15.3 22.9

226 26.5 21.4 32.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 40.876 ng/ul

RT: 21.433 min Scan# 3102

Delta R.T. 0.000 min

Lab File: BP015716.D

Acq: 26 Jun 2023 16:11




Tgt Ion:149 Resp: 1049785


Ion Ratio Lower Upper

149 100

167 27.2 21.0 31.6

279 4.3 3.8 5.6





#87

Chrysene

Concen: 39.794 ng/ul

RT: 21.604 min Scan# 3131

Delta R.T. 0.000 min

Lab File: BP015716.D

Acq: 26 Jun 2023 16:11

Instrument :

BNA_P

ClientSampleId :

SSTD040635





Tgt Ion:228 Resp: 1469588

Ion Ratio Lower Upper

228 100

226 29.7 24.3 36.5

229 19.9 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 41.874 ng/ul




RT: 22.398 min Scan# 3266


Delta R.T. 0.000 min

Lab File: BP015716.D

Acq: 26 Jun 2023 16:11

Tgt Ion:149 Resp: 1642812





#90

Benzo(b)fluoranthene

Concen: 41.177 ng/ul

RT: 23.327 min Scan# 3424

Delta R.T. 0.000 min

Lab File: BP015716.D

Acq: 26 Jun 2023 16:11

Instrument :

BNA_P

ClientSampleId :

SSTD040635





Tgt Ion:252 Resp: 1420328

Ion Ratio Lower Upper

252 100

253 22.7 17.3 25.9

125 8.6 7.5 11.3



#91

Benzo(k)fluoranthene

Concen: 40.997 ng/ul

RT: 23.380 min Scan# 3433

Delta R.T. 0.000 min

Lab File: BP015716.D

Acq: 26 Jun 2023 16:11




Tgt Ion:252 Resp: 1428795


Ion Ratio Lower Upper

252 100

253 21.6 17.0 25.4

125 8.4 7.3 10.9









#93
Benzo(a)pyrene
Concen: 40.754 ng/ul
RT: 24.004 min Scan# 3539
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11




Instrument :
BNA_P
ClientSampleId :
SSTD040635


Tgt Ion:252 Resp: 1228311
Ion Ratio Lower Upper
252 100
253 21.9 17.7 26.5
125 9.7 8.2 12.2



#94
Indeno(1,2,3-cd)pyrene
Concen: 40.481 ng/ul
RT: 26.815 min Scan# 4017
Delta R.T. 0.000 min
Lab File: BP015716.D
Acq: 26 Jun 2023 16:11

Tgt Ion:276 Resp: 1656303
Ion Ratio Lower Upper
276 100
138 21.3 15.7 23.5
277 25.0 19.6 29.4





#95

Dibenzo(a,h)anthracene

Concen: 40.511 ng/ul

RT: 26.827 min Scan# 4019

Delta R.T. 0.000 min

Lab File: BP015716.D

Acq: 26 Jun 2023 16:11

Instrument :

BNA_P

ClientSampleId :

SSTD040635





Tgt Ion:278 Resp: 1361461

Ion Ratio Lower Upper

278 100

139 13.4 10.5 15.7

279 23.0 18.4 27.6



#96

Benzo(g,h,i)perylene

Concen: 40.871 ng/ul

RT: 27.651 min Scan# 4159

Delta R.T. -0.006 min

Lab File: BP015716.D

Acq: 26 Jun 2023 16:11

Tgt Ion:276 Resp: 1369697

Ion Ratio Lower Upper

276 100

138 18.6 15.1 22.7

277 24.2 18.8 28.2

