

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062623\
 Data File : BP015718.D
 Acq On : 26 Jun 2023 17:19
 Operator : MA/JU
 Sample : SSTD16094
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD160637

Quant Time: Jun 27 00:02:58 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 26 23:54:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.063	152	152568	20.000	ng/ul	0.00
20) Naphthalene-d8	10.893	136	579145	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.716	164	321408	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.475	188	592228	20.000	ng/ul	0.00
79) Chrysene-d12	21.569	240	454237	20.000	ng/ul	0.00
88) Perylene-d12	24.116	264	570908	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.381	96	267438	63.067	ng/uL	0.00
4) Pyridine-d5	3.822	84	1728341	161.686	ng/ul	0.00
7) Phenol-d5	7.240	99	2008536	153.864	ng/ul	0.01
9) Bis-(2-Chloroethyl)eth...	7.399	67	1202475	148.892	ng/ul	0.01
11) 2-Chlorophenol-d4	7.599	132	1523399	154.597	ng/ul	0.00
15) 4-Methylphenol-d8	8.805	113	1531858	148.545	ng/ul	0.02
21) Nitrobenzene-d5	9.257	128	749004	169.227	ng/ul	0.01
24) 2-Nitrophenol-d4	9.981	143	856715	165.846	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.528	165	1467165	159.382	ng/ul	0.00
31) 4-Chloroaniline-d4	11.046	131	1744355	147.513	ng/ul	0.00
46) Dimethylphthalate-d6	14.128	166	3570690	143.734	ng/ul	0.00
49) Acenaphthylene-d8	14.416	160	4239618	151.815	ng/ul	0.00
54) 4-Nitrophenol-d4	14.957	143	558214	170.275	ng/ul	0.01
60) Fluorene-d10	15.716	176	2946336	144.039	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.863	200	617361	170.372	ng/ul	0.00
73) Anthracene-d10	17.581	188	4073972	150.599	ng/ul	0.00
81) Pyrene-d10	19.804	212	4060420	136.896	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.963	264	4468045	155.145	ng/ul	0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.422	88	283058	62.027	ng/uL	98
5) Pyridine	3.846	79	1784125	159.544	ng/ul	95
6) Benzaldehyde	7.205	77	726507	141.975	ng/ul	96
8) Phenol	7.269	94	2062048	152.221	ng/ul	99
10) Bis(2-Chloroethyl)ether	7.493	93	1604236	148.843	ng/ul	99
12) 2-Chlorophenol	7.634	128	1594618	152.318	ng/ul	100
13) 2-Methylphenol	8.528	108	1522287	148.839	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.605	45	2123644	144.260	ng/ul	99
16) Acetophenone	8.916	105	2413269	142.359	ng/ul	98
17) N-Nitroso-di-n-propyla...	8.922	70	1275385	135.460	ng/ul	98
18) 4-Methylphenol	8.881	108	1613903	146.863	ng/ul	95
19) Hexachloroethane	9.146	117	712951	147.451	ng/ul	98
22) Nitrobenzene	9.305	77	2001189	161.007	ng/ul	96
23) Isophorone	9.846	82	3556601	149.530	ng/ul	99
25) 2-Nitrophenol	10.016	139	895072	163.265	ng/ul	97
26) 2,4-Dimethylphenol	10.075	107	1851804	153.654	ng/ul	98
27) Bis(2-Chloroethoxy)met...	10.304	93	2066268	153.887	ng/ul	99
29) 2,4-Dichlorophenol	10.552	162	1447039	158.115	ng/ul	99
30) Naphthalene	10.952	128	4717917	149.853	ng/ul	98
32) 4-Chloroaniline	11.075	127	1819588	148.105	ng/ul	98
33) Hexachlorobutadiene	11.216	225	1083552	156.781	ng/ul	99
34) Caprolactam	11.910	113	252161	91.391	ng/ul	97
35) 4-Chloro-3-methylphenol	12.204	107	1614795	150.739	ng/ul	99
36) 2-Methylnaphthalene	12.551	142	3048630	147.095	ng/ul	98


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062623\
 Data File : BP015718.D
 Acq On : 26 Jun 2023 17:19
 Operator : MA/JU
 Sample : SSTD16094
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD160637

Quant Time: Jun 27 00:02:58 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 26 23:54:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.769	142	3025549	143.100	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.916	216	1727849	163.136	ng/ul	99
40) Hexachlorocyclopentadiene	12.881	237	713466	225.169	ng/ul	98
41) 2,4,6-Trichlorophenol	13.163	196	1112845	166.256	ng/ul	98
42) 2,4,5-Trichlorophenol	13.246	196	1195491	172.498	ng/ul	97
43) 1,1'-Biphenyl	13.551	154	3955535	155.378	ng/ul	100
44) 2-Chloronaphthalene	13.604	162	3120257	155.586	ng/ul	99
45) 2-Nitroaniline	13.822	65	1054286	168.872	ng/ul	97
47) Dimethylphthalate	14.181	163	3682520	143.235	ng/ul	99
48) 2,6-Dinitrotoluene	14.316	165	810646	155.442	ng/ul	97
50) Acenaphthylene	14.445	152	4568642	148.474	ng/ul	99
51) 3-Nitroaniline	14.645	138	635318	155.349	ng/ul	94
52) Acenaphthene	14.787	153	3180855	149.073	ng/ul	98
53) 2,4-Dinitrophenol	14.863	184	501424	182.438	ng/ul	96
55) 4-Nitrophenol	14.969	109	593279	163.030	ng/ul	96
56) Dibenzofuran	15.122	168	4194339	141.602	ng/ul	97
57) 2,4-Dinitrotoluene	15.110	165	1043931	147.605	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.351	232	936932	153.638	ng/ul	99
59) Diethylphthalate	15.534	149	3578651	137.550	ng/ul	99
61) Fluorene	15.775	166	3311371	137.825	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.757	204	1739373	141.484	ng/ul	100
63) 4-Nitroaniline	15.822	138	449880	162.701	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.881	198	618482	167.835	ng/ul#	96
67) N-Nitrosodiphenylamine	15.981	169	2777256	156.648	ng/ul	99
68) 4-Bromophenyl-phenylether	16.657	248	1104112	163.262	ng/ul	93
69) Hexachlorobenzene	16.781	284	1254697	157.235	ng/ul	100
70) Atrazine	16.939	200	1001364	147.551	ng/ul	99
71) Pentachlorophenol	17.134	266	688341	180.983	ng/ul	97
72) Phenanthrene	17.522	178	4806863	149.406	ng/ul	100
74) Anthracene	17.616	178	4775259	147.711	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.522	216	1736355	180.181	ng/ul	99
76) Pentachlorobenzene	15.040	250	1627574	167.343	ng/ul	99
77) Carbazole	17.892	167	4050289	148.645	ng/ul	100
78) Di-n-butylphthalate	18.410	149	5303705	146.821	ng/ul	99
80) Fluoranthene	19.480	202	5069422	137.523	ng/ul	99
82) Pyrene	19.839	202	5053960	134.405	ng/ul	99
83) Butylbenzylphthalate	20.686	149	2279318	148.579	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.480	252	1598022	155.626	ng/ul	97
85) Benzo(a)anthracene	21.551	228	4878367	152.671	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.439	149	3285859	155.856	ng/ul	97
87) Chrysene	21.610	228	4604130	151.871	ng/ul	99
89) Di-n-octyl phthalate	22.404	149	5785500	138.666	ng/ul	100
90) Benzo(b)fluoranthene	23.345	252	5516198	150.374	ng/ul	98
91) Benzo(k)fluoranthene	23.392	252	5424927	146.368	ng/ul	98
93) Benzo(a)pyrene	24.016	252	4970580	155.072	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.868	276	7363359	169.221	ng/ul	97
95) Dibenzo(a,h)anthracene	26.868	278	6060699	169.573	ng/ul	98
96) Benzo(g,h,i)perylene	27.686	276	6129106	171.972	ng/ul	98





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 62.027 ng/uL
RT: 3.422 min Scan# 40
Delta R.T. 0.000 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19




Instrument :
BNA_P
ClientSampleId :
SSTD160637


Tgt Ion: 88 Resp: 283058
Ion Ratio Lower Upper
88 100
43 31.8 27.4 41.2
58 74.0 59.8 89.8



#5
Pyridine
Concen: 159.544 ng/uL
RT: 3.846 min Scan# 112
Delta R.T. 0.000 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19

Tgt Ion: 79 Resp: 1784125
Ion Ratio Lower Upper
79 100
52 76.0 65.5 98.3
51 38.4 31.8 47.8









#6
Benzaldehyde
Concen: 141.975 ng/ul
RT: 7.205 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19




Instrument :
BNA_P
ClientSampleId :
SSTD160637


Tgt Ion: 77 Resp: 726507
Ion Ratio Lower Upper
77 100
105 93.4 73.4 110.0
106 89.3 66.6 100.0



#8
Phenol
Concen: 152.221 ng/ul
RT: 7.269 min Scan# 694
Delta R.T. 0.012 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19

Tgt Ion: 94 Resp: 2062048
Ion Ratio Lower Upper
94 100
65 32.3 26.4 39.6
66 46.6 37.3 55.9









#10
Bis(2-Chloroethyl)ether
Concen: 148.843 ng/ul
RT: 7.493 min Scan# 71
Delta R.T. 0.012 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19




Instrument :
BNA_P
ClientSampleId :
SSTD160637


Tgt Ion: 93 Resp: 1604236
Ion Ratio Lower Upper
93 100
63 77.8 62.8 94.2
95 32.5 25.6 38.4



#12
2-Chlorophenol
Concen: 152.318 ng/ul
RT: 7.634 min Scan# 756
Delta R.T. 0.006 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19

Tgt Ion:128 Resp: 1594618
Ion Ratio Lower Upper
128 100
64 51.9 41.2 61.8
130 32.6 26.0 39.0





#13

2-Methylphenol

Concen: 148.839 ng/ul

RT: 8.528 min Scan# 907

Delta R.T. 0.006 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19

Instrument :

BNA_P

ClientSampleId :





SSTD160637

Tgt Ion:108 Resp: 1522287

Ion Ratio Lower Upper

108 100

107 90.5 73.8 110.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 144.260 ng/ul

RT: 8.605 min Scan# 921

Delta R.T. 0.012 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19




Tgt Ion: 45 Resp: 2123644


Ion Ratio Lower Upper

45 100

77 15.7 12.4 18.6

79 12.3 9.3 13.9









#16
Acetophenone
Concen: 142.359 ng/ul
RT: 8.916 min Scan# 91
Delta R.T. 0.012 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19




Instrument :
BNA_P
ClientSampleId :
SSTD160637


Tgt Ion:105 Resp: 2413269
Ion Ratio Lower Upper
105 100
77 82.9 65.0 97.4
51 30.9 25.1 37.7



#17
N-Nitroso-di-n-propylamine
Concen: 135.460 ng/ul
RT: 8.922 min Scan# 975
Delta R.T. 0.041 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19

Tgt Ion: 70 Resp: 1275385
Ion Ratio Lower Upper
70 100
42 61.2 48.2 72.2
101 9.8 7.6 11.4
130 19.1 13.9 20.9





#18

4-Methylphenol

Concen: 146.863 ng/ul

RT: 8.881 min Scan# 968

Delta R.T. 0.024 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19

Instrument :

BNA_P

ClientSampleId :





SSTD160637

Tgt Ion:108 Resp: 1613903

Ion Ratio Lower Upper

108 100

107 109.0 91.3 136.9



#19

Hexachloroethane

Concen: 147.451 ng/ul

RT: 9.146 min Scan# 1013

Delta R.T. 0.000 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19




Tgt Ion:117 Resp: 712951


Ion Ratio Lower Upper

117 100

201 103.2 84.2 126.2

199 63.9 51.8 77.8





#22

Nitrobenzene

Concen: 161.007 ng/ul

RT: 9.305 min Scan# 1040

Delta R.T. 0.018 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19

Instrument :





BNA_P

ClientSampleId :

SSTD160637

Tgt Ion: 77 Resp: 2001189

Ion	Ratio	Lower	Upper
77	100		
123	41.4	31.0	46.4
65	14.5	11.3	16.9



#23

Isophorone

Concen: 149.530 ng/ul

RT: 9.846 min Scan# 1132




Delta R.T. 0.030 min


Lab File: BP015718.D

Acq: 26 Jun 2023 17:19

Tgt Ion: 82 Resp: 3556601

Ion	Ratio	Lower	Upper
82	100		
95	8.2	6.3	9.5
138	16.0	12.3	18.5









#25
2-Nitrophenol
Concen: 163.265 ng/ul
RT: 10.016 min Scan# 1161
Delta R.T. 0.012 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19




Instrument :
BNA_P
ClientSampleId :
SSTD160637


Tgt Ion:139 Resp: 895072
Ion Ratio Lower Upper
139 100
65 60.2 50.2 75.2
109 40.7 31.3 46.9



#26
2,4-Dimethylphenol
Concen: 153.654 ng/ul
RT: 10.075 min Scan# 1171
Delta R.T. 0.012 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19

Tgt Ion:107 Resp: 1851804
Ion Ratio Lower Upper
107 100
121 46.2 38.2 57.2
122 76.4 62.1 93.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 153.887 ng/ul
 RT: 10.304 min Scan# 1210
 Delta R.T. 0.012 min
 Lab File: BP015718.D
 Acq: 26 Jun 2023 17:19

Instrument :
 BNA_P
 ClientSampleId :
 SSTD160637

Tgt Ion: 93 Resp: 2066268




Ion	Ratio	Lower	Upper
93	100		
95	32.7	26.6	39.8
123	10.6	8.6	13.0




#29
 2,4-Dichlorophenol
 Concen: 158.115 ng/ul
 RT: 10.552 min Scan# 1252
 Delta R.T. 0.000 min
 Lab File: BP015718.D
 Acq: 26 Jun 2023 17:19

Tgt Ion:162 Resp: 1447039

Ion	Ratio	Lower	Upper
162	100		
164	65.9	53.0	79.6
98	39.7	31.3	46.9









#30
Naphthalene
Concen: 149.853 ng/ul
RT: 10.952 min Scan# 1320
Delta R.T. 0.012 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19




Instrument :
BNA_P
ClientSampleId :
SSTD160637


Tgt Ion:128 Resp: 4717917
Ion Ratio Lower Upper
128 100
129 11.0 8.6 12.8
127 13.8 10.3 15.5



#32
4-Chloroaniline
Concen: 148.105 ng/ul
RT: 11.075 min Scan# 1341
Delta R.T. 0.006 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19

Tgt Ion:127 Resp: 1819588
Ion Ratio Lower Upper
127 100
129 32.3 24.7 37.1









#33
Hexachlorobutadiene
Concen: 156.781 ng/ul
RT: 11.216 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19




Instrument :
BNA_P
ClientSampleId :
SSTD160637


Tgt Ion:225 Resp: 1083552
Ion Ratio Lower Upper
225 100
223 62.6 50.4 75.6
227 65.2 51.0 76.4



#34
Caprolactam
Concen: 91.391 ng/ul
RT: 11.910 min Scan# 1483
Delta R.T. 0.053 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19

Tgt Ion:113 Resp: 252161
Ion Ratio Lower Upper
113 100
55 175.9 142.8 214.2
56 141.4 116.9 175.3









#35
4-Chloro-3-methylphenol
Concen: 150.739 ng/ul
RT: 12.204 min Scan# 1533
Delta R.T. 0.018 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19




Instrument :
BNA_P
ClientSampleId :
SSTD160637


Tgt Ion:107 Resp: 1614795
Ion Ratio Lower Upper
107 100
144 23.6 18.4 27.6
142 72.6 58.6 87.8



#36
2-Methylnaphthalene
Concen: 147.095 ng/ul
RT: 12.551 min Scan# 1592
Delta R.T. 0.006 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19

Tgt Ion:142 Resp: 3048630
Ion Ratio Lower Upper
142 100
141 90.0 73.2 109.8





#37

1-Methylnaphthalene

Concen: 143.100 ng/ul

RT: 12.769 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19

Instrument :

BNA_P

ClientSampleId :

SSTD160637





Tgt Ion:142 Resp: 3025549

Ion Ratio Lower Upper

142 100

141 92.9 73.1 109.7

116 4.7 4.0 6.0



#39

1,2,4,5-Tetrachlorobenzene

Concen: 163.136 ng/ul

RT: 12.916 min Scan# 1654

Delta R.T. 0.006 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19

Tgt Ion:216 Resp: 1727849




Ion Ratio Lower Upper


216 100

214 78.8 62.4 93.6

179 22.5 17.6 26.4

108 18.3 15.5 23.3





#40

Hexachlorocyclopentadiene

Concen: 225.169 ng/ul

RT: 12.881 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19

Instrument :

BNA_P

ClientSampleId :

SSTD160637





Tgt Ion:237 Resp: 713466

Ion Ratio Lower Upper

237 100

235 62.6 48.6 73.0

272 12.5 10.5 15.7



#41

2,4,6-Trichlorophenol

Concen: 166.256 ng/ul

RT: 13.163 min Scan# 1696

Delta R.T. 0.006 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19




Tgt Ion:196 Resp: 1112845


Ion Ratio Lower Upper

196 100

198 96.6 75.4 113.2

200 30.9 23.6 35.4









#42
2,4,5-Trichlorophenol
Concen: 172.498 ng/ul
RT: 13.246 min Scan# 1710
Delta R.T. 0.006 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19




Instrument :
BNA_P
ClientSampleId :
SSTD160637


Tgt Ion:196 Resp: 1195491
Ion Ratio Lower Upper
196 100
198 96.5 74.9 112.3
200 32.0 24.3 36.5



#43
1,1'-Biphenyl
Concen: 155.378 ng/ul
RT: 13.551 min Scan# 1762
Delta R.T. 0.006 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19

Tgt Ion:154 Resp: 3955535
Ion Ratio Lower Upper
154 100
153 41.2 32.8 49.2
76 13.0 10.7 16.1





#44

2-Chloronaphthalene

Concen: 155.586 ng/ul

RT: 13.604 min Scan# 1771

Delta R.T. 0.012 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19

Instrument :

BNA_P

ClientSampleId :

SSTD160637





Tgt Ion:162 Resp: 3120257

Ion Ratio Lower Upper

162 100

164 33.4 26.1 39.1

127 42.3 33.4 50.0



#45

2-Nitroaniline

Concen: 168.872 ng/ul

RT: 13.822 min Scan# 1808

Delta R.T. 0.012 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19




Tgt Ion: 65 Resp: 1054286


Ion Ratio Lower Upper

65 100

92 63.6 49.5 74.3

138 93.1 71.4 107.2





#47

Dimethylphthalate

Concen: 143.235 ng/ul

RT: 14.181 min Scan# 1869

Delta R.T. 0.012 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19

Instrument :

BNA_P

ClientSampleId :

SSTD160637





Tgt Ion:163 Resp: 3682520

Ion Ratio Lower Upper

163 100

194 4.0 3.1 4.7

164 9.9 7.7 11.5



#48

2,6-Dinitrotoluene

Concen: 155.442 ng/ul

RT: 14.316 min Scan# 1892

Delta R.T. 0.012 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19




Tgt Ion:165 Resp: 810646


Ion Ratio Lower Upper

165 100

89 65.3 54.1 81.1

121 21.3 18.2 27.2





#50

Acenaphthylene

Concen: 148.474 ng/ul

RT: 14.445 min Scan# 1914

Delta R.T. 0.006 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19

Instrument :

BNA_P

ClientSampleId :

SSTD160637





Tgt Ion:152 Resp: 4568642

Ion Ratio Lower Upper

152 100

151 20.5 16.2 24.2

153 13.4 10.5 15.7



#51

3-Nitroaniline

Concen: 155.349 ng/ul

RT: 14.645 min Scan# 1948

Delta R.T. 0.000 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19




Tgt Ion:138 Resp: 635318


Ion Ratio Lower Upper

138 100

108 11.9 9.1 13.7

92 128.2 96.8 145.2





#52

Acenaphthene

Concen: 149.073 ng/ul

RT: 14.787 min Scan# 1972

Delta R.T. 0.012 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19

Instrument :

BNA_P

ClientSampleId :

SSTD160637





Tgt Ion:153 Resp: 3180855

Ion Ratio Lower Upper

153 100

152 47.3 39.4 59.2

154 87.7 71.4 107.2



#53

2,4-Dinitrophenol

Concen: 182.438 ng/ul

RT: 14.863 min Scan# 1985

Delta R.T. -0.006 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19




Tgt Ion:184 Resp: 501424


Ion Ratio Lower Upper

184 100

63 81.9 71.4 107.0

154 76.5 61.1 91.7









#55
4-Nitrophenol
Concen: 163.030 ng/ul
RT: 14.969 min Scan# 2001
Delta R.T. 0.012 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19




Instrument :
BNA_P
ClientSampleId :
SSTD160637


Tgt Ion:109 Resp: 593279
Ion Ratio Lower Upper
109 100
139 98.1 81.3 121.9
65 120.0 99.9 149.9



#56
Dibenzofuran
Concen: 141.602 ng/ul
RT: 15.122 min Scan# 2029
Delta R.T. 0.006 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19

Tgt Ion:168 Resp: 4194339
Ion Ratio Lower Upper
168 100
139 41.9 32.2 48.4





#57

2,4-Dinitrotoluene

Concen: 147.605 ng/ul

RT: 15.110 min Scan# 2026

Delta R.T. 0.006 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19

Instrument :

BNA_P

ClientSampleId :

SSTD160637





Tgt Ion:165 Resp: 1043931

Ion Ratio Lower Upper

165 100

63 66.5 56.8 85.2

182 2.8 2.0 3.0



#58

2,3,4,6-Tetrachlorophenol

Concen: 153.638 ng/ul

RT: 15.351 min Scan# 2068

Delta R.T. 0.006 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19

Tgt Ion:232 Resp: 936932




Ion Ratio Lower Upper


232 100

131 47.1 38.2 57.2

130 2.7 2.2 3.2

166 32.0 25.6 38.4









#59
Diethylphthalate
Concen: 137.550 ng/ul
RT: 15.534 min Scan# 2099
Delta R.T. 0.012 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19




Instrument :
BNA_P
ClientSampleId :
SSTD160637


Tgt Ion:149 Resp: 3578651
Ion Ratio Lower Upper
149 100
177 21.6 16.8 25.2
150 13.0 10.3 15.5



#61
Fluorene
Concen: 137.825 ng/ul
RT: 15.775 min Scan# 2140
Delta R.T. 0.012 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19

Tgt Ion:166 Resp: 3311371
Ion Ratio Lower Upper
166 100
165 98.4 78.7 118.1
167 13.4 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 141.484 ng/ul
RT: 15.757 min Scan# 2137
Delta R.T. 0.006 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19




Instrument :
BNA_P
ClientSampleId :
SSTD160637


Tgt Ion:204 Resp: 1739373
Ion Ratio Lower Upper
204 100
206 32.9 26.0 39.0
141 64.8 51.8 77.8



#63
4-Nitroaniline
Concen: 162.701 ng/ul
RT: 15.822 min Scan# 2148
Delta R.T. 0.006 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19

Tgt Ion:138 Resp: 449880
Ion Ratio Lower Upper
138 100
92 55.8 50.1 75.1
108 104.9 85.5 128.3









#66
4,6-Dinitro-2-methylphenol
Concen: 167.835 ng/ul
RT: 15.881 min Scan# 2158
Delta R.T. 0.006 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19




Instrument :
BNA_P
ClientSampleId :
SSTD160637


Tgt Ion:198 Resp: 618482
Ion Ratio Lower Upper
198 100
182 2.9 3.4 5.2#
77 28.6 24.4 36.6



#67
N-Nitrosodiphenylamine
Concen: 156.648 ng/ul
RT: 15.981 min Scan# 2175
Delta R.T. 0.012 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19

Tgt Ion:169 Resp: 2777256
Ion Ratio Lower Upper
169 100
168 67.4 54.2 81.4
167 36.3 29.7 44.5









#68
4-Bromophenyl-phenylether
Concen: 163.262 ng/ul
RT: 16.657 min Scan# 2111
Delta R.T. 0.006 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19




Instrument :
BNA_P
ClientSampleId :
SSTD160637


Tgt Ion:248 Resp: 1104112
Ion Ratio Lower Upper
248 100
250 97.5 84.6 126.8
141 80.3 67.8 101.6



#69
Hexachlorobenzene
Concen: 157.235 ng/ul
RT: 16.781 min Scan# 2311
Delta R.T. 0.006 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19

Tgt Ion:284 Resp: 1254697
Ion Ratio Lower Upper
284 100
142 34.6 27.8 41.8
249 32.4 26.1 39.1





#70

Atrazine

Concen: 147.551 ng/ul

RT: 16.939 min Scan# 21

Delta R.T. 0.012 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19

Instrument :

BNA_P

ClientSampleId :

SSTD160637





Tgt Ion:200 Resp: 1001364

Ion Ratio Lower Upper

200 100

173 26.0 19.8 29.8

215 46.0 36.8 55.2



#71

Pentachlorophenol

Concen: 180.983 ng/ul

RT: 17.134 min Scan# 2371

Delta R.T. 0.000 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19




Tgt Ion:266 Resp: 688341


Ion Ratio Lower Upper

266 100

264 62.4 53.4 80.2

268 63.6 51.4 77.2





#72

Phenanthrene

Concen: 149.406 ng/ul

RT: 17.522 min Scan# 2436

Delta R.T. 0.006 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19

Instrument :

BNA_P

ClientSampleId :

SSTD160637





Tgt Ion:178 Resp: 4806863

Ion Ratio Lower Upper

178 100

179 15.6 12.4 18.6

176 19.4 15.6 23.4



#74

Anthracene

Concen: 147.711 ng/ul

RT: 17.616 min Scan# 2453

Delta R.T. 0.006 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19




Tgt Ion:178 Resp: 4775259


Ion Ratio Lower Upper

178 100

179 15.6 12.1 18.1

176 19.1 14.6 22.0









#75
1,2,3,4-Tetrachlorobenzene
Concen: 180.181 ng/uL
RT: 13.522 min Scan# 1756
Delta R.T. 0.006 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19




Instrument :
BNA_P
ClientSampleId :
SSTD160637


Tgt Ion:216 Resp: 1736355
Ion Ratio Lower Upper
216 100
214 78.2 62.0 93.0
218 48.0 38.4 57.6



#76
Pentachlorobenzene
Concen: 167.343 ng/uL
RT: 15.040 min Scan# 2015
Delta R.T. 0.006 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19

Tgt Ion:250 Resp: 1627574
Ion Ratio Lower Upper
250 100
248 62.6 50.7 76.1
252 63.4 51.2 76.8









#77
Carbazole
Concen: 148.645 ng/ul
RT: 17.892 min Scan# 2499
Delta R.T. 0.006 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19




Instrument :
BNA_P
ClientSampleId :
SSTD160637


Tgt Ion:167 Resp: 4050289
Ion Ratio Lower Upper
167 100
166 21.7 17.4 26.0
139 13.6 11.0 16.4



#78
Di-n-butylphthalate
Concen: 146.821 ng/ul
RT: 18.410 min Scan# 2588
Delta R.T. 0.006 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19

Tgt Ion:149 Resp: 5303705
Ion Ratio Lower Upper
149 100
150 9.1 7.5 11.3
104 6.1 4.6 7.0





#80

Fluoranthene

Concen: 137.523 ng/ul

RT: 19.480 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19

Instrument :

BNA_P

ClientSampleId :

SSTD160637





Tgt Ion:202 Resp: 5069422

Ion Ratio Lower Upper

202 100

101 10.0 8.3 12.5

100 7.7 6.6 9.8



#82

Pyrene

Concen: 134.405 ng/ul

RT: 19.839 min Scan# 2831

Delta R.T. 0.006 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19




Tgt Ion:202 Resp: 5053960


Ion Ratio Lower Upper

202 100

101 11.0 9.3 13.9

100 9.1 7.6 11.4





#83

Butylbenzylphthalate

Concen: 148.579 ng/ul

RT: 20.686 min Scan# 2974

Delta R.T. 0.006 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19

Instrument :

BNA_P

ClientSampleId :

SSTD160637





Tgt Ion:149 Resp: 2279318

Ion Ratio Lower Upper

149 100

91 74.2 59.8 89.6

206 21.3 16.4 24.6



#84

3,3'-Dichlorobenzidine

Concen: 155.626 ng/ul

RT: 21.480 min Scan# 3110

Delta R.T. 0.006 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19




Tgt Ion:252 Resp: 1598022


Ion Ratio Lower Upper

252 100

254 66.3 51.0 76.6

126 10.6 8.6 13.0









#85
Benzo(a)anthracene
Concen: 152.671 ng/ul
RT: 21.551 min Scan# 3121
Delta R.T. 0.006 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19




Instrument :
BNA_P
ClientSampleId :
SSTD160637


Tgt Ion:228 Resp: 4878367
Ion Ratio Lower Upper
228 100
229 20.3 15.3 22.9
226 27.5 21.4 32.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 155.856 ng/ul
RT: 21.439 min Scan# 3103
Delta R.T. 0.006 min
Lab File: BP015718.D
Acq: 26 Jun 2023 17:19

Tgt Ion:149 Resp: 3285859
Ion Ratio Lower Upper
149 100
167 28.0 21.0 31.6
279 5.0 3.8 5.6





#87

Chrysene

Concen: 151.871 ng/ul

RT: 21.610 min Scan# 3131

Delta R.T. 0.006 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19

Instrument :

BNA_P

ClientSampleId :

SSTD160637





Tgt Ion:228 Resp: 4604130

Ion Ratio Lower Upper

228 100

226 30.5 24.3 36.5

229 20.0 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 138.666 ng/ul




RT: 22.404 min Scan# 3267


Delta R.T. 0.006 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19

Tgt Ion:149 Resp: 5785500





#90

Benzo(b)fluoranthene

Concen: 150.374 ng/ul

RT: 23.345 min Scan# 3424

Delta R.T. 0.018 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19

Instrument :

BNA_P

ClientSampleId :

SSTD160637





Tgt Ion:252 Resp: 5516198

Ion Ratio Lower Upper

252 100

253 21.9 17.3 25.9

125 7.7 7.5 11.3



#91

Benzo(k)fluoranthene

Concen: 146.368 ng/ul

RT: 23.392 min Scan# 3435

Delta R.T. 0.012 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19




Tgt Ion:252 Resp: 5424927


Ion Ratio Lower Upper

252 100

253 22.2 17.0 25.4

125 8.2 7.3 10.9





#93

Benzo(a)pyrene

Concen: 155.072 ng/ul

RT: 24.016 min Scan# 3

Delta R.T. 0.012 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19

Instrument :

BNA_P

ClientSampleId :

SSTD160637





Tgt Ion:252 Resp: 4970580

Ion Ratio Lower Upper

252 100

253 21.8 17.7 26.5

125 9.0 8.2 12.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 169.221 ng/ul

RT: 26.868 min Scan# 4026

Delta R.T. 0.053 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19




Tgt Ion:276 Resp: 7363359


Ion Ratio Lower Upper

276 100

138 20.6 15.7 23.5

277 26.1 19.6 29.4





#95

Dibenzo(a,h)anthracene

Concen: 169.573 ng/ul

RT: 26.868 min Scan# 4019

Delta R.T. 0.041 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19

Instrument :

BNA_P

ClientSampleId :

SSTD160637





Tgt Ion:278 Resp: 6060699

Ion Ratio Lower Upper

278 100

139 11.8 10.5 15.7

279 23.6 18.4 27.6



#96

Benzo(g,h,i)perylene

Concen: 171.972 ng/ul

RT: 27.686 min Scan# 4165

Delta R.T. 0.030 min

Lab File: BP015718.D

Acq: 26 Jun 2023 17:19

Tgt Ion:276 Resp: 6129106

Ion Ratio Lower Upper

276 100

138 17.3 15.1 22.7

277 23.9 18.8 28.2

