

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062623\
 Data File : BP015734.D
 Acq On : 27 Jun 2023 03:28
 Operator : MA/JU
 Sample : PB153446BS
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS446

Quant Time: Jun 27 04:17:54 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jun 27 00:56:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.063	152	116558	20.000	ng/u1	0.00
20) Naphthalene-d8	10.887	136	428223	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.710	164	248509	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.475	188	545522	20.000	ng/u1	0.00
79) Chrysene-d12	21.563	240	551717	20.000	ng/u1	0.00
88) Perylene-d12	24.110	264	685552	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.387	96	18563	5.730	ng/uL	0.00
4) Pyridine-d5	3.828	84	229588	28.113	ng/u1	0.00
7) Phenol-d5	7.228	99	284670	28.544	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.387	67	183140	29.682	ng/u1	0.00
11) 2-Chlorophenol-d4	7.593	132	235953	31.342	ng/u1	0.00
15) 4-Methylphenol-d8	8.787	113	225193	28.583	ng/u1	0.00
21) Nitrobenzene-d5	9.246	128	110803	33.857	ng/u1	0.00
24) 2-Nitrophenol-d4	9.969	143	122671	32.116	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.522	165	214284	31.482	ng/u1	0.00
31) 4-Chloroaniline-d4	11.040	131	274188	31.359	ng/u1	0.00
46) Dimethylphthalate-d6	14.116	166	627737	32.681	ng/u1	0.00
49) Acenaphthylene-d8	14.404	160	709686	32.868	ng/u1	0.00
54) 4-Nitrophenol-d4	14.945	143	93915	37.051	ng/u1	0.00
60) Fluorene-d10	15.704	176	515239	32.578	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.857	200	108735	32.411	ng/u1	0.00
73) Anthracene-d10	17.575	188	824929	33.105	ng/u1	0.00
81) Pyrene-d10	19.798	212	1036673	28.776	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.945	264	1182549	34.195	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.423	88	39462	11.319	ng/uL	99
5) Pyridine	3.852	79	229967	26.918	ng/u1	94
6) Benzaldehyde	7.199	77	179233	44.452	ng/u1	99
8) Phenol	7.258	94	288315	27.859	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.481	93	231164	28.074	ng/u1	99
12) 2-Chlorophenol	7.622	128	231467	28.941	ng/u1	98
13) 2-Methylphenol	8.516	108	213055	27.267	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.599	45	312848	27.818	ng/u1	99
16) Acetophenone	8.905	105	342887	26.476	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.881	70	182935	25.433	ng/u1	99
18) 4-Methylphenol	8.858	108	226289	26.954	ng/u1	96
19) Hexachloroethane	9.146	117	110490	29.911	ng/u1	99
22) Nitrobenzene	9.287	77	281435	30.623	ng/u1	98
23) Isophorone	9.810	82	499668	28.411	ng/u1	99
25) 2-Nitrophenol	10.005	139	121976	30.090	ng/u1	98
26) 2,4-Dimethylphenol	10.063	107	240809	27.023	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.293	93	288829	29.092	ng/u1	99
29) 2,4-Dichlorophenol	10.546	162	205470	30.364	ng/u1	99
30) Naphthalene	10.940	128	697185	29.949	ng/u1	98
32) 4-Chloroaniline	11.063	127	240291	26.452	ng/u1	99
33) Hexachlorobutadiene	11.210	225	156607	30.646	ng/u1	98
34) Caprolactam	11.857	113	63236	30.550	ng/u1	97
35) 4-Chloro-3-methylphenol	12.193	107	227929	28.776	ng/u1	97
36) 2-Methylnaphthalene	12.546	142	446741	29.152	ng/u1	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062623\
 Data File : BP015734.D
 Acq On : 27 Jun 2023 03:28
 Operator : MA/JU
 Sample : PB153446BS
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS446

Quant Time: Jun 27 04:17:54 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jun 27 00:56:58 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.763	142	452736	28.960	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.910	216	254134	31.033	ng/ul	98
40) Hexachlorocyclopentadiene	12.875	237	59760	24.393	ng/ul	99
41) 2,4,6-Trichlorophenol	13.157	196	157524	30.437	ng/ul	98
42) 2,4,5-Trichlorophenol	13.240	196	170087	30.984	ng/ul	97
43) 1,1'-Biphenyl	13.546	154	592299	30.091	ng/ul	99
44) 2-Chloronaphthalene	13.593	162	473286	30.522	ng/ul	99
45) 2-Nitroaniline	13.810	65	156067	31.987	ng/ul	97
47) Dimethylphthalate	14.163	163	597145	30.040	ng/ul	100
48) 2,6-Dinitrotoluene	14.304	165	126016	31.252	ng/ul	99
50) Acenaphthylene	14.434	152	721712	30.335	ng/ul	99
51) 3-Nitroaniline	14.640	138	119306	37.731	ng/ul	87
52) Acenaphthene	14.775	153	502745	30.473	ng/ul	98
53) 2,4-Dinitrophenol	14.869	184	62101	29.029	ng/ul	93
55) 4-Nitrophenol	14.963	109	87727	33.398	ng/ul	94
56) Dibenzofuran	15.110	168	696099	30.394	ng/ul	100
57) 2,4-Dinitrotoluene	15.104	165	183902	33.468	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.345	232	147250	31.229	ng/ul	99
59) Diethylphthalate	15.522	149	614929	30.569	ng/ul	98
61) Fluorene	15.763	166	570637	30.718	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.745	204	286562	30.147	ng/ul	99
63) 4-Nitroaniline	15.810	138	115179	53.221	ng/ul	94
66) 4,6-Dinitro-2-methylph...	15.869	198	101635	29.942	ng/ul#	98
67) N-Nitrosodiphenylamine	15.969	169	486109	29.766	ng/ul	100
68) 4-Bromophenyl-phenylether	16.651	248	182416	29.283	ng/ul	92
69) Hexachlorobenzene	16.775	284	220868	30.048	ng/ul	99
70) Atrazine	16.922	200	142266	22.758	ng/ul	100
71) Pentachlorophenol	17.128	266	110173	31.230	ng/ul	98
72) Phenanthrene	17.516	178	929835	31.375	ng/ul	99
74) Anthracene	17.610	178	924829	31.057	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.516	216	250831	28.257	ng/uL	99
76) Pentachlorobenzene	15.034	250	252568	28.192	ng/uL	99
77) Carbazole	17.886	167	846198	33.714	ng/ul	99
78) Di-n-butylphthalate	18.404	149	1101583	33.106	ng/ul	99
80) Fluoranthene	19.475	202	1166885	26.062	ng/ul	100
82) Pyrene	19.828	202	1215939	26.623	ng/ul	98
83) Butylbenzylphthalate	20.675	149	550902	29.566	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.474	252	437596	35.086	ng/ul	99
85) Benzo(a)anthracene	21.545	228	1224276	31.545	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.433	149	796697	31.112	ng/ul	99
87) Chrysene	21.604	228	1173398	31.867	ng/ul	99
89) Di-n-octyl phthalate	22.398	149	1428118	28.505	ng/ul	100
90) Benzo(b)fluoranthene	23.327	252	1354547	30.750	ng/ul	99
91) Benzo(k)fluoranthene	23.374	252	1311908	29.477	ng/ul	98
93) Benzo(a)pyrene	23.998	252	1197328	31.107	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.815	276	1656988	31.712	ng/ul	97
95) Dibenzo(a,h)anthracene	26.827	278	1377174	32.088	ng/ul	99
96) Benzo(g,h,i)perylene	27.651	276	1358452	31.602	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062623\
 Data File : BP015734.D
 Acq On : 27 Jun 2023 03:28
 Operator : MA/JU
 Sample : PB153446BS
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS446

Quant Time: Jun 27 04:17:54 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jun 27 00:56:58 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 11.319 ng/uL
RT: 3.423 min Scan# 40
Delta R.T. 0.001 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS446


Tgt Ion: 88 Resp: 39462
Ion Ratio Lower Upper
88 100
43 36.1 27.4 41.2
58 75.0 59.8 89.8



#5
Pyridine
Concen: 26.918 ng/uL
RT: 3.852 min Scan# 113
Delta R.T. 0.006 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28

Tgt Ion: 79 Resp: 229967
Ion Ratio Lower Upper
79 100
52 76.0 65.5 98.3
51 37.5 31.8 47.8









#6
Benzaldehyde
Concen: 44.452 ng/ul
RT: 7.199 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS446


Tgt Ion: 77 Resp: 179233
Ion Ratio Lower Upper
77 100
105 90.9 73.4 110.0
106 85.1 66.6 100.0



#8
Phenol
Concen: 27.859 ng/ul
RT: 7.258 min Scan# 692
Delta R.T. 0.000 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28

Tgt Ion: 94 Resp: 288315
Ion Ratio Lower Upper
94 100
65 32.0 26.4 39.6
66 45.9 37.3 55.9









#10
Bis(2-Chloroethyl)ether
Concen: 28.074 ng/ul
RT: 7.481 min Scan# 730
Delta R.T. 0.000 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS446


Tgt Ion: 93 Resp: 231164
Ion Ratio Lower Upper
93 100
63 77.3 62.8 94.2
95 31.3 25.6 38.4



#12
2-Chlorophenol
Concen: 28.941 ng/ul
RT: 7.622 min Scan# 754
Delta R.T. -0.006 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28

Tgt Ion:128 Resp: 231467
Ion Ratio Lower Upper
128 100
64 52.3 41.2 61.8
130 30.8 26.0 39.0









#13
2-Methylphenol
Concen: 27.267 ng/ul
RT: 8.516 min Scan# 907
Delta R.T. -0.006 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS446


Tgt Ion:108 Resp: 213055
Ion Ratio Lower Upper
108 100
107 90.3 73.8 110.6



#14
2,2'-oxybis(1-Chloropropane)
Concen: 27.818 ng/ul
RT: 8.599 min Scan# 920
Delta R.T. 0.006 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28

Tgt Ion: 45 Resp: 312848
Ion Ratio Lower Upper
45 100
77 15.1 12.4 18.6
79 11.4 9.3 13.9









#16
Acetophenone
Concen: 26.476 ng/ul
RT: 8.905 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS446


Tgt Ion:105 Resp: 342887
Ion Ratio Lower Upper
105 100
77 81.8 65.0 97.4
51 30.5 25.1 37.7



#17
N-Nitroso-di-n-propylamine
Concen: 25.433 ng/ul
RT: 8.881 min Scan# 968
Delta R.T. 0.000 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28

Tgt Ion: 70 Resp: 182935
Ion Ratio Lower Upper
70 100
42 60.8 48.2 72.2
101 10.2 7.6 11.4
130 18.9 13.9 20.9





#18

4-Methylphenol

Concen: 26.954 ng/ul

RT: 8.858 min Scan# 963

Delta R.T. 0.000 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28

Instrument :

BNA_P

ClientSampleId :





SLCS446

Tgt Ion:108 Resp: 226289

Ion Ratio Lower Upper

108 100

107 110.2 91.3 136.9



#19

Hexachloroethane

Concen: 29.911 ng/ul

RT: 9.146 min Scan# 1013

Delta R.T. 0.000 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28




Tgt Ion:117 Resp: 110490


Ion Ratio Lower Upper

117 100

201 104.3 84.2 126.2

199 65.5 51.8 77.8









#22
Nitrobenzene
Concen: 30.623 ng/ul
RT: 9.287 min Scan# 1037
Delta R.T. 0.000 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS446


Tgt Ion: 77 Resp: 281435
Ion Ratio Lower Upper
77 100
123 40.6 31.0 46.4
65 14.3 11.3 16.9



#23
Isophorone
Concen: 28.411 ng/ul
RT: 9.810 min Scan# 1126
Delta R.T. -0.006 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28

Tgt Ion: 82 Resp: 499668
Ion Ratio Lower Upper
82 100
95 8.2 6.3 9.5
138 16.0 12.3 18.5









#25
2-Nitrophenol
Concen: 30.090 ng/ul
RT: 10.005 min Scan# 1159
Delta R.T. 0.000 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS446


Tgt Ion:139 Resp: 121976
Ion Ratio Lower Upper
139 100
65 61.1 50.2 75.2
109 41.0 31.3 46.9



#26
2,4-Dimethylphenol
Concen: 27.023 ng/ul
RT: 10.063 min Scan# 1169
Delta R.T. 0.000 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28

Tgt Ion:107 Resp: 240809
Ion Ratio Lower Upper
107 100
121 47.0 38.2 57.2
122 78.0 62.1 93.1





#27

Bis(2-Chloroethoxy)methane

Concen: 29.092 ng/ul

RT: 10.293 min Scan# 1208

Delta R.T. 0.000 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28

Instrument :

BNA_P

ClientSampleId :

SLCS446





Tgt Ion: 93 Resp: 288829

Ion Ratio Lower Upper

93 100

95 32.7 26.6 39.8

123 10.7 8.6 13.0



#29

2,4-Dichlorophenol

Concen: 30.364 ng/ul

RT: 10.546 min Scan# 1251

Delta R.T. -0.006 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28




Tgt Ion:162 Resp: 205470


Ion Ratio Lower Upper

162 100

164 67.5 53.0 79.6

98 39.2 31.3 46.9









#30
Naphthalene
Concen: 29.949 ng/ul
RT: 10.940 min Scan# 1318
Delta R.T. 0.000 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS446


Tgt Ion:128 Resp: 697185
Ion Ratio Lower Upper
128 100
129 10.9 8.6 12.8
127 13.9 10.3 15.5



#32
4-Chloroaniline
Concen: 26.452 ng/ul
RT: 11.063 min Scan# 1339
Delta R.T. -0.006 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28

Tgt Ion:127 Resp: 240291
Ion Ratio Lower Upper
127 100
129 31.5 24.7 37.1





#33

Hexachlorobutadiene

Concen: 30.646 ng/ul

RT: 11.210 min Scan# 1364

Delta R.T. 0.000 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28

Instrument :

BNA_P

ClientSampleId :

SLCS446





Tgt Ion:225 Resp: 156607

Ion Ratio Lower Upper

225 100

223 63.7 50.4 75.6

227 66.1 51.0 76.4



#34

Caprolactam

Concen: 30.550 ng/ul

RT: 11.857 min Scan# 1474

Delta R.T. 0.000 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28




Tgt Ion:113 Resp: 63236


Ion Ratio Lower Upper

113 100

55 184.7 142.8 214.2

56 147.7 116.9 175.3









#35
4-Chloro-3-methylphenol
Concen: 28.776 ng/ul
RT: 12.193 min Scan# 1531
Delta R.T. 0.006 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS446


Tgt Ion:107 Resp: 227929
Ion Ratio Lower Upper
107 100
144 23.4 18.4 27.6
142 76.2 58.6 87.8



#36
2-Methylnaphthalene
Concen: 29.152 ng/ul
RT: 12.546 min Scan# 1591
Delta R.T. 0.000 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28

Tgt Ion:142 Resp: 446741
Ion Ratio Lower Upper
142 100
141 88.0 73.2 109.8









#37
1-Methylnaphthalene
Concen: 28.960 ng/ul
RT: 12.763 min Scan# 1628
Delta R.T. 0.000 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS446


Tgt Ion:142 Resp: 452736
Ion Ratio Lower Upper
142 100
141 94.1 73.1 109.7
116 4.4 4.0 6.0



#39
1,2,4,5-Tetrachlorobenzene
Concen: 31.033 ng/ul
RT: 12.910 min Scan# 1653
Delta R.T. 0.000 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28

Tgt Ion:216 Resp: 254134
Ion Ratio Lower Upper
216 100
214 76.5 62.4 93.6
179 22.4 17.6 26.4
108 18.8 15.5 23.3









#40
Hexachlorocyclopentadiene
Concen: 24.393 ng/ul
RT: 12.875 min Scan# 1647
Delta R.T. 0.000 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS446


Tgt Ion:237 Resp: 59760
Ion Ratio Lower Upper
237 100
235 61.2 48.6 73.0
272 12.8 10.5 15.7



#41
2,4,6-Trichlorophenol
Concen: 30.437 ng/ul
RT: 13.157 min Scan# 1695
Delta R.T. 0.000 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28

Tgt Ion:196 Resp: 157524
Ion Ratio Lower Upper
196 100
198 96.9 75.4 113.2
200 30.0 23.6 35.4





#42

2,4,5-Trichlorophenol

Concen: 30.984 ng/ul

RT: 13.240 min Scan# 1710

Delta R.T. 0.000 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28

Instrument :

BNA_P

ClientSampleId :

SLCS446





Tgt Ion:196 Resp: 170087

Ion Ratio Lower Upper

196 100

198 96.6 74.9 112.3

200 31.5 24.3 36.5



#43

1,1'-Biphenyl

Concen: 30.091 ng/ul

RT: 13.546 min Scan# 1761

Delta R.T. 0.000 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28




Tgt Ion:154 Resp: 592299


Ion Ratio Lower Upper

154 100

153 40.3 32.8 49.2

76 13.1 10.7 16.1









#44
2-Chloronaphthalene
Concen: 30.522 ng/ul
RT: 13.593 min Scan# 1769
Delta R.T. 0.000 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS446


Tgt Ion:162 Resp: 473286
Ion Ratio Lower Upper
162 100
164 32.0 26.1 39.1
127 41.5 33.4 50.0



#45
2-Nitroaniline
Concen: 31.987 ng/ul
RT: 13.810 min Scan# 1806
Delta R.T. 0.000 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28

Tgt Ion: 65 Resp: 156067
Ion Ratio Lower Upper
65 100
92 63.2 49.5 74.3
138 85.2 71.4 107.2









#47
Dimethylphthalate
Concen: 30.040 ng/ul
RT: 14.163 min Scan# 1866
Delta R.T. -0.006 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS446


Tgt Ion:163 Resp: 597145
Ion Ratio Lower Upper
163 100
194 3.8 3.1 4.7
164 9.7 7.7 11.5



#48
2,6-Dinitrotoluene
Concen: 31.252 ng/ul
RT: 14.304 min Scan# 1890
Delta R.T. 0.000 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28

Tgt Ion:165 Resp: 126016
Ion Ratio Lower Upper
165 100
89 66.9 54.1 81.1
121 22.4 18.2 27.2





#50

Acenaphthylene

Concen: 30.335 ng/ul

RT: 14.434 min Scan# 1912

Delta R.T. -0.006 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28

Instrument :

BNA_P

ClientSampleId :

SLCS446





Tgt Ion:152 Resp: 721712

Ion Ratio Lower Upper

152 100

151 20.6 16.2 24.2

153 13.1 10.5 15.7



#51

3-Nitroaniline

Concen: 37.731 ng/ul

RT: 14.640 min Scan# 1947

Delta R.T. -0.006 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28




Tgt Ion:138 Resp: 119306


Ion Ratio Lower Upper

138 100

108 13.6 9.1 13.7

92 136.9 96.8 145.2





#52

Acenaphthene

Concen: 30.473 ng/ul

RT: 14.775 min Scan# 1970

Delta R.T. 0.000 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28

Instrument :

BNA_P

ClientSampleId :

SLCS446





Tgt Ion:153 Resp: 502745

Ion Ratio Lower Upper

153 100

152 47.1 39.4 59.2

154 87.3 71.4 107.2



#53

2,4-Dinitrophenol

Concen: 29.029 ng/ul

RT: 14.869 min Scan# 1986

Delta R.T. 0.000 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28




Tgt Ion:184 Resp: 62101


Ion Ratio Lower Upper

184 100

63 80.0 71.4 107.0

154 72.9 61.1 91.7









#55
4-Nitrophenol
Concen: 33.398 ng/ul
RT: 14.963 min Scan# 2001
Delta R.T. 0.006 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS446


Tgt Ion:109 Resp: 87727
Ion Ratio Lower Upper
109 100
139 96.9 81.3 121.9
65 116.1 99.9 149.9



#56
Dibenzofuran
Concen: 30.394 ng/ul
RT: 15.110 min Scan# 2027
Delta R.T. -0.006 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28

Tgt Ion:168 Resp: 696099
Ion Ratio Lower Upper
168 100
139 40.1 32.2 48.4









#57
2,4-Dinitrotoluene
Concen: 33.468 ng/ul
RT: 15.104 min Scan# 2026
Delta R.T. 0.000 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS446


Tgt Ion:165 Resp: 183902
Ion Ratio Lower Upper
165 100
63 70.1 56.8 85.2
182 2.4 2.0 3.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 31.229 ng/ul
RT: 15.345 min Scan# 2067
Delta R.T. 0.000 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28

Tgt Ion:232 Resp: 147250
Ion Ratio Lower Upper
232 100
131 47.0 38.2 57.2
130 2.8 2.2 3.2
166 32.5 25.6 38.4





#59

Diethylphthalate

Concen: 30.569 ng/ul

RT: 15.522 min Scan# 2097

Delta R.T. 0.000 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28

Instrument :

BNA_P

ClientSampleId :

SLCS446





Tgt Ion:149 Resp: 614929

Ion Ratio Lower Upper

149 100

177 22.0 16.8 25.2

150 12.2 10.3 15.5



#61

Fluorene

Concen: 30.718 ng/ul

RT: 15.763 min Scan# 2138

Delta R.T. 0.000 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28




Tgt Ion:166 Resp: 570637


Ion Ratio Lower Upper

166 100

165 96.2 78.7 118.1

167 13.0 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 30.147 ng/ul
RT: 15.745 min Scan# 2135
Delta R.T. -0.006 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS446


Tgt Ion:204 Resp: 286562
Ion Ratio Lower Upper
204 100
206 32.6 26.0 39.0
141 65.7 51.8 77.8



#63
4-Nitroaniline
Concen: 53.221 ng/ul
RT: 15.810 min Scan# 2146
Delta R.T. -0.006 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28

Tgt Ion:138 Resp: 115179
Ion Ratio Lower Upper
138 100
92 57.5 50.1 75.1
108 100.6 85.5 128.3









#66
4,6-Dinitro-2-methylphenol
Concen: 29.942 ng/ul
RT: 15.869 min Scan# 2157
Delta R.T. -0.006 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS446


Tgt Ion:198 Resp: 101635
Ion Ratio Lower Upper
198 100
182 5.4 3.4 5.2#
77 31.3 24.4 36.6



#67
N-Nitrosodiphenylamine
Concen: 29.766 ng/ul
RT: 15.969 min Scan# 2173
Delta R.T. 0.000 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28

Tgt Ion:169 Resp: 486109
Ion Ratio Lower Upper
169 100
168 67.4 54.2 81.4
167 37.3 29.7 44.5









#68
4-Bromophenyl-phenylether
Concen: 29.283 ng/ul
RT: 16.651 min Scan# 2289
Delta R.T. 0.000 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS446


Tgt Ion:248 Resp: 182416
Ion Ratio Lower Upper
248 100
250 97.2 84.6 126.8
141 78.7 67.8 101.6



#69
Hexachlorobenzene
Concen: 30.048 ng/ul
RT: 16.775 min Scan# 2310
Delta R.T. 0.000 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28

Tgt Ion:284 Resp: 220868
Ion Ratio Lower Upper
284 100
142 33.8 27.8 41.8
249 32.0 26.1 39.1









#70
Atrazine
Concen: 22.758 ng/ul
RT: 16.922 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS446


Tgt Ion:200 Resp: 142266
Ion Ratio Lower Upper
200 100
173 24.6 19.8 29.8
215 45.8 36.8 55.2



#71
Pentachlorophenol
Concen: 31.230 ng/ul
RT: 17.128 min Scan# 2370
Delta R.T. -0.006 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28

Tgt Ion:266 Resp: 110173
Ion Ratio Lower Upper
266 100
264 64.0 53.4 80.2
268 64.7 51.4 77.2





#72

Phenanthrene

Concen: 31.375 ng/ul

RT: 17.516 min Scan# 2436

Delta R.T. 0.000 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28

Instrument :

BNA_P

ClientSampleId :

SLCS446





Tgt Ion:178 Resp: 929835

Ion Ratio Lower Upper

178 100

179 15.0 12.4 18.6

176 19.4 15.6 23.4



#74

Anthracene

Concen: 31.057 ng/ul

RT: 17.610 min Scan# 2452

Delta R.T. 0.000 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28




Tgt Ion:178 Resp: 924829


Ion Ratio Lower Upper

178 100

179 15.0 12.1 18.1

176 18.6 14.6 22.0









#75
1,2,3,4-Tetrachlorobenzene
Concen: 28.257 ng/uL
RT: 13.516 min Scan# 1756
Delta R.T. 0.000 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS446


Tgt Ion:216 Resp: 250831
Ion Ratio Lower Upper
216 100
214 76.6 62.0 93.0
218 48.8 38.4 57.6



#76
Pentachlorobenzene
Concen: 28.192 ng/uL
RT: 15.034 min Scan# 2014
Delta R.T. 0.000 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28

Tgt Ion:250 Resp: 252568
Ion Ratio Lower Upper
250 100
248 61.8 50.7 76.1
252 64.7 51.2 76.8





#77

Carbazole

Concen: 33.714 ng/ul

RT: 17.886 min Scan# 2499

Delta R.T. 0.000 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28

Instrument :

BNA_P

ClientSampleId :

SLCS446





Tgt Ion:167 Resp: 846198

Ion Ratio Lower Upper

167 100

166 21.5 17.4 26.0

139 13.3 11.0 16.4



#78

Di-n-butylphthalate

Concen: 33.106 ng/ul

RT: 18.404 min Scan# 2587

Delta R.T. 0.000 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28




Tgt Ion:149 Resp: 1101583


Ion Ratio Lower Upper

149 100

150 9.0 7.5 11.3

104 6.0 4.6 7.0









#80
Fluoranthene
Concen: 26.062 ng/ul
RT: 19.475 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28




Instrument :
BNA_P
ClientSampleId :
SLCS446


Tgt Ion:202 Resp: 1166885
Ion Ratio Lower Upper
202 100
101 10.4 8.3 12.5
100 7.9 6.6 9.8



#82
Pyrene
Concen: 26.623 ng/ul
RT: 19.828 min Scan# 2829
Delta R.T. -0.006 min
Lab File: BP015734.D
Acq: 27 Jun 2023 03:28

Tgt Ion:202 Resp: 1215939
Ion Ratio Lower Upper
202 100
101 12.0 9.3 13.9
100 10.3 7.6 11.4





#83

Butylbenzylphthalate

Concen: 29.566 ng/ul

RT: 20.675 min Scan# 2974

Delta R.T. -0.006 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28

Instrument :

BNA_P

ClientSampleId :

SLCS446





Tgt Ion:149 Resp: 550902

Ion Ratio Lower Upper

149 100

91 78.1 59.8 89.6

206 19.3 16.4 24.6



#84

3,3'-Dichlorobenzidine

Concen: 35.086 ng/ul

RT: 21.474 min Scan# 3109

Delta R.T. 0.000 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28




Tgt Ion:252 Resp: 437596


Ion Ratio Lower Upper

252 100

254 64.0 51.0 76.6

126 9.7 8.6 13.0





#85

Benzo(a)anthracene

Concen: 31.545 ng/ul

RT: 21.545 min Scan# 3121

Delta R.T. 0.000 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28

Instrument :

BNA_P

ClientSampleId :

SLCS446





Tgt Ion:228 Resp: 1224276

Ion Ratio Lower Upper

228 100

229 19.6 15.3 22.9

226 27.6 21.4 32.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 31.112 ng/ul

RT: 21.433 min Scan# 3102

Delta R.T. 0.000 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28




Tgt Ion:149 Resp: 796697


Ion Ratio Lower Upper

149 100

167 26.9 21.0 31.6

279 4.7 3.8 5.6





#87

Chrysene

Concen: 31.867 ng/ul

RT: 21.604 min Scan# 3131

Delta R.T. 0.000 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28

Instrument :

BNA_P

ClientSampleId :

SLCS446





Tgt Ion:228 Resp: 1173398

Ion Ratio Lower Upper

228 100

226 29.4 24.3 36.5

229 19.4 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 28.505 ng/ul




RT: 22.398 min Scan# 3266


Delta R.T. 0.000 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28

Tgt Ion:149 Resp: 1428118





#90

Benzo(b)fluoranthene

Concen: 30.750 ng/ul

RT: 23.327 min Scan# 3424

Delta R.T. 0.000 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28

Instrument :

BNA_P

ClientSampleId :

SLCS446





Tgt Ion:252 Resp: 1354547

Ion Ratio Lower Upper

252 100

253 21.8 17.3 25.9

125 8.6 7.5 11.3



#91

Benzo(k)fluoranthene

Concen: 29.477 ng/ul

RT: 23.374 min Scan# 3432

Delta R.T. -0.006 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28




Tgt Ion:252 Resp: 1311908


Ion Ratio Lower Upper

252 100

253 22.0 17.0 25.4

125 8.6 7.3 10.9





#93

Benzo(a)pyrene

Concen: 31.107 ng/ul

RT: 23.998 min Scan# 3538

Delta R.T. -0.006 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28

Instrument :

BNA_P

ClientSampleId :

SLCS446





Tgt Ion:252 Resp: 1197328

Ion Ratio Lower Upper

252 100

253 22.0 17.7 26.5

125 9.8 8.2 12.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 31.712 ng/ul

RT: 26.815 min Scan# 4017

Delta R.T. 0.000 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28




Tgt Ion:276 Resp: 1656988


Ion Ratio Lower Upper

276 100

138 21.2 15.7 23.5

277 25.7 19.6 29.4





#95

Dibenzo(a,h)anthracene

Concen: 32.088 ng/ul

RT: 26.827 min Scan# 4020

Delta R.T. 0.000 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28

Instrument :

BNA_P

ClientSampleId :

SLCS446





Tgt Ion:278 Resp: 1377174

Ion Ratio Lower Upper

278 100

139 13.2 10.5 15.7

279 23.6 18.4 27.6



#96

Benzo(g,h,i)perylene

Concen: 31.602 ng/ul

RT: 27.651 min Scan# 4159

Delta R.T. -0.006 min

Lab File: BP015734.D

Acq: 27 Jun 2023 03:28

Tgt Ion:276 Resp: 1358452

Ion Ratio Lower Upper

276 100

138 18.2 15.1 22.7

277 23.3 18.8 28.2

